

Alcohol Testing Program - Instrument Processing Sheet

Agency: Cape Coral Police Department Instrument Serial Number: 80-000859

Date In: 3/4/2026 DI Completion Date: 3/6/2026 ☐ Ship ☒ P/U ☐ H/D ☐ CMI ☐ EE

Intake By: <u>TDG</u> Date: <u>3/4/2026</u>		Quality Checks By: <u>TDG</u> Date: <u>3/6/2026</u>		Flow Adjustment By: _____																																																																																								
☒ Annual ☒ Dropped Off ☐ Registration ☐ Return from CMI / EE ☐ Training Instrument Visual Inspection ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/Accessories ☐ Power Cord ☐ Printer Cable ☐ Static Bag ☐ 12V DC Cable Notes: No box.		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value: <u>204</u> ☒ Flow Verification (L/s) Flow Column #: <u>ATP106</u> 32 mm 0.160 (.139-.169) 36 mm 0.171 (.156-.190) 53 mm 0.238 (.228-.278) 103 mm 0.492 (.447-.547) ☒ Barometric Pressure Check Gauge ID #: <u>33364</u> Gauge: <u>1023</u> Instrument: <u>1022</u> ☒ Stability Checks		Flow Column #: _____ ☐ 5L/min – 17mm ☐ 15L/min – 53mm ☐ 30L/min – 103mm ☐ R-Value: _____ ☐ Post Adjustment Verification (L/S) Flow Column #: _____ 32 mm (.139-.169) 36 mm (.156-.190) 53 mm (.228-.278) 103 mm (.447-.547)																																																																																								
						<table border="1" style="width:100%; border-collapse: collapse;"> <tr> <th style="width:15%;">Simulator</th> <th style="width:20%;">Serial #</th> <th style="width:20%;">Lot#/Exp</th> <th style="width:45%;">Maintenance By: _____ Date: _____</th> </tr> <tr> <td>0.050</td> <td>MP6286</td> <td>202406K 6/19/2026</td> <td rowspan="5"> ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Tank Sensor Tare ☐ Breath Tube Replacement ☐ Other: _____ </td> </tr> <tr> <td>0.080</td> <td>MP6287</td> <td>202406L 6/19/2026</td> </tr> <tr> <td>0.200</td> <td>MP6288</td> <td>202406N 6/20/2026</td> </tr> <tr> <td rowspan="2">0.080 DGS</td> <td rowspan="2">N/A</td> <td>AG429602</td> </tr> <tr> <td>10/22/2026</td> </tr> </table>		Simulator	Serial #	Lot#/Exp	Maintenance By: _____ Date: _____	0.050	MP6286	202406K 6/19/2026	☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Tank Sensor Tare ☐ Breath Tube Replacement ☐ Other: _____	0.080	MP6287	202406L 6/19/2026	0.200	MP6288	202406N 6/20/2026	0.080 DGS	N/A	AG429602	10/22/2026																																																																			
						Simulator	Serial #	Lot#/Exp	Maintenance By: _____ Date: _____																																																																																			
						0.050	MP6286	202406K 6/19/2026	☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Tank Sensor Tare ☐ Breath Tube Replacement ☐ Other: _____																																																																																			
						0.080	MP6287	202406L 6/19/2026																																																																																				
0.200	MP6288	202406N 6/20/2026																																																																																										
0.080 DGS	N/A	AG429602																																																																																										
		10/22/2026																																																																																										
<table border="1" style="width:100%; border-collapse: collapse;"> <tr> <th style="width:15%;">Simulator</th> <th style="width:20%;">Serial #</th> <th style="width:20%;">Lot #</th> <th style="width:20%;">Expiration</th> <th style="width:25%;">Gauge: <u>1023</u> Instrument: <u>1022</u></th> </tr> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> <td>Mouth Alcohol Solution Lot #: <u>2025-D</u> Exp: <u>9/25/2027</u></td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> <td>Acetone Stock Solution Lot #: <u>2024-D</u> Exp: <u>7/19/2026</u></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> <td style="background-color: #d3d3d3;"> <table border="1" style="width:100%; border-collapse: collapse;"> <tr> <th style="width:50%;">Simulator</th> <th style="width:50%;">Serial Number</th> </tr> <tr> <td>0.000</td> <td>MP6284</td> </tr> <tr> <td>0.200</td> <td>MP6285</td> </tr> <tr> <td>0.300</td> <td>MP6286</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> </tr> </table> </td> </tr> <tr> <td colspan="4"> ☐ Post Optical Bench Adjustment Stability Checks </td> <td>0.080</td> <td>MP6287</td> </tr> <tr> <td colspan="4" style="background-color: #d3d3d3;"> <table border="1" style="width:100%; border-collapse: collapse;"> <tr> <th style="width:15%;">Simulator</th> <th style="width:20%;">Serial #</th> <th style="width:20%;">Lot #</th> <th style="width:20%;">Expiration</th> <th style="width:25%;">0.200</th> </tr> <tr> <td>0.050</td> <td></td> <td></td> <td></td> <td>MP6288</td> </tr> <tr> <td>0.080</td> <td></td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> <td></td> </tr> </table> </td> <td colspan="2" style="background-color: #d3d3d3;"> <table border="1" style="width:100%; border-collapse: collapse;"> <tr> <th colspan="2" style="background-color: #d3d3d3;">Attachments</th> </tr> <tr> <td style="width:50%; vertical-align: top;"> ☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☐ Optical Bench Adjustment </td> <td style="width:50%; vertical-align: top;"> ☐ Post-Stability Checks ☐ Flow Adjustment ☐ Form 40 ☐ Other: _____ </td> </tr> </table> </td> </tr> <tr> <td colspan="2"> Gauge ID #: _____ Gauge: _____ Instrument: _____ </td> <td colspan="4"></td> </tr> <tr> <td colspan="2" rowspan="3">Notes/Suggested Service:</td> <td colspan="4" style="background-color: #d3d3d3;"> ☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use </td> </tr> <tr> <td colspan="4" style="background-color: #d3d3d3;"> Digitally signed by Shayla Platt Date: 2026.03.25 12:33:58 -04'00' Digitally signed by Shayla Platt Date: 2026.03.25 12:34:08 -04'00' </td> </tr> <tr> <td colspan="4" style="background-color: #d3d3d3;"> Shayla Platt Tech Review Shayla Platt Admin Review </td> </tr> </table>		Simulator	Serial #	Lot #	Expiration	Gauge: <u>1023</u> Instrument: <u>1022</u>	0.000		N/A	N/A	Mouth Alcohol Solution Lot #: <u>2025-D</u> Exp: <u>9/25/2027</u>	0.040				Acetone Stock Solution Lot #: <u>2024-D</u> Exp: <u>7/19/2026</u>	0.100				<table border="1" style="width:100%; border-collapse: collapse;"> <tr> <th style="width:50%;">Simulator</th> <th style="width:50%;">Serial Number</th> </tr> <tr> <td>0.000</td> <td>MP6284</td> </tr> <tr> <td>0.200</td> <td>MP6285</td> </tr> <tr> <td>0.300</td> <td>MP6286</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> </tr> </table>	Simulator	Serial Number	0.000	MP6284	0.200	MP6285	0.300	MP6286	0.080 DGS	N/A	☐ Post Optical Bench Adjustment Stability Checks				0.080	MP6287	<table border="1" style="width:100%; border-collapse: collapse;"> <tr> <th style="width:15%;">Simulator</th> <th style="width:20%;">Serial #</th> <th style="width:20%;">Lot #</th> <th style="width:20%;">Expiration</th> <th style="width:25%;">0.200</th> </tr> <tr> <td>0.050</td> <td></td> <td></td> <td></td> <td>MP6288</td> </tr> <tr> <td>0.080</td> <td></td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> <td></td> </tr> </table>				Simulator	Serial #	Lot #	Expiration	0.200	0.050				MP6288	0.080					0.200					0.080 DGS	N/A				<table border="1" style="width:100%; border-collapse: collapse;"> <tr> <th colspan="2" style="background-color: #d3d3d3;">Attachments</th> </tr> <tr> <td style="width:50%; vertical-align: top;"> ☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☐ Optical Bench Adjustment </td> <td style="width:50%; vertical-align: top;"> ☐ Post-Stability Checks ☐ Flow Adjustment ☐ Form 40 ☐ Other: _____ </td> </tr> </table>		Attachments		☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☐ Optical Bench Adjustment	☐ Post-Stability Checks ☐ Flow Adjustment ☐ Form 40 ☐ Other: _____	Gauge ID #: _____ Gauge: _____ Instrument: _____						Notes/Suggested Service:		☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use				Digitally signed by Shayla Platt Date: 2026.03.25 12:33:58 -04'00' Digitally signed by Shayla Platt Date: 2026.03.25 12:34:08 -04'00'				Shayla Platt Tech Review Shayla Platt Admin Review			
Simulator	Serial #	Lot #	Expiration	Gauge: <u>1023</u> Instrument: <u>1022</u>																																																																																								
0.000		N/A	N/A	Mouth Alcohol Solution Lot #: <u>2025-D</u> Exp: <u>9/25/2027</u>																																																																																								
0.040				Acetone Stock Solution Lot #: <u>2024-D</u> Exp: <u>7/19/2026</u>																																																																																								
0.100				<table border="1" style="width:100%; border-collapse: collapse;"> <tr> <th style="width:50%;">Simulator</th> <th style="width:50%;">Serial Number</th> </tr> <tr> <td>0.000</td> <td>MP6284</td> </tr> <tr> <td>0.200</td> <td>MP6285</td> </tr> <tr> <td>0.300</td> <td>MP6286</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> </tr> </table>	Simulator	Serial Number	0.000	MP6284	0.200	MP6285	0.300	MP6286	0.080 DGS	N/A																																																																														
Simulator	Serial Number																																																																																											
0.000	MP6284																																																																																											
0.200	MP6285																																																																																											
0.300	MP6286																																																																																											
0.080 DGS	N/A																																																																																											
☐ Post Optical Bench Adjustment Stability Checks				0.080	MP6287																																																																																							
<table border="1" style="width:100%; border-collapse: collapse;"> <tr> <th style="width:15%;">Simulator</th> <th style="width:20%;">Serial #</th> <th style="width:20%;">Lot #</th> <th style="width:20%;">Expiration</th> <th style="width:25%;">0.200</th> </tr> <tr> <td>0.050</td> <td></td> <td></td> <td></td> <td>MP6288</td> </tr> <tr> <td>0.080</td> <td></td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> <td></td> </tr> </table>				Simulator	Serial #	Lot #	Expiration	0.200	0.050				MP6288	0.080					0.200					0.080 DGS	N/A				<table border="1" style="width:100%; border-collapse: collapse;"> <tr> <th colspan="2" style="background-color: #d3d3d3;">Attachments</th> </tr> <tr> <td style="width:50%; vertical-align: top;"> ☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☐ Optical Bench Adjustment </td> <td style="width:50%; vertical-align: top;"> ☐ Post-Stability Checks ☐ Flow Adjustment ☐ Form 40 ☐ Other: _____ </td> </tr> </table>		Attachments		☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☐ Optical Bench Adjustment	☐ Post-Stability Checks ☐ Flow Adjustment ☐ Form 40 ☐ Other: _____																																																										
Simulator	Serial #	Lot #	Expiration	0.200																																																																																								
0.050				MP6288																																																																																								
0.080																																																																																												
0.200																																																																																												
0.080 DGS	N/A																																																																																											
Attachments																																																																																												
☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☐ Optical Bench Adjustment	☐ Post-Stability Checks ☐ Flow Adjustment ☐ Form 40 ☐ Other: _____																																																																																											
Gauge ID #: _____ Gauge: _____ Instrument: _____																																																																																												
Notes/Suggested Service:		☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use																																																																																										
		Digitally signed by Shayla Platt Date: 2026.03.25 12:33:58 -04'00' Digitally signed by Shayla Platt Date: 2026.03.25 12:34:08 -04'00'																																																																																										
		Shayla Platt Tech Review Shayla Platt Admin Review																																																																																										

Stability Checks

0.05g/210L		0.08g/210L		0.20g/210L		DGS 0.08g/210L																																																																																																																																																	
0.047 to 0.053		0.077 to 0.083		0.194 to 0.206		0.077 to 0.083																																																																																																																																																	
✓		✓		✓		✓																																																																																																																																																	
CAPE CORAL PD Intoxilyzer - Alcohol Analyzer Model 8000 03/06/2026 Software: 8100.27 SN 80-000859		CAPE CORAL PD Intoxilyzer - Alcohol Analyzer Model 8000 03/06/2026 Software: 8100.27 SN 80-000859		CAPE CORAL PD Intoxilyzer - Alcohol Analyzer Model 8000 03/06/2026 Software: 8100.27 SN 80-000859		CAPE CORAL PD Intoxilyzer - Alcohol Analyzer Model 8000 03/06/2026 Software: 8100.27 SN 80-000859																																																																																																																																																	
<table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>09:45</td></tr> <tr><td>Control Test</td><td>0.048</td><td>09:46</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:46</td></tr> <tr><td>Control Test</td><td>0.048</td><td>09:47</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:47</td></tr> <tr><td>Control Test</td><td>0.049</td><td>09:48</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:49</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0483</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.1945</td><td></td></tr> </tbody> </table>		Test	g/210L	Time	Air Blank	0.000	09:45	Control Test	0.048	09:46	Air Blank	0.000	09:46	Control Test	0.048	09:47	Air Blank	0.000	09:47	Control Test	0.049	09:48	Air Blank	0.000	09:49	Control Test Stats			Average	0.0483		Std Dev	0.0006		Rel Std Dev(%)	1.1945		<table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>09:52</td></tr> <tr><td>Control Test</td><td>0.079</td><td>09:53</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:53</td></tr> <tr><td>Control Test</td><td>0.079</td><td>09:54</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:54</td></tr> <tr><td>Control Test</td><td>0.079</td><td>09:55</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:56</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0790</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table>		Test	g/210L	Time	Air Blank	0.000	09:52	Control Test	0.079	09:53	Air Blank	0.000	09:53	Control Test	0.079	09:54	Air Blank	0.000	09:54	Control Test	0.079	09:55	Air Blank	0.000	09:56	Control Test Stats			Average	0.0790		Std Dev	0.0000		Rel Std Dev(%)	0.0000		<table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>09:59</td></tr> <tr><td>Control Test</td><td>0.200</td><td>10:00</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:00</td></tr> <tr><td>Control Test</td><td>0.200</td><td>10:01</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:01</td></tr> <tr><td>Control Test</td><td>0.199</td><td>10:02</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:03</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1997</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.2892</td><td></td></tr> </tbody> </table>		Test	g/210L	Time	Air Blank	0.000	09:59	Control Test	0.200	10:00	Air Blank	0.000	10:00	Control Test	0.200	10:01	Air Blank	0.000	10:01	Control Test	0.199	10:02	Air Blank	0.000	10:03	Control Test Stats			Average	0.1997		Std Dev	0.0006		Rel Std Dev(%)	0.2892		<table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>10:05</td></tr> <tr><td>Control Test</td><td>0.082</td><td>10:05</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:06</td></tr> <tr><td>Control Test</td><td>0.080</td><td>10:06</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:06</td></tr> <tr><td>Control Test</td><td>0.081</td><td>10:07</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:07</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0810</td><td></td></tr> <tr><td>Std Dev</td><td>0.0010</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.2346</td><td></td></tr> </tbody> </table>		Test	g/210L	Time	Air Blank	0.000	10:05	Control Test	0.082	10:05	Air Blank	0.000	10:06	Control Test	0.080	10:06	Air Blank	0.000	10:06	Control Test	0.081	10:07	Air Blank	0.000	10:07	Control Test Stats			Average	0.0810		Std Dev	0.0010		Rel Std Dev(%)	1.2346	
Test	g/210L	Time																																																																																																																																																					
Air Blank	0.000	09:45																																																																																																																																																					
Control Test	0.048	09:46																																																																																																																																																					
Air Blank	0.000	09:46																																																																																																																																																					
Control Test	0.048	09:47																																																																																																																																																					
Air Blank	0.000	09:47																																																																																																																																																					
Control Test	0.049	09:48																																																																																																																																																					
Air Blank	0.000	09:49																																																																																																																																																					
Control Test Stats																																																																																																																																																							
Average	0.0483																																																																																																																																																						
Std Dev	0.0006																																																																																																																																																						
Rel Std Dev(%)	1.1945																																																																																																																																																						
Test	g/210L	Time																																																																																																																																																					
Air Blank	0.000	09:52																																																																																																																																																					
Control Test	0.079	09:53																																																																																																																																																					
Air Blank	0.000	09:53																																																																																																																																																					
Control Test	0.079	09:54																																																																																																																																																					
Air Blank	0.000	09:54																																																																																																																																																					
Control Test	0.079	09:55																																																																																																																																																					
Air Blank	0.000	09:56																																																																																																																																																					
Control Test Stats																																																																																																																																																							
Average	0.0790																																																																																																																																																						
Std Dev	0.0000																																																																																																																																																						
Rel Std Dev(%)	0.0000																																																																																																																																																						
Test	g/210L	Time																																																																																																																																																					
Air Blank	0.000	09:59																																																																																																																																																					
Control Test	0.200	10:00																																																																																																																																																					
Air Blank	0.000	10:00																																																																																																																																																					
Control Test	0.200	10:01																																																																																																																																																					
Air Blank	0.000	10:01																																																																																																																																																					
Control Test	0.199	10:02																																																																																																																																																					
Air Blank	0.000	10:03																																																																																																																																																					
Control Test Stats																																																																																																																																																							
Average	0.1997																																																																																																																																																						
Std Dev	0.0006																																																																																																																																																						
Rel Std Dev(%)	0.2892																																																																																																																																																						
Test	g/210L	Time																																																																																																																																																					
Air Blank	0.000	10:05																																																																																																																																																					
Control Test	0.082	10:05																																																																																																																																																					
Air Blank	0.000	10:06																																																																																																																																																					
Control Test	0.080	10:06																																																																																																																																																					
Air Blank	0.000	10:06																																																																																																																																																					
Control Test	0.081	10:07																																																																																																																																																					
Air Blank	0.000	10:07																																																																																																																																																					
Control Test Stats																																																																																																																																																							
Average	0.0810																																																																																																																																																						
Std Dev	0.0010																																																																																																																																																						
Rel Std Dev(%)	1.2346																																																																																																																																																						
Operator's Signature <i>ML</i>		Operator's Signature <i>ML</i>		Operator's Signature <i>ML</i>		Operator's Signature <i>ML</i>																																																																																																																																																	

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: CAPE CORAL PD
Time of Inspection: 13:20

Date of Inspection: 03/06/2026

Serial Number: 80-000859
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202406K Exp: 06/19/2026	0.08g/210L Test (g/210L) Lot#:202406L Exp: 06/19/2026	0.20g/210L Test (g/210L) Lot#:202406N Exp: 06/20/2026	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG429602 Exp: 10/22/2026
0.000	0.049	0.079	0.199	0.081
0.000	0.049	0.079	0.199	0.081
0.000	0.049	0.079	0.199	0.082
0.000	0.049	0.079	0.198	0.081
0.000	0.049	0.079	0.199	0.082
0.000	0.049	0.079	0.199	0.082
0.000	0.049	0.079	0.199	0.082
0.000	0.049	0.079	0.199	0.081
0.000	0.049	0.079	0.199	0.082
0.000	0.049	0.080	0.199	0.082

Standard Deviations	0.0000	0.0003	0.0003	0.0005
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0002 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

03/06/2026
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000859, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-000859</u>	UNCERTAINTY* $\pm$
Owning Agency:	<u>CAPE CORAL PD</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>03/06/2026</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>13:20</u>	0.200 g/ 210 L 0.008
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.
This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Taylor
Gutschow
Digitally signed by Taylor
Gutschow
Date: 2026.03.06 16:13:29
-05'00'

03/06/2026

Date

TAYLOR D GUTSCHOW,
Department Inspector

FDLE/ATP Form 69 January 2026

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality