

Alcohol Testing Program - Instrument Processing Sheet

Agency: LEON COUNTY SO Instrument Serial Number: 80-000958
 Date In: 1/6/2026 DI Completion Date: 01/08/2026 Ship P/U H/D CMI EE

Intake By: <u>WKP</u> Date: <u>1/6/2026</u> <input checked="" type="checkbox"/> Annual <input checked="" type="checkbox"/> Dropped Off <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE <input type="checkbox"/> Training Instrument Visual Inspection <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/Accessories <input type="checkbox"/> Power Cord <input type="checkbox"/> Printer Cable <input type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes:	Quality Checks By: <u>WKP</u> Date: <u>01/08/2026</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value: <u>135</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column #: <u>ATP 105</u> 32 mm <u>0.160</u> (.139-.169) 36 mm <u>0.179</u> (.156-.190) 53 mm <u>0.250</u> (.228-.278) 103 mm <u>0.511</u> (.447-.547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID #: <u>28662</u> Gauge: <u>1017</u> Instrument: <u>1013</u> <input checked="" type="checkbox"/> Stability Checks	Flow Adjustment By: _____ Flow Column #: _____ <input type="checkbox"/> 5L/min – 17mm <input type="checkbox"/> 15L/min – 53mm <input type="checkbox"/> 30L/min – 103mm <input type="checkbox"/> R-Value: _____ <input type="checkbox"/> Post Adjustment Verification (L/S) Flow Column #: <u>0.250</u> 32 mm _____ (.139-.169) 36 mm _____ (.156-.190) 53 mm _____ (.228-.278) 103 mm _____ (.447-.547)
--	--	---

Simulator	Serial #	Lot#/Exp	Maintenance By:	Date:
0.050	MP5088	202406K	<input type="checkbox"/> Battery Replacement	
		06/19/2026		
0.080	MP5089	202406L	<input type="checkbox"/> Dry Gas Regulator Replacement	
		06/19/2026		
0.200	MP5090	202406N	<input type="checkbox"/> Tank Sensor Tare	
		06/20/2026		
0.080 DGS	N/A	AG510701	<input type="checkbox"/> Breath Tube Replacement	
		04/17/2027		

Optical Bench Adjustment By: _____	Department Inspection By: <u>WKP</u>
---	---

Barometric Pressure Gauge: _____ ID#: _____	Barometric Pressure ID#: <u>28427</u>
---	---------------------------------------

Simulator	Serial #	Lot #	Expiration	Gauge: <u>1016</u>	Instrument: <u>1012</u>
0.000		N/A	N/A	Mouth Alcohol Solution Lot #: <u>2025-D</u>	Exp: <u>09/25/2027</u>
0.040				Acetone Stock Solution Lot #: <u>2025-B</u>	Exp: <u>09/22/2027</u>
0.100				Simulator	Serial Number
0.200				0.000	MP6289
0.300				Interferent	MP6290
0.080 DGS	N/A			0.050	MP5088
<input type="checkbox"/> Post Optical Bench Adjustment Stability Checks				0.080	MP5089
0.200				0.200	MP5090

Simulator Serial # Lot # Expiration 0.050 0.080 0.200 0.080 DGS N/A	Attachments <input checked="" type="checkbox"/> Form 41 <input type="checkbox"/> Post-Stability Checks <input checked="" type="checkbox"/> Stability Checks <input type="checkbox"/> Flow Adjustment <input checked="" type="checkbox"/> Calibration Certificate <input type="checkbox"/> Form 40 <input type="checkbox"/> Optical Bench Adjustment <input type="checkbox"/> Other:
---	--

Gauge ID #: _____ Instrument: _____

Notes/Suggested Service:

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC
 Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use
 Conduct an Agency Inspection Before Evidentiary Use

Digitally signed by Taylor Gutschow Date: 2026.01.27 13:13:55 -05'00'	LeAndra Higginbotham	Digitally signed by LeAndra Higginbotham Date: 2026.01.28 12:31:07 -05'00'
Tech Review		Admin Review

Stability Checks

0.05g/210L 0.047 to 0.053 ✓	0.08g/210L 0.077 to 0.083 ✓	0.20g/210L 0.194 to 0.206 ✓	DGS 0.08g/210L 0.077 to 0.083 ✓ ≤0.003 of Wet ✓																																																																																																																																																
<p>LEON COUNTY SO Intoxilyzer - Alconol Analyzer Model: 8000 SN 80-000958 01/08/2026 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>10:53</td></tr> <tr><td>Control Test</td><td>0.048</td><td>10:53</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:54</td></tr> <tr><td>Control Test</td><td>0.048</td><td>10:55</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:55</td></tr> <tr><td>Control Test</td><td>0.048</td><td>10:56</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:56</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0480</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> <p style="text-align: right;">  ----- Operator's Signature </p>	Test	g/210L	Time	Air Blank	0.000	10:53	Control Test	0.048	10:53	Air Blank	0.000	10:54	Control Test	0.048	10:55	Air Blank	0.000	10:55	Control Test	0.048	10:56	Air Blank	0.000	10:56	Control Test Stats			Average	0.0480		Std Dev	0.0000		Rel Std Dev(%)	0.0000		<p>LEON COUNTY SO Intoxilyzer - Alconol Analyzer Model: 8000 SN 80-000958 01/08/2026 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>11:01</td></tr> <tr><td>Control Test</td><td>0.079</td><td>11:02</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:03</td></tr> <tr><td>Control Test</td><td>0.079</td><td>11:03</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:04</td></tr> <tr><td>Control Test</td><td>0.079</td><td>11:05</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:05</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0790</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> <p style="text-align: right;">  ----- Operator's Signature </p>	Test	g/210L	Time	Air Blank	0.000	11:01	Control Test	0.079	11:02	Air Blank	0.000	11:03	Control Test	0.079	11:03	Air Blank	0.000	11:04	Control Test	0.079	11:05	Air Blank	0.000	11:05	Control Test Stats			Average	0.0790		Std Dev	0.0000		Rel Std Dev(%)	0.0000		<p>LEON COUNTY SO Intoxilyzer - Alconol Analyzer Model: 8000 SN 80-000958 01/08/2026 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>11:08</td></tr> <tr><td>Control Test</td><td>0.202</td><td>11:08</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:09</td></tr> <tr><td>Control Test</td><td>0.201</td><td>11:10</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:10</td></tr> <tr><td>Control Test</td><td>0.201</td><td>11:11</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:12</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.2013</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.2868</td><td></td></tr> </tbody> </table> <p style="text-align: right;">  ----- Operator's Signature </p>	Test	g/210L	Time	Air Blank	0.000	11:08	Control Test	0.202	11:08	Air Blank	0.000	11:09	Control Test	0.201	11:10	Air Blank	0.000	11:10	Control Test	0.201	11:11	Air Blank	0.000	11:12	Control Test Stats			Average	0.2013		Std Dev	0.0006		Rel Std Dev(%)	0.2868		<p>LEON COUNTY SO Intoxilyzer - Alconol Analyzer Model: 8000 SN 80-000958 01/08/2026 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>11:13</td></tr> <tr><td>Control Test</td><td>0.081</td><td>11:14</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:14</td></tr> <tr><td>Control Test</td><td>0.082</td><td>11:15</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:15</td></tr> <tr><td>Control Test</td><td>0.082</td><td>11:15</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:16</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0817</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7070</td><td></td></tr> </tbody> </table> <p style="text-align: right;">  ----- Operator's Signature </p>	Test	g/210L	Time	Air Blank	0.000	11:13	Control Test	0.081	11:14	Air Blank	0.000	11:14	Control Test	0.082	11:15	Air Blank	0.000	11:15	Control Test	0.082	11:15	Air Blank	0.000	11:16	Control Test Stats			Average	0.0817		Std Dev	0.0006		Rel Std Dev(%)	0.7070	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	10:53																																																																																																																																																	
Control Test	0.048	10:53																																																																																																																																																	
Air Blank	0.000	10:54																																																																																																																																																	
Control Test	0.048	10:55																																																																																																																																																	
Air Blank	0.000	10:55																																																																																																																																																	
Control Test	0.048	10:56																																																																																																																																																	
Air Blank	0.000	10:56																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0480																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	11:01																																																																																																																																																	
Control Test	0.079	11:02																																																																																																																																																	
Air Blank	0.000	11:03																																																																																																																																																	
Control Test	0.079	11:03																																																																																																																																																	
Air Blank	0.000	11:04																																																																																																																																																	
Control Test	0.079	11:05																																																																																																																																																	
Air Blank	0.000	11:05																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0790																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	11:08																																																																																																																																																	
Control Test	0.202	11:08																																																																																																																																																	
Air Blank	0.000	11:09																																																																																																																																																	
Control Test	0.201	11:10																																																																																																																																																	
Air Blank	0.000	11:10																																																																																																																																																	
Control Test	0.201	11:11																																																																																																																																																	
Air Blank	0.000	11:12																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.2013																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.2868																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	11:13																																																																																																																																																	
Control Test	0.081	11:14																																																																																																																																																	
Air Blank	0.000	11:14																																																																																																																																																	
Control Test	0.082	11:15																																																																																																																																																	
Air Blank	0.000	11:15																																																																																																																																																	
Control Test	0.082	11:15																																																																																																																																																	
Air Blank	0.000	11:16																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0817																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7070																																																																																																																																																		

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: LEON COUNTY SO
Time of Inspection: 13:25

Date of Inspection: 01/08/2026

Serial Number: 80-000958
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202406K Exp: 06/19/2026	0.08g/210L Test (g/210L) Lot#:202406L Exp: 06/19/2026	0.20g/210L Test (g/210L) Lot#:202406N Exp: 06/20/2026	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG510701 Exp: 04/17/2027
0.000	0.048	0.079	0.202	0.082
0.000	0.048	0.079	0.202	0.082
0.000	0.048	0.079	0.202	0.082
0.000	0.048	0.079	0.203	0.082
0.000	0.048	0.079	0.203	0.082
0.000	0.048	0.079	0.203	0.082
0.000	0.048	0.079	0.203	0.082
0.000	0.049	0.079	0.203	0.082
0.000	0.048	0.079	0.203	0.082
0.000	0.048	0.079	0.203	0.082

Standard Deviations	0.0003	0.0000	0.0004	0.0000
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0001 Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Wen-Chi K Pierson
Signature and Printed Name

01/08/2026
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000958, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-000958</u>	UNCERTAINTY* \pm	
Owning Agency:	<u>LEON COUNTY SO</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>01/08/2026</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>13:25</u>	0.200 g/ 210 L	0.008
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Wen-Chi
Pierson
Digitally signed by Wen-Chi
Pierson
Date: 2026.01.26 10:55:00
-05'00'

01/08/2026

Date

WEN-CHI K PIERSON,
Department Inspector

FDLE/ATP Form 69 January 2026
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality