

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LITZKOW,CARLTON,D  
BREATH TEST OPERATOR AGENCY: Alachua County SO  
ARRESTING OFFICER: FREEMAN,MATT  
ARRESTING OFFICER AGENCY: Alachua County SO  
LAST AGENCY INSPECTION: 05/16/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26422080A1  
DRY GAS EXPIRY DATE: 11/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 03:41  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:11

SUBJECT'S LAST NAME: HEATH

DATE/TIME OF TEST: 05/20/2024 04:51:49

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:44 |
| Air Blank          | 0.000 | 04:45 |
| Control Test       | 0.082 | 04:45 |
| Air Blank          | 0.000 | 04:45 |
| Subject Sample # 1 | 0.189 | 04:46 |
| Breath Volume      | 3.695 | 04:46 |
| Air Blank          | 0.000 | 04:47 |
| Air Blank          | 0.000 | 04:49 |
| Subject Sample # 2 | 0.186 | 04:50 |
| Breath Volume      | 4.183 | 04:50 |
| Air Blank          | 0.000 | 04:50 |
| Control Test       | 0.082 | 04:51 |
| Air Blank          | 0.000 | 04:51 |
| Diagnostics Check  | OK    | 04:51 |

---

Data Download Date/Time: 06/28/2024 11:43

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TELLE,MADDISON,J  
BREATH TEST OPERATOR AGENCY: Alachua County SO  
ARRESTING OFFICER: BOIVIN,NICHOLAS  
ARRESTING OFFICER AGENCY: Alachua County SO  
LAST AGENCY INSPECTION: 05/16/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26422080A1  
DRY GAS EXPIRY DATE: 11/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 19:12  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 19:59

SUBJECT'S LAST NAME: COMPTON

DATE/TIME OF TEST: 05/22/2024 20:28:14

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 20:22 |
| Air Blank          | 0.000 | 20:22 |
| Control Test       | 0.082 | 20:22 |
| Air Blank          | 0.000 | 20:23 |
| Subject Sample # 1 | 0.173 | 20:23 |
| Breath Volume      | 3.597 | 20:23 |
| Air Blank          | 0.000 | 20:24 |
| Air Blank          | 0.000 | 20:26 |
| Subject Sample # 2 | 0.186 | 20:26 |
| Breath Volume      | 4.101 | 20:26 |
| Air Blank          | 0.000 | 20:27 |
| Control Test       | 0.081 | 20:27 |
| Air Blank          | 0.000 | 20:28 |
| Diagnostics Check  | OK    | 20:28 |

---

Data Download Date/Time: 06/28/2024 11:43

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TELLE,MADDISON,J  
BREATH TEST OPERATOR AGENCY: Alachua County SO  
ARRESTING OFFICER: GARRIDO,ANTHONY  
ARRESTING OFFICER AGENCY: Alachua PD  
LAST AGENCY INSPECTION: 05/16/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26422080A1  
DRY GAS EXPIRY DATE: 11/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 00:50  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:21

SUBJECT'S LAST NAME: MALDONADO SANCHEZ

DATE/TIME OF TEST: 05/23/2024 03:04:34

|                   |           |       |                    |
|-------------------|-----------|-------|--------------------|
| Diagnostics Check | OK        | 02:48 |                    |
| Air Blank         | 0.000     | 02:49 |                    |
| Control Test      | 0.082     | 02:49 |                    |
| Air Blank         | 0.000     | 02:50 |                    |
| Subject Sample    | # 1 0.107 | 02:51 |                    |
| Breath Volume     | 1.273     | 02:51 |                    |
| Air Blank         | 0.000     | 02:52 |                    |
| Air Blank         | 0.000     | 02:54 |                    |
| Subject Sample    | # 2       | 02:57 | No Sample Provided |
| Breath Volume     | 0.000     | 02:57 |                    |
| Air Blank         | 0.000     | 02:58 |                    |
| Air Blank         | 0.000     | 03:00 |                    |
| Subject Sample    | # 3       | 03:03 | No Sample Provided |
| Breath Volume     | 0.000     | 03:03 |                    |
| Air Blank         | 0.000     | 03:03 |                    |
| Control Test      | 0.083     | 03:03 |                    |
| Air Blank         | 0.000     | 03:04 |                    |
| Diagnostics Check | OK        | 03:04 |                    |

Data Download Date/Time: 06/28/2024 11:43

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROBINSON,DAVIS,H  
BREATH TEST OPERATOR AGENCY: U of FI PD  
ARRESTING OFFICER: ROBINSON,DAVIS  
ARRESTING OFFICER AGENCY: U of FI PD  
LAST AGENCY INSPECTION: 05/16/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26422080A1  
DRY GAS EXPIRY DATE: 11/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 03:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: MERCADO-LOPEZ

DATE/TIME OF TEST: 05/25/2024 04:15:18

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:07 |
| Air Blank          | 0.000 | 04:08 |
| Control Test       | 0.082 | 04:08 |
| Air Blank          | 0.000 | 04:09 |
| Subject Sample # 1 | 0.147 | 04:10 |
| Breath Volume      | 2.199 | 04:10 |
| Air Blank          | 0.000 | 04:11 |
| Air Blank          | 0.000 | 04:12 |
| Subject Sample # 2 | 0.142 | 04:13 |
| Breath Volume      | 2.207 | 04:13 |
| Air Blank          | 0.000 | 04:14 |
| Control Test       | 0.081 | 04:14 |
| Air Blank          | 0.000 | 04:15 |
| Diagnostics Check  | OK    | 04:15 |

---

Data Download Date/Time: 06/28/2024 11:43

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: EVANS,R,J  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: EVANS,R  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 05/16/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26422080A1  
DRY GAS EXPIRY DATE: 11/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 03:26  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: PARRA VEGA

DATE/TIME OF TEST: 05/25/2024 05:06:40

|                   |           |       |                |
|-------------------|-----------|-------|----------------|
| Diagnostics Check | OK        | 04:53 |                |
| Air Blank         | 0.000     | 04:54 |                |
| Control Test      | 0.082     | 04:54 |                |
| Air Blank         | 0.000     | 04:55 |                |
| Subject Sample    | # 1 0.058 | 04:58 | Volume Not Met |
| Breath Volume     | 0.191     | 04:58 |                |
| Air Blank         | 0.000     | 04:58 |                |
| Air Blank         | 0.000     | 05:00 |                |
| Subject Sample    | # 2 0.092 | 05:01 |                |
| Breath Volume     | 2.261     | 05:01 |                |
| Air Blank         | 0.000     | 05:02 |                |
| Air Blank         | 0.000     | 05:04 |                |
| Subject Sample    | # 3 0.099 | 05:05 |                |
| Breath Volume     | 2.832     | 05:05 |                |
| Air Blank         | 0.000     | 05:05 |                |
| Control Test      | 0.082     | 05:06 |                |
| Air Blank         | 0.000     | 05:06 |                |
| Diagnostics Check | OK        | 05:06 |                |

Data Download Date/Time: 06/28/2024 11:43

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: EVANS,R,J  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: EVANS,R  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 05/16/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26422080A1  
DRY GAS EXPIRY DATE: 11/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 19:57  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: CROUCH

DATE/TIME OF TEST: 05/25/2024 21:08:06

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 21:01 |
| Air Blank          | 0.000 | 21:01 |
| Control Test       | 0.082 | 21:01 |
| Air Blank          | 0.000 | 21:02 |
| Subject Sample # 1 | 0.211 | 21:03 |
| Breath Volume      | 3.785 | 21:03 |
| Air Blank          | 0.000 | 21:04 |
| Air Blank          | 0.000 | 21:05 |
| Subject Sample # 2 | 0.217 | 21:06 |
| Breath Volume      | 3.933 | 21:06 |
| Air Blank          | 0.000 | 21:07 |
| Control Test       | 0.081 | 21:07 |
| Air Blank          | 0.000 | 21:08 |
| Diagnostics Check  | OK    | 21:08 |

---

Data Download Date/Time: 06/28/2024 11:44

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARRIS,DILLON,J  
BREATH TEST OPERATOR AGENCY: Alachua County SO  
ARRESTING OFFICER: MINER,MONTAVIS  
ARRESTING OFFICER AGENCY: High Springs PD  
LAST AGENCY INSPECTION: 05/16/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26422080A1  
DRY GAS EXPIRY DATE: 11/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 07:02  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 08:00

SUBJECT'S LAST NAME: RINGENBERG

DATE/TIME OF TEST: 05/26/2024 08:32:59

---

|                   |       |       |                    |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK    | 08:21 |                    |
| Air Blank         | 0.000 | 08:22 |                    |
| Control Test      | 0.082 | 08:22 |                    |
| Air Blank         | 0.000 | 08:22 |                    |
| Subject Sample    | # 1   | 08:26 | No Sample Provided |
| Breath Volume     | 0.000 | 08:26 |                    |
| Air Blank         | 0.000 | 08:26 |                    |
| Air Blank         | 0.000 | 08:28 |                    |
| Subject Sample    | # 2   | 08:31 | No Sample Provided |
| Breath Volume     | 0.000 | 08:31 |                    |
| Air Blank         | 0.000 | 08:32 |                    |
| Control Test      | 0.082 | 08:32 |                    |
| Air Blank         | 0.000 | 08:32 |                    |
| Diagnostics Check | OK    | 08:32 |                    |

---

Data Download Date/Time: 06/28/2024 11:44

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HAYES,JARED,A  
BREATH TEST OPERATOR AGENCY: FFWCC (Lake City)  
ARRESTING OFFICER: GUIRATE,ANTHONY  
ARRESTING OFFICER AGENCY: FFWCC (Lake City)  
LAST AGENCY INSPECTION: 05/16/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26422080A1  
DRY GAS EXPIRY DATE: 11/05/2024  
VIOLATION CODE: BUI  
ARREST TIME: 17:41  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: PARDUE

DATE/TIME OF TEST: 05/26/2024 19:49:14

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 19:43 |
| Air Blank          | 0.000 | 19:43 |
| Control Test       | 0.081 | 19:43 |
| Air Blank          | 0.000 | 19:44 |
| Subject Sample # 1 | 0.129 | 19:44 |
| Breath Volume      | 4.238 | 19:44 |
| Air Blank          | 0.000 | 19:45 |
| Air Blank          | 0.000 | 19:47 |
| Subject Sample # 2 | 0.135 | 19:47 |
| Breath Volume      | 4.523 | 19:47 |
| Air Blank          | 0.000 | 19:48 |
| Control Test       | 0.079 | 19:48 |
| Air Blank          | 0.000 | 19:49 |
| Diagnostics Check  | OK    | 19:49 |

---

Data Download Date/Time: 06/28/2024 11:44

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: EVANS,R,J  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: EVANS,R  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 05/16/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26422080A1  
DRY GAS EXPIRY DATE: 11/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 01:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: HARVEY

DATE/TIME OF TEST: 05/27/2024 03:23:35

|                   |       |       |                    |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK    | 03:11 |                    |
| Air Blank         | 0.000 | 03:11 |                    |
| Control Test      | 0.082 | 03:12 |                    |
| Air Blank         | 0.000 | 03:12 |                    |
| Subject Sample    | # 1   | 03:15 | No Sample Provided |
| Breath Volume     | 0.000 | 03:15 |                    |
| Air Blank         | 0.000 | 03:16 |                    |
| Air Blank         | 0.000 | 03:18 |                    |
| Subject Sample    | # 2   | 0.117 | 03:18              |
| Breath Volume     | 3.667 | 03:18 |                    |
| Air Blank         | 0.000 | 03:19 |                    |
| Air Blank         | 0.000 | 03:21 |                    |
| Subject Sample    | # 3   | 0.129 | 03:22              |
| Breath Volume     | 3.781 | 03:22 |                    |
| Air Blank         | 0.000 | 03:22 |                    |
| Control Test      | 0.082 | 03:22 |                    |
| Air Blank         | 0.000 | 03:23 |                    |
| Diagnostics Check | OK    | 03:23 |                    |

Data Download Date/Time: 06/28/2024 11:44

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CRANDALL,DOUGLAS,A  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: CRANDALL,DOUGLAS  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 05/16/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26422080A1  
DRY GAS EXPIRY DATE: 11/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 18:06  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 05/27/2024 19:45:38

---

|                   |       |       |            |
|-------------------|-------|-------|------------|
| Diagnostics Check | OK    | 19:44 |            |
| Air Blank         |       | 19:45 | RFI Detect |
| Air Blank         | 0.000 | 19:45 |            |

---

Data Download Date/Time: 06/28/2024 11:44

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CRANDALL,DOUGLAS,A  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: CRANDALL,DOUGLAS  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 05/16/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26422080A1  
DRY GAS EXPIRY DATE: 11/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 18:06  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 05/27/2024 19:54:59

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 19:48 |
| Air Blank          | 0.000 | 19:48 |
| Control Test       | 0.082 | 19:49 |
| Air Blank          | 0.000 | 19:49 |
| Subject Sample # 1 | 0.291 | 19:50 |
| Breath Volume      | 1.800 | 19:50 |
| Air Blank          | 0.000 | 19:50 |
| Air Blank          | 0.000 | 19:52 |
| Subject Sample # 2 | 0.274 | 19:53 |
| Breath Volume      | 1.148 | 19:53 |
| Air Blank          | 0.000 | 19:54 |
| Control Test       | 0.080 | 19:54 |
| Air Blank          | 0.000 | 19:54 |
| Diagnostics Check  | OK    | 19:54 |

---

Data Download Date/Time: 06/28/2024 11:44

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: EVANS,R,J  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: EVANS,R  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 05/16/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26422080A1  
DRY GAS EXPIRY DATE: 11/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 12:07  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: FRANKLIN

DATE/TIME OF TEST: 05/30/2024 04:01:00

---

|                   |       |       |                    |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK    | 03:49 |                    |
| Air Blank         | 0.000 | 03:50 |                    |
| Control Test      | 0.082 | 03:50 |                    |
| Air Blank         | 0.000 | 03:50 |                    |
| Subject Sample    | # 1   | 03:54 | No Sample Provided |
| Breath Volume     | 0.000 | 03:54 |                    |
| Air Blank         | 0.000 | 03:54 |                    |
| Air Blank         | 0.000 | 03:56 |                    |
| Subject Sample    | # 2   | 03:59 | No Sample Provided |
| Breath Volume     | 0.000 | 03:59 |                    |
| Air Blank         | 0.000 | 04:00 |                    |
| Control Test      | 0.083 | 04:00 |                    |
| Air Blank         | 0.000 | 04:00 |                    |
| Diagnostics Check | OK    | 04:01 |                    |

---

Data Download Date/Time: 06/28/2024 11:44

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JACOBSON,KARSON,J  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: JACOBSON,KARSON  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 04/29/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402623828  
DRY GAS EXPIRY DATE: 12/22/2025  
VIOLATION CODE: DUI  
ARREST TIME: 02:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:09

SUBJECT'S LAST NAME: JOYNER

DATE/TIME OF TEST: 05/03/2024 02:42:54

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:36 |
| Air Blank          | 0.000 | 02:37 |
| Control Test       | 0.082 | 02:37 |
| Air Blank          | 0.000 | 02:38 |
| Subject Sample # 1 | 0.061 | 02:38 |
| Breath Volume      | 2.148 | 02:38 |
| Air Blank          | 0.000 | 02:39 |
| Air Blank          | 0.000 | 02:40 |
| Subject Sample # 2 | 0.060 | 02:41 |
| Breath Volume      | 2.859 | 02:41 |
| Air Blank          | 0.000 | 02:41 |
| Control Test       | 0.082 | 02:42 |
| Air Blank          | 0.000 | 02:42 |
| Diagnostics Check  | OK    | 02:42 |

---

Data Download Date/Time: 05/28/2024 14:48

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WARD,RYAN,D  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: WARD,RYAN  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 04/29/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402623828  
DRY GAS EXPIRY DATE: 12/22/2025  
VIOLATION CODE: DUI  
ARREST TIME: 22:02  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: NAUSLEY

DATE/TIME OF TEST: 05/18/2024 23:07:34

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 23:03 |                      |
| Air Blank         | 0.000 | 23:03 |                      |
| Control Test      | 0.081 | 23:04 |                      |
| Air Blank         | 0.000 | 23:04 |                      |
| Subject Sample    | # 1   | 23:06 | Subject Test Refused |
| Breath Volume     | 0.000 | 23:06 |                      |
| Air Blank         | 0.000 | 23:06 |                      |
| Control Test      | 0.081 | 23:06 |                      |
| Air Blank         | 0.000 | 23:07 |                      |
| Diagnostics Check | OK    | 23:07 |                      |

Data Download Date/Time: 05/28/2024 14:48

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: REGISTER,JASON,H  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: REGISTER,JASON  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 04/29/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402623828  
DRY GAS EXPIRY DATE: 12/22/2025  
VIOLATION CODE: DUI  
ARREST TIME: 20:43  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: NASH

DATE/TIME OF TEST: 05/21/2024 22:28:39

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:20 |
| Air Blank          | 0.000 | 22:21 |
| Control Test       | 0.082 | 22:21 |
| Air Blank          | 0.000 | 22:22 |
| Subject Sample # 1 | 0.241 | 22:24 |
| Breath Volume      | 1.167 | 22:24 |
| Air Blank          | 0.000 | 22:24 |
| Air Blank          | 0.000 | 22:26 |
| Subject Sample # 2 | 0.249 | 22:27 |
| Breath Volume      | 1.496 | 22:27 |
| Air Blank          | 0.000 | 22:27 |
| Control Test       | 0.079 | 22:28 |
| Air Blank          | 0.000 | 22:28 |
| Diagnostics Check  | OK    | 22:28 |

---

Data Download Date/Time: 05/28/2024 14:48

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SAMPSON,DAVID,A  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: SAMPSON,DAVID  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 04/29/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402623828  
DRY GAS EXPIRY DATE: 12/22/2025  
VIOLATION CODE: DUI  
ARREST TIME: 18:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:03

SUBJECT'S LAST NAME: KEIRSEY

DATE/TIME OF TEST: 05/24/2024 20:01:55

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 19:54 |
| Air Blank          | 0.000 | 19:54 |
| Control Test       | 0.081 | 19:55 |
| Air Blank          | 0.000 | 19:55 |
| Subject Sample # 1 | 0.103 | 19:56 |
| Breath Volume      | 1.601 | 19:56 |
| Air Blank          | 0.000 | 19:57 |
| Air Blank          | 0.000 | 19:59 |
| Subject Sample # 2 | 0.112 | 20:00 |
| Breath Volume      | 2.628 | 20:00 |
| Air Blank          | 0.000 | 20:00 |
| Control Test       | 0.083 | 20:01 |
| Air Blank          | 0.000 | 20:01 |
| Diagnostics Check  | OK    | 20:01 |

---

Data Download Date/Time: 05/28/2024 14:48

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LANGER,MATTHEW,D  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: DETWEILER,CONNOR  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 04/29/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402623828  
DRY GAS EXPIRY DATE: 12/22/2025  
VIOLATION CODE: DUI  
ARREST TIME: 03:17  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:43

SUBJECT'S LAST NAME: KREIDER

DATE/TIME OF TEST: 05/25/2024 04:27:00

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:20 |
| Air Blank          | 0.000 | 04:21 |
| Control Test       | 0.082 | 04:21 |
| Air Blank          | 0.000 | 04:21 |
| Subject Sample # 1 | 0.274 | 04:22 |
| Breath Volume      | 1.550 | 04:22 |
| Air Blank          | 0.000 | 04:23 |
| Air Blank          | 0.000 | 04:24 |
| Subject Sample # 2 | 0.260 | 04:25 |
| Breath Volume      | 2.023 | 04:25 |
| Air Blank          | 0.000 | 04:26 |
| Control Test       | 0.081 | 04:26 |
| Air Blank          | 0.000 | 04:26 |
| Diagnostics Check  | OK    | 04:27 |

---

Data Download Date/Time: 05/28/2024 14:48

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000742

INSTRUMENT REGISTERED TO: VALPARAISO PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RINALDI,JOSEPH,E  
BREATH TEST OPERATOR AGENCY: Eglin AFB PD  
ARRESTING OFFICER: RINALDI,JOSEPH  
ARRESTING OFFICER AGENCY: Eglin AFB PD  
LAST AGENCY INSPECTION: 04/15/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402477284  
DRY GAS EXPIRY DATE: 06/27/2025  
VIOLATION CODE: DUI  
ARREST TIME: 03:14  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:16

SUBJECT'S LAST NAME: ADAMES

DATE/TIME OF TEST: 05/04/2024 04:32:53

---

|                   |           |       |                |
|-------------------|-----------|-------|----------------|
| Diagnostics Check | OK        | 04:21 |                |
| Air Blank         | 0.000     | 04:21 |                |
| Control Test      | 0.081     | 04:22 |                |
| Air Blank         | 0.000     | 04:22 |                |
| Subject Sample    | # 1 0.175 | 04:25 | Volume Not Met |
| Breath Volume     | 0.050     | 04:25 |                |
| Air Blank         | 0.000     | 04:26 |                |
| Air Blank         | 0.000     | 04:28 |                |
| Subject Sample    | # 2 0.000 | 04:31 | Volume Not Met |
| Breath Volume     | 0.015     | 04:31 |                |
| Air Blank         | 0.000     | 04:31 |                |
| Control Test      | 0.079     | 04:32 |                |
| Air Blank         | 0.000     | 04:32 |                |
| Diagnostics Check | OK        | 04:32 |                |

---

Data Download Date/Time: 05/18/2024 00:30

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000742

INSTRUMENT REGISTERED TO: VALPARAISO PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RINALDI,JOSEPH,E  
BREATH TEST OPERATOR AGENCY: Eglin AFB PD  
ARRESTING OFFICER: RINALDI,JOSEPH  
ARRESTING OFFICER AGENCY: Eglin AFB PD  
LAST AGENCY INSPECTION: 04/15/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402477284  
DRY GAS EXPIRY DATE: 06/27/2025  
VIOLATION CODE: DUI  
ARREST TIME: 03:14  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:16

SUBJECT'S LAST NAME: ADAMES

DATE/TIME OF TEST: 05/04/2024 04:38:39

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 04:35 |                      |
| Air Blank         | 0.000 | 04:36 |                      |
| Control Test      | 0.078 | 04:36 |                      |
| Air Blank         | 0.000 | 04:37 |                      |
| Subject Sample    | # 1   | 04:37 | Subject Test Refused |
| Breath Volume     | 0.000 | 04:37 |                      |
| Air Blank         | 0.000 | 04:37 |                      |
| Control Test      | 0.080 | 04:38 |                      |
| Air Blank         | 0.000 | 04:38 |                      |
| Diagnostics Check | OK    | 04:38 |                      |

---

Data Download Date/Time: 05/18/2024 00:30

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000749

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FORCE,TAYLOR,B  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: FORCE,TAYLOR  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 04/23/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402690226  
DRY GAS EXPIRY DATE: 03/09/2026  
VIOLATION CODE: DUI  
ARREST TIME: 21:51  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:12

SUBJECT'S LAST NAME: POWELL

DATE/TIME OF TEST: 05/04/2024 22:44:33

---

|                   |           |       |                |
|-------------------|-----------|-------|----------------|
| Diagnostics Check | OK        | 22:32 |                |
| Air Blank         | 0.000     | 22:33 |                |
| Control Test      | 0.080     | 22:33 |                |
| Air Blank         | 0.000     | 22:34 |                |
| Subject Sample    | # 1 0.221 | 22:37 | Volume Not Met |
| Breath Volume     | 0.863     | 22:37 |                |
| Air Blank         | 0.000     | 22:37 |                |
| Air Blank         | 0.000     | 22:39 |                |
| Subject Sample    | # 2 0.210 | 22:42 | Volume Not Met |
| Breath Volume     | 1.000     | 22:42 |                |
| Air Blank         | 0.000     | 22:43 |                |
| Control Test      | 0.078     | 22:43 |                |
| Air Blank         | 0.000     | 22:44 |                |
| Diagnostics Check | OK        | 22:44 |                |

---

Data Download Date/Time: 05/06/2024 14:02

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000749

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ORELLANA, ADALBERTO, L  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: ORELLANA, ADALBERTO  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 04/23/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402690226  
DRY GAS EXPIRY DATE: 03/09/2026  
VIOLATION CODE: DUI  
ARREST TIME: 19:54  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:27

SUBJECT'S LAST NAME: BRAKOVEC

DATE/TIME OF TEST: 05/05/2024 21:14:05

|                   |       |       |                    |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK    | 20:58 |                    |
| Air Blank         | 0.000 | 20:58 |                    |
| Control Test      | 0.080 | 20:59 |                    |
| Air Blank         | 0.000 | 20:59 |                    |
| Subject Sample    | # 1   | 21:02 | No Sample Provided |
| Breath Volume     | 0.000 | 21:02 |                    |
| Air Blank         | 0.000 | 21:03 |                    |
| Air Blank         | 0.000 | 21:05 |                    |
| Subject Sample    | # 2   | 21:06 |                    |
| Breath Volume     | 1.125 | 21:06 |                    |
| Air Blank         | 0.000 | 21:07 |                    |
| Air Blank         | 0.000 | 21:09 |                    |
| Subject Sample    | # 3   | 21:12 |                    |
| Breath Volume     | 1.621 | 21:12 |                    |
| Air Blank         | 0.000 | 21:13 |                    |
| Control Test      | 0.081 | 21:13 |                    |
| Air Blank         | 0.000 | 21:13 |                    |
| Diagnostics Check | OK    | 21:14 |                    |

Data Download Date/Time: 05/06/2024 14:02

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000759

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: THOMPSON,JOSEPH,M  
BREATH TEST OPERATOR AGENCY: Washington County SO  
ARRESTING OFFICER: SMITH,SCOTT  
ARRESTING OFFICER AGENCY: Washington County SO  
LAST AGENCY INSPECTION: 04/25/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402843437  
DRY GAS EXPIRY DATE: 09/08/2026  
VIOLATION CODE: DUI  
ARREST TIME: 23:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: HOUSTON

DATE/TIME OF TEST: 05/02/2024 00:37:56

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 00:35 |                      |
| Air Blank         | 0.000 | 00:35 |                      |
| Control Test      | 0.078 | 00:35 |                      |
| Air Blank         | 0.000 | 00:36 |                      |
| Subject Sample    | # 1   | 00:36 | Subject Test Refused |
| Breath Volume     | 0.000 | 00:36 |                      |
| Air Blank         | 0.000 | 00:37 |                      |
| Control Test      | 0.078 | 00:37 |                      |
| Air Blank         | 0.000 | 00:37 |                      |
| Diagnostics Check | OK    | 00:37 |                      |

---

Data Download Date/Time: 06/05/2024 10:40

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000759

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LENDERMAN,TYLER,K  
BREATH TEST OPERATOR AGENCY: Chipley PD  
ARRESTING OFFICER: LENDERMAN,TYLER  
ARRESTING OFFICER AGENCY: Chipley PD  
LAST AGENCY INSPECTION: 04/25/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402843437  
DRY GAS EXPIRY DATE: 09/08/2026  
VIOLATION CODE: DUI  
ARREST TIME: 03:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: MORALES

DATE/TIME OF TEST: 05/04/2024 03:43:11

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:36 |
| Air Blank          | 0.000 | 03:37 |
| Control Test       | 0.079 | 03:37 |
| Air Blank          | 0.000 | 03:38 |
| Subject Sample # 1 | 0.258 | 03:38 |
| Breath Volume      | 1.531 | 03:38 |
| Air Blank          | 0.000 | 03:39 |
| Air Blank          | 0.000 | 03:40 |
| Subject Sample # 2 | 0.238 | 03:41 |
| Breath Volume      | 1.453 | 03:41 |
| Air Blank          | 0.000 | 03:42 |
| Control Test       | 0.078 | 03:42 |
| Air Blank          | 0.000 | 03:43 |
| Diagnostics Check  | OK    | 03:43 |

---

Data Download Date/Time: 06/05/2024 10:40

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000759

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FOWLER,ERIC,C  
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)  
ARRESTING OFFICER: FOWLER,ERIC  
ARRESTING OFFICER AGENCY: FFWCC (Panama City)  
LAST AGENCY INSPECTION: 04/25/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402843437  
DRY GAS EXPIRY DATE: 09/08/2026  
VIOLATION CODE: BUI  
ARREST TIME: 17:52  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:09

SUBJECT'S LAST NAME: REEVES

DATE/TIME OF TEST: 05/04/2024 18:56:08

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 18:53 |                      |
| Air Blank         | 0.000 | 18:53 |                      |
| Control Test      | 0.079 | 18:54 |                      |
| Air Blank         | 0.000 | 18:54 |                      |
| Subject Sample    | # 1   | 18:54 | Subject Test Refused |
| Breath Volume     | 0.000 | 18:54 |                      |
| Air Blank         | 0.000 | 18:55 |                      |
| Control Test      | 0.079 | 18:55 |                      |
| Air Blank         | 0.000 | 18:56 |                      |
| Diagnostics Check | OK    | 18:56 |                      |

---

Data Download Date/Time: 06/05/2024 10:40

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000759

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ALLEN,FORREST,H  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: ALLEN,FORREST  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 04/25/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402843437  
DRY GAS EXPIRY DATE: 09/08/2026  
VIOLATION CODE: DUI  
ARREST TIME: 11:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 13:47

SUBJECT'S LAST NAME: DEVITO

DATE/TIME OF TEST: 05/19/2024 14:13:37

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 14:07 |
| Air Blank          | 0.000 | 14:07 |
| Control Test       | 0.078 | 14:08 |
| Air Blank          | 0.000 | 14:08 |
| Subject Sample # 1 | 0.000 | 14:09 |
| Breath Volume      | 2.425 | 14:09 |
| Air Blank          | 0.000 | 14:09 |
| Air Blank          | 0.000 | 14:11 |
| Subject Sample # 2 | 0.000 | 14:12 |
| Breath Volume      | 2.640 | 14:12 |
| Air Blank          | 0.000 | 14:12 |
| Control Test       | 0.079 | 14:13 |
| Air Blank          | 0.000 | 14:13 |
| Diagnostics Check  | OK    | 14:13 |

---

Data Download Date/Time: 06/05/2024 10:40

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000759

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: THOMPSON,JOSEPH,M  
BREATH TEST OPERATOR AGENCY: Washington County SO  
ARRESTING OFFICER: TATE,AYDEN  
ARRESTING OFFICER AGENCY: Washington County SO  
LAST AGENCY INSPECTION: 04/25/2024

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 402843437  
DRY GAS EXPIRY DATE: 09/08/2026  
VIOLATION CODE: DUI  
ARREST TIME: 22:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: RYALS

DATE/TIME OF TEST: 05/20/2024 22:55:13

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:48 |
| Air Blank          | 0.000 | 22:48 |
| Control Test       | 0.078 | 22:49 |
| Air Blank          | 0.000 | 22:49 |
| Subject Sample # 1 | 0.000 | 22:50 |
| Breath Volume      | 2.285 | 22:50 |
| Air Blank          | 0.000 | 22:51 |
| Air Blank          | 0.000 | 22:53 |
| Subject Sample # 2 | 0.000 | 22:53 |
| Breath Volume      | 2.152 | 22:53 |
| Air Blank          | 0.000 | 22:54 |
| Control Test       | 0.077 | 22:54 |
| Air Blank          | 0.000 | 22:55 |
| Diagnostics Check  | OK    | 22:55 |

---

Data Download Date/Time: 06/05/2024 10:40

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000759

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SOLOMON,JR,  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: ELMORE,TT  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 04/25/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402843437  
DRY GAS EXPIRY DATE: 09/08/2026  
VIOLATION CODE: DUI  
ARREST TIME: 19:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:50

SUBJECT'S LAST NAME: ALLIET

DATE/TIME OF TEST: 05/28/2024 20:21:20

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 20:18 |                      |
| Air Blank         | 0.000 | 20:18 |                      |
| Control Test      | 0.078 | 20:19 |                      |
| Air Blank         | 0.000 | 20:19 |                      |
| Subject Sample    | # 1   | 20:19 | Subject Test Refused |
| Breath Volume     | 0.000 | 20:19 |                      |
| Air Blank         | 0.000 | 20:20 |                      |
| Control Test      | 0.078 | 20:20 |                      |
| Air Blank         | 0.000 | 20:21 |                      |
| Diagnostics Check | OK    | 20:21 |                      |

---

Data Download Date/Time: 06/05/2024 10:40

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TELLE,MADDISON,J  
BREATH TEST OPERATOR AGENCY: Alachua County SO  
ARRESTING OFFICER: GARRIDO,ANTHONY  
ARRESTING OFFICER AGENCY: Alachua PD  
LAST AGENCY INSPECTION: 04/19/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26422080A1  
DRY GAS EXPIRY DATE: 11/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 00:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:34

SUBJECT'S LAST NAME: DOUET

DATE/TIME OF TEST: 05/05/2024 02:10:27

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:03 |
| Air Blank          | 0.000 | 02:04 |
| Control Test       | 0.080 | 02:04 |
| Air Blank          | 0.000 | 02:04 |
| Subject Sample # 1 | 0.162 | 02:06 |
| Breath Volume      | 2.402 | 02:06 |
| Air Blank          | 0.000 | 02:06 |
| Air Blank          | 0.000 | 02:08 |
| Subject Sample # 2 | 0.151 | 02:08 |
| Breath Volume      | 2.433 | 02:08 |
| Air Blank          | 0.000 | 02:09 |
| Control Test       | 0.081 | 02:09 |
| Air Blank          | 0.000 | 02:10 |
| Diagnostics Check  | OK    | 02:10 |

---

Data Download Date/Time: 05/17/2024 09:39

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LITZKOW,CARLTON,D  
BREATH TEST OPERATOR AGENCY: Alachua County SO  
ARRESTING OFFICER: BAKER,WILLIAM  
ARRESTING OFFICER AGENCY: Alachua County SO  
LAST AGENCY INSPECTION: 04/19/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26422080A1  
DRY GAS EXPIRY DATE: 11/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 03:38  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:11

SUBJECT'S LAST NAME: FERGUSON

DATE/TIME OF TEST: 05/06/2024 04:52:06

|                   |       |       |                    |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK    | 04:40 |                    |
| Air Blank         | 0.000 | 04:41 |                    |
| Control Test      | 0.080 | 04:41 |                    |
| Air Blank         | 0.000 | 04:42 |                    |
| Subject Sample    | # 1   | 04:45 | No Sample Provided |
| Breath Volume     | 0.000 | 04:45 |                    |
| Air Blank         | 0.000 | 04:45 |                    |
| Air Blank         | 0.000 | 04:47 |                    |
| Subject Sample    | # 2   | 04:50 | No Sample Provided |
| Breath Volume     | 0.000 | 04:50 |                    |
| Air Blank         | 0.000 | 04:51 |                    |
| Control Test      | 0.081 | 04:51 |                    |
| Air Blank         | 0.000 | 04:52 |                    |
| Diagnostics Check | OK    | 04:52 |                    |

Data Download Date/Time: 05/17/2024 09:39

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SMITH,SHELAIN,M  
BREATH TEST OPERATOR AGENCY: Alachua County SO  
ARRESTING OFFICER: VALDEZ,ALEXANDER  
ARRESTING OFFICER AGENCY: Alachua County SO  
LAST AGENCY INSPECTION: 04/19/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26422080A1  
DRY GAS EXPIRY DATE: 11/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 01:38  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: RAY

DATE/TIME OF TEST: 05/08/2024 02:56:00

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:49 |
| Air Blank          | 0.000 | 02:50 |
| Control Test       | 0.080 | 02:50 |
| Air Blank          | 0.000 | 02:51 |
| Subject Sample # 1 | 0.199 | 02:51 |
| Breath Volume      | 1.347 | 02:51 |
| Air Blank          | 0.000 | 02:52 |
| Air Blank          | 0.000 | 02:54 |
| Subject Sample # 2 | 0.191 | 02:54 |
| Breath Volume      | 1.441 | 02:54 |
| Air Blank          | 0.000 | 02:55 |
| Control Test       | 0.080 | 02:55 |
| Air Blank          | 0.000 | 02:55 |
| Diagnostics Check  | OK    | 02:56 |

---

Data Download Date/Time: 05/17/2024 09:39

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RENSLEN,SCOTT,M  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: BENNETT,JONATHON  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 04/19/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26422080A1  
DRY GAS EXPIRY DATE: 11/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 11:54  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: RENSLEN

DATE/TIME OF TEST: 05/09/2024 01:17:56

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:11 |
| Air Blank          | 0.000 | 01:12 |
| Control Test       | 0.080 | 01:12 |
| Air Blank          | 0.000 | 01:12 |
| Subject Sample # 1 | 0.176 | 01:13 |
| Breath Volume      | 2.437 | 01:13 |
| Air Blank          | 0.000 | 01:13 |
| Air Blank          | 0.000 | 01:15 |
| Subject Sample # 2 | 0.183 | 01:16 |
| Breath Volume      | 2.460 | 01:16 |
| Air Blank          | 0.000 | 01:17 |
| Control Test       | 0.080 | 01:17 |
| Air Blank          | 0.000 | 01:17 |
| Diagnostics Check  | OK    | 01:17 |

---

Data Download Date/Time: 05/17/2024 09:39

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: COPPOCK,JOEL,S  
BREATH TEST OPERATOR AGENCY: Bay County SO  
ARRESTING OFFICER: TILLMAN,JOE  
ARRESTING OFFICER AGENCY: High Springs PD  
LAST AGENCY INSPECTION: 05/17/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26422080A1  
DRY GAS EXPIRY DATE: 11/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 14:43  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 14:45

SUBJECT'S LAST NAME: JENKINS

DATE/TIME OF TEST: 05/26/2024 15:25:15

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 15:18 |
| Air Blank          | 0.000 | 15:19 |
| Control Test       | 0.080 | 15:19 |
| Air Blank          | 0.000 | 15:20 |
| Subject Sample # 1 | 0.151 | 15:20 |
| Breath Volume      | 1.636 | 15:20 |
| Air Blank          | 0.000 | 15:21 |
| Air Blank          | 0.000 | 15:23 |
| Subject Sample # 2 | 0.150 | 15:23 |
| Breath Volume      | 1.511 | 15:23 |
| Air Blank          | 0.000 | 15:24 |
| Control Test       | 0.080 | 15:24 |
| Air Blank          | 0.000 | 15:25 |
| Diagnostics Check  | OK    | 15:25 |

---

Data Download Date/Time: 06/01/2024 07:21

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LAPEE,WYATT,T  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: LEE,TIMMY  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 04/23/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402623828  
DRY GAS EXPIRY DATE: 12/22/2025  
VIOLATION CODE: DUI  
ARREST TIME: 23:49  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:44

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 05/02/2024 01:10:33

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:04 |
| Air Blank          | 0.000 | 01:04 |
| Control Test       | 0.080 | 01:05 |
| Air Blank          | 0.000 | 01:05 |
| Subject Sample # 1 | 0.158 | 01:06 |
| Breath Volume      | 2.828 | 01:06 |
| Air Blank          | 0.000 | 01:06 |
| Air Blank          | 0.000 | 01:08 |
| Subject Sample # 2 | 0.162 | 01:08 |
| Breath Volume      | 2.718 | 01:08 |
| Air Blank          | 0.000 | 01:09 |
| Control Test       | 0.079 | 01:09 |
| Air Blank          | 0.000 | 01:10 |
| Diagnostics Check  | OK    | 01:10 |

---

Data Download Date/Time: 05/08/2024 10:17

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LAPEE,WYATT,T  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: SCRUGGS,JADEN  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 04/23/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402623828  
DRY GAS EXPIRY DATE: 12/22/2025  
VIOLATION CODE: DUI  
ARREST TIME: 02:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:21

SUBJECT'S LAST NAME: SUMEREL

DATE/TIME OF TEST: 05/02/2024 03:44:36

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 03:41 |                      |
| Air Blank         | 0.000 | 03:41 |                      |
| Control Test      | 0.080 | 03:42 |                      |
| Air Blank         | 0.000 | 03:42 |                      |
| Subject Sample    | # 1   | 03:43 | Subject Test Refused |
| Breath Volume     | 0.000 | 03:43 |                      |
| Air Blank         | 0.000 | 03:43 |                      |
| Control Test      | 0.079 | 03:44 |                      |
| Air Blank         | 0.000 | 03:44 |                      |
| Diagnostics Check | OK    | 03:44 |                      |

Data Download Date/Time: 05/08/2024 10:17

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ESNEUL,MICHAEL,J  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: ESNEUL,MICHAEL  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 04/23/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402623828  
DRY GAS EXPIRY DATE: 12/22/2025  
VIOLATION CODE: DUI  
ARREST TIME: 21:42  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: CALLENDER

DATE/TIME OF TEST: 05/02/2024 23:02:41

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:56 |
| Air Blank          | 0.000 | 22:56 |
| Control Test       | 0.079 | 22:56 |
| Air Blank          | 0.000 | 22:57 |
| Subject Sample # 1 | 0.239 | 22:57 |
| Breath Volume      | 3.132 | 22:57 |
| Air Blank          | 0.000 | 22:58 |
| Air Blank          | 0.000 | 23:00 |
| Subject Sample # 2 | 0.246 | 23:01 |
| Breath Volume      | 1.425 | 23:01 |
| Air Blank          | 0.000 | 23:01 |
| Control Test       | 0.078 | 23:02 |
| Air Blank          | 0.000 | 23:02 |
| Diagnostics Check  | OK    | 23:02 |

---

Data Download Date/Time: 05/08/2024 10:17

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ESNEUL,MICHAEL,J  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: ESNEUL,MICHAEL  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 04/23/2024

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 402623828  
DRY GAS EXPIRY DATE: 12/22/2025  
VIOLATION CODE: DUI  
ARREST TIME: 20:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: FEKETY

DATE/TIME OF TEST: 05/03/2024 20:59:25

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 20:51 |
| Air Blank          | 0.000 | 20:52 |
| Control Test       | 0.079 | 20:52 |
| Air Blank          | 0.000 | 20:52 |
| Subject Sample # 1 | 0.156 | 20:53 |
| Breath Volume      | 2.027 | 20:53 |
| Air Blank          | 0.000 | 20:54 |
| Air Blank          | 0.000 | 20:56 |
| Subject Sample # 2 | 0.162 | 20:57 |
| Breath Volume      | 1.593 | 20:57 |
| Air Blank          | 0.000 | 20:58 |
| Control Test       | 0.080 | 20:58 |
| Air Blank          | 0.000 | 20:59 |
| Diagnostics Check  | OK    | 20:59 |

---

Data Download Date/Time: 05/08/2024 10:17

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ATKINS,JAMES,D  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: ATKINS,JAMES  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 04/23/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402623828  
DRY GAS EXPIRY DATE: 12/22/2025  
VIOLATION CODE: DUI  
ARREST TIME: 16:06  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 16:33

SUBJECT'S LAST NAME: JARUSHI

DATE/TIME OF TEST: 05/05/2024 17:06:45

---

|                   |           |       |                |
|-------------------|-----------|-------|----------------|
| Diagnostics Check | OK        | 16:53 |                |
| Air Blank         | 0.000     | 16:53 |                |
| Control Test      | 0.080     | 16:54 |                |
| Air Blank         | 0.000     | 16:54 |                |
| Subject Sample    | # 1 0.000 | 16:56 |                |
| Breath Volume     | 1.476     | 16:56 |                |
| Air Blank         | 0.000     | 16:57 |                |
| Air Blank         | 0.000     | 16:59 |                |
| Subject Sample    | # 2 0.000 | 17:02 | Volume Not Met |
| Breath Volume     | 0.363     | 17:02 |                |
| Air Blank         | 0.000     | 17:02 |                |
| Air Blank         | 0.000     | 17:04 |                |
| Subject Sample    | # 3 0.000 | 17:05 |                |
| Breath Volume     | 1.500     | 17:05 |                |
| Air Blank         | 0.000     | 17:05 |                |
| Control Test      | 0.078     | 17:06 |                |
| Air Blank         | 0.000     | 17:06 |                |
| Diagnostics Check | OK        | 17:06 |                |

---

Data Download Date/Time: 05/08/2024 10:17

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BEHRENS,JACOB,K  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: BEHRENS,JACOB  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 04/23/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402623828  
DRY GAS EXPIRY DATE: 12/22/2025  
VIOLATION CODE: DUI  
ARREST TIME: 19:12  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 19:56

SUBJECT'S LAST NAME: HOLAHAN

DATE/TIME OF TEST: 05/05/2024 20:21:17

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 20:18 |                      |
| Air Blank         | 0.000 | 20:18 |                      |
| Control Test      | 0.080 | 20:19 |                      |
| Air Blank         | 0.000 | 20:19 |                      |
| Subject Sample    | # 1   | 20:19 | Subject Test Refused |
| Breath Volume     | 0.000 | 20:19 |                      |
| Air Blank         | 0.000 | 20:20 |                      |
| Control Test      | 0.079 | 20:20 |                      |
| Air Blank         | 0.000 | 20:21 |                      |
| Diagnostics Check | OK    | 20:21 |                      |

---

Data Download Date/Time: 05/08/2024 10:17

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NASH,DAVID,A  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: MILLER,ADAM  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 05/08/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402623828  
DRY GAS EXPIRY DATE: 12/22/2025  
VIOLATION CODE: DUI  
ARREST TIME: 03:52  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 05:40

SUBJECT'S LAST NAME: HIGGINBOTHAM

DATE/TIME OF TEST: 05/11/2024 06:08:46

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 06:04 |                      |
| Air Blank         | 0.000 | 06:05 |                      |
| Control Test      | 0.079 | 06:05 |                      |
| Air Blank         | 0.000 | 06:05 |                      |
| Subject Sample    | # 1   | 06:07 | Subject Test Refused |
| Breath Volume     | 0.000 | 06:07 |                      |
| Air Blank         | 0.000 | 06:07 |                      |
| Control Test      | 0.079 | 06:08 |                      |
| Air Blank         | 0.000 | 06:08 |                      |
| Diagnostics Check | OK    | 06:08 |                      |

---

Data Download Date/Time: 06/24/2024 10:19

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: OVERBAY, WILLIAM, M  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: OVERBAY, WILLIAM  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 05/08/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402623828  
DRY GAS EXPIRY DATE: 12/22/2025  
VIOLATION CODE: DUI  
ARREST TIME: 22:17  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:52

SUBJECT'S LAST NAME: THIBODEAUX

DATE/TIME OF TEST: 05/12/2024 23:24:55

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 23:21 |                      |
| Air Blank         | 0.000 | 23:22 |                      |
| Control Test      | 0.079 | 23:22 |                      |
| Air Blank         | 0.000 | 23:23 |                      |
| Subject Sample    | # 1   | 23:23 | Subject Test Refused |
| Breath Volume     | 0.000 | 23:23 |                      |
| Air Blank         | 0.000 | 23:24 |                      |
| Control Test      | 0.079 | 23:24 |                      |
| Air Blank         | 0.000 | 23:24 |                      |
| Diagnostics Check | OK    | 23:24 |                      |

---

Data Download Date/Time: 06/24/2024 10:19

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WALKER,LANDON,L  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: GRUNDEN,GABE  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 05/08/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402623828  
DRY GAS EXPIRY DATE: 12/22/2025  
VIOLATION CODE: DUI  
ARREST TIME: 22:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: SAMAYOA DIAZ

DATE/TIME OF TEST: 05/15/2024 23:21:08

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:14 |
| Air Blank          | 0.000 | 23:15 |
| Control Test       | 0.080 | 23:15 |
| Air Blank          | 0.000 | 23:16 |
| Subject Sample # 1 | 0.194 | 23:16 |
| Breath Volume      | 3.222 | 23:16 |
| Air Blank          | 0.000 | 23:17 |
| Air Blank          | 0.000 | 23:19 |
| Subject Sample # 2 | 0.183 | 23:19 |
| Breath Volume      | 2.738 | 23:19 |
| Air Blank          | 0.000 | 23:20 |
| Control Test       | 0.077 | 23:20 |
| Air Blank          | 0.000 | 23:21 |
| Diagnostics Check  | OK    | 23:21 |

---

Data Download Date/Time: 06/24/2024 10:20

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MCFARLAND,JEREMIAH,M  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: LARSEN,ROB  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 05/08/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402623828  
DRY GAS EXPIRY DATE: 12/22/2025  
VIOLATION CODE: DUI  
ARREST TIME: 11:46  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: DAWSON

DATE/TIME OF TEST: 05/17/2024 01:33:27

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 01:30 |                      |
| Air Blank         | 0.000 | 01:31 |                      |
| Control Test      | 0.080 | 01:31 |                      |
| Air Blank         | 0.000 | 01:31 |                      |
| Subject Sample    | # 1   | 01:32 | Subject Test Refused |
| Breath Volume     | 0.000 | 01:32 |                      |
| Air Blank         | 0.000 | 01:32 |                      |
| Control Test      | 0.079 | 01:32 |                      |
| Air Blank         | 0.000 | 01:33 |                      |
| Diagnostics Check | OK    | 01:33 |                      |

---

Data Download Date/Time: 06/24/2024 10:20

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: STINSON,LACY,R  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: JENNINGS,ANTHONY  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 05/08/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402623828  
DRY GAS EXPIRY DATE: 12/22/2025  
VIOLATION CODE: DUI  
ARREST TIME: 20:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:09

SUBJECT'S LAST NAME: MALEY

DATE/TIME OF TEST: 05/19/2024 21:21:38

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 21:15 |
| Air Blank          | 0.000 | 21:15 |
| Control Test       | 0.080 | 21:16 |
| Air Blank          | 0.000 | 21:16 |
| Subject Sample # 1 | 0.151 | 21:17 |
| Breath Volume      | 2.125 | 21:17 |
| Air Blank          | 0.000 | 21:17 |
| Air Blank          | 0.000 | 21:19 |
| Subject Sample # 2 | 0.149 | 21:20 |
| Breath Volume      | 1.601 | 21:20 |
| Air Blank          | 0.000 | 21:20 |
| Control Test       | 0.077 | 21:21 |
| Air Blank          | 0.000 | 21:21 |
| Diagnostics Check  | OK    | 21:21 |

---

Data Download Date/Time: 06/24/2024 10:20

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WALKER,LANDON,L  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: WALKER,LANDON  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 05/08/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402623828  
DRY GAS EXPIRY DATE: 12/22/2025  
VIOLATION CODE: DUI  
ARREST TIME: 00:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: CAAL CU

DATE/TIME OF TEST: 05/25/2024 01:43:14

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:36 |
| Air Blank          | 0.000 | 01:37 |
| Control Test       | 0.080 | 01:37 |
| Air Blank          | 0.000 | 01:37 |
| Subject Sample # 1 | 0.194 | 01:38 |
| Breath Volume      | 2.320 | 01:38 |
| Air Blank          | 0.000 | 01:39 |
| Air Blank          | 0.000 | 01:41 |
| Subject Sample # 2 | 0.180 | 01:41 |
| Breath Volume      | 1.132 | 01:41 |
| Air Blank          | 0.000 | 01:42 |
| Control Test       | 0.078 | 01:42 |
| Air Blank          | 0.000 | 01:43 |
| Diagnostics Check  | OK    | 01:43 |

---

Data Download Date/Time: 06/24/2024 10:20

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MCFARLAND,JEREMIAH,M  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: SCRUGGS,JADEN  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 05/08/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402623828  
DRY GAS EXPIRY DATE: 12/22/2025  
VIOLATION CODE: DUI  
ARREST TIME: 21:51  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:47

SUBJECT'S LAST NAME: PINTO WANDERLEYS

DATE/TIME OF TEST: 05/26/2024 23:14:06

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:07 |
| Air Blank          | 0.000 | 23:08 |
| Control Test       | 0.080 | 23:08 |
| Air Blank          | 0.000 | 23:09 |
| Subject Sample # 1 | 0.185 | 23:09 |
| Breath Volume      | 3.156 | 23:09 |
| Air Blank          | 0.000 | 23:10 |
| Air Blank          | 0.000 | 23:12 |
| Subject Sample # 2 | 0.186 | 23:12 |
| Breath Volume      | 3.386 | 23:12 |
| Air Blank          | 0.000 | 23:13 |
| Control Test       | 0.079 | 23:13 |
| Air Blank          | 0.000 | 23:14 |
| Diagnostics Check  | OK    | 23:14 |

---

Data Download Date/Time: 06/24/2024 10:20

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MCFARLAND,JEREMIAH,M  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: SPENCE,JUSTIN  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 05/08/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402623828  
DRY GAS EXPIRY DATE: 12/22/2025  
VIOLATION CODE: DUI  
ARREST TIME: 12:51  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 05/27/2024 01:55:38

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 01:52 |                      |
| Air Blank         | 0.000 | 01:53 |                      |
| Control Test      | 0.080 | 01:53 |                      |
| Air Blank         | 0.000 | 01:54 |                      |
| Subject Sample    | # 1   | 01:54 | Subject Test Refused |
| Breath Volume     | 0.000 | 01:54 |                      |
| Air Blank         | 0.000 | 01:54 |                      |
| Control Test      | 0.080 | 01:55 |                      |
| Air Blank         | 0.000 | 01:55 |                      |
| Diagnostics Check | OK    | 01:55 |                      |

---

Data Download Date/Time: 06/24/2024 10:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MCFARLAND,JEREMIAH,M  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: LARSEN,ROB  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 05/08/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402623828  
DRY GAS EXPIRY DATE: 12/22/2025  
VIOLATION CODE: DUI  
ARREST TIME: 01:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:26

SUBJECT'S LAST NAME: EVERETT

DATE/TIME OF TEST: 05/27/2024 02:52:59

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:46 |
| Air Blank          | 0.000 | 02:47 |
| Control Test       | 0.080 | 02:47 |
| Air Blank          | 0.000 | 02:47 |
| Subject Sample # 1 | 0.121 | 02:48 |
| Breath Volume      | 2.378 | 02:48 |
| Air Blank          | 0.000 | 02:49 |
| Air Blank          | 0.000 | 02:50 |
| Subject Sample # 2 | 0.115 | 02:51 |
| Breath Volume      | 1.937 | 02:51 |
| Air Blank          | 0.000 | 02:52 |
| Control Test       | 0.079 | 02:52 |
| Air Blank          | 0.000 | 02:52 |
| Diagnostics Check  | OK    | 02:52 |

---

Data Download Date/Time: 06/24/2024 10:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WALKER, LANDON, L  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: RUNGE, J  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 05/08/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402623828  
DRY GAS EXPIRY DATE: 12/22/2025  
VIOLATION CODE: DUI  
ARREST TIME: 04:13  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:01

SUBJECT'S LAST NAME: BARTON

DATE/TIME OF TEST: 05/27/2024 05:32:48

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 05:29 |                      |
| Air Blank         | 0.000 | 05:30 |                      |
| Control Test      | 0.079 | 05:30 |                      |
| Air Blank         | 0.000 | 05:31 |                      |
| Subject Sample    | # 1   | 05:31 | Subject Test Refused |
| Breath Volume     | 0.000 | 05:31 |                      |
| Air Blank         | 0.000 | 05:31 |                      |
| Control Test      | 0.079 | 05:32 |                      |
| Air Blank         | 0.000 | 05:32 |                      |
| Diagnostics Check | OK    | 05:32 |                      |

Data Download Date/Time: 06/24/2024 10:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: REYNA,JAVIER,O  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: REYNA,JAVIER  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 05/08/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402623828  
DRY GAS EXPIRY DATE: 12/22/2025  
VIOLATION CODE: DUI  
ARREST TIME: 11:12  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:32

SUBJECT'S LAST NAME: GRIFFIN

DATE/TIME OF TEST: 05/28/2024 00:05:05

|                   |           |       |                |
|-------------------|-----------|-------|----------------|
| Diagnostics Check | OK        | 23:53 |                |
| Air Blank         | 0.000     | 23:53 |                |
| Control Test      | 0.079     | 23:54 |                |
| Air Blank         | 0.000     | 23:54 |                |
| Subject Sample    | # 2 0.153 | 23:57 | Volume Not Met |
| Breath Volume     | 0.105     | 23:57 |                |
| Air Blank         | 0.000     | 23:58 |                |
| Air Blank         | 0.000     | 00:00 |                |
| Subject Sample    | # 1 0.136 | 00:03 | Volume Not Met |
| Breath Volume     | 0.457     | 00:03 |                |
| Air Blank         | 0.000     | 00:04 |                |
| Control Test      | 0.077     | 00:04 |                |
| Air Blank         | 0.000     | 00:04 |                |
| Diagnostics Check | OK        | 00:05 |                |

Data Download Date/Time: 06/24/2024 10:20

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BEHRENS,JACOB,K  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: BEHRENS,JACOB  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 05/08/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402623828  
DRY GAS EXPIRY DATE: 12/22/2025  
VIOLATION CODE: DUI  
ARREST TIME: 03:11  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:52

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 05/28/2024 04:19:38

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:12 |
| Air Blank          | 0.000 | 04:12 |
| Control Test       | 0.079 | 04:13 |
| Air Blank          | 0.000 | 04:13 |
| Subject Sample # 1 | 0.191 | 04:15 |
| Breath Volume      | 1.355 | 04:15 |
| Air Blank          | 0.000 | 04:15 |
| Air Blank          | 0.000 | 04:17 |
| Subject Sample # 2 | 0.194 | 04:18 |
| Breath Volume      | 1.296 | 04:18 |
| Air Blank          | 0.000 | 04:18 |
| Control Test       | 0.078 | 04:19 |
| Air Blank          | 0.000 | 04:19 |
| Diagnostics Check  | OK    | 04:19 |

---

Data Download Date/Time: 06/24/2024 10:20

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000768

INSTRUMENT REGISTERED TO: SUWANNEE CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARDEN, BRANDIE, L  
BREATH TEST OPERATOR AGENCY: Live Oak PD  
ARRESTING OFFICER: AGAPITO, S  
ARRESTING OFFICER AGENCY: Suwannee County SO  
LAST AGENCY INSPECTION: 04/28/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14323080A3  
DRY GAS EXPIRY DATE: 09/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 12:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: HOOPER

DATE/TIME OF TEST: 05/18/2024 00:47:58

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:41 |
| Air Blank          | 0.000 | 00:42 |
| Control Test       | 0.080 | 00:42 |
| Air Blank          | 0.000 | 00:43 |
| Subject Sample # 1 | 0.007 | 00:43 |
| Breath Volume      | 1.867 | 00:43 |
| Air Blank          | 0.000 | 00:44 |
| Air Blank          | 0.000 | 00:46 |
| Subject Sample # 2 | 0.000 | 00:46 |
| Breath Volume      | 2.523 | 00:46 |
| Air Blank          | 0.000 | 00:47 |
| Control Test       | 0.078 | 00:47 |
| Air Blank          | 0.000 | 00:47 |
| Diagnostics Check  | OK    | 00:47 |

---

Data Download Date/Time: 05/22/2024 11:35

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000768

INSTRUMENT REGISTERED TO: SUWANNEE CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WAGNER,DUSTIN,R  
BREATH TEST OPERATOR AGENCY: Suwannee County SO  
ARRESTING OFFICER: WAGNER,DUSTIN  
ARRESTING OFFICER AGENCY: Suwannee County SO  
LAST AGENCY INSPECTION: 04/28/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14323080A3  
DRY GAS EXPIRY DATE: 09/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 02:11  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:11

SUBJECT'S LAST NAME: MURRAY

DATE/TIME OF TEST: 05/19/2024 04:49:07

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 04:46 |                      |
| Air Blank         | 0.000 | 04:46 |                      |
| Control Test      | 0.080 | 04:47 |                      |
| Air Blank         | 0.000 | 04:47 |                      |
| Subject Sample    | # 1   | 04:47 | Subject Test Refused |
| Breath Volume     | 0.000 | 04:47 |                      |
| Air Blank         | 0.000 | 04:48 |                      |
| Control Test      | 0.080 | 04:48 |                      |
| Air Blank         | 0.000 | 04:49 |                      |
| Diagnostics Check | OK    | 04:49 |                      |

---

Data Download Date/Time: 05/22/2024 11:35

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000768

INSTRUMENT REGISTERED TO: SUWANNEE CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BUTLER,GERALD,V  
BREATH TEST OPERATOR AGENCY: Live Oak PD  
ARRESTING OFFICER: RECCA,MICHAEL  
ARRESTING OFFICER AGENCY: Suwannee County SO  
LAST AGENCY INSPECTION: 05/22/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14323080A3  
DRY GAS EXPIRY DATE: 09/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 13:14  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 13:33

SUBJECT'S LAST NAME: FEDRICK

DATE/TIME OF TEST: 05/25/2024 14:05:16

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 13:58 |
| Air Blank          | 0.000 | 13:58 |
| Control Test       | 0.081 | 13:59 |
| Air Blank          | 0.000 | 13:59 |
| Subject Sample # 1 | 0.221 | 14:00 |
| Breath Volume      | 2.066 | 14:00 |
| Air Blank          | 0.000 | 14:01 |
| Air Blank          | 0.000 | 14:03 |
| Subject Sample # 2 | 0.215 | 14:03 |
| Breath Volume      | 2.066 | 14:03 |
| Air Blank          | 0.000 | 14:04 |
| Control Test       | 0.077 | 14:04 |
| Air Blank          | 0.000 | 14:05 |
| Diagnostics Check  | OK    | 14:05 |

---

Data Download Date/Time: 06/19/2024 07:57

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000768

INSTRUMENT REGISTERED TO: SUWANNEE CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: STEVENSON,DAVID,R  
BREATH TEST OPERATOR AGENCY: Suwannee County SO  
ARRESTING OFFICER: LANDIS,MIKE  
ARRESTING OFFICER AGENCY: Suwannee County SO  
LAST AGENCY INSPECTION: 05/22/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14323080A3  
DRY GAS EXPIRY DATE: 09/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 08:38  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 08:53

SUBJECT'S LAST NAME: SHEFFIELD

DATE/TIME OF TEST: 05/29/2024 09:45:35

---

|                   |       |       |                    |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK    | 09:33 |                    |
| Air Blank         | 0.000 | 09:33 |                    |
| Control Test      | 0.080 | 09:34 |                    |
| Air Blank         | 0.000 | 09:34 |                    |
| Subject Sample    | # 1   | 09:37 | No Sample Provided |
| Breath Volume     | 0.000 | 09:37 |                    |
| Air Blank         | 0.000 | 09:38 |                    |
| Air Blank         | 0.000 | 09:40 |                    |
| Subject Sample    | # 2   | 0.135 | 09:40              |
| Breath Volume     | 1.605 | 09:40 |                    |
| Air Blank         | 0.000 | 09:41 |                    |
| Air Blank         | 0.000 | 09:43 |                    |
| Subject Sample    | # 3   | 0.132 | 09:43              |
| Breath Volume     | 1.996 | 09:43 |                    |
| Air Blank         | 0.000 | 09:44 |                    |
| Control Test      | 0.079 | 09:44 |                    |
| Air Blank         | 0.000 | 09:45 |                    |
| Diagnostics Check | OK    | 09:45 |                    |

---

Data Download Date/Time: 06/19/2024 07:57

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000770

INSTRUMENT REGISTERED TO: HAMILTON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KOSKO,KRISTOPHER,R  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: KOSKO,KRISTOPHER  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 04/28/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 29122080A1  
DRY GAS EXPIRY DATE: 12/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 23:57  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: HERNANDEZ MEZA

DATE/TIME OF TEST: 05/13/2024 00:47:36

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:40 |
| Air Blank          | 0.000 | 00:40 |
| Control Test       | 0.081 | 00:41 |
| Air Blank          | 0.000 | 00:41 |
| Subject Sample # 1 | 0.141 | 00:42 |
| Breath Volume      | 1.488 | 00:42 |
| Air Blank          | 0.000 | 00:43 |
| Air Blank          | 0.000 | 00:45 |
| Subject Sample # 2 | 0.134 | 00:45 |
| Breath Volume      | 1.460 | 00:45 |
| Air Blank          | 0.000 | 00:46 |
| Control Test       | 0.078 | 00:46 |
| Air Blank          | 0.000 | 00:47 |
| Diagnostics Check  | OK    | 00:47 |

---

Data Download Date/Time: 05/31/2024 01:29

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BICKNELL,BRANDON,M  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: FOSTER,COLE  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 05/05/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402558895  
DRY GAS EXPIRY DATE: 10/03/2025  
VIOLATION CODE: DUI  
ARREST TIME: 18:50  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 19:33

SUBJECT'S LAST NAME: BAYNARD

DATE/TIME OF TEST: 05/05/2024 20:00:39

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 19:53 |
| Air Blank          | 0.000 | 19:54 |
| Control Test       | 0.080 | 19:54 |
| Air Blank          | 0.000 | 19:55 |
| Subject Sample # 1 | 0.235 | 19:56 |
| Breath Volume      | 1.472 | 19:56 |
| Air Blank          | 0.000 | 19:56 |
| Air Blank          | 0.000 | 19:58 |
| Subject Sample # 2 | 0.226 | 19:59 |
| Breath Volume      | 1.816 | 19:59 |
| Air Blank          | 0.000 | 19:59 |
| Control Test       | 0.080 | 20:00 |
| Air Blank          | 0.000 | 20:00 |
| Diagnostics Check  | OK    | 20:00 |

---

Data Download Date/Time: 06/07/2024 02:50

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BLACKWOOD,PATRICK,A  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: BLACKWOOD,PATRICK  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 05/05/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402558895  
DRY GAS EXPIRY DATE: 10/03/2025  
VIOLATION CODE: DUI  
ARREST TIME: 23:51  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:51

SUBJECT'S LAST NAME: BYNUM 3

DATE/TIME OF TEST: 05/07/2024 02:30:16

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 02:27 |                      |
| Air Blank         | 0.000 | 02:27 |                      |
| Control Test      | 0.080 | 02:28 |                      |
| Air Blank         | 0.000 | 02:28 |                      |
| Subject Sample    | # 1   | 02:28 | Subject Test Refused |
| Breath Volume     | 0.000 | 02:28 |                      |
| Air Blank         | 0.000 | 02:29 |                      |
| Control Test      | 0.082 | 02:29 |                      |
| Air Blank         | 0.000 | 02:30 |                      |
| Diagnostics Check | OK    | 02:30 |                      |

---

Data Download Date/Time: 06/07/2024 02:51

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CUPPLES,ZACHARY,L  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: CUPPLES,ZACHARY  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 05/05/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402558895  
DRY GAS EXPIRY DATE: 10/03/2025  
VIOLATION CODE: DUI  
ARREST TIME: 12:30  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 13:20

SUBJECT'S LAST NAME: CLEVELAND

DATE/TIME OF TEST: 05/08/2024 13:47:26

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 13:40 |
| Air Blank          | 0.000 | 13:41 |
| Control Test       | 0.081 | 13:41 |
| Air Blank          | 0.000 | 13:42 |
| Subject Sample # 1 | 0.000 | 13:42 |
| Breath Volume      | 1.773 | 13:42 |
| Air Blank          | 0.000 | 13:42 |
| Air Blank          | 0.000 | 13:44 |
| Subject Sample # 2 | 0.000 | 13:46 |
| Breath Volume      | 1.914 | 13:46 |
| Air Blank          | 0.000 | 13:46 |
| Control Test       | 0.081 | 13:46 |
| Air Blank          | 0.000 | 13:47 |
| Diagnostics Check  | OK    | 13:47 |

---

Data Download Date/Time: 06/07/2024 02:51

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MOYE,JAMES,M  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: CATES,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 05/05/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402558895  
DRY GAS EXPIRY DATE: 10/03/2025  
VIOLATION CODE: DUI  
ARREST TIME: 22:20  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:49

SUBJECT'S LAST NAME: HEDMAN

DATE/TIME OF TEST: 05/08/2024 23:16:32

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:09 |
| Air Blank          | 0.000 | 23:09 |
| Control Test       | 0.080 | 23:10 |
| Air Blank          | 0.000 | 23:10 |
| Subject Sample # 1 | 0.054 | 23:12 |
| Breath Volume      | 1.882 | 23:12 |
| Air Blank          | 0.000 | 23:12 |
| Air Blank          | 0.000 | 23:14 |
| Subject Sample # 2 | 0.052 | 23:15 |
| Breath Volume      | 1.933 | 23:15 |
| Air Blank          | 0.000 | 23:15 |
| Control Test       | 0.079 | 23:15 |
| Air Blank          | 0.000 | 23:16 |
| Diagnostics Check  | OK    | 23:16 |

---

Data Download Date/Time: 06/07/2024 02:51

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: POWELL,JESSIE,E  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: POWELL,JESSIE  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 05/05/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402558895  
DRY GAS EXPIRY DATE: 10/03/2025  
VIOLATION CODE: DUI  
ARREST TIME: 14:11  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 14:43

SUBJECT'S LAST NAME: EDWARDS

DATE/TIME OF TEST: 05/09/2024 15:11:06

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 15:04 |
| Air Blank          | 0.000 | 15:05 |
| Control Test       | 0.081 | 15:05 |
| Air Blank          | 0.000 | 15:05 |
| Subject Sample # 1 | 0.007 | 15:06 |
| Breath Volume      | 2.910 | 15:06 |
| Air Blank          | 0.000 | 15:07 |
| Air Blank          | 0.000 | 15:09 |
| Subject Sample # 2 | 0.000 | 15:09 |
| Breath Volume      | 3.941 | 15:09 |
| Air Blank          | 0.000 | 15:10 |
| Control Test       | 0.079 | 15:10 |
| Air Blank          | 0.000 | 15:11 |
| Diagnostics Check  | OK    | 15:11 |

---

Data Download Date/Time: 06/07/2024 02:51

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MACPHERSON,ZACHARY,R  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: MACPHERSON,ZACHARY  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 05/05/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402558895  
DRY GAS EXPIRY DATE: 10/03/2025  
VIOLATION CODE: DUI  
ARREST TIME: 20:33  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: KING

DATE/TIME OF TEST: 05/09/2024 21:44:30

|                    |       |       |                 |
|--------------------|-------|-------|-----------------|
| Diagnostics Check  | OK    | 21:29 |                 |
| Air Blank          | 0.000 | 21:30 |                 |
| Control Test       | 0.081 | 21:30 |                 |
| Air Blank          | 0.000 | 21:31 |                 |
| Subject Sample # 1 | 0.093 | 21:34 |                 |
| Breath Volume      | 1.320 | 21:34 |                 |
| Air Blank          | 0.000 | 21:34 |                 |
| Air Blank          | 0.000 | 21:36 |                 |
| Subject Sample # 2 | 0.084 | 21:39 | Slope Not Level |
| Breath Volume      | 1.277 | 21:39 |                 |
| Air Blank          | 0.000 | 21:40 |                 |
| Air Blank          | 0.000 | 21:42 |                 |
| Subject Sample # 3 | 0.084 | 21:42 |                 |
| Breath Volume      | 1.203 | 21:42 |                 |
| Air Blank          | 0.000 | 21:43 |                 |
| Control Test       | 0.080 | 21:43 |                 |
| Air Blank          | 0.000 | 21:44 |                 |
| Diagnostics Check  | OK    | 21:44 |                 |

Data Download Date/Time: 06/07/2024 02:51

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MACPHERSON,ZACHARY,R  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: MACPHERSON,ZACHARY  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 05/05/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402558895  
DRY GAS EXPIRY DATE: 10/03/2025  
VIOLATION CODE: DUI  
ARREST TIME: 20:33  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: KING

DATE/TIME OF TEST: 05/09/2024 21:55:02

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 21:47 |
| Air Blank          | 0.000 | 21:47 |
| Control Test       | 0.081 | 21:47 |
| Air Blank          | 0.000 | 21:48 |
| Subject Sample # 1 | 0.084 | 21:50 |
| Breath Volume      | 1.789 | 21:50 |
| Air Blank          | 0.000 | 21:51 |
| Air Blank          | 0.000 | 21:52 |
| Subject Sample # 2 | 0.091 | 21:53 |
| Breath Volume      | 1.531 | 21:53 |
| Air Blank          | 0.000 | 21:54 |
| Control Test       | 0.079 | 21:54 |
| Air Blank          | 0.000 | 21:54 |
| Diagnostics Check  | OK    | 21:55 |

---

Data Download Date/Time: 06/07/2024 02:51

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MACPHERSON,ZACHARY,R  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: WATERS,CALEB  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 05/05/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402558895  
DRY GAS EXPIRY DATE: 10/03/2025  
VIOLATION CODE: DUI  
ARREST TIME: 03:19  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:57

SUBJECT'S LAST NAME: MOYSTNER

DATE/TIME OF TEST: 05/10/2024 04:20:48

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 04:17 |                      |
| Air Blank         | 0.000 | 04:18 |                      |
| Control Test      | 0.080 | 04:18 |                      |
| Air Blank         | 0.000 | 04:19 |                      |
| Subject Sample    | # 1   | 04:19 | Subject Test Refused |
| Breath Volume     | 0.000 | 04:19 |                      |
| Air Blank         | 0.000 | 04:19 |                      |
| Control Test      | 0.081 | 04:20 |                      |
| Air Blank         | 0.000 | 04:20 |                      |
| Diagnostics Check | OK    | 04:20 |                      |

---

Data Download Date/Time: 06/07/2024 02:51

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SEIB,ALEXANDER,A  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: SEIB,ALEXANDER  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 05/05/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402558895  
DRY GAS EXPIRY DATE: 10/03/2025  
VIOLATION CODE: DUI  
ARREST TIME: 18:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:30

SUBJECT'S LAST NAME: BECK

DATE/TIME OF TEST: 05/10/2024 19:56:40

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 19:50 |
| Air Blank          | 0.000 | 19:50 |
| Control Test       | 0.080 | 19:51 |
| Air Blank          | 0.000 | 19:51 |
| Subject Sample # 1 | 0.029 | 19:52 |
| Breath Volume      | 3.253 | 19:52 |
| Air Blank          | 0.000 | 19:52 |
| Air Blank          | 0.000 | 19:54 |
| Subject Sample # 2 | 0.027 | 19:55 |
| Breath Volume      | 2.957 | 19:55 |
| Air Blank          | 0.000 | 19:55 |
| Control Test       | 0.079 | 19:56 |
| Air Blank          | 0.000 | 19:56 |
| Diagnostics Check  | OK    | 19:56 |

---

Data Download Date/Time: 06/07/2024 02:51

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LOPEZ,BRIAN,T  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: RYAN,JUSTIN  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 05/05/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402558895  
DRY GAS EXPIRY DATE: 10/03/2025  
VIOLATION CODE: DUI  
ARREST TIME: 03:07  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 05/11/2024 04:18:36

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:10 |
| Air Blank          | 0.000 | 04:11 |
| Control Test       | 0.079 | 04:11 |
| Air Blank          | 0.000 | 04:11 |
| Subject Sample # 1 | 0.191 | 04:13 |
| Breath Volume      | 2.035 | 04:13 |
| Air Blank          | 0.000 | 04:14 |
| Air Blank          | 0.000 | 04:16 |
| Subject Sample # 2 | 0.187 | 04:17 |
| Breath Volume      | 1.175 | 04:17 |
| Air Blank          | 0.000 | 04:17 |
| Control Test       | 0.078 | 04:18 |
| Air Blank          | 0.000 | 04:18 |
| Diagnostics Check  | OK    | 04:18 |

---

Data Download Date/Time: 06/07/2024 02:51

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOWARD,NOAH,M  
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD  
ARRESTING OFFICER: HOWARD,NOAH  
ARRESTING OFFICER AGENCY: Gulf Breeze PD  
LAST AGENCY INSPECTION: 05/05/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402558895  
DRY GAS EXPIRY DATE: 10/02/2025  
VIOLATION CODE: DUI  
ARREST TIME: 02:54  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:03

SUBJECT'S LAST NAME: AKIN

DATE/TIME OF TEST: 05/11/2024 04:30:38

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:24 |
| Air Blank          | 0.000 | 04:24 |
| Control Test       | 0.079 | 04:25 |
| Air Blank          | 0.000 | 04:25 |
| Subject Sample # 1 | 0.147 | 04:26 |
| Breath Volume      | 3.140 | 04:26 |
| Air Blank          | 0.000 | 04:26 |
| Air Blank          | 0.000 | 04:28 |
| Subject Sample # 2 | 0.141 | 04:29 |
| Breath Volume      | 4.000 | 04:29 |
| Air Blank          | 0.000 | 04:29 |
| Control Test       | 0.079 | 04:30 |
| Air Blank          | 0.000 | 04:30 |
| Diagnostics Check  | OK    | 04:30 |

---

Data Download Date/Time: 06/07/2024 02:51

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CAREY,ANDREW,S  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: DENNIS,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 05/05/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402558895  
DRY GAS EXPIRY DATE: 10/03/2025  
VIOLATION CODE: DUI  
ARREST TIME: 00:18  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:38

SUBJECT'S LAST NAME: CHAVEZ

DATE/TIME OF TEST: 05/12/2024 01:06:50

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:00 |
| Air Blank          | 0.000 | 01:00 |
| Control Test       | 0.081 | 01:01 |
| Air Blank          | 0.000 | 01:01 |
| Subject Sample # 1 | 0.000 | 01:02 |
| Breath Volume      | 3.257 | 01:02 |
| Air Blank          | 0.000 | 01:03 |
| Air Blank          | 0.000 | 01:05 |
| Subject Sample # 2 | 0.000 | 01:05 |
| Breath Volume      | 2.671 | 01:05 |
| Air Blank          | 0.000 | 01:05 |
| Control Test       | 0.082 | 01:06 |
| Air Blank          | 0.000 | 01:06 |
| Diagnostics Check  | OK    | 01:06 |

---

Data Download Date/Time: 06/07/2024 02:51

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MUISE,CHRISTOPHER,D  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: MUISE,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 05/05/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402558895  
DRY GAS EXPIRY DATE: 10/02/2025  
VIOLATION CODE: DUI  
ARREST TIME: 01:14  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 05/12/2024 02:08:02

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:00 |
| Air Blank          | 0.000 | 02:00 |
| Control Test       | 0.080 | 02:01 |
| Air Blank          | 0.000 | 02:01 |
| Subject Sample # 1 | 0.148 | 02:03 |
| Breath Volume      | 2.164 | 02:03 |
| Air Blank          | 0.000 | 02:04 |
| Air Blank          | 0.000 | 02:06 |
| Subject Sample # 2 | 0.149 | 02:06 |
| Breath Volume      | 1.292 | 02:06 |
| Air Blank          | 0.000 | 02:07 |
| Control Test       | 0.079 | 02:07 |
| Air Blank          | 0.000 | 02:07 |
| Diagnostics Check  | OK    | 02:08 |

---

Data Download Date/Time: 06/07/2024 02:51

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAVIS,KAYLAN,A  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: DAVIS,KAYLAN  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 05/05/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402558895  
DRY GAS EXPIRY DATE: 10/03/2025  
VIOLATION CODE: DUI  
ARREST TIME: 04:33  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:51

SUBJECT'S LAST NAME: GOMEZ

DATE/TIME OF TEST: 05/12/2024 06:17:41

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 06:11 |
| Air Blank          | 0.000 | 06:12 |
| Control Test       | 0.081 | 06:12 |
| Air Blank          | 0.000 | 06:12 |
| Subject Sample # 1 | 0.136 | 06:13 |
| Breath Volume      | 1.929 | 06:13 |
| Air Blank          | 0.000 | 06:13 |
| Air Blank          | 0.000 | 06:15 |
| Subject Sample # 2 | 0.130 | 06:16 |
| Breath Volume      | 1.714 | 06:16 |
| Air Blank          | 0.000 | 06:16 |
| Control Test       | 0.080 | 06:17 |
| Air Blank          | 0.000 | 06:17 |
| Diagnostics Check  | OK    | 06:17 |

---

Data Download Date/Time: 06/07/2024 02:51

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CAREY,ANDREW,S  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: CORNMAN,KEVIN  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 05/05/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402558895  
DRY GAS EXPIRY DATE: 10/03/2025  
VIOLATION CODE: DUI  
ARREST TIME: 22:11  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:29

SUBJECT'S LAST NAME: PHIPPS

DATE/TIME OF TEST: 05/12/2024 22:56:27

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:49 |
| Air Blank          | 0.000 | 22:49 |
| Control Test       | 0.081 | 22:49 |
| Air Blank          | 0.000 | 22:50 |
| Subject Sample # 1 | 0.219 | 22:52 |
| Breath Volume      | 2.121 | 22:52 |
| Air Blank          | 0.000 | 22:52 |
| Air Blank          | 0.000 | 22:54 |
| Subject Sample # 2 | 0.238 | 22:54 |
| Breath Volume      | 1.726 | 22:54 |
| Air Blank          | 0.000 | 22:55 |
| Control Test       | 0.080 | 22:55 |
| Air Blank          | 0.000 | 22:56 |
| Diagnostics Check  | OK    | 22:56 |

---

Data Download Date/Time: 06/07/2024 02:51

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ZOLLER,J,G  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: ZOLLER,J  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 05/05/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402558895  
DRY GAS EXPIRY DATE: 10/03/2025  
VIOLATION CODE: DUI  
ARREST TIME: 05:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:15

SUBJECT'S LAST NAME: SWITZER

DATE/TIME OF TEST: 05/17/2024 06:41:59

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 06:35 |
| Air Blank          | 0.000 | 06:36 |
| Control Test       | 0.081 | 06:36 |
| Air Blank          | 0.000 | 06:36 |
| Subject Sample # 1 | 0.000 | 06:37 |
| Breath Volume      | 2.128 | 06:37 |
| Air Blank          | 0.000 | 06:38 |
| Air Blank          | 0.000 | 06:40 |
| Subject Sample # 2 | 0.000 | 06:40 |
| Breath Volume      | 1.687 | 06:40 |
| Air Blank          | 0.000 | 06:41 |
| Control Test       | 0.080 | 06:41 |
| Air Blank          | 0.000 | 06:41 |
| Diagnostics Check  | OK    | 06:41 |

---

Data Download Date/Time: 06/07/2024 02:51

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MACPHERSON,ZACHARY,R  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: MACPHERSON,ZACHARY  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 05/05/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402558895  
DRY GAS EXPIRY DATE: 10/03/2025  
VIOLATION CODE: DUI  
ARREST TIME: 19:43  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: PAGE

DATE/TIME OF TEST: 05/19/2024 20:33:55

---

|                   |           |       |                  |
|-------------------|-----------|-------|------------------|
| Diagnostics Check | OK        | 20:28 |                  |
| Air Blank         | 0.000     | 20:29 |                  |
| Control Test      | 0.081     | 20:29 |                  |
| Air Blank         | 0.000     | 20:30 |                  |
| Subject Sample    | # 1 0.144 | 20:33 | Volume Not Met   |
| Breath Volume     | 0.535     | 20:33 |                  |
| Air Blank         |           | 20:33 | Sequence Aborted |
| Air Blank         | 0.000     | 20:33 |                  |

---

Data Download Date/Time: 06/07/2024 02:51

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MACPHERSON,ZACHARY,R  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: MACPHERSON,ZACHARY  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 05/05/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402558895  
DRY GAS EXPIRY DATE: 10/03/2025  
VIOLATION CODE: DUI  
ARREST TIME: 19:43  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: PAGE

DATE/TIME OF TEST: 05/19/2024 20:45:10

---

|                   |           |       |                      |
|-------------------|-----------|-------|----------------------|
| Diagnostics Check | OK        | 20:34 |                      |
| Air Blank         | 0.000     | 20:35 |                      |
| Control Test      | 0.080     | 20:35 |                      |
| Air Blank         | 0.000     | 20:36 |                      |
| Subject Sample    | # 1 0.173 | 20:38 |                      |
| Breath Volume     | 1.636     | 20:38 |                      |
| Air Blank         | 0.000     | 20:39 |                      |
| Air Blank         | 0.000     | 20:40 |                      |
| Subject Sample    | # 2       | 20:43 | Subject Test Refused |
| Breath Volume     | 0.566     | 20:43 |                      |
| Air Blank         | 0.000     | 20:44 |                      |
| Control Test      | 0.078     | 20:44 |                      |
| Air Blank         | 0.000     | 20:45 |                      |
| Diagnostics Check | OK        | 20:45 |                      |

---

Data Download Date/Time: 06/07/2024 02:51

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MACPHERSON,ZACHARY,R  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: COOLEY,LOGAN  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 05/05/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402558895  
DRY GAS EXPIRY DATE: 10/03/2025  
VIOLATION CODE: DUI  
ARREST TIME: 01:09  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: WATSON

DATE/TIME OF TEST: 05/20/2024 02:14:42

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:07 |
| Air Blank          | 0.000 | 02:08 |
| Control Test       | 0.080 | 02:08 |
| Air Blank          | 0.000 | 02:08 |
| Subject Sample # 1 | 0.163 | 02:09 |
| Breath Volume      | 1.613 | 02:09 |
| Air Blank          | 0.000 | 02:10 |
| Air Blank          | 0.000 | 02:11 |
| Subject Sample # 2 | 0.156 | 02:13 |
| Breath Volume      | 1.437 | 02:13 |
| Air Blank          | 0.000 | 02:13 |
| Control Test       | 0.080 | 02:14 |
| Air Blank          | 0.000 | 02:14 |
| Diagnostics Check  | OK    | 02:14 |

---

Data Download Date/Time: 06/07/2024 02:51

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ZOLLER,J,G  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: ZOLLER,J  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 05/05/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402558895  
DRY GAS EXPIRY DATE: 10/03/2025  
VIOLATION CODE: DUI  
ARREST TIME: 18:57  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 19:54

SUBJECT'S LAST NAME: BROGDON

DATE/TIME OF TEST: 05/21/2024 20:21:08

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 20:14 |
| Air Blank          | 0.000 | 20:15 |
| Control Test       | 0.082 | 20:15 |
| Air Blank          | 0.000 | 20:16 |
| Subject Sample # 1 | 0.000 | 20:16 |
| Breath Volume      | 1.171 | 20:16 |
| Air Blank          | 0.000 | 20:17 |
| Air Blank          | 0.000 | 20:19 |
| Subject Sample # 2 | 0.000 | 20:19 |
| Breath Volume      | 1.894 | 20:19 |
| Air Blank          | 0.000 | 20:20 |
| Control Test       | 0.080 | 20:20 |
| Air Blank          | 0.000 | 20:21 |
| Diagnostics Check  | OK    | 20:21 |

---

Data Download Date/Time: 06/07/2024 02:51

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NELSON,JESSICA,R  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: NELSON,JESSICA  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 05/05/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402558895  
DRY GAS EXPIRY DATE: 10/03/2025  
VIOLATION CODE: DUI  
ARREST TIME: 00:38  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:04

SUBJECT'S LAST NAME: SANFORD

DATE/TIME OF TEST: 05/22/2024 01:31:13

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:24 |
| Air Blank          | 0.000 | 01:25 |
| Control Test       | 0.081 | 01:25 |
| Air Blank          | 0.000 | 01:25 |
| Subject Sample # 1 | 0.200 | 01:26 |
| Breath Volume      | 2.058 | 01:26 |
| Air Blank          | 0.000 | 01:27 |
| Air Blank          | 0.000 | 01:28 |
| Subject Sample # 2 | 0.196 | 01:29 |
| Breath Volume      | 2.441 | 01:29 |
| Air Blank          | 0.000 | 01:30 |
| Control Test       | 0.080 | 01:30 |
| Air Blank          | 0.000 | 01:31 |
| Diagnostics Check  | OK    | 01:31 |

---

Data Download Date/Time: 06/07/2024 02:51

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NELSON,JESSICA,R  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: NELSON,JESSICA  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 05/05/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402558895  
DRY GAS EXPIRY DATE: 10/03/2025  
VIOLATION CODE: DUI  
ARREST TIME: 22:12  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:28

SUBJECT'S LAST NAME: GREEN

DATE/TIME OF TEST: 05/24/2024 22:56:23

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:49 |
| Air Blank          | 0.000 | 22:49 |
| Control Test       | 0.080 | 22:49 |
| Air Blank          | 0.000 | 22:50 |
| Subject Sample # 1 | 0.156 | 22:51 |
| Breath Volume      | 1.550 | 22:51 |
| Air Blank          | 0.000 | 22:52 |
| Air Blank          | 0.000 | 22:53 |
| Subject Sample # 2 | 0.157 | 22:54 |
| Breath Volume      | 1.851 | 22:54 |
| Air Blank          | 0.000 | 22:55 |
| Control Test       | 0.080 | 22:55 |
| Air Blank          | 0.000 | 22:56 |
| Diagnostics Check  | OK    | 22:56 |

---

Data Download Date/Time: 06/07/2024 02:51

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LOPER,CADEN,D  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: LOPER,CADEN  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 05/05/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402558895  
DRY GAS EXPIRY DATE: 10/03/2025  
VIOLATION CODE: DUI  
ARREST TIME: 12:20  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 12:57

SUBJECT'S LAST NAME: YORK

DATE/TIME OF TEST: 05/25/2024 13:26:28

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 13:18 |
| Air Blank          | 0.000 | 13:18 |
| Control Test       | 0.080 | 13:18 |
| Air Blank          | 0.000 | 13:19 |
| Subject Sample # 1 | 0.232 | 13:20 |
| Breath Volume      | 1.335 | 13:20 |
| Air Blank          | 0.000 | 13:21 |
| Air Blank          | 0.000 | 13:23 |
| Subject Sample # 2 | 0.241 | 13:24 |
| Breath Volume      | 2.023 | 13:24 |
| Air Blank          | 0.000 | 13:25 |
| Control Test       | 0.079 | 13:25 |
| Air Blank          | 0.000 | 13:26 |
| Diagnostics Check  | OK    | 13:26 |

---

Data Download Date/Time: 06/07/2024 02:51

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LOPER,CADEN,D  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: LOPER,CADEN  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 05/05/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402558895  
DRY GAS EXPIRY DATE: 10/03/2025  
VIOLATION CODE: DUI  
ARREST TIME: 16:10  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 16:50

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 05/25/2024 17:19:39

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 17:11 |
| Air Blank          | 0.000 | 17:11 |
| Control Test       | 0.081 | 17:11 |
| Air Blank          | 0.000 | 17:12 |
| Subject Sample # 1 | 0.148 | 17:14 |
| Breath Volume      | 1.351 | 17:14 |
| Air Blank          | 0.000 | 17:15 |
| Air Blank          | 0.000 | 17:16 |
| Subject Sample # 2 | 0.149 | 17:18 |
| Breath Volume      | 1.203 | 17:18 |
| Air Blank          | 0.000 | 17:18 |
| Control Test       | 0.079 | 17:19 |
| Air Blank          | 0.000 | 17:19 |
| Diagnostics Check  | OK    | 17:19 |

---

Data Download Date/Time: 06/07/2024 02:51

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CAREY,ANDREW,S  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: CAREY,ANDREW  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 05/05/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402558895  
DRY GAS EXPIRY DATE: 10/03/2025  
VIOLATION CODE: DUI  
ARREST TIME: 23:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: SEVERINO PEREZ

DATE/TIME OF TEST: 05/26/2024 00:29:33

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:22 |
| Air Blank          | 0.000 | 00:23 |
| Control Test       | 0.080 | 00:23 |
| Air Blank          | 0.000 | 00:23 |
| Subject Sample # 1 | 0.046 | 00:25 |
| Breath Volume      | 1.332 | 00:25 |
| Air Blank          | 0.000 | 00:25 |
| Air Blank          | 0.000 | 00:27 |
| Subject Sample # 2 | 0.043 | 00:28 |
| Breath Volume      | 1.550 | 00:28 |
| Air Blank          | 0.000 | 00:28 |
| Control Test       | 0.080 | 00:28 |
| Air Blank          | 0.000 | 00:29 |
| Diagnostics Check  | OK    | 00:29 |

---

Data Download Date/Time: 06/07/2024 02:51

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BLACKWOOD,PATRICK,A  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: RAMIREZ,BRYANT  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 05/05/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402558895  
DRY GAS EXPIRY DATE: 10/03/2025  
VIOLATION CODE: DUI  
ARREST TIME: 21:04  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:32

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 05/26/2024 21:58:04

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 21:53 |                      |
| Air Blank         | 0.000 | 21:54 |                      |
| Control Test      | 0.080 | 21:54 |                      |
| Air Blank         | 0.000 | 21:55 |                      |
| Subject Sample    | # 1   | 21:56 | Subject Test Refused |
| Breath Volume     | 0.000 | 21:56 |                      |
| Air Blank         | 0.000 | 21:57 |                      |
| Control Test      | 0.080 | 21:57 |                      |
| Air Blank         | 0.000 | 21:57 |                      |
| Diagnostics Check | OK    | 21:58 |                      |

---

Data Download Date/Time: 06/07/2024 02:51

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CAREY,ANDREW,S  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: FELDHAUS,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 05/05/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402558895  
DRY GAS EXPIRY DATE: 10/03/2025  
VIOLATION CODE: DUI  
ARREST TIME: 01:26  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:57

SUBJECT'S LAST NAME: MCQUAID

DATE/TIME OF TEST: 05/27/2024 02:37:19

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:30 |
| Air Blank          | 0.000 | 02:30 |
| Control Test       | 0.080 | 02:30 |
| Air Blank          | 0.000 | 02:31 |
| Subject Sample # 1 | 0.096 | 02:32 |
| Breath Volume      | 1.429 | 02:32 |
| Air Blank          | 0.000 | 02:33 |
| Air Blank          | 0.000 | 02:34 |
| Subject Sample # 2 | 0.091 | 02:35 |
| Breath Volume      | 2.433 | 02:35 |
| Air Blank          | 0.000 | 02:36 |
| Control Test       | 0.078 | 02:36 |
| Air Blank          | 0.000 | 02:37 |
| Diagnostics Check  | OK    | 02:37 |

---

Data Download Date/Time: 06/07/2024 02:52

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MILLIGAN,CONNIE,A  
BREATH TEST OPERATOR AGENCY: Lake City PD  
ARRESTING OFFICER: WATTS,D  
ARRESTING OFFICER AGENCY: Lake City PD  
LAST AGENCY INSPECTION: 04/29/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08622080A3  
DRY GAS EXPIRY DATE: 06/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 23:06  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: KENDRICK

DATE/TIME OF TEST: 05/04/2024 00:11:40

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:05 |
| Air Blank          | 0.000 | 00:05 |
| Control Test       | 0.079 | 00:05 |
| Air Blank          | 0.000 | 00:06 |
| Subject Sample # 1 | 0.135 | 00:07 |
| Breath Volume      | 1.523 | 00:07 |
| Air Blank          | 0.000 | 00:07 |
| Air Blank          | 0.000 | 00:09 |
| Subject Sample # 2 | 0.132 | 00:10 |
| Breath Volume      | 1.984 | 00:10 |
| Air Blank          | 0.000 | 00:10 |
| Control Test       | 0.075 | 00:11 |
| Air Blank          | 0.000 | 00:11 |
| Diagnostics Check  | OK    | 00:11 |

---

Data Download Date/Time: 06/03/2024 11:10

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LORTON,CORNELIUS,Q  
BREATH TEST OPERATOR AGENCY: Columbia County SO  
ARRESTING OFFICER: LORTON,CORNELIUS  
ARRESTING OFFICER AGENCY: Columbia County SO  
LAST AGENCY INSPECTION: 04/29/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08622080A3  
DRY GAS EXPIRY DATE: 06/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 03:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 05/05/2024 04:27:41

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:20 |
| Air Blank          | 0.000 | 04:21 |
| Control Test       | 0.078 | 04:21 |
| Air Blank          | 0.000 | 04:22 |
| Subject Sample # 1 | 0.159 | 04:23 |
| Breath Volume      | 2.406 | 04:23 |
| Air Blank          | 0.000 | 04:23 |
| Air Blank          | 0.000 | 04:25 |
| Subject Sample # 2 | 0.144 | 04:26 |
| Breath Volume      | 1.281 | 04:26 |
| Air Blank          | 0.000 | 04:26 |
| Control Test       | 0.077 | 04:27 |
| Air Blank          | 0.000 | 04:27 |
| Diagnostics Check  | OK    | 04:27 |

---

Data Download Date/Time: 06/03/2024 11:10

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LOVICK-PLEMMONS,L,M  
BREATH TEST OPERATOR AGENCY: Lake City PD  
ARRESTING OFFICER: BURNSED,J  
ARRESTING OFFICER AGENCY: Lake City PD  
LAST AGENCY INSPECTION: 04/29/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08622080A3  
DRY GAS EXPIRY DATE: 06/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 16:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:35

SUBJECT'S LAST NAME: MATTIMORE

DATE/TIME OF TEST: 05/05/2024 17:15:38

---

|                   |           |       |                           |
|-------------------|-----------|-------|---------------------------|
| Diagnostics Check | OK        | 17:06 |                           |
| Air Blank         | 0.000     | 17:06 |                           |
| Control Test      | 0.078     | 17:06 |                           |
| Air Blank         | 0.000     | 17:07 |                           |
| Subject Sample    | # 1 0.370 | 17:07 |                           |
| Breath Volume     | 2.394     | 17:07 |                           |
| Air Blank         | 0.000     | 17:08 |                           |
| Air Blank         | 0.000     | 17:10 |                           |
| Subject Sample    | # 2 0.402 | 17:10 |                           |
| Breath Volume     | 2.757     | 17:10 |                           |
| Air Blank         | 0.000     | 17:11 |                           |
| Air Blank         | 0.000     | 17:13 |                           |
| Subject Sample    | # 3 0.390 | 17:13 |                           |
| Breath Volume     | 1.882     | 17:13 |                           |
| Air Blank         | 0.000     | 17:14 |                           |
| Control Test      | 0.073     | 17:15 | Control Outside Tolerance |
| Air Blank         | 0.000     | 17:15 |                           |

---

Data Download Date/Time: 06/03/2024 11:10

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SUMNER,NATHAN,J  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: SUMNER,NATHAN  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 04/29/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08622080A3  
DRY GAS EXPIRY DATE: 06/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 19:26  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 19:49

SUBJECT'S LAST NAME: MESSERVY

DATE/TIME OF TEST: 05/10/2024 20:16:33

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 20:10 |
| Air Blank          | 0.000 | 20:10 |
| Control Test       | 0.078 | 20:11 |
| Air Blank          | 0.000 | 20:11 |
| Subject Sample # 1 | 0.092 | 20:12 |
| Breath Volume      | 1.171 | 20:12 |
| Air Blank          | 0.000 | 20:12 |
| Air Blank          | 0.000 | 20:14 |
| Subject Sample # 2 | 0.092 | 20:14 |
| Breath Volume      | 1.761 | 20:14 |
| Air Blank          | 0.000 | 20:15 |
| Control Test       | 0.078 | 20:15 |
| Air Blank          | 0.000 | 20:16 |
| Diagnostics Check  | OK    | 20:16 |

---

Data Download Date/Time: 06/03/2024 11:10

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MORGAN,DARRAN,M  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: BURK,COREY  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 04/29/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08622080A3  
DRY GAS EXPIRY DATE: 06/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 13:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 14:08

SUBJECT'S LAST NAME: ORTEGA DE GRACIA

DATE/TIME OF TEST: 05/14/2024 14:38:06

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 14:31 |
| Air Blank          | 0.000 | 14:31 |
| Control Test       | 0.078 | 14:32 |
| Air Blank          | 0.000 | 14:32 |
| Subject Sample # 1 | 0.170 | 14:33 |
| Breath Volume      | 2.101 | 14:33 |
| Air Blank          | 0.000 | 14:34 |
| Air Blank          | 0.000 | 14:35 |
| Subject Sample # 2 | 0.168 | 14:36 |
| Breath Volume      | 1.437 | 14:36 |
| Air Blank          | 0.000 | 14:37 |
| Control Test       | 0.078 | 14:37 |
| Air Blank          | 0.000 | 14:38 |
| Diagnostics Check  | OK    | 14:38 |

---

Data Download Date/Time: 06/03/2024 11:10

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SUMNER,NATHAN,J  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: SUMNER,NATHAN  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 04/29/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08622080A3  
DRY GAS EXPIRY DATE: 06/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 22:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 05/20/2024 23:53:05

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:47 |
| Air Blank          | 0.000 | 23:47 |
| Control Test       | 0.080 | 23:48 |
| Air Blank          | 0.000 | 23:48 |
| Subject Sample # 1 | 0.016 | 23:48 |
| Breath Volume      | 4.144 | 23:48 |
| Air Blank          | 0.000 | 23:49 |
| Air Blank          | 0.000 | 23:51 |
| Subject Sample # 2 | 0.013 | 23:51 |
| Breath Volume      | 3.648 | 23:51 |
| Air Blank          | 0.000 | 23:52 |
| Control Test       | 0.077 | 23:52 |
| Air Blank          | 0.000 | 23:52 |
| Diagnostics Check  | OK    | 23:53 |

---

Data Download Date/Time: 06/03/2024 11:10

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SUMNER,NATHAN,J  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: SUMNER,NATHAN  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 04/29/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08622080A3  
DRY GAS EXPIRY DATE: 06/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 02:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:31

SUBJECT'S LAST NAME: MCCULLAR

DATE/TIME OF TEST: 05/25/2024 04:01:39

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:55 |
| Air Blank          | 0.000 | 03:55 |
| Control Test       | 0.080 | 03:56 |
| Air Blank          | 0.000 | 03:56 |
| Subject Sample # 1 | 0.178 | 03:57 |
| Breath Volume      | 2.488 | 03:57 |
| Air Blank          | 0.000 | 03:57 |
| Air Blank          | 0.000 | 03:59 |
| Subject Sample # 2 | 0.177 | 03:59 |
| Breath Volume      | 2.484 | 03:59 |
| Air Blank          | 0.000 | 04:00 |
| Control Test       | 0.079 | 04:01 |
| Air Blank          | 0.000 | 04:01 |
| Diagnostics Check  | OK    | 04:01 |

---

Data Download Date/Time: 06/03/2024 11:10

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KOSKO,KRISTOPHER,R  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: KOSKO,KRISTOPHER  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 04/29/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08622080A3  
DRY GAS EXPIRY DATE: 06/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 00:11  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: BREVARD

DATE/TIME OF TEST: 05/30/2024 01:02:52

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:56 |
| Air Blank          | 0.000 | 00:57 |
| Control Test       | 0.080 | 00:57 |
| Air Blank          | 0.000 | 00:57 |
| Subject Sample # 1 | 0.133 | 00:58 |
| Breath Volume      | 1.507 | 00:58 |
| Air Blank          | 0.000 | 00:59 |
| Air Blank          | 0.000 | 01:00 |
| Subject Sample # 2 | 0.129 | 01:01 |
| Breath Volume      | 1.730 | 01:01 |
| Air Blank          | 0.000 | 01:01 |
| Control Test       | 0.078 | 01:02 |
| Air Blank          | 0.000 | 01:02 |
| Diagnostics Check  | OK    | 01:02 |

---

Data Download Date/Time: 06/03/2024 11:10

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HANSEN,ERIC,M  
BREATH TEST OPERATOR AGENCY: Lake City PD  
ARRESTING OFFICER: HANSEN,ERIC  
ARRESTING OFFICER AGENCY: Lake City PD  
LAST AGENCY INSPECTION: 04/29/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08622080A3  
DRY GAS EXPIRY DATE: 06/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 21:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:21

SUBJECT'S LAST NAME: BRADLEY

DATE/TIME OF TEST: 05/31/2024 22:09:27

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 22:06 |                      |
| Air Blank         | 0.000 | 22:06 |                      |
| Control Test      | 0.080 | 22:07 |                      |
| Air Blank         | 0.000 | 22:07 |                      |
| Subject Sample    | # 1   | 22:08 | Subject Test Refused |
| Breath Volume     | 0.000 | 22:08 |                      |
| Air Blank         | 0.000 | 22:08 |                      |
| Control Test      | 0.080 | 22:08 |                      |
| Air Blank         | 0.000 | 22:09 |                      |
| Diagnostics Check | OK    | 22:09 |                      |

---

Data Download Date/Time: 06/03/2024 11:10

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000779

INSTRUMENT REGISTERED TO: MADISON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SMITH,JIMMY,L  
BREATH TEST OPERATOR AGENCY: Madison County SO  
ARRESTING OFFICER: JOOST,MARK  
ARRESTING OFFICER AGENCY: Madison County SO  
LAST AGENCY INSPECTION: 05/08/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33922080A1  
DRY GAS EXPIRY DATE: 01/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 16:38  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:20

SUBJECT'S LAST NAME: MURRAY

DATE/TIME OF TEST: 05/27/2024 17:58:42

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 17:52 |
| Air Blank          | 0.000 | 17:53 |
| Control Test       | 0.079 | 17:53 |
| Air Blank          | 0.000 | 17:53 |
| Subject Sample # 1 | 0.161 | 17:54 |
| Breath Volume      | 1.230 | 17:54 |
| Air Blank          | 0.000 | 17:55 |
| Air Blank          | 0.000 | 17:56 |
| Subject Sample # 2 | 0.153 | 17:57 |
| Breath Volume      | 1.593 | 17:57 |
| Air Blank          | 0.000 | 17:57 |
| Control Test       | 0.077 | 17:58 |
| Air Blank          | 0.000 | 17:58 |
| Diagnostics Check  | OK    | 17:58 |

---

Data Download Date/Time: 06/11/2024 14:14

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000785

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HERRINGTON,STEPHEN,L  
BREATH TEST OPERATOR AGENCY: Holmes County SO  
ARRESTING OFFICER: HERRINGTON,STEPHEN  
ARRESTING OFFICER AGENCY: Holmes County SO  
LAST AGENCY INSPECTION: 05/21/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402328118  
DRY GAS EXPIRY DATE: 01/20/2025  
VIOLATION CODE: DUI  
ARREST TIME: 02:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: ARNETTE

DATE/TIME OF TEST: 05/26/2024 02:40:19

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:31 |
| Air Blank          | 0.000 | 02:31 |
| Control Test       | 0.080 | 02:31 |
| Air Blank          | 0.000 | 02:32 |
| Subject Sample # 1 | 0.208 | 02:35 |
| Breath Volume      | 1.953 | 02:35 |
| Air Blank          | 0.000 | 02:35 |
| Air Blank          | 0.000 | 02:37 |
| Subject Sample # 2 | 0.210 | 02:38 |
| Breath Volume      | 2.585 | 02:38 |
| Air Blank          | 0.000 | 02:39 |
| Control Test       | 0.080 | 02:39 |
| Air Blank          | 0.000 | 02:40 |
| Diagnostics Check  | OK    | 02:40 |

---

Data Download Date/Time: 06/05/2024 10:30

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CHANTHAVONG,VIRGINA,A  
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD  
ARRESTING OFFICER: ERLEMANN,JOSHUA  
ARRESTING OFFICER AGENCY: Gulf Breeze PD  
LAST AGENCY INSPECTION: 04/09/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402367487  
DRY GAS EXPIRY DATE: 04/28/2025  
VIOLATION CODE: DUI  
ARREST TIME: 03:12  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: CHANTHAVONG

DATE/TIME OF TEST: 05/08/2024 03:48:26

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:42 |
| Air Blank          | 0.000 | 03:42 |
| Control Test       | 0.078 | 03:43 |
| Air Blank          | 0.000 | 03:43 |
| Subject Sample # 1 | 0.069 | 03:44 |
| Breath Volume      | 2.027 | 03:44 |
| Air Blank          | 0.000 | 03:44 |
| Air Blank          | 0.000 | 03:46 |
| Subject Sample # 2 | 0.064 | 03:46 |
| Breath Volume      | 2.503 | 03:46 |
| Air Blank          | 0.000 | 03:47 |
| Control Test       | 0.078 | 03:47 |
| Air Blank          | 0.000 | 03:48 |
| Diagnostics Check  | OK    | 03:48 |

---

Data Download Date/Time: 05/26/2024 01:17

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ALEXANDER,RICHARD,D  
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD  
ARRESTING OFFICER: ALEXANDER,RICHARD  
ARRESTING OFFICER AGENCY: Gulf Breeze PD  
LAST AGENCY INSPECTION: 04/09/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402367487  
DRY GAS EXPIRY DATE: 04/28/2025  
VIOLATION CODE: DUI  
ARREST TIME: 13:35  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 13:55

SUBJECT'S LAST NAME: PARKER

DATE/TIME OF TEST: 05/11/2024 14:27:39

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 14:20 |
| Air Blank          | 0.000 | 14:21 |
| Control Test       | 0.078 | 14:21 |
| Air Blank          | 0.000 | 14:22 |
| Subject Sample # 1 | 0.000 | 14:23 |
| Breath Volume      | 3.398 | 14:23 |
| Air Blank          | 0.000 | 14:23 |
| Air Blank          | 0.000 | 14:25 |
| Subject Sample # 2 | 0.000 | 14:26 |
| Breath Volume      | 3.359 | 14:26 |
| Air Blank          | 0.000 | 14:26 |
| Control Test       | 0.080 | 14:27 |
| Air Blank          | 0.000 | 14:27 |
| Diagnostics Check  | OK    | 14:27 |

---

Data Download Date/Time: 05/26/2024 01:18

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAVIS,NATE,K  
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD  
ARRESTING OFFICER: DAVIS,NATE  
ARRESTING OFFICER AGENCY: Gulf Breeze PD  
LAST AGENCY INSPECTION: 04/09/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402367487  
DRY GAS EXPIRY DATE: 04/28/2025  
VIOLATION CODE: DUI  
ARREST TIME: 16:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:15

SUBJECT'S LAST NAME: EUBANK

DATE/TIME OF TEST: 05/18/2024 17:44:39

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 17:37 |
| Air Blank          | 0.000 | 17:37 |
| Control Test       | 0.079 | 17:37 |
| Air Blank          | 0.000 | 17:38 |
| Subject Sample # 1 | 0.168 | 17:39 |
| Breath Volume      | 1.457 | 17:39 |
| Air Blank          | 0.000 | 17:40 |
| Air Blank          | 0.000 | 17:42 |
| Subject Sample # 2 | 0.177 | 17:43 |
| Breath Volume      | 2.875 | 17:43 |
| Air Blank          | 0.000 | 17:43 |
| Control Test       | 0.076 | 17:44 |
| Air Blank          | 0.000 | 17:44 |
| Diagnostics Check  | OK    | 17:44 |

---

Data Download Date/Time: 05/26/2024 01:18

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000791

INSTRUMENT REGISTERED TO: LARGO PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ASBURY,CHRISTOPHER,M  
BREATH TEST OPERATOR AGENCY: Largo PD  
ARRESTING OFFICER: LOUX,SAGE  
ARRESTING OFFICER AGENCY: Largo PD  
LAST AGENCY INSPECTION: 04/04/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15922080A4  
DRY GAS EXPIRY DATE: 08/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 16:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 15:38

SUBJECT'S LAST NAME: DELISE

DATE/TIME OF TEST: 05/04/2024 17:08:12

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 17:01 |
| Air Blank          | 0.000 | 17:01 |
| Control Test       | 0.080 | 17:02 |
| Air Blank          | 0.000 | 17:02 |
| Subject Sample # 1 | 0.295 | 17:03 |
| Breath Volume      | 3.058 | 17:03 |
| Air Blank          | 0.000 | 17:04 |
| Air Blank          | 0.000 | 17:06 |
| Subject Sample # 2 | 0.276 | 17:06 |
| Breath Volume      | 2.390 | 17:06 |
| Air Blank          | 0.000 | 17:07 |
| Control Test       | 0.078 | 17:07 |
| Air Blank          | 0.000 | 17:08 |
| Diagnostics Check  | OK    | 17:08 |

---

Data Download Date/Time: 05/30/2024 15:07

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000791

INSTRUMENT REGISTERED TO: LARGO PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,  
BREATH TEST OPERATOR AGENCY: Largo PD  
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA  
ARRESTING OFFICER AGENCY: Largo PD  
LAST AGENCY INSPECTION: 04/04/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15922080A4  
DRY GAS EXPIRY DATE: 08/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 02:34  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: SAVAGE

DATE/TIME OF TEST: 05/05/2024 03:38:40

---

|                   |           |       |                 |
|-------------------|-----------|-------|-----------------|
| Diagnostics Check | OK        | 03:25 |                 |
| Air Blank         | 0.000     | 03:26 |                 |
| Control Test      | 0.080     | 03:26 |                 |
| Air Blank         | 0.000     | 03:27 |                 |
| Subject Sample    | # 1 0.078 | 03:30 | Slope Not Level |
| Breath Volume     | 1.343     | 03:30 |                 |
| Air Blank         | 0.000     | 03:31 |                 |
| Air Blank         | 0.000     | 03:32 |                 |
| Subject Sample    | # 2 0.077 | 03:33 |                 |
| Breath Volume     | 1.511     | 03:33 |                 |
| Air Blank         | 0.000     | 03:34 |                 |
| Air Blank         | 0.000     | 03:35 |                 |
| Subject Sample    | # 3 0.076 | 03:37 |                 |
| Breath Volume     | 1.269     | 03:37 |                 |
| Air Blank         | 0.000     | 03:37 |                 |
| Control Test      | 0.080     | 03:38 |                 |
| Air Blank         | 0.000     | 03:38 |                 |
| Diagnostics Check | OK        | 03:38 |                 |

---

Data Download Date/Time: 05/30/2024 15:07

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000791

INSTRUMENT REGISTERED TO: LARGO PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CLARK,RYAN,N  
BREATH TEST OPERATOR AGENCY: Largo PD  
ARRESTING OFFICER: TUMICKI,SAMANTHA  
ARRESTING OFFICER AGENCY: Largo PD  
LAST AGENCY INSPECTION: 04/04/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15922080A4  
DRY GAS EXPIRY DATE: 08/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 23:18  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:28

SUBJECT'S LAST NAME: OLIVARES

DATE/TIME OF TEST: 05/05/2024 23:51:29

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 23:48 |                      |
| Air Blank         | 0.000 | 23:48 |                      |
| Control Test      | 0.080 | 23:49 |                      |
| Air Blank         | 0.000 | 23:49 |                      |
| Subject Sample    | # 1   | 23:50 | Subject Test Refused |
| Breath Volume     | 0.000 | 23:50 |                      |
| Air Blank         | 0.000 | 23:50 |                      |
| Control Test      | 0.079 | 23:50 |                      |
| Air Blank         | 0.000 | 23:51 |                      |
| Diagnostics Check | OK    | 23:51 |                      |

---

Data Download Date/Time: 05/30/2024 15:07

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000791

INSTRUMENT REGISTERED TO: LARGO PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,  
BREATH TEST OPERATOR AGENCY: Largo PD  
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA  
ARRESTING OFFICER AGENCY: Largo PD  
LAST AGENCY INSPECTION: 04/04/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15922080A4  
DRY GAS EXPIRY DATE: 08/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 00:27  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:46

SUBJECT'S LAST NAME: KIKIC

DATE/TIME OF TEST: 05/11/2024 01:09:43

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 01:06 |                      |
| Air Blank         | 0.000 | 01:07 |                      |
| Control Test      | 0.079 | 01:07 |                      |
| Air Blank         | 0.000 | 01:07 |                      |
| Subject Sample    | # 1   | 01:08 | Subject Test Refused |
| Breath Volume     | 0.000 | 01:08 |                      |
| Air Blank         | 0.000 | 01:08 |                      |
| Control Test      | 0.080 | 01:09 |                      |
| Air Blank         | 0.000 | 01:09 |                      |
| Diagnostics Check | OK    | 01:09 |                      |

---

Data Download Date/Time: 05/30/2024 15:07

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000791

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SINGSON,ROLAND,A

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: SINGSON,ROLAND

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 04/04/2024

Date or Time Changed: No

DRY GAS STANDARD LOT #: 15922080A4

DRY GAS EXPIRY DATE: 08/05/2024

VIOLATION CODE: DUI

ARREST TIME: 15:13

GENDER: Male

OBSERVATION PERIOD BEGAN: 15:35

SUBJECT'S LAST NAME: ROEBUCK

DATE/TIME OF TEST: 05/11/2024 15:59:59

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 15:56 |                      |
| Air Blank         | 0.000 | 15:57 |                      |
| Control Test      | 0.080 | 15:57 |                      |
| Air Blank         | 0.000 | 15:58 |                      |
| Subject Sample    | # 1   | 15:58 | Subject Test Refused |
| Breath Volume     | 0.000 | 15:58 |                      |
| Air Blank         | 0.000 | 15:59 |                      |
| Control Test      | 0.079 | 15:59 |                      |
| Air Blank         | 0.000 | 15:59 |                      |
| Diagnostics Check | OK    | 15:59 |                      |

---

Data Download Date/Time: 05/30/2024 15:07

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000791

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: GOSIER,MATTHEW,D

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: GOSIER,MATTHEW

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 04/04/2024

Date or Time Changed: No

DRY GAS STANDARD LOT #: 15922080A4

DRY GAS EXPIRY DATE: 08/05/2024

VIOLATION CODE: DUI

ARREST TIME: 03:10

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:33

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 05/12/2024 04:01:42

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 03:58 |                      |
| Air Blank         | 0.000 | 03:59 |                      |
| Control Test      | 0.079 | 03:59 |                      |
| Air Blank         | 0.000 | 04:00 |                      |
| Subject Sample    | # 1   | 04:00 | Subject Test Refused |
| Breath Volume     | 0.000 | 04:00 |                      |
| Air Blank         | 0.000 | 04:00 |                      |
| Control Test      | 0.079 | 04:01 |                      |
| Air Blank         | 0.000 | 04:01 |                      |
| Diagnostics Check | OK    | 04:01 |                      |

---

Data Download Date/Time: 05/30/2024 15:07

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000791

INSTRUMENT REGISTERED TO: LARGO PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,  
BREATH TEST OPERATOR AGENCY: Largo PD  
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA  
ARRESTING OFFICER AGENCY: Largo PD  
LAST AGENCY INSPECTION: 04/04/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15922080A4  
DRY GAS EXPIRY DATE: 08/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 03:13  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:33

SUBJECT'S LAST NAME: ROMAN

DATE/TIME OF TEST: 05/16/2024 03:59:57

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:53 |
| Air Blank          | 0.000 | 03:54 |
| Control Test       | 0.080 | 03:54 |
| Air Blank          | 0.000 | 03:54 |
| Subject Sample # 1 | 0.112 | 03:55 |
| Breath Volume      | 2.265 | 03:55 |
| Air Blank          | 0.000 | 03:56 |
| Air Blank          | 0.000 | 03:57 |
| Subject Sample # 2 | 0.113 | 03:58 |
| Breath Volume      | 1.910 | 03:58 |
| Air Blank          | 0.000 | 03:59 |
| Control Test       | 0.080 | 03:59 |
| Air Blank          | 0.000 | 03:59 |
| Diagnostics Check  | OK    | 03:59 |

---

Data Download Date/Time: 05/30/2024 15:08

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000791

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: GOSIER,MATTHEW,D

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: GOSIER,MATTHEW

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 04/04/2024

Date or Time Changed: No

DRY GAS STANDARD LOT #: 15922080A4

DRY GAS EXPIRY DATE: 08/05/2024

VIOLATION CODE: DUI

ARREST TIME: 18:48

GENDER: Male

OBSERVATION PERIOD BEGAN: 19:11

SUBJECT'S LAST NAME: LOWE

DATE/TIME OF TEST: 05/16/2024 19:39:17

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 19:32 |
| Air Blank          | 0.000 | 19:33 |
| Control Test       | 0.079 | 19:33 |
| Air Blank          | 0.000 | 19:34 |
| Subject Sample # 1 | 0.172 | 19:34 |
| Breath Volume      | 1.191 | 19:34 |
| Air Blank          | 0.000 | 19:35 |
| Air Blank          | 0.000 | 19:37 |
| Subject Sample # 2 | 0.168 | 19:37 |
| Breath Volume      | 1.214 | 19:37 |
| Air Blank          | 0.000 | 19:38 |
| Control Test       | 0.080 | 19:38 |
| Air Blank          | 0.000 | 19:39 |
| Diagnostics Check  | OK    | 19:39 |

---

Data Download Date/Time: 05/30/2024 15:08

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000791

INSTRUMENT REGISTERED TO: LARGO PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,  
BREATH TEST OPERATOR AGENCY: Largo PD  
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA  
ARRESTING OFFICER AGENCY: Largo PD  
LAST AGENCY INSPECTION: 04/04/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15922080A4  
DRY GAS EXPIRY DATE: 08/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 22:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:52

SUBJECT'S LAST NAME: CASSANELLI

DATE/TIME OF TEST: 05/16/2024 23:19:52

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:13 |
| Air Blank          | 0.000 | 23:14 |
| Control Test       | 0.080 | 23:14 |
| Air Blank          | 0.000 | 23:14 |
| Subject Sample # 1 | 0.036 | 23:15 |
| Breath Volume      | 2.027 | 23:15 |
| Air Blank          | 0.000 | 23:15 |
| Air Blank          | 0.000 | 23:17 |
| Subject Sample # 2 | 0.036 | 23:18 |
| Breath Volume      | 2.847 | 23:18 |
| Air Blank          | 0.000 | 23:18 |
| Control Test       | 0.079 | 23:19 |
| Air Blank          | 0.000 | 23:19 |
| Diagnostics Check  | OK    | 23:19 |

---

Data Download Date/Time: 05/30/2024 15:08

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000791

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: YEATES,SCOTT

ARRESTING OFFICER AGENCY: Clearwater PD

LAST AGENCY INSPECTION: 04/04/2024

Date or Time Changed: No

DRY GAS STANDARD LOT #: 15922080A4

DRY GAS EXPIRY DATE: 08/05/2024

VIOLATION CODE: DUI

ARREST TIME: 01:26

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: MOSLEY

DATE/TIME OF TEST: 05/17/2024 03:17:34

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:10 |
| Air Blank          | 0.000 | 03:11 |
| Control Test       | 0.081 | 03:11 |
| Air Blank          | 0.000 | 03:12 |
| Subject Sample # 1 | 0.117 | 03:12 |
| Breath Volume      | 3.492 | 03:12 |
| Air Blank          | 0.000 | 03:13 |
| Air Blank          | 0.000 | 03:15 |
| Subject Sample # 2 | 0.121 | 03:15 |
| Breath Volume      | 3.964 | 03:15 |
| Air Blank          | 0.000 | 03:16 |
| Control Test       | 0.080 | 03:16 |
| Air Blank          | 0.000 | 03:17 |
| Diagnostics Check  | OK    | 03:17 |

---

Data Download Date/Time: 05/30/2024 15:08

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000791

INSTRUMENT REGISTERED TO: LARGO PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,  
BREATH TEST OPERATOR AGENCY: Largo PD  
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA  
ARRESTING OFFICER AGENCY: Largo PD  
LAST AGENCY INSPECTION: 04/04/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15922080A4  
DRY GAS EXPIRY DATE: 08/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 00:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: HUYNH

DATE/TIME OF TEST: 05/18/2024 01:36:23

---

|                   |       |       |                    |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK    | 01:23 |                    |
| Air Blank         | 0.000 | 01:23 |                    |
| Control Test      | 0.080 | 01:24 |                    |
| Air Blank         | 0.000 | 01:24 |                    |
| Subject Sample    | # 1   | 01:27 | No Sample Provided |
| Breath Volume     | 0.000 | 01:27 |                    |
| Air Blank         | 0.000 | 01:28 |                    |
| Air Blank         | 0.000 | 01:30 |                    |
| Subject Sample    | # 2   | 0.132 | 01:31              |
| Breath Volume     | 1.351 | 01:31 |                    |
| Air Blank         | 0.000 | 01:31 |                    |
| Air Blank         | 0.000 | 01:33 |                    |
| Subject Sample    | # 3   | 0.143 | 01:34              |
| Breath Volume     | 2.410 | 01:34 |                    |
| Air Blank         | 0.000 | 01:35 |                    |
| Control Test      | 0.078 | 01:35 |                    |
| Air Blank         | 0.000 | 01:36 |                    |
| Diagnostics Check | OK    | 01:36 |                    |

---

Data Download Date/Time: 05/30/2024 15:08

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000791

INSTRUMENT REGISTERED TO: LARGO PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GOSIER,MATTHEW,D  
BREATH TEST OPERATOR AGENCY: Largo PD  
ARRESTING OFFICER: GOSIER,MATTHEW  
ARRESTING OFFICER AGENCY: Largo PD  
LAST AGENCY INSPECTION: 04/04/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15922080A4  
DRY GAS EXPIRY DATE: 08/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 01:31  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: ARENT

DATE/TIME OF TEST: 05/19/2024 02:23:10

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:16 |
| Air Blank          | 0.000 | 02:17 |
| Control Test       | 0.080 | 02:17 |
| Air Blank          | 0.000 | 02:17 |
| Subject Sample # 1 | 0.186 | 02:18 |
| Breath Volume      | 2.347 | 02:18 |
| Air Blank          | 0.000 | 02:19 |
| Air Blank          | 0.000 | 02:21 |
| Subject Sample # 2 | 0.188 | 02:21 |
| Breath Volume      | 1.507 | 02:21 |
| Air Blank          | 0.000 | 02:22 |
| Control Test       | 0.078 | 02:22 |
| Air Blank          | 0.000 | 02:23 |
| Diagnostics Check  | OK    | 02:23 |

---

Data Download Date/Time: 05/30/2024 15:08

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000791

INSTRUMENT REGISTERED TO: LARGO PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GOSIER,MATTHEW,D  
BREATH TEST OPERATOR AGENCY: Largo PD  
ARRESTING OFFICER: GOSIER,MATTHEW  
ARRESTING OFFICER AGENCY: Largo PD  
LAST AGENCY INSPECTION: 04/04/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15922080A4  
DRY GAS EXPIRY DATE: 08/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 23:19  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: WOLF

DATE/TIME OF TEST: 05/20/2024 00:08:24

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:02 |
| Air Blank          | 0.000 | 00:02 |
| Control Test       | 0.080 | 00:02 |
| Air Blank          | 0.000 | 00:03 |
| Subject Sample # 1 | 0.195 | 00:03 |
| Breath Volume      | 1.886 | 00:03 |
| Air Blank          | 0.000 | 00:04 |
| Air Blank          | 0.000 | 00:06 |
| Subject Sample # 2 | 0.196 | 00:06 |
| Breath Volume      | 1.976 | 00:06 |
| Air Blank          | 0.000 | 00:07 |
| Control Test       | 0.079 | 00:07 |
| Air Blank          | 0.000 | 00:08 |
| Diagnostics Check  | OK    | 00:08 |

---

Data Download Date/Time: 05/30/2024 15:08

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000791

INSTRUMENT REGISTERED TO: LARGO PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GOSIER,MATTHEW,D  
BREATH TEST OPERATOR AGENCY: Largo PD  
ARRESTING OFFICER: GOSIER,MATTHEW  
ARRESTING OFFICER AGENCY: Largo PD  
LAST AGENCY INSPECTION: 04/04/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15922080A4  
DRY GAS EXPIRY DATE: 08/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 01:24  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:41

SUBJECT'S LAST NAME: PLUMRIDGE

DATE/TIME OF TEST: 05/20/2024 02:11:23

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:04 |
| Air Blank          | 0.000 | 02:04 |
| Control Test       | 0.080 | 02:05 |
| Air Blank          | 0.000 | 02:05 |
| Subject Sample # 1 | 0.093 | 02:06 |
| Breath Volume      | 2.019 | 02:06 |
| Air Blank          | 0.000 | 02:06 |
| Air Blank          | 0.000 | 02:08 |
| Subject Sample # 2 | 0.093 | 02:09 |
| Breath Volume      | 1.644 | 02:09 |
| Air Blank          | 0.000 | 02:10 |
| Control Test       | 0.080 | 02:10 |
| Air Blank          | 0.000 | 02:11 |
| Diagnostics Check  | OK    | 02:11 |

---

Data Download Date/Time: 05/30/2024 15:08

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000791

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: UPCHURCH,JERRY,S

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: DEPIERRE,JONATHAN

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 04/04/2024

Date or Time Changed: No

DRY GAS STANDARD LOT #: 15922080A4

DRY GAS EXPIRY DATE: 08/05/2024

VIOLATION CODE: DUI

ARREST TIME: 03:50

GENDER: Male

OBSERVATION PERIOD BEGAN: 05:05

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 05/23/2024 05:37:34

---

|                   |       |       |                    |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK    | 05:26 |                    |
| Air Blank         | 0.000 | 05:26 |                    |
| Control Test      | 0.080 | 05:26 |                    |
| Air Blank         | 0.000 | 05:27 |                    |
| Subject Sample    | # 1   | 05:30 | No Sample Provided |
| Breath Volume     | 0.000 | 05:30 |                    |
| Air Blank         | 0.000 | 05:31 |                    |
| Air Blank         | 0.000 | 05:33 |                    |
| Subject Sample    | # 2   | 05:36 | No Sample Provided |
| Breath Volume     | 0.000 | 05:36 |                    |
| Air Blank         | 0.000 | 05:36 |                    |
| Control Test      | 0.081 | 05:36 |                    |
| Air Blank         | 0.000 | 05:37 |                    |
| Diagnostics Check | OK    | 05:37 |                    |

---

Data Download Date/Time: 05/30/2024 15:08

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000791

INSTRUMENT REGISTERED TO: LARGO PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GOSIER,MATTHEW,D  
BREATH TEST OPERATOR AGENCY: Largo PD  
ARRESTING OFFICER: GOSIER,MATTHEW  
ARRESTING OFFICER AGENCY: Largo PD  
LAST AGENCY INSPECTION: 04/04/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15922080A4  
DRY GAS EXPIRY DATE: 08/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 03:23  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:48

SUBJECT'S LAST NAME: NOLASCO

DATE/TIME OF TEST: 05/26/2024 04:21:07

---

|                   |       |       |                    |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK    | 04:09 |                    |
| Air Blank         | 0.000 | 04:10 |                    |
| Control Test      | 0.080 | 04:10 |                    |
| Air Blank         | 0.000 | 04:11 |                    |
| Subject Sample    | # 1   | 04:14 | No Sample Provided |
| Breath Volume     | 0.000 | 04:14 |                    |
| Air Blank         | 0.000 | 04:14 |                    |
| Air Blank         | 0.000 | 04:16 |                    |
| Subject Sample    | # 2   | 04:19 | No Sample Provided |
| Breath Volume     | 0.000 | 04:19 |                    |
| Air Blank         | 0.000 | 04:20 |                    |
| Control Test      | 0.079 | 04:20 |                    |
| Air Blank         | 0.000 | 04:21 |                    |
| Diagnostics Check | OK    | 04:21 |                    |

---

Data Download Date/Time: 05/30/2024 15:08

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000791

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: GOSIER,MATTHEW,D

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: GOSIER,MATTHEW

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 04/04/2024

Date or Time Changed: No

DRY GAS STANDARD LOT #: 15922080A4

DRY GAS EXPIRY DATE: 08/05/2024

VIOLATION CODE: DUI

ARREST TIME: 01:08

GENDER: Female

OBSERVATION PERIOD BEGAN: 01:33

SUBJECT'S LAST NAME: BONANNO

DATE/TIME OF TEST: 05/27/2024 01:59:39

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:53 |
| Air Blank          | 0.000 | 01:53 |
| Control Test       | 0.080 | 01:54 |
| Air Blank          | 0.000 | 01:54 |
| Subject Sample # 1 | 0.000 | 01:55 |
| Breath Volume      | 1.453 | 01:55 |
| Air Blank          | 0.000 | 01:55 |
| Air Blank          | 0.000 | 01:57 |
| Subject Sample # 2 | 0.000 | 01:58 |
| Breath Volume      | 1.285 | 01:58 |
| Air Blank          | 0.000 | 01:58 |
| Control Test       | 0.080 | 01:59 |
| Air Blank          | 0.000 | 01:59 |
| Diagnostics Check  | OK    | 01:59 |

---

Data Download Date/Time: 05/30/2024 15:08

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000791

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: UPCHURCH,JERRY,S

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: GUILLORY,J

ARRESTING OFFICER AGENCY: Indian Shores PD

LAST AGENCY INSPECTION: 04/04/2024

Date or Time Changed: No

DRY GAS STANDARD LOT #: 15922080A4

DRY GAS EXPIRY DATE: 08/05/2024

VIOLATION CODE: DUI

ARREST TIME: 01:32

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: CHUILLI

DATE/TIME OF TEST: 05/28/2024 02:28:30

---

|                   |       |       |                    |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK    | 02:17 |                    |
| Air Blank         | 0.000 | 02:17 |                    |
| Control Test      | 0.080 | 02:17 |                    |
| Air Blank         | 0.000 | 02:18 |                    |
| Subject Sample    | # 1   | 02:21 | No Sample Provided |
| Breath Volume     | 0.000 | 02:21 |                    |
| Air Blank         | 0.000 | 02:21 |                    |
| Air Blank         | 0.000 | 02:23 |                    |
| Subject Sample    | # 2   | 02:27 | No Sample Provided |
| Breath Volume     | 0.000 | 02:27 |                    |
| Air Blank         | 0.000 | 02:27 |                    |
| Control Test      | 0.080 | 02:27 |                    |
| Air Blank         | 0.000 | 02:28 |                    |
| Diagnostics Check | OK    | 02:28 |                    |

---

Data Download Date/Time: 05/30/2024 15:08

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000791

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: UPCHURCH,JERRY,S

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: GUILLORY,J

ARRESTING OFFICER AGENCY: Indian Shores PD

LAST AGENCY INSPECTION: 04/04/2024

Date or Time Changed: No

DRY GAS STANDARD LOT #: 15922080A4

DRY GAS EXPIRY DATE: 08/05/2024

VIOLATION CODE: DUI

ARREST TIME: 01:32

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: CHUILLI

DATE/TIME OF TEST: 05/28/2024 02:36:19

---

|                   |       |       |                  |
|-------------------|-------|-------|------------------|
| Diagnostics Check | OK    | 02:34 |                  |
| Air Blank         | 0.000 | 02:34 |                  |
| Control Test      | 0.080 | 02:35 |                  |
| Air Blank         | 0.000 | 02:35 |                  |
| Subject Sample    | # 1   | 02:35 | Sequence Aborted |
| Breath Volume     | 0.000 | 02:35 |                  |
| Air Blank         | 0.000 | 02:36 |                  |

---

Data Download Date/Time: 05/30/2024 15:08

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000791

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SINGSON,ROLAND,A

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: SINGSON,ROLAND

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 04/04/2024

Date or Time Changed: No

DRY GAS STANDARD LOT #: 15922080A4

DRY GAS EXPIRY DATE: 08/05/2024

VIOLATION CODE: DUI

ARREST TIME: 17:42

GENDER: Male

OBSERVATION PERIOD BEGAN: 18:23

SUBJECT'S LAST NAME: BURGOS

DATE/TIME OF TEST: 05/28/2024 18:51:24

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 18:45 |
| Air Blank          | 0.000 | 18:45 |
| Control Test       | 0.080 | 18:45 |
| Air Blank          | 0.000 | 18:46 |
| Subject Sample # 1 | 0.000 | 18:46 |
| Breath Volume      | 3.027 | 18:46 |
| Air Blank          | 0.000 | 18:47 |
| Air Blank          | 0.000 | 18:49 |
| Subject Sample # 2 | 0.000 | 18:50 |
| Breath Volume      | 3.125 | 18:50 |
| Air Blank          | 0.000 | 18:50 |
| Control Test       | 0.080 | 18:50 |
| Air Blank          | 0.000 | 18:51 |
| Diagnostics Check  | OK    | 18:51 |

---

Data Download Date/Time: 05/30/2024 15:08

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CHUNN,PETER,E  
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD  
ARRESTING OFFICER: HOLLIGAN,ROMEO  
ARRESTING OFFICER AGENCY: Port St. Lucie PD  
LAST AGENCY INSPECTION: 04/25/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08223080A4  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 05:49  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:00

SUBJECT'S LAST NAME: HOLMES

DATE/TIME OF TEST: 05/01/2024 10:36:16

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 10:29 |
| Air Blank          | 0.000 | 10:30 |
| Control Test       | 0.081 | 10:30 |
| Air Blank          | 0.000 | 10:30 |
| Subject Sample # 1 | 0.000 | 10:31 |
| Breath Volume      | 4.722 | 10:31 |
| Air Blank          | 0.000 | 10:32 |
| Air Blank          | 0.000 | 10:34 |
| Subject Sample # 2 | 0.000 | 10:34 |
| Breath Volume      | 2.191 | 10:34 |
| Air Blank          | 0.000 | 10:35 |
| Control Test       | 0.081 | 10:35 |
| Air Blank          | 0.000 | 10:36 |
| Diagnostics Check  | OK    | 10:36 |

---

Data Download Date/Time: 05/27/2024 16:31

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JONES,CORTLAND,T  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: MCWILLIAMS,NICHOLAS  
ARRESTING OFFICER AGENCY: Fort Pierce PD  
LAST AGENCY INSPECTION: 04/25/2024

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 08223080A4  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 10:47  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 11:25

SUBJECT'S LAST NAME: SLOAN

DATE/TIME OF TEST: 05/03/2024 11:52:18

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 11:46 |
| Air Blank          | 0.000 | 11:46 |
| Control Test       | 0.081 | 11:46 |
| Air Blank          | 0.000 | 11:47 |
| Subject Sample # 1 | 0.000 | 11:47 |
| Breath Volume      | 3.003 | 11:47 |
| Air Blank          | 0.000 | 11:48 |
| Air Blank          | 0.000 | 11:50 |
| Subject Sample # 2 | 0.000 | 11:50 |
| Breath Volume      | 2.628 | 11:50 |
| Air Blank          | 0.000 | 11:51 |
| Control Test       | 0.081 | 11:51 |
| Air Blank          | 0.000 | 11:52 |
| Diagnostics Check  | OK    | 11:52 |

---

Data Download Date/Time: 05/27/2024 16:31

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JONES,CORTLAND,T  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: JONES,CORTLAND  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 04/25/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08223080A4  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 13:02  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 13:42

SUBJECT'S LAST NAME: MARICHAL GUERRA

DATE/TIME OF TEST: 05/03/2024 14:08:18

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 14:05 |                      |
| Air Blank         | 0.000 | 14:05 |                      |
| Control Test      | 0.081 | 14:06 |                      |
| Air Blank         | 0.000 | 14:06 |                      |
| Subject Sample    | # 1   | 14:06 | Subject Test Refused |
| Breath Volume     | 0.000 | 14:06 |                      |
| Air Blank         | 0.000 | 14:07 |                      |
| Control Test      | 0.082 | 14:07 |                      |
| Air Blank         | 0.000 | 14:08 |                      |
| Diagnostics Check | OK    | 14:08 |                      |

---

Data Download Date/Time: 05/27/2024 16:32

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BROCATO,SHAWN,V  
BREATH TEST OPERATOR AGENCY: Fort Pierce PD  
ARRESTING OFFICER: LAPOLLA,BRANDON  
ARRESTING OFFICER AGENCY: Fort Pierce PD  
LAST AGENCY INSPECTION: 04/25/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08223080A4  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 03:16  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: SANTIAGO

DATE/TIME OF TEST: 05/04/2024 04:23:08

---

|                   |           |       |                   |
|-------------------|-----------|-------|-------------------|
| Diagnostics Check | OK        | 04:08 |                   |
| Air Blank         | 0.000     | 04:08 |                   |
| Control Test      | 0.082     | 04:09 |                   |
| Air Blank         | 0.000     | 04:09 |                   |
| Subject Sample    | # 1 0.157 | 04:12 | No .020 agreement |
| Breath Volume     | 2.007     | 04:12 |                   |
| Air Blank         | 0.000     | 04:13 |                   |
| Air Blank         | 0.000     | 04:15 |                   |
| Subject Sample    | # 2 0.137 | 04:18 | Volume Not Met    |
| Breath Volume     | 0.875     | 04:18 |                   |
| Air Blank         | 0.000     | 04:19 |                   |
| Air Blank         | 0.000     | 04:21 |                   |
| Subject Sample    | # 3 0.129 | 04:21 | No .020 agreement |
| Breath Volume     | 1.777     | 04:21 |                   |
| Air Blank         | 0.000     | 04:22 |                   |
| Control Test      | 0.076     | 04:22 |                   |
| Air Blank         | 0.000     | 04:23 |                   |
| Diagnostics Check | OK        | 04:23 |                   |

---

Data Download Date/Time: 05/27/2024 16:32

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NIGH, ADAM, R  
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD  
ARRESTING OFFICER: CHALTOFF, SCOTT  
ARRESTING OFFICER AGENCY: Port St. Lucie PD  
LAST AGENCY INSPECTION: 04/25/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08223080A4  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 23:23  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: THIELMAN

DATE/TIME OF TEST: 05/05/2024 03:34:53

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:25 |
| Air Blank          | 0.000 | 03:25 |
| Control Test       | 0.080 | 03:25 |
| Air Blank          | 0.000 | 03:26 |
| Subject Sample # 1 | 0.218 | 03:26 |
| Breath Volume      | 2.746 | 03:26 |
| Air Blank          | 0.000 | 03:27 |
| Air Blank          | 0.000 | 03:29 |
| Subject Sample # 2 | 0.192 | 03:30 |
| Breath Volume      | 1.804 | 03:30 |
| Air Blank          | 0.000 | 03:31 |
| Air Blank          | 0.000 | 03:32 |
| Subject Sample # 3 | 0.206 | 03:33 |
| Breath Volume      | 1.707 | 03:33 |
| Air Blank          | 0.000 | 03:33 |
| Control Test       | 0.079 | 03:34 |
| Air Blank          | 0.000 | 03:34 |
| Diagnostics Check  | OK    | 03:34 |

---

Data Download Date/Time: 05/27/2024 16:32

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WAITES,DEJAS,A  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: BELL,WILLIAM  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 04/25/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08223080A4  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 00:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:37

SUBJECT'S LAST NAME: HOLLEY

DATE/TIME OF TEST: 05/09/2024 02:07:24

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:01 |
| Air Blank          | 0.000 | 02:01 |
| Control Test       | 0.082 | 02:01 |
| Air Blank          | 0.000 | 02:02 |
| Subject Sample # 1 | 0.089 | 02:02 |
| Breath Volume      | 3.453 | 02:02 |
| Air Blank          | 0.000 | 02:03 |
| Air Blank          | 0.000 | 02:05 |
| Subject Sample # 2 | 0.084 | 02:05 |
| Breath Volume      | 1.343 | 02:05 |
| Air Blank          | 0.000 | 02:06 |
| Control Test       | 0.081 | 02:06 |
| Air Blank          | 0.000 | 02:07 |
| Diagnostics Check  | OK    | 02:07 |

---

Data Download Date/Time: 05/27/2024 16:32

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BROCATO,SHAWN,V  
BREATH TEST OPERATOR AGENCY: Fort Pierce PD  
ARRESTING OFFICER: LAPOLLA,BRANDON  
ARRESTING OFFICER AGENCY: Fort Pierce PD  
LAST AGENCY INSPECTION: 04/25/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08223080A4  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 21:26  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: JIMINEZ-DEGANTE

DATE/TIME OF TEST: 05/10/2024 22:58:15

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:51 |
| Air Blank          | 0.000 | 22:52 |
| Control Test       | 0.082 | 22:52 |
| Air Blank          | 0.000 | 22:52 |
| Subject Sample # 1 | 0.263 | 22:53 |
| Breath Volume      | 2.121 | 22:53 |
| Air Blank          | 0.000 | 22:54 |
| Air Blank          | 0.000 | 22:55 |
| Subject Sample # 2 | 0.251 | 22:56 |
| Breath Volume      | 1.375 | 22:56 |
| Air Blank          | 0.000 | 22:57 |
| Control Test       | 0.079 | 22:57 |
| Air Blank          | 0.000 | 22:58 |
| Diagnostics Check  | OK    | 22:58 |

---

Data Download Date/Time: 05/27/2024 16:32

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BROCATO,SHAWN,V  
BREATH TEST OPERATOR AGENCY: Fort Pierce PD  
ARRESTING OFFICER: MEMON,MATTHEW  
ARRESTING OFFICER AGENCY: Fort Pierce PD  
LAST AGENCY INSPECTION: 04/25/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08223080A4  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 01:08  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:27

SUBJECT'S LAST NAME: FEBUS

DATE/TIME OF TEST: 05/11/2024 01:54:09

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:47 |
| Air Blank          | 0.000 | 01:47 |
| Control Test       | 0.081 | 01:48 |
| Air Blank          | 0.000 | 01:48 |
| Subject Sample # 1 | 0.155 | 01:49 |
| Breath Volume      | 1.273 | 01:49 |
| Air Blank          | 0.000 | 01:50 |
| Air Blank          | 0.000 | 01:52 |
| Subject Sample # 2 | 0.157 | 01:52 |
| Breath Volume      | 1.460 | 01:52 |
| Air Blank          | 0.000 | 01:53 |
| Control Test       | 0.080 | 01:53 |
| Air Blank          | 0.000 | 01:54 |
| Diagnostics Check  | OK    | 01:54 |

---

Data Download Date/Time: 05/27/2024 16:32

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RICCARDI,A,  
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD  
ARRESTING OFFICER: SANTAMARIA,J  
ARRESTING OFFICER AGENCY: Port St. Lucie PD  
LAST AGENCY INSPECTION: 04/25/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08223080A4  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 02:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: BELONY

DATE/TIME OF TEST: 05/11/2024 04:09:31

---

|                   |       |       |                    |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK    | 03:58 |                    |
| Air Blank         | 0.000 | 03:58 |                    |
| Control Test      | 0.081 | 03:58 |                    |
| Air Blank         | 0.000 | 03:59 |                    |
| Subject Sample    | # 1   | 04:02 | No Sample Provided |
| Breath Volume     | 0.000 | 04:02 |                    |
| Air Blank         | 0.000 | 04:03 |                    |
| Air Blank         | 0.000 | 04:05 |                    |
| Subject Sample    | # 2   | 04:08 | No Sample Provided |
| Breath Volume     | 0.000 | 04:08 |                    |
| Air Blank         | 0.000 | 04:08 |                    |
| Control Test      | 0.082 | 04:08 |                    |
| Air Blank         | 0.000 | 04:09 |                    |
| Diagnostics Check | OK    | 04:09 |                    |

---

Data Download Date/Time: 05/27/2024 16:32

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PEREZ,JUSTIN,J  
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD  
ARRESTING OFFICER: WILSON,TANERIA  
ARRESTING OFFICER AGENCY: Port St. Lucie PD  
LAST AGENCY INSPECTION: 04/25/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08223080A4  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 02:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: FERNANDEZ

DATE/TIME OF TEST: 05/12/2024 03:35:13

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 03:30 |                      |
| Air Blank         | 0.000 | 03:31 |                      |
| Control Test      | 0.081 | 03:31 |                      |
| Air Blank         | 0.000 | 03:32 |                      |
| Subject Sample    | # 1   | 03:33 | Subject Test Refused |
| Breath Volume     | 0.023 | 03:33 |                      |
| Air Blank         | 0.000 | 03:34 |                      |
| Control Test      | 0.082 | 03:34 |                      |
| Air Blank         | 0.000 | 03:35 |                      |
| Diagnostics Check | OK    | 03:35 |                      |

---

Data Download Date/Time: 05/27/2024 16:32

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RESTO,LUIS,A  
BREATH TEST OPERATOR AGENCY: FHP Troop K  
ARRESTING OFFICER: RESTO,LUIS  
ARRESTING OFFICER AGENCY: FHP Troop K  
LAST AGENCY INSPECTION: 04/25/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08223080A4  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 05:07  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:43

SUBJECT'S LAST NAME: CONTRERAS AVILA

DATE/TIME OF TEST: 05/12/2024 06:19:09

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 06:12 |
| Air Blank          | 0.000 | 06:13 |
| Control Test       | 0.081 | 06:13 |
| Air Blank          | 0.000 | 06:14 |
| Subject Sample # 1 | 0.222 | 06:14 |
| Breath Volume      | 2.484 | 06:14 |
| Air Blank          | 0.000 | 06:15 |
| Air Blank          | 0.000 | 06:17 |
| Subject Sample # 2 | 0.229 | 06:17 |
| Breath Volume      | 4.308 | 06:17 |
| Air Blank          | 0.000 | 06:18 |
| Control Test       | 0.080 | 06:18 |
| Air Blank          | 0.000 | 06:19 |
| Diagnostics Check  | OK    | 06:19 |

---

Data Download Date/Time: 05/27/2024 16:32

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BROCATO,SHAWN,V  
BREATH TEST OPERATOR AGENCY: Fort Pierce PD  
ARRESTING OFFICER: WOOD,KIMBERLY  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 04/25/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08223080A4  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 23:12  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:17

SUBJECT'S LAST NAME: TUBBY

DATE/TIME OF TEST: 05/17/2024 00:16:05

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:09 |
| Air Blank          | 0.000 | 00:10 |
| Control Test       | 0.082 | 00:10 |
| Air Blank          | 0.000 | 00:10 |
| Subject Sample # 1 | 0.249 | 00:11 |
| Breath Volume      | 2.730 | 00:11 |
| Air Blank          | 0.000 | 00:12 |
| Air Blank          | 0.000 | 00:14 |
| Subject Sample # 2 | 0.254 | 00:14 |
| Breath Volume      | 2.707 | 00:14 |
| Air Blank          | 0.000 | 00:15 |
| Control Test       | 0.078 | 00:15 |
| Air Blank          | 0.000 | 00:15 |
| Diagnostics Check  | OK    | 00:16 |

---

Data Download Date/Time: 05/27/2024 16:32

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WAITES,DEJAS,A  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: BENNING,CHEYENNE  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 04/25/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08223080A4  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 21:42  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:27

SUBJECT'S LAST NAME: PEREZ SUAREZ

DATE/TIME OF TEST: 05/18/2024 22:57:29

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:50 |
| Air Blank          | 0.000 | 22:50 |
| Control Test       | 0.082 | 22:51 |
| Air Blank          | 0.000 | 22:51 |
| Subject Sample # 1 | 0.242 | 22:52 |
| Breath Volume      | 2.230 | 22:52 |
| Air Blank          | 0.000 | 22:53 |
| Air Blank          | 0.000 | 22:55 |
| Subject Sample # 2 | 0.228 | 22:55 |
| Breath Volume      | 1.988 | 22:55 |
| Air Blank          | 0.000 | 22:56 |
| Control Test       | 0.081 | 22:56 |
| Air Blank          | 0.000 | 22:57 |
| Diagnostics Check  | OK    | 22:57 |

---

Data Download Date/Time: 05/27/2024 16:32

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SWIERKOWSKI,CHRISTIAN,T  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: ETIENNE,EMMANUEL  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 04/25/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08223080A4  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 00:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: PEREZ LARA

DATE/TIME OF TEST: 05/22/2024 01:19:31

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:12 |
| Air Blank          | 0.000 | 01:13 |
| Control Test       | 0.082 | 01:13 |
| Air Blank          | 0.000 | 01:14 |
| Subject Sample # 1 | 0.139 | 01:14 |
| Breath Volume      | 1.734 | 01:14 |
| Air Blank          | 0.000 | 01:15 |
| Air Blank          | 0.000 | 01:17 |
| Subject Sample # 2 | 0.134 | 01:17 |
| Breath Volume      | 2.343 | 01:17 |
| Air Blank          | 0.000 | 01:18 |
| Control Test       | 0.080 | 01:18 |
| Air Blank          | 0.000 | 01:19 |
| Diagnostics Check  | OK    | 01:19 |

---

Data Download Date/Time: 05/27/2024 16:32

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LIMB,HUNTER,  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: LIMB,HUNTER  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 04/25/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08223080A4  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 23:34  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:06

SUBJECT'S LAST NAME: GOMEZ AGUILAR

DATE/TIME OF TEST: 05/26/2024 00:32:35

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:26 |
| Air Blank          | 0.000 | 00:26 |
| Control Test       | 0.081 | 00:26 |
| Air Blank          | 0.000 | 00:27 |
| Subject Sample # 1 | 0.151 | 00:28 |
| Breath Volume      | 1.917 | 00:28 |
| Air Blank          | 0.000 | 00:28 |
| Air Blank          | 0.000 | 00:30 |
| Subject Sample # 2 | 0.143 | 00:31 |
| Breath Volume      | 3.371 | 00:31 |
| Air Blank          | 0.000 | 00:31 |
| Control Test       | 0.081 | 00:31 |
| Air Blank          | 0.000 | 00:32 |
| Diagnostics Check  | OK    | 00:32 |

---

Data Download Date/Time: 05/27/2024 16:32

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GRAZIANO,STEVEN,L  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: ETIENNE,EMMANUEL  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 04/25/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08223080A4  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 23:10  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 05/27/2024 00:08:02

---

|                   |           |       |                    |
|-------------------|-----------|-------|--------------------|
| Diagnostics Check | OK        | 23:53 |                    |
| Air Blank         | 0.000     | 23:54 |                    |
| Control Test      | 0.081     | 23:54 |                    |
| Air Blank         | 0.000     | 23:55 |                    |
| Subject Sample    | # 3 0.105 | 23:55 |                    |
| Breath Volume     | 1.898     | 23:55 |                    |
| Air Blank         | 0.000     | 23:56 |                    |
| Air Blank         | 0.000     | 23:58 |                    |
| Subject Sample    | # 1       | 00:01 | No Sample Provided |
| Breath Volume     | 0.000     | 00:01 |                    |
| Air Blank         | 0.000     | 00:01 |                    |
| Air Blank         | 0.000     | 00:03 |                    |
| Subject Sample    | # 2 0.097 | 00:06 |                    |
| Breath Volume     | 1.785     | 00:06 |                    |
| Air Blank         | 0.000     | 00:07 |                    |
| Control Test      | 0.081     | 00:07 |                    |
| Air Blank         | 0.000     | 00:07 |                    |
| Diagnostics Check | OK        | 00:08 |                    |

---

Data Download Date/Time: 05/27/2024 16:32

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CAGLIONI,DAVID,A  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: WOOD,KIM  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 05/27/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08223080A4  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 18:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:35

SUBJECT'S LAST NAME: LEE

DATE/TIME OF TEST: 05/27/2024 20:01:43

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 19:55 |
| Air Blank          | 0.000 | 19:55 |
| Control Test       | 0.081 | 19:56 |
| Air Blank          | 0.000 | 19:56 |
| Subject Sample # 1 | 0.208 | 19:57 |
| Breath Volume      | 1.707 | 19:57 |
| Air Blank          | 0.000 | 19:58 |
| Air Blank          | 0.000 | 19:59 |
| Subject Sample # 2 | 0.191 | 20:00 |
| Breath Volume      | 2.046 | 20:00 |
| Air Blank          | 0.000 | 20:00 |
| Control Test       | 0.080 | 20:01 |
| Air Blank          | 0.000 | 20:01 |
| Diagnostics Check  | OK    | 20:01 |

---

Data Download Date/Time: 06/25/2024 15:45

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JONES,CORTLAND,T  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: JONES,CORTLAND  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 05/27/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08223080A4  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 20:51  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:16

SUBJECT'S LAST NAME: ORGILL

DATE/TIME OF TEST: 05/27/2024 21:42:59

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 21:36 |
| Air Blank          | 0.000 | 21:37 |
| Control Test       | 0.081 | 21:37 |
| Air Blank          | 0.000 | 21:37 |
| Subject Sample # 1 | 0.168 | 21:38 |
| Breath Volume      | 1.628 | 21:38 |
| Air Blank          | 0.000 | 21:39 |
| Air Blank          | 0.000 | 21:41 |
| Subject Sample # 2 | 0.176 | 21:41 |
| Breath Volume      | 1.332 | 21:41 |
| Air Blank          | 0.000 | 21:42 |
| Control Test       | 0.080 | 21:42 |
| Air Blank          | 0.000 | 21:42 |
| Diagnostics Check  | OK    | 21:42 |

---

Data Download Date/Time: 06/25/2024 15:45

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CAGLIONI,DAVID,A  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: WOOD,KIMBERLY  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 05/27/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08223080A4  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 11:10  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 11:30

SUBJECT'S LAST NAME: SPENCE

DATE/TIME OF TEST: 05/28/2024 11:54:38

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 11:50 |                      |
| Air Blank         | 0.000 | 11:50 |                      |
| Control Test      | 0.081 | 11:51 |                      |
| Air Blank         | 0.000 | 11:51 |                      |
| Subject Sample    | # 1   | 11:53 | Subject Test Refused |
| Breath Volume     | 0.000 | 11:53 |                      |
| Air Blank         | 0.000 | 11:53 |                      |
| Control Test      | 0.081 | 11:54 |                      |
| Air Blank         | 0.000 | 11:54 |                      |
| Diagnostics Check | OK    | 11:54 |                      |

---

Data Download Date/Time: 06/25/2024 15:45

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KUBITSCHEK,LEONARD,H  
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD  
ARRESTING OFFICER: KUBITSCHEK,LEONARD  
ARRESTING OFFICER AGENCY: Port St. Lucie PD  
LAST AGENCY INSPECTION: 05/27/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08223080A4  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 14:33  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 15:00

SUBJECT'S LAST NAME: HUYNH

DATE/TIME OF TEST: 05/28/2024 15:57:32

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 15:50 |
| Air Blank          | 0.000 | 15:50 |
| Control Test       | 0.081 | 15:51 |
| Air Blank          | 0.000 | 15:51 |
| Subject Sample # 1 | 0.220 | 15:52 |
| Breath Volume      | 1.886 | 15:52 |
| Air Blank          | 0.000 | 15:52 |
| Air Blank          | 0.000 | 15:54 |
| Subject Sample # 2 | 0.206 | 15:55 |
| Breath Volume      | 2.000 | 15:55 |
| Air Blank          | 0.000 | 15:56 |
| Control Test       | 0.079 | 15:56 |
| Air Blank          | 0.000 | 15:57 |
| Diagnostics Check  | OK    | 15:57 |

---

Data Download Date/Time: 06/25/2024 15:45

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ALBA,CHRISTOPHER,  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: MEMON,MATTHEW  
ARRESTING OFFICER AGENCY: Fort Pierce PD  
LAST AGENCY INSPECTION: 05/27/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08223080A4  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 23:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:43

SUBJECT'S LAST NAME: DURDEN

DATE/TIME OF TEST: 05/31/2024 00:24:12

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:15 |
| Air Blank          | 0.000 | 00:16 |
| Control Test       | 0.081 | 00:16 |
| Air Blank          | 0.000 | 00:17 |
| Subject Sample # 1 | 0.245 | 00:19 |
| Breath Volume      | 1.953 | 00:19 |
| Air Blank          | 0.000 | 00:20 |
| Air Blank          | 0.000 | 00:21 |
| Subject Sample # 2 | 0.238 | 00:22 |
| Breath Volume      | 1.468 | 00:22 |
| Air Blank          | 0.000 | 00:23 |
| Control Test       | 0.079 | 00:23 |
| Air Blank          | 0.000 | 00:24 |
| Diagnostics Check  | OK    | 00:24 |

---

Data Download Date/Time: 06/25/2024 15:46

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BROCATO,SHAWN,V  
BREATH TEST OPERATOR AGENCY: Fort Pierce PD  
ARRESTING OFFICER: PARRETT,AUSTIN  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 05/27/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08223080A4  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 02:08  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: YOUSSEF

DATE/TIME OF TEST: 05/31/2024 03:06:56

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:00 |
| Air Blank          | 0.000 | 03:00 |
| Control Test       | 0.081 | 03:01 |
| Air Blank          | 0.000 | 03:01 |
| Subject Sample # 1 | 0.227 | 03:02 |
| Breath Volume      | 1.460 | 03:02 |
| Air Blank          | 0.000 | 03:02 |
| Air Blank          | 0.000 | 03:04 |
| Subject Sample # 2 | 0.237 | 03:05 |
| Breath Volume      | 1.921 | 03:05 |
| Air Blank          | 0.000 | 03:06 |
| Control Test       | 0.081 | 03:06 |
| Air Blank          | 0.000 | 03:06 |
| Diagnostics Check  | OK    | 03:06 |

---

Data Download Date/Time: 06/25/2024 15:46

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000795

INSTRUMENT REGISTERED TO: GULF COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SWIFT,MICHAEL,S  
BREATH TEST OPERATOR AGENCY: Gulf County SO  
ARRESTING OFFICER: RANDIG,MATTHEW  
ARRESTING OFFICER AGENCY: Gulf County SO  
LAST AGENCY INSPECTION: 05/15/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402607186  
DRY GAS EXPIRY DATE: 12/02/2025  
VIOLATION CODE: DUI  
ARREST TIME: 00:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:26

SUBJECT'S LAST NAME: COURSEY

DATE/TIME OF TEST: 05/20/2024 01:56:10

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:48 |
| Air Blank          | 0.000 | 01:49 |
| Control Test       | 0.080 | 01:49 |
| Air Blank          | 0.000 | 01:49 |
| Subject Sample # 1 | 0.322 | 01:51 |
| Breath Volume      | 2.699 | 01:51 |
| Air Blank          | 0.000 | 01:52 |
| Air Blank          | 0.000 | 01:54 |
| Subject Sample # 2 | 0.320 | 01:54 |
| Breath Volume      | 1.660 | 01:54 |
| Air Blank          | 0.000 | 01:55 |
| Control Test       | 0.080 | 01:55 |
| Air Blank          | 0.000 | 01:56 |
| Diagnostics Check  | OK    | 01:56 |

---

Data Download Date/Time: 06/14/2024 16:22

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000795

INSTRUMENT REGISTERED TO: GULF COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SWIFT,MICHAEL,S  
BREATH TEST OPERATOR AGENCY: Gulf County SO  
ARRESTING OFFICER: SWIFT,MICHAEL  
ARRESTING OFFICER AGENCY: Gulf County SO  
LAST AGENCY INSPECTION: 05/15/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402607186  
DRY GAS EXPIRY DATE: 12/02/2025  
VIOLATION CODE: DUI  
ARREST TIME: 01:24  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:14

SUBJECT'S LAST NAME: CHAMBLISS

DATE/TIME OF TEST: 05/24/2024 02:45:17

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:35 |
| Air Blank          | 0.000 | 02:35 |
| Control Test       | 0.079 | 02:35 |
| Air Blank          | 0.000 | 02:36 |
| Subject Sample # 1 | 0.181 | 02:39 |
| Breath Volume      | 1.265 | 02:39 |
| Air Blank          | 0.000 | 02:39 |
| Air Blank          | 0.000 | 02:41 |
| Subject Sample # 2 | 0.188 | 02:43 |
| Breath Volume      | 2.062 | 02:43 |
| Air Blank          | 0.000 | 02:44 |
| Control Test       | 0.080 | 02:44 |
| Air Blank          | 0.000 | 02:45 |
| Diagnostics Check  | OK    | 02:45 |

---

Data Download Date/Time: 06/14/2024 16:22

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GREEN,OMARIA  
BREATH TEST OPERATOR AGENCY: Haines City PD  
ARRESTING OFFICER: GREEN,OMARI  
ARRESTING OFFICER AGENCY: Haines City PD  
LAST AGENCY INSPECTION: 05/17/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 10912080A1  
DRY GAS EXPIRY DATE: 06/01/2014  
VIOLATION CODE: DUI  
ARREST TIME: 05:52  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:35

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 05/19/2024 07:02:05

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 06:55 |
| Air Blank          | 0.000 | 06:56 |
| Control Test       | 0.082 | 06:56 |
| Air Blank          | 0.000 | 06:57 |
| Subject Sample # 1 | 0.091 | 06:57 |
| Breath Volume      | 1.894 | 06:57 |
| Air Blank          | 0.000 | 06:58 |
| Air Blank          | 0.000 | 07:00 |
| Subject Sample # 2 | 0.093 | 07:00 |
| Breath Volume      | 2.234 | 07:00 |
| Air Blank          | 0.000 | 07:01 |
| Control Test       | 0.082 | 07:01 |
| Air Blank          | 0.000 | 07:02 |
| Diagnostics Check  | OK    | 07:02 |

---

Data Download Date/Time: 06/17/2024 07:28

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TURNER,LATORIAN,T  
BREATH TEST OPERATOR AGENCY: Haines City PD  
ARRESTING OFFICER: JANIAK,MALIKHI  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 05/17/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 10912080A1  
DRY GAS EXPIRY DATE: 12/22/2025  
VIOLATION CODE: DUI  
ARREST TIME: 21:03  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:23

SUBJECT'S LAST NAME: HUGHEY

DATE/TIME OF TEST: 05/30/2024 21:55:00

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 21:48 |
| Air Blank          | 0.000 | 21:48 |
| Control Test       | 0.083 | 21:48 |
| Air Blank          | 0.000 | 21:49 |
| Subject Sample # 1 | 0.054 | 21:50 |
| Breath Volume      | 1.773 | 21:50 |
| Air Blank          | 0.000 | 21:51 |
| Air Blank          | 0.000 | 21:52 |
| Subject Sample # 2 | 0.056 | 21:53 |
| Breath Volume      | 2.031 | 21:53 |
| Air Blank          | 0.000 | 21:54 |
| Control Test       | 0.082 | 21:54 |
| Air Blank          | 0.000 | 21:54 |
| Diagnostics Check  | OK    | 21:55 |

---

Data Download Date/Time: 06/17/2024 07:28

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: YOUNG,ROBERT,A  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: MOORE,MICHA  
ARRESTING OFFICER AGENCY: Marion County SO  
LAST AGENCY INSPECTION: 04/26/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402872586  
DRY GAS EXPIRY DATE: 10/12/2026  
VIOLATION CODE: DUI  
ARREST TIME: 23:36  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:28

SUBJECT'S LAST NAME: VERNON

DATE/TIME OF TEST: 05/03/2024 23:48:35

---

|                   |    |       |            |
|-------------------|----|-------|------------|
| Diagnostics Check | OK | 23:48 |            |
| Air Blank         |    | 23:48 | RFI Detect |
| Air Blank         |    | 23:48 | RFI Detect |

---

Data Download Date/Time: 05/29/2024 07:54

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: YOUNG,ROBERT,A  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: MOORE,MICHA  
ARRESTING OFFICER AGENCY: Marion County SO  
LAST AGENCY INSPECTION: 04/26/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402872586  
DRY GAS EXPIRY DATE: 10/12/2026  
VIOLATION CODE: DUI  
ARREST TIME: 23:36  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:28

SUBJECT'S LAST NAME: VERNON

DATE/TIME OF TEST: 05/03/2024 23:51:46

---

|                   |    |       |            |
|-------------------|----|-------|------------|
| Diagnostics Check | OK | 23:51 |            |
| Air Blank         |    | 23:51 | RFI Detect |
| Air Blank         |    | 23:51 | RFI Detect |

---

Data Download Date/Time: 05/29/2024 07:54

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: YOUNG,ROBERT,A  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: LATHAM,JAMIE  
ARRESTING OFFICER AGENCY: Marion County SO  
LAST AGENCY INSPECTION: 04/26/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402872586  
DRY GAS EXPIRY DATE: 10/12/2026  
VIOLATION CODE: DUI  
ARREST TIME: 01:33  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:21

SUBJECT'S LAST NAME: ESPINAL PEREZ

DATE/TIME OF TEST: 05/04/2024 01:42:46

---

|                   |       |       |            |
|-------------------|-------|-------|------------|
| Diagnostics Check | OK    | 01:41 |            |
| Air Blank         | 0.000 | 01:41 |            |
| Control Test      | 0.081 | 01:42 |            |
| Air Blank         |       | 01:42 | RFI Detect |
| Air Blank         | 0.000 | 01:42 |            |

---

Data Download Date/Time: 05/29/2024 07:54

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: YOUNG,ROBERT,A  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: LATHAM,JAMIE  
ARRESTING OFFICER AGENCY: Marion County SO  
LAST AGENCY INSPECTION: 04/26/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402872586  
DRY GAS EXPIRY DATE: 10/12/2026  
VIOLATION CODE: DUI  
ARREST TIME: 01:33  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:21

SUBJECT'S LAST NAME: ESPINAL PEREZ

DATE/TIME OF TEST: 05/04/2024 01:46:41

---

|                   |       |       |            |
|-------------------|-------|-------|------------|
| Diagnostics Check | OK    | 01:43 |            |
| Air Blank         | 0.000 | 01:44 |            |
| Control Test      | 0.081 | 01:44 |            |
| Air Blank         | 0.000 | 01:45 |            |
| Subject Sample    | # 1   | 01:45 | RFI Detect |
| Breath Volume     | 0.000 | 01:45 |            |
| Air Blank         | 0.000 | 01:45 |            |
| Control Test      | 0.081 | 01:46 |            |
| Air Blank         | 0.000 | 01:46 |            |
| Diagnostics Check | OK    | 01:46 |            |

---

Data Download Date/Time: 05/29/2024 07:54

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: YOUNG,ROBERT,A  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: LATHAM,JAMIE  
ARRESTING OFFICER AGENCY: Marion County SO  
LAST AGENCY INSPECTION: 04/26/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402872586  
DRY GAS EXPIRY DATE: 10/12/2026  
VIOLATION CODE: DUI  
ARREST TIME: 01:33  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:21

SUBJECT'S LAST NAME: ESPINAL PEREZ

DATE/TIME OF TEST: 05/04/2024 01:48:09

---

|                   |    |       |            |
|-------------------|----|-------|------------|
| Diagnostics Check | OK | 01:47 |            |
| Air Blank         |    | 01:48 | RFI Detect |
| Air Blank         |    | 01:48 | RFI Detect |

---

Data Download Date/Time: 05/29/2024 07:54

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: YOUNG,ROBERT,A  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: YOUNG,ROBERT  
ARRESTING OFFICER AGENCY: Marion County SO  
LAST AGENCY INSPECTION: 04/26/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402872586  
DRY GAS EXPIRY DATE: 10/12/2026  
VIOLATION CODE: DUI  
ARREST TIME: 01:03  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: MACKEY

DATE/TIME OF TEST: 05/06/2024 01:45:41

---

|                   |       |       |            |
|-------------------|-------|-------|------------|
| Diagnostics Check | OK    | 01:40 |            |
| Air Blank         | 0.000 | 01:41 |            |
| Control Test      | 0.080 | 01:41 |            |
| Air Blank         | 0.000 | 01:41 |            |
| Subject Sample    | # 1   | 01:44 | RFI Detect |
| Breath Volume     | 1.734 | 01:44 |            |
| Air Blank         | 0.000 | 01:44 |            |
| Control Test      | 0.080 | 01:45 |            |
| Air Blank         | 0.000 | 01:45 |            |
| Diagnostics Check | OK    | 01:45 |            |

---

Data Download Date/Time: 05/29/2024 07:54

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: YOUNG,ROBERT,A  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: YOUNG,ROBERT  
ARRESTING OFFICER AGENCY: Marion County SO  
LAST AGENCY INSPECTION: 04/26/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402872586  
DRY GAS EXPIRY DATE: 10/12/2026  
VIOLATION CODE: DUI  
ARREST TIME: 01:03  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: MACKEY

DATE/TIME OF TEST: 05/06/2024 01:56:27

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:46 |
| Air Blank          | 0.000 | 01:47 |
| Control Test       | 0.080 | 01:47 |
| Air Blank          | 0.000 | 01:48 |
| Subject Sample # 1 | 0.201 | 01:50 |
| Breath Volume      | 1.843 | 01:50 |
| Air Blank          | 0.000 | 01:51 |
| Air Blank          | 0.000 | 01:53 |
| Subject Sample # 2 | 0.200 | 01:54 |
| Breath Volume      | 1.667 | 01:54 |
| Air Blank          | 0.000 | 01:55 |
| Control Test       | 0.080 | 01:55 |
| Air Blank          | 0.000 | 01:56 |
| Diagnostics Check  | OK    | 01:56 |

---

Data Download Date/Time: 05/29/2024 07:54

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LABRAKE,DEREK,D  
BREATH TEST OPERATOR AGENCY: Hernando County SO  
ARRESTING OFFICER: LOCKE,JOSEPH  
ARRESTING OFFICER AGENCY: Hernando County SO  
LAST AGENCY INSPECTION: 04/16/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08223080A3  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 10:51  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 11:38

SUBJECT'S LAST NAME: STACY

DATE/TIME OF TEST: 05/01/2024 12:05:42

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 11:59 |
| Air Blank          | 0.000 | 11:59 |
| Control Test       | 0.081 | 11:59 |
| Air Blank          | 0.000 | 12:00 |
| Subject Sample # 1 | 0.000 | 12:01 |
| Breath Volume      | 1.222 | 12:01 |
| Air Blank          | 0.000 | 12:01 |
| Air Blank          | 0.000 | 12:03 |
| Subject Sample # 2 | 0.000 | 12:04 |
| Breath Volume      | 1.203 | 12:04 |
| Air Blank          | 0.000 | 12:04 |
| Control Test       | 0.079 | 12:05 |
| Air Blank          | 0.000 | 12:05 |
| Diagnostics Check  | OK    | 12:05 |

---

Data Download Date/Time: 05/28/2024 21:26

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SPENCER, COLLEEN, E  
BREATH TEST OPERATOR AGENCY: Hernando County SO  
ARRESTING OFFICER: CLAYTON, AUSTIN  
ARRESTING OFFICER AGENCY: Hernando County SO  
LAST AGENCY INSPECTION: 04/16/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08223080A3  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 17:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:54

SUBJECT'S LAST NAME: CONROY

DATE/TIME OF TEST: 05/03/2024 18:21:33

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 18:14 |
| Air Blank          | 0.000 | 18:14 |
| Control Test       | 0.080 | 18:15 |
| Air Blank          | 0.000 | 18:15 |
| Subject Sample # 1 | 0.292 | 18:16 |
| Breath Volume      | 2.609 | 18:16 |
| Air Blank          | 0.000 | 18:17 |
| Air Blank          | 0.000 | 18:19 |
| Subject Sample # 2 | 0.279 | 18:19 |
| Breath Volume      | 1.898 | 18:19 |
| Air Blank          | 0.000 | 18:20 |
| Control Test       | 0.078 | 18:20 |
| Air Blank          | 0.000 | 18:21 |
| Diagnostics Check  | OK    | 18:21 |

---

Data Download Date/Time: 05/28/2024 21:26

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CARR,SHARLENE,  
BREATH TEST OPERATOR AGENCY: Hernando County SO  
ARRESTING OFFICER: 2147,TREMBLAY  
ARRESTING OFFICER AGENCY: Hernando County SO  
LAST AGENCY INSPECTION: 04/16/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08223080A3  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 21:57  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:54

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 05/03/2024 23:18:41

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 23:14 |                      |
| Air Blank         | 0.000 | 23:14 |                      |
| Control Test      | 0.079 | 23:15 |                      |
| Air Blank         | 0.000 | 23:15 |                      |
| Subject Sample    | # 1   | 23:17 | Subject Test Refused |
| Breath Volume     | 0.000 | 23:17 |                      |
| Air Blank         | 0.000 | 23:17 |                      |
| Control Test      | 0.079 | 23:18 |                      |
| Air Blank         | 0.000 | 23:18 |                      |
| Diagnostics Check | OK    | 23:18 |                      |

---

Data Download Date/Time: 05/28/2024 21:26

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CARR,SHARLENE,  
BREATH TEST OPERATOR AGENCY: Hernando County SO  
ARRESTING OFFICER: TREIBLAY,F  
ARRESTING OFFICER AGENCY: Hernando County SO  
LAST AGENCY INSPECTION: 04/16/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08223080A3  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 18:09  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 18:30

SUBJECT'S LAST NAME: AUTY

DATE/TIME OF TEST: 05/04/2024 18:53:25

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 18:50 |                      |
| Air Blank         | 0.000 | 18:50 |                      |
| Control Test      | 0.079 | 18:51 |                      |
| Air Blank         | 0.000 | 18:51 |                      |
| Subject Sample    | # 1   | 18:52 | Subject Test Refused |
| Breath Volume     | 0.000 | 18:52 |                      |
| Air Blank         | 0.000 | 18:52 |                      |
| Control Test      | 0.080 | 18:52 |                      |
| Air Blank         | 0.000 | 18:53 |                      |
| Diagnostics Check | OK    | 18:53 |                      |

---

Data Download Date/Time: 05/28/2024 21:26

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CARR,SHARLENE,  
BREATH TEST OPERATOR AGENCY: Hernando County SO  
ARRESTING OFFICER: CLAYTON,A  
ARRESTING OFFICER AGENCY: Hernando County SO  
LAST AGENCY INSPECTION: 04/16/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08223080A3  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 23:57  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: GINAN

DATE/TIME OF TEST: 05/05/2024 00:53:56

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 00:51 |                      |
| Air Blank         | 0.000 | 00:51 |                      |
| Control Test      | 0.080 | 00:51 |                      |
| Air Blank         | 0.000 | 00:52 |                      |
| Subject Sample    | # 1   | 00:52 | Subject Test Refused |
| Breath Volume     | 0.000 | 00:52 |                      |
| Air Blank         | 0.000 | 00:52 |                      |
| Control Test      | 0.080 | 00:53 |                      |
| Air Blank         | 0.000 | 00:53 |                      |
| Diagnostics Check | OK    | 00:53 |                      |

---

Data Download Date/Time: 05/28/2024 21:26

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CARR,SHARLENE,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: TREMBLAY,F  
ARRESTING OFFICER AGENCY: Hernando County SO  
LAST AGENCY INSPECTION: 04/16/2024

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 08223080A3  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 02:34  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 05/05/2024 03:32:37

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:25 |
| Air Blank          | 0.000 | 03:25 |
| Control Test       | 0.080 | 03:26 |
| Air Blank          | 0.000 | 03:26 |
| Subject Sample # 1 | 0.088 | 03:27 |
| Breath Volume      | 1.441 | 03:27 |
| Air Blank          | 0.000 | 03:27 |
| Air Blank          | 0.000 | 03:29 |
| Subject Sample # 2 | 0.091 | 03:30 |
| Breath Volume      | 1.871 | 03:30 |
| Air Blank          | 0.000 | 03:31 |
| Control Test       | 0.080 | 03:31 |
| Air Blank          | 0.000 | 03:32 |
| Diagnostics Check  | OK    | 03:32 |

---

Data Download Date/Time: 05/28/2024 21:26

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BERLYNE,R,D  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: BERLYNE,R  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 04/16/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08223080A3  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 21:07  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: KELLY

DATE/TIME OF TEST: 05/05/2024 22:23:44

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:16 |
| Air Blank          | 0.000 | 22:16 |
| Control Test       | 0.080 | 22:16 |
| Air Blank          | 0.000 | 22:17 |
| Subject Sample # 1 | 0.216 | 22:19 |
| Breath Volume      | 2.867 | 22:19 |
| Air Blank          | 0.000 | 22:19 |
| Air Blank          | 0.000 | 22:21 |
| Subject Sample # 2 | 0.221 | 22:22 |
| Breath Volume      | 1.914 | 22:22 |
| Air Blank          | 0.000 | 22:22 |
| Control Test       | 0.079 | 22:23 |
| Air Blank          | 0.000 | 22:23 |
| Diagnostics Check  | OK    | 22:23 |

---

Data Download Date/Time: 05/28/2024 21:26

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ALIAGA,JULIO,A  
BREATH TEST OPERATOR AGENCY: Hernando County SO  
ARRESTING OFFICER: ALIAGA,JULIO  
ARRESTING OFFICER AGENCY: Hernando County SO  
LAST AGENCY INSPECTION: 04/16/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08223080A3  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 23:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:04

SUBJECT'S LAST NAME: HUGHES

DATE/TIME OF TEST: 05/07/2024 00:28:13

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 00:25 |                      |
| Air Blank         | 0.000 | 00:25 |                      |
| Control Test      | 0.080 | 00:26 |                      |
| Air Blank         | 0.000 | 00:26 |                      |
| Subject Sample    | # 1   | 00:26 | Subject Test Refused |
| Breath Volume     | 0.000 | 00:26 |                      |
| Air Blank         | 0.000 | 00:27 |                      |
| Control Test      | 0.079 | 00:27 |                      |
| Air Blank         | 0.000 | 00:28 |                      |
| Diagnostics Check | OK    | 00:28 |                      |

---

Data Download Date/Time: 05/28/2024 21:26

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ALIAGA,JULIO,A  
BREATH TEST OPERATOR AGENCY: Hernando County SO  
ARRESTING OFFICER: ALIAGA,JULIO  
ARRESTING OFFICER AGENCY: Hernando County SO  
LAST AGENCY INSPECTION: 04/16/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08223080A3  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 21:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:39

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 05/10/2024 22:02:51

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 21:59 |                      |
| Air Blank         | 0.000 | 22:00 |                      |
| Control Test      | 0.080 | 22:00 |                      |
| Air Blank         | 0.000 | 22:01 |                      |
| Subject Sample    | # 1   | 22:01 | Subject Test Refused |
| Breath Volume     | 0.000 | 22:01 |                      |
| Air Blank         | 0.000 | 22:01 |                      |
| Control Test      | 0.079 | 22:02 |                      |
| Air Blank         | 0.000 | 22:02 |                      |
| Diagnostics Check | OK    | 22:02 |                      |

---

Data Download Date/Time: 05/28/2024 21:26

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LABRAKE,DEREK,D  
BREATH TEST OPERATOR AGENCY: Hernando County SO  
ARRESTING OFFICER: KELLY,J  
ARRESTING OFFICER AGENCY: Hernando County SO  
LAST AGENCY INSPECTION: 04/16/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08223080A3  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 12:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 13:20

SUBJECT'S LAST NAME: MARIANI

DATE/TIME OF TEST: 05/11/2024 13:52:22

---

|                   |       |       |                    |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK    | 13:40 |                    |
| Air Blank         | 0.000 | 13:41 |                    |
| Control Test      | 0.081 | 13:41 |                    |
| Air Blank         | 0.000 | 13:42 |                    |
| Subject Sample    | # 1   | 13:45 | No Sample Provided |
| Breath Volume     | 0.000 | 13:45 |                    |
| Air Blank         | 0.000 | 13:45 |                    |
| Air Blank         | 0.000 | 13:47 |                    |
| Subject Sample    | # 2   | 13:50 | No Sample Provided |
| Breath Volume     | 0.000 | 13:50 |                    |
| Air Blank         | 0.000 | 13:51 |                    |
| Control Test      | 0.080 | 13:51 |                    |
| Air Blank         | 0.000 | 13:52 |                    |
| Diagnostics Check | OK    | 13:52 |                    |

---

Data Download Date/Time: 05/28/2024 21:26

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BERLYNE,R,D  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: BERLYNE,R  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 04/16/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08223080A3  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 22:38  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:07

SUBJECT'S LAST NAME: MCKINLEY

DATE/TIME OF TEST: 05/13/2024 23:34:11

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:27 |
| Air Blank          | 0.000 | 23:27 |
| Control Test       | 0.081 | 23:28 |
| Air Blank          | 0.000 | 23:28 |
| Subject Sample # 1 | 0.117 | 23:29 |
| Breath Volume      | 2.246 | 23:29 |
| Air Blank          | 0.000 | 23:30 |
| Air Blank          | 0.000 | 23:31 |
| Subject Sample # 2 | 0.112 | 23:32 |
| Breath Volume      | 1.855 | 23:32 |
| Air Blank          | 0.000 | 23:33 |
| Control Test       | 0.079 | 23:33 |
| Air Blank          | 0.000 | 23:34 |
| Diagnostics Check  | OK    | 23:34 |

---

Data Download Date/Time: 05/28/2024 21:26

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BERLYNE,R,D

BREATH TEST OPERATOR AGENCY: FHP Troop C

ARRESTING OFFICER: BERLYNE,R

ARRESTING OFFICER AGENCY: FHP Troop C

LAST AGENCY INSPECTION: 04/16/2024

Date or Time Changed: No

DRY GAS STANDARD LOT #: 08223080A3

DRY GAS EXPIRY DATE: 06/05/2025

VIOLATION CODE: DUI

ARREST TIME: 12:19

GENDER: Female

OBSERVATION PERIOD BEGAN: 12:46

SUBJECT'S LAST NAME: FUNK

DATE/TIME OF TEST: 05/14/2024 13:18:08

---

|                   |           |       |                |
|-------------------|-----------|-------|----------------|
| Diagnostics Check | OK        | 13:06 |                |
| Air Blank         | 0.000     | 13:06 |                |
| Control Test      | 0.080     | 13:07 |                |
| Air Blank         | 0.000     | 13:07 |                |
| Subject Sample    | # 1 0.086 | 13:10 | Volume Not Met |
| Breath Volume     | 0.906     | 13:10 |                |
| Air Blank         | 0.000     | 13:11 |                |
| Air Blank         | 0.000     | 13:13 |                |
| Subject Sample    | # 2 0.087 | 13:16 | Volume Not Met |
| Breath Volume     | 0.652     | 13:16 |                |
| Air Blank         | 0.000     | 13:17 |                |
| Control Test      | 0.080     | 13:17 |                |
| Air Blank         | 0.000     | 13:18 |                |
| Diagnostics Check | OK        | 13:18 |                |

---

Data Download Date/Time: 05/28/2024 21:26

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CARR,SHARLENE,  
BREATH TEST OPERATOR AGENCY: Hernando County SO  
ARRESTING OFFICER: CLAYTON,A  
ARRESTING OFFICER AGENCY: Hernando County SO  
LAST AGENCY INSPECTION: 04/16/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08223080A3  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 00:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 05/20/2024 02:05:32

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 02:01 |                      |
| Air Blank         | 0.000 | 02:02 |                      |
| Control Test      | 0.080 | 02:02 |                      |
| Air Blank         | 0.000 | 02:03 |                      |
| Subject Sample    | # 1   | 02:04 | Subject Test Refused |
| Breath Volume     | 0.000 | 02:04 |                      |
| Air Blank         | 0.000 | 02:04 |                      |
| Control Test      | 0.080 | 02:04 |                      |
| Air Blank         | 0.000 | 02:05 |                      |
| Diagnostics Check | OK    | 02:05 |                      |

---

Data Download Date/Time: 05/28/2024 21:27

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WEST,GHRISTOPHER,S  
BREATH TEST OPERATOR AGENCY: Hernando County SO  
ARRESTING OFFICER: CLAYTON,AUSTIN  
ARRESTING OFFICER AGENCY: Hernando County SO  
LAST AGENCY INSPECTION: 04/16/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08223080A3  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 16:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:25

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 05/22/2024 17:51:55

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 17:45 |
| Air Blank          | 0.000 | 17:46 |
| Control Test       | 0.080 | 17:46 |
| Air Blank          | 0.000 | 17:47 |
| Subject Sample # 1 | 0.000 | 17:47 |
| Breath Volume      | 2.218 | 17:47 |
| Air Blank          | 0.000 | 17:48 |
| Air Blank          | 0.000 | 17:50 |
| Subject Sample # 2 | 0.000 | 17:50 |
| Breath Volume      | 2.050 | 17:50 |
| Air Blank          | 0.000 | 17:50 |
| Control Test       | 0.080 | 17:51 |
| Air Blank          | 0.000 | 17:51 |
| Diagnostics Check  | OK    | 17:51 |

---

Data Download Date/Time: 05/28/2024 21:27

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JOHNSTON,STACY,M  
BREATH TEST OPERATOR AGENCY: Hernando County SO  
ARRESTING OFFICER: CLAYTON,A  
ARRESTING OFFICER AGENCY: Hernando County SO  
LAST AGENCY INSPECTION: 04/16/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08223080A3  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 00:46  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: DOLAN

DATE/TIME OF TEST: 05/24/2024 01:31:53

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 01:28 |                      |
| Air Blank         | 0.000 | 01:29 |                      |
| Control Test      | 0.079 | 01:29 |                      |
| Air Blank         | 0.000 | 01:30 |                      |
| Subject Sample    | # 1   | 01:30 | Subject Test Refused |
| Breath Volume     | 0.000 | 01:30 |                      |
| Air Blank         | 0.000 | 01:30 |                      |
| Control Test      | 0.080 | 01:31 |                      |
| Air Blank         | 0.000 | 01:31 |                      |
| Diagnostics Check | OK    | 01:31 |                      |

---

Data Download Date/Time: 05/28/2024 21:27

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ALIAGA,JULIO,A  
BREATH TEST OPERATOR AGENCY: Hernando County SO  
ARRESTING OFFICER: ALIAGA,JULIO  
ARRESTING OFFICER AGENCY: Hernando County SO  
LAST AGENCY INSPECTION: 04/16/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08223080A3  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 22:29  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:59

SUBJECT'S LAST NAME: DANVERS

DATE/TIME OF TEST: 05/24/2024 23:27:50

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:19 |
| Air Blank          | 0.000 | 23:20 |
| Control Test       | 0.079 | 23:20 |
| Air Blank          | 0.000 | 23:21 |
| Subject Sample # 1 | 0.256 | 23:23 |
| Breath Volume      | 2.062 | 23:23 |
| Air Blank          | 0.000 | 23:24 |
| Air Blank          | 0.000 | 23:25 |
| Subject Sample # 2 | 0.237 | 23:26 |
| Breath Volume      | 1.582 | 23:26 |
| Air Blank          | 0.000 | 23:26 |
| Control Test       | 0.076 | 23:27 |
| Air Blank          | 0.000 | 23:27 |
| Diagnostics Check  | OK    | 23:27 |

---

Data Download Date/Time: 05/28/2024 21:27

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RODRIGUEZ,MARCUS,J  
BREATH TEST OPERATOR AGENCY: Hernando County SO  
ARRESTING OFFICER: DAVILA,A  
ARRESTING OFFICER AGENCY: Hernando County SO  
LAST AGENCY INSPECTION: 04/16/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08223080A3  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 19:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:46

SUBJECT'S LAST NAME: MORALES

DATE/TIME OF TEST: 05/25/2024 20:13:33

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 20:07 |
| Air Blank          | 0.000 | 20:07 |
| Control Test       | 0.080 | 20:07 |
| Air Blank          | 0.000 | 20:08 |
| Subject Sample # 1 | 0.000 | 20:09 |
| Breath Volume      | 1.945 | 20:09 |
| Air Blank          | 0.000 | 20:09 |
| Air Blank          | 0.000 | 20:11 |
| Subject Sample # 2 | 0.000 | 20:12 |
| Breath Volume      | 2.328 | 20:12 |
| Air Blank          | 0.000 | 20:12 |
| Control Test       | 0.081 | 20:12 |
| Air Blank          | 0.000 | 20:13 |
| Diagnostics Check  | OK    | 20:13 |

---

Data Download Date/Time: 05/28/2024 21:27

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RODRIGUEZ,MARCUS,J  
BREATH TEST OPERATOR AGENCY: Hernando County SO  
ARRESTING OFFICER: CLAYTON,A  
ARRESTING OFFICER AGENCY: Hernando County SO  
LAST AGENCY INSPECTION: 04/16/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08223080A3  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 03:04  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: SAVELL

DATE/TIME OF TEST: 05/26/2024 04:10:23

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:03 |
| Air Blank          | 0.000 | 04:04 |
| Control Test       | 0.080 | 04:04 |
| Air Blank          | 0.000 | 04:04 |
| Subject Sample # 1 | 0.125 | 04:05 |
| Breath Volume      | 2.496 | 04:05 |
| Air Blank          | 0.000 | 04:06 |
| Air Blank          | 0.000 | 04:08 |
| Subject Sample # 2 | 0.130 | 04:08 |
| Breath Volume      | 3.097 | 04:08 |
| Air Blank          | 0.000 | 04:09 |
| Control Test       | 0.080 | 04:09 |
| Air Blank          | 0.000 | 04:10 |
| Diagnostics Check  | OK    | 04:10 |

---

Data Download Date/Time: 05/28/2024 21:27

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WEST,CHRISTOPHER,S  
BREATH TEST OPERATOR AGENCY: Hernando County SO  
ARRESTING OFFICER: TREBLEY,F  
ARRESTING OFFICER AGENCY: Hernando County SO  
LAST AGENCY INSPECTION: 04/16/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08223080A3  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 19:44  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: YOUNG

DATE/TIME OF TEST: 05/27/2024 21:03:59

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 21:01 |                      |
| Air Blank         | 0.000 | 21:01 |                      |
| Control Test      | 0.079 | 21:01 |                      |
| Air Blank         | 0.000 | 21:02 |                      |
| Subject Sample    | # 1   | 21:02 | Subject Test Refused |
| Breath Volume     | 0.000 | 21:02 |                      |
| Air Blank         | 0.000 | 21:03 |                      |
| Control Test      | 0.078 | 21:03 |                      |
| Air Blank         | 0.000 | 21:03 |                      |
| Diagnostics Check | OK    | 21:03 |                      |

---

Data Download Date/Time: 05/28/2024 21:27

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CARR,SHARLENE,  
BREATH TEST OPERATOR AGENCY: Hernando County SO  
ARRESTING OFFICER: CLAYTON,A  
ARRESTING OFFICER AGENCY: Hernando County SO  
LAST AGENCY INSPECTION: 05/28/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08223080A3  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 00:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: REUEL

DATE/TIME OF TEST: 05/29/2024 01:31:11

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:20 |
| Air Blank          | 0.000 | 01:20 |
| Control Test       | 0.079 | 01:21 |
| Air Blank          | 0.000 | 01:21 |
| Subject Sample # 1 | 0.169 | 01:23 |
| Breath Volume      | 1.320 | 01:23 |
| Air Blank          | 0.000 | 01:24 |
| Air Blank          | 0.000 | 01:26 |
| Subject Sample # 2 | 0.164 | 01:29 |
| Breath Volume      | 2.171 | 01:29 |
| Air Blank          | 0.000 | 01:30 |
| Control Test       | 0.079 | 01:30 |
| Air Blank          | 0.000 | 01:31 |
| Diagnostics Check  | OK    | 01:31 |

---

Data Download Date/Time: 06/25/2024 22:02

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RODRIGUEZ,FRANCESCA,M  
BREATH TEST OPERATOR AGENCY: Hernando County SO  
ARRESTING OFFICER: DAVILA,A  
ARRESTING OFFICER AGENCY: Hernando County SO  
LAST AGENCY INSPECTION: 05/28/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08223080A3  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 20:10  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: ZECCHIN

DATE/TIME OF TEST: 05/30/2024 21:27:01

|                   |           |       |                           |
|-------------------|-----------|-------|---------------------------|
| Diagnostics Check | OK        | 21:11 |                           |
| Air Blank         | 0.000     | 21:11 |                           |
| Control Test      | 0.080     | 21:11 |                           |
| Air Blank         | 0.000     | 21:12 |                           |
| Subject Sample    | # 1 0.211 | 21:14 |                           |
| Breath Volume     | 1.300     | 21:14 |                           |
| Air Blank         | 0.000     | 21:15 |                           |
| Air Blank         | 0.000     | 21:16 |                           |
| Subject Sample    | # 2 0.166 | 21:19 | Volume Not Met            |
| Breath Volume     | 0.742     | 21:19 |                           |
| Air Blank         | 0.000     | 21:21 |                           |
| Air Blank         | 0.000     | 21:22 |                           |
| Subject Sample    | # 3 0.202 | 21:25 |                           |
| Breath Volume     | 1.394     | 21:25 |                           |
| Air Blank         | 0.000     | 21:26 |                           |
| Control Test      | 0.074     | 21:26 | Control Outside Tolerance |
| Air Blank         | 0.000     | 21:27 |                           |

Data Download Date/Time: 06/25/2024 22:02

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ALIAGA,JULIO,A  
BREATH TEST OPERATOR AGENCY: Hernando County SO  
ARRESTING OFFICER: ALIAGA,JULIO  
ARRESTING OFFICER AGENCY: Hernando County SO  
LAST AGENCY INSPECTION: 05/28/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08223080A3  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 00:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: WATKINS

DATE/TIME OF TEST: 05/31/2024 01:26:23

---

|                   |       |       |            |
|-------------------|-------|-------|------------|
| Diagnostics Check | OK    | 01:25 |            |
| Air Blank         |       | 01:25 | RFI Detect |
| Air Blank         | 0.000 | 01:26 |            |

---

Data Download Date/Time: 06/25/2024 22:02

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ALIAGA,JULIO,A  
BREATH TEST OPERATOR AGENCY: Hernando County SO  
ARRESTING OFFICER: ALIAGA,JULIO  
ARRESTING OFFICER AGENCY: Hernando County SO  
LAST AGENCY INSPECTION: 05/28/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08223080A3  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 00:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: WATKINS

DATE/TIME OF TEST: 05/31/2024 01:35:02

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:27 |
| Air Blank          | 0.000 | 01:28 |
| Control Test       | 0.080 | 01:28 |
| Air Blank          | 0.000 | 01:29 |
| Subject Sample # 1 | 0.186 | 01:30 |
| Breath Volume      | 2.773 | 01:30 |
| Air Blank          | 0.000 | 01:31 |
| Air Blank          | 0.000 | 01:32 |
| Subject Sample # 2 | 0.169 | 01:33 |
| Breath Volume      | 1.972 | 01:33 |
| Air Blank          | 0.000 | 01:34 |
| Control Test       | 0.078 | 01:34 |
| Air Blank          | 0.000 | 01:34 |
| Diagnostics Check  | OK    | 01:35 |

---

Data Download Date/Time: 06/25/2024 22:02

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: REED,ERNEST,L  
BREATH TEST OPERATOR AGENCY: Sumter County SO  
ARRESTING OFFICER: GOEHRING,JONATHAN  
ARRESTING OFFICER AGENCY: Sumter County SO  
LAST AGENCY INSPECTION: 05/31/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 221303  
DRY GAS EXPIRY DATE: 01/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 15:57  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:43

SUBJECT'S LAST NAME: BOLEY

DATE/TIME OF TEST: 05/31/2024 17:11:09

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 17:04 |
| Air Blank          | 0.000 | 17:04 |
| Control Test       | 0.080 | 17:04 |
| Air Blank          | 0.000 | 17:05 |
| Subject Sample # 1 | 0.000 | 17:06 |
| Breath Volume      | 2.210 | 17:06 |
| Air Blank          | 0.000 | 17:06 |
| Air Blank          | 0.000 | 17:08 |
| Subject Sample # 2 | 0.000 | 17:09 |
| Breath Volume      | 1.484 | 17:09 |
| Air Blank          | 0.000 | 17:10 |
| Control Test       | 0.079 | 17:10 |
| Air Blank          | 0.000 | 17:11 |
| Diagnostics Check  | OK    | 17:11 |

---

Data Download Date/Time: 06/30/2024 22:38

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: STRICKLAND,KAVEN,L  
BREATH TEST OPERATOR AGENCY: Sumter County SO  
ARRESTING OFFICER: HRADECKY,C  
ARRESTING OFFICER AGENCY: Sumter County SO  
LAST AGENCY INSPECTION: 05/31/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 221303  
DRY GAS EXPIRY DATE: 01/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 19:42  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:51

SUBJECT'S LAST NAME: STRICKLAND

DATE/TIME OF TEST: 05/31/2024 22:18:23

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:11 |
| Air Blank          | 0.000 | 22:11 |
| Control Test       | 0.081 | 22:12 |
| Air Blank          | 0.000 | 22:12 |
| Subject Sample # 1 | 0.051 | 22:13 |
| Breath Volume      | 1.531 | 22:13 |
| Air Blank          | 0.000 | 22:14 |
| Air Blank          | 0.000 | 22:16 |
| Subject Sample # 2 | 0.047 | 22:16 |
| Breath Volume      | 1.437 | 22:16 |
| Air Blank          | 0.000 | 22:17 |
| Control Test       | 0.080 | 22:17 |
| Air Blank          | 0.000 | 22:18 |
| Diagnostics Check  | OK    | 22:18 |

---

Data Download Date/Time: 06/30/2024 22:38

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MOORE,MARCUS,A  
BREATH TEST OPERATOR AGENCY: Sumter County SO  
ARRESTING OFFICER: GOEHRING,J  
ARRESTING OFFICER AGENCY: Sumter County SO  
LAST AGENCY INSPECTION: 04/29/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 221303  
DRY GAS EXPIRY DATE: 01/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 01:38  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:34

SUBJECT'S LAST NAME: SEIBERT

DATE/TIME OF TEST: 05/04/2024 02:58:10

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 02:55 |                      |
| Air Blank         | 0.000 | 02:55 |                      |
| Control Test      | 0.079 | 02:56 |                      |
| Air Blank         | 0.000 | 02:56 |                      |
| Subject Sample    | # 1   | 02:56 | Subject Test Refused |
| Breath Volume     | 0.000 | 02:56 |                      |
| Air Blank         | 0.000 | 02:57 |                      |
| Control Test      | 0.078 | 02:57 |                      |
| Air Blank         | 0.000 | 02:58 |                      |
| Diagnostics Check | OK    | 02:58 |                      |

---

Data Download Date/Time: 05/31/2024 14:11

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: STICKLAND,KAVEN,L  
BREATH TEST OPERATOR AGENCY: Sumter County SO  
ARRESTING OFFICER: MALESKE,R  
ARRESTING OFFICER AGENCY: Sumter County SO  
LAST AGENCY INSPECTION: 04/29/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 221303  
DRY GAS EXPIRY DATE: 01/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 23:54  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:32

SUBJECT'S LAST NAME: LAUREANO

DATE/TIME OF TEST: 05/05/2024 00:55:06

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 00:52 |                      |
| Air Blank         | 0.000 | 00:52 |                      |
| Control Test      | 0.078 | 00:53 |                      |
| Air Blank         | 0.000 | 00:53 |                      |
| Subject Sample    | # 1   | 00:53 | Subject Test Refused |
| Breath Volume     | 0.000 | 00:53 |                      |
| Air Blank         | 0.000 | 00:54 |                      |
| Control Test      | 0.078 | 00:54 |                      |
| Air Blank         | 0.000 | 00:55 |                      |
| Diagnostics Check | OK    | 00:55 |                      |

---

Data Download Date/Time: 05/31/2024 14:28

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MOORE,MARCUS,A  
BREATH TEST OPERATOR AGENCY: Sumter County SO  
ARRESTING OFFICER: GOEHRING,J  
ARRESTING OFFICER AGENCY: Sumter County SO  
LAST AGENCY INSPECTION: 04/29/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 221303  
DRY GAS EXPIRY DATE: 01/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 23:29  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:39

SUBJECT'S LAST NAME: SHOEMAKER

DATE/TIME OF TEST: 05/10/2024 01:03:05

---

|                   |       |       |            |
|-------------------|-------|-------|------------|
| Diagnostics Check | OK    | 01:00 |            |
| Air Blank         | 0.000 | 01:00 |            |
| Control Test      | 0.078 | 01:00 |            |
| Air Blank         | 0.000 | 01:01 |            |
| Subject Sample    | # 1   | 01:01 | RFI Detect |
| Breath Volume     | 0.000 | 01:01 |            |
| Air Blank         | 0.000 | 01:02 |            |
| Control Test      | 0.077 | 01:02 |            |
| Air Blank         | 0.000 | 01:03 |            |
| Diagnostics Check | OK    | 01:03 |            |

---

Data Download Date/Time: 05/31/2024 14:28

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MOORE,M,A  
BREATH TEST OPERATOR AGENCY: Sumter County SO  
ARRESTING OFFICER: GOEHRING,J  
ARRESTING OFFICER AGENCY: Sumter County SO  
LAST AGENCY INSPECTION: 04/29/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 221303  
DRY GAS EXPIRY DATE: 01/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 23:29  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:39

SUBJECT'S LAST NAME: SHOEMAKER

DATE/TIME OF TEST: 05/10/2024 01:11:43

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:04 |
| Air Blank          | 0.000 | 01:05 |
| Control Test       | 0.078 | 01:05 |
| Air Blank          | 0.000 | 01:06 |
| Subject Sample # 1 | 0.250 | 01:07 |
| Breath Volume      | 1.933 | 01:07 |
| Air Blank          | 0.000 | 01:07 |
| Air Blank          | 0.000 | 01:09 |
| Subject Sample # 2 | 0.254 | 01:10 |
| Breath Volume      | 1.675 | 01:10 |
| Air Blank          | 0.000 | 01:10 |
| Control Test       | 0.078 | 01:11 |
| Air Blank          | 0.000 | 01:11 |
| Diagnostics Check  | OK    | 01:11 |

---

Data Download Date/Time: 05/31/2024 14:28

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: STRICKLAND,KAVEN,L  
BREATH TEST OPERATOR AGENCY: Sumter County SO  
ARRESTING OFFICER: HARDECKY,C  
ARRESTING OFFICER AGENCY: Sumter County SO  
LAST AGENCY INSPECTION: 04/29/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 221303  
DRY GAS EXPIRY DATE: 01/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 20:06  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:45

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 05/13/2024 21:08:53

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 21:06 |                      |
| Air Blank         | 0.000 | 21:06 |                      |
| Control Test      | 0.078 | 21:06 |                      |
| Air Blank         | 0.000 | 21:07 |                      |
| Subject Sample    | # 1   | 21:07 | Subject Test Refused |
| Breath Volume     | 0.000 | 21:07 |                      |
| Air Blank         | 0.000 | 21:07 |                      |
| Control Test      | 0.079 | 21:08 |                      |
| Air Blank         | 0.000 | 21:08 |                      |
| Diagnostics Check | OK    | 21:08 |                      |

---

Data Download Date/Time: 05/31/2024 14:28

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MOORE,REBECCA,G  
BREATH TEST OPERATOR AGENCY: Sumter County SO  
ARRESTING OFFICER: GOEHRING,JONATHAN  
ARRESTING OFFICER AGENCY: Sumter County SO  
LAST AGENCY INSPECTION: 04/29/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 221303  
DRY GAS EXPIRY DATE: 01/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 00:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:57

SUBJECT'S LAST NAME: BERGERON

DATE/TIME OF TEST: 05/17/2024 01:24:33

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:18 |
| Air Blank          | 0.000 | 01:18 |
| Control Test       | 0.078 | 01:19 |
| Air Blank          | 0.000 | 01:19 |
| Subject Sample # 1 | 0.141 | 01:20 |
| Breath Volume      | 1.367 | 01:20 |
| Air Blank          | 0.000 | 01:20 |
| Air Blank          | 0.000 | 01:22 |
| Subject Sample # 2 | 0.125 | 01:23 |
| Breath Volume      | 1.105 | 01:23 |
| Air Blank          | 0.000 | 01:23 |
| Control Test       | 0.078 | 01:23 |
| Air Blank          | 0.000 | 01:24 |
| Diagnostics Check  | OK    | 01:24 |

---

Data Download Date/Time: 05/31/2024 14:28

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: STRICKLAND,KAVEN,L  
BREATH TEST OPERATOR AGENCY: Sumter County SO  
ARRESTING OFFICER: RAMOS,D  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 04/29/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 221303  
DRY GAS EXPIRY DATE: 01/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 23:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:26

SUBJECT'S LAST NAME: FRIAS ARAIZA

DATE/TIME OF TEST: 05/18/2024 00:54:34

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:48 |
| Air Blank          | 0.000 | 00:48 |
| Control Test       | 0.078 | 00:48 |
| Air Blank          | 0.000 | 00:49 |
| Subject Sample # 1 | 0.194 | 00:49 |
| Breath Volume      | 3.039 | 00:49 |
| Air Blank          | 0.000 | 00:50 |
| Air Blank          | 0.000 | 00:52 |
| Subject Sample # 2 | 0.184 | 00:53 |
| Breath Volume      | 2.671 | 00:53 |
| Air Blank          | 0.000 | 00:53 |
| Control Test       | 0.078 | 00:53 |
| Air Blank          | 0.000 | 00:54 |
| Diagnostics Check  | OK    | 00:54 |

---

Data Download Date/Time: 05/31/2024 14:28

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: STRICKLAND,KAVEN,L  
BREATH TEST OPERATOR AGENCY: Sumter County SO  
ARRESTING OFFICER: HOSKEY,N  
ARRESTING OFFICER AGENCY: Sumter County SO  
LAST AGENCY INSPECTION: 04/29/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 221303  
DRY GAS EXPIRY DATE: 01/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 00:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:58

SUBJECT'S LAST NAME: FLANAGAN-FISHER

DATE/TIME OF TEST: 05/19/2024 01:31:27

---

|                   |           |       |                |
|-------------------|-----------|-------|----------------|
| Diagnostics Check | OK        | 01:18 |                |
| Air Blank         | 0.000     | 01:18 |                |
| Control Test      | 0.078     | 01:19 |                |
| Air Blank         | 0.000     | 01:19 |                |
| Subject Sample    | # 1 0.134 | 01:22 | Volume Not Met |
| Breath Volume     | 0.492     | 01:22 |                |
| Air Blank         | 0.000     | 01:23 |                |
| Air Blank         | 0.000     | 01:25 |                |
| Subject Sample    | # 2 0.155 | 01:26 |                |
| Breath Volume     | 2.472     | 01:26 |                |
| Air Blank         | 0.000     | 01:26 |                |
| Air Blank         | 0.000     | 01:28 |                |
| Subject Sample    | # 3 0.141 | 01:29 |                |
| Breath Volume     | 1.457     | 01:29 |                |
| Air Blank         | 0.000     | 01:30 |                |
| Control Test      | 0.078     | 01:30 |                |
| Air Blank         | 0.000     | 01:31 |                |
| Diagnostics Check | OK        | 01:31 |                |

---

Data Download Date/Time: 05/31/2024 14:29

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GAVIN, DEMERIS, P  
BREATH TEST OPERATOR AGENCY: Sumter County SO  
ARRESTING OFFICER: COSSEBOOM, M  
ARRESTING OFFICER AGENCY: Wildwood PD  
LAST AGENCY INSPECTION: 04/29/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 221303  
DRY GAS EXPIRY DATE: 01/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 14:52  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 15:20

SUBJECT'S LAST NAME: ENGSTROM

DATE/TIME OF TEST: 05/24/2024 15:47:31

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 15:40 |
| Air Blank          | 0.000 | 15:40 |
| Control Test       | 0.079 | 15:41 |
| Air Blank          | 0.000 | 15:41 |
| Subject Sample # 1 | 0.172 | 15:42 |
| Breath Volume      | 3.210 | 15:42 |
| Air Blank          | 0.000 | 15:43 |
| Air Blank          | 0.000 | 15:45 |
| Subject Sample # 2 | 0.173 | 15:45 |
| Breath Volume      | 3.117 | 15:45 |
| Air Blank          | 0.000 | 15:46 |
| Control Test       | 0.078 | 15:46 |
| Air Blank          | 0.000 | 15:47 |
| Diagnostics Check  | OK    | 15:47 |

---

Data Download Date/Time: 05/31/2024 14:29

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ALVARADO,RODERICK,A  
BREATH TEST OPERATOR AGENCY: Sumter County SO  
ARRESTING OFFICER: GOEHRING,J  
ARRESTING OFFICER AGENCY: Sumter County SO  
LAST AGENCY INSPECTION: 04/29/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 221303  
DRY GAS EXPIRY DATE: 01/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 01:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: EDWARDS

DATE/TIME OF TEST: 05/25/2024 03:08:17

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:01 |
| Air Blank          | 0.000 | 03:01 |
| Control Test       | 0.078 | 03:01 |
| Air Blank          | 0.000 | 03:02 |
| Subject Sample # 1 | 0.177 | 03:03 |
| Breath Volume      | 1.718 | 03:03 |
| Air Blank          | 0.000 | 03:04 |
| Air Blank          | 0.000 | 03:05 |
| Subject Sample # 2 | 0.182 | 03:06 |
| Breath Volume      | 2.402 | 03:06 |
| Air Blank          | 0.000 | 03:07 |
| Control Test       | 0.078 | 03:07 |
| Air Blank          | 0.000 | 03:08 |
| Diagnostics Check  | OK    | 03:08 |

---

Data Download Date/Time: 05/31/2024 14:29

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MITCHELL,KENNETH,D  
BREATH TEST OPERATOR AGENCY: Sumter County SO  
ARRESTING OFFICER: GOHRIG,T  
ARRESTING OFFICER AGENCY: Sumter County SO  
LAST AGENCY INSPECTION: 04/29/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 221303  
DRY GAS EXPIRY DATE: 01/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 22:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: HOWARD

DATE/TIME OF TEST: 05/26/2024 00:09:02

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:02 |
| Air Blank          | 0.000 | 00:03 |
| Control Test       | 0.078 | 00:03 |
| Air Blank          | 0.000 | 00:04 |
| Subject Sample # 1 | 0.163 | 00:04 |
| Breath Volume      | 1.570 | 00:04 |
| Air Blank          | 0.000 | 00:05 |
| Air Blank          | 0.000 | 00:07 |
| Subject Sample # 2 | 0.150 | 00:07 |
| Breath Volume      | 1.140 | 00:07 |
| Air Blank          | 0.000 | 00:08 |
| Control Test       | 0.078 | 00:08 |
| Air Blank          | 0.000 | 00:08 |
| Diagnostics Check  | OK    | 00:09 |

---

Data Download Date/Time: 05/31/2024 14:29

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ALVARADO,RODERICK,A  
BREATH TEST OPERATOR AGENCY: Sumter County SO  
ARRESTING OFFICER: ST CLAIR,J  
ARRESTING OFFICER AGENCY: Sumter County SO  
LAST AGENCY INSPECTION: 04/29/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 221303  
DRY GAS EXPIRY DATE: 01/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 00:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: PEREZ SANTIZ

DATE/TIME OF TEST: 05/26/2024 01:05:54

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:58 |
| Air Blank          | 0.000 | 00:58 |
| Control Test       | 0.079 | 00:59 |
| Air Blank          | 0.000 | 00:59 |
| Subject Sample # 1 | 0.112 | 01:00 |
| Breath Volume      | 2.714 | 01:00 |
| Air Blank          | 0.000 | 01:01 |
| Air Blank          | 0.000 | 01:03 |
| Subject Sample # 2 | 0.109 | 01:04 |
| Breath Volume      | 2.914 | 01:04 |
| Air Blank          | 0.000 | 01:04 |
| Control Test       | 0.078 | 01:05 |
| Air Blank          | 0.000 | 01:05 |
| Diagnostics Check  | OK    | 01:05 |

---

Data Download Date/Time: 05/31/2024 14:29

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MOORE,MARCUS,A  
BREATH TEST OPERATOR AGENCY: Sumter County SO  
ARRESTING OFFICER: HINDERHOFER,E  
ARRESTING OFFICER AGENCY: Sumter County SO  
LAST AGENCY INSPECTION: 04/29/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 221303  
DRY GAS EXPIRY DATE: 01/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 22:13  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:42

SUBJECT'S LAST NAME: JUDD

DATE/TIME OF TEST: 05/28/2024 23:05:40

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 23:02 |                      |
| Air Blank         | 0.000 | 23:03 |                      |
| Control Test      | 0.078 | 23:03 |                      |
| Air Blank         | 0.000 | 23:04 |                      |
| Subject Sample    | # 1   | 23:04 | Subject Test Refused |
| Breath Volume     | 0.000 | 23:04 |                      |
| Air Blank         | 0.000 | 23:04 |                      |
| Control Test      | 0.078 | 23:05 |                      |
| Air Blank         | 0.000 | 23:05 |                      |
| Diagnostics Check | OK    | 23:05 |                      |

---

Data Download Date/Time: 05/31/2024 14:29

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ARNOLD,SHAWN,D  
BREATH TEST OPERATOR AGENCY: CCA Citrus  
ARRESTING OFFICER: TAYOR,T  
ARRESTING OFFICER AGENCY: Citrus County SO  
LAST AGENCY INSPECTION: 04/23/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 231043  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 07:54  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 08:05

SUBJECT'S LAST NAME: NILE

DATE/TIME OF TEST: 05/01/2024 09:03:55

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 09:00 |                      |
| Air Blank         | 0.000 | 09:01 |                      |
| Control Test      | 0.080 | 09:01 |                      |
| Air Blank         | 0.000 | 09:02 |                      |
| Subject Sample    | # 1   | 09:02 | Subject Test Refused |
| Breath Volume     | 0.000 | 09:02 |                      |
| Air Blank         | 0.000 | 09:03 |                      |
| Control Test      | 0.080 | 09:03 |                      |
| Air Blank         | 0.000 | 09:03 |                      |
| Diagnostics Check | OK    | 09:03 |                      |

---

Data Download Date/Time: 05/30/2024 12:51

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DOTSON,ANTHONY,T  
BREATH TEST OPERATOR AGENCY: CCA Citrus  
ARRESTING OFFICER: HOWARD,T  
ARRESTING OFFICER AGENCY: Citrus County SO  
LAST AGENCY INSPECTION: 04/23/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 231043  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 19:53  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:34

SUBJECT'S LAST NAME: HOWLAND

DATE/TIME OF TEST: 05/02/2024 21:06:06

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 20:58 |
| Air Blank          | 0.000 | 20:58 |
| Control Test       | 0.080 | 20:58 |
| Air Blank          | 0.000 | 20:59 |
| Subject Sample # 1 | 0.341 | 21:00 |
| Breath Volume      | 1.230 | 21:00 |
| Air Blank          | 0.000 | 21:01 |
| Air Blank          | 0.000 | 21:03 |
| Subject Sample # 2 | 0.333 | 21:04 |
| Breath Volume      | 1.386 | 21:04 |
| Air Blank          | 0.000 | 21:05 |
| Control Test       | 0.078 | 21:05 |
| Air Blank          | 0.000 | 21:06 |
| Diagnostics Check  | OK    | 21:06 |

---

Data Download Date/Time: 05/30/2024 12:51

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HATFIELD, PAIGE, S  
BREATH TEST OPERATOR AGENCY: Citrus County SO  
ARRESTING OFFICER: ROBINSON, KYLE  
ARRESTING OFFICER AGENCY: Citrus County SO  
LAST AGENCY INSPECTION: 04/23/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 231043  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 13:26  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 13:26

SUBJECT'S LAST NAME: HATFIELD

DATE/TIME OF TEST: 05/03/2024 14:48:03

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 14:39 |
| Air Blank          | 0.000 | 14:40 |
| Control Test       | 0.080 | 14:40 |
| Air Blank          | 0.000 | 14:40 |
| Subject Sample # 1 | 0.089 | 14:43 |
| Breath Volume      | 1.976 | 14:43 |
| Air Blank          | 0.000 | 14:44 |
| Air Blank          | 0.000 | 14:46 |
| Subject Sample # 2 | 0.088 | 14:46 |
| Breath Volume      | 1.828 | 14:46 |
| Air Blank          | 0.000 | 14:47 |
| Control Test       | 0.078 | 14:47 |
| Air Blank          | 0.000 | 14:47 |
| Diagnostics Check  | OK    | 14:48 |

---

Data Download Date/Time: 05/30/2024 12:51

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MOORE,LENNARD,  
BREATH TEST OPERATOR AGENCY: CCA Citrus  
ARRESTING OFFICER: PERMENTER,M  
ARRESTING OFFICER AGENCY: Citrus County SO  
LAST AGENCY INSPECTION: 04/23/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 231043  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 02:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 05/04/2024 03:38:17

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:30 |
| Air Blank          | 0.000 | 03:31 |
| Control Test       | 0.080 | 03:31 |
| Air Blank          | 0.000 | 03:32 |
| Subject Sample # 1 | 0.154 | 03:33 |
| Breath Volume      | 1.835 | 03:33 |
| Air Blank          | 0.000 | 03:33 |
| Air Blank          | 0.000 | 03:35 |
| Subject Sample # 2 | 0.169 | 03:36 |
| Breath Volume      | 4.480 | 03:36 |
| Air Blank          | 0.000 | 03:37 |
| Control Test       | 0.079 | 03:37 |
| Air Blank          | 0.000 | 03:38 |
| Diagnostics Check  | OK    | 03:38 |

---

Data Download Date/Time: 05/30/2024 12:51

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MOORE,LENNARD,  
BREATH TEST OPERATOR AGENCY: CCA Citrus  
ARRESTING OFFICER: GARRISON,T  
ARRESTING OFFICER AGENCY: Citrus County SO  
LAST AGENCY INSPECTION: 04/23/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 231043  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 17:01  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 17:37

SUBJECT'S LAST NAME: HARDEN

DATE/TIME OF TEST: 05/04/2024 18:00:28

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 17:57 |                      |
| Air Blank         | 0.000 | 17:58 |                      |
| Control Test      | 0.080 | 17:58 |                      |
| Air Blank         | 0.000 | 17:58 |                      |
| Subject Sample    | # 1   | 17:59 | Subject Test Refused |
| Breath Volume     | 0.000 | 17:59 |                      |
| Air Blank         | 0.000 | 17:59 |                      |
| Control Test      | 0.080 | 17:59 |                      |
| Air Blank         | 0.000 | 18:00 |                      |
| Diagnostics Check | OK    | 18:00 |                      |

---

Data Download Date/Time: 05/30/2024 12:51

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BROWN,JEFFERY,P  
BREATH TEST OPERATOR AGENCY: CCA Citrus  
ARRESTING OFFICER: GARRISON,T  
ARRESTING OFFICER AGENCY: Citrus County SO  
LAST AGENCY INSPECTION: 04/23/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 231043  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 21:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:33

SUBJECT'S LAST NAME: ROMAN-HAMMER

DATE/TIME OF TEST: 05/04/2024 22:07:34

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 21:59 |
| Air Blank          | 0.000 | 22:00 |
| Control Test       | 0.080 | 22:00 |
| Air Blank          | 0.000 | 22:01 |
| Subject Sample # 1 | 0.214 | 22:03 |
| Breath Volume      | 1.785 | 22:03 |
| Air Blank          | 0.000 | 22:03 |
| Air Blank          | 0.000 | 22:05 |
| Subject Sample # 2 | 0.207 | 22:06 |
| Breath Volume      | 1.308 | 22:06 |
| Air Blank          | 0.000 | 22:06 |
| Control Test       | 0.080 | 22:06 |
| Air Blank          | 0.000 | 22:07 |
| Diagnostics Check  | OK    | 22:07 |

---

Data Download Date/Time: 05/30/2024 12:51

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MOORE,LENNARD,  
BREATH TEST OPERATOR AGENCY: CCA Citrus  
ARRESTING OFFICER: PERMENTER,N  
ARRESTING OFFICER AGENCY: Citrus County SO  
LAST AGENCY INSPECTION: 04/23/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 231043  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 23:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:47

SUBJECT'S LAST NAME: BALL

DATE/TIME OF TEST: 05/06/2024 00:10:18

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 00:07 |                      |
| Air Blank         | 0.000 | 00:07 |                      |
| Control Test      | 0.080 | 00:08 |                      |
| Air Blank         | 0.000 | 00:08 |                      |
| Subject Sample    | # 1   | 00:08 | Subject Test Refused |
| Breath Volume     | 0.000 | 00:08 |                      |
| Air Blank         | 0.000 | 00:09 |                      |
| Control Test      | 0.080 | 00:09 |                      |
| Air Blank         | 0.000 | 00:10 |                      |
| Diagnostics Check | OK    | 00:10 |                      |

---

Data Download Date/Time: 05/30/2024 12:52

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DOTSON,ANTHONY,T  
BREATH TEST OPERATOR AGENCY: CCA Citrus  
ARRESTING OFFICER: D,D  
ARRESTING OFFICER AGENCY: Citrus County SO  
LAST AGENCY INSPECTION: 04/23/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 231043  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 18:30  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: SHEFFIELD

DATE/TIME OF TEST: 05/11/2024 19:31:32

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 19:28 |                      |
| Air Blank         | 0.000 | 19:28 |                      |
| Control Test      | 0.081 | 19:28 |                      |
| Air Blank         | 0.000 | 19:29 |                      |
| Subject Sample    | # 1   | 19:30 | Subject Test Refused |
| Breath Volume     | 0.000 | 19:30 |                      |
| Air Blank         | 0.000 | 19:30 |                      |
| Control Test      | 0.080 | 19:30 |                      |
| Air Blank         | 0.000 | 19:31 |                      |
| Diagnostics Check | OK    | 19:31 |                      |

Data Download Date/Time: 05/30/2024 12:52

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DOTSON,ANTHONY,T  
BREATH TEST OPERATOR AGENCY: CCA Citrus  
ARRESTING OFFICER: TAYLOR,T  
ARRESTING OFFICER AGENCY: Citrus County SO  
LAST AGENCY INSPECTION: 04/23/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 231043  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 20:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:14

SUBJECT'S LAST NAME: GLYNN

DATE/TIME OF TEST: 05/11/2024 20:41:34

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 20:34 |
| Air Blank          | 0.000 | 20:35 |
| Control Test       | 0.080 | 20:35 |
| Air Blank          | 0.000 | 20:36 |
| Subject Sample # 1 | 0.326 | 20:36 |
| Breath Volume      | 1.144 | 20:36 |
| Air Blank          | 0.000 | 20:37 |
| Air Blank          | 0.000 | 20:39 |
| Subject Sample # 2 | 0.312 | 20:39 |
| Breath Volume      | 1.335 | 20:39 |
| Air Blank          | 0.000 | 20:40 |
| Control Test       | 0.078 | 20:40 |
| Air Blank          | 0.000 | 20:41 |
| Diagnostics Check  | OK    | 20:41 |

---

Data Download Date/Time: 05/30/2024 12:52

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DOTSON,ANTHONY,T  
BREATH TEST OPERATOR AGENCY: CCA Citrus  
ARRESTING OFFICER: TAYLOR,T  
ARRESTING OFFICER AGENCY: Citrus County SO  
LAST AGENCY INSPECTION: 04/23/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 231043  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 23:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:43

SUBJECT'S LAST NAME: LACEY

DATE/TIME OF TEST: 05/12/2024 00:12:47

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:05 |
| Air Blank          | 0.000 | 00:06 |
| Control Test       | 0.080 | 00:06 |
| Air Blank          | 0.000 | 00:06 |
| Subject Sample # 1 | 0.282 | 00:07 |
| Breath Volume      | 1.222 | 00:07 |
| Air Blank          | 0.000 | 00:08 |
| Air Blank          | 0.000 | 00:10 |
| Subject Sample # 2 | 0.277 | 00:11 |
| Breath Volume      | 1.324 | 00:11 |
| Air Blank          | 0.000 | 00:11 |
| Control Test       | 0.078 | 00:12 |
| Air Blank          | 0.000 | 00:12 |
| Diagnostics Check  | OK    | 00:12 |

---

Data Download Date/Time: 05/30/2024 12:52

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DOTSON,ANTHONY,T  
BREATH TEST OPERATOR AGENCY: CCA Citrus  
ARRESTING OFFICER: GARRISON,T  
ARRESTING OFFICER AGENCY: Citrus County SO  
LAST AGENCY INSPECTION: 04/23/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 231043  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 17:20  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 17:58

SUBJECT'S LAST NAME: CLADER

DATE/TIME OF TEST: 05/21/2024 18:25:38

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 18:18 |
| Air Blank          | 0.000 | 18:19 |
| Control Test       | 0.080 | 18:19 |
| Air Blank          | 0.000 | 18:20 |
| Subject Sample # 1 | 0.000 | 18:21 |
| Breath Volume      | 1.687 | 18:21 |
| Air Blank          | 0.000 | 18:21 |
| Air Blank          | 0.000 | 18:23 |
| Subject Sample # 2 | 0.000 | 18:24 |
| Breath Volume      | 1.843 | 18:24 |
| Air Blank          | 0.000 | 18:24 |
| Control Test       | 0.081 | 18:25 |
| Air Blank          | 0.000 | 18:25 |
| Diagnostics Check  | OK    | 18:25 |

---

Data Download Date/Time: 05/30/2024 12:52

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RANSOM,AUBREY,G  
BREATH TEST OPERATOR AGENCY: FFWCC (Lake City)  
ARRESTING OFFICER: RANSOM,AUBREY  
ARRESTING OFFICER AGENCY: FFWCC (Lake City)  
LAST AGENCY INSPECTION: 04/23/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 231043  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: BUI  
ARREST TIME: 23:34  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:39

SUBJECT'S LAST NAME: TROY

DATE/TIME OF TEST: 05/25/2024 01:02:08

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 00:59 |                      |
| Air Blank         | 0.000 | 00:59 |                      |
| Control Test      | 0.081 | 01:00 |                      |
| Air Blank         | 0.000 | 01:00 |                      |
| Subject Sample    | # 1   | 01:00 | Subject Test Refused |
| Breath Volume     | 0.000 | 01:00 |                      |
| Air Blank         | 0.000 | 01:01 |                      |
| Control Test      | 0.080 | 01:01 |                      |
| Air Blank         | 0.000 | 01:02 |                      |
| Diagnostics Check | OK    | 01:02 |                      |

---

Data Download Date/Time: 05/30/2024 12:52

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROGERS,ELIZABETH,R  
BREATH TEST OPERATOR AGENCY: Citrus County SO  
ARRESTING OFFICER: JONES,BRANDON  
ARRESTING OFFICER AGENCY: Citrus County SO  
LAST AGENCY INSPECTION: 04/23/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 231043  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 21:43  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:12

SUBJECT'S LAST NAME: ROGERS

DATE/TIME OF TEST: 05/25/2024 22:38:10

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 22:35 |                      |
| Air Blank         | 0.000 | 22:35 |                      |
| Control Test      | 0.080 | 22:35 |                      |
| Air Blank         | 0.000 | 22:36 |                      |
| Subject Sample    | # 1   | 22:36 | Subject Test Refused |
| Breath Volume     | 0.000 | 22:36 |                      |
| Air Blank         | 0.000 | 22:37 |                      |
| Control Test      | 0.079 | 22:37 |                      |
| Air Blank         | 0.000 | 22:38 |                      |
| Diagnostics Check | OK    | 22:38 |                      |

Data Download Date/Time: 05/30/2024 12:52

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DOTSON,ANTHONY,T  
BREATH TEST OPERATOR AGENCY: CCA Citrus  
ARRESTING OFFICER: DEP,T  
ARRESTING OFFICER AGENCY: Citrus County SO  
LAST AGENCY INSPECTION: 05/29/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 231043  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 23:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:18

SUBJECT'S LAST NAME: DOE

DATE/TIME OF TEST: 05/29/2024 23:45:29

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 23:42 |                      |
| Air Blank         | 0.000 | 23:42 |                      |
| Control Test      | 0.079 | 23:43 |                      |
| Air Blank         | 0.000 | 23:43 |                      |
| Subject Sample    | # 1   | 23:44 | Subject Test Refused |
| Breath Volume     | 0.000 | 23:44 |                      |
| Air Blank         | 0.000 | 23:44 |                      |
| Control Test      | 0.079 | 23:44 |                      |
| Air Blank         | 0.000 | 23:45 |                      |
| Diagnostics Check | OK    | 23:45 |                      |

---

Data Download Date/Time: 05/30/2024 12:52

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: POTH,LAUREN,A  
BREATH TEST OPERATOR AGENCY: Mount Dora PD  
ARRESTING OFFICER: POTH,LAUREN  
ARRESTING OFFICER AGENCY: Mount Dora PD  
LAST AGENCY INSPECTION: 04/09/2024

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 33522080A5  
DRY GAS EXPIRY DATE: 01/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 06:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 07:03

SUBJECT'S LAST NAME: MICHELS

DATE/TIME OF TEST: 05/03/2024 07:30:36

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 07:24 |
| Air Blank          | 0.000 | 07:24 |
| Control Test       | 0.081 | 07:24 |
| Air Blank          | 0.000 | 07:25 |
| Subject Sample # 1 | 0.000 | 07:26 |
| Breath Volume      | 2.914 | 07:26 |
| Air Blank          | 0.000 | 07:26 |
| Air Blank          | 0.000 | 07:28 |
| Subject Sample # 2 | 0.000 | 07:29 |
| Breath Volume      | 2.839 | 07:29 |
| Air Blank          | 0.000 | 07:29 |
| Control Test       | 0.081 | 07:30 |
| Air Blank          | 0.000 | 07:30 |
| Diagnostics Check  | OK    | 07:30 |

---

Data Download Date/Time: 05/30/2024 10:56

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RICHETTI,KALEIGH,W  
BREATH TEST OPERATOR AGENCY: Fruitland Park PD  
ARRESTING OFFICER: RICHETTI,KALEIGH  
ARRESTING OFFICER AGENCY: Fruitland Park PD  
LAST AGENCY INSPECTION: 04/09/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33522080A5  
DRY GAS EXPIRY DATE: 01/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 22:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: COB ARANA

DATE/TIME OF TEST: 05/03/2024 23:48:40

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:42 |
| Air Blank          | 0.000 | 23:42 |
| Control Test       | 0.080 | 23:42 |
| Air Blank          | 0.000 | 23:43 |
| Subject Sample # 1 | 0.401 | 23:44 |
| Breath Volume      | 1.410 | 23:44 |
| Air Blank          | 0.000 | 23:44 |
| Air Blank          | 0.000 | 23:46 |
| Subject Sample # 2 | 0.408 | 23:47 |
| Breath Volume      | 1.804 | 23:47 |
| Air Blank          | 0.000 | 23:47 |
| Control Test       | 0.080 | 23:48 |
| Air Blank          | 0.000 | 23:48 |
| Diagnostics Check  | OK    | 23:48 |

---

Data Download Date/Time: 05/30/2024 10:56

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CHESSHER,SANDRA,W  
BREATH TEST OPERATOR AGENCY: Lake County SO  
ARRESTING OFFICER: DURYE,M  
ARRESTING OFFICER AGENCY: Lake County SO  
LAST AGENCY INSPECTION: 04/09/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33522080A5  
DRY GAS EXPIRY DATE: 01/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 02:34  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 05/04/2024 03:17:04

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:10 |
| Air Blank          | 0.000 | 03:11 |
| Control Test       | 0.081 | 03:11 |
| Air Blank          | 0.000 | 03:11 |
| Subject Sample # 1 | 0.153 | 03:12 |
| Breath Volume      | 4.468 | 03:12 |
| Air Blank          | 0.000 | 03:13 |
| Air Blank          | 0.000 | 03:14 |
| Subject Sample # 2 | 0.151 | 03:15 |
| Breath Volume      | 3.023 | 03:15 |
| Air Blank          | 0.000 | 03:16 |
| Control Test       | 0.081 | 03:16 |
| Air Blank          | 0.000 | 03:16 |
| Diagnostics Check  | OK    | 03:17 |

---

Data Download Date/Time: 05/30/2024 10:56

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RICHETTI, KALEIGH, W  
BREATH TEST OPERATOR AGENCY: Fruitland Park PD  
ARRESTING OFFICER: RICHETTI, K  
ARRESTING OFFICER AGENCY: Fruitland Park PD  
LAST AGENCY INSPECTION: 04/09/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33522080A5  
DRY GAS EXPIRY DATE: 01/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 11:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: COB ARANA

DATE/TIME OF TEST: 05/04/2024 03:48:31

---

|                   |       |       |            |
|-------------------|-------|-------|------------|
| Diagnostics Check | OK    | 03:47 |            |
| Air Blank         |       | 03:48 | RFI Detect |
| Air Blank         | 0.000 | 03:48 |            |

---

Data Download Date/Time: 05/30/2024 10:56

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RICHETTI,KALEIGH,W  
BREATH TEST OPERATOR AGENCY: Fruitland Park PD  
ARRESTING OFFICER: RICHETTI,KALEIGH  
ARRESTING OFFICER AGENCY: Fruitland Park PD  
LAST AGENCY INSPECTION: 04/09/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33522080A5  
DRY GAS EXPIRY DATE: 01/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 23:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: COB ARANA

DATE/TIME OF TEST: 05/04/2024 03:56:58

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:50 |
| Air Blank          | 0.000 | 03:51 |
| Control Test       | 0.081 | 03:51 |
| Air Blank          | 0.000 | 03:52 |
| Subject Sample # 1 | 0.296 | 03:52 |
| Breath Volume      | 1.968 | 03:52 |
| Air Blank          | 0.000 | 03:53 |
| Air Blank          | 0.000 | 03:55 |
| Subject Sample # 2 | 0.295 | 03:55 |
| Breath Volume      | 2.187 | 03:55 |
| Air Blank          | 0.000 | 03:56 |
| Control Test       | 0.080 | 03:56 |
| Air Blank          | 0.000 | 03:56 |
| Diagnostics Check  | OK    | 03:56 |

---

Data Download Date/Time: 05/30/2024 10:56

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILDEMAN-STUART,NATASHA,L  
BREATH TEST OPERATOR AGENCY: Fruitland Park PD  
ARRESTING OFFICER: WILDEMAN-STUART,NATASH  
ARRESTING OFFICER AGENCY: Fruitland Park PD  
LAST AGENCY INSPECTION: 04/09/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33522080A5  
DRY GAS EXPIRY DATE: 01/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 21:01  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: GUILLEMETTE

DATE/TIME OF TEST: 05/04/2024 21:56:39

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 21:49 |
| Air Blank          | 0.000 | 21:50 |
| Control Test       | 0.081 | 21:50 |
| Air Blank          | 0.000 | 21:51 |
| Subject Sample # 1 | 0.142 | 21:51 |
| Breath Volume      | 2.726 | 21:51 |
| Air Blank          | 0.000 | 21:52 |
| Air Blank          | 0.000 | 21:54 |
| Subject Sample # 2 | 0.141 | 21:55 |
| Breath Volume      | 2.464 | 21:55 |
| Air Blank          | 0.000 | 21:55 |
| Control Test       | 0.079 | 21:56 |
| Air Blank          | 0.000 | 21:56 |
| Diagnostics Check  | OK    | 21:56 |

---

Data Download Date/Time: 05/30/2024 10:56

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LOUISSAINT,ROBERT,P  
BREATH TEST OPERATOR AGENCY: Clermont PD  
ARRESTING OFFICER: PRADO,JOSE  
ARRESTING OFFICER AGENCY: Lake County SO  
LAST AGENCY INSPECTION: 04/09/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33522080A5  
DRY GAS EXPIRY DATE: 01/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 02:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: CRUS ROJAS

DATE/TIME OF TEST: 05/12/2024 03:56:03

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 03:53 |                      |
| Air Blank         | 0.000 | 03:53 |                      |
| Control Test      | 0.081 | 03:54 |                      |
| Air Blank         | 0.000 | 03:54 |                      |
| Subject Sample    | # 1   | 03:54 | Subject Test Refused |
| Breath Volume     | 0.000 | 03:54 |                      |
| Air Blank         | 0.000 | 03:55 |                      |
| Control Test      | 0.081 | 03:55 |                      |
| Air Blank         | 0.000 | 03:55 |                      |
| Diagnostics Check | OK    | 03:56 |                      |

---

Data Download Date/Time: 05/30/2024 10:56

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PARKER,STEPHEN,V  
BREATH TEST OPERATOR AGENCY: Eustis PD  
ARRESTING OFFICER: PARKER,STEPHEN  
ARRESTING OFFICER AGENCY: Eustis PD  
LAST AGENCY INSPECTION: 04/09/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33522080A5  
DRY GAS EXPIRY DATE: 01/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 13:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 13:45

SUBJECT'S LAST NAME: GRIGSBY

DATE/TIME OF TEST: 05/13/2024 14:13:26

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 14:10 |                      |
| Air Blank         | 0.000 | 14:11 |                      |
| Control Test      | 0.081 | 14:11 |                      |
| Air Blank         | 0.000 | 14:11 |                      |
| Subject Sample    | # 1   | 14:12 | Subject Test Refused |
| Breath Volume     | 0.000 | 14:12 |                      |
| Air Blank         | 0.000 | 14:12 |                      |
| Control Test      | 0.081 | 14:12 |                      |
| Air Blank         | 0.000 | 14:13 |                      |
| Diagnostics Check | OK    | 14:13 |                      |

---

Data Download Date/Time: 05/30/2024 10:56

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: POTH,LAUREN,A  
BREATH TEST OPERATOR AGENCY: Mount Dora PD  
ARRESTING OFFICER: MORRISON,K  
ARRESTING OFFICER AGENCY: Mount Dora PD  
LAST AGENCY INSPECTION: 04/09/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33522080A5  
DRY GAS EXPIRY DATE: 01/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 13:16  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 14:10

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 05/15/2024 14:36:55

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 14:34 |                      |
| Air Blank         | 0.000 | 14:34 |                      |
| Control Test      | 0.080 | 14:34 |                      |
| Air Blank         | 0.000 | 14:35 |                      |
| Subject Sample    | # 1   | 14:35 | Subject Test Refused |
| Breath Volume     | 0.000 | 14:35 |                      |
| Air Blank         | 0.000 | 14:36 |                      |
| Control Test      | 0.080 | 14:36 |                      |
| Air Blank         | 0.000 | 14:36 |                      |
| Diagnostics Check | OK    | 14:36 |                      |

---

Data Download Date/Time: 05/30/2024 10:56

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RILEY,BRANDON,J  
BREATH TEST OPERATOR AGENCY: FHP Troop D  
ARRESTING OFFICER: RILEY,BRANDON  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 04/09/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33522080A5  
DRY GAS EXPIRY DATE: 01/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 22:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:24

SUBJECT'S LAST NAME: HORN

DATE/TIME OF TEST: 05/18/2024 23:55:41

---

|                   |       |       |                    |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK    | 23:44 |                    |
| Air Blank         | 0.000 | 23:44 |                    |
| Control Test      | 0.081 | 23:45 |                    |
| Air Blank         | 0.000 | 23:45 |                    |
| Subject Sample    | # 1   | 23:48 | No Sample Provided |
| Breath Volume     | 0.000 | 23:48 |                    |
| Air Blank         | 0.000 | 23:49 |                    |
| Air Blank         | 0.000 | 23:51 |                    |
| Subject Sample    | # 2   | 23:54 | No Sample Provided |
| Breath Volume     | 0.000 | 23:54 |                    |
| Air Blank         | 0.000 | 23:54 |                    |
| Control Test      | 0.081 | 23:55 |                    |
| Air Blank         | 0.000 | 23:55 |                    |
| Diagnostics Check | OK    | 23:55 |                    |

---

Data Download Date/Time: 05/30/2024 10:56

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MASON,BROOKE,M  
BREATH TEST OPERATOR AGENCY: Mount Dora PD  
ARRESTING OFFICER: OWENS,DENNIS  
ARRESTING OFFICER AGENCY: Mount Dora PD  
LAST AGENCY INSPECTION: 04/09/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33522080A5  
DRY GAS EXPIRY DATE: 01/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 23:06  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: PHILLIPS

DATE/TIME OF TEST: 05/22/2024 23:52:01

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:45 |
| Air Blank          | 0.000 | 23:46 |
| Control Test       | 0.080 | 23:46 |
| Air Blank          | 0.000 | 23:47 |
| Subject Sample # 1 | 0.203 | 23:47 |
| Breath Volume      | 1.472 | 23:47 |
| Air Blank          | 0.000 | 23:48 |
| Air Blank          | 0.000 | 23:50 |
| Subject Sample # 2 | 0.199 | 23:50 |
| Breath Volume      | 1.343 | 23:50 |
| Air Blank          | 0.000 | 23:51 |
| Control Test       | 0.081 | 23:51 |
| Air Blank          | 0.000 | 23:51 |
| Diagnostics Check  | OK    | 23:52 |

---

Data Download Date/Time: 05/30/2024 10:56

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BURTZLAFF, SYDNEE, M  
BREATH TEST OPERATOR AGENCY: Eustis PD  
ARRESTING OFFICER: REGO, BRYAN  
ARRESTING OFFICER AGENCY: Eustis PD  
LAST AGENCY INSPECTION: 04/09/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33522080A5  
DRY GAS EXPIRY DATE: 01/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 23:54  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: POE

DATE/TIME OF TEST: 05/24/2024 00:54:06

---

|                   |           |       |                |
|-------------------|-----------|-------|----------------|
| Diagnostics Check | OK        | 00:41 |                |
| Air Blank         | 0.000     | 00:42 |                |
| Control Test      | 0.081     | 00:42 |                |
| Air Blank         | 0.000     | 00:43 |                |
| Subject Sample    | # 1 0.222 | 00:46 | Volume Not Met |
| Breath Volume     | 0.195     | 00:46 |                |
| Air Blank         | 0.000     | 00:47 |                |
| Air Blank         | 0.000     | 00:48 |                |
| Subject Sample    | # 2 0.215 | 00:49 |                |
| Breath Volume     | 1.605     | 00:49 |                |
| Air Blank         | 0.000     | 00:50 |                |
| Air Blank         | 0.000     | 00:52 |                |
| Subject Sample    | # 3 0.205 | 00:52 |                |
| Breath Volume     | 1.304     | 00:52 |                |
| Air Blank         | 0.000     | 00:53 |                |
| Control Test      | 0.077     | 00:53 |                |
| Air Blank         | 0.000     | 00:54 |                |
| Diagnostics Check | OK        | 00:54 |                |

---

Data Download Date/Time: 05/30/2024 10:56

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROUBICEK,J,G  
BREATH TEST OPERATOR AGENCY: Eustis PD  
ARRESTING OFFICER: ROUBICEK,J  
ARRESTING OFFICER AGENCY: Eustis PD  
LAST AGENCY INSPECTION: 04/09/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33522080A5  
DRY GAS EXPIRY DATE: 01/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 21:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: PROUT

DATE/TIME OF TEST: 05/24/2024 22:22:42

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:16 |
| Air Blank          | 0.000 | 22:16 |
| Control Test       | 0.081 | 22:17 |
| Air Blank          | 0.000 | 22:17 |
| Subject Sample # 1 | 0.192 | 22:18 |
| Breath Volume      | 1.183 | 22:18 |
| Air Blank          | 0.000 | 22:18 |
| Air Blank          | 0.000 | 22:20 |
| Subject Sample # 2 | 0.185 | 22:21 |
| Breath Volume      | 1.503 | 22:21 |
| Air Blank          | 0.000 | 22:21 |
| Control Test       | 0.081 | 22:22 |
| Air Blank          | 0.000 | 22:22 |
| Diagnostics Check  | OK    | 22:22 |

---

Data Download Date/Time: 05/30/2024 10:56

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: STAMP,KYLE,A  
BREATH TEST OPERATOR AGENCY: FFWCC (Ocala)  
ARRESTING OFFICER: STAMP,KYLE  
ARRESTING OFFICER AGENCY: FFWCC (Ocala)  
LAST AGENCY INSPECTION: 04/09/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33522080A5  
DRY GAS EXPIRY DATE: 01/05/2025  
VIOLATION CODE: BUI  
ARREST TIME: 18:33  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:20

SUBJECT'S LAST NAME: LOWE

DATE/TIME OF TEST: 05/26/2024 19:48:17

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 19:45 |                      |
| Air Blank         | 0.000 | 19:45 |                      |
| Control Test      | 0.081 | 19:46 |                      |
| Air Blank         | 0.000 | 19:46 |                      |
| Subject Sample    | # 1   | 19:46 | Subject Test Refused |
| Breath Volume     | 0.000 | 19:46 |                      |
| Air Blank         | 0.000 | 19:47 |                      |
| Control Test      | 0.080 | 19:47 |                      |
| Air Blank         | 0.000 | 19:48 |                      |
| Diagnostics Check | OK    | 19:48 |                      |

---

Data Download Date/Time: 05/30/2024 10:56

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BURTZLAFF, SYDNEE, M  
BREATH TEST OPERATOR AGENCY: Eustis PD  
ARRESTING OFFICER: ROBINSON, MELISSA  
ARRESTING OFFICER AGENCY: Tavares PD  
LAST AGENCY INSPECTION: 04/09/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33522080A5  
DRY GAS EXPIRY DATE: 01/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 03:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: EZZELL

DATE/TIME OF TEST: 05/29/2024 04:01:29

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:55 |
| Air Blank          | 0.000 | 03:55 |
| Control Test       | 0.080 | 03:56 |
| Air Blank          | 0.000 | 03:56 |
| Subject Sample # 1 | 0.219 | 03:57 |
| Breath Volume      | 1.980 | 03:57 |
| Air Blank          | 0.000 | 03:57 |
| Air Blank          | 0.000 | 03:59 |
| Subject Sample # 2 | 0.215 | 03:59 |
| Breath Volume      | 2.480 | 03:59 |
| Air Blank          | 0.000 | 04:00 |
| Control Test       | 0.081 | 04:00 |
| Air Blank          | 0.000 | 04:01 |
| Diagnostics Check  | OK    | 04:01 |

---

Data Download Date/Time: 05/30/2024 10:56

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOOVER,DAVID,W  
BREATH TEST OPERATOR AGENCY: Eustis PD  
ARRESTING OFFICER: IMM,MICHAEL  
ARRESTING OFFICER AGENCY: Mount Dora PD  
LAST AGENCY INSPECTION: 04/09/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33522080A5  
DRY GAS EXPIRY DATE: 01/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 16:16  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:46

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 05/29/2024 17:13:49

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 17:06 |
| Air Blank          | 0.000 | 17:06 |
| Control Test       | 0.081 | 17:07 |
| Air Blank          | 0.000 | 17:07 |
| Subject Sample # 1 | 0.181 | 17:09 |
| Breath Volume      | 1.589 | 17:09 |
| Air Blank          | 0.000 | 17:10 |
| Air Blank          | 0.000 | 17:11 |
| Subject Sample # 2 | 0.176 | 17:12 |
| Breath Volume      | 2.058 | 17:12 |
| Air Blank          | 0.000 | 17:12 |
| Control Test       | 0.080 | 17:13 |
| Air Blank          | 0.000 | 17:13 |
| Diagnostics Check  | OK    | 17:13 |

---

Data Download Date/Time: 05/30/2024 10:56

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: POST,HAYDEN,L

BREATH TEST OPERATOR AGENCY: Leesburg PD

ARRESTING OFFICER: EVANS,BRENDIN

ARRESTING OFFICER AGENCY: Leesburg PD

LAST AGENCY INSPECTION: 04/29/2024

Date or Time Changed: No

DRY GAS STANDARD LOT #: 08223080A5

DRY GAS EXPIRY DATE: 06/05/2025

VIOLATION CODE: DUI

ARREST TIME: 15:02

GENDER: Male

OBSERVATION PERIOD BEGAN: 15:15

SUBJECT'S LAST NAME: BILLINGS

DATE/TIME OF TEST: 05/01/2024 15:48:53

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 15:46 |                      |
| Air Blank         | 0.000 | 15:46 |                      |
| Control Test      | 0.080 | 15:46 |                      |
| Air Blank         | 0.000 | 15:47 |                      |
| Subject Sample    | # 1   | 15:47 | Subject Test Refused |
| Breath Volume     | 0.000 | 15:47 |                      |
| Air Blank         | 0.000 | 15:47 |                      |
| Control Test      | 0.081 | 15:48 |                      |
| Air Blank         | 0.000 | 15:48 |                      |
| Diagnostics Check | OK    | 15:48 |                      |

---

Data Download Date/Time: 05/21/2024 14:59

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SOMMERSDORF,JOHN,G  
BREATH TEST OPERATOR AGENCY: Leesburg PD  
ARRESTING OFFICER: SOMMERSDORF,JOHN  
ARRESTING OFFICER AGENCY: Leesburg PD  
LAST AGENCY INSPECTION: 04/29/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08223080A5  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 06:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 07:14

SUBJECT'S LAST NAME: HORWEDEL

DATE/TIME OF TEST: 05/02/2024 07:58:38

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 07:52 |
| Air Blank          | 0.000 | 07:53 |
| Control Test       | 0.080 | 07:53 |
| Air Blank          | 0.000 | 07:53 |
| Subject Sample # 1 | 0.186 | 07:54 |
| Breath Volume      | 2.695 | 07:54 |
| Air Blank          | 0.000 | 07:54 |
| Air Blank          | 0.000 | 07:56 |
| Subject Sample # 2 | 0.182 | 07:57 |
| Breath Volume      | 3.074 | 07:57 |
| Air Blank          | 0.000 | 07:57 |
| Control Test       | 0.077 | 07:58 |
| Air Blank          | 0.000 | 07:58 |
| Diagnostics Check  | OK    | 07:58 |

---

Data Download Date/Time: 05/21/2024 14:59

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILDEMAN-STUART,JORDAN,R  
BREATH TEST OPERATOR AGENCY: Leesburg PD  
ARRESTING OFFICER: ROMAN,ALONZO  
ARRESTING OFFICER AGENCY: Leesburg PD  
LAST AGENCY INSPECTION: 04/29/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08223080A5  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 21:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:11

SUBJECT'S LAST NAME: MCCORMICK

DATE/TIME OF TEST: 05/03/2024 22:37:52

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:31 |
| Air Blank          | 0.000 | 22:32 |
| Control Test       | 0.080 | 22:32 |
| Air Blank          | 0.000 | 22:33 |
| Subject Sample # 1 | 0.281 | 22:33 |
| Breath Volume      | 3.875 | 22:33 |
| Air Blank          | 0.000 | 22:34 |
| Air Blank          | 0.000 | 22:35 |
| Subject Sample # 2 | 0.301 | 22:36 |
| Breath Volume      | 3.128 | 22:36 |
| Air Blank          | 0.000 | 22:36 |
| Control Test       | 0.079 | 22:37 |
| Air Blank          | 0.000 | 22:37 |
| Diagnostics Check  | OK    | 22:37 |

---

Data Download Date/Time: 05/21/2024 14:59

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: D'AGOSTINO,STEFANO,V  
BREATH TEST OPERATOR AGENCY: Leesburg PD  
ARRESTING OFFICER: LENFANT,BENJAMIN  
ARRESTING OFFICER AGENCY: Leesburg PD  
LAST AGENCY INSPECTION: 04/29/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08223080A5  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 00:04  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:21

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 05/12/2024 00:56:16

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 00:53 |                      |
| Air Blank         | 0.000 | 00:53 |                      |
| Control Test      | 0.081 | 00:54 |                      |
| Air Blank         | 0.000 | 00:54 |                      |
| Subject Sample    | # 1   | 00:54 | Subject Test Refused |
| Breath Volume     | 0.000 | 00:54 |                      |
| Air Blank         | 0.000 | 00:55 |                      |
| Control Test      | 0.081 | 00:55 |                      |
| Air Blank         | 0.000 | 00:56 |                      |
| Diagnostics Check | OK    | 00:56 |                      |

---

Data Download Date/Time: 05/21/2024 14:59

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: POST,HAYDEN,L  
BREATH TEST OPERATOR AGENCY: Leesburg PD  
ARRESTING OFFICER: POST,HAYDEN  
ARRESTING OFFICER AGENCY: Leesburg PD  
LAST AGENCY INSPECTION: 04/29/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08223080A5  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 16:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:09

SUBJECT'S LAST NAME: TRAINER

DATE/TIME OF TEST: 05/12/2024 17:39:35

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 17:33 |
| Air Blank          | 0.000 | 17:33 |
| Control Test       | 0.080 | 17:33 |
| Air Blank          | 0.000 | 17:34 |
| Subject Sample # 1 | 0.157 | 17:34 |
| Breath Volume      | 1.968 | 17:34 |
| Air Blank          | 0.000 | 17:35 |
| Air Blank          | 0.000 | 17:37 |
| Subject Sample # 2 | 0.164 | 17:38 |
| Breath Volume      | 3.160 | 17:38 |
| Air Blank          | 0.000 | 17:38 |
| Control Test       | 0.081 | 17:39 |
| Air Blank          | 0.000 | 17:39 |
| Diagnostics Check  | OK    | 17:39 |

---

Data Download Date/Time: 05/21/2024 14:59

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: D'AGOSTINO,STEFANO,V  
BREATH TEST OPERATOR AGENCY: Leesburg PD  
ARRESTING OFFICER: MEYER,PHILIP  
ARRESTING OFFICER AGENCY: Fruitland Park PD  
LAST AGENCY INSPECTION: 04/29/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08223080A5  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 19:13  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 19:45

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 05/12/2024 20:12:29

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 20:06 |
| Air Blank          | 0.000 | 20:06 |
| Control Test       | 0.080 | 20:06 |
| Air Blank          | 0.000 | 20:07 |
| Subject Sample # 1 | 0.137 | 20:08 |
| Breath Volume      | 2.761 | 20:08 |
| Air Blank          | 0.000 | 20:08 |
| Air Blank          | 0.000 | 20:10 |
| Subject Sample # 2 | 0.144 | 20:11 |
| Breath Volume      | 2.964 | 20:11 |
| Air Blank          | 0.000 | 20:11 |
| Control Test       | 0.080 | 20:11 |
| Air Blank          | 0.000 | 20:12 |
| Diagnostics Check  | OK    | 20:12 |

---

Data Download Date/Time: 05/21/2024 14:59

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MCCLAIN, DAVID, R  
BREATH TEST OPERATOR AGENCY: Leesburg PD  
ARRESTING OFFICER: PEREZ, EDWIN  
ARRESTING OFFICER AGENCY: Leesburg PD  
LAST AGENCY INSPECTION: 04/29/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08223080A5  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 13:33  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 13:49

SUBJECT'S LAST NAME: SANCHEZ SANCHEZ

DATE/TIME OF TEST: 05/19/2024 14:17:35

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 14:10 |
| Air Blank          | 0.000 | 14:11 |
| Control Test       | 0.079 | 14:11 |
| Air Blank          | 0.000 | 14:12 |
| Subject Sample # 1 | 0.157 | 14:12 |
| Breath Volume      | 1.351 | 14:12 |
| Air Blank          | 0.000 | 14:12 |
| Air Blank          | 0.000 | 14:14 |
| Subject Sample # 2 | 0.166 | 14:16 |
| Breath Volume      | 2.179 | 14:16 |
| Air Blank          | 0.000 | 14:16 |
| Control Test       | 0.079 | 14:17 |
| Air Blank          | 0.000 | 14:17 |
| Diagnostics Check  | OK    | 14:17 |

---

Data Download Date/Time: 05/21/2024 14:59

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: D'AGOSTINO,STEFANO,V  
BREATH TEST OPERATOR AGENCY: Leesburg PD  
ARRESTING OFFICER: VELEZ,JORDAN  
ARRESTING OFFICER AGENCY: Leesburg PD  
LAST AGENCY INSPECTION: 04/29/2024

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 08223080A5  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 01:03  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:29

SUBJECT'S LAST NAME: GODFREY

DATE/TIME OF TEST: 05/21/2024 01:56:43

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:50 |
| Air Blank          | 0.000 | 01:50 |
| Control Test       | 0.080 | 01:51 |
| Air Blank          | 0.000 | 01:51 |
| Subject Sample # 1 | 0.119 | 01:52 |
| Breath Volume      | 2.402 | 01:52 |
| Air Blank          | 0.000 | 01:52 |
| Air Blank          | 0.000 | 01:54 |
| Subject Sample # 2 | 0.122 | 01:55 |
| Breath Volume      | 2.082 | 01:55 |
| Air Blank          | 0.000 | 01:55 |
| Control Test       | 0.080 | 01:56 |
| Air Blank          | 0.000 | 01:56 |
| Diagnostics Check  | OK    | 01:56 |

---

Data Download Date/Time: 05/21/2024 15:00

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILDEMAN-STUART,JORDAN,R  
BREATH TEST OPERATOR AGENCY: Leesburg PD  
ARRESTING OFFICER: WILDEMAN-STUART,JORDAN  
ARRESTING OFFICER AGENCY: Leesburg PD  
LAST AGENCY INSPECTION: 05/21/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08223080A5  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 19:36  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 19:53

SUBJECT'S LAST NAME: MCCLURE

DATE/TIME OF TEST: 05/23/2024 20:18:19

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 20:14 |                      |
| Air Blank         | 0.000 | 20:15 |                      |
| Control Test      | 0.080 | 20:15 |                      |
| Air Blank         | 0.000 | 20:16 |                      |
| Subject Sample    | # 1   | 20:16 | Subject Test Refused |
| Breath Volume     | 0.000 | 20:16 |                      |
| Air Blank         | 0.000 | 20:17 |                      |
| Control Test      | 0.080 | 20:17 |                      |
| Air Blank         | 0.000 | 20:18 |                      |
| Diagnostics Check | OK    | 20:18 |                      |

---

Data Download Date/Time: 06/25/2024 16:00

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: YOUNG,ROBERT,A  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: YOUNG,ROBERT  
ARRESTING OFFICER AGENCY: Marion County SO  
LAST AGENCY INSPECTION: 04/26/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402872586  
DRY GAS EXPIRY DATE: 10/12/2026  
VIOLATION CODE: DUI  
ARREST TIME: 03:16  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:06

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 05/25/2024 03:36:40

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:26 |
| Air Blank          | 0.000 | 03:27 |
| Control Test       | 0.081 | 03:27 |
| Air Blank          | 0.000 | 03:27 |
| Subject Sample # 1 | 0.099 | 03:29 |
| Breath Volume      | 1.242 | 03:29 |
| Air Blank          | 0.000 | 03:30 |
| Air Blank          | 0.000 | 03:32 |
| Subject Sample # 2 | 0.088 | 03:35 |
| Breath Volume      | 1.488 | 03:35 |
| Air Blank          | 0.000 | 03:35 |
| Control Test       | 0.079 | 03:36 |
| Air Blank          | 0.000 | 03:36 |
| Diagnostics Check  | OK    | 03:36 |

---

Data Download Date/Time: 05/29/2024 07:35

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: LEISCH, B  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28423080A2  
DRY GAS EXPIRY DATE: 11/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 21:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:02

SUBJECT'S LAST NAME: GOMEZ RAMIREZ

DATE/TIME OF TEST: 05/03/2024 22:32:52

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:22 |
| Air Blank          | 0.000 | 22:23 |
| Control Test       | 0.081 | 22:23 |
| Air Blank          | 0.000 | 22:24 |
| Subject Sample # 1 | 0.203 | 22:25 |
| Breath Volume      | 2.351 | 22:25 |
| Air Blank          | 0.000 | 22:26 |
| Air Blank          | 0.000 | 22:27 |
| Subject Sample # 2 | 0.180 | 22:28 |
| Breath Volume      | 1.257 | 22:28 |
| Air Blank          | 0.000 | 22:29 |
| Air Blank          | 0.000 | 22:30 |
| Subject Sample # 3 | 0.181 | 22:31 |
| Breath Volume      | 1.464 | 22:31 |
| Air Blank          | 0.000 | 22:31 |
| Control Test       | 0.080 | 22:32 |
| Air Blank          | 0.000 | 22:32 |
| Diagnostics Check  | OK    | 22:32 |

---

Data Download Date/Time: 05/08/2024 11:55

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: ALVA-ARIAS,K  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28423080A2  
DRY GAS EXPIRY DATE: 11/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 03:34  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:13

SUBJECT'S LAST NAME: MARTINEZ ARRELLIN

DATE/TIME OF TEST: 05/04/2024 04:47:00

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:38 |
| Air Blank          | 0.000 | 04:39 |
| Control Test       | 0.081 | 04:39 |
| Air Blank          | 0.000 | 04:40 |
| Subject Sample # 1 | 0.200 | 04:42 |
| Breath Volume      | 1.523 | 04:42 |
| Air Blank          | 0.000 | 04:43 |
| Air Blank          | 0.000 | 04:44 |
| Subject Sample # 2 | 0.185 | 04:45 |
| Breath Volume      | 2.031 | 04:45 |
| Air Blank          | 0.000 | 04:46 |
| Control Test       | 0.077 | 04:46 |
| Air Blank          | 0.000 | 04:46 |
| Diagnostics Check  | OK    | 04:47 |

---

Data Download Date/Time: 05/08/2024 11:55

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ORTEGA,JESSICA,M  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: WILLIAMSON,A  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28423080A2  
DRY GAS EXPIRY DATE: 11/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 21:04  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:21

SUBJECT'S LAST NAME: LOVE

DATE/TIME OF TEST: 05/04/2024 21:49:39

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 21:42 |
| Air Blank          | 0.000 | 21:43 |
| Control Test       | 0.081 | 21:43 |
| Air Blank          | 0.000 | 21:44 |
| Subject Sample # 1 | 0.000 | 21:45 |
| Breath Volume      | 2.062 | 21:45 |
| Air Blank          | 0.000 | 21:45 |
| Air Blank          | 0.000 | 21:47 |
| Subject Sample # 2 | 0.000 | 21:48 |
| Breath Volume      | 1.679 | 21:48 |
| Air Blank          | 0.000 | 21:48 |
| Control Test       | 0.081 | 21:49 |
| Air Blank          | 0.000 | 21:49 |
| Diagnostics Check  | OK    | 21:49 |

---

Data Download Date/Time: 05/08/2024 11:55

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ORTEGA,JESSICA,M  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: ALVA-ARIAS,K  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28423080A2  
DRY GAS EXPIRY DATE: 11/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 00:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:18

SUBJECT'S LAST NAME: PAZ SANTIAGO

DATE/TIME OF TEST: 05/06/2024 01:54:03

|                   |           |       |                |
|-------------------|-----------|-------|----------------|
| Diagnostics Check | OK        | 01:38 |                |
| Air Blank         | 0.000     | 01:39 |                |
| Control Test      | 0.081     | 01:39 |                |
| Air Blank         | 0.000     | 01:40 |                |
| Subject Sample    | # 1 0.167 | 01:41 |                |
| Breath Volume     | 1.417     | 01:41 |                |
| Air Blank         | 0.000     | 01:41 |                |
| Air Blank         | 0.000     | 01:43 |                |
| Subject Sample    | # 2 0.162 | 01:46 | Volume Not Met |
| Breath Volume     | 0.058     | 01:46 |                |
| Air Blank         | 0.000     | 01:47 |                |
| Air Blank         | 0.000     | 01:49 |                |
| Subject Sample    | # 3 0.156 | 01:52 | Volume Not Met |
| Breath Volume     | 0.246     | 01:52 |                |
| Air Blank         | 0.000     | 01:53 |                |
| Control Test      | 0.079     | 01:53 |                |
| Air Blank         | 0.000     | 01:53 |                |
| Diagnostics Check | OK        | 01:54 |                |

Data Download Date/Time: 05/08/2024 11:55

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PIERCEY,JANICE,F  
BREATH TEST OPERATOR AGENCY: Tampa PD  
ARRESTING OFFICER: FRUEH,D  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28423080A2  
DRY GAS EXPIRY DATE: 11/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 02:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:37

SUBJECT'S LAST NAME: OTIS

DATE/TIME OF TEST: 05/08/2024 03:05:48

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:58 |
| Air Blank          | 0.000 | 02:58 |
| Control Test       | 0.081 | 02:59 |
| Air Blank          | 0.000 | 02:59 |
| Subject Sample # 1 | 0.136 | 03:00 |
| Breath Volume      | 1.589 | 03:00 |
| Air Blank          | 0.000 | 03:01 |
| Air Blank          | 0.000 | 03:03 |
| Subject Sample # 2 | 0.141 | 03:04 |
| Breath Volume      | 1.898 | 03:04 |
| Air Blank          | 0.000 | 03:04 |
| Control Test       | 0.079 | 03:05 |
| Air Blank          | 0.000 | 03:05 |
| Diagnostics Check  | OK    | 03:05 |

---

Data Download Date/Time: 05/08/2024 11:55

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PIERCEY,JANICE,F  
BREATH TEST OPERATOR AGENCY: Tampa PD  
ARRESTING OFFICER: SARRASIN,R  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 05/08/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28423080A2  
DRY GAS EXPIRY DATE: 11/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 01:18  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:31

SUBJECT'S LAST NAME: GALLANT

DATE/TIME OF TEST: 05/09/2024 01:58:28

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:51 |
| Air Blank          | 0.000 | 01:52 |
| Control Test       | 0.081 | 01:52 |
| Air Blank          | 0.000 | 01:53 |
| Subject Sample # 1 | 0.121 | 01:53 |
| Breath Volume      | 2.625 | 01:53 |
| Air Blank          | 0.000 | 01:54 |
| Air Blank          | 0.000 | 01:56 |
| Subject Sample # 2 | 0.115 | 01:56 |
| Breath Volume      | 2.218 | 01:56 |
| Air Blank          | 0.000 | 01:57 |
| Control Test       | 0.080 | 01:57 |
| Air Blank          | 0.000 | 01:58 |
| Diagnostics Check  | OK    | 01:58 |

---

Data Download Date/Time: 06/05/2024 16:16

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: GARNER,W  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 05/08/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28423080A2  
DRY GAS EXPIRY DATE: 11/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 03:12  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:37

SUBJECT'S LAST NAME: JACOBSEN

DATE/TIME OF TEST: 05/09/2024 04:04:38

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:58 |
| Air Blank          | 0.000 | 03:58 |
| Control Test       | 0.080 | 03:58 |
| Air Blank          | 0.000 | 03:59 |
| Subject Sample # 1 | 0.129 | 04:00 |
| Breath Volume      | 2.687 | 04:00 |
| Air Blank          | 0.000 | 04:00 |
| Air Blank          | 0.000 | 04:02 |
| Subject Sample # 2 | 0.137 | 04:02 |
| Breath Volume      | 2.835 | 04:02 |
| Air Blank          | 0.000 | 04:03 |
| Control Test       | 0.079 | 04:04 |
| Air Blank          | 0.000 | 04:04 |
| Diagnostics Check  | OK    | 04:04 |

---

Data Download Date/Time: 06/05/2024 16:16

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JONES,MINDY,L  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: VAN-TREESE,S  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 05/08/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28423080A2  
DRY GAS EXPIRY DATE: 11/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 00:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: CUTLER

DATE/TIME OF TEST: 05/10/2024 01:47:18

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:41 |
| Air Blank          | 0.000 | 01:41 |
| Control Test       | 0.081 | 01:41 |
| Air Blank          | 0.000 | 01:42 |
| Subject Sample # 1 | 0.000 | 01:42 |
| Breath Volume      | 3.242 | 01:42 |
| Air Blank          | 0.000 | 01:43 |
| Air Blank          | 0.000 | 01:45 |
| Subject Sample # 2 | 0.000 | 01:45 |
| Breath Volume      | 3.335 | 01:45 |
| Air Blank          | 0.000 | 01:46 |
| Control Test       | 0.080 | 01:46 |
| Air Blank          | 0.000 | 01:47 |
| Diagnostics Check  | OK    | 01:47 |

---

Data Download Date/Time: 06/05/2024 16:16

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ORTEGA,JESSICA,M  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: HERNANDEZ,S  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 05/08/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28423080A2  
DRY GAS EXPIRY DATE: 11/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 20:54  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:39

SUBJECT'S LAST NAME: CRECENCIO LOPEZ

DATE/TIME OF TEST: 05/10/2024 22:11:39

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 21:59 |
| Air Blank          | 0.000 | 22:00 |
| Control Test       | 0.081 | 22:00 |
| Air Blank          | 0.000 | 22:01 |
| Subject Sample # 1 | 0.253 | 22:01 |
| Breath Volume      | 1.671 | 22:01 |
| Air Blank          | 0.000 | 22:02 |
| Air Blank          | 0.000 | 22:04 |
| Subject Sample # 2 | 0.220 | 22:05 |
| Breath Volume      | 1.164 | 22:05 |
| Air Blank          | 0.000 | 22:05 |
| Air Blank          | 0.000 | 22:07 |
| Subject Sample # 3 | 0.241 | 22:09 |
| Breath Volume      | 1.753 | 22:09 |
| Air Blank          | 0.000 | 22:10 |
| Control Test       | 0.080 | 22:11 |
| Air Blank          | 0.000 | 22:11 |
| Diagnostics Check  | OK    | 22:11 |

---

Data Download Date/Time: 06/05/2024 16:16

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JONES,MINDY,L  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: VAN-TREESE,S  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 05/08/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28423080A2  
DRY GAS EXPIRY DATE: 11/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 21:38  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:04

SUBJECT'S LAST NAME: DEBREUS

DATE/TIME OF TEST: 05/10/2024 22:36:28

---

|                   |           |       |                |
|-------------------|-----------|-------|----------------|
| Diagnostics Check | OK        | 22:24 |                |
| Air Blank         | 0.000     | 22:24 |                |
| Control Test      | 0.079     | 22:25 |                |
| Air Blank         | 0.000     | 22:25 |                |
| Subject Sample    | # 1 0.246 | 22:28 | Volume Not Met |
| Breath Volume     | 0.558     | 22:28 |                |
| Air Blank         | 0.000     | 22:29 |                |
| Air Blank         | 0.000     | 22:31 |                |
| Subject Sample    | # 2 0.247 | 22:34 | Volume Not Met |
| Breath Volume     | 0.031     | 22:34 |                |
| Air Blank         | 0.000     | 22:35 |                |
| Control Test      | 0.078     | 22:35 |                |
| Air Blank         | 0.000     | 22:36 |                |
| Diagnostics Check | OK        | 22:36 |                |

---

Data Download Date/Time: 06/05/2024 16:16

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JONES,MINDY,L  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: VAN-TREESE,S  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 05/08/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28423080A2  
DRY GAS EXPIRY DATE: 11/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 21:38  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:04

SUBJECT'S LAST NAME: DEBREUS

DATE/TIME OF TEST: 05/10/2024 22:48:01

---

|                   |           |       |                      |
|-------------------|-----------|-------|----------------------|
| Diagnostics Check | OK        | 22:39 |                      |
| Air Blank         | 0.000     | 22:39 |                      |
| Control Test      | 0.078     | 22:40 |                      |
| Air Blank         | 0.000     | 22:40 |                      |
| Subject Sample    | # 1 0.245 | 22:43 | Volume Not Met       |
| Breath Volume     | 0.640     | 22:43 |                      |
| Air Blank         | 0.000     | 22:44 |                      |
| Air Blank         | 0.000     | 22:46 |                      |
| Subject Sample    | # 2       | 22:46 | Subject Test Refused |
| Breath Volume     | 0.000     | 22:46 |                      |
| Air Blank         | 0.000     | 22:47 |                      |
| Control Test      | 0.076     | 22:47 |                      |
| Air Blank         | 0.000     | 22:47 |                      |
| Diagnostics Check | OK        | 22:48 |                      |

---

Data Download Date/Time: 06/05/2024 16:16

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JONES,MINDY,L  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: HOTHEM,A  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 05/08/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28423080A2  
DRY GAS EXPIRY DATE: 11/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 03:37  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:07

SUBJECT'S LAST NAME: PERKINS

DATE/TIME OF TEST: 05/11/2024 04:34:37

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:28 |
| Air Blank          | 0.000 | 04:28 |
| Control Test       | 0.081 | 04:28 |
| Air Blank          | 0.000 | 04:29 |
| Subject Sample # 1 | 0.041 | 04:30 |
| Breath Volume      | 1.933 | 04:30 |
| Air Blank          | 0.000 | 04:30 |
| Air Blank          | 0.000 | 04:32 |
| Subject Sample # 2 | 0.041 | 04:33 |
| Breath Volume      | 1.613 | 04:33 |
| Air Blank          | 0.000 | 04:33 |
| Control Test       | 0.082 | 04:34 |
| Air Blank          | 0.000 | 04:34 |
| Diagnostics Check  | OK    | 04:34 |

---

Data Download Date/Time: 06/05/2024 16:16

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ORTEGA,JESSICA,M  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: MILILLO,M  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 05/08/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28423080A2  
DRY GAS EXPIRY DATE: 11/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 05:13  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:37

SUBJECT'S LAST NAME: MARTINEZ CASTRO

DATE/TIME OF TEST: 05/11/2024 06:04:42

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 05:57 |
| Air Blank          | 0.000 | 05:58 |
| Control Test       | 0.081 | 05:58 |
| Air Blank          | 0.000 | 05:58 |
| Subject Sample # 1 | 0.145 | 06:00 |
| Breath Volume      | 1.812 | 06:00 |
| Air Blank          | 0.000 | 06:00 |
| Air Blank          | 0.000 | 06:02 |
| Subject Sample # 2 | 0.145 | 06:03 |
| Breath Volume      | 1.781 | 06:03 |
| Air Blank          | 0.000 | 06:03 |
| Control Test       | 0.079 | 06:04 |
| Air Blank          | 0.000 | 06:04 |
| Diagnostics Check  | OK    | 06:04 |

---

Data Download Date/Time: 06/05/2024 16:17

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: LAM,K  
ARRESTING OFFICER AGENCY: Plant City PD  
LAST AGENCY INSPECTION: 05/08/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28423080A2  
DRY GAS EXPIRY DATE: 11/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 21:43  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: FRANCISCO PEDRO

DATE/TIME OF TEST: 05/11/2024 23:24:54

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:17 |
| Air Blank          | 0.000 | 23:18 |
| Control Test       | 0.081 | 23:18 |
| Air Blank          | 0.000 | 23:18 |
| Subject Sample # 1 | 0.253 | 23:20 |
| Breath Volume      | 1.480 | 23:20 |
| Air Blank          | 0.000 | 23:21 |
| Air Blank          | 0.000 | 23:22 |
| Subject Sample # 2 | 0.257 | 23:23 |
| Breath Volume      | 2.308 | 23:23 |
| Air Blank          | 0.000 | 23:23 |
| Control Test       | 0.079 | 23:24 |
| Air Blank          | 0.000 | 23:24 |
| Diagnostics Check  | OK    | 23:24 |

---

Data Download Date/Time: 06/05/2024 16:17

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GANT, TIFFANY,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: LANDRY, G  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 05/08/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28423080A2  
DRY GAS EXPIRY DATE: 11/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 22:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:38

SUBJECT'S LAST NAME: ORTEGA-RAMDON

DATE/TIME OF TEST: 05/12/2024 00:06:40

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:59 |
| Air Blank          | 0.000 | 00:00 |
| Control Test       | 0.080 | 00:00 |
| Air Blank          | 0.000 | 00:01 |
| Subject Sample # 1 | 0.119 | 00:02 |
| Breath Volume      | 2.765 | 00:02 |
| Air Blank          | 0.000 | 00:02 |
| Air Blank          | 0.000 | 00:04 |
| Subject Sample # 2 | 0.129 | 00:05 |
| Breath Volume      | 2.839 | 00:05 |
| Air Blank          | 0.000 | 00:05 |
| Control Test       | 0.080 | 00:06 |
| Air Blank          | 0.000 | 00:06 |
| Diagnostics Check  | OK    | 00:06 |

---

Data Download Date/Time: 06/05/2024 16:17

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: RENDON, D  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 05/08/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28423080A2  
DRY GAS EXPIRY DATE: 11/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 02:53  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:24

SUBJECT'S LAST NAME: MARTINEZ BAUTISTA

DATE/TIME OF TEST: 05/12/2024 03:51:39

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:44 |
| Air Blank          | 0.000 | 03:44 |
| Control Test       | 0.080 | 03:44 |
| Air Blank          | 0.000 | 03:45 |
| Subject Sample # 1 | 0.189 | 03:46 |
| Breath Volume      | 3.472 | 03:46 |
| Air Blank          | 0.000 | 03:47 |
| Air Blank          | 0.000 | 03:49 |
| Subject Sample # 2 | 0.185 | 03:49 |
| Breath Volume      | 2.183 | 03:49 |
| Air Blank          | 0.000 | 03:50 |
| Control Test       | 0.079 | 03:51 |
| Air Blank          | 0.000 | 03:51 |
| Diagnostics Check  | OK    | 03:51 |

---

Data Download Date/Time: 06/05/2024 16:17

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PIERCEY,JANICE,F  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BURKE,B  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 05/08/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28423080A2  
DRY GAS EXPIRY DATE: 11/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 04:03  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:35

SUBJECT'S LAST NAME: PIDOCHEV

DATE/TIME OF TEST: 05/12/2024 05:04:55

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:58 |
| Air Blank          | 0.000 | 04:58 |
| Control Test       | 0.080 | 04:58 |
| Air Blank          | 0.000 | 04:59 |
| Subject Sample # 1 | 0.194 | 05:00 |
| Breath Volume      | 2.789 | 05:00 |
| Air Blank          | 0.000 | 05:01 |
| Air Blank          | 0.000 | 05:02 |
| Subject Sample # 2 | 0.186 | 05:03 |
| Breath Volume      | 1.777 | 05:03 |
| Air Blank          | 0.000 | 05:03 |
| Control Test       | 0.079 | 05:04 |
| Air Blank          | 0.000 | 05:04 |
| Diagnostics Check  | OK    | 05:04 |

---

Data Download Date/Time: 06/05/2024 16:17

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: MILILLO,M  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 05/08/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28423080A2  
DRY GAS EXPIRY DATE: 11/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 04:27  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:55

SUBJECT'S LAST NAME: ZARINANA

DATE/TIME OF TEST: 05/12/2024 05:20:23

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 05:15 |                      |
| Air Blank         | 0.000 | 05:15 |                      |
| Control Test      | 0.080 | 05:16 |                      |
| Air Blank         | 0.000 | 05:16 |                      |
| Subject Sample    | # 1   | 05:19 | Subject Test Refused |
| Breath Volume     | 0.000 | 05:19 |                      |
| Air Blank         | 0.000 | 05:19 |                      |
| Control Test      | 0.080 | 05:19 |                      |
| Air Blank         | 0.000 | 05:20 |                      |
| Diagnostics Check | OK    | 05:20 |                      |

---

Data Download Date/Time: 06/05/2024 16:17

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GANT, TIFFANY,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BADEN, T  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 05/08/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28423080A2  
DRY GAS EXPIRY DATE: 11/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 02:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:46

SUBJECT'S LAST NAME: LOPEZ LOPEZ

DATE/TIME OF TEST: 05/13/2024 03:14:22

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:06 |
| Air Blank          | 0.000 | 03:07 |
| Control Test       | 0.082 | 03:07 |
| Air Blank          | 0.000 | 03:08 |
| Subject Sample # 1 | 0.159 | 03:09 |
| Breath Volume      | 1.515 | 03:09 |
| Air Blank          | 0.000 | 03:10 |
| Air Blank          | 0.000 | 03:12 |
| Subject Sample # 2 | 0.152 | 03:12 |
| Breath Volume      | 1.480 | 03:12 |
| Air Blank          | 0.000 | 03:13 |
| Control Test       | 0.079 | 03:13 |
| Air Blank          | 0.000 | 03:14 |
| Diagnostics Check  | OK    | 03:14 |

---

Data Download Date/Time: 06/05/2024 16:17

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PIERCEY,JANICE,F  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BURKE,B  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 05/08/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28423080A2  
DRY GAS EXPIRY DATE: 11/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 03:24  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: BROOKS

DATE/TIME OF TEST: 05/13/2024 04:32:50

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:25 |
| Air Blank          | 0.000 | 04:26 |
| Control Test       | 0.081 | 04:26 |
| Air Blank          | 0.000 | 04:26 |
| Subject Sample # 1 | 0.130 | 04:27 |
| Breath Volume      | 1.285 | 04:27 |
| Air Blank          | 0.000 | 04:28 |
| Air Blank          | 0.000 | 04:29 |
| Subject Sample # 2 | 0.131 | 04:31 |
| Breath Volume      | 1.976 | 04:31 |
| Air Blank          | 0.000 | 04:31 |
| Control Test       | 0.080 | 04:32 |
| Air Blank          | 0.000 | 04:32 |
| Diagnostics Check  | OK    | 04:32 |

---

Data Download Date/Time: 06/05/2024 16:17

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GANT, TIFFANY,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BAILEY, D  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 05/08/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28423080A2  
DRY GAS EXPIRY DATE: 11/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 01:13  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: SCHULTZ

DATE/TIME OF TEST: 05/14/2024 01:55:31

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 01:52 |                      |
| Air Blank         | 0.000 | 01:52 |                      |
| Control Test      | 0.080 | 01:53 |                      |
| Air Blank         | 0.000 | 01:53 |                      |
| Subject Sample    | # 1   | 01:54 | Subject Test Refused |
| Breath Volume     | 0.000 | 01:54 |                      |
| Air Blank         | 0.000 | 01:54 |                      |
| Control Test      | 0.079 | 01:54 |                      |
| Air Blank         | 0.000 | 01:55 |                      |
| Diagnostics Check | OK    | 01:55 |                      |

---

Data Download Date/Time: 06/05/2024 16:17

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: FLINT, J  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 05/08/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28423080A2  
DRY GAS EXPIRY DATE: 11/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 03:13  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:04

SUBJECT'S LAST NAME: FATHALLA

DATE/TIME OF TEST: 05/14/2024 04:31:22

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:24 |
| Air Blank          | 0.000 | 04:24 |
| Control Test       | 0.081 | 04:25 |
| Air Blank          | 0.000 | 04:25 |
| Subject Sample # 1 | 0.047 | 04:26 |
| Breath Volume      | 2.574 | 04:26 |
| Air Blank          | 0.000 | 04:27 |
| Air Blank          | 0.000 | 04:28 |
| Subject Sample # 2 | 0.045 | 04:29 |
| Breath Volume      | 2.773 | 04:29 |
| Air Blank          | 0.000 | 04:30 |
| Control Test       | 0.079 | 04:30 |
| Air Blank          | 0.000 | 04:31 |
| Diagnostics Check  | OK    | 04:31 |

---

Data Download Date/Time: 06/05/2024 16:17

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PIERCEY,JANICE,F  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: GARNER,W  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 05/08/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28423080A2  
DRY GAS EXPIRY DATE: 11/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 02:36  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 05/15/2024 03:25:54

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:19 |
| Air Blank          | 0.000 | 03:19 |
| Control Test       | 0.080 | 03:19 |
| Air Blank          | 0.000 | 03:20 |
| Subject Sample # 1 | 0.202 | 03:21 |
| Breath Volume      | 3.187 | 03:21 |
| Air Blank          | 0.000 | 03:22 |
| Air Blank          | 0.000 | 03:23 |
| Subject Sample # 2 | 0.208 | 03:24 |
| Breath Volume      | 2.433 | 03:24 |
| Air Blank          | 0.000 | 03:24 |
| Control Test       | 0.080 | 03:25 |
| Air Blank          | 0.000 | 03:25 |
| Diagnostics Check  | OK    | 03:25 |

---

Data Download Date/Time: 06/05/2024 16:17

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JONES,MINDY,L  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: DIBIASE,J  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 05/08/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28423080A2  
DRY GAS EXPIRY DATE: 11/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 03:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:49

SUBJECT'S LAST NAME: CAMPBELL

DATE/TIME OF TEST: 05/15/2024 04:16:41

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:09 |
| Air Blank          | 0.000 | 04:10 |
| Control Test       | 0.080 | 04:10 |
| Air Blank          | 0.000 | 04:11 |
| Subject Sample # 1 | 0.098 | 04:12 |
| Breath Volume      | 2.375 | 04:12 |
| Air Blank          | 0.000 | 04:12 |
| Air Blank          | 0.000 | 04:14 |
| Subject Sample # 2 | 0.096 | 04:15 |
| Breath Volume      | 1.230 | 04:15 |
| Air Blank          | 0.000 | 04:15 |
| Control Test       | 0.079 | 04:16 |
| Air Blank          | 0.000 | 04:16 |
| Diagnostics Check  | OK    | 04:16 |

---

Data Download Date/Time: 06/05/2024 16:17

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ORTEGA,JESSICA,M  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: SMITH,D  
ARRESTING OFFICER AGENCY: Plant City PD  
LAST AGENCY INSPECTION: 05/08/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28423080A2  
DRY GAS EXPIRY DATE: 11/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 06:23  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 07:37

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 05/16/2024 08:04:33

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 07:57 |
| Air Blank          | 0.000 | 07:58 |
| Control Test       | 0.081 | 07:58 |
| Air Blank          | 0.000 | 07:59 |
| Subject Sample # 1 | 0.147 | 07:59 |
| Breath Volume      | 3.203 | 07:59 |
| Air Blank          | 0.000 | 08:00 |
| Air Blank          | 0.000 | 08:02 |
| Subject Sample # 2 | 0.151 | 08:02 |
| Breath Volume      | 3.480 | 08:02 |
| Air Blank          | 0.000 | 08:03 |
| Control Test       | 0.079 | 08:03 |
| Air Blank          | 0.000 | 08:04 |
| Diagnostics Check  | OK    | 08:04 |

---

Data Download Date/Time: 06/05/2024 16:17

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GANT, TIFFANY,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: RENDON, D  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 05/08/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28423080A2  
DRY GAS EXPIRY DATE: 11/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 03:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: RADA OROZCO

DATE/TIME OF TEST: 05/17/2024 04:47:24

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:41 |
| Air Blank          | 0.000 | 04:41 |
| Control Test       | 0.081 | 04:41 |
| Air Blank          | 0.000 | 04:42 |
| Subject Sample # 1 | 0.204 | 04:42 |
| Breath Volume      | 2.406 | 04:42 |
| Air Blank          | 0.000 | 04:43 |
| Air Blank          | 0.000 | 04:45 |
| Subject Sample # 2 | 0.199 | 04:45 |
| Breath Volume      | 2.070 | 04:45 |
| Air Blank          | 0.000 | 04:46 |
| Control Test       | 0.080 | 04:46 |
| Air Blank          | 0.000 | 04:47 |
| Diagnostics Check  | OK    | 04:47 |

---

Data Download Date/Time: 06/05/2024 16:17

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BADEN, T  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 05/08/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28423080A2  
DRY GAS EXPIRY DATE: 11/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 21:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:34

SUBJECT'S LAST NAME: CLAUDIO MORALES

DATE/TIME OF TEST: 05/17/2024 22:00:54

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 21:54 |
| Air Blank          | 0.000 | 21:54 |
| Control Test       | 0.081 | 21:55 |
| Air Blank          | 0.000 | 21:55 |
| Subject Sample # 1 | 0.147 | 21:56 |
| Breath Volume      | 3.218 | 21:56 |
| Air Blank          | 0.000 | 21:57 |
| Air Blank          | 0.000 | 21:58 |
| Subject Sample # 2 | 0.152 | 21:59 |
| Breath Volume      | 2.925 | 21:59 |
| Air Blank          | 0.000 | 21:59 |
| Control Test       | 0.079 | 22:00 |
| Air Blank          | 0.000 | 22:00 |
| Diagnostics Check  | OK    | 22:00 |

---

Data Download Date/Time: 06/05/2024 16:17

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GANT, TIFFANY,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: GARNER, W  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 05/08/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28423080A2  
DRY GAS EXPIRY DATE: 11/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 23:02  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:24

SUBJECT'S LAST NAME: WOLFRAM

DATE/TIME OF TEST: 05/17/2024 23:51:22

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:45 |
| Air Blank          | 0.000 | 23:45 |
| Control Test       | 0.080 | 23:45 |
| Air Blank          | 0.000 | 23:46 |
| Subject Sample # 1 | 0.224 | 23:46 |
| Breath Volume      | 1.425 | 23:46 |
| Air Blank          | 0.000 | 23:47 |
| Air Blank          | 0.000 | 23:49 |
| Subject Sample # 2 | 0.218 | 23:49 |
| Breath Volume      | 1.875 | 23:49 |
| Air Blank          | 0.000 | 23:50 |
| Control Test       | 0.079 | 23:50 |
| Air Blank          | 0.000 | 23:51 |
| Diagnostics Check  | OK    | 23:51 |

---

Data Download Date/Time: 06/05/2024 16:17

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: FLINT,J  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 05/08/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28423080A2  
DRY GAS EXPIRY DATE: 11/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 23:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:26

SUBJECT'S LAST NAME: PATISHTAN GOMEZ

DATE/TIME OF TEST: 05/18/2024 00:53:57

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:47 |
| Air Blank          | 0.000 | 00:47 |
| Control Test       | 0.079 | 00:47 |
| Air Blank          | 0.000 | 00:48 |
| Subject Sample # 1 | 0.178 | 00:49 |
| Breath Volume      | 1.753 | 00:49 |
| Air Blank          | 0.000 | 00:50 |
| Air Blank          | 0.000 | 00:51 |
| Subject Sample # 2 | 0.177 | 00:52 |
| Breath Volume      | 1.480 | 00:52 |
| Air Blank          | 0.000 | 00:53 |
| Control Test       | 0.079 | 00:53 |
| Air Blank          | 0.000 | 00:53 |
| Diagnostics Check  | OK    | 00:53 |

---

Data Download Date/Time: 06/05/2024 16:17

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: SNAGGS, S  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 05/08/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28423080A2  
DRY GAS EXPIRY DATE: 11/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 03:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:12

SUBJECT'S LAST NAME: MORATAYA GUTIERREZ

DATE/TIME OF TEST: 05/18/2024 04:40:24

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:33 |
| Air Blank          | 0.000 | 04:33 |
| Control Test       | 0.081 | 04:33 |
| Air Blank          | 0.000 | 04:34 |
| Subject Sample # 1 | 0.116 | 04:35 |
| Breath Volume      | 1.722 | 04:35 |
| Air Blank          | 0.000 | 04:36 |
| Air Blank          | 0.000 | 04:38 |
| Subject Sample # 2 | 0.118 | 04:38 |
| Breath Volume      | 2.085 | 04:38 |
| Air Blank          | 0.000 | 04:39 |
| Control Test       | 0.079 | 04:39 |
| Air Blank          | 0.000 | 04:40 |
| Diagnostics Check  | OK    | 04:40 |

---

Data Download Date/Time: 06/05/2024 16:17

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KROME,BRITTNEY,M  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: KISS,L  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 05/08/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28423080A2  
DRY GAS EXPIRY DATE: 11/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 20:02  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:47

SUBJECT'S LAST NAME: LEONARD JR

DATE/TIME OF TEST: 05/18/2024 21:13:59

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 21:07 |
| Air Blank          | 0.000 | 21:07 |
| Control Test       | 0.081 | 21:08 |
| Air Blank          | 0.000 | 21:08 |
| Subject Sample # 1 | 0.241 | 21:09 |
| Breath Volume      | 3.312 | 21:09 |
| Air Blank          | 0.000 | 21:09 |
| Air Blank          | 0.000 | 21:11 |
| Subject Sample # 2 | 0.230 | 21:12 |
| Breath Volume      | 1.949 | 21:12 |
| Air Blank          | 0.000 | 21:13 |
| Control Test       | 0.078 | 21:13 |
| Air Blank          | 0.000 | 21:13 |
| Diagnostics Check  | OK    | 21:13 |

---

Data Download Date/Time: 06/05/2024 16:17

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JONES,MINDY,L  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BAILEY,D  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 05/08/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28423080A2  
DRY GAS EXPIRY DATE: 11/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 03:40  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:59

SUBJECT'S LAST NAME: BARGO

DATE/TIME OF TEST: 05/19/2024 04:25:57

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:19 |
| Air Blank          | 0.000 | 04:19 |
| Control Test       | 0.080 | 04:20 |
| Air Blank          | 0.000 | 04:20 |
| Subject Sample # 1 | 0.211 | 04:21 |
| Breath Volume      | 2.933 | 04:21 |
| Air Blank          | 0.000 | 04:21 |
| Air Blank          | 0.000 | 04:23 |
| Subject Sample # 2 | 0.212 | 04:24 |
| Breath Volume      | 2.781 | 04:24 |
| Air Blank          | 0.000 | 04:25 |
| Control Test       | 0.080 | 04:25 |
| Air Blank          | 0.000 | 04:25 |
| Diagnostics Check  | OK    | 04:25 |

---

Data Download Date/Time: 06/05/2024 16:18

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PIERCEY,JANICE,F  
BREATH TEST OPERATOR AGENCY: Tampa PD  
ARRESTING OFFICER: BLANCHARD,J  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 05/08/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28423080A2  
DRY GAS EXPIRY DATE: 11/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 05:57  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 06:21

SUBJECT'S LAST NAME: OTERO

DATE/TIME OF TEST: 05/19/2024 06:47:00

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 06:41 |                      |
| Air Blank         | 0.000 | 06:42 |                      |
| Control Test      | 0.081 | 06:42 |                      |
| Air Blank         | 0.000 | 06:42 |                      |
| Subject Sample    | # 1   | 06:45 | Subject Test Refused |
| Breath Volume     | 0.000 | 06:45 |                      |
| Air Blank         | 0.000 | 06:46 |                      |
| Control Test      | 0.079 | 06:46 |                      |
| Air Blank         | 0.000 | 06:46 |                      |
| Diagnostics Check | OK    | 06:47 |                      |

---

Data Download Date/Time: 06/05/2024 16:18

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KROME,BRITTNEY,M  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: VAN-TREESE,S  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 05/08/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28423080A2  
DRY GAS EXPIRY DATE: 11/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 21:02  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:27

SUBJECT'S LAST NAME: GARDNER

DATE/TIME OF TEST: 05/19/2024 21:54:00

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 21:47 |
| Air Blank          | 0.000 | 21:47 |
| Control Test       | 0.080 | 21:48 |
| Air Blank          | 0.000 | 21:48 |
| Subject Sample # 1 | 0.112 | 21:49 |
| Breath Volume      | 2.074 | 21:49 |
| Air Blank          | 0.000 | 21:50 |
| Air Blank          | 0.000 | 21:51 |
| Subject Sample # 2 | 0.114 | 21:52 |
| Breath Volume      | 2.046 | 21:52 |
| Air Blank          | 0.000 | 21:53 |
| Control Test       | 0.079 | 21:53 |
| Air Blank          | 0.000 | 21:53 |
| Diagnostics Check  | OK    | 21:54 |

---

Data Download Date/Time: 06/05/2024 16:18

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JONES,MINDY,L  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: ALVA-ARIAS,K  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 05/08/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28423080A2  
DRY GAS EXPIRY DATE: 11/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 22:39  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:18

SUBJECT'S LAST NAME: PARDO

DATE/TIME OF TEST: 05/19/2024 23:45:31

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:38 |
| Air Blank          | 0.000 | 23:39 |
| Control Test       | 0.080 | 23:39 |
| Air Blank          | 0.000 | 23:39 |
| Subject Sample # 1 | 0.036 | 23:40 |
| Breath Volume      | 1.246 | 23:40 |
| Air Blank          | 0.000 | 23:41 |
| Air Blank          | 0.000 | 23:43 |
| Subject Sample # 2 | 0.035 | 23:43 |
| Breath Volume      | 1.304 | 23:43 |
| Air Blank          | 0.000 | 23:44 |
| Control Test       | 0.079 | 23:44 |
| Air Blank          | 0.000 | 23:45 |
| Diagnostics Check  | OK    | 23:45 |

---

Data Download Date/Time: 06/05/2024 16:18

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KROME,BRITTNEY,M  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: LANDRY,G  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 05/08/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28423080A2  
DRY GAS EXPIRY DATE: 11/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 01:23  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:46

SUBJECT'S LAST NAME: EBERHARDT

DATE/TIME OF TEST: 05/20/2024 02:12:58

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:06 |
| Air Blank          | 0.000 | 02:06 |
| Control Test       | 0.079 | 02:07 |
| Air Blank          | 0.000 | 02:07 |
| Subject Sample # 1 | 0.157 | 02:08 |
| Breath Volume      | 2.250 | 02:08 |
| Air Blank          | 0.000 | 02:09 |
| Air Blank          | 0.000 | 02:10 |
| Subject Sample # 2 | 0.167 | 02:11 |
| Breath Volume      | 2.324 | 02:11 |
| Air Blank          | 0.000 | 02:12 |
| Control Test       | 0.079 | 02:12 |
| Air Blank          | 0.000 | 02:12 |
| Diagnostics Check  | OK    | 02:12 |

---

Data Download Date/Time: 06/05/2024 16:18

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: FRUEH, D  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 05/08/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28423080A2  
DRY GAS EXPIRY DATE: 11/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 02:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: JACKSON III

DATE/TIME OF TEST: 05/22/2024 02:51:52

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:45 |
| Air Blank          | 0.000 | 02:45 |
| Control Test       | 0.081 | 02:46 |
| Air Blank          | 0.000 | 02:46 |
| Subject Sample # 1 | 0.078 | 02:47 |
| Breath Volume      | 2.617 | 02:47 |
| Air Blank          | 0.000 | 02:48 |
| Air Blank          | 0.000 | 02:49 |
| Subject Sample # 2 | 0.077 | 02:50 |
| Breath Volume      | 3.769 | 02:50 |
| Air Blank          | 0.000 | 02:50 |
| Control Test       | 0.079 | 02:51 |
| Air Blank          | 0.000 | 02:51 |
| Diagnostics Check  | OK    | 02:51 |

---

Data Download Date/Time: 06/05/2024 16:18

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ORTEGA,JESSICA,M  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: CORRAO,M  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 05/08/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28423080A2  
DRY GAS EXPIRY DATE: 11/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 10:14  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 11:19

SUBJECT'S LAST NAME: WATKINS

DATE/TIME OF TEST: 05/22/2024 11:46:37

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 11:39 |
| Air Blank          | 0.000 | 11:40 |
| Control Test       | 0.080 | 11:40 |
| Air Blank          | 0.000 | 11:41 |
| Subject Sample # 1 | 0.245 | 11:41 |
| Breath Volume      | 2.453 | 11:41 |
| Air Blank          | 0.000 | 11:42 |
| Air Blank          | 0.000 | 11:44 |
| Subject Sample # 2 | 0.257 | 11:44 |
| Breath Volume      | 1.773 | 11:44 |
| Air Blank          | 0.000 | 11:45 |
| Control Test       | 0.079 | 11:45 |
| Air Blank          | 0.000 | 11:46 |
| Diagnostics Check  | OK    | 11:46 |

---

Data Download Date/Time: 06/05/2024 16:18

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: SNAGGS, S  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 05/08/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28423080A2  
DRY GAS EXPIRY DATE: 11/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 19:18  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:45

SUBJECT'S LAST NAME: COMNES

DATE/TIME OF TEST: 05/22/2024 20:11:58

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 20:05 |
| Air Blank          | 0.000 | 20:06 |
| Control Test       | 0.081 | 20:06 |
| Air Blank          | 0.000 | 20:06 |
| Subject Sample # 1 | 0.000 | 20:07 |
| Breath Volume      | 2.000 | 20:07 |
| Air Blank          | 0.000 | 20:07 |
| Air Blank          | 0.000 | 20:09 |
| Subject Sample # 2 | 0.000 | 20:10 |
| Breath Volume      | 1.792 | 20:10 |
| Air Blank          | 0.000 | 20:11 |
| Control Test       | 0.080 | 20:11 |
| Air Blank          | 0.000 | 20:11 |
| Diagnostics Check  | OK    | 20:11 |

---

Data Download Date/Time: 06/05/2024 16:18

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BADEN,T  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 05/08/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28423080A2  
DRY GAS EXPIRY DATE: 11/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 03:26  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:47

SUBJECT'S LAST NAME: ELEABY II

DATE/TIME OF TEST: 05/23/2024 04:14:43

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:07 |
| Air Blank          | 0.000 | 04:08 |
| Control Test       | 0.081 | 04:08 |
| Air Blank          | 0.000 | 04:09 |
| Subject Sample # 1 | 0.178 | 04:10 |
| Breath Volume      | 1.605 | 04:10 |
| Air Blank          | 0.000 | 04:10 |
| Air Blank          | 0.000 | 04:12 |
| Subject Sample # 2 | 0.176 | 04:13 |
| Breath Volume      | 3.695 | 04:13 |
| Air Blank          | 0.000 | 04:13 |
| Control Test       | 0.080 | 04:14 |
| Air Blank          | 0.000 | 04:14 |
| Diagnostics Check  | OK    | 04:14 |

---

Data Download Date/Time: 06/05/2024 16:18

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KROME,BRITTNEY,M  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: PORTER,S  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 05/08/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28423080A2  
DRY GAS EXPIRY DATE: 11/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 18:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:28

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 05/23/2024 19:55:08

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 19:48 |
| Air Blank          | 0.000 | 19:49 |
| Control Test       | 0.081 | 19:49 |
| Air Blank          | 0.000 | 19:50 |
| Subject Sample # 1 | 0.173 | 19:50 |
| Breath Volume      | 2.218 | 19:50 |
| Air Blank          | 0.000 | 19:51 |
| Air Blank          | 0.000 | 19:53 |
| Subject Sample # 2 | 0.164 | 19:53 |
| Breath Volume      | 2.179 | 19:53 |
| Air Blank          | 0.000 | 19:54 |
| Control Test       | 0.079 | 19:54 |
| Air Blank          | 0.000 | 19:55 |
| Diagnostics Check  | OK    | 19:55 |

---

Data Download Date/Time: 06/05/2024 16:18

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JONES,MINDY,L  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: SNAGGS,S  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 05/08/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28423080A2  
DRY GAS EXPIRY DATE: 11/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 21:51  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:19

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 05/23/2024 22:47:59

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:39 |
| Air Blank          | 0.000 | 22:40 |
| Control Test       | 0.081 | 22:40 |
| Air Blank          | 0.000 | 22:40 |
| Subject Sample # 1 | 0.138 | 22:43 |
| Breath Volume      | 3.652 | 22:43 |
| Air Blank          | 0.000 | 22:44 |
| Air Blank          | 0.000 | 22:45 |
| Subject Sample # 2 | 0.128 | 22:46 |
| Breath Volume      | 2.328 | 22:46 |
| Air Blank          | 0.000 | 22:47 |
| Control Test       | 0.079 | 22:47 |
| Air Blank          | 0.000 | 22:47 |
| Diagnostics Check  | OK    | 22:47 |

---

Data Download Date/Time: 06/05/2024 16:18

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JONES,MINDY,L  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: TELLEZ-ALFARO,D  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 05/08/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28423080A2  
DRY GAS EXPIRY DATE: 11/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 02:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: CERRO TETHU

DATE/TIME OF TEST: 05/24/2024 03:05:44

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:58 |
| Air Blank          | 0.000 | 02:59 |
| Control Test       | 0.081 | 02:59 |
| Air Blank          | 0.000 | 02:59 |
| Subject Sample # 1 | 0.180 | 03:01 |
| Breath Volume      | 2.441 | 03:01 |
| Air Blank          | 0.000 | 03:01 |
| Air Blank          | 0.000 | 03:03 |
| Subject Sample # 2 | 0.186 | 03:04 |
| Breath Volume      | 2.250 | 03:04 |
| Air Blank          | 0.000 | 03:04 |
| Control Test       | 0.079 | 03:05 |
| Air Blank          | 0.000 | 03:05 |
| Diagnostics Check  | OK    | 03:05 |

---

Data Download Date/Time: 06/05/2024 16:18

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PIERCEY,JANICE,F  
BREATH TEST OPERATOR AGENCY: Tampa PD  
ARRESTING OFFICER: BADEN,T  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 05/08/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28423080A2  
DRY GAS EXPIRY DATE: 11/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 01:57  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:21

SUBJECT'S LAST NAME: EGUES LEON

DATE/TIME OF TEST: 05/25/2024 02:52:31

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:44 |
| Air Blank          | 0.000 | 02:45 |
| Control Test       | 0.080 | 02:45 |
| Air Blank          | 0.000 | 02:46 |
| Subject Sample # 1 | 0.241 | 02:47 |
| Breath Volume      | 2.347 | 02:47 |
| Air Blank          | 0.000 | 02:47 |
| Air Blank          | 0.000 | 02:49 |
| Subject Sample # 2 | 0.231 | 02:50 |
| Breath Volume      | 1.695 | 02:50 |
| Air Blank          | 0.000 | 02:51 |
| Control Test       | 0.079 | 02:51 |
| Air Blank          | 0.000 | 02:52 |
| Diagnostics Check  | OK    | 02:52 |

---

Data Download Date/Time: 06/05/2024 16:18

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PIERCEY,JANICE,F  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: CLARK,I  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 05/08/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28423080A2  
DRY GAS EXPIRY DATE: 11/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 04:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:10

SUBJECT'S LAST NAME: WALTHOUR

DATE/TIME OF TEST: 05/25/2024 05:38:09

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 05:30 |
| Air Blank          | 0.000 | 05:30 |
| Control Test       | 0.079 | 05:31 |
| Air Blank          | 0.000 | 05:31 |
| Subject Sample # 1 | 0.112 | 05:32 |
| Breath Volume      | 2.542 | 05:32 |
| Air Blank          | 0.000 | 05:33 |
| Air Blank          | 0.000 | 05:34 |
| Subject Sample # 2 | 0.123 | 05:36 |
| Breath Volume      | 2.503 | 05:36 |
| Air Blank          | 0.000 | 05:37 |
| Control Test       | 0.079 | 05:37 |
| Air Blank          | 0.000 | 05:38 |
| Diagnostics Check  | OK    | 05:38 |

---

Data Download Date/Time: 06/05/2024 16:18

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: FRUEH,D  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 05/08/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28423080A2  
DRY GAS EXPIRY DATE: 11/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 04:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:22

SUBJECT'S LAST NAME: QUINONES

DATE/TIME OF TEST: 05/26/2024 04:49:56

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:42 |
| Air Blank          | 0.000 | 04:43 |
| Control Test       | 0.081 | 04:43 |
| Air Blank          | 0.000 | 04:43 |
| Subject Sample # 1 | 0.118 | 04:45 |
| Breath Volume      | 3.472 | 04:45 |
| Air Blank          | 0.000 | 04:45 |
| Air Blank          | 0.000 | 04:47 |
| Subject Sample # 2 | 0.124 | 04:48 |
| Breath Volume      | 2.324 | 04:48 |
| Air Blank          | 0.000 | 04:49 |
| Control Test       | 0.079 | 04:49 |
| Air Blank          | 0.000 | 04:49 |
| Diagnostics Check  | OK    | 04:49 |

---

Data Download Date/Time: 06/05/2024 16:18

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: JOHNSON, B  
ARRESTING OFFICER AGENCY: Macdill Air Force Ba  
LAST AGENCY INSPECTION: 05/08/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28423080A2  
DRY GAS EXPIRY DATE: 11/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 01:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: HILL

DATE/TIME OF TEST: 05/27/2024 02:13:32

---

|                   |       |       |            |
|-------------------|-------|-------|------------|
| Diagnostics Check | OK    | 02:10 |            |
| Air Blank         | 0.000 | 02:11 |            |
| Control Test      | 0.081 | 02:11 |            |
| Air Blank         | 0.000 | 02:11 |            |
| Subject Sample    | # 1   | 02:12 | RFI Detect |
| Breath Volume     | 0.000 | 02:12 |            |
| Air Blank         | 0.000 | 02:12 |            |
| Control Test      | 0.080 | 02:12 |            |
| Air Blank         | 0.000 | 02:13 |            |
| Diagnostics Check | OK    | 02:13 |            |

---

Data Download Date/Time: 06/05/2024 16:18

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: JOHNSON, B  
ARRESTING OFFICER AGENCY: Macdill Air Force Ba  
LAST AGENCY INSPECTION: 05/08/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28423080A2  
DRY GAS EXPIRY DATE: 11/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 01:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: HILL

DATE/TIME OF TEST: 05/27/2024 02:22:17

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:15 |
| Air Blank          | 0.000 | 02:15 |
| Control Test       | 0.081 | 02:16 |
| Air Blank          | 0.000 | 02:16 |
| Subject Sample # 1 | 0.189 | 02:17 |
| Breath Volume      | 2.835 | 02:17 |
| Air Blank          | 0.000 | 02:18 |
| Air Blank          | 0.000 | 02:20 |
| Subject Sample # 2 | 0.174 | 02:20 |
| Breath Volume      | 1.609 | 02:20 |
| Air Blank          | 0.000 | 02:21 |
| Control Test       | 0.079 | 02:21 |
| Air Blank          | 0.000 | 02:22 |
| Diagnostics Check  | OK    | 02:22 |

---

Data Download Date/Time: 06/05/2024 16:18

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: HERNANDEZ,S  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 05/08/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28423080A2  
DRY GAS EXPIRY DATE: 11/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 05:08  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:40

SUBJECT'S LAST NAME: MENA HERNANDEZ

DATE/TIME OF TEST: 05/27/2024 06:18:13

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 06:11 |
| Air Blank          | 0.000 | 06:12 |
| Control Test       | 0.080 | 06:12 |
| Air Blank          | 0.000 | 06:12 |
| Subject Sample # 1 | 0.150 | 06:13 |
| Breath Volume      | 2.195 | 06:13 |
| Air Blank          | 0.000 | 06:14 |
| Air Blank          | 0.000 | 06:15 |
| Subject Sample # 2 | 0.152 | 06:16 |
| Breath Volume      | 2.234 | 06:16 |
| Air Blank          | 0.000 | 06:17 |
| Control Test       | 0.079 | 06:17 |
| Air Blank          | 0.000 | 06:18 |
| Diagnostics Check  | OK    | 06:18 |

---

Data Download Date/Time: 06/05/2024 16:18

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PIERCEY,JANICE,F  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: THORSEN,A  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 05/08/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28423080A2  
DRY GAS EXPIRY DATE: 11/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 20:27  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: HUDSON

DATE/TIME OF TEST: 05/27/2024 22:03:05

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 21:55 |
| Air Blank          | 0.000 | 21:56 |
| Control Test       | 0.080 | 21:56 |
| Air Blank          | 0.000 | 21:57 |
| Subject Sample # 1 | 0.133 | 21:58 |
| Breath Volume      | 1.339 | 21:58 |
| Air Blank          | 0.000 | 21:59 |
| Air Blank          | 0.000 | 22:01 |
| Subject Sample # 2 | 0.120 | 22:01 |
| Breath Volume      | 1.441 | 22:01 |
| Air Blank          | 0.000 | 22:02 |
| Control Test       | 0.080 | 22:02 |
| Air Blank          | 0.000 | 22:03 |
| Diagnostics Check  | OK    | 22:03 |

---

Data Download Date/Time: 06/05/2024 16:18

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: PORTER,S  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 05/08/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28423080A2  
DRY GAS EXPIRY DATE: 11/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 23:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:02

SUBJECT'S LAST NAME: CORTEZ

DATE/TIME OF TEST: 05/28/2024 00:37:03

|                   |           |       |                |
|-------------------|-----------|-------|----------------|
| Diagnostics Check | OK        | 00:22 |                |
| Air Blank         | 0.000     | 00:22 |                |
| Control Test      | 0.081     | 00:23 |                |
| Air Blank         | 0.000     | 00:23 |                |
| Subject Sample    | # 1 0.175 | 00:26 |                |
| Breath Volume     | 1.406     | 00:26 |                |
| Air Blank         | 0.000     | 00:26 |                |
| Air Blank         | 0.000     | 00:28 |                |
| Subject Sample    | # 2 0.160 | 00:31 | Volume Not Met |
| Breath Volume     | 0.941     | 00:31 |                |
| Air Blank         | 0.000     | 00:32 |                |
| Air Blank         | 0.000     | 00:34 |                |
| Subject Sample    | # 3 0.184 | 00:35 |                |
| Breath Volume     | 1.238     | 00:35 |                |
| Air Blank         | 0.000     | 00:36 |                |
| Control Test      | 0.077     | 00:36 |                |
| Air Blank         | 0.000     | 00:36 |                |
| Diagnostics Check | OK        | 00:37 |                |

Data Download Date/Time: 06/05/2024 16:19

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: ALVA ARIAS,K  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 05/08/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28423080A2  
DRY GAS EXPIRY DATE: 11/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 23:33  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: RIVERA SANCHEZ

DATE/TIME OF TEST: 05/28/2024 00:49:18

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:42 |
| Air Blank          | 0.000 | 00:42 |
| Control Test       | 0.078 | 00:43 |
| Air Blank          | 0.000 | 00:43 |
| Subject Sample # 1 | 0.223 | 00:44 |
| Breath Volume      | 2.042 | 00:44 |
| Air Blank          | 0.000 | 00:45 |
| Air Blank          | 0.000 | 00:47 |
| Subject Sample # 2 | 0.227 | 00:47 |
| Breath Volume      | 2.433 | 00:47 |
| Air Blank          | 0.000 | 00:48 |
| Control Test       | 0.078 | 00:48 |
| Air Blank          | 0.000 | 00:49 |
| Diagnostics Check  | OK    | 00:49 |

---

Data Download Date/Time: 06/05/2024 16:19

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JONES,MINDY,L  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: CARRASCO,A  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 05/08/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28423080A2  
DRY GAS EXPIRY DATE: 11/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 01:39  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 05/30/2024 02:40:02

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:33 |
| Air Blank          | 0.000 | 02:34 |
| Control Test       | 0.081 | 02:34 |
| Air Blank          | 0.000 | 02:35 |
| Subject Sample # 1 | 0.108 | 02:35 |
| Breath Volume      | 2.933 | 02:35 |
| Air Blank          | 0.000 | 02:36 |
| Air Blank          | 0.000 | 02:38 |
| Subject Sample # 2 | 0.107 | 02:38 |
| Breath Volume      | 2.976 | 02:38 |
| Air Blank          | 0.000 | 02:39 |
| Control Test       | 0.079 | 02:39 |
| Air Blank          | 0.000 | 02:39 |
| Diagnostics Check  | OK    | 02:40 |

---

Data Download Date/Time: 06/05/2024 16:19

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: EARLYWINE,G  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 05/08/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28423080A2  
DRY GAS EXPIRY DATE: 11/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 01:07  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:33

SUBJECT'S LAST NAME: PEACOCK

DATE/TIME OF TEST: 05/31/2024 02:01:58

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:54 |
| Air Blank          | 0.000 | 01:55 |
| Control Test       | 0.081 | 01:55 |
| Air Blank          | 0.000 | 01:55 |
| Subject Sample # 1 | 0.063 | 01:57 |
| Breath Volume      | 2.261 | 01:57 |
| Air Blank          | 0.000 | 01:58 |
| Air Blank          | 0.000 | 01:59 |
| Subject Sample # 2 | 0.070 | 02:00 |
| Breath Volume      | 2.097 | 02:00 |
| Air Blank          | 0.000 | 02:01 |
| Control Test       | 0.080 | 02:01 |
| Air Blank          | 0.000 | 02:01 |
| Diagnostics Check  | OK    | 02:01 |

---

Data Download Date/Time: 06/05/2024 16:19

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PIERCEY,JANICE,F  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: ALVA-ARIAS,K  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 05/08/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28423080A2  
DRY GAS EXPIRY DATE: 11/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 21:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: AGUILAR PEREZ

DATE/TIME OF TEST: 05/31/2024 22:58:53

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:50 |
| Air Blank          | 0.000 | 22:51 |
| Control Test       | 0.081 | 22:51 |
| Air Blank          | 0.000 | 22:52 |
| Subject Sample # 1 | 0.156 | 22:54 |
| Breath Volume      | 1.531 | 22:54 |
| Air Blank          | 0.000 | 22:54 |
| Air Blank          | 0.000 | 22:56 |
| Subject Sample # 2 | 0.145 | 22:57 |
| Breath Volume      | 1.457 | 22:57 |
| Air Blank          | 0.000 | 22:57 |
| Control Test       | 0.079 | 22:58 |
| Air Blank          | 0.000 | 22:58 |
| Diagnostics Check  | OK    | 22:58 |

---

Data Download Date/Time: 06/05/2024 16:19

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L  
BREATH TEST OPERATOR AGENCY: Martin County SO  
ARRESTING OFFICER: THOMAS,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Martin County SO  
LAST AGENCY INSPECTION: 04/20/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15922080A5  
DRY GAS EXPIRY DATE: 08/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 21:27  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:17

SUBJECT'S LAST NAME: SLOVAK

DATE/TIME OF TEST: 05/05/2024 21:44:38

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 21:37 |
| Air Blank          | 0.000 | 21:38 |
| Control Test       | 0.079 | 21:38 |
| Air Blank          | 0.000 | 21:39 |
| Subject Sample # 1 | 0.074 | 21:39 |
| Breath Volume      | 1.132 | 21:39 |
| Air Blank          | 0.000 | 21:40 |
| Air Blank          | 0.000 | 21:42 |
| Subject Sample # 2 | 0.070 | 21:43 |
| Breath Volume      | 1.691 | 21:43 |
| Air Blank          | 0.000 | 21:43 |
| Control Test       | 0.080 | 21:44 |
| Air Blank          | 0.000 | 21:44 |
| Diagnostics Check  | OK    | 21:44 |

---

Data Download Date/Time: 05/21/2024 00:33

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L  
BREATH TEST OPERATOR AGENCY: Martin County SO  
ARRESTING OFFICER: THOMAS,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Martin County SO  
LAST AGENCY INSPECTION: 05/20/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15922080A5  
DRY GAS EXPIRY DATE: 08/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 19:02  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 19:36

SUBJECT'S LAST NAME: FOSTER

DATE/TIME OF TEST: 05/23/2024 20:19:12

---

|                    |       |       |              |
|--------------------|-------|-------|--------------|
| Diagnostics Check  | OK    | 20:13 |              |
| Air Blank          | 0.000 | 20:14 |              |
| Control Test       | 0.076 | 20:14 |              |
| Air Blank          | 0.000 | 20:15 |              |
| Subject Sample # 1 | 0.295 | 20:16 |              |
| Breath Volume      | 1.425 | 20:16 |              |
| Air Blank          | 0.000 | 20:16 |              |
| Air Blank          |       | 20:18 | Ambient Fail |
| Air Blank          | 0.000 | 20:19 |              |

---

Data Download Date/Time: 06/25/2024 14:43

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L  
BREATH TEST OPERATOR AGENCY: Martin County SO  
ARRESTING OFFICER: THOMAS,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Martin County SO  
LAST AGENCY INSPECTION: 05/20/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15922080A5  
DRY GAS EXPIRY DATE: 08/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 19:02  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 19:36

SUBJECT'S LAST NAME: FOSTER

DATE/TIME OF TEST: 05/23/2024 20:28:34

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 20:20 |
| Air Blank          | 0.000 | 20:20 |
| Control Test       | 0.076 | 20:21 |
| Air Blank          | 0.000 | 20:21 |
| Subject Sample # 1 | 0.292 | 20:23 |
| Breath Volume      | 1.578 | 20:23 |
| Air Blank          | 0.000 | 20:24 |
| Air Blank          | 0.000 | 20:25 |
| Subject Sample # 2 | 0.285 | 20:26 |
| Breath Volume      | 1.355 | 20:26 |
| Air Blank          | 0.000 | 20:27 |
| Control Test       | 0.075 | 20:27 |
| Air Blank          | 0.000 | 20:28 |
| Diagnostics Check  | OK    | 20:28 |

---

Data Download Date/Time: 06/25/2024 14:43

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILKES,JOHN,D  
BREATH TEST OPERATOR AGENCY: Martin County SO  
ARRESTING OFFICER: DELANEY,ALLISON  
ARRESTING OFFICER AGENCY: FFWCC (West Palm)  
LAST AGENCY INSPECTION: 04/20/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15922080A5  
DRY GAS EXPIRY DATE: 08/05/2024  
VIOLATION CODE: BUI  
ARREST TIME: 18:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:30

SUBJECT'S LAST NAME: PANACCIONE

DATE/TIME OF TEST: 05/11/2024 19:17:24

---

|                   |       |       |            |
|-------------------|-------|-------|------------|
| Diagnostics Check | OK    | 19:15 |            |
| Air Blank         | 0.000 | 19:16 |            |
| Control Test      | 0.080 | 19:16 |            |
| Air Blank         | 0.000 | 19:17 |            |
|                   |       | 19:17 | RFI Detect |
| Air Blank         |       | 19:17 | RFI Detect |

---

Data Download Date/Time: 05/21/2024 00:51

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILKES,JOHN,D  
BREATH TEST OPERATOR AGENCY: Martin County SO  
ARRESTING OFFICER: DELANEY,ALLISON  
ARRESTING OFFICER AGENCY: FFWCC (West Palm)  
LAST AGENCY INSPECTION: 04/20/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15922080A5  
DRY GAS EXPIRY DATE: 08/05/2024  
VIOLATION CODE: BUI  
ARREST TIME: 18:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:30

SUBJECT'S LAST NAME: PANACCIONE

DATE/TIME OF TEST: 05/11/2024 19:25:49

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 19:19 |
| Air Blank          | 0.000 | 19:19 |
| Control Test       | 0.079 | 19:20 |
| Air Blank          | 0.000 | 19:20 |
| Subject Sample # 1 | 0.192 | 19:21 |
| Breath Volume      | 1.863 | 19:21 |
| Air Blank          | 0.000 | 19:21 |
| Air Blank          | 0.000 | 19:23 |
| Subject Sample # 2 | 0.190 | 19:24 |
| Breath Volume      | 2.113 | 19:24 |
| Air Blank          | 0.000 | 19:24 |
| Control Test       | 0.079 | 19:25 |
| Air Blank          | 0.000 | 19:25 |
| Diagnostics Check  | OK    | 19:25 |

---

Data Download Date/Time: 05/21/2024 00:51

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SPIRITO,GEORGE,W  
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)  
ARRESTING OFFICER: SPIRITO,GEORGE  
ARRESTING OFFICER AGENCY: FFWCC (West Palm)  
LAST AGENCY INSPECTION: 04/20/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15922080A5  
DRY GAS EXPIRY DATE: 08/05/2024  
VIOLATION CODE: BUI  
ARREST TIME: 13:11  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 13:05

SUBJECT'S LAST NAME: ASH

DATE/TIME OF TEST: 05/17/2024 13:43:24

---

|                   |       |       |            |
|-------------------|-------|-------|------------|
| Diagnostics Check | OK    | 13:42 |            |
| Air Blank         | 0.000 | 13:42 |            |
| Control Test      | 0.079 | 13:42 |            |
| Air Blank         |       | 13:43 | RFI Detect |
| Air Blank         |       | 13:43 | RFI Detect |

---

Data Download Date/Time: 05/21/2024 00:51

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SPIRITO,GEORGE,W  
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)  
ARRESTING OFFICER: SPIRITO,GEORGE  
ARRESTING OFFICER AGENCY: FFWCC (West Palm)  
LAST AGENCY INSPECTION: 04/20/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15922080A5  
DRY GAS EXPIRY DATE: 08/05/2024  
VIOLATION CODE: BUI  
ARREST TIME: 13:11  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 13:01

SUBJECT'S LAST NAME: ASH

DATE/TIME OF TEST: 05/17/2024 13:52:19

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 13:45 |
| Air Blank          | 0.000 | 13:46 |
| Control Test       | 0.079 | 13:46 |
| Air Blank          | 0.000 | 13:46 |
| Subject Sample # 1 | 0.054 | 13:47 |
| Breath Volume      | 3.187 | 13:47 |
| Air Blank          | 0.000 | 13:48 |
| Air Blank          | 0.000 | 13:50 |
| Subject Sample # 2 | 0.051 | 13:50 |
| Breath Volume      | 1.750 | 13:50 |
| Air Blank          | 0.000 | 13:51 |
| Control Test       | 0.079 | 13:51 |
| Air Blank          | 0.000 | 13:52 |
| Diagnostics Check  | OK    | 13:52 |

---

Data Download Date/Time: 05/21/2024 00:51

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000834

INSTRUMENT REGISTERED TO: GULF COAST STATE COL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MOLINA,JOSE,A  
BREATH TEST OPERATOR AGENCY: FHP Troop E  
ARRESTING OFFICER: MOLINA,JOSE  
ARRESTING OFFICER AGENCY: FHP Troop E  
LAST AGENCY INSPECTION: 05/06/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 00919080A2  
DRY GAS EXPIRY DATE: 03/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 11:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 12:30

SUBJECT'S LAST NAME: WHITTEN

DATE/TIME OF TEST: 05/06/2024 13:08:06

---

|                   |       |       |                 |
|-------------------|-------|-------|-----------------|
| Diagnostics Check | OK    | 13:05 |                 |
| Air Blank         | 0.000 | 13:05 |                 |
| Control Test      | 0.079 | 13:06 |                 |
| Air Blank         | 0.000 | 13:06 |                 |
| Subject Sample    | # 1   | 13:07 | Slope Not Met   |
| Breath Volume     | 1.539 | 13:07 |                 |
| Air Blank         |       | 13:07 | Improper Sample |
| Air Blank         | 0.000 | 13:08 |                 |

---

Data Download Date/Time: 10/16/2024 15:14

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000834

INSTRUMENT REGISTERED TO: GULF COAST STATE COL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MOLINA,JOSE,A  
BREATH TEST OPERATOR AGENCY: FHP Troop E  
ARRESTING OFFICER: MOLINA,JOSE  
ARRESTING OFFICER AGENCY: FHP Troop E  
LAST AGENCY INSPECTION: 05/06/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 00919080A2  
DRY GAS EXPIRY DATE: 03/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 11:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 12:30

SUBJECT'S LAST NAME: WHITTEN

DATE/TIME OF TEST: 05/06/2024 13:17:08

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 13:11 |
| Air Blank          | 0.000 | 13:11 |
| Control Test       | 0.078 | 13:12 |
| Air Blank          | 0.000 | 13:12 |
| Subject Sample # 1 | 0.000 | 13:12 |
| Breath Volume      | 2.488 | 13:12 |
| Air Blank          | 0.000 | 13:13 |
| Air Blank          | 0.000 | 13:15 |
| Subject Sample # 2 | 0.000 | 13:15 |
| Breath Volume      | 2.523 | 13:15 |
| Air Blank          | 0.000 | 13:16 |
| Control Test       | 0.079 | 13:16 |
| Air Blank          | 0.000 | 13:17 |
| Diagnostics Check  | OK    | 13:17 |

---

Data Download Date/Time: 10/16/2024 15:14

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000834

INSTRUMENT REGISTERED TO: GULF COAST STATE COL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ARIAS,MANUEL,S  
BREATH TEST OPERATOR AGENCY: FHP Troop E  
ARRESTING OFFICER: ARIAS,MANUEL  
ARRESTING OFFICER AGENCY: FHP Troop E  
LAST AGENCY INSPECTION: 05/06/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 00919080A2  
DRY GAS EXPIRY DATE: 03/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 11:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 12:30

SUBJECT'S LAST NAME: RONTERIUS

DATE/TIME OF TEST: 05/06/2024 13:47:55

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 13:24 |
| Air Blank          | 0.000 | 13:25 |
| Control Test       | 0.078 | 13:25 |
| Air Blank          | 0.000 | 13:25 |
| Subject Sample # 1 | 0.000 | 13:26 |
| Breath Volume      | 2.671 | 13:26 |
| Air Blank          | 0.000 | 13:26 |
| Air Blank          | 0.000 | 13:28 |
| Subject Sample # 2 | 0.000 | 13:29 |
| Breath Volume      | 3.281 | 13:29 |
| Air Blank          | 0.000 | 13:29 |
| Control Test       | 0.082 | 13:30 |
| Air Blank          | 0.000 | 13:30 |
| Diagnostics Check  | OK    | 13:47 |

---

Data Download Date/Time: 10/16/2024 15:14

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000834

INSTRUMENT REGISTERED TO: GULF COAST STATE COL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ELLISON,DAVID,J  
BREATH TEST OPERATOR AGENCY: FHP Troop E  
ARRESTING OFFICER: ELLISON,DAVID  
ARRESTING OFFICER AGENCY: FHP Troop E  
LAST AGENCY INSPECTION: 05/06/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 00919080A2  
DRY GAS EXPIRY DATE: 03/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 10:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 11:30

SUBJECT'S LAST NAME: HUBBARD

DATE/TIME OF TEST: 05/06/2024 14:07:32

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 14:01 |
| Air Blank          | 0.000 | 14:02 |
| Control Test       | 0.078 | 14:02 |
| Air Blank          | 0.000 | 14:02 |
| Subject Sample # 1 | 0.007 | 14:03 |
| Breath Volume      | 3.343 | 14:03 |
| Air Blank          | 0.000 | 14:03 |
| Air Blank          | 0.000 | 14:05 |
| Subject Sample # 2 | 0.000 | 14:06 |
| Breath Volume      | 3.714 | 14:06 |
| Air Blank          | 0.000 | 14:06 |
| Control Test       | 0.084 | 14:06 |
| Air Blank          | 0.000 | 14:07 |
| Diagnostics Check  | OK    | 14:07 |

---

Data Download Date/Time: 10/16/2024 15:14

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000834

INSTRUMENT REGISTERED TO: GULF COAST STATE COL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LUNA,ANGEL,  
BREATH TEST OPERATOR AGENCY: FHP Troop E  
ARRESTING OFFICER: LUNA,ANGEL  
ARRESTING OFFICER AGENCY: FHP Troop E  
LAST AGENCY INSPECTION: 05/06/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 00919080A2  
DRY GAS EXPIRY DATE: 03/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 11:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 11:00

SUBJECT'S LAST NAME: WHITTEN

DATE/TIME OF TEST: 05/06/2024 14:18:12

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 14:12 |
| Air Blank          | 0.000 | 14:12 |
| Control Test       | 0.078 | 14:12 |
| Air Blank          | 0.000 | 14:13 |
| Subject Sample # 1 | 0.000 | 14:13 |
| Breath Volume      | 2.621 | 14:13 |
| Air Blank          | 0.000 | 14:14 |
| Air Blank          | 0.000 | 14:16 |
| Subject Sample # 2 | 0.000 | 14:16 |
| Breath Volume      | 3.503 | 14:16 |
| Air Blank          | 0.000 | 14:17 |
| Control Test       | 0.081 | 14:17 |
| Air Blank          | 0.000 | 14:18 |
| Diagnostics Check  | OK    | 14:18 |

---

Data Download Date/Time: 10/16/2024 15:14

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000834

INSTRUMENT REGISTERED TO: GULF COAST STATE COL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DE LEON,ANTONIO,  
BREATH TEST OPERATOR AGENCY: FHP Troop E  
ARRESTING OFFICER: DE LEON,ANTONIO  
ARRESTING OFFICER AGENCY: FHP Troop E  
LAST AGENCY INSPECTION: 05/06/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 00919080A2  
DRY GAS EXPIRY DATE: 03/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 14:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 14:00

SUBJECT'S LAST NAME: WHITTEN

DATE/TIME OF TEST: 05/06/2024 14:29:12

---

|                    |       |       |                    |
|--------------------|-------|-------|--------------------|
| Diagnostics Check  | OK    | 14:22 |                    |
| Air Blank          | 0.000 | 14:23 |                    |
| Control Test       | 0.077 | 14:23 |                    |
| Air Blank          | 0.000 | 14:24 |                    |
| Subject Sample # 1 | 0.000 | 14:24 |                    |
| Breath Volume      | 1.910 | 14:24 |                    |
| Air Blank          | 0.000 | 14:25 |                    |
| Air Blank          | 0.000 | 14:27 |                    |
| Subject Sample # 2 | 0.000 | 14:27 |                    |
| Breath Volume      | 2.398 | 14:27 |                    |
| Air Blank          | 0.000 | 14:28 |                    |
| Control Test       | 0.085 | 14:28 | Interferent Detect |
| Air Blank          | 0.000 | 14:29 |                    |

---

Data Download Date/Time: 10/16/2024 15:14

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000834

INSTRUMENT REGISTERED TO: GULF COAST STATE COL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DE LEON,ANTONIO,  
BREATH TEST OPERATOR AGENCY: FHP Troop E  
ARRESTING OFFICER: DE LEON,ANTONIO  
ARRESTING OFFICER AGENCY: FHP Troop E  
LAST AGENCY INSPECTION: 05/06/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 00919080A2  
DRY GAS EXPIRY DATE: 03/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 14:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 14:00

SUBJECT'S LAST NAME: WHITTEN

DATE/TIME OF TEST: 05/06/2024 14:37:16

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 14:31 |
| Air Blank          | 0.000 | 14:31 |
| Control Test       | 0.077 | 14:31 |
| Air Blank          | 0.000 | 14:32 |
| Subject Sample # 1 | 0.000 | 14:32 |
| Breath Volume      | 2.507 | 14:32 |
| Air Blank          | 0.000 | 14:33 |
| Air Blank          | 0.000 | 14:35 |
| Subject Sample # 2 | 0.000 | 14:35 |
| Breath Volume      | 2.578 | 14:35 |
| Air Blank          | 0.000 | 14:36 |
| Control Test       | 0.080 | 14:36 |
| Air Blank          | 0.000 | 14:37 |
| Diagnostics Check  | OK    | 14:37 |

---

Data Download Date/Time: 10/16/2024 15:14

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000834

INSTRUMENT REGISTERED TO: GULF COAST STATE COL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HICKS,TERENCE,E  
BREATH TEST OPERATOR AGENCY: FHP Troop E  
ARRESTING OFFICER: HICKS,TERENCE  
ARRESTING OFFICER AGENCY: FHP Troop E  
LAST AGENCY INSPECTION: 05/06/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 00919080A2  
DRY GAS EXPIRY DATE: 03/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 11:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 12:30

SUBJECT'S LAST NAME: WHITTEN

DATE/TIME OF TEST: 05/06/2024 14:45:43

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 14:39 |
| Air Blank          | 0.000 | 14:40 |
| Control Test       | 0.078 | 14:40 |
| Air Blank          | 0.000 | 14:40 |
| Subject Sample # 1 | 0.000 | 14:41 |
| Breath Volume      | 1.726 | 14:41 |
| Air Blank          | 0.000 | 14:41 |
| Air Blank          | 0.000 | 14:43 |
| Subject Sample # 2 | 0.000 | 14:44 |
| Breath Volume      | 2.710 | 14:44 |
| Air Blank          | 0.000 | 14:44 |
| Control Test       | 0.084 | 14:45 |
| Air Blank          | 0.000 | 14:45 |
| Diagnostics Check  | OK    | 14:45 |

---

Data Download Date/Time: 10/16/2024 15:14

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000834

INSTRUMENT REGISTERED TO: GULF COAST STATE COL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GREISHEIMER,JOHN,V  
BREATH TEST OPERATOR AGENCY: FHP Troop E  
ARRESTING OFFICER: GREISHEIMER,JOHN  
ARRESTING OFFICER AGENCY: FHP Troop E  
LAST AGENCY INSPECTION: 05/06/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 00919080A2  
DRY GAS EXPIRY DATE: 03/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 13:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 14:00

SUBJECT'S LAST NAME: WHITTEN

DATE/TIME OF TEST: 05/06/2024 14:54:13

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 14:48 |
| Air Blank          | 0.000 | 14:48 |
| Control Test       | 0.077 | 14:48 |
| Air Blank          | 0.000 | 14:49 |
| Subject Sample # 1 | 0.000 | 14:49 |
| Breath Volume      | 2.644 | 14:49 |
| Air Blank          | 0.000 | 14:50 |
| Air Blank          | 0.000 | 14:52 |
| Subject Sample # 2 | 0.000 | 14:52 |
| Breath Volume      | 2.890 | 14:52 |
| Air Blank          | 0.000 | 14:53 |
| Control Test       | 0.081 | 14:53 |
| Air Blank          | 0.000 | 14:54 |
| Diagnostics Check  | OK    | 14:54 |

---

Data Download Date/Time: 10/16/2024 15:14

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000834

INSTRUMENT REGISTERED TO: GULF COAST STATE COL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PERDIGON,JACQUELINE,  
BREATH TEST OPERATOR AGENCY: South Miami PD  
ARRESTING OFFICER: PERDIGON,JACQUELINE  
ARRESTING OFFICER AGENCY: South Miami PD  
LAST AGENCY INSPECTION: 05/06/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 00919080A2  
DRY GAS EXPIRY DATE: 03/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 13:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 14:00

SUBJECT'S LAST NAME: WHITTEN

DATE/TIME OF TEST: 05/06/2024 15:03:22

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 14:57 |
| Air Blank          | 0.000 | 14:57 |
| Control Test       | 0.077 | 14:58 |
| Air Blank          | 0.000 | 14:58 |
| Subject Sample # 1 | 0.000 | 14:59 |
| Breath Volume      | 1.406 | 14:59 |
| Air Blank          | 0.000 | 14:59 |
| Air Blank          | 0.000 | 15:01 |
| Subject Sample # 2 | 0.000 | 15:01 |
| Breath Volume      | 1.406 | 15:01 |
| Air Blank          | 0.000 | 15:02 |
| Control Test       | 0.077 | 15:02 |
| Air Blank          | 0.000 | 15:03 |
| Diagnostics Check  | OK    | 15:03 |

---

Data Download Date/Time: 10/16/2024 15:14

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000834

INSTRUMENT REGISTERED TO: GULF COAST STATE COL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DIAZ,MISAEI,  
BREATH TEST OPERATOR AGENCY: FHP Troop E  
ARRESTING OFFICER: DIAZ,MISAEI  
ARRESTING OFFICER AGENCY: FHP Troop E  
LAST AGENCY INSPECTION: 05/06/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 00919080A2  
DRY GAS EXPIRY DATE: 03/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 13:02  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 14:02

SUBJECT'S LAST NAME: HUBBARD SR

DATE/TIME OF TEST: 05/06/2024 15:13:52

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 15:07 |
| Air Blank          | 0.000 | 15:08 |
| Control Test       | 0.078 | 15:08 |
| Air Blank          | 0.000 | 15:09 |
| Subject Sample # 1 | 0.000 | 15:09 |
| Breath Volume      | 2.230 | 15:09 |
| Air Blank          | 0.000 | 15:10 |
| Air Blank          | 0.000 | 15:12 |
| Subject Sample # 2 | 0.000 | 15:12 |
| Breath Volume      | 2.058 | 15:12 |
| Air Blank          | 0.000 | 15:12 |
| Control Test       | 0.079 | 15:13 |
| Air Blank          | 0.000 | 15:13 |
| Diagnostics Check  | OK    | 15:13 |

---

Data Download Date/Time: 10/16/2024 15:14

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000834

INSTRUMENT REGISTERED TO: GULF COAST STATE COL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FALCAO,THIAGO,  
BREATH TEST OPERATOR AGENCY: Columbia County SO  
ARRESTING OFFICER: FALCAO,THIAGO  
ARRESTING OFFICER AGENCY: Columbia County SO  
LAST AGENCY INSPECTION: 05/06/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 00919080A2  
DRY GAS EXPIRY DATE: 03/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 12:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 12:30

SUBJECT'S LAST NAME: WHITTEN

DATE/TIME OF TEST: 05/06/2024 15:22:19

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 15:16 |
| Air Blank          | 0.000 | 15:16 |
| Control Test       | 0.078 | 15:16 |
| Air Blank          | 0.000 | 15:17 |
| Subject Sample # 1 | 0.000 | 15:17 |
| Breath Volume      | 1.570 | 15:17 |
| Air Blank          | 0.000 | 15:18 |
| Air Blank          | 0.000 | 15:20 |
| Subject Sample # 2 | 0.000 | 15:20 |
| Breath Volume      | 2.601 | 15:20 |
| Air Blank          | 0.000 | 15:21 |
| Control Test       | 0.080 | 15:21 |
| Air Blank          | 0.000 | 15:22 |
| Diagnostics Check  | OK    | 15:22 |

---

Data Download Date/Time: 10/16/2024 15:14

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000834

INSTRUMENT REGISTERED TO: GULF COAST STATE COL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DIAZ,MISAEI,  
BREATH TEST OPERATOR AGENCY: FHP Troop E  
ARRESTING OFFICER: DIAZ,MISAEI  
ARRESTING OFFICER AGENCY: FHP Troop E  
LAST AGENCY INSPECTION: 05/06/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 00919080A2  
DRY GAS EXPIRY DATE: 03/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 13:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 12:46

SUBJECT'S LAST NAME: WHITTEN

DATE/TIME OF TEST: 05/06/2024 15:33:33

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 15:27 |
| Air Blank          | 0.000 | 15:27 |
| Control Test       | 0.077 | 15:27 |
| Air Blank          | 0.000 | 15:28 |
| Subject Sample # 1 | 0.000 | 15:29 |
| Breath Volume      | 1.957 | 15:29 |
| Air Blank          | 0.000 | 15:29 |
| Air Blank          | 0.000 | 15:31 |
| Subject Sample # 2 | 0.000 | 15:32 |
| Breath Volume      | 1.917 | 15:32 |
| Air Blank          | 0.000 | 15:32 |
| Control Test       | 0.079 | 15:32 |
| Air Blank          | 0.000 | 15:33 |
| Diagnostics Check  | OK    | 15:33 |

---

Data Download Date/Time: 10/16/2024 15:14

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000834

INSTRUMENT REGISTERED TO: GULF COAST STATE COL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WASHINGTON, RONALD, M  
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)  
ARRESTING OFFICER: WASHINGTON, RONALD  
ARRESTING OFFICER AGENCY: FFWCC (West Palm)  
LAST AGENCY INSPECTION: 05/06/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 00919080A2  
DRY GAS EXPIRY DATE: 03/05/2025  
VIOLATION CODE: BUI  
ARREST TIME: 14:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 15:00

SUBJECT'S LAST NAME: BREITENFIELD

DATE/TIME OF TEST: 05/06/2024 15:41:56

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 15:35 |
| Air Blank          | 0.000 | 15:36 |
| Control Test       | 0.078 | 15:36 |
| Air Blank          | 0.000 | 15:37 |
| Subject Sample # 1 | 0.000 | 15:37 |
| Breath Volume      | 1.449 | 15:37 |
| Air Blank          | 0.000 | 15:38 |
| Air Blank          | 0.000 | 15:40 |
| Subject Sample # 2 | 0.000 | 15:40 |
| Breath Volume      | 2.359 | 15:40 |
| Air Blank          | 0.000 | 15:41 |
| Control Test       | 0.079 | 15:41 |
| Air Blank          | 0.000 | 15:41 |
| Diagnostics Check  | OK    | 15:41 |

---

Data Download Date/Time: 10/16/2024 15:14

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000834

INSTRUMENT REGISTERED TO: GULF COAST STATE COL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MARIMON,GABRIEL,  
BREATH TEST OPERATOR AGENCY: FHP Troop E  
ARRESTING OFFICER: MARIMON,GABRIEL  
ARRESTING OFFICER AGENCY: FHP Troop E  
LAST AGENCY INSPECTION: 05/06/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 00919080A2  
DRY GAS EXPIRY DATE: 03/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 15:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 15:15

SUBJECT'S LAST NAME: BREITENFIELD

DATE/TIME OF TEST: 05/06/2024 15:52:37

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 15:46 |
| Air Blank          | 0.000 | 15:47 |
| Control Test       | 0.077 | 15:47 |
| Air Blank          | 0.000 | 15:47 |
| Subject Sample # 1 | 0.000 | 15:48 |
| Breath Volume      | 2.160 | 15:48 |
| Air Blank          | 0.000 | 15:48 |
| Air Blank          | 0.000 | 15:50 |
| Subject Sample # 2 | 0.000 | 15:51 |
| Breath Volume      | 2.132 | 15:51 |
| Air Blank          | 0.000 | 15:51 |
| Control Test       | 0.079 | 15:52 |
| Air Blank          | 0.000 | 15:52 |
| Diagnostics Check  | OK    | 15:52 |

---

Data Download Date/Time: 10/16/2024 15:14

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000834

INSTRUMENT REGISTERED TO: GULF COAST STATE COL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HICKS,TERENCE,E  
BREATH TEST OPERATOR AGENCY: FHP Troop E  
ARRESTING OFFICER: HICKS,TERENCE  
ARRESTING OFFICER AGENCY: FHP Troop E  
LAST AGENCY INSPECTION: 05/06/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 00919080A2  
DRY GAS EXPIRY DATE: 03/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 04:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:30

SUBJECT'S LAST NAME: BREITENFIELD

DATE/TIME OF TEST: 05/07/2024 08:33:35

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 08:27 |
| Air Blank          | 0.000 | 08:27 |
| Control Test       | 0.078 | 08:28 |
| Air Blank          | 0.000 | 08:28 |
| Subject Sample # 1 | 0.000 | 08:29 |
| Breath Volume      | 2.355 | 08:29 |
| Air Blank          | 0.000 | 08:29 |
| Air Blank          | 0.000 | 08:31 |
| Subject Sample # 2 | 0.000 | 08:32 |
| Breath Volume      | 1.890 | 08:32 |
| Air Blank          | 0.000 | 08:32 |
| Control Test       | 0.081 | 08:32 |
| Air Blank          | 0.000 | 08:33 |
| Diagnostics Check  | OK    | 08:33 |

---

Data Download Date/Time: 10/16/2024 15:15

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000834

INSTRUMENT REGISTERED TO: GULF COAST STATE COL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CLEARE,PAMELA,F  
BREATH TEST OPERATOR AGENCY: FHP Troop E  
ARRESTING OFFICER: CLEARE,PAMELA  
ARRESTING OFFICER AGENCY: FHP Troop E  
LAST AGENCY INSPECTION: 05/06/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 00919080A2  
DRY GAS EXPIRY DATE: 03/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 07:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 08:30

SUBJECT'S LAST NAME: WHITTEN

DATE/TIME OF TEST: 05/07/2024 08:56:57

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 08:51 |
| Air Blank          | 0.000 | 08:51 |
| Control Test       | 0.078 | 08:51 |
| Air Blank          | 0.000 | 08:52 |
| Subject Sample # 1 | 0.000 | 08:52 |
| Breath Volume      | 2.332 | 08:52 |
| Air Blank          | 0.000 | 08:53 |
| Air Blank          | 0.000 | 08:55 |
| Subject Sample # 2 | 0.000 | 08:55 |
| Breath Volume      | 2.121 | 08:55 |
| Air Blank          | 0.000 | 08:56 |
| Control Test       | 0.081 | 08:56 |
| Air Blank          | 0.000 | 08:56 |
| Diagnostics Check  | OK    | 08:56 |

---

Data Download Date/Time: 10/16/2024 15:15

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000834

INSTRUMENT REGISTERED TO: GULF COAST STATE COL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: OLDS,LEE,V  
BREATH TEST OPERATOR AGENCY: FHP Troop E  
ARRESTING OFFICER: OLDS,LEE  
ARRESTING OFFICER AGENCY: FHP Troop E  
LAST AGENCY INSPECTION: 05/06/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 00919080A2  
DRY GAS EXPIRY DATE: 03/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 07:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 08:30

SUBJECT'S LAST NAME: OLDS

DATE/TIME OF TEST: 05/07/2024 09:06:44

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 09:00 |
| Air Blank          | 0.000 | 09:01 |
| Control Test       | 0.078 | 09:01 |
| Air Blank          | 0.000 | 09:02 |
| Subject Sample # 1 | 0.000 | 09:02 |
| Breath Volume      | 2.445 | 09:02 |
| Air Blank          | 0.000 | 09:02 |
| Air Blank          | 0.000 | 09:04 |
| Subject Sample # 2 | 0.000 | 09:05 |
| Breath Volume      | 2.234 | 09:05 |
| Air Blank          | 0.000 | 09:05 |
| Control Test       | 0.080 | 09:06 |
| Air Blank          | 0.000 | 09:06 |
| Diagnostics Check  | OK    | 09:06 |

---

Data Download Date/Time: 10/16/2024 15:15

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000834

INSTRUMENT REGISTERED TO: GULF COAST STATE COL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CRUZ,RUSSELL,W  
BREATH TEST OPERATOR AGENCY: FHP Troop E  
ARRESTING OFFICER: CRUZ,RUSSELL  
ARRESTING OFFICER AGENCY: FHP Troop E  
LAST AGENCY INSPECTION: 05/06/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 00919080A2  
DRY GAS EXPIRY DATE: 03/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 09:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 09:21

SUBJECT'S LAST NAME: WHITTEN

DATE/TIME OF TEST: 05/07/2024 09:48:56

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 09:43 |
| Air Blank          | 0.000 | 09:43 |
| Control Test       | 0.079 | 09:43 |
| Air Blank          | 0.000 | 09:44 |
| Subject Sample # 1 | 0.000 | 09:44 |
| Breath Volume      | 1.527 | 09:44 |
| Air Blank          | 0.000 | 09:45 |
| Air Blank          | 0.000 | 09:47 |
| Subject Sample # 2 | 0.000 | 09:47 |
| Breath Volume      | 1.167 | 09:47 |
| Air Blank          | 0.000 | 09:48 |
| Control Test       | 0.079 | 09:48 |
| Air Blank          | 0.000 | 09:48 |
| Diagnostics Check  | OK    | 09:48 |

---

Data Download Date/Time: 10/16/2024 15:15

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000834

INSTRUMENT REGISTERED TO: GULF COAST STATE COL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FALCAO,THIAGO,  
BREATH TEST OPERATOR AGENCY: Suwannee County SO  
ARRESTING OFFICER: FALCAO,THIAGO  
ARRESTING OFFICER AGENCY: Suwannee County SO  
LAST AGENCY INSPECTION: 05/06/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 00919080A2  
DRY GAS EXPIRY DATE: 03/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 09:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 09:30

SUBJECT'S LAST NAME: WHITTEN

DATE/TIME OF TEST: 05/07/2024 10:27:41

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 10:21 |
| Air Blank          | 0.000 | 10:22 |
| Control Test       | 0.079 | 10:22 |
| Air Blank          | 0.000 | 10:22 |
| Subject Sample # 1 | 0.000 | 10:23 |
| Breath Volume      | 2.574 | 10:23 |
| Air Blank          | 0.000 | 10:23 |
| Air Blank          | 0.000 | 10:25 |
| Subject Sample # 2 | 0.000 | 10:26 |
| Breath Volume      | 3.179 | 10:26 |
| Air Blank          | 0.000 | 10:26 |
| Control Test       | 0.085 | 10:27 |
| Air Blank          | 0.000 | 10:27 |
| Diagnostics Check  | OK    | 10:27 |

---

Data Download Date/Time: 10/16/2024 15:15

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000834

INSTRUMENT REGISTERED TO: GULF COAST STATE COL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: OLDS,LEE,V  
BREATH TEST OPERATOR AGENCY: FHP Troop E  
ARRESTING OFFICER: OLDS,LEE  
ARRESTING OFFICER AGENCY: FHP Troop E  
LAST AGENCY INSPECTION: 05/06/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 00919080A2  
DRY GAS EXPIRY DATE: 03/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 11:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 12:00

SUBJECT'S LAST NAME: HUBBARD

DATE/TIME OF TEST: 05/07/2024 12:44:17

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 12:38 |
| Air Blank          | 0.000 | 12:38 |
| Control Test       | 0.078 | 12:39 |
| Air Blank          | 0.000 | 12:39 |
| Subject Sample # 1 | 0.000 | 12:40 |
| Breath Volume      | 2.476 | 12:40 |
| Air Blank          | 0.000 | 12:40 |
| Air Blank          | 0.000 | 12:42 |
| Subject Sample # 2 | 0.000 | 12:42 |
| Breath Volume      | 2.343 | 12:42 |
| Air Blank          | 0.000 | 12:43 |
| Control Test       | 0.085 | 12:43 |
| Air Blank          | 0.000 | 12:44 |
| Diagnostics Check  | OK    | 12:44 |

---

Data Download Date/Time: 10/16/2024 15:15

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000834

INSTRUMENT REGISTERED TO: GULF COAST STATE COL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SHIPP,TOM,B  
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)  
ARRESTING OFFICER: SHIPP,TOM  
ARRESTING OFFICER AGENCY: FFWCC (West Palm)  
LAST AGENCY INSPECTION: 05/06/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 00919080A2  
DRY GAS EXPIRY DATE: 03/05/2025  
VIOLATION CODE: BUI  
ARREST TIME: 12:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 12:30

SUBJECT'S LAST NAME: HUBBARD

DATE/TIME OF TEST: 05/07/2024 13:08:03

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 13:01 |
| Air Blank          | 0.000 | 13:02 |
| Control Test       | 0.078 | 13:02 |
| Air Blank          | 0.000 | 13:03 |
| Subject Sample # 1 | 0.000 | 13:03 |
| Breath Volume      | 2.136 | 13:03 |
| Air Blank          | 0.000 | 13:04 |
| Air Blank          | 0.000 | 13:06 |
| Subject Sample # 2 | 0.000 | 13:06 |
| Breath Volume      | 2.417 | 13:06 |
| Air Blank          | 0.000 | 13:07 |
| Control Test       | 0.084 | 13:07 |
| Air Blank          | 0.000 | 13:07 |
| Diagnostics Check  | OK    | 13:08 |

---

Data Download Date/Time: 10/16/2024 15:15

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000834

INSTRUMENT REGISTERED TO: GULF COAST STATE COL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BENITEZ,NORLAN,  
BREATH TEST OPERATOR AGENCY: Bay County SO  
ARRESTING OFFICER: BENITEZ,NORLAN  
ARRESTING OFFICER AGENCY: Palatka PD  
LAST AGENCY INSPECTION: 05/06/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 00919080A2  
DRY GAS EXPIRY DATE: 03/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 06:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 07:59

SUBJECT'S LAST NAME: WHITTEN

DATE/TIME OF TEST: 05/08/2024 08:27:05

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 08:21 |
| Air Blank          | 0.000 | 08:21 |
| Control Test       | 0.078 | 08:21 |
| Air Blank          | 0.000 | 08:22 |
| Subject Sample # 1 | 0.000 | 08:22 |
| Breath Volume      | 2.011 | 08:22 |
| Air Blank          | 0.000 | 08:23 |
| Air Blank          | 0.000 | 08:25 |
| Subject Sample # 2 | 0.000 | 08:25 |
| Breath Volume      | 2.515 | 08:25 |
| Air Blank          | 0.000 | 08:26 |
| Control Test       | 0.085 | 08:26 |
| Air Blank          | 0.000 | 08:27 |
| Diagnostics Check  | OK    | 08:27 |

---

Data Download Date/Time: 10/16/2024 15:15

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MOSS,ROBERT,R  
BREATH TEST OPERATOR AGENCY: Panama City Beach PD  
ARRESTING OFFICER: MOSS,ROBERT  
ARRESTING OFFICER AGENCY: Panama City Beach PD  
LAST AGENCY INSPECTION: 04/05/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402806079  
DRY GAS EXPIRY DATE: 07/26/2026  
VIOLATION CODE: DUI  
ARREST TIME: 20:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:28

SUBJECT'S LAST NAME: MCWENDELLSON

DATE/TIME OF TEST: 05/03/2024 20:55:33

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 20:48 |
| Air Blank          | 0.000 | 20:49 |
| Control Test       | 0.081 | 20:49 |
| Air Blank          | 0.000 | 20:49 |
| Subject Sample # 1 | 0.144 | 20:51 |
| Breath Volume      | 1.394 | 20:51 |
| Air Blank          | 0.000 | 20:51 |
| Air Blank          | 0.000 | 20:53 |
| Subject Sample # 2 | 0.143 | 20:54 |
| Breath Volume      | 1.734 | 20:54 |
| Air Blank          | 0.000 | 20:54 |
| Control Test       | 0.080 | 20:54 |
| Air Blank          | 0.000 | 20:55 |
| Diagnostics Check  | OK    | 20:55 |

---

Data Download Date/Time: 05/08/2024 19:33

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HATLEY,KENDALL,L  
BREATH TEST OPERATOR AGENCY: Panama City Beach PD  
ARRESTING OFFICER: MARSHALL,LEWIS  
ARRESTING OFFICER AGENCY: Panama City Beach PD  
LAST AGENCY INSPECTION: 04/05/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402806079  
DRY GAS EXPIRY DATE: 07/26/2026  
VIOLATION CODE: DUI  
ARREST TIME: 03:10  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: MASTIO

DATE/TIME OF TEST: 05/04/2024 03:59:11

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:50 |
| Air Blank          | 0.000 | 03:51 |
| Control Test       | 0.078 | 03:51 |
| Air Blank          | 0.000 | 03:52 |
| Subject Sample # 1 | 0.180 | 03:54 |
| Breath Volume      | 1.753 | 03:54 |
| Air Blank          | 0.000 | 03:54 |
| Air Blank          | 0.000 | 03:56 |
| Subject Sample # 2 | 0.198 | 03:57 |
| Breath Volume      | 2.015 | 03:57 |
| Air Blank          | 0.000 | 03:58 |
| Control Test       | 0.081 | 03:58 |
| Air Blank          | 0.000 | 03:59 |
| Diagnostics Check  | OK    | 03:59 |

---

Data Download Date/Time: 05/08/2024 19:33

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HATLEY,KENDALL,L  
BREATH TEST OPERATOR AGENCY: Panama City Beach PD  
ARRESTING OFFICER: HATLEY,KENDALL  
ARRESTING OFFICER AGENCY: Panama City Beach PD  
LAST AGENCY INSPECTION: 04/05/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402806079  
DRY GAS EXPIRY DATE: 07/26/2026  
VIOLATION CODE: DUI  
ARREST TIME: 18:35  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 19:45

SUBJECT'S LAST NAME: KING

DATE/TIME OF TEST: 05/04/2024 20:10:30

---

|                   |       |       |            |
|-------------------|-------|-------|------------|
| Diagnostics Check | OK    | 20:08 |            |
| Air Blank         | 0.000 | 20:08 |            |
| Control Test      | 0.080 | 20:09 |            |
| Air Blank         | 0.000 | 20:09 |            |
| Subject Sample    | # 1   | 20:09 | RFI Detect |
| Breath Volume     | 0.000 | 20:09 |            |
| Air Blank         |       | 20:10 | RFI Detect |
| Air Blank         | 0.000 | 20:10 |            |

---

Data Download Date/Time: 05/08/2024 19:33

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HATLEY,KENDALL,L  
BREATH TEST OPERATOR AGENCY: Panama City Beach PD  
ARRESTING OFFICER: HATLEY,KENDALL  
ARRESTING OFFICER AGENCY: Panama City Beach PD  
LAST AGENCY INSPECTION: 04/05/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402806079  
DRY GAS EXPIRY DATE: 07/26/2026  
VIOLATION CODE: DUI  
ARREST TIME: 18:35  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 19:45

SUBJECT'S LAST NAME: KING

DATE/TIME OF TEST: 05/04/2024 20:19:39

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 20:12 |
| Air Blank          | 0.000 | 20:13 |
| Control Test       | 0.081 | 20:13 |
| Air Blank          | 0.000 | 20:13 |
| Subject Sample # 1 | 0.081 | 20:14 |
| Breath Volume      | 2.113 | 20:14 |
| Air Blank          | 0.000 | 20:15 |
| Air Blank          | 0.000 | 20:17 |
| Subject Sample # 2 | 0.081 | 20:18 |
| Breath Volume      | 2.773 | 20:18 |
| Air Blank          | 0.000 | 20:18 |
| Control Test       | 0.080 | 20:19 |
| Air Blank          | 0.000 | 20:19 |
| Diagnostics Check  | OK    | 20:19 |

---

Data Download Date/Time: 05/08/2024 19:33

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LERNER,MATTHEW,H  
BREATH TEST OPERATOR AGENCY: Panama City Beach PD  
ARRESTING OFFICER: ANDREWS,SHAAMAR  
ARRESTING OFFICER AGENCY: Panama City Beach PD  
LAST AGENCY INSPECTION: 04/05/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402806079  
DRY GAS EXPIRY DATE: 07/26/2026  
VIOLATION CODE: DUI  
ARREST TIME: 22:15  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:29

SUBJECT'S LAST NAME: YEAGER

DATE/TIME OF TEST: 05/04/2024 23:01:40

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:52 |
| Air Blank          | 0.000 | 22:53 |
| Control Test       | 0.080 | 22:53 |
| Air Blank          | 0.000 | 22:54 |
| Subject Sample # 1 | 0.268 | 22:56 |
| Breath Volume      | 2.457 | 22:56 |
| Air Blank          | 0.000 | 22:56 |
| Air Blank          | 0.000 | 22:58 |
| Subject Sample # 2 | 0.263 | 23:00 |
| Breath Volume      | 2.097 | 23:00 |
| Air Blank          | 0.000 | 23:00 |
| Control Test       | 0.080 | 23:01 |
| Air Blank          | 0.000 | 23:01 |
| Diagnostics Check  | OK    | 23:01 |

---

Data Download Date/Time: 05/08/2024 19:33

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HATLEY,KENDALL,L  
BREATH TEST OPERATOR AGENCY: Panama City Beach PD  
ARRESTING OFFICER: HATLEY,KENDALL  
ARRESTING OFFICER AGENCY: Panama City Beach PD  
LAST AGENCY INSPECTION: 04/05/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402806079  
DRY GAS EXPIRY DATE: 07/26/2026  
VIOLATION CODE: DUI  
ARREST TIME: 18:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:18

SUBJECT'S LAST NAME: CHRISTIE

DATE/TIME OF TEST: 05/05/2024 19:45:26

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 19:38 |
| Air Blank          | 0.000 | 19:39 |
| Control Test       | 0.080 | 19:39 |
| Air Blank          | 0.000 | 19:40 |
| Subject Sample # 1 | 0.041 | 19:40 |
| Breath Volume      | 3.355 | 19:40 |
| Air Blank          | 0.000 | 19:41 |
| Air Blank          | 0.000 | 19:43 |
| Subject Sample # 2 | 0.041 | 19:43 |
| Breath Volume      | 3.675 | 19:43 |
| Air Blank          | 0.000 | 19:44 |
| Control Test       | 0.080 | 19:44 |
| Air Blank          | 0.000 | 19:45 |
| Diagnostics Check  | OK    | 19:45 |

---

Data Download Date/Time: 05/08/2024 19:33

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HATLEY,KENDALL,L  
BREATH TEST OPERATOR AGENCY: Panama City Beach PD  
ARRESTING OFFICER: DAVIS,LEANDRA  
ARRESTING OFFICER AGENCY: Panama City Beach PD  
LAST AGENCY INSPECTION: 04/05/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402806079  
DRY GAS EXPIRY DATE: 07/26/2026  
VIOLATION CODE: DUI  
ARREST TIME: 20:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: RYAN

DATE/TIME OF TEST: 05/05/2024 20:54:11

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 20:50 |                      |
| Air Blank         | 0.000 | 20:51 |                      |
| Control Test      | 0.079 | 20:51 |                      |
| Air Blank         | 0.000 | 20:52 |                      |
| Subject Sample    | # 1   | 20:52 | Subject Test Refused |
| Breath Volume     | 0.000 | 20:52 |                      |
| Air Blank         | 0.000 | 20:53 |                      |
| Control Test      | 0.080 | 20:53 |                      |
| Air Blank         | 0.000 | 20:54 |                      |
| Diagnostics Check | OK    | 20:54 |                      |

---

Data Download Date/Time: 05/08/2024 19:33

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MELCHER,ANDREW,M  
BREATH TEST OPERATOR AGENCY: Panama City Beach PD  
ARRESTING OFFICER: NIELSEN,YELENA  
ARRESTING OFFICER AGENCY: Panama City Beach PD  
LAST AGENCY INSPECTION: 05/06/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402806079  
DRY GAS EXPIRY DATE: 07/26/2026  
VIOLATION CODE: DUI  
ARREST TIME: 13:24  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 13:45

SUBJECT'S LAST NAME: PIERSON

DATE/TIME OF TEST: 05/07/2024 14:12:11

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 14:06 |
| Air Blank          | 0.000 | 14:06 |
| Control Test       | 0.081 | 14:06 |
| Air Blank          | 0.000 | 14:07 |
| Subject Sample # 1 | 0.193 | 14:07 |
| Breath Volume      | 2.640 | 14:07 |
| Air Blank          | 0.000 | 14:08 |
| Air Blank          | 0.000 | 14:10 |
| Subject Sample # 2 | 0.194 | 14:10 |
| Breath Volume      | 2.808 | 14:10 |
| Air Blank          | 0.000 | 14:11 |
| Control Test       | 0.081 | 14:11 |
| Air Blank          | 0.000 | 14:12 |
| Diagnostics Check  | OK    | 14:12 |

---

Data Download Date/Time: 05/08/2024 19:33

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MOSS,ROBERT,R  
BREATH TEST OPERATOR AGENCY: Panama City Beach PD  
ARRESTING OFFICER: MOSS,ROBERT  
ARRESTING OFFICER AGENCY: Panama City Beach PD  
LAST AGENCY INSPECTION: 05/06/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402806079  
DRY GAS EXPIRY DATE: 07/26/2026  
VIOLATION CODE: DUI  
ARREST TIME: 23:51  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:04

SUBJECT'S LAST NAME: RAMOS JIMENEZ

DATE/TIME OF TEST: 05/09/2024 00:31:29

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:25 |
| Air Blank          | 0.000 | 00:25 |
| Control Test       | 0.081 | 00:25 |
| Air Blank          | 0.000 | 00:26 |
| Subject Sample # 1 | 0.119 | 00:26 |
| Breath Volume      | 1.667 | 00:26 |
| Air Blank          | 0.000 | 00:27 |
| Air Blank          | 0.000 | 00:29 |
| Subject Sample # 2 | 0.123 | 00:30 |
| Breath Volume      | 1.574 | 00:30 |
| Air Blank          | 0.000 | 00:30 |
| Control Test       | 0.081 | 00:30 |
| Air Blank          | 0.000 | 00:31 |
| Diagnostics Check  | OK    | 00:31 |

---

Data Download Date/Time: 06/07/2024 15:02

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MELCHER,ANDREW,M  
BREATH TEST OPERATOR AGENCY: Panama City Beach PD  
ARRESTING OFFICER: MONROE,JOEL  
ARRESTING OFFICER AGENCY: Panama City Beach PD  
LAST AGENCY INSPECTION: 05/06/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402806079  
DRY GAS EXPIRY DATE: 07/26/2026  
VIOLATION CODE: DUI  
ARREST TIME: 17:49  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 18:22

SUBJECT'S LAST NAME: IAMS

DATE/TIME OF TEST: 05/12/2024 18:48:39

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 18:42 |
| Air Blank          | 0.000 | 18:43 |
| Control Test       | 0.081 | 18:43 |
| Air Blank          | 0.000 | 18:44 |
| Subject Sample # 1 | 0.177 | 18:44 |
| Breath Volume      | 2.492 | 18:44 |
| Air Blank          | 0.000 | 18:44 |
| Air Blank          | 0.000 | 18:46 |
| Subject Sample # 2 | 0.157 | 18:47 |
| Breath Volume      | 2.019 | 18:47 |
| Air Blank          | 0.000 | 18:47 |
| Control Test       | 0.081 | 18:48 |
| Air Blank          | 0.000 | 18:48 |
| Diagnostics Check  | OK    | 18:48 |

---

Data Download Date/Time: 06/07/2024 15:02

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LERNER,MATTHEW,H  
BREATH TEST OPERATOR AGENCY: Panama City Beach PD  
ARRESTING OFFICER: ANDREWS,SHAAMAR  
ARRESTING OFFICER AGENCY: Panama City Beach PD  
LAST AGENCY INSPECTION: 05/06/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402806079  
DRY GAS EXPIRY DATE: 07/26/2026  
VIOLATION CODE: DUI  
ARREST TIME: 03:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: BERNARD

DATE/TIME OF TEST: 05/14/2024 04:31:45

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:22 |
| Air Blank          | 0.000 | 04:23 |
| Control Test       | 0.080 | 04:23 |
| Air Blank          | 0.000 | 04:23 |
| Subject Sample # 1 | 0.201 | 04:24 |
| Breath Volume      | 1.921 | 04:24 |
| Air Blank          | 0.000 | 04:25 |
| Air Blank          | 0.000 | 04:26 |
| Subject Sample # 2 | 0.172 | 04:27 |
| Breath Volume      | 2.320 | 04:27 |
| Air Blank          | 0.000 | 04:27 |
| Air Blank          | 0.000 | 04:29 |
| Subject Sample # 3 | 0.198 | 04:30 |
| Breath Volume      | 2.472 | 04:30 |
| Air Blank          | 0.000 | 04:30 |
| Control Test       | 0.081 | 04:31 |
| Air Blank          | 0.000 | 04:31 |
| Diagnostics Check  | OK    | 04:31 |

---

Data Download Date/Time: 06/07/2024 15:02

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MOSS,ROBERT,R  
BREATH TEST OPERATOR AGENCY: Panama City Beach PD  
ARRESTING OFFICER: MOSS,ROBERT  
ARRESTING OFFICER AGENCY: Panama City Beach PD  
LAST AGENCY INSPECTION: 05/06/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402806079  
DRY GAS EXPIRY DATE: 07/26/2026  
VIOLATION CODE: DUI  
ARREST TIME: 21:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: TEEGARDEN

DATE/TIME OF TEST: 05/17/2024 21:46:48

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 21:40 |
| Air Blank          | 0.000 | 21:41 |
| Control Test       | 0.080 | 21:41 |
| Air Blank          | 0.000 | 21:42 |
| Subject Sample # 1 | 0.000 | 21:42 |
| Breath Volume      | 2.335 | 21:42 |
| Air Blank          | 0.000 | 21:43 |
| Air Blank          | 0.000 | 21:45 |
| Subject Sample # 2 | 0.000 | 21:45 |
| Breath Volume      | 1.210 | 21:45 |
| Air Blank          | 0.000 | 21:45 |
| Control Test       | 0.081 | 21:46 |
| Air Blank          | 0.000 | 21:46 |
| Diagnostics Check  | OK    | 21:46 |

---

Data Download Date/Time: 06/07/2024 15:02

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LERNER,MATTHEW,H  
BREATH TEST OPERATOR AGENCY: Panama City Beach PD  
ARRESTING OFFICER: ANDREWS,SHAAMAR  
ARRESTING OFFICER AGENCY: Panama City Beach PD  
LAST AGENCY INSPECTION: 05/06/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402806079  
DRY GAS EXPIRY DATE: 07/26/2026  
VIOLATION CODE: DUI  
ARREST TIME: 03:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 05/19/2024 04:05:36

---

|                   |           |       |                |
|-------------------|-----------|-------|----------------|
| Diagnostics Check | OK        | 03:54 |                |
| Air Blank         | 0.000     | 03:54 |                |
| Control Test      | 0.080     | 03:54 |                |
| Air Blank         | 0.000     | 03:55 |                |
| Subject Sample    | # 1 0.152 | 03:58 | Volume Not Met |
| Breath Volume     | 0.082     | 03:58 |                |
| Air Blank         | 0.000     | 03:59 |                |
| Air Blank         | 0.000     | 04:00 |                |
| Subject Sample    | # 2 0.178 | 04:04 | Volume Not Met |
| Breath Volume     | 0.164     | 04:04 |                |
| Air Blank         | 0.000     | 04:04 |                |
| Control Test      | 0.081     | 04:05 |                |
| Air Blank         | 0.000     | 04:05 |                |
| Diagnostics Check | OK        | 04:05 |                |

---

Data Download Date/Time: 06/07/2024 15:02

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KIGHT,BRANDON,T  
BREATH TEST OPERATOR AGENCY: Panama City Beach PD  
ARRESTING OFFICER: COZAD,AUNDREA  
ARRESTING OFFICER AGENCY: Panama City Beach PD  
LAST AGENCY INSPECTION: 05/06/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402806079  
DRY GAS EXPIRY DATE: 07/26/2026  
VIOLATION CODE: DUI  
ARREST TIME: 06:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 07:00

SUBJECT'S LAST NAME: ABREU DE JESUS

DATE/TIME OF TEST: 05/20/2024 07:27:18

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 07:20 |
| Air Blank          | 0.000 | 07:21 |
| Control Test       | 0.081 | 07:21 |
| Air Blank          | 0.000 | 07:22 |
| Subject Sample # 1 | 0.153 | 07:22 |
| Breath Volume      | 1.898 | 07:22 |
| Air Blank          | 0.000 | 07:23 |
| Air Blank          | 0.000 | 07:25 |
| Subject Sample # 2 | 0.159 | 07:25 |
| Breath Volume      | 2.656 | 07:25 |
| Air Blank          | 0.000 | 07:26 |
| Control Test       | 0.081 | 07:26 |
| Air Blank          | 0.000 | 07:27 |
| Diagnostics Check  | OK    | 07:27 |

---

Data Download Date/Time: 06/07/2024 15:02

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HATLEY,KENDALL,L  
BREATH TEST OPERATOR AGENCY: Panama City Beach PD  
ARRESTING OFFICER: HATLEY,KENDALL  
ARRESTING OFFICER AGENCY: Panama City Beach PD  
LAST AGENCY INSPECTION: 05/06/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402806079  
DRY GAS EXPIRY DATE: 07/26/2026  
VIOLATION CODE: DUI  
ARREST TIME: 05:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:13

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 05/28/2024 05:39:57

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 05:33 |
| Air Blank          | 0.000 | 05:34 |
| Control Test       | 0.081 | 05:34 |
| Air Blank          | 0.000 | 05:35 |
| Subject Sample # 1 | 0.000 | 05:35 |
| Breath Volume      | 1.816 | 05:35 |
| Air Blank          | 0.000 | 05:36 |
| Air Blank          | 0.000 | 05:38 |
| Subject Sample # 2 | 0.000 | 05:38 |
| Breath Volume      | 1.839 | 05:38 |
| Air Blank          | 0.000 | 05:39 |
| Control Test       | 0.080 | 05:39 |
| Air Blank          | 0.000 | 05:39 |
| Diagnostics Check  | OK    | 05:39 |

---

Data Download Date/Time: 06/07/2024 15:02

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MCDOWELL,MICHAEL,J  
BREATH TEST OPERATOR AGENCY: Panama City Beach PD  
ARRESTING OFFICER: COZAD,AUNDREA  
ARRESTING OFFICER AGENCY: Panama City Beach PD  
LAST AGENCY INSPECTION: 05/06/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402806079  
DRY GAS EXPIRY DATE: 07/26/2026  
VIOLATION CODE: DUI  
ARREST TIME: 13:38  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 14:14

SUBJECT'S LAST NAME: MERRING

DATE/TIME OF TEST: 05/29/2024 14:41:09

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 14:34 |
| Air Blank          | 0.000 | 14:35 |
| Control Test       | 0.081 | 14:35 |
| Air Blank          | 0.000 | 14:35 |
| Subject Sample # 1 | 0.000 | 14:36 |
| Breath Volume      | 2.460 | 14:36 |
| Air Blank          | 0.000 | 14:37 |
| Air Blank          | 0.000 | 14:39 |
| Subject Sample # 2 | 0.000 | 14:39 |
| Breath Volume      | 2.398 | 14:39 |
| Air Blank          | 0.000 | 14:40 |
| Control Test       | 0.081 | 14:40 |
| Air Blank          | 0.000 | 14:41 |
| Diagnostics Check  | OK    | 14:41 |

---

Data Download Date/Time: 06/07/2024 15:02

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WEBER,JEFFREY,J  
BREATH TEST OPERATOR AGENCY: Panama City Beach PD  
ARRESTING OFFICER: WEBER,JEFFREY  
ARRESTING OFFICER AGENCY: Panama City Beach PD  
LAST AGENCY INSPECTION: 05/06/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402806079  
DRY GAS EXPIRY DATE: 07/26/2026  
VIOLATION CODE: DUI  
ARREST TIME: 02:37  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:46

SUBJECT'S LAST NAME: MCCRAY

DATE/TIME OF TEST: 05/30/2024 03:15:42

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:07 |
| Air Blank          | 0.000 | 03:07 |
| Control Test       | 0.080 | 03:08 |
| Air Blank          | 0.000 | 03:08 |
| Subject Sample # 1 | 0.137 | 03:10 |
| Breath Volume      | 1.558 | 03:10 |
| Air Blank          | 0.000 | 03:11 |
| Air Blank          | 0.000 | 03:13 |
| Subject Sample # 2 | 0.136 | 03:14 |
| Breath Volume      | 1.527 | 03:14 |
| Air Blank          | 0.000 | 03:14 |
| Control Test       | 0.080 | 03:15 |
| Air Blank          | 0.000 | 03:15 |
| Diagnostics Check  | OK    | 03:15 |

---

Data Download Date/Time: 06/07/2024 15:02

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MELCHER,ANDREW,M  
BREATH TEST OPERATOR AGENCY: Panama City Beach PD  
ARRESTING OFFICER: MONROE,JOEL  
ARRESTING OFFICER AGENCY: Panama City Beach PD  
LAST AGENCY INSPECTION: 05/06/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402806079  
DRY GAS EXPIRY DATE: 07/26/2026  
VIOLATION CODE: DUI  
ARREST TIME: 22:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:58

SUBJECT'S LAST NAME: RANGEL

DATE/TIME OF TEST: 05/30/2024 23:30:52

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:24 |
| Air Blank          | 0.000 | 23:25 |
| Control Test       | 0.080 | 23:25 |
| Air Blank          | 0.000 | 23:26 |
| Subject Sample # 1 | 0.142 | 23:26 |
| Breath Volume      | 1.640 | 23:26 |
| Air Blank          | 0.000 | 23:27 |
| Air Blank          | 0.000 | 23:29 |
| Subject Sample # 2 | 0.141 | 23:29 |
| Breath Volume      | 1.417 | 23:29 |
| Air Blank          | 0.000 | 23:29 |
| Control Test       | 0.080 | 23:30 |
| Air Blank          | 0.000 | 23:30 |
| Diagnostics Check  | OK    | 23:30 |

---

Data Download Date/Time: 06/07/2024 15:02

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MELCHER,ANDREW,M  
BREATH TEST OPERATOR AGENCY: Panama City Beach PD  
ARRESTING OFFICER: MELCHER,ANDREW  
ARRESTING OFFICER AGENCY: Panama City Beach PD  
LAST AGENCY INSPECTION: 05/06/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402806079  
DRY GAS EXPIRY DATE: 07/26/2026  
VIOLATION CODE: DUI  
ARREST TIME: 00:46  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: NORGAN

DATE/TIME OF TEST: 05/31/2024 01:37:46

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:30 |
| Air Blank          | 0.000 | 01:31 |
| Control Test       | 0.080 | 01:31 |
| Air Blank          | 0.000 | 01:32 |
| Subject Sample # 1 | 0.178 | 01:33 |
| Breath Volume      | 2.527 | 01:33 |
| Air Blank          | 0.000 | 01:33 |
| Air Blank          | 0.000 | 01:35 |
| Subject Sample # 2 | 0.174 | 01:36 |
| Breath Volume      | 2.652 | 01:36 |
| Air Blank          | 0.000 | 01:36 |
| Control Test       | 0.080 | 01:37 |
| Air Blank          | 0.000 | 01:37 |
| Diagnostics Check  | OK    | 01:37 |

---

Data Download Date/Time: 06/07/2024 15:02

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: PANAMA CITY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONAHAN,TIMOTHY,J  
BREATH TEST OPERATOR AGENCY: Panama City PD  
ARRESTING OFFICER: LEE,NICHOLAS  
ARRESTING OFFICER AGENCY: Panama City PD  
LAST AGENCY INSPECTION: 04/27/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 2402946336  
DRY GAS EXPIRY DATE: 01/17/2027  
VIOLATION CODE: DUI  
ARREST TIME: 23:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:06

SUBJECT'S LAST NAME: MILLIRONS

DATE/TIME OF TEST: 05/02/2024 00:33:20

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:26 |
| Air Blank          | 0.000 | 00:27 |
| Control Test       | 0.081 | 00:27 |
| Air Blank          | 0.000 | 00:27 |
| Subject Sample # 1 | 0.022 | 00:28 |
| Breath Volume      | 1.843 | 00:28 |
| Air Blank          | 0.000 | 00:28 |
| Air Blank          | 0.000 | 00:30 |
| Subject Sample # 2 | 0.021 | 00:31 |
| Breath Volume      | 2.601 | 00:31 |
| Air Blank          | 0.000 | 00:32 |
| Control Test       | 0.079 | 00:32 |
| Air Blank          | 0.000 | 00:33 |
| Diagnostics Check  | OK    | 00:33 |

---

Data Download Date/Time: 05/26/2024 23:23

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: PANAMA CITY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GRISOFF,BRENDEN,L  
BREATH TEST OPERATOR AGENCY: Panama City PD  
ARRESTING OFFICER: CADY,ELISE  
ARRESTING OFFICER AGENCY: Panama City PD  
LAST AGENCY INSPECTION: 04/27/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 2402946336  
DRY GAS EXPIRY DATE: 01/17/2027  
VIOLATION CODE: DUI  
ARREST TIME: 19:22  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 19:56

SUBJECT'S LAST NAME: DUDLEY

DATE/TIME OF TEST: 05/08/2024 20:22:18

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 20:16 |
| Air Blank          | 0.000 | 20:16 |
| Control Test       | 0.080 | 20:17 |
| Air Blank          | 0.000 | 20:17 |
| Subject Sample # 1 | 0.000 | 20:18 |
| Breath Volume      | 3.570 | 20:18 |
| Air Blank          | 0.000 | 20:18 |
| Air Blank          | 0.000 | 20:20 |
| Subject Sample # 2 | 0.000 | 20:20 |
| Breath Volume      | 3.500 | 20:20 |
| Air Blank          | 0.000 | 20:21 |
| Control Test       | 0.079 | 20:21 |
| Air Blank          | 0.000 | 20:22 |
| Diagnostics Check  | OK    | 20:22 |

---

Data Download Date/Time: 05/26/2024 23:23

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: PANAMA CITY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JONES,DANIEL,L  
BREATH TEST OPERATOR AGENCY: Panama City PD  
ARRESTING OFFICER: JONES,DANIEL  
ARRESTING OFFICER AGENCY: Panama City PD  
LAST AGENCY INSPECTION: 04/27/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 2402946336  
DRY GAS EXPIRY DATE: 01/17/2027  
VIOLATION CODE: DUI  
ARREST TIME: 01:36  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 05/10/2024 02:11:33

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:05 |
| Air Blank          | 0.000 | 02:05 |
| Control Test       | 0.080 | 02:06 |
| Air Blank          | 0.000 | 02:06 |
| Subject Sample # 1 | 0.196 | 02:07 |
| Breath Volume      | 2.414 | 02:07 |
| Air Blank          | 0.000 | 02:07 |
| Air Blank          | 0.000 | 02:09 |
| Subject Sample # 2 | 0.186 | 02:10 |
| Breath Volume      | 1.566 | 02:10 |
| Air Blank          | 0.000 | 02:10 |
| Control Test       | 0.080 | 02:10 |
| Air Blank          | 0.000 | 02:11 |
| Diagnostics Check  | OK    | 02:11 |

---

Data Download Date/Time: 05/26/2024 23:23

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: PANAMA CITY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JONES,DANIEL,L  
BREATH TEST OPERATOR AGENCY: Panama City PD  
ARRESTING OFFICER: SCOTT,DYLAN  
ARRESTING OFFICER AGENCY: Panama City PD  
LAST AGENCY INSPECTION: 04/27/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 2402946336  
DRY GAS EXPIRY DATE: 01/17/2027  
VIOLATION CODE: DUI  
ARREST TIME: 02:44  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:57

SUBJECT'S LAST NAME: ALVARADO

DATE/TIME OF TEST: 05/11/2024 03:24:29

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:17 |
| Air Blank          | 0.000 | 03:18 |
| Control Test       | 0.081 | 03:18 |
| Air Blank          | 0.000 | 03:19 |
| Subject Sample # 1 | 0.149 | 03:20 |
| Breath Volume      | 1.683 | 03:20 |
| Air Blank          | 0.000 | 03:20 |
| Air Blank          | 0.000 | 03:22 |
| Subject Sample # 2 | 0.146 | 03:23 |
| Breath Volume      | 1.898 | 03:23 |
| Air Blank          | 0.000 | 03:23 |
| Control Test       | 0.081 | 03:23 |
| Air Blank          | 0.000 | 03:24 |
| Diagnostics Check  | OK    | 03:24 |

---

Data Download Date/Time: 05/26/2024 23:23

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: PANAMA CITY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ALLYN, WILLIAM, P  
BREATH TEST OPERATOR AGENCY: Panama City PD  
ARRESTING OFFICER: GAY, JON  
ARRESTING OFFICER AGENCY: Panama City PD  
LAST AGENCY INSPECTION: 04/27/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 2402946336  
DRY GAS EXPIRY DATE: 01/17/2027  
VIOLATION CODE: DUI  
ARREST TIME: 08:35  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 08:54

SUBJECT'S LAST NAME: TOWNE

DATE/TIME OF TEST: 05/14/2024 09:24:38

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 09:18 |
| Air Blank          | 0.000 | 09:18 |
| Control Test       | 0.080 | 09:18 |
| Air Blank          | 0.000 | 09:19 |
| Subject Sample # 1 | 0.000 | 09:20 |
| Breath Volume      | 1.105 | 09:20 |
| Air Blank          | 0.000 | 09:20 |
| Air Blank          | 0.000 | 09:22 |
| Subject Sample # 2 | 0.000 | 09:23 |
| Breath Volume      | 1.484 | 09:23 |
| Air Blank          | 0.000 | 09:23 |
| Control Test       | 0.080 | 09:24 |
| Air Blank          | 0.000 | 09:24 |
| Diagnostics Check  | OK    | 09:24 |

---

Data Download Date/Time: 05/26/2024 23:23

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: PANAMA CITY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GRISOFF,BRENDEN,L  
BREATH TEST OPERATOR AGENCY: Panama City PD  
ARRESTING OFFICER: ZIPF,RICHARD  
ARRESTING OFFICER AGENCY: Panama City PD  
LAST AGENCY INSPECTION: 04/27/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 2402946336  
DRY GAS EXPIRY DATE: 01/17/2027  
VIOLATION CODE: DUI  
ARREST TIME: 00:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: MCGHEE

DATE/TIME OF TEST: 05/18/2024 01:33:05

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:26 |
| Air Blank          | 0.000 | 01:27 |
| Control Test       | 0.080 | 01:27 |
| Air Blank          | 0.000 | 01:27 |
| Subject Sample # 1 | 0.085 | 01:28 |
| Breath Volume      | 2.378 | 01:28 |
| Air Blank          | 0.000 | 01:29 |
| Air Blank          | 0.000 | 01:31 |
| Subject Sample # 2 | 0.081 | 01:31 |
| Breath Volume      | 2.035 | 01:31 |
| Air Blank          | 0.000 | 01:32 |
| Control Test       | 0.080 | 01:32 |
| Air Blank          | 0.000 | 01:32 |
| Diagnostics Check  | OK    | 01:33 |

---

Data Download Date/Time: 05/26/2024 23:23

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: PANAMA CITY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ALLYN,WILLIAM,P  
BREATH TEST OPERATOR AGENCY: Panama City PD  
ARRESTING OFFICER: ALLYN,WILLIAM  
ARRESTING OFFICER AGENCY: Panama City PD  
LAST AGENCY INSPECTION: 04/27/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 2402946336  
DRY GAS EXPIRY DATE: 01/17/2027  
VIOLATION CODE: DUI  
ARREST TIME: 05:42  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:00

SUBJECT'S LAST NAME: BARFIELD

DATE/TIME OF TEST: 05/18/2024 06:23:26

---

|                   |       |       |            |
|-------------------|-------|-------|------------|
| Diagnostics Check | OK    | 06:20 |            |
| Air Blank         | 0.000 | 06:21 |            |
| Control Test      | 0.080 | 06:21 |            |
| Air Blank         | 0.000 | 06:21 |            |
| Subject Sample    | # 1   | 06:22 | RFI Detect |
| Breath Volume     | 0.000 | 06:22 |            |
| Air Blank         | 0.000 | 06:22 |            |
| Control Test      | 0.081 | 06:22 |            |
| Air Blank         | 0.000 | 06:23 |            |
| Diagnostics Check | OK    | 06:23 |            |

---

Data Download Date/Time: 05/26/2024 23:23

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: PANAMA CITY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ALLYN,WILLIAM,P  
BREATH TEST OPERATOR AGENCY: Panama City PD  
ARRESTING OFFICER: ALLYN,WILLIAM  
ARRESTING OFFICER AGENCY: Panama City PD  
LAST AGENCY INSPECTION: 04/27/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 2402946336  
DRY GAS EXPIRY DATE: 01/17/2027  
VIOLATION CODE: DUI  
ARREST TIME: 05:42  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:00

SUBJECT'S LAST NAME: BARFIELD

DATE/TIME OF TEST: 05/18/2024 06:31:34

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 06:25 |
| Air Blank          | 0.000 | 06:25 |
| Control Test       | 0.081 | 06:25 |
| Air Blank          | 0.000 | 06:26 |
| Subject Sample # 1 | 0.057 | 06:26 |
| Breath Volume      | 1.445 | 06:26 |
| Air Blank          | 0.000 | 06:27 |
| Air Blank          | 0.000 | 06:29 |
| Subject Sample # 2 | 0.055 | 06:30 |
| Breath Volume      | 1.937 | 06:30 |
| Air Blank          | 0.000 | 06:30 |
| Control Test       | 0.080 | 06:31 |
| Air Blank          | 0.000 | 06:31 |
| Diagnostics Check  | OK    | 06:31 |

---

Data Download Date/Time: 05/26/2024 23:23

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: PANAMA CITY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JONES,DANIEL,L  
BREATH TEST OPERATOR AGENCY: Panama City PD  
ARRESTING OFFICER: JONES,DANIEL  
ARRESTING OFFICER AGENCY: Panama City PD  
LAST AGENCY INSPECTION: 04/27/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 2402946336  
DRY GAS EXPIRY DATE: 01/17/2027  
VIOLATION CODE: DUI  
ARREST TIME: 21:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:13

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 05/19/2024 21:40:14

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 21:33 |
| Air Blank          | 0.000 | 21:34 |
| Control Test       | 0.080 | 21:34 |
| Air Blank          | 0.000 | 21:35 |
| Subject Sample # 1 | 0.000 | 21:35 |
| Breath Volume      | 3.726 | 21:35 |
| Air Blank          | 0.000 | 21:36 |
| Air Blank          | 0.000 | 21:38 |
| Subject Sample # 2 | 0.000 | 21:38 |
| Breath Volume      | 3.238 | 21:38 |
| Air Blank          | 0.000 | 21:39 |
| Control Test       | 0.082 | 21:39 |
| Air Blank          | 0.000 | 21:40 |
| Diagnostics Check  | OK    | 21:40 |

---

Data Download Date/Time: 05/26/2024 23:23

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: PANAMA CITY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JONES,DANIEL,L  
BREATH TEST OPERATOR AGENCY: Panama City PD  
ARRESTING OFFICER: JONES,DANIEL  
ARRESTING OFFICER AGENCY: Panama City PD  
LAST AGENCY INSPECTION: 04/27/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 2402946336  
DRY GAS EXPIRY DATE: 01/17/2027  
VIOLATION CODE: DUI  
ARREST TIME: 23:27  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: CITALAN

DATE/TIME OF TEST: 05/20/2024 00:01:00

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:50 |
| Air Blank          | 0.000 | 23:51 |
| Control Test       | 0.080 | 23:51 |
| Air Blank          | 0.000 | 23:52 |
| Subject Sample # 1 | 0.240 | 23:52 |
| Breath Volume      | 1.343 | 23:52 |
| Air Blank          | 0.000 | 23:53 |
| Air Blank          | 0.000 | 23:55 |
| Subject Sample # 2 | 0.262 | 23:56 |
| Breath Volume      | 1.511 | 23:56 |
| Air Blank          | 0.000 | 23:56 |
| Air Blank          | 0.000 | 23:58 |
| Subject Sample # 3 | 0.250 | 23:59 |
| Breath Volume      | 1.550 | 23:59 |
| Air Blank          | 0.000 | 00:00 |
| Control Test       | 0.080 | 00:00 |
| Air Blank          | 0.000 | 00:00 |
| Diagnostics Check  | OK    | 00:01 |

---

Data Download Date/Time: 05/26/2024 23:23

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: PANAMA CITY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JONES,DANIEL,L  
BREATH TEST OPERATOR AGENCY: Panama City PD  
ARRESTING OFFICER: JONES,DANIEL  
ARRESTING OFFICER AGENCY: Panama City PD  
LAST AGENCY INSPECTION: 04/27/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 2402946336  
DRY GAS EXPIRY DATE: 01/17/2027  
VIOLATION CODE: DUI  
ARREST TIME: 05:31  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:32

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 05/21/2024 05:59:25

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 05:53 |
| Air Blank          | 0.000 | 05:53 |
| Control Test       | 0.081 | 05:53 |
| Air Blank          | 0.000 | 05:54 |
| Subject Sample # 1 | 0.237 | 05:54 |
| Breath Volume      | 1.558 | 05:54 |
| Air Blank          | 0.000 | 05:55 |
| Air Blank          | 0.000 | 05:57 |
| Subject Sample # 2 | 0.252 | 05:57 |
| Breath Volume      | 2.234 | 05:57 |
| Air Blank          | 0.000 | 05:58 |
| Control Test       | 0.079 | 05:58 |
| Air Blank          | 0.000 | 05:59 |
| Diagnostics Check  | OK    | 05:59 |

---

Data Download Date/Time: 05/26/2024 23:23

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: PANAMA CITY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NELSON,GARRETT,R  
BREATH TEST OPERATOR AGENCY: Panama City PD  
ARRESTING OFFICER: SOUTH,DAMIAN  
ARRESTING OFFICER AGENCY: Panama City PD  
LAST AGENCY INSPECTION: 04/27/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 2402946336  
DRY GAS EXPIRY DATE: 01/17/2027  
VIOLATION CODE: DUI  
ARREST TIME: 22:03  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: FAOUTAS

DATE/TIME OF TEST: 05/22/2024 22:47:37

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:41 |
| Air Blank          | 0.000 | 22:41 |
| Control Test       | 0.081 | 22:41 |
| Air Blank          | 0.000 | 22:42 |
| Subject Sample # 1 | 0.257 | 22:43 |
| Breath Volume      | 2.648 | 22:43 |
| Air Blank          | 0.000 | 22:43 |
| Air Blank          | 0.000 | 22:45 |
| Subject Sample # 2 | 0.256 | 22:46 |
| Breath Volume      | 3.000 | 22:46 |
| Air Blank          | 0.000 | 22:46 |
| Control Test       | 0.080 | 22:47 |
| Air Blank          | 0.000 | 22:47 |
| Diagnostics Check  | OK    | 22:47 |

---

Data Download Date/Time: 05/26/2024 23:23

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: PANAMA CITY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NELSON,GARRETT,R  
BREATH TEST OPERATOR AGENCY: Panama City PD  
ARRESTING OFFICER: OQUENDO,JOEL  
ARRESTING OFFICER AGENCY: Springfield PD  
LAST AGENCY INSPECTION: 04/27/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 2402946336  
DRY GAS EXPIRY DATE: 01/17/2027  
VIOLATION CODE: DUI  
ARREST TIME: 01:08  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:36

SUBJECT'S LAST NAME: BIAS

DATE/TIME OF TEST: 05/25/2024 02:02:28

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:56 |
| Air Blank          | 0.000 | 01:56 |
| Control Test       | 0.080 | 01:56 |
| Air Blank          | 0.000 | 01:57 |
| Subject Sample # 1 | 0.149 | 01:58 |
| Breath Volume      | 1.347 | 01:58 |
| Air Blank          | 0.000 | 01:58 |
| Air Blank          | 0.000 | 02:00 |
| Subject Sample # 2 | 0.154 | 02:01 |
| Breath Volume      | 1.292 | 02:01 |
| Air Blank          | 0.000 | 02:01 |
| Control Test       | 0.080 | 02:01 |
| Air Blank          | 0.000 | 02:02 |
| Diagnostics Check  | OK    | 02:02 |

---

Data Download Date/Time: 05/26/2024 23:23

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: PANAMA CITY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GRISOFF,BRENDEN,L  
BREATH TEST OPERATOR AGENCY: Panama City PD  
ARRESTING OFFICER: ZIPF,RICHARD  
ARRESTING OFFICER AGENCY: Panama City PD  
LAST AGENCY INSPECTION: 05/26/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 2402946336  
DRY GAS EXPIRY DATE: 01/17/2027  
VIOLATION CODE: DUI  
ARREST TIME: 23:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: ORELLANA

DATE/TIME OF TEST: 05/27/2024 00:22:19

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:15 |
| Air Blank          | 0.000 | 00:15 |
| Control Test       | 0.080 | 00:16 |
| Air Blank          | 0.000 | 00:16 |
| Subject Sample # 1 | 0.197 | 00:17 |
| Breath Volume      | 1.449 | 00:17 |
| Air Blank          | 0.000 | 00:18 |
| Air Blank          | 0.000 | 00:20 |
| Subject Sample # 2 | 0.203 | 00:20 |
| Breath Volume      | 1.804 | 00:20 |
| Air Blank          | 0.000 | 00:21 |
| Control Test       | 0.079 | 00:21 |
| Air Blank          | 0.000 | 00:22 |
| Diagnostics Check  | OK    | 00:22 |

---

Data Download Date/Time: 06/30/2024 23:12

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: PANAMA CITY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JONES,DANIEL,L  
BREATH TEST OPERATOR AGENCY: Panama City PD  
ARRESTING OFFICER: JONES,DANIEL  
ARRESTING OFFICER AGENCY: Panama City PD  
LAST AGENCY INSPECTION: 05/26/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 2402946336  
DRY GAS EXPIRY DATE: 01/17/2027  
VIOLATION CODE: DUI  
ARREST TIME: 22:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:17

SUBJECT'S LAST NAME: HEADLEY

DATE/TIME OF TEST: 05/27/2024 22:47:51

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:37 |
| Air Blank          | 0.000 | 22:38 |
| Control Test       | 0.079 | 22:38 |
| Air Blank          | 0.000 | 22:39 |
| Subject Sample # 1 | 0.204 | 22:40 |
| Breath Volume      | 1.578 | 22:40 |
| Air Blank          | 0.000 | 22:40 |
| Air Blank          | 0.000 | 22:42 |
| Subject Sample # 2 | 0.226 | 22:43 |
| Breath Volume      | 1.832 | 22:43 |
| Air Blank          | 0.000 | 22:43 |
| Air Blank          | 0.000 | 22:45 |
| Subject Sample # 3 | 0.232 | 22:46 |
| Breath Volume      | 2.742 | 22:46 |
| Air Blank          | 0.000 | 22:46 |
| Control Test       | 0.079 | 22:47 |
| Air Blank          | 0.000 | 22:47 |
| Diagnostics Check  | OK    | 22:47 |

---

Data Download Date/Time: 06/30/2024 23:12

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: PANAMA CITY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JONES,DANIEL,L  
BREATH TEST OPERATOR AGENCY: Panama City PD  
ARRESTING OFFICER: JONES,DANIEL  
ARRESTING OFFICER AGENCY: Panama City PD  
LAST AGENCY INSPECTION: 05/26/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 2402946336  
DRY GAS EXPIRY DATE: 01/17/2027  
VIOLATION CODE: DUI  
ARREST TIME: 01:43  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:52

SUBJECT'S LAST NAME: HUSBANDS

DATE/TIME OF TEST: 05/31/2024 02:21:15

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 02:12 |                      |
| Air Blank         | 0.000 | 02:13 |                      |
| Control Test      | 0.080 | 02:13 |                      |
| Air Blank         | 0.000 | 02:14 |                      |
| Subject Sample    | # 1   | 02:17 | No Sample Provided   |
| Breath Volume     | 0.000 | 02:17 |                      |
| Air Blank         | 0.000 | 02:17 |                      |
| Air Blank         | 0.000 | 02:19 |                      |
| Subject Sample    | # 2   | 02:19 | Subject Test Refused |
| Breath Volume     | 0.000 | 02:19 |                      |
| Air Blank         | 0.000 | 02:20 |                      |
| Control Test      | 0.080 | 02:20 |                      |
| Air Blank         | 0.000 | 02:21 |                      |
| Diagnostics Check | OK    | 02:21 |                      |

Data Download Date/Time: 06/30/2024 23:12

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000860

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ALCALDE,ALEXANDER,A  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: JANIAK,MALIKHI  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 04/16/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33522080A5  
DRY GAS EXPIRY DATE: 01/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 22:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:32

SUBJECT'S LAST NAME: WOODINGTON

DATE/TIME OF TEST: 05/01/2024 23:02:57

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:56 |
| Air Blank          | 0.000 | 22:57 |
| Control Test       | 0.080 | 22:57 |
| Air Blank          | 0.000 | 22:58 |
| Subject Sample # 1 | 0.000 | 22:58 |
| Breath Volume      | 2.890 | 22:58 |
| Air Blank          | 0.000 | 22:59 |
| Air Blank          | 0.000 | 23:01 |
| Subject Sample # 2 | 0.000 | 23:01 |
| Breath Volume      | 2.406 | 23:01 |
| Air Blank          | 0.000 | 23:02 |
| Control Test       | 0.082 | 23:02 |
| Air Blank          | 0.000 | 23:02 |
| Diagnostics Check  | OK    | 23:02 |

---

Data Download Date/Time: 05/09/2024 10:47

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000860

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ALCALDE,ALEXANDER,A  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: ALCALDE,ALEXANDER  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 04/16/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33522080A5  
DRY GAS EXPIRY DATE: 01/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 21:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:29

SUBJECT'S LAST NAME: HARRIS JR

DATE/TIME OF TEST: 05/02/2024 21:57:35

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 21:51 |
| Air Blank          | 0.000 | 21:51 |
| Control Test       | 0.080 | 21:52 |
| Air Blank          | 0.000 | 21:52 |
| Subject Sample # 1 | 0.007 | 21:53 |
| Breath Volume      | 1.453 | 21:53 |
| Air Blank          | 0.000 | 21:53 |
| Air Blank          | 0.000 | 21:55 |
| Subject Sample # 2 | 0.009 | 21:56 |
| Breath Volume      | 2.785 | 21:56 |
| Air Blank          | 0.000 | 21:56 |
| Control Test       | 0.081 | 21:56 |
| Air Blank          | 0.000 | 21:57 |
| Diagnostics Check  | OK    | 21:57 |

---

Data Download Date/Time: 05/09/2024 10:47

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000860

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ALCALDE,ALEXANDER,A  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: JANIAK,MALIKHI  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 04/16/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33522080A5  
DRY GAS EXPIRY DATE: 01/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 17:41  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:42

SUBJECT'S LAST NAME: ORTIZ

DATE/TIME OF TEST: 05/06/2024 18:12:47

---

|                   |           |       |            |
|-------------------|-----------|-------|------------|
| Diagnostics Check | OK        | 18:09 |            |
| Air Blank         | 0.000     | 18:09 |            |
| Control Test      | 0.081     | 18:10 |            |
| Air Blank         | 0.000     | 18:10 |            |
| Subject Sample    | # 1 0.355 | 18:11 |            |
| Breath Volume     | 1.332     | 18:11 |            |
| Air Blank         |           | 18:12 | Purge Fail |
| Air Blank         | 0.000     | 18:12 |            |

---

Data Download Date/Time: 05/09/2024 10:47

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000860

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ALCALDE,ALEXANDER,A  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: ALCALDE,ALEXANDER  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 04/16/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33522080A5  
DRY GAS EXPIRY DATE: 01/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 20:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:05

SUBJECT'S LAST NAME: RIAVEZ

DATE/TIME OF TEST: 05/06/2024 20:42:34

---

|                   |           |       |                |
|-------------------|-----------|-------|----------------|
| Diagnostics Check | OK        | 20:28 |                |
| Air Blank         | 0.000     | 20:29 |                |
| Control Test      | 0.080     | 20:29 |                |
| Air Blank         | 0.000     | 20:29 |                |
| Subject Sample    | # 1 0.053 | 20:32 | Volume Not Met |
| Breath Volume     | 0.417     | 20:32 |                |
| Air Blank         | 0.000     | 20:33 |                |
| Air Blank         | 0.000     | 20:35 |                |
| Subject Sample    | # 2 0.060 | 20:37 |                |
| Breath Volume     | 1.582     | 20:37 |                |
| Air Blank         | 0.000     | 20:38 |                |
| Air Blank         | 0.000     | 20:39 |                |
| Subject Sample    | # 3 0.067 | 20:41 |                |
| Breath Volume     | 2.089     | 20:41 |                |
| Air Blank         | 0.000     | 20:41 |                |
| Control Test      | 0.081     | 20:41 |                |
| Air Blank         | 0.000     | 20:42 |                |
| Diagnostics Check | OK        | 20:42 |                |

---

Data Download Date/Time: 05/09/2024 10:47

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000860

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ALCALDE,ALEXANDER,A  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: ALCALDE,ALEXANDER  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 05/07/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33522080A5  
DRY GAS EXPIRY DATE: 01/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 07:02  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 07:02

SUBJECT'S LAST NAME: CEDILLO

DATE/TIME OF TEST: 05/26/2024 07:31:32

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 07:25 |
| Air Blank          | 0.000 | 07:25 |
| Control Test       | 0.079 | 07:25 |
| Air Blank          | 0.000 | 07:26 |
| Subject Sample # 1 | 0.168 | 07:27 |
| Breath Volume      | 1.730 | 07:27 |
| Air Blank          | 0.000 | 07:27 |
| Air Blank          | 0.000 | 07:29 |
| Subject Sample # 2 | 0.186 | 07:29 |
| Breath Volume      | 1.242 | 07:29 |
| Air Blank          | 0.000 | 07:30 |
| Control Test       | 0.079 | 07:30 |
| Air Blank          | 0.000 | 07:31 |
| Diagnostics Check  | OK    | 07:31 |

---

Data Download Date/Time: 06/11/2024 11:27