

Instrument Processing Sheet

Agency: NASSAU COUNTY SO Instrument Serial Number: 80-001281
 Date In: 12/16/2025 DI Completion Date: 12/17/25 Ship P/U H/D CMI EE


Intake By: <u>KTS</u> Date: <u>12/16/25</u> <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input checked="" type="checkbox"/> Return from CMI / EE <input type="checkbox"/> Return unworked <input type="checkbox"/> Training Visual Inspection <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/Accessories <input type="checkbox"/> Power Cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes:	Quality Checks By: <u>KTS</u> Date: <u>12/17/25</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value: <u>158</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column #: <u>ATP103</u> 32 mm <u>0.148</u> (.139-.169) 36 mm <u>0.160</u> (.156-.190) 53 mm <u>0.238</u> (.228-.278) 103 mm <u>0.507</u> (.447-.547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID #: <u>28427</u> Gauge: <u>1017</u> Instrument: <u>1017</u> <input checked="" type="checkbox"/> Stability Checks	Flow Adjustment By: _____ Date: _____ Flow Column #: _____ <input type="checkbox"/> 5L/min – 17mm <input type="checkbox"/> 15L/min – 53mm <input type="checkbox"/> 30L/min – 103mm <input type="checkbox"/> R-Value: _____ <input type="checkbox"/> Post Adjustment Verification (L/S) Flow Column #: _____ 32 mm _____ (.139-.169) 36 mm _____ (.156-.190) 53 mm _____ (.228-.278) 103 mm _____ (.447-.547)
---	--	--

Simulator	Serial #	Lot#/Exp	Maintenance	By:	Date:
0.050	MP5088	202406K 6/19/2026	<input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement and Tank Sensor Tare <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other:		
0.080	MP5089	202406L 6/19/2026			
0.200	MP5090	202406N 6/20/2026			
0.080 DGS	N/A	AG510701			
		4/17/2027			

Optical Bench Adjustment By: _____	Department Inspection By: <u>SP</u>																																								
Barometric Pressure Gauge: _____ ID#: _____	Barometric Pressure ID#: <u>28421</u>																																								
<table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>	Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Gauge: <u>1019</u> Instrument: <u>1017</u> Mouth Alcohol Solution Lot #: <u>2025-D</u> Exp: <u>09/25/27</u> Acetone Stock Solution Lot #: <u>2025-B</u> Exp: <u>09/22/27</u> <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>MP5086</td> </tr> <tr> <td>Interferent</td> <td>MP5087</td> </tr> <tr> <td>0.050</td> <td>MP5088</td> </tr> <tr> <td>0.080</td> <td>MP5089</td> </tr> <tr> <td>0.200</td> <td>MP5090</td> </tr> </tbody> </table>	Simulator	Serial Number	0.000	MP5086	Interferent	MP5087	0.050	MP5088	0.080	MP5089	0.200	MP5090
Simulator	Serial #	Lot #	Expiration																																						
0.000		N/A	N/A																																						
0.040																																									
0.100																																									
0.200																																									
0.300																																									
0.080 DGS	N/A																																								
Simulator	Serial Number																																								
0.000	MP5086																																								
Interferent	MP5087																																								
0.050	MP5088																																								
0.080	MP5089																																								
0.200	MP5090																																								
<input type="checkbox"/> Post Optical Bench Adjustment Stability Checks	Attachments <input checked="" type="checkbox"/> Form 41 <input type="checkbox"/> Post-Stability Checks <input checked="" type="checkbox"/> Stability Checks <input type="checkbox"/> Flow Adjustment <input checked="" type="checkbox"/> Calibration Certificate <input type="checkbox"/> Form 40 <input type="checkbox"/> Optical Bench Adjustment <input type="checkbox"/> Other:																																								
Barometric Pressure Gauge: _____ ID#: _____																																									

Notes/Suggested Service:	<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <hr/> <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <hr/> <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use
	Digitally signed by Taylor Gutschow Date: 2025.12.21 12:57:23 -05'00'
	Digitally signed by Wen-Chi Pierson Date: 2025.12.22 08:22:22 -05'00'
	Taylor Gutschow Wen-Chi Pierson Tech Review Admin Review

Stability Checks

0.050 g/210L 0.047 to 0.053 g/210L	0.080 g/210L 0.077 to 0.083 g/210L	0.200 g/210L 0.194 to 0.206 g/210L	DGS 0.080 g/210L 0.077 to 0.083 g/210L ±0.003 g/210L of Wet																																																																																																																																				
<p>MASSAU COUNTY SO Intoxilyzer - Alcohol Analyzer Model: 8000 SN 80-001281 12/17/2025 Software: 8100.27</p> <p>Test g/210L Time</p> <table border="1"> <tr><td>Air Blank</td><td>1.100</td><td>09:03</td></tr> <tr><td>Control Test</td><td>1.049</td><td>09:04</td></tr> <tr><td>Air Blank</td><td>1.000</td><td>09:04</td></tr> <tr><td>Control Test</td><td>1.048</td><td>09:05</td></tr> <tr><td>Air Blank</td><td>1.000</td><td>09:06</td></tr> <tr><td>Control Test</td><td>1.048</td><td>09:06</td></tr> <tr><td>Air Blank</td><td>1.000</td><td>09:07</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>1.0483</td><td></td></tr> <tr><td>Std Dev</td><td>0.0036</td><td></td></tr> <tr><td>Rel. Std Dev(%)</td><td>1.1945</td><td></td></tr> </table> <p>Operator's Signature: </p>	Air Blank	1.100	09:03	Control Test	1.049	09:04	Air Blank	1.000	09:04	Control Test	1.048	09:05	Air Blank	1.000	09:06	Control Test	1.048	09:06	Air Blank	1.000	09:07	Control Test Stats			Average	1.0483		Std Dev	0.0036		Rel. Std Dev(%)	1.1945		<p>MASSAU COUNTY SO Intoxilyzer - Alcohol Analyzer Model: 8000 SN 80-001281 12/17/2025 Software: 8100.27</p> <p>Test g/210L Time</p> <table border="1"> <tr><td>Air Blank</td><td>1.000</td><td>09:08</td></tr> <tr><td>Control Test</td><td>1.078</td><td>09:09</td></tr> <tr><td>Air Blank</td><td>1.000</td><td>09:09</td></tr> <tr><td>Control Test</td><td>1.078</td><td>09:10</td></tr> <tr><td>Air Blank</td><td>1.000</td><td>09:11</td></tr> <tr><td>Control Test</td><td>1.079</td><td>09:12</td></tr> <tr><td>Air Blank</td><td>1.000</td><td>09:12</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>1.0783</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel. Std Dev(%)</td><td>0.7370</td><td></td></tr> </table> <p>Operator's Signature: </p>	Air Blank	1.000	09:08	Control Test	1.078	09:09	Air Blank	1.000	09:09	Control Test	1.078	09:10	Air Blank	1.000	09:11	Control Test	1.079	09:12	Air Blank	1.000	09:12	Control Test Stats			Average	1.0783		Std Dev	0.0006		Rel. Std Dev(%)	0.7370		<p>MASSAU COUNTY SO Intoxilyzer - Alcohol Analyzer Model: 8000 SN 80-001281 12/17/2025 Software: 8100.27</p> <p>Test g/210L Time</p> <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>09:13</td></tr> <tr><td>Control Test</td><td>0.199</td><td>09:14</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:14</td></tr> <tr><td>Control Test</td><td>1.198</td><td>09:15</td></tr> <tr><td>Air Blank</td><td>1.000</td><td>09:16</td></tr> <tr><td>Control Test</td><td>0.199</td><td>09:16</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:17</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>1.1987</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel. Std Dev(%)</td><td>0.2906</td><td></td></tr> </table> <p>Operator's Signature: </p>	Air Blank	0.000	09:13	Control Test	0.199	09:14	Air Blank	0.000	09:14	Control Test	1.198	09:15	Air Blank	1.000	09:16	Control Test	0.199	09:16	Air Blank	0.000	09:17	Control Test Stats			Average	1.1987		Std Dev	0.0006		Rel. Std Dev(%)	0.2906		<p>MASSAU COUNTY SO Intoxilyzer - Alcohol Analyzer Model: 8000 SN 80-001281 12/17/2025 Software: 8100.27</p> <p>Test g/210L Time</p> <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>08:59</td></tr> <tr><td>Control Test</td><td>0.081</td><td>08:59</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:00</td></tr> <tr><td>Control Test</td><td>0.081</td><td>09:00</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:01</td></tr> <tr><td>Control Test</td><td>0.081</td><td>09:01</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:02</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0810</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel. Std Dev(%)</td><td>0.0000</td><td></td></tr> </table> <p>Operator's Signature: </p>	Air Blank	0.000	08:59	Control Test	0.081	08:59	Air Blank	0.000	09:00	Control Test	0.081	09:00	Air Blank	0.000	09:01	Control Test	0.081	09:01	Air Blank	0.000	09:02	Control Test Stats			Average	0.0810		Std Dev	0.0000		Rel. Std Dev(%)	0.0000	
Air Blank	1.100	09:03																																																																																																																																					
Control Test	1.049	09:04																																																																																																																																					
Air Blank	1.000	09:04																																																																																																																																					
Control Test	1.048	09:05																																																																																																																																					
Air Blank	1.000	09:06																																																																																																																																					
Control Test	1.048	09:06																																																																																																																																					
Air Blank	1.000	09:07																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	1.0483																																																																																																																																						
Std Dev	0.0036																																																																																																																																						
Rel. Std Dev(%)	1.1945																																																																																																																																						
Air Blank	1.000	09:08																																																																																																																																					
Control Test	1.078	09:09																																																																																																																																					
Air Blank	1.000	09:09																																																																																																																																					
Control Test	1.078	09:10																																																																																																																																					
Air Blank	1.000	09:11																																																																																																																																					
Control Test	1.079	09:12																																																																																																																																					
Air Blank	1.000	09:12																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	1.0783																																																																																																																																						
Std Dev	0.0006																																																																																																																																						
Rel. Std Dev(%)	0.7370																																																																																																																																						
Air Blank	0.000	09:13																																																																																																																																					
Control Test	0.199	09:14																																																																																																																																					
Air Blank	0.000	09:14																																																																																																																																					
Control Test	1.198	09:15																																																																																																																																					
Air Blank	1.000	09:16																																																																																																																																					
Control Test	0.199	09:16																																																																																																																																					
Air Blank	0.000	09:17																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	1.1987																																																																																																																																						
Std Dev	0.0006																																																																																																																																						
Rel. Std Dev(%)	0.2906																																																																																																																																						
Air Blank	0.000	08:59																																																																																																																																					
Control Test	0.081	08:59																																																																																																																																					
Air Blank	0.000	09:00																																																																																																																																					
Control Test	0.081	09:00																																																																																																																																					
Air Blank	0.000	09:01																																																																																																																																					
Control Test	0.081	09:01																																																																																																																																					
Air Blank	0.000	09:02																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.0810																																																																																																																																						
Std Dev	0.0000																																																																																																																																						
Rel. Std Dev(%)	0.0000																																																																																																																																						
Performed Root Case Analysis																																																																																																																																							

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: NASSAU COUNTY SO
Time of Inspection: 16:19

Date of Inspection: 12/17/2025

Serial Number: 80-001281
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202406K Exp: 06/19/2026	0.08g/210L Test (g/210L) Lot#:202406L Exp: 06/19/2026	0.20g/210L Test (g/210L) Lot#:202406N Exp: 06/20/2026	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG510701 Exp: 04/17/2027
0.000	0.049	0.078	0.199	0.080
0.000	0.048	0.078	0.199	0.080
0.000	0.048	0.078	0.199	0.079
0.000	0.048	0.078	0.200	0.079
0.000	0.048	0.078	0.199	0.079
0.000	0.048	0.078	0.199	0.078
0.000	0.048	0.078	0.199	0.079
0.000	0.048	0.078	0.199	0.079
0.000	0.049	0.078	0.199	0.078
0.000	0.049	0.078	0.199	0.079

Standard Deviations	0.0004	0.0000	0.0003	0.0006
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

A F / M A: RAN OUT OF BREATH.

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Shayla Platt

SHAYLA D PLATT

Signature and Printed Name

12/17/2025
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001281, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001281</u>	UNCERTAINTY* \pm	
Owning Agency:	<u>NASSAU COUNTY SO</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>12/17/2025</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>16:19</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Shayla Platt
Digitally signed by Shayla Platt
Date: 2025.12.19 15:53:52 -05'00'

12/17/2025

Date

SHAYLA D PLATT,
Department Inspector

FDLE/ATP Form 69 October 2024

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality



INSTRUMENT PROCESSING SHEET

Agency Nassau County Sheriffs Office

S/N 80-001281

Florida Department of Law Enforcement

Date In 05/16/2025 DI Completion Date

Ship P/U H/D CMI EE

Intake By SLH Date 5/19/2025 Quality Checks By SLH Date 5/19/2025 Flow Calibration By Date

Calibration Adjustment By SP Department Inspection By

Table with 4 columns: Simulator, Serial #, Lot #, Expiration. Contains data for Barometric Pressure Gauge 1010 ID #30793.

Table with 4 columns: Simulator, Serial #, Lot #, Expiration. Contains data for Post Calibration Adjustment Stability Checks.

Notes/Suggested Service: Used spare keyboard to operate instrument. SLH 5/19/25. Sending to repair due to outside tolerance values. Keyboard also not operating. SP 9/17/25

Department Inspection section including Barometric Pressure ID#, Gauge, Mouth Alcohol Solution Lot #, Acetone Stock Solution Lot #, and Simulator/Serial Number table.

Attachments section with checkboxes for Form 41, Stability Checks, Calibration Certificate, Calibration Adjustment, Post-Stability Checks, Flow Calibration, Form 40, and Other Form 51.

Compliance checkboxes: Instrument Complies with Chapter 11D-8, FAC; Instrument Does Not Comply with Chapter 11D-8, FAC; Return to/Place into Evidentiary Use; Remain Out of Evidentiary Use; Conduct an Agency Inspection Before Evidentiary Use.

Taylor Gutschow Tech Review / Date and Phil Nicodemo Admin Review / Date with digital signatures.

NASSAU COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001281 ✓
05/19/2025
Software: 8100.27


Test	g/210L	Time
Air Blank	0.000	14:09
Control Test	0.197	14:10
Air Blank	0.000	14:11
Control Test	0.199	14:11
Air Blank	0.000	14:12
Control Test	0.199	14:13
Air Blank	0.000	14:13
Control Test Stats		
Average	0.1983	
Std Dev	0.0012	
Rel Std Dev(%)	0.5822	


Operator's Signature

NASSAU COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001281
05/19/2025
Software: 8100.27

#1

Test	g/210L	Time
Air Blank	0.000	14:17
Control Test	0.069	14:18
Air Blank	0.000	14:19
Control Test	0.070	14:19
Air Blank	0.000	14:20
Control Test	0.071	14:21
Air Blank	0.000	14:21
Control Test Stats		
Average	0.0700	
Std Dev	0.0010	
Rel Std Dev(%)	1.4286	




Operator's Signature

RCA performed - tightened lid
and checked for leaks and re-ran.
SUT

NASSAU COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001281
05/19/2025
Software: 8100.27

#2

Test	g/210L	Time
Air Blank	0.000	14:24
Control Test	0.071	14:25
Air Blank	0.000	14:26
Control Test	0.072	14:26
Air Blank	0.000	14:27
Control Test	0.073	14:27
Air Blank	0.000	14:28
Control Test Stats		
Average	0.0720	
Std Dev	0.0010	
Rel Std Dev(%)	1.3889	


Operator's Signature

no additional RCA needed.
SUT

NASSAU COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001281
05/19/2025
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	14:45
Control Test	0.037	14:46
Air Blank	0.000	14:46
Control Test	0.040	14:47
Air Blank	0.000	14:47
Control Test	0.040	14:48
Air Blank	0.000	14:49
Control Test Stats		
Average	0.0390	
Std Dev	0.0017	
Rel Std Dev(%)	4.4412	


Operator's Signature

RCA - no user or equipment issue,
did not re-run.



NASSAU COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001281
05/19/2025
Software: 8100.27

Dgs

Test	g/210L	Time
-----	-----	-----
Air Blank	0.000	14:54
Control Test	0.077	14:54
Air Blank	0.000	14:55
Control Test	0.078	14:55
Air Blank	0.000	14:56
Control Test	0.077	14:56
Air Blank	0.000	14:57
Control Test Stats		
Average	0.0773	
Std Dev	0.0006	
Rel Std Dev(%)	0.7466	

Mujjal

Operator's Signature

Post Cal Adjust Stability Checks

Met

DGS

NASSAU COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001281
05/21/2025
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:05
Control Test	0.051	11:05
Air Blank	0.000	11:06
Control Test	0.051	11:06
Air Blank	0.000	11:07
Control Test	0.050	11:08
Air Blank	0.000	11:08
Control Test Stats		
Average	0.0507	
Std Dev	0.0006	
Rel Std Dev(%)	1.1395	

SP

Operator's Signature

NASSAU COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001281
05/21/2025
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:30
Control Test	0.084	11:30
Air Blank	0.000	11:31
Control Test	0.084	11:31
Air Blank	0.000	11:32
Control Test	0.086	11:33
Air Blank	0.000	11:33
Control Test Stats		
Average	0.0847	
Std Dev	0.0012	
Rel Std Dev(%)	1.3638	

SP

Operator's Signature

NASSAU COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001281
05/21/2025
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:03
Control Test	0.203	12:04
Air Blank	0.000	12:04
Control Test	0.204	12:05
Air Blank	0.000	12:05
Control Test	0.203	12:06
Air Blank	0.000	12:06
Control Test Stats		
Average	0.2033	
Std Dev	0.0006	
Rel Std Dev(%)	0.2839	

SP

Operator's Signature

NASSAU COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001281
05/21/2025
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:11
Control Test	0.079	11:11
Air Blank	0.000	11:12
Control Test	0.079	11:12
Air Blank	0.000	11:13
Control Test	0.080	11:13
Air Blank	0.000	11:14
Control Test Stats		
Average	0.0793	
Std Dev	0.0006	
Rel Std Dev(%)	0.7277	

SP

Operator's Signature

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: NASSAU COUNTY SO
Time of Inspection: 14:07

Date of Inspection: 05/19/2025

Serial Number: 80-001281 ✓
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08g/210L Test (g/210L) Lot#: _____ Exp: _____	0.20g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: _____ Exp: _____

Number of Simulators Used: _____

Remarks:
BYPASS AI FOR OPERATION. COMPLIANCE UNDETERMINED

N/A
SMB

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Leandra Higginbotham

LEANDRA HIGGINBOTHAM

Signature and Printed Name

05/19/2025
Date

Return Material Authorization

Ship to: CMI, Inc.

Enforcement Electronics

Shipment to repair facility authorized by: B. Higginbotham on 8/18/25

Items Returned: Instrument Supplies Other Describe: _____

Instrument Model: Intoxilyzer 8000 Serial Number: 80-001281

Bill To Address:
Nassau County SO

Ship to Address:
FDLE-TALLAHASSEE

Reason for Return:
Unable to bring values w/in tolerance post
cal adjustment. Keyboard not operational

Please choose one of the following options:

- 1. I _____, authorize all repairs.
- 2. I _____, authorize repairs up to \$ _____.
- 3. I require an estimate **BEFORE** any repairs will be authorized and/ or conducted.

Please contact: Name: _____

Phone #: _____ Email: _____

ATP Contact Name: Shayla Platt ATP Email: shaylaplatt@fdle.state.

fl.us