

Agency: LAKE COUNTY SO Instrument Serial Number: 80-000823
Date In: 7/14/2026 DI Completion Date: 7/22/26 ☐ Ship ☒ P/U ☐ H/D ☐ CMI ☐ EE

*All printed or downloaded copies are uncontrolled
Confirm version prior to use*

Issued By: Forensic Quality Manager
Issue Date: December 22, 2025

062926

111
1 1

DPS

FAIL

SERIAL #
0823

80-000823

7/14/26 ~~RA~~

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKE COUNTY SO

Time of Inspection: 12:06

Date of Inspection: 07/14/2026

Serial Number: 80-000823

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:


Number of Simulators Used: _____

Remarks:

BYPASSED AI TO OPERATE INSTRUMENT COMPLIANCE NOT DETERMINED

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



KATIE T SPEARIN
Signature and Printed Name

07/14/2026
Date

LAKE COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000823
07/14/2026
Software: 8100.27

DIAGNOSTICS

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Performed diagnostic check in attempt to replicate AI's reported issue.
KTS 7/15/26

Stability Checks

0.050 g/210L 0.047 to 0.053 g/210L	0.080 g/210L 0.077 to 0.083 g/210L	0.200 g/210L 0.194 to 0.206 g/210L	DGS 0.080 g/210L 0.077 to 0.083 g/210L ≤0.003 g/210L of Wet																																																																																																																																																
LAKE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000823 07/15/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>12:36</td></tr> <tr><td>Control Test</td><td>0.047</td><td>12:37</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:37</td></tr> <tr><td>Control Test</td><td>0.047</td><td>12:38</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:39</td></tr> <tr><td>Control Test</td><td>0.050</td><td>12:39</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:40</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0480</td><td></td></tr> <tr><td>Std Dev</td><td>0.0017</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>3.6084</td><td></td></tr> </tbody> </table> Operator's Signature	Test	g/210L	Time	Air Blank	0.000	12:36	Control Test	0.047	12:37	Air Blank	0.000	12:37	Control Test	0.047	12:38	Air Blank	0.000	12:39	Control Test	0.050	12:39	Air Blank	0.000	12:40	Control Test Stats			Average	0.0480		Std Dev	0.0017		Rel Std Dev(%)	3.6084		LAKE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000823 07/15/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>12:42</td></tr> <tr><td>Control Test</td><td>0.077</td><td>12:43</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:44</td></tr> <tr><td>Control Test</td><td>0.076</td><td>12:44</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:45</td></tr> <tr><td>Control Test</td><td>0.077</td><td>12:45</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:46</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0767</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7531</td><td></td></tr> </tbody> </table> Operator's Signature	Test	g/210L	Time	Air Blank	0.000	12:42	Control Test	0.077	12:43	Air Blank	0.000	12:44	Control Test	0.076	12:44	Air Blank	0.000	12:45	Control Test	0.077	12:45	Air Blank	0.000	12:46	Control Test Stats			Average	0.0767		Std Dev	0.0006		Rel Std Dev(%)	0.7531		LAKE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000823 07/15/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>12:48</td></tr> <tr><td>Control Test</td><td>0.198</td><td>12:49</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:49</td></tr> <tr><td>Control Test</td><td>0.199</td><td>12:50</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:51</td></tr> <tr><td>Control Test</td><td>0.200</td><td>12:51</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:52</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1990</td><td></td></tr> <tr><td>Std Dev</td><td>0.0010</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.5025</td><td></td></tr> </tbody> </table> Operator's Signature	Test	g/210L	Time	Air Blank	0.000	12:48	Control Test	0.198	12:49	Air Blank	0.000	12:49	Control Test	0.199	12:50	Air Blank	0.000	12:51	Control Test	0.200	12:51	Air Blank	0.000	12:52	Control Test Stats			Average	0.1990		Std Dev	0.0010		Rel Std Dev(%)	0.5025		LAKE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000823 07/15/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>12:33</td></tr> <tr><td>Control Test</td><td>0.079</td><td>12:33</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:33</td></tr> <tr><td>Control Test</td><td>0.078</td><td>12:34</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:34</td></tr> <tr><td>Control Test</td><td>0.077</td><td>12:35</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:35</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0780</td><td></td></tr> <tr><td>Std Dev</td><td>0.0010</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.2821</td><td></td></tr> </tbody> </table> Operator's Signature	Test	g/210L	Time	Air Blank	0.000	12:33	Control Test	0.079	12:33	Air Blank	0.000	12:33	Control Test	0.078	12:34	Air Blank	0.000	12:34	Control Test	0.077	12:35	Air Blank	0.000	12:35	Control Test Stats			Average	0.0780		Std Dev	0.0010		Rel Std Dev(%)	1.2821	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	12:36																																																																																																																																																	
Control Test	0.047	12:37																																																																																																																																																	
Air Blank	0.000	12:37																																																																																																																																																	
Control Test	0.047	12:38																																																																																																																																																	
Air Blank	0.000	12:39																																																																																																																																																	
Control Test	0.050	12:39																																																																																																																																																	
Air Blank	0.000	12:40																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0480																																																																																																																																																		
Std Dev	0.0017																																																																																																																																																		
Rel Std Dev(%)	3.6084																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	12:42																																																																																																																																																	
Control Test	0.077	12:43																																																																																																																																																	
Air Blank	0.000	12:44																																																																																																																																																	
Control Test	0.076	12:44																																																																																																																																																	
Air Blank	0.000	12:45																																																																																																																																																	
Control Test	0.077	12:45																																																																																																																																																	
Air Blank	0.000	12:46																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0767																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7531																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	12:48																																																																																																																																																	
Control Test	0.198	12:49																																																																																																																																																	
Air Blank	0.000	12:49																																																																																																																																																	
Control Test	0.199	12:50																																																																																																																																																	
Air Blank	0.000	12:51																																																																																																																																																	
Control Test	0.200	12:51																																																																																																																																																	
Air Blank	0.000	12:52																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1990																																																																																																																																																		
Std Dev	0.0010																																																																																																																																																		
Rel Std Dev(%)	0.5025																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	12:33																																																																																																																																																	
Control Test	0.079	12:33																																																																																																																																																	
Air Blank	0.000	12:33																																																																																																																																																	
Control Test	0.078	12:34																																																																																																																																																	
Air Blank	0.000	12:34																																																																																																																																																	
Control Test	0.077	12:35																																																																																																																																																	
Air Blank	0.000	12:35																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0780																																																																																																																																																		
Std Dev	0.0010																																																																																																																																																		
Rel Std Dev(%)	1.2821																																																																																																																																																		

80-000823
Optical Bench
Calibration Adjustment
WKP 7/21/2020

Post-Cal Stability Checks

0.05g/210L 0.047 to 0.053	0.08g/210L 0.077 to 0.083	0.20g/210L 0.194 to 0.206	DGS 0.08g/210L 0.077 to 0.083 ≤0.003 of Wet																																																																																																																																																
LAKE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000823 07/21/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>13:39</td></tr> <tr><td>Control Test</td><td>0.050</td><td>13:40</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:40</td></tr> <tr><td>Control Test</td><td>0.049</td><td>13:41</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:42</td></tr> <tr><td>Control Test</td><td>0.049</td><td>13:42</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:43</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0493</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.1703</td><td></td></tr> </tbody> </table> Operator's Signature: <i>menpi</i>	Test	g/210L	Time	Air Blank	0.000	13:39	Control Test	0.050	13:40	Air Blank	0.000	13:40	Control Test	0.049	13:41	Air Blank	0.000	13:42	Control Test	0.049	13:42	Air Blank	0.000	13:43	Control Test Stats			Average	0.0493		Std Dev	0.0006		Rel Std Dev(%)	1.1703		LAKE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000823 07/21/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>13:52</td></tr> <tr><td>Control Test</td><td>0.079</td><td>13:53</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:54</td></tr> <tr><td>Control Test</td><td>0.078</td><td>13:54</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:55</td></tr> <tr><td>Control Test</td><td>0.078</td><td>13:55</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:56</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0783</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7370</td><td></td></tr> </tbody> </table> Operator's Signature: <i>menpi</i>	Test	g/210L	Time	Air Blank	0.000	13:52	Control Test	0.079	13:53	Air Blank	0.000	13:54	Control Test	0.078	13:54	Air Blank	0.000	13:55	Control Test	0.078	13:55	Air Blank	0.000	13:56	Control Test Stats			Average	0.0783		Std Dev	0.0006		Rel Std Dev(%)	0.7370		LAKE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000823 07/21/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>14:06</td></tr> <tr><td>Control Test</td><td>0.197</td><td>14:07</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:08</td></tr> <tr><td>Control Test</td><td>0.196</td><td>14:08</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:09</td></tr> <tr><td>Control Test</td><td>0.196</td><td>14:10</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:10</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1963</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.2941</td><td></td></tr> </tbody> </table> Operator's Signature: <i>menpi</i>	Test	g/210L	Time	Air Blank	0.000	14:06	Control Test	0.197	14:07	Air Blank	0.000	14:08	Control Test	0.196	14:08	Air Blank	0.000	14:09	Control Test	0.196	14:10	Air Blank	0.000	14:10	Control Test Stats			Average	0.1963		Std Dev	0.0006		Rel Std Dev(%)	0.2941		LAKE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000823 07/21/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>13:01</td></tr> <tr><td>Control Test</td><td>0.081</td><td>13:01</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:02</td></tr> <tr><td>Control Test</td><td>0.081</td><td>13:02</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:03</td></tr> <tr><td>Control Test</td><td>0.081</td><td>13:03</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:03</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0810</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> Operator's Signature: <i>menpi</i>	Test	g/210L	Time	Air Blank	0.000	13:01	Control Test	0.081	13:01	Air Blank	0.000	13:02	Control Test	0.081	13:02	Air Blank	0.000	13:03	Control Test	0.081	13:03	Air Blank	0.000	13:03	Control Test Stats			Average	0.0810		Std Dev	0.0000		Rel Std Dev(%)	0.0000	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	13:39																																																																																																																																																	
Control Test	0.050	13:40																																																																																																																																																	
Air Blank	0.000	13:40																																																																																																																																																	
Control Test	0.049	13:41																																																																																																																																																	
Air Blank	0.000	13:42																																																																																																																																																	
Control Test	0.049	13:42																																																																																																																																																	
Air Blank	0.000	13:43																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0493																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	1.1703																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	13:52																																																																																																																																																	
Control Test	0.079	13:53																																																																																																																																																	
Air Blank	0.000	13:54																																																																																																																																																	
Control Test	0.078	13:54																																																																																																																																																	
Air Blank	0.000	13:55																																																																																																																																																	
Control Test	0.078	13:55																																																																																																																																																	
Air Blank	0.000	13:56																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0783																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7370																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	14:06																																																																																																																																																	
Control Test	0.197	14:07																																																																																																																																																	
Air Blank	0.000	14:08																																																																																																																																																	
Control Test	0.196	14:08																																																																																																																																																	
Air Blank	0.000	14:09																																																																																																																																																	
Control Test	0.196	14:10																																																																																																																																																	
Air Blank	0.000	14:10																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1963																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.2941																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	13:01																																																																																																																																																	
Control Test	0.081	13:01																																																																																																																																																	
Air Blank	0.000	13:02																																																																																																																																																	
Control Test	0.081	13:02																																																																																																																																																	
Air Blank	0.000	13:03																																																																																																																																																	
Control Test	0.081	13:03																																																																																																																																																	
Air Blank	0.000	13:03																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0810																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKE COUNTY SO
Time of Inspection: 12:32

Date of Inspection: 07/22/2026

Serial Number: 80-000823
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202603A Exp: 03/03/2028	0.08g/210L Test (g/210L) Lot#:202603B Exp: 03/03/2028	0.20g/210L Test (g/210L) Lot#:202603D Exp: 03/04/2028	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:23825080A1 Exp: 10/05/2027
0.000	0.048	0.078	0.197	0.080
0.000	0.049	0.078	0.197	0.081
0.000	0.049	0.078	0.197	0.080
0.000	0.049	0.078	0.197	0.080
0.000	0.049	0.078	0.197	0.080
0.000	0.049	0.078	0.197	0.080
0.000	0.048	0.078	0.197	0.080
0.000	0.049	0.078	0.197	0.080
0.000	0.048	0.078	0.197	0.080
0.000	0.049	0.077	0.196	0.079


Standard Deviations	0.0004	0.0003	0.0003	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



KATIE T SPEARIN
Signature and Printed Name

07/22/2026
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000823, manufactured by CML, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-000823</u>	UNCERTAINTY * $\pm$
Owning Agency:	<u>LAKE COUNTY SO</u>	0.050 g/210 L 0.004
Calibration Date:	<u>07/22/2026</u>	0.080 g/210 L 0.004
Calibration Time:	<u>12:32</u>	0.200 g/210 L 0.008
		0.080 g/210 L Dry Gas Control 0.005

All results are reported in g/210 L.
Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).
The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/IEC 17025 standards.
This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

FDLE/ATP Form 69 January 2026
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Date 07/22/2026

KATIE T SPEARIN,
Department Inspector