



INSTRUMENT PROCESSING SHEET

Agency LAFAYETTE COUNTY SO

S/N 80-000773

Florida Department of Law Enforcement

Date In 10/6/2025

DI Completion Date 10/28/2025

Ship P/U H/D CMI EE

Intake By WKP Date 10/6/2025 Quality Checks By SLH Date 10/22/2025 Flow Calibration By Date
Annual Registration Return from CMI / EE
Visual Inspection: Case Handle Keyboard Dry Gas Shelf Feet Breath Tube Ports Screws Tight
Other Equipment/ Accessories: Power cord Printer Cable Static Bag 12V DC Cable
Notes: Received with hole in exhaust cover. WKP 10/6/2025 Dropped off by agency. SLH 10/22/2025
Breath Tube Screen Replace External O-Rings Instrument Set Up Verified R-Value 157 Flow Verification (L/s) Flow Column # ATP 102
32 mm 0.148 (.139 - .169) 36 mm 0.160 (.156 - .190) 53 mm 0.238 (.228 - .278) 103 mm 0.511 (.447 - .547)
Barometric Pressure Check Gauge ID # 28421 Stability Checks
Simulator Serial # Lot #/Exp
0.050 MP5088 202406K 06/19/2026
0.080 MP5089 202406L 06/19/2026
0.200 MP5090 202406N 06/20/2026
0.080 DGS N/A AG510701 04/17/2027
Flow Column # 5L/min - 17mm 15L/min - 53mm 30L/min - 103mm
R-Value Post Calibration Verification (L/s)
Flow Column # 32 mm (.139 - .169) 36 mm (.156 - .190) 53 mm (.228 - .278) 103 mm (.447 - .547)
Maintenance By SLH Date 10/27/2025
Battery Replacement Dry Gas Regulator Replacement Breath Tube Replacement Other Replaced exhaust cover.

Calibration Adjustment By Department Inspection By SLH
Barometric Pressure Gauge ID #
Simulator Serial # Lot # Expiration
0.000 N/A N/A
0.040
0.100
0.200
0.300
0.080 DGS N/A
Post Calibration Adjustment Stability Checks
Simulator Serial # Lot # Expiration
0.050
0.080
0.200
0.080 DGS N/A
Barometric Pressure ID# 28421
Gauge 1010 Instrument 1004
Mouth Alcohol Solution Lot # 2025C
Acetone Stock Solution Lot # 2025B
Simulator Serial Number
0.000 MP5086
Interferent MP5087
0.050 MP6291
0.080 MP6292
0.200 MP5090
Attachments
Form 41 Post-Stability Checks
Stability Checks Flow Calibration
Calibration Certificate Form 40
Calibration Adjustment Other

Notes/Suggested Service: Pressure gauge read 1015 and instrument read 1009 - added info for updated IPS. SLH 10/22/2025
Instrument Complies with Chapter 11D-8, FAC
Instrument Does Not Comply with Chapter 11D-8, FAC
Return to/Place into Evidentiary Use
Remain Out of Evidentiary Use
Conduct an Agency Inspection Before Evidentiary Use
Shayla Platt Digitally signed by Shayla Platt Date: 2025.12.04 14:40:19-05'00'
Taylor Gutschow Digitally signed by Taylor Gutschow Date: 2025.12.03 14:50:39 -05'00'
Tech Review / Date Admin Review / Date

# Stability Checks

86-000773  
 10/22/2025 sub  
 pns 5202/ee/01

0.050 g/210L 0.047 to 0.053 g/210L	0.080 g/210L 0.077 to 0.083 g/210L	0.200 g/210L 0.194 to 0.206 g/210L	DGS 0.080 g/210L 0.077 to 0.083 g/210L 50.003 g/210L of Wet																																																																																																																																																
<p>Performed Root Case Analysis</p> <p>LAFAYETTE CO SO            Intoxilyzer - Alcohol Analyzer            Model 8000 SN 80-000773            10/22/2025            Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.003</td><td>15:06</td></tr> <tr><td>Control Test</td><td>0.047</td><td>15:06</td></tr> <tr><td>Air Blank</td><td>0.040</td><td>15:07</td></tr> <tr><td>Control Test</td><td>0.049</td><td>15:08</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:08</td></tr> <tr><td>Control Test</td><td>0.048</td><td>15:09</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:09</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0480</td><td></td></tr> <tr><td>Std Dev</td><td>0.0010</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>2.0833</td><td></td></tr> </tbody> </table> <p>Operator's Signature: <i>M. Sigel</i></p>	Test	g/210L	Time	Air Blank	0.003	15:06	Control Test	0.047	15:06	Air Blank	0.040	15:07	Control Test	0.049	15:08	Air Blank	0.000	15:08	Control Test	0.048	15:09	Air Blank	0.000	15:09	Control Test Stats			Average	0.0480		Std Dev	0.0010		Rel Std Dev(%)	2.0833		<p>Performed Root Case Analysis</p> <p>LAFAYETTE CO SO            Intoxilyzer - Alcohol Analyzer            Model 8000 SN 80-000773            10/22/2025            Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>15:12</td></tr> <tr><td>Control Test</td><td>0.078</td><td>15:13</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:13</td></tr> <tr><td>Control Test</td><td>0.078</td><td>15:14</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:14</td></tr> <tr><td>Control Test</td><td>0.078</td><td>15:15</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:16</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0780</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> <p>Operator's Signature: <i>M. Sigel</i></p>	Test	g/210L	Time	Air Blank	0.000	15:12	Control Test	0.078	15:13	Air Blank	0.000	15:13	Control Test	0.078	15:14	Air Blank	0.000	15:14	Control Test	0.078	15:15	Air Blank	0.000	15:16	Control Test Stats			Average	0.0780		Std Dev	0.0000		Rel Std Dev(%)	0.0000		<p>Performed Root Case Analysis</p> <p>LAFAYETTE CO SO            Intoxilyzer - Alcohol Analyzer            Model 8000 SN 80-000773            10/22/2025            Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>15:24</td></tr> <tr><td>Control Test</td><td>0.198</td><td>15:24</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:25</td></tr> <tr><td>Control Test</td><td>0.199</td><td>15:26</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:26</td></tr> <tr><td>Control Test</td><td>0.198</td><td>15:27</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:27</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1983</td><td></td></tr> <tr><td>Std Dev</td><td>0.0005</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.2911</td><td></td></tr> </tbody> </table> <p>Operator's Signature: <i>M. Sigel</i></p>	Test	g/210L	Time	Air Blank	0.000	15:24	Control Test	0.198	15:24	Air Blank	0.000	15:25	Control Test	0.199	15:26	Air Blank	0.000	15:26	Control Test	0.198	15:27	Air Blank	0.000	15:27	Control Test Stats			Average	0.1983		Std Dev	0.0005		Rel Std Dev(%)	0.2911		<p>Performed Root Case Analysis</p> <p>LAFAYETTE CO SO            Intoxilyzer - Alcohol Analyzer            Model 8000 SN 80-000773            10/22/2025            Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>15:31</td></tr> <tr><td>Control Test</td><td>0.081</td><td>15:31</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:32</td></tr> <tr><td>Control Test</td><td>0.080</td><td>15:32</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:32</td></tr> <tr><td>Control Test</td><td>0.080</td><td>15:33</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:33</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0803</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7187</td><td></td></tr> </tbody> </table> <p>Operator's Signature: <i>M. Sigel</i></p>	Test	g/210L	Time	Air Blank	0.000	15:31	Control Test	0.081	15:31	Air Blank	0.000	15:32	Control Test	0.080	15:32	Air Blank	0.000	15:32	Control Test	0.080	15:33	Air Blank	0.000	15:33	Control Test Stats			Average	0.0803		Std Dev	0.0006		Rel Std Dev(%)	0.7187	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.003	15:06																																																																																																																																																	
Control Test	0.047	15:06																																																																																																																																																	
Air Blank	0.040	15:07																																																																																																																																																	
Control Test	0.049	15:08																																																																																																																																																	
Air Blank	0.000	15:08																																																																																																																																																	
Control Test	0.048	15:09																																																																																																																																																	
Air Blank	0.000	15:09																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0480																																																																																																																																																		
Std Dev	0.0010																																																																																																																																																		
Rel Std Dev(%)	2.0833																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	15:12																																																																																																																																																	
Control Test	0.078	15:13																																																																																																																																																	
Air Blank	0.000	15:13																																																																																																																																																	
Control Test	0.078	15:14																																																																																																																																																	
Air Blank	0.000	15:14																																																																																																																																																	
Control Test	0.078	15:15																																																																																																																																																	
Air Blank	0.000	15:16																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0780																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	15:24																																																																																																																																																	
Control Test	0.198	15:24																																																																																																																																																	
Air Blank	0.000	15:25																																																																																																																																																	
Control Test	0.199	15:26																																																																																																																																																	
Air Blank	0.000	15:26																																																																																																																																																	
Control Test	0.198	15:27																																																																																																																																																	
Air Blank	0.000	15:27																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1983																																																																																																																																																		
Std Dev	0.0005																																																																																																																																																		
Rel Std Dev(%)	0.2911																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	15:31																																																																																																																																																	
Control Test	0.081	15:31																																																																																																																																																	
Air Blank	0.000	15:32																																																																																																																																																	
Control Test	0.080	15:32																																																																																																																																																	
Air Blank	0.000	15:32																																																																																																																																																	
Control Test	0.080	15:33																																																																																																																																																	
Air Blank	0.000	15:33																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0803																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7187																																																																																																																																																		

# Florida Department of Law Enforcement Alcohol Testing Program

## DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: LAFAYETTE CO SO  
Time of Inspection: 13:11

Date of Inspection: 10/28/2025

Serial Number: 80-000773  
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202406K Exp: 06/19/2026	0.08g/210L Test (g/210L) Lot#:202406L Exp: 06/19/2026	0.20g/210L Test (g/210L) Lot#:202406N Exp: 06/20/2026	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG510701 Exp: 04/17/2027
0.000	0.047	0.077	0.194	0.079
0.000	0.048	0.077	0.195	0.080
0.000	0.048	0.078	0.195	0.080
0.000	0.048	0.078	0.195	0.080
0.000	0.048	0.077	0.194	0.080
0.000	0.048	0.077	0.194	0.080
0.000	0.047	0.077	0.194	0.079
0.000	0.048	0.078	0.195	0.079
0.000	0.048	0.077	0.195	0.080
0.000	0.048	0.077	0.194	0.080

Standard Deviations	0.0004	0.0004	0.0005	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.



I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



LEANDRA HIGGINBOTHAM

Signature and Printed Name

10/28/2025  
Date



Florida Department of Law Enforcement  
Alcohol Testing Program  
2331 Phillips Road  
Tallahassee, FL 32308

# Calibration Certificate

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000773, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-000773</u>	UNCERTAINTY* ±	
Owning Agency:	<u>LAFAYETTE CO SO</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>10/28/2025</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>13:11</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.  
Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.  
\*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3)  
The instrument results before and after any adjustment are found in the associated pre and post stability checks

### TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.  
This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

**LEANDRA HIGGINBOTHAM,**  
Department Inspector

10/28/2025  
Date

FDLE/ATP Form 69 October 2024  
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality