



Agency OKALOOSA COUNTY SO

S/N 80-000763

Date In 9/17/2025

DI Completion Date 9/24/2025

☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Calibration Adjustment

By

☐ Post Calibration Adjustment Stability Checks

Notes/Suggested Service:

Department Inspection

By SP

Attachments

- ☒ Instrument Complies with Chapter 11D-8, FAC
- ☐ Instrument Does Not Comply with Chapter 11D-8, FAC
- ☒ Return to/Place into Evidentiary Use
- ☐ Remain Out of Evidentiary Use
- ☒ Conduct an Agency Inspection Before Evidentiary Use

Taylor
Gutschow

Digitally signed by Taylor Gutschon
Date: 2025.09.24 16:08:03 -04'00'

Phil Nicodem

Digitally signed by Phil Nicodemo
Date: 2025.09.26 16:02:28 -04'00'

Tech Review / Date

Admin Review / Date

Stability Checks

[illegible]

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: OKALOOSA COUNTY SO
Time of Inspection: 13:37

Date of Inspection: 09/24/2025

Serial Number: 80-000763
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202406K Exp: 06/19/2026	0.08g/210L Test (g/210L) Lot#:202406L Exp: 06/19/2026	0.20g/210L Test (g/210L) Lot#:202406N Exp: 06/20/2026	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG510701 Exp: 04/17/2027
0.000	0.049	0.078	0.201	0.080
0.000	0.048	0.078	0.201	0.079
0.000	0.048	0.078	0.201	0.079
0.000	0.048	0.079	0.201	0.080
0.000	0.048	0.079	0.201	0.080
0.000	0.048	0.078	0.201	0.079
0.000	0.048	0.078	0.201	0.080
0.000	0.048	0.079	0.201	0.079
0.000	0.049	0.078	0.201	0.079
0.000	0.048	0.079	0.201	0.080

Standard Deviations	0.0004	0.0005	0.0000	0.0005
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Shayla Platt

SHAYLA D PLATT

Signature and Printed Name

09/24/2025
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000763, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-000763</u>	UNCERTAINTY* $\pm$
Owning Agency:	<u>OKALOOSA COUNTY SO</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>09/24/2025</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>13:37</u>	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Shayla Platt
Digitally signed by
Shayla Platt
Date: 2025.09.24
13:52:55 -04'00'

09/24/2025

Date

SHAYLA D PLATT,

FDLE/ATP Form 69 October 2024

Issuing Authority: Alcohol Testing Program

Department Inspector

Service • Integrity • Respect • Quality



INSTRUMENT PROCESSING SHEET

Agency Okaloosa County SOS/N 80-000763Florida Department of
Law EnforcementDate In 5/8/2025DI Completion Date 5/21/25☐ Ship ☐ P/U ☐ H/D ☒ CMI ☐ EE

Intake	By SP	Date <u>5/8/25</u>	Quality Checks	By SLH	Date <u>05/19/2025</u>	Flow Calibration	By	Date																																																																	
☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☐ Power cord ☐ Printer Cable ☐ Static Bag ☐ 12V DC Cable Notes: Instrument arrived in FDLE box. Modem port pushed into instrument. Instrument manually uploaded.			☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>204</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP105</u> 32 mm <u>0.144</u> (.139 - .169) 36 mm <u>0.160</u> (.156 - .190) 53 mm <u>0.234</u> (.228 - .278) 103 mm <u>0.511</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>28427</u> ☒ Stability Checks			Flow Column # _____ ☐ 5L/min – 17mm ☐ 15L/min – 53mm ☐ 30L/min – 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																																																			
			<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP6295</td><td>202406K 06/19/2026</td></tr><tr><td>0.080</td><td>MP6296</td><td>202406L 06/19/2026</td></tr><tr><td>0.200</td><td>MP5090</td><td>202406N 06/20/2026</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG429602 10/22/2026</td></tr></tbody></table>			Simulator	Serial #	Lot #/Exp	0.050	MP6295	202406K 06/19/2026	0.080	MP6296	202406L 06/19/2026	0.200	MP5090	202406N 06/20/2026	0.080 DGS	N/A	AG429602 10/22/2026	<table border="1"><thead><tr><th colspan="2">Maintenance By _____ Date _____</th></tr><tr><td>☐ Battery Replacement</td><td></td></tr><tr><td>☐ Dry Gas Regulator Replacement</td><td></td></tr><tr><td>☐ Breath Tube Replacement</td><td></td></tr><tr><td>☐ Other _____</td><td></td></tr></thead></table>			Maintenance By _____ Date _____		☐ Battery Replacement		☐ Dry Gas Regulator Replacement		☐ Breath Tube Replacement		☐ Other _____																																									
Simulator	Serial #	Lot #/Exp																																																																							
0.050	MP6295	202406K 06/19/2026																																																																							
0.080	MP6296	202406L 06/19/2026																																																																							
0.200	MP5090	202406N 06/20/2026																																																																							
0.080 DGS	N/A	AG429602 10/22/2026																																																																							
Maintenance By _____ Date _____																																																																									
☐ Battery Replacement																																																																									
☐ Dry Gas Regulator Replacement																																																																									
☐ Breath Tube Replacement																																																																									
☐ Other _____																																																																									
Calibration Adjustment By _____			Department Inspection By <u>SP</u>																																																																						
Barometric Pressure Gauge _____ ID # _____ <table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td></td><td></td><td></td></tr><tr><td>0.100</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.300</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table> ☐ Post Calibration Adjustment Stability Checks <table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td></td><td></td><td></td></tr><tr><td>0.080</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>			Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			Barometric Pressure ID# <u>28427</u> Gauge <u>1009</u> Instrument <u>1010</u> Mouth Alcohol Solution Lot # <u>2024-A</u> Acetone Stock Solution Lot # <u>2024-B</u> <table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>MP5086</td></tr><tr><td>Interferent</td><td>MP5087</td></tr><tr><td>0.050</td><td>MP5088</td></tr><tr><td>0.080</td><td>MP5089</td></tr><tr><td>0.200</td><td>MP5090</td></tr></tbody></table> Attachments <table border="1"><tbody><tr><td>☒ Form 41</td><td>☐ Post-Stability Checks</td></tr><tr><td>☒ Stability Checks</td><td>☐ Flow Calibration</td></tr><tr><td>☒ Calibration Certificate</td><td>☒ Form 40</td></tr><tr><td>☐ Calibration Adjustment</td><td>☒ Other <u>Note from agency Form 51 SP 5/27/25</u></td></tr></tbody></table>			Simulator	Serial Number	0.000	MP5086	Interferent	MP5087	0.050	MP5088	0.080	MP5089	0.200	MP5090	☒ Form 41	☐ Post-Stability Checks	☒ Stability Checks	☐ Flow Calibration	☒ Calibration Certificate	☒ Form 40	☐ Calibration Adjustment	☒ Other <u>Note from agency Form 51 SP 5/27/25</u>
Simulator	Serial #	Lot #	Expiration																																																																						
0.000		N/A	N/A																																																																						
0.040																																																																									
0.100																																																																									
0.200																																																																									
0.300																																																																									
0.080 DGS	N/A																																																																								
Simulator	Serial #	Lot #	Expiration																																																																						
0.050																																																																									
0.080																																																																									
0.200																																																																									
0.080 DGS	N/A																																																																								
Simulator	Serial Number																																																																								
0.000	MP5086																																																																								
Interferent	MP5087																																																																								
0.050	MP5088																																																																								
0.080	MP5089																																																																								
0.200	MP5090																																																																								
☒ Form 41	☐ Post-Stability Checks																																																																								
☒ Stability Checks	☐ Flow Calibration																																																																								
☒ Calibration Certificate	☒ Form 40																																																																								
☐ Calibration Adjustment	☒ Other <u>Note from agency Form 51 SP 5/27/25</u>																																																																								
Notes/Suggested Service: _____ 5/21/25 SP 5/28/25 Sending to CMI for modem port repair. SP Tech Review: Removed number from repair form and added Form 51 to attachments. SP 5/27/25 Tech Review: Added date to repair comment. SP 5/28/25			☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☐ Return to/Place into Evidentiary Use ☒ Remain Out of Evidentiary Use ☐ Conduct an Agency Inspection Before Evidentiary Use																																																																						
Taylor Gutschow Digitally signed by Taylor Gutschow Date: 2025.05.28 11:16:42 -0400			Phil Nicodemo Digitally signed by Phil Nicodemo Date: 2025.05.29 09:35:27 -0400																																																																						
Tech Review / Date			Admin Review / Date																																																																						

Stability Checks

5/19/25 Sub
SN: 80-000763

0.050 g/210L 0.047 to 0.053 g/210L	0.080 g/210L 0.077 to 0.083 g/210L	0.200 g/210L 0.194 to 0.206 g/210L	DGS 0.080 g/210L 0.077 to 0.083 g/210L 50.003 g/210L of Wet																																																															
OKALOSA COUNTY SO Intoxilyzer - Alcomol Analyzer Model 8000 05/19/2025 Software: 8100.27 Test 9/210L Time <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>13:13</td></tr> <tr><td>Control Test</td><td>0.049</td><td>13:14</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:14</td></tr> <tr><td>Control Test</td><td>0.049</td><td>13:15</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:15</td></tr> <tr><td>Control Test</td><td>0.049</td><td>13:16</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:17</td></tr> </table> Control Test Stats Average 0.0490 Std Dev 0.0000 Rel Std Dev(%) 0.0000	Air Blank	0.000	13:13	Control Test	0.049	13:14	Air Blank	0.000	13:14	Control Test	0.049	13:15	Air Blank	0.000	13:15	Control Test	0.049	13:16	Air Blank	0.000	13:17	OKALOSA COUNTY SO Intoxilyzer - Alcomol Analyzer Model 8000 05/19/2025 Software: 8100.27 Test 9/210L Time <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>13:19</td></tr> <tr><td>Control Test</td><td>0.200</td><td>13:20</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:20</td></tr> <tr><td>Control Test</td><td>0.199</td><td>13:21</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:22</td></tr> <tr><td>Control Test</td><td>0.198</td><td>13:22</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:23</td></tr> </table> Control Test Stats Average 0.1990 Std Dev 0.0010 Rel Std Dev(%) 0.5025	Air Blank	0.000	13:19	Control Test	0.200	13:20	Air Blank	0.000	13:20	Control Test	0.199	13:21	Air Blank	0.000	13:22	Control Test	0.198	13:22	Air Blank	0.000	13:23	OKALOSA COUNTY SO Intoxilyzer - Alcomol Analyzer Model 8000 05/19/2025 Software: 8100.27 Test 9/210L Time <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>13:19</td></tr> <tr><td>Control Test</td><td>0.200</td><td>13:20</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:20</td></tr> <tr><td>Control Test</td><td>0.199</td><td>13:21</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:22</td></tr> <tr><td>Control Test</td><td>0.198</td><td>13:22</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:23</td></tr> </table> Control Test Stats Average 0.1990 Std Dev 0.0010 Rel Std Dev(%) 0.5025	Air Blank	0.000	13:19	Control Test	0.200	13:20	Air Blank	0.000	13:20	Control Test	0.199	13:21	Air Blank	0.000	13:22	Control Test	0.198	13:22	Air Blank	0.000	13:23	0.077 to 0.083 g/210L 0.077 to 0.083 g/210L 50.003 g/210L of Wet
Air Blank	0.000	13:13																																																																
Control Test	0.049	13:14																																																																
Air Blank	0.000	13:14																																																																
Control Test	0.049	13:15																																																																
Air Blank	0.000	13:15																																																																
Control Test	0.049	13:16																																																																
Air Blank	0.000	13:17																																																																
Air Blank	0.000	13:19																																																																
Control Test	0.200	13:20																																																																
Air Blank	0.000	13:20																																																																
Control Test	0.199	13:21																																																																
Air Blank	0.000	13:22																																																																
Control Test	0.198	13:22																																																																
Air Blank	0.000	13:23																																																																
Air Blank	0.000	13:19																																																																
Control Test	0.200	13:20																																																																
Air Blank	0.000	13:20																																																																
Control Test	0.199	13:21																																																																
Air Blank	0.000	13:22																																																																
Control Test	0.198	13:22																																																																
Air Blank	0.000	13:23																																																																
Performed Root Case Analysis print paper tore - 5/19/25 SN 80-000763 Sub	Performed Root Case Analysis	Performed Root Case Analysis	Performed Root Case Analysis																																																															
Operator's Signature 	Operator's Signature 	Operator's Signature 																																																																

see
additional
page
Sub

OKALOOSA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000763
05/19/2025
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	14:11
Control Test	0.080	14:11
Air Blank	0.000	14:12
Control Test	0.080	14:12
Air Blank	0.000	14:12
Control Test	0.080	14:13
Air Blank	0.000	14:13
Control Test Stats		
Average	0.0800	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	


Operator's Signature

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: OKALOOSA COUNTY SO
Time of Inspection: 10:36

Date of Inspection: 05/21/2025

Serial Number: 80-000763
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202406K Exp: 06/19/2026	0.08g/210L Test (g/210L) Lot#:202406L Exp: 06/19/2026	0.20g/210L Test (g/210L) Lot#:202406N Exp: 06/20/2026	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG429602 Exp: 10/22/2026
0.000	0.044 / 0.047	0.074 / 0.079	0.178 / 0.199	0.079
0.000	0.046 / 0.047	0.074 / 0.079	0.197 / 0.199	0.079
0.000	0.046 / 0.047	0.070 / 0.079	0.197 / 0.199	0.079
0.000	0.046 / 0.047	0.075 / 0.079	0.197 / 0.199	0.079
0.000	0.047 / 0.047	0.074 / 0.079	0.198 / 0.199	0.079
0.000	0.047 / 0.046	0.074 / 0.079	0.198 / 0.199	0.079
0.000	0.047 / 0.047	0.075 / 0.079	0.197 / 0.199	0.079
0.000	0.047 / 0.046	0.074 / 0.079	0.197 / 0.199	0.079
0.000	0.047 / 0.047	0.074 / 0.079	0.198 / 0.200	0.078
0.000	0.047 / 0.046	0.075 / 0.079	0.198 / 0.199	0.078

Standard Deviations	0.0009 / 0.0004	0.0014 / 0.0000	0.0061 / 0.0003	0.0004
---------------------	-----------------	-----------------	-----------------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0002 Number of Simulators Used: 5

Remarks:

05: Control Outside Tolerance- TIGHTEN SIM HEAD. 08: Control Outside Tolerance- CALIBRATION PORT UNS
CREWRD. 20: Control Outside Tolerance. *Checked sim connections and recal. SP*

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Shayla Platt

Signature and Printed Name SHAYLA D PLATT

05/21/2025
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000763, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-000763</u>	UNCERTAINTY* $\pm$
Owning Agency:	<u>OKALOOSA COUNTY SO</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>05/21/2025</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>10:36</u>	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Shayla Platt

Digitally signed by

Shayla Platt

Date: 2025.05.22

13:19:07 -04'00'

05/21/2025

Date

SHAYLA D PLATT,

Department Inspector

FDLE/ATP Form 69 October 2024

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Florida Department of Law Enforcement

Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: OKALOOSA COUNTY SO

Serial Number: 80-000763

Time of Inspection: 12:17

Date of Inspection: 05/19/2025

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot #: Exp:


Number of Simulators Used: _____

Remarks:

BYPASS AI FOR OPERATION, COMPLIANCE UNDETERMINED

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



LEANDRA HIGGINBOTHAM

Signature and Printed Name

05/19/2025

Date

763

Needs an annual inspection.

Please download because we weren't able to get it to connect to download.

Also the phone connection got pushed inside of the instrument and we believe it need to go back to CMI to repair. We currently have one at CMI now for the same exact issue.

Inspector Bill Jerard

Please ship this back to:

Okaloosa County Sheriff's Office

650 Chappie James Street SW

Crestview, Florida 32536

Return Material Authorization

Ship to: ☒ CMI, Inc.

☐ Enforcement Electronics

Shipment to repair facility authorized by: Bill Jerard on 5/8/25

Items Returned: Instrument ☒ Supplies ☐ Other ☐ Describe: 204-SP 5/27/25

Instrument Model: Intoxilyzer 8000 Serial Number: 80-000763

Bill To Address:

Okaloosa County SO

Ship to Address:

FDLE Tallahassee

Reason for Return:

Modem port pushed into instrument

Please choose one of the following options:

- ☐ 1. I _____, authorize all repairs.
- ☐ 2. I _____, authorize repairs up to \$_____.
- ☒ 3. I require an estimate **BEFORE** any repairs will be authorized and/ or conducted.

Please contact: Name: Bill Jerard

Phone #: _____ Email: bjerard@sheriff-okaloosa.org

ATP Contact Name: Shayla Platt ATP Email: _____