



AGENCY INSPECTION AND OTHER ELECTRONIC DATA REVIEW

Agency: FHP Troop E Miami

Instrument Serial Number: 80-003348

AGENCY INSPECTION DISCREPANCY

Agency Inspector: Yosniel Lorenzo

Inspection Date: 07/01/2023

- Agency inspection discrepancy options including test repeats, calendar month, expired solutions, and lot numbers.

REQUESTED CORRECTIVE ACTION

- Requested corrective action options including amendments, written explanations, and removal of instruments.

Taylor Gutschow

Digitally signed by Taylor Gutschow Date: 2023.09.22 18:02:47 -04'00'

09/22/2023

Signature of Alcohol Testing Program Staff Member

Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FHP TROOP E MIAMI
Time of Inspection: 00:10

Date of Inspection: 07/01/2023

Serial Number: 80-003348
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202112A Exp: 12/07/2023	0.08g/210L Test (g/210L) Lot#:202108D Exp: 08/17/2023	0.20g/210L Test (g/210L) Lot#:202106B Exp: 06/22/2023	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:402310313 Exp: 01/15/2025
0.000	0.049	0.078	0.196	0.080
0.000	0.049	0.078	0.197	0.080
0.000	0.049	0.078	0.197	0.079

Number of Simulators Used: 5

Remarks:

A F / M A: RFI Detect RFI Detect. Int Det: RFI Detect RADIO NEAR BY .

EXPIRE ARS USED

Y.L. 10/31/23

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I held a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

YOSNIEL LORENZO

Signature and Printed Name

07/01/2023
Date