

Alcohol Testing Program - Instrument Processing Sheet

Agency: CLAY COUNTY SO

Instrument Serial Number: 80-006237

Date In: 7/20/2026

DI Completion Date: 7/24/2026

☒ Ship

☐ P/U

☐ H/D

☐ CMI

☐ EE

Intake By: <u>WKP</u> Date: <u>7/21/2026</u>		Quality Checks By: <u>SLH</u> Date: <u>7/22/2026</u>		Flow Adjustment By: _____																																									
☒ Annual ☐ Dropped Off ☐ Registration ☒ Return from CMI / EE ☐ Training Instrument Visual Inspection ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/Accessories ☐ Power Cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes:		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value: <u>222</u> ☒ Flow Verification (L/s) Flow Column #: <u>ATP 102</u> 32 mm <u>0.156</u> (.139-.169) 36 mm <u>0.171</u> (.156-.190) 53 mm <u>0.238</u> (.228-.278) 103 mm <u>0.496</u> (.447-.547) ☒ Barometric Pressure Check Gauge ID #: <u>28427</u> Gauge: <u>1010</u> Instrument: <u>1013</u> ☒ Stability Checks		Flow Column #: _____ ☐ 5L/min – 17mm ☐ 15L/min – 53mm ☐ 30L/min – 103mm ☐ R-Value: _____ ☐ Post Adjustment Verification (L/S) Flow Column #: _____ 32 mm _____ (.139-.169) 36 mm _____ (.156-.190) 53 mm _____ (.228-.278) 103 mm _____ (.447-.547)																																									
						<table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th style="width:15%;">Simulator</th> <th style="width:25%;">Serial #</th> <th style="width:20%;">Lot#/Exp</th> <th style="width:40%;">Maintenance By: _____ Date: _____</th> </tr> </thead> <tbody> <tr> <td rowspan="2">0.050</td> <td rowspan="2">MP5088</td> <td>202603A</td> <td>☐ Battery Replacement</td> </tr> <tr> <td>03/03/2028</td> <td>☐ Dry Gas Regulator Replacement</td> </tr> <tr> <td rowspan="2">0.080</td> <td rowspan="2">MP5089</td> <td>202603B</td> <td>☐ Tank Sensor Tare</td> </tr> <tr> <td>03/03/2028</td> <td>☐ Breath Tube Replacement</td> </tr> <tr> <td rowspan="2">0.200</td> <td rowspan="2">MP5090</td> <td>202603D</td> <td>☐ Other:</td> </tr> <tr> <td>03/04/2028</td> <td></td> </tr> <tr> <td rowspan="2">0.080 DGS</td> <td rowspan="2">N/A</td> <td>23825080A1</td> <td></td> </tr> <tr> <td>10/5/2027</td> <td></td> </tr> </tbody> </table>		Simulator	Serial #	Lot#/Exp	Maintenance By: _____ Date: _____	0.050	MP5088	202603A	☐ Battery Replacement	03/03/2028	☐ Dry Gas Regulator Replacement	0.080	MP5089	202603B	☐ Tank Sensor Tare	03/03/2028	☐ Breath Tube Replacement	0.200	MP5090	202603D	☐ Other:	03/04/2028		0.080 DGS	N/A	23825080A1		10/5/2027											
						Simulator	Serial #	Lot#/Exp	Maintenance By: _____ Date: _____																																				
						0.050	MP5088	202603A	☐ Battery Replacement																																				
								03/03/2028	☐ Dry Gas Regulator Replacement																																				
0.080	MP5089	202603B	☐ Tank Sensor Tare																																										
		03/03/2028	☐ Breath Tube Replacement																																										
0.200	MP5090	202603D	☐ Other:																																										
		03/04/2028																																											
0.080 DGS	N/A	23825080A1																																											
		10/5/2027																																											
<table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th style="width:15%;">Simulator</th> <th style="width:25%;">Serial #</th> <th style="width:20%;">Lot #</th> <th style="width:40%;">Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>		Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Department Inspection By: <u>SLH</u> Barometric Pressure Gauge: _____ ID#: _____ Barometric Pressure ID#: <u>28427</u> Gauge: <u>1014</u> Instrument: <u>1016</u> Mouth Alcohol Solution Lot #: <u>2025-D</u> Exp: <u>09/25/2027</u> Acetone Stock Solution Lot #: <u>2025-B</u> Exp: <u>09/22/2027</u>															
Simulator	Serial #	Lot #	Expiration																																										
0.000		N/A	N/A																																										
0.040																																													
0.100																																													
0.200																																													
0.300																																													
0.080 DGS	N/A																																												
☐ Post Optical Bench Adjustment Stability Checks		<table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th style="width:15%;">Simulator</th> <th style="width:25%;">Serial #</th> <th style="width:20%;">Lot #</th> <th style="width:40%;">Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>		Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			<table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th style="width:15%;">Simulator</th> <th style="width:25%;">Serial #</th> <th style="width:20%;">Lot #</th> <th style="width:40%;">Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>		Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A		
Simulator	Serial #	Lot #	Expiration																																										
0.050																																													
0.080																																													
0.200																																													
0.080 DGS	N/A																																												
Simulator	Serial #	Lot #	Expiration																																										
0.050																																													
0.080																																													
0.200																																													
0.080 DGS	N/A																																												
Gauge ID #: _____ Gauge: _____ Instrument: _____		Attachments ☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☐ Optical Bench Adjustment		☐ Post-Stability Checks ☐ Flow Adjustment ☐ Form 40 ☐ Other:																																									
Notes/Suggested Service:		☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC																																											
		☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use																																											
		☒ Conduct an Agency Inspection Before Evidentiary Use																																											
Digitally signed by Wen-Chi Pierson Date: 2026.07.27 15:52:46 -04'00'		Digitally signed by Shayla Platt Date: 2026.07.29 12:55:41 -04'00'																																											
Tech Review		Admin Review																																											

Stability Checks

80-00 6237

0.050 g/210L 0.047 to 0.053 g/210L	0.080 g/210L 0.077 to 0.083 g/210L	0.200 g/210L 0.194 to 0.206 g/210L	DGS 0.080 g/210L 0.077 to 0.083 g/210L 50.003 g/210L of Wet																																																																																																																																				
CLAY COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 07/22/2026 Software: 8100.27 SN 80-006237 Test g/210L Time <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>13:00</td></tr> <tr><td>Control Test</td><td>0.079</td><td>13:01</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:01</td></tr> <tr><td>Control Test</td><td>0.078</td><td>13:02</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:02</td></tr> <tr><td>Control Test</td><td>0.078</td><td>13:03</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:04</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0783</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7370</td><td></td></tr> </table> Operator Signature <i>M. Miller</i>	Air Blank	0.000	13:00	Control Test	0.079	13:01	Air Blank	0.000	13:01	Control Test	0.078	13:02	Air Blank	0.000	13:02	Control Test	0.078	13:03	Air Blank	0.000	13:04	Control Test Stats			Average	0.0783		Std Dev	0.0006		Rel Std Dev(%)	0.7370		CLAY COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 07/22/2026 Software: 8100.27 SN 80-006237 Test g/210L Time <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>13:00</td></tr> <tr><td>Control Test</td><td>0.079</td><td>13:01</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:01</td></tr> <tr><td>Control Test</td><td>0.078</td><td>13:02</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:02</td></tr> <tr><td>Control Test</td><td>0.078</td><td>13:03</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:04</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0783</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7370</td><td></td></tr> </table> Operator Signature <i>M. Miller</i>	Air Blank	0.000	13:00	Control Test	0.079	13:01	Air Blank	0.000	13:01	Control Test	0.078	13:02	Air Blank	0.000	13:02	Control Test	0.078	13:03	Air Blank	0.000	13:04	Control Test Stats			Average	0.0783		Std Dev	0.0006		Rel Std Dev(%)	0.7370		CLAY COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 07/22/2026 Software: 8100.27 SN 80-006237 Test g/210L Time <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>13:06</td></tr> <tr><td>Control Test</td><td>0.203</td><td>13:07</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:07</td></tr> <tr><td>Control Test</td><td>0.203</td><td>13:08</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:09</td></tr> <tr><td>Control Test</td><td>0.202</td><td>13:09</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:10</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.2027</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.2849</td><td></td></tr> </table> Operator Signature <i>M. Miller</i>	Air Blank	0.000	13:06	Control Test	0.203	13:07	Air Blank	0.000	13:07	Control Test	0.203	13:08	Air Blank	0.000	13:09	Control Test	0.202	13:09	Air Blank	0.000	13:10	Control Test Stats			Average	0.2027		Std Dev	0.0006		Rel Std Dev(%)	0.2849		CLAY COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 07/22/2026 Software: 8100.27 SN 80-006237 Test g/210L Time <table border="1"> <tr><td>Air Blank</td><td>1.000</td><td>13:14</td></tr> <tr><td>Control Test</td><td>1.079</td><td>13:15</td></tr> <tr><td>Air Blank</td><td>1.000</td><td>13:15</td></tr> <tr><td>Control Test</td><td>0.079</td><td>13:15</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:16</td></tr> <tr><td>Control Test</td><td>0.079</td><td>13:16</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:17</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>1.0790</td><td></td></tr> <tr><td>Std Dev</td><td>1.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </table> Operator Signature <i>M. Miller</i>	Air Blank	1.000	13:14	Control Test	1.079	13:15	Air Blank	1.000	13:15	Control Test	0.079	13:15	Air Blank	0.000	13:16	Control Test	0.079	13:16	Air Blank	0.000	13:17	Control Test Stats			Average	1.0790		Std Dev	1.0000		Rel Std Dev(%)	0.0000	
Air Blank	0.000	13:00																																																																																																																																					
Control Test	0.079	13:01																																																																																																																																					
Air Blank	0.000	13:01																																																																																																																																					
Control Test	0.078	13:02																																																																																																																																					
Air Blank	0.000	13:02																																																																																																																																					
Control Test	0.078	13:03																																																																																																																																					
Air Blank	0.000	13:04																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.0783																																																																																																																																						
Std Dev	0.0006																																																																																																																																						
Rel Std Dev(%)	0.7370																																																																																																																																						
Air Blank	0.000	13:00																																																																																																																																					
Control Test	0.079	13:01																																																																																																																																					
Air Blank	0.000	13:01																																																																																																																																					
Control Test	0.078	13:02																																																																																																																																					
Air Blank	0.000	13:02																																																																																																																																					
Control Test	0.078	13:03																																																																																																																																					
Air Blank	0.000	13:04																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.0783																																																																																																																																						
Std Dev	0.0006																																																																																																																																						
Rel Std Dev(%)	0.7370																																																																																																																																						
Air Blank	0.000	13:06																																																																																																																																					
Control Test	0.203	13:07																																																																																																																																					
Air Blank	0.000	13:07																																																																																																																																					
Control Test	0.203	13:08																																																																																																																																					
Air Blank	0.000	13:09																																																																																																																																					
Control Test	0.202	13:09																																																																																																																																					
Air Blank	0.000	13:10																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.2027																																																																																																																																						
Std Dev	0.0006																																																																																																																																						
Rel Std Dev(%)	0.2849																																																																																																																																						
Air Blank	1.000	13:14																																																																																																																																					
Control Test	1.079	13:15																																																																																																																																					
Air Blank	1.000	13:15																																																																																																																																					
Control Test	0.079	13:15																																																																																																																																					
Air Blank	0.000	13:16																																																																																																																																					
Control Test	0.079	13:16																																																																																																																																					
Air Blank	0.000	13:17																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	1.0790																																																																																																																																						
Std Dev	1.0000																																																																																																																																						
Rel Std Dev(%)	0.0000																																																																																																																																						

Verified instrument was able to read gas pressure. ☒

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: CLAY COUNTY SO
Time of Inspection: 12:58

Date of Inspection: 07/24/2026

Serial Number: 80-006237
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202603A Exp: 03/03/2028	0.08g/210L Test (g/210L) Lot#:202603B Exp: 03/03/2028	0.20g/210L Test (g/210L) Lot#:202603D Exp: 03/04/2028	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:23825080A1 Exp: 10/05/2027
0.000	0.049	0.079	0.202	0.079
0.000	0.049	0.080	0.203	0.078
0.000	0.050	0.079	0.203	0.079
0.000	0.049	0.079	0.203	0.078
0.000	0.049	0.079	0.204	0.078
0.000	0.049	0.078	0.204	0.078
0.000	0.049	0.079	0.203	0.079
0.000	0.050	0.079	0.203	0.078
0.000	0.049	0.079	0.204	0.078
0.000	0.050	0.079	0.204	0.078

Standard Deviations	0.0004	0.0004	0.0006	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5


Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.


LEANDRA HIGGINBOTHAM
Signature and Printed Name

07/24/2026
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-006237, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-006237</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>CLAY COUNTY SO</u>	0.050 g/210 L	0.004
Calibration Date:	<u>07/24/2026</u>	0.080 g/210 L	0.004
Calibration Time:	<u>12:58</u>	0.200 g/210 L	0.008
		0.080 g/210 L Dry Gas Control	0.005

All results are reported in g/210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

07/24/2026

Date


LEANDRA HIGGINBOTHAM,
Department Inspector

FDLE/ATP Form 69 January 2026
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality