

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: BROWARD COUNTY SO
Time of Inspection: 12:10

Date of Inspection: 07/11/2025

Serial Number: 80-006925
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08g/210L Test (g/210L) Lot#: _____ Exp: _____	0.20g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: _____ Exp: _____

Number of Simulators Used: _____

Remarks:

AI BYPASSED TO OPERATE INSTRUMENT. COMPLIANCE UNDETERMINED.

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

07/11/2025
Date

Stability Checks

0.05g/210L 0.047 to 0.053	0.08g/210L 0.077 to 0.083	0.20g/210L 0.194 to 0.206	DGS 0.08g/210L 0.077 to 0.083																																																																																																																																																
<p><input checked="" type="checkbox"/></p>	<p><input checked="" type="checkbox"/></p>	<p><input checked="" type="checkbox"/></p>	<p><input checked="" type="checkbox"/> ≤ 0.003 of Wet <input checked="" type="checkbox"/></p>																																																																																																																																																
<p>BROWARD COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-016925 07/11/2025 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>12:33</td></tr> <tr><td>Control Test</td><td>0.051</td><td>12:34</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:34</td></tr> <tr><td>Control Test</td><td>0.051</td><td>12:35</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:36</td></tr> <tr><td>Control Test</td><td>0.051</td><td>12:36</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:37</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0510</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> <p>Operator's Signature <i>MC</i></p>	Test	g/210L	Time	Air Blank	0.000	12:33	Control Test	0.051	12:34	Air Blank	0.000	12:34	Control Test	0.051	12:35	Air Blank	0.000	12:36	Control Test	0.051	12:36	Air Blank	0.000	12:37	Control Test Stats			Average	0.0510		Std Dev	0.0000		Rel Std Dev(%)	0.0000		<p>BROWARD COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-016925 07/11/2025 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>12:41</td></tr> <tr><td>Control Test</td><td>0.081</td><td>12:42</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:42</td></tr> <tr><td>Control Test</td><td>0.081</td><td>12:43</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:43</td></tr> <tr><td>Control Test</td><td>0.081</td><td>12:44</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:45</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0810</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> <p>Operator's Signature <i>MC</i></p>	Test	g/210L	Time	Air Blank	0.000	12:41	Control Test	0.081	12:42	Air Blank	0.000	12:42	Control Test	0.081	12:43	Air Blank	0.000	12:43	Control Test	0.081	12:44	Air Blank	0.000	12:45	Control Test Stats			Average	0.0810		Std Dev	0.0000		Rel Std Dev(%)	0.0000		<p>BROWARD COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-016925 07/11/2025 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>12:48</td></tr> <tr><td>Control Test</td><td>0.203</td><td>12:49</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:49</td></tr> <tr><td>Control Test</td><td>0.203</td><td>12:50</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:50</td></tr> <tr><td>Control Test</td><td>0.202</td><td>12:51</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:52</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.2027</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.2849</td><td></td></tr> </tbody> </table> <p>Operator's Signature <i>MC</i></p>	Test	g/210L	Time	Air Blank	0.000	12:48	Control Test	0.203	12:49	Air Blank	0.000	12:49	Control Test	0.203	12:50	Air Blank	0.000	12:50	Control Test	0.202	12:51	Air Blank	0.000	12:52	Control Test Stats			Average	0.2027		Std Dev	0.0006		Rel Std Dev(%)	0.2849		<p>BROWARD COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-016925 07/11/2025 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>12:52</td></tr> <tr><td>Control Test</td><td>0.078</td><td>12:53</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:53</td></tr> <tr><td>Control Test</td><td>0.077</td><td>12:54</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:54</td></tr> <tr><td>Control Test</td><td>0.077</td><td>12:54</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:55</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0773</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7456</td><td></td></tr> </tbody> </table> <p>Operator's Signature <i>MC</i></p>	Test	g/210L	Time	Air Blank	0.000	12:52	Control Test	0.078	12:53	Air Blank	0.000	12:53	Control Test	0.077	12:54	Air Blank	0.000	12:54	Control Test	0.077	12:54	Air Blank	0.000	12:55	Control Test Stats			Average	0.0773		Std Dev	0.0006		Rel Std Dev(%)	0.7456	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	12:33																																																																																																																																																	
Control Test	0.051	12:34																																																																																																																																																	
Air Blank	0.000	12:34																																																																																																																																																	
Control Test	0.051	12:35																																																																																																																																																	
Air Blank	0.000	12:36																																																																																																																																																	
Control Test	0.051	12:36																																																																																																																																																	
Air Blank	0.000	12:37																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0510																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	12:41																																																																																																																																																	
Control Test	0.081	12:42																																																																																																																																																	
Air Blank	0.000	12:42																																																																																																																																																	
Control Test	0.081	12:43																																																																																																																																																	
Air Blank	0.000	12:43																																																																																																																																																	
Control Test	0.081	12:44																																																																																																																																																	
Air Blank	0.000	12:45																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0810																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	12:48																																																																																																																																																	
Control Test	0.203	12:49																																																																																																																																																	
Air Blank	0.000	12:49																																																																																																																																																	
Control Test	0.203	12:50																																																																																																																																																	
Air Blank	0.000	12:50																																																																																																																																																	
Control Test	0.202	12:51																																																																																																																																																	
Air Blank	0.000	12:52																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.2027																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.2849																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	12:52																																																																																																																																																	
Control Test	0.078	12:53																																																																																																																																																	
Air Blank	0.000	12:53																																																																																																																																																	
Control Test	0.077	12:54																																																																																																																																																	
Air Blank	0.000	12:54																																																																																																																																																	
Control Test	0.077	12:54																																																																																																																																																	
Air Blank	0.000	12:55																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0773																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7456																																																																																																																																																		

Post-Cal Stability Checks

0.05g/210L 0.047 to 0.053	0.08g/210L 0.077 to 0.083	0.20g/210L 0.194 to 0.206	DGS 0.08g/210L 0.077 to 0.083																																																																																																																																																
<p>BROWARD COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-006925 07/23/2025 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>11:31</td></tr> <tr><td>Control Test</td><td>0.049</td><td>11:32</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:32</td></tr> <tr><td>Control Test</td><td>0.049</td><td>11:33</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:34</td></tr> <tr><td>Control Test</td><td>0.049</td><td>11:34</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:35</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0490</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> <p>Operator's Signature _____</p>	Test	g/210L	Time	Air Blank	0.000	11:31	Control Test	0.049	11:32	Air Blank	0.000	11:32	Control Test	0.049	11:33	Air Blank	0.000	11:34	Control Test	0.049	11:34	Air Blank	0.000	11:35	Control Test Stats			Average	0.0490		Std Dev	0.0000		Rel Std Dev(%)	0.0000		<p>BROWARD COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-006925 07/23/2025 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>11:58</td></tr> <tr><td>Control Test</td><td>0.079</td><td>11:58</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:59</td></tr> <tr><td>Control Test</td><td>0.079</td><td>11:59</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:00</td></tr> <tr><td>Control Test</td><td>0.078</td><td>12:01</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:01</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0787</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7339</td><td></td></tr> </tbody> </table> <p>Operator's Signature _____</p>	Test	g/210L	Time	Air Blank	0.000	11:58	Control Test	0.079	11:58	Air Blank	0.000	11:59	Control Test	0.079	11:59	Air Blank	0.000	12:00	Control Test	0.078	12:01	Air Blank	0.000	12:01	Control Test Stats			Average	0.0787		Std Dev	0.0006		Rel Std Dev(%)	0.7339		<p>BROWARD COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-006925 07/23/2025 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>12:04</td></tr> <tr><td>Control Test</td><td>0.198</td><td>12:05</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:06</td></tr> <tr><td>Control Test</td><td>0.197</td><td>12:06</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:07</td></tr> <tr><td>Control Test</td><td>0.198</td><td>12:08</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:08</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1977</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.2921</td><td></td></tr> </tbody> </table> <p>Operator's Signature _____</p>	Test	g/210L	Time	Air Blank	0.000	12:04	Control Test	0.198	12:05	Air Blank	0.000	12:06	Control Test	0.197	12:06	Air Blank	0.000	12:07	Control Test	0.198	12:08	Air Blank	0.000	12:08	Control Test Stats			Average	0.1977		Std Dev	0.0006		Rel Std Dev(%)	0.2921		<p>BROWARD COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-006925 07/23/2025 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>11:14</td></tr> <tr><td>Control Test</td><td>0.078</td><td>11:14</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:15</td></tr> <tr><td>Control Test</td><td>0.078</td><td>11:15</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:15</td></tr> <tr><td>Control Test</td><td>0.078</td><td>11:16</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:17</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0780</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> <p>Operator's Signature _____</p>	Test	g/210L	Time	Air Blank	0.000	11:14	Control Test	0.078	11:14	Air Blank	0.000	11:15	Control Test	0.078	11:15	Air Blank	0.000	11:15	Control Test	0.078	11:16	Air Blank	0.000	11:17	Control Test Stats			Average	0.0780		Std Dev	0.0000		Rel Std Dev(%)	0.0000	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	11:31																																																																																																																																																	
Control Test	0.049	11:32																																																																																																																																																	
Air Blank	0.000	11:32																																																																																																																																																	
Control Test	0.049	11:33																																																																																																																																																	
Air Blank	0.000	11:34																																																																																																																																																	
Control Test	0.049	11:34																																																																																																																																																	
Air Blank	0.000	11:35																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0490																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	11:58																																																																																																																																																	
Control Test	0.079	11:58																																																																																																																																																	
Air Blank	0.000	11:59																																																																																																																																																	
Control Test	0.079	11:59																																																																																																																																																	
Air Blank	0.000	12:00																																																																																																																																																	
Control Test	0.078	12:01																																																																																																																																																	
Air Blank	0.000	12:01																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0787																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7339																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	12:04																																																																																																																																																	
Control Test	0.198	12:05																																																																																																																																																	
Air Blank	0.000	12:06																																																																																																																																																	
Control Test	0.197	12:06																																																																																																																																																	
Air Blank	0.000	12:07																																																																																																																																																	
Control Test	0.198	12:08																																																																																																																																																	
Air Blank	0.000	12:08																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1977																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.2921																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	11:14																																																																																																																																																	
Control Test	0.078	11:14																																																																																																																																																	
Air Blank	0.000	11:15																																																																																																																																																	
Control Test	0.078	11:15																																																																																																																																																	
Air Blank	0.000	11:15																																																																																																																																																	
Control Test	0.078	11:16																																																																																																																																																	
Air Blank	0.000	11:17																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0780																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: BROWARD COUNTY SO
Time of Inspection: 13:37

Date of Inspection: 07/29/2025

Serial Number: 80-006925
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202406K Exp: 06/19/2026	0.08g/210L Test (g/210L) Lot#:202406L Exp: 06/19/2026	0.20g/210L Test (g/210L) Lot#:202406N Exp: 06/20/2026	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG429602 Exp: 10/22/2026
0.000	0.049	0.078	0.197	0.078
0.000	0.048	0.078	0.197	0.080
0.000	0.048	0.078	0.197	0.080
0.000	0.048	0.077	0.197	0.080
0.000	0.049	0.077	0.196	0.080
0.000	0.048	0.078	0.196	0.080
0.000	0.048	0.078	0.197	0.080
0.000	0.048	0.078	0.197	0.080
0.000	0.049	0.078	0.197	0.080
0.000	0.049	0.078	0.197	0.080


Standard Deviations	0.0005	0.0004	0.0004	0.0006
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

07/29/2025
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-006925, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-006925</u>	UNCERTAINTY* \pm	
Owning Agency:	<u>BROWARD COUNTY SO</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>07/29/2025</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>13:37</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Taylor
Gutschow
Digitally signed by Taylor Gutschow
Date: 2025.07.29 15:48:16 -0400'

07/29/2025

Date

TAYLOR D GUTSCHOW,

Department Inspector

FDLE/ATP Form 69 December 2021

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality





Flow Calibration Adjustment(s)

Performed by TDG

BROWARD COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006925
02/05/2025
Software: 8100.27

Flow Rate Calibration*****
1: Rate (Liters/min) = 5
SQRT(Diff)) = 5.383
2: Rate (Liters/min) = 15
SQRT(Diff)) = 10.953
3: Rate (Liters/min) = 30
SQRT(Diff)) = 20.270
Dependent Data Scale Factor = 100000 L/min
Independent Data Scale Factor = 256
Rounded Slope = 653
Rounded Intercept = -373752
Correlation = 0.99957

Stability Checks

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
<p>0.047 to 0.053</p> <p>BROWARD COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-006925 02/05/2025 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 11:21 Control Test 0.000 11:22 Air Blank 0.000 11:22 Control Test 0.000 11:23 Air Blank 0.000 11:24 Control Test 0.000 11:24 Air Blank 0.000 11:25 Control Test Stats Average 0.0000 Std Dev 0.0000 Rel Std Dev(%) 0.0000</p> <p>Operator's Signature </p>	<p>0.077 to 0.083</p> <p>BROWARD COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-006925 02/05/2025 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank* 0.000 11:28 Control Test 0.000 11:28 Air Blank 0.000 11:29 Control Test 0.000 11:29 Air Blank 0.000 11:30 Control Test 0.000 11:31 Air Blank 0.000 11:31 Control Test Stats Average 0.0000 Std Dev 0.0000 Rel Std Dev(%) 0.0000</p> <p>Operator's Signature </p>	<p>0.194 to 0.206</p> <p>BROWARD COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-006925 02/05/2025 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 11:15 Control Test 0.000 11:15 Air Blank 0.000 11:16 Control Test 0.000 11:16 Air Blank 0.000 11:17 Control Test 0.000 11:18 Air Blank 0.000 11:18 Control Test Stats Average 0.0000 Std Dev 0.0000 Rel Std Dev(%) 0.0000</p> <p>Operator's Signature </p>	<p>0.077 to 0.083</p> <p>BROWARD COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-006925 02/05/2025 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 11:03 Control Test 0.076 11:04 Air Blank 0.000 11:04 Control Test 0.076 11:04 Air Blank 0.000 11:05 Control Test 0.076 11:05 Air Blank 0.000 11:06 Control Test Stats Average 0.0760 Std Dev 0.0000 Rel Std Dev(%) 0.0000</p> <p>Operator's Signature </p>

NS

Return Material Authorization

Ship to: CMI, Inc.

Enforcement Electronics

Shipment to repair facility authorized by: Anaya Frazier on 2/5/2025

Items Returned: Instrument Supplies Other Describe: _____
Instrument Model: Intoxilyzer 8000 Serial Number: 80-006925

Bill To Address: <u>Broward County Sheriff's Office</u> <u>Attn: Anaya Frazier</u>	Ship to Address: <u>Florida Department of Law Enforcement</u> <u>Fort Myers Regional Operations Center</u> <u>Attn: Taylor Gutschow</u> <u>4700 Terminal Drive, Suite 1</u> <u>Fort Myers, FL 33907</u>
-------------------------------------------------------------------------------------------------	-------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------

Reason for Return:
The instrument just returned from repair (see work order 409021). It reads simulator solutions as 0 (there are no bubbles in the simulator when a sample is collected). The R-value is close to 100. Instrument also required a flow calibration adjustment, which was performed by FDLE.

Please choose one of the following options:

1. I _____, authorize all repairs.

2. I _____, authorize repairs up to \$ _____.

3. I require an estimate **BEFORE** any repairs will be authorized and/ or conducted.

Please contact: Name: Anaya Frazier
Phone #: 1-305-218-6752 Email: Anaya_Frazier@sheriff.org
ATP Contact Name: Taylor Gutschow ATP Email: TaylorGutschow@fdle.state.fl.us