

## Instrument Processing Sheet

Agency: Miami Beach Police Department Instrument Serial Number: 80-002232  
 Date In: 10/7/2025 DI Completion Date: 11/14/2025  Ship  P/U  H/D  CMI  EE

<b>Intake</b> By: <u>TDG</u> Date: <u>10/28/25</u>	<b>Quality Checks</b> By: <u>TDG</u> Date: <u>(See Notes)</u>	<b>Flow Adjustment</b> By: <u>TDG</u> Date: <u>11/13/25</u>															
<input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input checked="" type="checkbox"/> Return from CMI / EE <input type="checkbox"/> Return unworked <input type="checkbox"/> Training Visual Inspection <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/Accessories <input type="checkbox"/> Power Cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes:	<input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value: <u>198</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column #: <u>ATP104</u> 32 mm <u>0.171*</u> (.139-.169) 36 mm <u>0.187</u> (.156-.190) 53 mm <u>0.253</u> (.228-.278) 103 mm <u>0.507</u> (.447-.547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID #: <u>33364</u> Gauge: <u>1025</u> Instrument: <u>1022</u> <input checked="" type="checkbox"/> Stability Checks <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot#/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>MP6286</td> <td>202406K 6/19/2026</td> </tr> <tr> <td>0.080</td> <td>MP6287</td> <td>202406L 6/19/2026</td> </tr> <tr> <td>0.200</td> <td>MP6288</td> <td>202406N 6/20/2026</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG429602 10/22/2026</td> </tr> </tbody> </table>	Simulator	Serial #	Lot#/Exp	0.050	MP6286	202406K 6/19/2026	0.080	MP6287	202406L 6/19/2026	0.200	MP6288	202406N 6/20/2026	0.080 DGS	N/A	AG429602 10/22/2026	Flow Column #: <u>ATP101</u> <input checked="" type="checkbox"/> 5L/min – 17mm <input checked="" type="checkbox"/> 15L/min – 53mm <input checked="" type="checkbox"/> 30L/min – 103mm <input checked="" type="checkbox"/> R-Value: <u>198</u> <input checked="" type="checkbox"/> Post Adjustment Verification (L/S) Flow Column #: <u>ATP104</u> 32 mm <u>0.144</u> (.139-.169) 36 mm <u>0.160</u> (.156-.190) 53 mm <u>0.230</u> (.228-.278) 103 mm <u>0.500</u> (.447-.547)
Simulator	Serial #	Lot#/Exp															
0.050	MP6286	202406K 6/19/2026															
0.080	MP6287	202406L 6/19/2026															
0.200	MP6288	202406N 6/20/2026															
0.080 DGS	N/A	AG429602 10/22/2026															
		<b>Maintenance</b> By: _____ Date: _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement and Tank Sensor Tare <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other:															

<b>Optical Bench Adjustment</b> By: _____	<b>Department Inspection</b> By: <u>TDG</u>																																								
Barometric Pressure Gauge: _____ ID#: _____	Barometric Pressure ID#: <u>33364</u>																																								
<table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> <input type="checkbox"/> Post Optical Bench Adjustment Stability Checks	Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Gauge: <u>1022</u> Instrument: <u>1019</u> Mouth Alcohol Solution Lot #: <u>2025-C</u> Exp: <u>9/25/2027</u> Acetone Stock Solution Lot #: <u>2024-B</u> Exp: <u>7/19/2026</u> <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>MP6284</td> </tr> <tr> <td>Interferent</td> <td>MP6285</td> </tr> <tr> <td>0.050</td> <td>MP6286</td> </tr> <tr> <td>0.080</td> <td>MP6287</td> </tr> <tr> <td>0.200</td> <td>MP6288</td> </tr> </tbody> </table>	Simulator	Serial Number	0.000	MP6284	Interferent	MP6285	0.050	MP6286	0.080	MP6287	0.200	MP6288
Simulator	Serial #	Lot #	Expiration																																						
0.000		N/A	N/A																																						
0.040																																									
0.100																																									
0.200																																									
0.300																																									
0.080 DGS	N/A																																								
Simulator	Serial Number																																								
0.000	MP6284																																								
Interferent	MP6285																																								
0.050	MP6286																																								
0.080	MP6287																																								
0.200	MP6288																																								
<table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>	Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			<b>Attachments</b> <input checked="" type="checkbox"/> Form 41 <input type="checkbox"/> Post-Stability Checks <input checked="" type="checkbox"/> Stability Checks <input checked="" type="checkbox"/> Flow Adjustment <input checked="" type="checkbox"/> Calibration Certificate <input checked="" type="checkbox"/> Form 40 <input type="checkbox"/> Optical Bench Adjustment <input type="checkbox"/> Other:																				
Simulator	Serial #	Lot #	Expiration																																						
0.050																																									
0.080																																									
0.200																																									
0.080 DGS	N/A																																								
Barometric Pressure Gauge: _____ ID#: _____																																									

Notes/Suggested Service: Checked breath tube screen and replaced o-rings on 10/28. Verified settings, recorded R-value, and performed flow verification and pressure check on 11/13. Performed Stability Checks on 11/14. (TDG 11/14/25) *Outside nominal. Root cause analysis did not identify any user or equipment error. (TDG 11/13/25)	<input checked="" type="checkbox"/> <b>Instrument Complies with Chapter 11D-8, FAC</b> <input type="checkbox"/> <b>Instrument Does Not Comply with Chapter 11D-8, FAC</b> <hr/> <input checked="" type="checkbox"/> <b>Return to/Place into Evidentiary Use</b> <input type="checkbox"/> <b>Remain Out of Evidentiary Use</b> <hr/> <input checked="" type="checkbox"/> <b>Conduct an Agency Inspection Before Evidentiary Use</b>
Shayla Platt Tech Review	Digitally signed by Shayla Platt Date: 2025.11.17 20:43:48 -05'00'
	LeAndra Higginbotham Admin Review
	Digitally signed by LeAndra Higginbotham Date: 2025.11.18 19:08:59 -05'00'

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: MIAMI BEACH PD  
Time of Inspection: 09:48

Date of Inspection: 11/13/2025

Serial Number: 80-002232  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08g/210L Test (g/210L) Lot#: _____ Exp: _____	0.20g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: _____ Exp: _____


Number of Simulators Used: \_\_\_\_\_

Remarks:

AI not conducted. Compliance not determined. ML  
11/13/25

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

11/13/2025  
Date

# Flow Calibration Adjustment(s)

Performed by TDG

MIAMI BEACH PD  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-002232  
11/13/2025  
Software: 8100.27

Flow Rate Calibration\*\*\*\*\*  
1: Rate (Liters/min) = 5  
   SQRT(Diff) ) = 7.211  
2: Rate (Liters/min) = 15  
   SQRT(Diff) ) = 12.039  
3: Rate (Liters/min) = 30  
   SQRT(Diff) ) = 21.164  
Dependent Data Scale Factor = 100000 L/min  
Independent Data Scale Factor = 256  
Rounded Slope = 692  
Rounded Intercept = -721086  
Correlation = 0.99814

# Stability Checks

0.05g/210L 0.047 to 0.053	0.08g/210L 0.077 to 0.083	0.20g/210L 0.194 to 0.206	DGS 0.08g/210L 0.077 to 0.083	≤0.003 of Wet																																																																																																																																															
<p>MIAMI BEACH PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-002232 11/14/2025 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>09:46</td></tr> <tr><td>Control Test</td><td>0.049</td><td>09:47</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:47</td></tr> <tr><td>Control Test</td><td>0.048</td><td>09:48</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:49</td></tr> <tr><td>Control Test</td><td>0.049</td><td>09:49</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:50</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0487</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel. Std Dev(%)</td><td>1.1863</td><td></td></tr> </tbody> </table> <p>Operator's Signature <i>JMC</i></p>	Test	g/210L	Time	Air Blank	0.000	09:46	Control Test	0.049	09:47	Air Blank	0.000	09:47	Control Test	0.048	09:48	Air Blank	0.000	09:49	Control Test	0.049	09:49	Air Blank	0.000	09:50	Control Test Stats			Average	0.0487		Std Dev	0.0006		Rel. Std Dev(%)	1.1863		<p>MIAMI BEACH PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-002232 11/14/2025 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>09:57</td></tr> <tr><td>Control Test</td><td>0.079</td><td>09:58</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:58</td></tr> <tr><td>Control Test</td><td>0.079</td><td>09:59</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:00</td></tr> <tr><td>Control Test</td><td>0.079</td><td>10:00</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:01</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0790</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel. Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> <p>Operator's Signature <i>JMC</i></p>	Test	g/210L	Time	Air Blank	0.000	09:57	Control Test	0.079	09:58	Air Blank	0.000	09:58	Control Test	0.079	09:59	Air Blank	0.000	10:00	Control Test	0.079	10:00	Air Blank	0.000	10:01	Control Test Stats			Average	0.0790		Std Dev	0.0000		Rel. Std Dev(%)	0.0000		<p>MIAMI BEACH PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-002232 11/14/2025 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>10:03</td></tr> <tr><td>Control Test</td><td>0.199</td><td>10:04</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:05</td></tr> <tr><td>Control Test</td><td>0.200</td><td>10:05</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:06</td></tr> <tr><td>Control Test</td><td>0.200</td><td>10:06</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:07</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1997</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel. Std Dev(%)</td><td>0.2892</td><td></td></tr> </tbody> </table> <p>Operator's Signature <i>JMC</i></p>	Test	g/210L	Time	Air Blank	0.000	10:03	Control Test	0.199	10:04	Air Blank	0.000	10:05	Control Test	0.200	10:05	Air Blank	0.000	10:06	Control Test	0.200	10:06	Air Blank	0.000	10:07	Control Test Stats			Average	0.1997		Std Dev	0.0006		Rel. Std Dev(%)	0.2892		<p>MIAMI BEACH PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-002232 11/14/2025 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>09:21</td></tr> <tr><td>Control Test</td><td>0.078</td><td>09:22</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:22</td></tr> <tr><td>Control Test</td><td>0.077</td><td>09:22</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:23</td></tr> <tr><td>Control Test</td><td>0.078</td><td>09:23</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:24</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0777</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel. Std Dev(%)</td><td>0.7434</td><td></td></tr> </tbody> </table> <p>Operator's Signature <i>JMC</i></p>	Test	g/210L	Time	Air Blank	0.000	09:21	Control Test	0.078	09:22	Air Blank	0.000	09:22	Control Test	0.077	09:22	Air Blank	0.000	09:23	Control Test	0.078	09:23	Air Blank	0.000	09:24	Control Test Stats			Average	0.0777		Std Dev	0.0006		Rel. Std Dev(%)	0.7434	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	09:46																																																																																																																																																	
Control Test	0.049	09:47																																																																																																																																																	
Air Blank	0.000	09:47																																																																																																																																																	
Control Test	0.048	09:48																																																																																																																																																	
Air Blank	0.000	09:49																																																																																																																																																	
Control Test	0.049	09:49																																																																																																																																																	
Air Blank	0.000	09:50																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0487																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel. Std Dev(%)	1.1863																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	09:57																																																																																																																																																	
Control Test	0.079	09:58																																																																																																																																																	
Air Blank	0.000	09:58																																																																																																																																																	
Control Test	0.079	09:59																																																																																																																																																	
Air Blank	0.000	10:00																																																																																																																																																	
Control Test	0.079	10:00																																																																																																																																																	
Air Blank	0.000	10:01																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0790																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel. Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	10:03																																																																																																																																																	
Control Test	0.199	10:04																																																																																																																																																	
Air Blank	0.000	10:05																																																																																																																																																	
Control Test	0.200	10:05																																																																																																																																																	
Air Blank	0.000	10:06																																																																																																																																																	
Control Test	0.200	10:06																																																																																																																																																	
Air Blank	0.000	10:07																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1997																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel. Std Dev(%)	0.2892																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	09:21																																																																																																																																																	
Control Test	0.078	09:22																																																																																																																																																	
Air Blank	0.000	09:22																																																																																																																																																	
Control Test	0.077	09:22																																																																																																																																																	
Air Blank	0.000	09:23																																																																																																																																																	
Control Test	0.078	09:23																																																																																																																																																	
Air Blank	0.000	09:24																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0777																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel. Std Dev(%)	0.7434																																																																																																																																																		

# Florida Department of Law Enforcement Alcohol Testing Program

## DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: MIAMI BEACH PD  
Time of Inspection: 13:32

Date of Inspection: 11/14/2025

Serial Number: 80-002232  
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202406K Exp: 06/19/2026	0.08g/210L Test (g/210L) Lot#:202406L Exp: 06/19/2026	0.20g/210L Test (g/210L) Lot#:202406N Exp: 06/20/2026	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG429602 Exp: 10/22/2026
0.000	0.051	0.082	0.201	0.076
0.000	0.051	0.082	0.201	0.076
0.000	0.051	0.082	0.201	0.075
0.000	0.051	0.081	0.202	0.076
0.000	0.051	0.081	0.202	0.075
0.000	0.051	0.081	0.201	0.076
0.000	0.051	0.081	0.201	0.076
0.000	0.051	0.081	0.201	0.075
0.000	0.051	0.081	0.202	0.076
0.000	0.051	0.081	0.201	0.075

Standard Deviations	0.0000	0.0004	0.0004	0.0005
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

  
 \_\_\_\_\_  
 Signature and Printed Name

TAYLOR D GUTSCHOW

11/14/2025  
 Date



# Calibration Certificate

Florida Department of Law Enforcement  
Alcohol Testing Program  
4700 Terminal Drive, Suite 1  
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-002232, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-002232</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>MIAMI BEACH PD</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>11/14/2025</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>13:32</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within  $\pm 0.005$  or 5%, whichever is greater, of the target alcohol concentration.

\*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

## TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Taylor  
Gutschow  
Date: 2025.11.17 12:14:17  
-05'00'

11/14/2025

Date

TAYLOR D GUTSCHOW,

Department Inspector

FDLE/ATP Form 69 December 2021

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality



# INSTRUMENT PROCESSING SHEET

Agency Miami Beach PDS/N 80-002232Florida Department of  
Law EnforcementDate In 03/03/2025 DI Completion Date N/A Ship  P/U  H/D  CMI  EE

<b>Intake</b> By <u>TDG</u> Date <u>03/04/2025</u> <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: <u>Dropped off with box.</u>	<b>Quality Checks</b> By <u>TDG</u> Date <u>03/04/2025</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>81*</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP104</u> 32 mm <u>0.093**</u> (.139 - .169) 36 mm <u>0.109**</u> (.156 - .190) 53 mm <u>0.191**</u> (.228 - .278) 103 mm <u>0.472</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>26932</u> <input checked="" type="checkbox"/> Stability Checks	<b>Flow Calibration</b> By _____ Date _____ Flow Column # _____ <input type="checkbox"/> 5L/min – 17mm <input type="checkbox"/> 15L/min – 53mm <input type="checkbox"/> 30L/min – 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																									
<table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>MP6286</td> <td>202303K 03/29/2025</td> </tr> <tr> <td>0.080</td> <td>MP6287</td> <td>202303L 03/29/2025</td> </tr> <tr> <td>0.200</td> <td>MP6288</td> <td>202304C 04/05/2025</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG429602 10/22/2026</td> </tr> </tbody> </table>		Simulator	Serial #	Lot #/Exp	0.050	MP6286	202303K 03/29/2025	0.080	MP6287	202303L 03/29/2025	0.200	MP6288	202304C 04/05/2025	0.080 DGS	N/A	AG429602 10/22/2026	<b>Maintenance</b> By _____ Date _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____																										
Simulator	Serial #	Lot #/Exp																																									
0.050	MP6286	202303K 03/29/2025																																									
0.080	MP6287	202303L 03/29/2025																																									
0.200	MP6288	202304C 04/05/2025																																									
0.080 DGS	N/A	AG429602 10/22/2026																																									
<b>Calibration Adjustment</b> By _____ Barometric Pressure Gauge _____ ID # _____		<b>Department Inspection</b> By _____ Barometric Pressure ID# _____ Gauge _____ Instrument _____ Mouth Alcohol Solution Lot # _____ Acetone Stock Solution Lot # _____																																									
<table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>		Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			<table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> </tr> <tr> <td>Interferent</td> <td></td> </tr> <tr> <td>0.050</td> <td></td> </tr> <tr> <td>0.080</td> <td></td> </tr> <tr> <td>0.200</td> <td></td> </tr> </tbody> </table>		Simulator	Serial Number	0.000		Interferent		0.050		0.080		0.200	
Simulator	Serial #	Lot #	Expiration																																								
0.000		N/A	N/A																																								
0.040																																											
0.100																																											
0.200																																											
0.300																																											
0.080 DGS	N/A																																										
Simulator	Serial Number																																										
0.000																																											
Interferent																																											
0.050																																											
0.080																																											
0.200																																											
<input type="checkbox"/> Post Calibration Adjustment Stability Checks		<b>Attachments</b> <input type="checkbox"/> Form 41 <input type="checkbox"/> Post-Stability Checks <input checked="" type="checkbox"/> Stability Checks <input type="checkbox"/> Flow Calibration <input type="checkbox"/> Calibration Certificate <input type="checkbox"/> Form 40 <input type="checkbox"/> Calibration Adjustment <input checked="" type="checkbox"/> Other <u>Form 51</u>																																									
Notes/Suggested Service: <u>*R-value is less than 100. (TDG)</u> <u>**Outside nominal range. (TDG)</u> <u>Performed root cause analyses on flow verification and stability checks. Ruled out external equipment and user error. (TDG)</u>		<input type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input type="checkbox"/> Return to/Place into Evidentiary Use <input checked="" type="checkbox"/> Remain Out of Evidentiary Use <input type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use																																									
		Digitally signed by <u>Shayla Platt</u> <small>Destinee Armstrong Date: 2025.03.05 14:28:12 -0500</small>																																									
		Digitally signed by <u>Shayla Platt</u> <small>Shayla Platt Date: 2025.03.06 11:01:21 -05'00'</small>																																									
		Tech Review / Date _____ Admin Review / Date _____																																									

# Stability Checks

0.05g/210L 0.047 to 0.053	0.08g/210L 0.077 to 0.083	0.20g/210L 0.194 to 0.206	DGS 0.08g/210L 0.077 to 0.083																																																																																																																																																
<p>MIAMI BEACH PD Intoxilyzer - Alcohol Analyzer Model 8000 03/04/2025 Software: 8100.27</p> <p>SN 80-002232</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>12:37</td></tr> <tr><td>Control Test</td><td>0.048</td><td>12:38</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:38</td></tr> <tr><td>Control Test</td><td>0.048</td><td>12:39</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:40</td></tr> <tr><td>Control Test</td><td>0.048</td><td>12:40</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:41</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0480</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	12:37	Control Test	0.048	12:38	Air Blank	0.000	12:38	Control Test	0.048	12:39	Air Blank	0.000	12:40	Control Test	0.048	12:40	Air Blank	0.000	12:41	Control Test Stats			Average	0.0480		Std Dev	0.0000		Rel Std Dev(%)	0.0000		<p>MIAMI BEACH PD Intoxilyzer - Alcohol Analyzer Model 8000 03/04/2025 Software: 8100.27</p> <p>SN 80-002232</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>12:16</td></tr> <tr><td>Control Test</td><td>0.078</td><td>12:17</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:17</td></tr> <tr><td>Control Test</td><td>0.078</td><td>12:18</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:19</td></tr> <tr><td>Control Test</td><td>0.077</td><td>12:19</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:20</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0777</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7434</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	12:16	Control Test	0.078	12:17	Air Blank	0.000	12:17	Control Test	0.078	12:18	Air Blank	0.000	12:19	Control Test	0.077	12:19	Air Blank	0.000	12:20	Control Test Stats			Average	0.0777		Std Dev	0.0006		Rel Std Dev(%)	0.7434		<p>MIAMI BEACH PD Intoxilyzer - Alcohol Analyzer Model 8000 03/04/2025 Software: 8100.27</p> <p>SN 80-002232</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>12:24</td></tr> <tr><td>Control Test</td><td>0.196</td><td>12:25</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:25</td></tr> <tr><td>Control Test</td><td>0.194</td><td>12:26</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:26</td></tr> <tr><td>Control Test</td><td>0.193</td><td>12:27</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:27</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1943</td><td></td></tr> <tr><td>Std Dev</td><td>0.0015</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7860</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	12:24	Control Test	0.196	12:25	Air Blank	0.000	12:25	Control Test	0.194	12:26	Air Blank	0.000	12:26	Control Test	0.193	12:27	Air Blank	0.000	12:27	Control Test Stats			Average	0.1943		Std Dev	0.0015		Rel Std Dev(%)	0.7860		<p>MIAMI BEACH PD Intoxilyzer - Alcohol Analyzer Model 8000 03/04/2025 Software: 8100.27</p> <p>SN 80-002232</p> <p>DGS</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>12:53</td></tr> <tr><td>Control Test</td><td>0.077</td><td>12:53</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:54</td></tr> <tr><td>Control Test</td><td>0.077</td><td>12:54</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:54</td></tr> <tr><td>Control Test</td><td>0.077</td><td>12:55</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:55</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0770</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	12:53	Control Test	0.077	12:53	Air Blank	0.000	12:54	Control Test	0.077	12:54	Air Blank	0.000	12:54	Control Test	0.077	12:55	Air Blank	0.000	12:55	Control Test Stats			Average	0.0770		Std Dev	0.0000		Rel Std Dev(%)	0.0000	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	12:37																																																																																																																																																	
Control Test	0.048	12:38																																																																																																																																																	
Air Blank	0.000	12:38																																																																																																																																																	
Control Test	0.048	12:39																																																																																																																																																	
Air Blank	0.000	12:40																																																																																																																																																	
Control Test	0.048	12:40																																																																																																																																																	
Air Blank	0.000	12:41																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0480																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	12:16																																																																																																																																																	
Control Test	0.078	12:17																																																																																																																																																	
Air Blank	0.000	12:17																																																																																																																																																	
Control Test	0.078	12:18																																																																																																																																																	
Air Blank	0.000	12:19																																																																																																																																																	
Control Test	0.077	12:19																																																																																																																																																	
Air Blank	0.000	12:20																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0777																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7434																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	12:24																																																																																																																																																	
Control Test	0.196	12:25																																																																																																																																																	
Air Blank	0.000	12:25																																																																																																																																																	
Control Test	0.194	12:26																																																																																																																																																	
Air Blank	0.000	12:26																																																																																																																																																	
Control Test	0.193	12:27																																																																																																																																																	
Air Blank	0.000	12:27																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1943																																																																																																																																																		
Std Dev	0.0015																																																																																																																																																		
Rel Std Dev(%)	0.7860																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	12:53																																																																																																																																																	
Control Test	0.077	12:53																																																																																																																																																	
Air Blank	0.000	12:54																																																																																																																																																	
Control Test	0.077	12:54																																																																																																																																																	
Air Blank	0.000	12:54																																																																																																																																																	
Control Test	0.077	12:55																																																																																																																																																	
Air Blank	0.000	12:55																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0770																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
<p>Operator's Signature</p> 	<p>Operator's Signature</p> 	<p>Operator's Signature</p> 	<p>Operator's Signature</p> 																																																																																																																																																

**Return Material Authorization**

**Ship to:**     CMI, Inc.  
                   Enforcement Electronics

Shipment to repair facility authorized by: Pedro Socarras on 03/04/2025

**Items Returned:**    Instrument     Supplies     Other  Describe: \_\_\_\_\_

Instrument Model: Intoxilyzer 8000                      Serial Number: 80-002232

**Bill To Address:**  
Miami Beach Police Department  
Attn: Pedro Socarras  
\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_

**Ship to Address:**  
Florida Department of Law Enforcement  
Fort Myers Regional Operations Center  
Attn: Taylor Gutschow  
4700 Terminal Drive, Suite 1  
Fort Myers, FL 33907

**Reason for Return:**  
R-value is less than 100. Instrument needs a flow cal adjust and an optical cal adjust.  
\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_

**Please choose one of the following options:**

1. I \_\_\_\_\_, authorize all repairs.

2. I \_\_\_\_\_, authorize repairs up to \$\_\_\_\_\_.

3. I require an estimate **BEFORE** any repairs will be authorized and/ or conducted.

                  Please contact: Name: Pedro Socarras  
                  Phone #: 1-305-345-6526                      Email: pedrosocarras@miamibeachfl.gov  
ATP Contact Name: Taylor Gutschow                      ATP Email: TaylorGutschow@fdle.state.fl.us