

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

**Agency** ST PETERSBURG PD                      **Date** 03/12/2026                      **Type** Department Inspection  
**Inspector** GUTSCHOW,TAYLOR,D              **Time** 12:00:10                      **Software** 8100.27

**0.05 Lot #** 202406K                      **0.08 Lot #** 202406L                      **0.20 Lot #** 202406N                      **0.08 Gas Lot #** AG526603  
**Expiration** 06/19/2026                      **Expiration** 06/19/2026                      **Expiration** 06/20/2026                      **Expiration** 09/23/2027

**Number of Simulators** 5                      **Average Standard Deviation** 0.0004                      **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.051
		0.08 g/210L Test # 4	0.080
		0.20 g/210L Test # 4	0.201
		0.08 g/210L Dry Gas Std Test # 4	0.078
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.051
		0.08 g/210L Test # 5	0.080
		0.20 g/210L Test # 5	0.201
		0.08 g/210L Dry Gas Std Test # 5	0.078
		Interferent Detect Test # 5	INT

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Inspection Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.050
0.08 g/210L Test # 6	0.080
0.20 g/210L Test # 6	0.201
0.08 g/210L Dry Gas Std Test # 6	0.078
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.051
0.08 g/210L Test # 7	0.080
0.20 g/210L Test # 7	0.201
0.08 g/210L Dry Gas Std Test # 7	0.078
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.051
0.08 g/210L Test # 8	0.080
0.20 g/210L Test # 8	0.201
0.08 g/210L Dry Gas Std Test # 8	0.079
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.050
0.08 g/210L Test # 9	0.081
0.20 g/210L Test # 9	0.201
0.08 g/210L Dry Gas Std Test # 9	0.078
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.080
0.20 g/210L Test # 10	0.200
0.08 g/210L Dry Gas Std Test # 10	0.078
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0005
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0005
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

**Remarks:**

**Data Download Date/Time** 03/12/2026 12:16

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

<b>Agency</b> ST PETERSBURG PD	<b>Date</b> 04/27/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> WILLMES,AUSTIN,J	<b>Time</b> 13:36:30	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 04125080A5
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 03/05/2027

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	Yes	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.052
		0.08 g/210L Test # 3	0.082
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT

**Remarks:**

Time-Date changed. IN COMPLIANCE

**Data Download Date/Time** 04/27/2026 13:44

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

<b>Agency</b> NASSAU COUNTY SO	<b>Date</b> 02/28/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> HIGGINBOTHAM, BRIAN, K	<b>Time</b> 21:48:17	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 04924080A4
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 03/05/2026

**Number of Simulators** 4

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 02/28/2026 23:48

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

<b>Agency</b> NASSAU COUNTY SO	<b>Date</b> 03/30/2026	<b>Type</b> Agency Inspection	
<b>Inspector</b> KUTZ, THOMAS, E	<b>Time</b> 15:00:51	<b>Software</b> 8100.27	
<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202510E	<b>0.20 Lot #</b> 202510F	<b>0.08 Gas Lot #</b> 26-4437
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/28/2027	<b>Expiration</b> 10/29/2027	<b>Expiration</b> 01/29/2028
<b>Number of Simulators</b> 4		<b>Compliance</b> Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 03/30/2026 15:03

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

<b>Agency</b> NASSAU COUNTY SO	<b>Date</b> 04/26/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> HIGGINBOTHAM, BRIAN, K	<b>Time</b> 22:32:55	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202510E	<b>0.20 Lot #</b> 202510F	<b>0.08 Gas Lot #</b> 26-4437
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/28/2027	<b>Expiration</b> 10/29/2027	<b>Expiration</b> 01/29/2028

**Number of Simulators** 4

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 04/26/2026 22:39

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

<b>Agency</b> POLK COUNTY SO	<b>Date</b> 01/20/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> DAWSON, MICHAEL, E	<b>Time</b> 05:58:53	<b>Software</b> 8100.27

<b>0.05 Lot #</b>	<b>0.08 Lot #</b>	<b>0.20 Lot #</b>	<b>0.08 Gas Lot #</b>
<b>Expiration</b>	<b>Expiration</b>	<b>Expiration</b>	<b>Expiration</b>

**Number of Simulators** 5 **Compliance** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	Yes		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

**Remarks:**

A F / M A: . Y.

**Data Download Date/Time** 01/20/2026 06:39

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

<b>Agency</b> POLK COUNTY SO	<b>Date</b> 01/20/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> DAWSON, MICHAEL, E	<b>Time</b> 06:35:18	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202510E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 01524080A2
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/28/2027	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 03/05/2026

**Number of Simulators** 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 01/20/2026 06:39

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

<b>Agency</b> POLK COUNTY SO	<b>Date</b> 02/24/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> DAWSON, MICHAEL, E	<b>Time</b> 07:46:19	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202510E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 15025080A2
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/28/2027	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 07/05/2027

**Number of Simulators** 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.082
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 02/24/2026 07:49

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

<b>Agency</b> POLK COUNTY SO	<b>Date</b> 03/24/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> DAWSON, MICHAEL, E	<b>Time</b> 07:21:06	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202510E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 15025080A2
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/28/2027	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 07/05/2027

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 03/24/2026 07:23

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

<b>Agency</b> POLK COUNTY SO	<b>Date</b> 04/28/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> DAWSON, MICHAEL, E	<b>Time</b> 09:02:38	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202510E	<b>0.08 Lot #</b> 202506E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 15025080A2
<b>Expiration</b> 10/28/2027	<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 07/05/2027

**Number of Simulators** 5

**Compliance** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # (Repeat) 1	0.080
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.048
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.08 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # (Repeat) 2	0.080
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.080
		0.08 g/210L Dry Gas Std Test # (Repeat) 3	0.080
		Interferent Detect Test # 3	INT

**Remarks:**

08:DID NOT ENSURE HOSE WAS CONNECTED FROM DRY GAS CYLINDER Non-compliance: DRY GAS CYLINDER OUT OF TOLERANCE AS NOTED .

**Data Download Date/Time** 04/28/2026 09:46

**Printed: 28 April 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

<b>Agency</b> POLK COUNTY SO	<b>Date</b> 04/28/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> DAWSON, MICHAEL, E	<b>Time</b> 09:43:31	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202510E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 15025080A2
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/28/2027	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 07/05/2027

**Number of Simulators** 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 04/28/2026 09:46

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001056

<b>Agency</b> DAVIE PD	<b>Date</b> 01/09/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> YOUNG,JOSHUA,E	<b>Time</b> 21:30:05	<b>Software</b> 8100.27

<b>0.05 Lot #</b>	<b>0.08 Lot #</b>	<b>0.20 Lot #</b>	<b>0.08 Gas Lot #</b>
<b>Expiration</b>	<b>Expiration</b>	<b>Expiration</b>	<b>Expiration</b>

**Number of Simulators** 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	Interferent Detect Test # 1	0.000
Minimum Sample Volume Check: OK		Alcohol Free Test # 2	0.000
Barometric Pressure Sensor Check: OK		Interferent Detect Test # 2	0.000
Alcohol Free Subject Test: 0.000	Yes	Alcohol Free Test # 3	0.000
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 3	0.000
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	Yes		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

**Remarks:**

Int Det: .

**Data Download Date/Time** 01/09/2026 22:02

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001056

<b>Agency</b> DAVIE PD	<b>Date</b> 01/09/2026	<b>Type</b> Agency Inspection	
<b>Inspector</b> YOUNG,JOSHUA,E	<b>Time</b> 22:03:05	<b>Software</b> 8100.27	
<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 402649340
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 01/27/2026
<b>Number of Simulators</b> 5		<b>Compliance</b> Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 01/09/2026 22:02

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001056

<b>Agency</b> DAVIE PD	<b>Date</b> 02/26/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> YOUNG,JOSHUA,E	<b>Time</b> 03:41:25	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 402649340
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 01/27/2026

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.052
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.203
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 02/26/2026 03:43

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001056

<b>Agency</b> DAVIE PD	<b>Date</b> 03/25/2026	<b>Type</b> Agency Inspection	
<b>Inspector</b> HAINES, DAVID, K	<b>Time</b> 14:51:34	<b>Software</b> 8100.27	
<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 402872586
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 10/12/2026
<b>Number of Simulators</b> 5		<b>Compliance</b> Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.052
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 03/25/2026 14:51

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

<b>Agency</b> LAKELAND PD	<b>Date</b> 01/28/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> DELGADO, RAMON, J	<b>Time</b> 13:17:54	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 04125080A5
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 03/05/2027

**Number of Simulators** 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.191
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.191
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.192
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 01/28/2026 13:28

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

<b>Agency</b> LAKELAND PD	<b>Date</b> 02/25/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> DELGADO, RAMON, J	<b>Time</b> 09:52:12	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202510E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 04125080A5
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/28/2027	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 03/05/2027

**Number of Simulators** 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.192
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.193
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 02/25/2026 09:57

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Inspection Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

<b>Agency</b> LAKELAND PD	<b>Date</b> 04/02/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> PIERSON, WEN-CHI, K	<b>Time</b> 10:59:47	<b>Software</b> 8100.27

<b>0.05 Lot #</b>	<b>0.08 Lot #</b>	<b>0.20 Lot #</b>	<b>0.08 Gas Lot #</b>
<b>Expiration</b>	<b>Expiration</b>	<b>Expiration</b>	<b>Expiration</b>

**Number of Simulators** **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

**Remarks:**

BYPASS AI TO OPERATE INSTRUMENT. COMPLIANCE NOT DETERMINED.

**Data Download Date/Time** 04/16/2026 11:14

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

<b>Agency</b> LAKELAND PD	<b>Date</b> 04/16/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> SPEARIN,KATIE,T	<b>Time</b> 08:58:10	<b>Software</b> 8100.27

<b>0.05 Lot #</b>	<b>0.08 Lot #</b>	<b>0.20 Lot #</b>	<b>0.08 Gas Lot #</b>
<b>Expiration</b>	<b>Expiration</b>	<b>Expiration</b>	<b>Expiration</b>

**Number of Simulators** **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

**Remarks:**

---

**Data Download Date/Time** 04/16/2026 11:14

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

<b>Agency</b> LAKELAND PD	<b>Date</b> 04/16/2026	<b>Type</b> Department Inspection	
<b>Inspector</b> SPEARIN,KATIE,T	<b>Time</b> 11:12:23	<b>Software</b> 8100.27	
<b>0.05 Lot #</b> 202406K	<b>0.08 Lot #</b> 202603B	<b>0.20 Lot #</b> 202406N	<b>0.08 Gas Lot #</b> 23825080A1
<b>Expiration</b> 06/19/2026	<b>Expiration</b> 03/03/2028	<b>Expiration</b> 06/20/2026	<b>Expiration</b> 10/05/2027
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b> 0.0003	<b>Compliance</b> Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.049
		0.08 g/210L Test # 4	0.078
		0.20 g/210L Test # 4	0.198
		0.08 g/210L Dry Gas Std Test # 4	0.078
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.049
		0.08 g/210L Test # 5	0.078
		0.20 g/210L Test # 5	0.197
		0.08 g/210L Dry Gas Std Test # 5	0.078
		Interferent Detect Test # 5	INT

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.049
0.08 g/210L Test # 6	0.078
0.20 g/210L Test # 6	0.197
0.08 g/210L Dry Gas Std Test # 6	0.078
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.049
0.08 g/210L Test # 7	0.078
0.20 g/210L Test # 7	0.197
0.08 g/210L Dry Gas Std Test # 7	0.078
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.050
0.08 g/210L Test # 8	0.078
0.20 g/210L Test # 8	0.198
0.08 g/210L Dry Gas Std Test # 8	0.078
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.049
0.08 g/210L Test # 9	0.078
0.20 g/210L Test # 9	0.198
0.08 g/210L Dry Gas Std Test # 9	0.077
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.049
0.08 g/210L Test # 10	0.078
0.20 g/210L Test # 10	0.198
0.08 g/210L Dry Gas Std Test # 10	0.077
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0000
0.20 g/210L Test (Std Dev)	0.0006
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

**Remarks:**

---

Data Download Date/Time 04/16/2026 11:14

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001058

<b>Agency</b> POLK CSO	<b>Date</b> 01/20/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> LUCAS,NATHANIEL,	<b>Time</b> 09:01:28	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202510E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 15025080A1
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/28/2027	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 07/05/2027

**Number of Simulators** 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT

**Remarks:**

N/A

**Data Download Date/Time** 01/20/2026 08:59

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001058

<b>Agency</b> POLK CSO	<b>Date</b> 02/24/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> LUCAS,NATHANIEL,	<b>Time</b> 08:52:50	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202510E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 15025080A1
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/28/2027	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 07/05/2027

**Number of Simulators** 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.203
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.076
		0.20 g/210L Test # 3	0.203
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

**Remarks:**

N/A

**Data Download Date/Time** 02/24/2026 08:51

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001058

<b>Agency</b> POLK CSO	<b>Date</b> 03/24/2026	<b>Type</b> Agency Inspection	
<b>Inspector</b> WALLISER,SHAUN,J	<b>Time</b> 10:28:01	<b>Software</b> 8100.27	
<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202510E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 15025080A1
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/28/2027	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 07/05/2027

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.203
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 03/24/2026 10:28

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001059

<b>Agency</b> FERNANDINA BEACH PD	<b>Date</b> 01/05/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> SMITH,JASON,E	<b>Time</b> 15:25:11	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 25024080A3
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 10/05/2026

**Number of Simulators** 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.076
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 01/05/2026 15:26

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001059

<b>Agency</b> FERNANDINA BEACH PD	<b>Date</b> 02/04/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> SMITH,JASON,E	<b>Time</b> 17:43:23	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 25024080A3
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 10/05/2026

**Number of Simulators** 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT

**Remarks:**

A F / M A: RFI Detect.

**Data Download Date/Time** 02/04/2026 17:42

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001059

<b>Agency</b> FERNANDINA BEACH PD	<b>Date</b> 03/05/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> SMITH,JASON,E	<b>Time</b> 15:01:14	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 25024080A3
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 10/05/2026

**Number of Simulators** 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.076
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.076
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.076
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 03/05/2026 14:59

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001059

<b>Agency</b> FERNANDINA BEACH PD	<b>Date</b> 03/18/2026	<b>Type</b> Department Inspection	
<b>Inspector</b> HIGGINBOTHAM, LEANDRA,	<b>Time</b> 13:43:02	<b>Software</b> 8100.27	
<b>0.05 Lot #</b> 202406K	<b>0.08 Lot #</b> 202406L	<b>0.20 Lot #</b> 202406N	<b>0.08 Gas Lot #</b> 23825080A1
<b>Expiration</b> 06/19/2026	<b>Expiration</b> 06/19/2026	<b>Expiration</b> 06/20/2026	<b>Expiration</b> 10/05/2027
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b> 0.0004	<b>Compliance</b> Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.195
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.049
		0.08 g/210L Test # 4	0.078
		0.20 g/210L Test # 4	0.195
		0.08 g/210L Dry Gas Std Test # 4	0.080
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.048
		0.08 g/210L Test # 5	0.078
		0.20 g/210L Test # 5	0.195
		0.08 g/210L Dry Gas Std Test # 5	0.080
		Interferent Detect Test # 5	INT

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Inspection Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001059

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.048
0.08 g/210L Test # 6	0.078
0.20 g/210L Test # 6	0.196
0.08 g/210L Dry Gas Std Test # 6	0.081
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.049
0.08 g/210L Test # 7	0.078
0.20 g/210L Test # 7	0.196
0.08 g/210L Dry Gas Std Test # 7	0.081
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.048
0.08 g/210L Test # 8	0.078
0.20 g/210L Test # 8	0.196
0.08 g/210L Dry Gas Std Test # 8	0.080
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.048
0.08 g/210L Test # 9	0.078
0.20 g/210L Test # 9	0.196
0.08 g/210L Dry Gas Std Test # 9	0.081
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.048
0.08 g/210L Test # 10	0.078
0.20 g/210L Test # 10	0.195
0.08 g/210L Dry Gas Std Test # 10	0.081
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0003
0.20 g/210L Test (Std Dev)	0.0005
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

**Remarks:**

**Data Download Date/Time** 03/18/2026 13:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001059

Agency FERNANDINA BEACH PD      Date 04/13/2026      Type Agency Inspection  
 Inspector SMITH,JASON,E      Time 14:51:13      Software 8100.27

0.05 Lot #      0.08 Lot #      0.20 Lot #      0.08 Gas Lot #  
 Expiration      Expiration      Expiration      Expiration

Number of Simulators 5      Compliance No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

**Remarks:**

A F / M A: Improper Sample. Non-compliance:IMPROPER SAMPLE.

Data Download Date/Time 04/13/2026 15:36

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001059

<b>Agency</b> FERNANDINA BEACH PD	<b>Date</b> 04/13/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> SMITH,JASON,E	<b>Time</b> 15:39:12	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 25024080A3
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 10/05/2026

**Number of Simulators** 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.076
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.076
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.076
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 04/13/2026 15:36

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001060

<b>Agency</b> WINTER HAVEN PD	<b>Date</b> 02/02/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> LANGO, MICHAEL, T	<b>Time</b> 17:25:42	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 19825080A1
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 09/05/2027

**Number of Simulators** 4

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	Alcohol Free Test # (Repeat) 1	0.000
Minimum Sample Volume Check: OK		0.05 g/210L Test # 1	0.050
Barometric Pressure Sensor Check: OK		0.08 g/210L Test # 1	0.079
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # 1	0.196
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	Alcohol Free Test # (Repeat) 2	0.000
0.05 g/210L Test: OK	Yes	0.05 g/210L Test # 2	0.051
0.08 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.20 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
		Interferent Detect Test # 2	INT
		Alcohol Free Test # (Repeat) 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

**Remarks:**

00: RFI Detect.

**Data Download Date/Time** 02/02/2026 17:24

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001060

<b>Agency</b> WINTER HAVEN PD	<b>Date</b> 03/10/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> WOMACK,ERIC,	<b>Time</b> 14:25:33	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 19825080A1
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 09/05/2027

**Number of Simulators** 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 03/10/2026 14:23

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001061

<b>Agency</b> GULFPORT PD	<b>Date</b> 01/27/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> MAROTTA, MICHAEL, S	<b>Time</b> 03:07:06	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 04125080A5
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 03/05/2027

**Number of Simulators** 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 01/27/2026 03:02

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001061

<b>Agency</b> GULFPORT PD	<b>Date</b> 02/18/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> MAROTTA, MICHAEL, S	<b>Time</b> 01:58:08	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 04125080A5
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 03/05/2027

**Number of Simulators** 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 02/18/2026 01:54

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001061

<b>Agency</b> GULFPORT PD	<b>Date</b> 03/28/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> MAROTTA, MICHAEL, S	<b>Time</b> 01:53:30	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 04125080A5
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 03/05/2027

**Number of Simulators** 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 04/02/2026 09:30

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001063

<b>Agency</b> HOLLYWOOD PD	<b>Date</b> 01/14/2026	<b>Type</b> Agency Inspection	
<b>Inspector</b> DERIET, RAFAEL,	<b>Time</b> 10:58:54	<b>Software</b> 8100.27	
<b>0.05 Lot #</b> 202402E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 403141809
<b>Expiration</b> 02/27/2026	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 09/10/2027
<b>Number of Simulators</b> 5		<b>Compliance</b> Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.203
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.204
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

**Remarks:**

BSO A. FRAZIER WAS PRESENT.

**Data Download Date/Time** 01/14/2026 11:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001063

<b>Agency</b> HOLLYWOOD PD	<b>Date</b> 02/04/2026	<b>Type</b> Agency Inspection	
<b>Inspector</b> DERIET, RAFAEL,	<b>Time</b> 10:10:32	<b>Software</b> 8100.27	
<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 403141809
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 09/10/2027
<b>Number of Simulators</b> 5		<b>Compliance</b> Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.205
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.205
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.204
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

**Remarks:**

BSO A. FRAZIER WAS PRESENT.

**Data Download Date/Time** 02/04/2026 10:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001063

<b>Agency</b> HOLLYWOOD PD	<b>Date</b> 03/04/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> DERIET,RAFAEL,	<b>Time</b> 10:30:36	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 403141809
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 09/10/2027

  

**Number of Simulators** 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.204
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.204
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.204
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

**Remarks:**

BSO A. FRAZIER WAS PRESENT.

**Data Download Date/Time** 03/04/2026 11:14

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001063

<b>Agency</b> HOLLYWOOD PD	<b>Date</b> 04/01/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> DERIET,RAFAEL,	<b>Time</b> 10:02:19	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 403141809
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 09/10/2027

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.203
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 04/01/2026 10:55

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

<b>Agency</b> NEW PORT RICHEY PD	<b>Date</b> 01/28/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> ENGEL,JASON,S	<b>Time</b> 10:34:55	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202510E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 25925080A1
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/28/2027	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 11/05/2027

**Number of Simulators** 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

**Remarks:**

WIPED INSTRUMENT DOWN

**Data Download Date/Time** 01/28/2026 10:38

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

<b>Agency</b> NEW PORT RICHEY PD	<b>Date</b> 02/21/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> ENGEL,JASON,S	<b>Time</b> 20:07:59	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202510E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 25925080A1
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/28/2027	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 11/05/2027

**Number of Simulators** 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

**Remarks:**

CHECKED TUBING

**Data Download Date/Time** 02/25/2026 18:23

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

<b>Agency</b> NEW PORT RICHEY PD	<b>Date</b> 03/16/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> ENGEL,JASON,S	<b>Time</b> 21:45:28	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202510E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 25925080A1
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/28/2027	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 11/05/2027

**Number of Simulators** 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

**Remarks:**

WIPED DOWN INS

**Data Download Date/Time** 03/16/2026 21:44

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

<b>Agency</b> NEW PORT RICHEY PD	<b>Date</b> 04/13/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> ENGEL,JASON,S	<b>Time</b> 21:22:55	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202510E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 25925080A1
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/28/2027	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 11/05/2027

**Number of Simulators** 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

**Remarks:**

CHECKED TUBING

**Data Download Date/Time** 04/13/2026 21:45

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001065

<b>Agency</b> TARPON SPRINGS PD	<b>Date</b> 01/23/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> BROWN,JEFFREY,F	<b>Time</b> 14:40:06	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202402E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 31124080A1
<b>Expiration</b> 02/27/2026	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 12/05/2026

**Number of Simulators** 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

**Remarks:**

CLEANED AREA

**Data Download Date/Time** 01/23/2026 14:41

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001065

<b>Agency</b> TARPON SPRINGS PD	<b>Date</b> 02/25/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> BROWN,JEFFREY,F	<b>Time</b> 13:55:47	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202402E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 31124080A1
<b>Expiration</b> 02/27/2026	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 12/05/2026

**Number of Simulators** 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT

**Remarks:**

CLEANED AREA.

**Data Download Date/Time** 02/25/2026 13:56



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001065

<b>Agency</b> TARPON SPRINGS PD	<b>Date</b> 03/25/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> BROWN,JEFFREY,F	<b>Time</b> 11:00:52	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 3112408A1
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 12/05/2026

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT

**Remarks:**

CHANGED OUT .20 SIM. WITHIN RANGE. PASSED INSPECTION.

**Data Download Date/Time** 03/25/2026 11:00

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

Agency PINELLAS COUNTY SO	Date 01/07/2026	Type Agency Inspection	
Inspector DEANE, M,	Time 14:24:40	Software 8100.27	

  

0.05 Lot # 202506E	0.08 Lot # 202510E	0.20 Lot # 202410F	0.08 Gas Lot # AG409204
Expiration 06/24/2027	Expiration 10/28/2027	Expiration 10/30/2026	Expiration 04/01/2026

Number of Simulators 5 Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

**Remarks:**

Data Download Date/Time 01/07/2026 14:22

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

<b>Agency</b> PINELLAS COUNTY SO	<b>Date</b> 02/04/2026	<b>Type</b> Agency Inspection	
<b>Inspector</b> DEANE, M,	<b>Time</b> 10:41:03	<b>Software</b> 8100.27	

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202510E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> AG409204
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/28/2027	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 04/01/2026

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 02/04/2026 10:38

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

**Agency** PINELLAS COUNTY SO                      **Date** 02/12/2026                      **Type** Department Inspection  
**Inspector** SPEARIN,KATIE,T                      **Time** 12:19:53                      **Software** 8100.27

**0.05 Lot #** 202406K                      **0.08 Lot #** 202406L                      **0.20 Lot #** 202406N                      **0.08 Gas Lot #** AG510701  
**Expiration** 06/19/2026                      **Expiration** 06/19/2026                      **Expiration** 06/20/2026                      **Expiration** 04/17/2027

**Number of Simulators** 5                      **Average Standard Deviation** 0.0004                      **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.049
		0.08 g/210L Test # 4	0.080
		0.20 g/210L Test # 4	0.200
		0.08 g/210L Dry Gas Std Test # 4	0.080
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.049
		0.08 g/210L Test # 5	0.079
		0.20 g/210L Test # 5	0.199
		0.08 g/210L Dry Gas Std Test # 5	0.080
		Interferent Detect Test # 5	INT

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Inspection Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.049
0.08 g/210L Test # 6	0.079
0.20 g/210L Test # 6	0.200
0.08 g/210L Dry Gas Std Test # 6	0.080
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.080
0.20 g/210L Test # 7	0.200
0.08 g/210L Dry Gas Std Test # 7	0.080
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.050
0.08 g/210L Test # 8	0.080
0.20 g/210L Test # 8	0.200
0.08 g/210L Dry Gas Std Test # 8	0.081
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.050
0.08 g/210L Test # 9	0.079
0.20 g/210L Test # 9	0.200
0.08 g/210L Dry Gas Std Test # 9	0.080
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.080
0.20 g/210L Test # 10	0.200
0.08 g/210L Dry Gas Std Test # 10	0.081
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0005
0.08 g/210L Test (Std Dev)	0.0005
0.20 g/210L Test (Std Dev)	0.0004
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

**Remarks:**

---

**Data Download Date/Time** 02/12/2026 12:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

<b>Agency</b> PINELLAS COUNTY SO	<b>Date</b> 03/04/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> DEANE, M,	<b>Time</b> 09:32:05	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202510E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> AG525301
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/28/2027	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 09/10/2027

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 03/04/2026 09:27

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

Agency PINELLAS COUNTY SO	Date 04/01/2026	Type Agency Inspection	
Inspector DEANE, M,	Time 09:33:49	Software 8100.27	
<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202510E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> AG525301
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/28/2027	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 09/10/2027

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

**Remarks:**

Data Download Date/Time 04/01/2026 09:35

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

<b>Agency</b> CITY OF MIRAMAR POL	<b>Date</b> 01/14/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> HOLLOWAY,DEVARIOUS,J	<b>Time</b> 10:54:22	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202402E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 403345775
<b>Expiration</b> 02/27/2026	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 05/13/2028

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

**Remarks:**

A. FRAZIER PRESENT.

**Data Download Date/Time** 01/14/2026 11:22

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

<b>Agency</b> CITY OF MIRAMAR POL	<b>Date</b> 02/04/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> HOLLOWAY,DEVARIOUS,J	<b>Time</b> 10:03:10	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 403345775
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 05/13/2028

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.08 g/210L Test # (Repeat) 1	0.080
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # 1	0.200
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.051
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.08 g/210L Test: OK	Yes	0.08 g/210L Test # (Repeat) 2	0.080
0.20 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # (Repeat) 3	0.080
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

**Remarks:**

08: RFI Detect.RADIO IN THE ROOM.

**Data Download Date/Time** 02/04/2026 10:34

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

<b>Agency</b> CITY OF MIRAMAR POL	<b>Date</b> 03/04/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> HOLLOWAY,DEVARIOUS,J	<b>Time</b> 10:22:11	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 403345775
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 05/13/2028

**Number of Simulators** 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

**Remarks:**

A FRAIZER PRESENT

**Data Download Date/Time** 03/05/2026 02:13

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

<b>Agency</b> CITY OF MIRAMAR POL	<b>Date</b> 04/01/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> HOLLOWAY,DEVARIOUS,J	<b>Time</b> 09:44:53	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 403345775
<b>Expiration</b> 06/24/2026	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 05/13/2028

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

**Remarks:**

A FRAIZER PRESENT

**Data Download Date/Time** 04/01/2026 10:58

**Printed: 28 April 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

<b>Agency</b> CLEARWATER PD	<b>Date</b> 01/23/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> YEATES,SCOTT,W	<b>Time</b> 18:15:52	<b>Software</b> 8100.27

<b>0.05 Lot #</b>	<b>0.08 Lot #</b>	<b>0.20 Lot #</b>	<b>0.08 Gas Lot #</b>
<b>Expiration</b>	<b>Expiration</b>	<b>Expiration</b>	<b>Expiration</b>

**Number of Simulators** 5 **Compliance** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	Interferent Detect Test # 1	0.000
Minimum Sample Volume Check: OK		Interferent Detect Test # (Repeat) 1	INT
Barometric Pressure Sensor Check: OK		Alcohol Free Test # 2	0.000
Alcohol Free Subject Test: 0.000	Yes	Interferent Detect Test # 2	INT
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # (Repeat) 2	0.000
Interferent Detect Test: Interferent Detect	No	Alcohol Free Test # 3	0.000
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test # 3	INT
Alcohol Free Test: OK	Yes	Interferent Detect Test # (Repeat) 3	INT
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

**Remarks:**

2 ATTEMPT DID NOT INT. Non-compliance: .

**Data Download Date/Time** 01/23/2026 19:32

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

<b>Agency</b> CLEARWATER PD	<b>Date</b> 01/23/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> YEATES,SCOTT,W	<b>Time</b> 18:54:57	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 29425080A2
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 12/05/2027

**Number of Simulators** 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 01/23/2026 19:32

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

<b>Agency</b> CLEARWATER PD	<b>Date</b> 02/19/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> YEATES,SCOTT,W	<b>Time</b> 17:24:19	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 29425080A2
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 12/05/2027

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 02/19/2026 23:04

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

<b>Agency</b> CLEARWATER PD	<b>Date</b> 03/05/2026	<b>Type</b> Agency Inspection	
<b>Inspector</b> YEATES,SCOTT,W	<b>Time</b> 19:40:59	<b>Software</b> 8100.27	
<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 29425080A2
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 12/05/2027
<b>Number of Simulators</b> 5		<b>Compliance</b> Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # (Repeat) 1	0.082
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.049
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.08 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.000
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # (Repeat) 2	0.081
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.000
		0.08 g/210L Dry Gas Std Test # (Repeat) 3	0.082
		Interferent Detect Test # 3	INT

**Remarks:**

08: Control Outside Tolerance.

**Data Download Date/Time** 03/05/2026 19:46

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

<b>Agency</b> CLEARWATER PD	<b>Date</b> 04/16/2026	<b>Type</b> Department Inspection
<b>Inspector</b> SPEARIN,KATIE,T	<b>Time</b> 11:25:17	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202406K	<b>0.08 Lot #</b> 202603B	<b>0.20 Lot #</b> 202406N	<b>0.08 Gas Lot #</b> 23825080A1
<b>Expiration</b> 06/19/2026	<b>Expiration</b> 03/03/2028	<b>Expiration</b> 06/20/2026	<b>Expiration</b> 10/05/2027

  

<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b> 0.0002	<b>Compliance</b> Yes
-------------------------------	--	-----------------------

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.048
		0.08 g/210L Test # 4	0.078
		0.20 g/210L Test # 4	0.198
		0.08 g/210L Dry Gas Std Test # 4	0.081
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.048
		0.08 g/210L Test # 5	0.078
		0.20 g/210L Test # 5	0.198
		0.08 g/210L Dry Gas Std Test # 5	0.081
		Interferent Detect Test # 5	INT

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.048
0.08 g/210L Test # 6	0.078
0.20 g/210L Test # 6	0.199
0.08 g/210L Dry Gas Std Test # 6	0.081
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.048
0.08 g/210L Test # 7	0.078
0.20 g/210L Test # 7	0.199
0.08 g/210L Dry Gas Std Test # 7	0.081
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.048
0.08 g/210L Test # 8	0.078
0.20 g/210L Test # 8	0.199
0.08 g/210L Dry Gas Std Test # 8	0.081
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.048
0.08 g/210L Test # 9	0.078
0.20 g/210L Test # 9	0.199
0.08 g/210L Dry Gas Std Test # 9	0.081
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.049
0.08 g/210L Test # 10	0.078
0.20 g/210L Test # 10	0.198
0.08 g/210L Dry Gas Std Test # 10	0.081
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0003
0.08 g/210L Test (Std Dev)	0.0000
0.20 g/210L Test (Std Dev)	0.0005
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0000

**Remarks:**

Data Download Date/Time 04/16/2026 11:25

Printed: 28 April 2026

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

<b>Agency</b> JACKSONVILLE SO	<b>Date</b> 01/07/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> LIGHTSEY,ROBERT,D	<b>Time</b> 12:38:33	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202402E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 13225080A2
<b>Expiration</b> 02/27/2026	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 06/05/2027

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.074
Barometric Pressure Sensor Check: OK		0.08 g/210L Test # (Repeat) 1	0.077
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # 1	0.192
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.047
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.075
0.08 g/210L Test: OK	Yes	0.08 g/210L Test # (Repeat) 2	0.077
0.20 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.076
		0.08 g/210L Test # (Repeat) 3	0.077
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

**Remarks:**

08: Control Outside CHECKED CONECTIONS RE-RAN OK

**Data Download Date/Time** 01/07/2026 12:38

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

<b>Agency</b> JACKSONVILLE SO	<b>Date</b> 02/20/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> POTTER, MICHAEL, R	<b>Time</b> 12:28:05	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202402E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 13225080A2
<b>Expiration</b> 02/27/2026	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 06/05/2027

**Number of Simulators** 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.045
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.045
		0.08 g/210L Test # 3	0.076
		0.20 g/210L Test # 3	0.195
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 02/20/2026 12:30

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

<b>Agency</b> JACKSONVILLE SO	<b>Date</b> 03/20/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> LIGHTSEY,ROBERT,D	<b>Time</b> 11:49:20	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 36525080A4
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 02/05/2028

**Number of Simulators** 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.083
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.076
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.083
		Interferent Detect Test # 3	INT

**Remarks:**

IN SERV @ PTDF

**Data Download Date/Time** 03/20/2026 11:49

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

<b>Agency</b> JACKSONVILLE SO	<b>Date</b> 04/23/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> LIGHTSEY,ROBERT,D	<b>Time</b> 12:44:50	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 36525080A4
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 02/05/2028

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

**Remarks:**

IN SERV @ PTDF

**Data Download Date/Time** 04/23/2026 13:00

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

<b>Agency</b> CLAY COUNTY SO	<b>Date</b> 01/29/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> JOHNSON, MATTHEW, P	<b>Time</b> 13:01:59	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202402E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 01524080A1
<b>Expiration</b> 02/27/2026	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 03/05/2026

**Number of Simulators** 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

**Remarks:**

MONTHLY INSPECTION

**Data Download Date/Time** 03/06/2026 09:36

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

<b>Agency</b> CLAY COUNTY SO	<b>Date</b> 03/06/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> PIERSON, WEN-CHI, K	<b>Time</b> 09:47:30	<b>Software</b> 8100.27

<b>0.05 Lot #</b>	<b>0.08 Lot #</b>	<b>0.20 Lot #</b>	<b>0.08 Gas Lot #</b>
<b>Expiration</b>	<b>Expiration</b>	<b>Expiration</b>	<b>Expiration</b>

**Number of Simulators** **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

**Remarks:**

BYPASS AI TO OPERATE INSTRUMENT. COMPLIANCE NOT DETERMINED.

**Data Download Date/Time** 04/27/2026 15:38

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001073

<b>Agency</b> DADE CITY PD	<b>Date</b> 01/29/2026	<b>Type</b> Agency Inspection	
<b>Inspector</b> DEL VALLE,M,	<b>Time</b> 18:50:50	<b>Software</b> 8100.27	
<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202510E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 19825080A1
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/28/2027	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 09/05/2027
<b>Number of Simulators</b> 5		<b>Compliance</b> Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

**Remarks:**

NO VISIBLE DAMAGE TO INSTRUMENT

**Data Download Date/Time** 02/02/2026 17:57

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001073

<b>Agency</b> DADE CITY PD	<b>Date</b> 02/27/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> DEL VALLE, M,	<b>Time</b> 05:56:59	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202510E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 19825080A1
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/28/2027	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 09/05/2027

  

<b>Number of Simulators</b> 5	<b>Compliance</b> Yes
-------------------------------	-----------------------

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

**Remarks:**

NO VISIBLE DAMAGE TO INSTRUMENT

**Data Download Date/Time** 03/02/2026 18:39

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001073

<b>Agency</b> DADE CITY PD	<b>Date</b> 03/31/2026	<b>Type</b> Agency Inspection	
<b>Inspector</b> DEL VALLE, M,	<b>Time</b> 01:40:32	<b>Software</b> 8100.27	
<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202510E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 19825080A1
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/28/2027	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 09/05/2027
<b>Number of Simulators</b> 5		<b>Compliance</b> Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

**Remarks:**

NO VISIBLE DAMAGE TO INSTRUMENT

**Data Download Date/Time** 04/03/2026 18:48

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

**Agency** PASCO COUNTY SO                      **Date** 01/15/2026                      **Type** Department Inspection  
**Inspector** SPEARIN,KATIE,T                      **Time** 14:45:38                      **Software** 8100.27

**0.05 Lot #** 202406K                      **0.08 Lot #** 202406L                      **0.20 Lot #** 202406N                      **0.08 Gas Lot #** AG510701  
**Expiration** 06/19/2026                      **Expiration** 06/19/2026                      **Expiration** 06/20/2026                      **Expiration** 04/17/2027

**Number of Simulators** 5                      **Average Standard Deviation** 0.0002                      **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.050
		0.08 g/210L Test # 4	0.079
		0.20 g/210L Test # 4	0.202
		0.08 g/210L Dry Gas Std Test # 4	0.080
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.049
		0.08 g/210L Test # 5	0.079
		0.20 g/210L Test # 5	0.202
		0.08 g/210L Dry Gas Std Test # 5	0.080
		Interferent Detect Test # 5	INT

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.050
0.08 g/210L Test # 6	0.079
0.20 g/210L Test # 6	0.203
0.08 g/210L Dry Gas Std Test # 6	0.080
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.079
0.20 g/210L Test # 7	0.202
0.08 g/210L Dry Gas Std Test # 7	0.080
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.050
0.08 g/210L Test # 8	0.079
0.20 g/210L Test # 8	0.202
0.08 g/210L Dry Gas Std Test # 8	0.080
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.050
0.08 g/210L Test # 9	0.079
0.20 g/210L Test # 9	0.202
0.08 g/210L Dry Gas Std Test # 9	0.080
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.079
0.20 g/210L Test # 10	0.202
0.08 g/210L Dry Gas Std Test # 10	0.080
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0003
0.20 g/210L Test (Std Dev)	0.0004
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0000

#### Remarks:

A F / M A:.

Data Download Date/Time 01/15/2026 15:06

Printed: 28 April 2026

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

<b>Agency</b> PASCO COUNTY SO	<b>Date</b> 02/27/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> FRISCO,RICKY,A	<b>Time</b> 13:24:22	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 10124080A2
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 06/05/2026

**Number of Simulators** 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 02/27/2026 13:33

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

<b>Agency</b> PASCO COUNTY SO	<b>Date</b> 02/27/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> 918504108677,FRISCO,RC	<b>Time</b> 13:27:32	<b>Software</b> 8100.27

<b>0.05 Lot #</b>	<b>0.08 Lot #</b>	<b>0.20 Lot #</b>	<b>0.08 Gas Lot #</b>
<b>Expiration</b>	<b>Expiration</b>	<b>Expiration</b>	<b>Expiration</b>

**Number of Simulators** **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

**Remarks:**

**Data Download Date/Time** 02/27/2026 13:33

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

<b>Agency</b> PASCO COUNTY SO	<b>Date</b> 03/20/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> FRISCO,RICKY,A	<b>Time</b> 12:35:02	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 10124080A2
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 06/05/2026

**Number of Simulators** 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 03/20/2026 13:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001077

<b>Agency</b> HARDEE COUNTY SO	<b>Date</b> 02/01/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> ALBRITTON,GARRETT,M	<b>Time</b> 01:00:36	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 403345775
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 05/13/2028

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 02/01/2026 00:59

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001077

<b>Agency</b> HARDEE COUNTY SO	<b>Date</b> 02/28/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> ALBRITTON,GARRETT,M	<b>Time</b> 15:34:29	<b>Software</b> 8100.27

<b>0.05 Lot #</b>	<b>0.08 Lot #</b>	<b>0.20 Lot #</b>	<b>0.08 Gas Lot #</b>
<b>Expiration</b>	<b>Expiration</b>	<b>Expiration</b>	<b>Expiration</b>

**Number of Simulators** 5 **Compliance** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	Interferent Detect Test # 1	0.000
Minimum Sample Volume Check: OK		Alcohol Free Test # 2	0.000
Barometric Pressure Sensor Check: OK		Interferent Detect Test # 2	0.000
Alcohol Free Subject Test: 0.000	Yes	Alcohol Free Test # 3	0.000
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 3	0.000
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	Yes		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

**Remarks:**

PURGE FAIL. Non-compliance: PURGE FAIL. RESET.

**Data Download Date/Time** 02/28/2026 16:07

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001077

<b>Agency</b> HARDEE COUNTY SO	<b>Date</b> 02/28/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> ALBRITTON,GARRETT,M	<b>Time</b> 16:12:42	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202510E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 403345775
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/28/2027	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 05/13/2028

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 02/28/2026 16:07

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001077

<b>Agency</b> HARDEE COUNTY SO	<b>Date</b> 03/31/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> ALBRITTON,GARRETT,M	<b>Time</b> 20:58:27	<b>Software</b> 8100.27

<b>0.05 Lot #</b>	<b>0.08 Lot #</b>	<b>0.20 Lot #</b>	<b>0.08 Gas Lot #</b>
<b>Expiration</b>	<b>Expiration</b>	<b>Expiration</b>	<b>Expiration</b>

**Number of Simulators** 5 **Compliance** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

**Remarks:**

A F / M A: Ambient Fail Ambient Fail. Non-compliance: .

**Data Download Date/Time** 03/31/2026 21:14



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001077

<b>Agency</b> HARDEE COUNTY SO	<b>Date</b> 03/31/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> ALBRITTON,GARRETT,M	<b>Time</b> 21:12:10	<b>Software</b> 8100.27

<b>0.05 Lot #</b>	<b>0.08 Lot #</b>	<b>0.20 Lot #</b>	<b>0.08 Gas Lot #</b>
<b>Expiration</b>	<b>Expiration</b>	<b>Expiration</b>	<b>Expiration</b>

**Number of Simulators** 5 **Compliance** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

**Remarks:**

A F / M A: Ambient Fail Ambient Fail. Non-compliance: .

**Data Download Date/Time** 03/31/2026 21:14

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

<b>Agency</b> ST PETERSBURG PD	<b>Date</b> 01/21/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> CHOUINARD, MITCHELL, J	<b>Time</b> 08:51:58	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 31124080A1
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 12/05/2026

**Number of Simulators** 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.205
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.204
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.203
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

**Remarks:**

A F / M A: INSUFFICIENT SOLUTION

**Data Download Date/Time** 01/21/2026 09:00

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

<b>Agency</b> ST PETERSBURG PD	<b>Date</b> 02/23/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> CHOUINARD, MITCHELL, J	<b>Time</b> 08:34:31	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 31124080A1
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 12/05/2026

**Number of Simulators** 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

**Remarks:**

IN COMPLIANCE

**Data Download Date/Time** 02/26/2026 08:05

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

<b>Agency</b> ST PETERSBURG PD	<b>Date</b> 03/26/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> WILLMES,AUSTIN,J	<b>Time</b> 14:26:25	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 31124080A1
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 12/05/2026

**Number of Simulators** 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	Yes	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

**Remarks:**

Time-Date changed.

**Data Download Date/Time** 03/26/2026 14:28

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

<b>Agency</b> ST PETERSBURG PD	<b>Date</b> 04/27/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> WILLMES,AUSTIN,J	<b>Time</b> 13:37:15	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 31124080A1
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 12/05/2026

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

**Remarks:**

IN COMPLIANCE

**Data Download Date/Time** 04/27/2026 15:02

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

<b>Agency</b> LARGO PD	<b>Date</b> 01/20/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> GOSIER,MATTHEW,D	<b>Time</b> 10:20:25	<b>Software</b> 8100.27

<b>0.05 Lot #</b>	<b>0.08 Lot #</b>	<b>0.20 Lot #</b>	<b>0.08 Gas Lot #</b>
<b>Expiration</b>	<b>Expiration</b>	<b>Expiration</b>	<b>Expiration</b>

**Number of Simulators** 5 **Compliance** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	Yes		
Mouth Alcohol Test: Slope Not Met	Yes		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

**Remarks:**

USER ERROR.

**Data Download Date/Time** 01/20/2026 12:06

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

<b>Agency</b> LARGO PD	<b>Date</b> 01/20/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> GOSIER, MATTHEW, D	<b>Time</b> 11:06:49	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 23725080A2
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 10/05/2027

**Number of Simulators** 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 01/20/2026 12:06

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

<b>Agency</b> LARGO PD	<b>Date</b> 02/11/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> GOSIER,MATTHEW,D	<b>Time</b> 10:38:54	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 23725080A2
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 10/05/2027

**Number of Simulators** 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.205
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.207
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.207
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

**Remarks:**

NA

**Data Download Date/Time** 02/11/2026 10:36

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

<b>Agency</b> LARGO PD	<b>Date</b> 03/09/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> GOSIER, MATTHEW, D	<b>Time</b> 10:39:19	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 23725080A2
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 10/05/2027

**Number of Simulators** 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.207
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.208
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.209
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 03/09/2026 10:46

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

<b>Agency</b> LARGO PD	<b>Date</b> 04/15/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> GOSIER, MATTHEW, D	<b>Time</b> 09:53:54	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202510E	<b>0.20 Lot #</b> 202510F	<b>0.08 Gas Lot #</b> 23725080A2
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/28/2027	<b>Expiration</b> 10/29/2027	<b>Expiration</b> 10/05/2027

**Number of Simulators** 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

**Remarks:**

N/A

**Data Download Date/Time** 04/15/2026 09:54

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

Agency POLK COUNTY SO	Date 01/30/2026	Type Department Inspection
Inspector SPEARIN,KATIE,T	Time 09:44:31	Software 8100.27

0.05 Lot # 202406K	0.08 Lot # 202406L	0.20 Lot # 202406N	0.08 Gas Lot # AG510701
Expiration 06/19/2026	Expiration 06/19/2026	Expiration 06/20/2026	Expiration 04/17/2027

Number of Simulators 5                      Average Standard Deviation 0.0004                      Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.195
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.049
		0.08 g/210L Test # 4	0.079
		0.20 g/210L Test # 4	0.195
		0.08 g/210L Dry Gas Std Test # 4	0.080
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.049
		0.08 g/210L Test # 5	0.079
		0.20 g/210L Test # 5	0.195
		0.08 g/210L Dry Gas Std Test # 5	0.080
		Interferent Detect Test # 5	INT

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Inspection Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.049
0.08 g/210L Test # 6	0.079
0.20 g/210L Test # 6	0.195
0.08 g/210L Dry Gas Std Test # 6	0.080
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.049
0.08 g/210L Test # 7	0.079
0.20 g/210L Test # 7	0.197
0.08 g/210L Dry Gas Std Test # 7	0.080
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.048
0.08 g/210L Test # 8	0.079
0.20 g/210L Test # 8	0.196
0.08 g/210L Dry Gas Std Test # 8	0.080
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.049
0.08 g/210L Test # 9	0.079
0.20 g/210L Test # 9	0.196
0.08 g/210L Dry Gas Std Test # 9	0.080
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.049
0.08 g/210L Test # 10	0.079
0.20 g/210L Test # 10	0.197
0.08 g/210L Dry Gas Std Test # 10	0.080
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0009
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0000

**Remarks:**

**Data Download Date/Time** 01/30/2026 09:53

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

<b>Agency</b> PASCO COUNTY SO	<b>Date</b> 02/04/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> SPEARIN,KATIE,T	<b>Time</b> 08:41:35	<b>Software</b> 8100.27

<b>0.05 Lot #</b>	<b>0.08 Lot #</b>	<b>0.20 Lot #</b>	<b>0.08 Gas Lot #</b>
<b>Expiration</b>	<b>Expiration</b>	<b>Expiration</b>	<b>Expiration</b>

**Number of Simulators** **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

**Remarks:**

BYPASSED TO OPERATE INSTRUMENT.COMPLIANCE NOT DETERMINED.

**Data Download Date/Time** 02/05/2026 14:38

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

**Agency** PASCO COUNTY SO                      **Date** 02/05/2026                      **Type** Department Inspection  
**Inspector** SPEARIN,KATIE,T                      **Time** 12:11:20                      **Software** 8100.27

**0.05 Lot #** 202406K                      **0.08 Lot #** 202406L                      **0.20 Lot #** 202406N                      **0.08 Gas Lot #** AG510701  
**Expiration** 06/19/2026                      **Expiration** 06/19/2026                      **Expiration** 06/20/2026                      **Expiration** 04/17/2027

**Number of Simulators** 5                      **Average Standard Deviation** 0.0004                      **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.050
		0.08 g/210L Test # 4	0.079
		0.20 g/210L Test # 4	0.201
		0.08 g/210L Dry Gas Std Test # 4	0.079
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.049
		0.08 g/210L Test # 5	0.079
		0.20 g/210L Test # 5	0.201
		0.08 g/210L Dry Gas Std Test # 5	0.080
		Interferent Detect Test # 5	INT

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Inspection Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.050
0.08 g/210L Test # 6	0.080
0.20 g/210L Test # 6	0.200
0.08 g/210L Dry Gas Std Test # 6	0.080
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.079
0.20 g/210L Test # 7	0.201
0.08 g/210L Dry Gas Std Test # 7	0.079
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.050
0.08 g/210L Test # 8	0.080
0.20 g/210L Test # 8	0.201
0.08 g/210L Dry Gas Std Test # 8	0.080
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.050
0.08 g/210L Test # 9	0.080
0.20 g/210L Test # 9	0.200
0.08 g/210L Dry Gas Std Test # 9	0.079
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.080
0.20 g/210L Test # 10	0.200
0.08 g/210L Dry Gas Std Test # 10	0.080
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0005
0.20 g/210L Test (Std Dev)	0.0005
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

**Remarks:**

**Data Download Date/Time** 02/05/2026 14:38

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

<b>Agency</b> PASCO COUNTY SO	<b>Date</b> 02/27/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> FRISCO,RICKY,A	<b>Time</b> 13:13:14	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 10124080A2
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 06/05/2026

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 02/27/2026 13:39

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

<b>Agency</b> PASCO COUNTY SO	<b>Date</b> 03/20/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> FRISCO,RICKY,A	<b>Time</b> 12:37:41	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 10124080A2
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 06/05/2026

**Number of Simulators** 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 03/20/2026 13:04

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

<b>Agency</b> JACKSONVILLE SO	<b>Date</b> 01/07/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> LIGHTSEY,ROBERT,D	<b>Time</b> 10:38:32	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202402E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 23725080A1
<b>Expiration</b> 02/27/2026	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 10/05/2027

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT

**Remarks:**

IN SERV @ PTDF

**Data Download Date/Time** 01/07/2026 10:41

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

<b>Agency</b> JACKSONVILLE SO	<b>Date</b> 02/20/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> POTTER, MICHAEL, R	<b>Time</b> 09:59:22	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202402E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 23725080A1
<b>Expiration</b> 02/27/2026	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 10/05/2027

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.045
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.046
		0.08 g/210L Test # 3	0.076
		0.20 g/210L Test # 3	0.194
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT

**Remarks:**

IN SERV @ PTDF

**Data Download Date/Time** 02/20/2026 09:59

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

<b>Agency</b> JACKSONVILLE SO	<b>Date</b> 03/19/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> LIGHTSEY,ROBERT,D	<b>Time</b> 11:06:15	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 23725080A1
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 10/05/2027

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

**Remarks:**

IN SERV @ PTDF

**Data Download Date/Time** 03/20/2026 07:13

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

<b>Agency</b> FL HIGHWAY PATROL	<b>Date</b> 01/03/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> ABRAHAM,ERIC,D	<b>Time</b> 11:25:56	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202510E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 402843437
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/28/2027	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 09/08/2026

**Number of Simulators** 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.055
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 01/03/2026 12:24

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

<b>Agency</b> FL HIGHWAY PATROL	<b>Date</b> 02/10/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> ABRAHAM,ERIC,D	<b>Time</b> 09:26:02	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202510E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 402843437
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/28/2027	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 09/08/2026

**Number of Simulators** 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 02/10/2026 10:24

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

<b>Agency</b> FL HIGHWAY PATROL	<b>Date</b> 03/06/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> ABRAHAM,ERIC,D	<b>Time</b> 04:49:09	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202510E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 402843437
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/28/2027	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 09/08/2026

**Number of Simulators** 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.187
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # (Repeat) 1	0.190
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.048
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.08 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.188
0.20 g/210L Test: OK	Yes	0.20 g/210L Test # (Repeat) 2	0.191
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.189
		0.20 g/210L Test # (Repeat) 3	0.191
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

**Remarks:**

20: Control Outside Tolerance.

**Data Download Date/Time** 03/06/2026 05:46

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

<b>Agency</b> FL HIGHWAY PATROL	<b>Date</b> 04/11/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> ABRAHAM,ERIC,D	<b>Time</b> 08:19:00	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202510E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 402843437
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/28/2027	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 09/08/2026

**Number of Simulators** 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 04/11/2026 09:28

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001112

<b>Agency</b> FL HIGHWAY PATROL	<b>Date</b> 01/30/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> HITCHCOCK, WILLIAM, D	<b>Time</b> 07:03:29	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 403207507
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 11/26/2027

**Number of Simulators** 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.040
Minimum Sample Volume Check: OK		0.05 g/210L Test # (Repeat) 1	0.050
Barometric Pressure Sensor Check: OK		0.08 g/210L Test # 1	0.081
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # 1	0.202
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.040
0.05 g/210L Test: OK	Yes	0.05 g/210L Test # (Repeat) 2	0.050
0.08 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.20 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.041
		0.05 g/210L Test # (Repeat) 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

**Remarks:**

05: Control Outside REPEATED OK

**Data Download Date/Time** 02/02/2026 10:56

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001112

<b>Agency</b> FL HIGHWAY PATROL	<b>Date</b> 02/28/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> HITCHCOCK, WILLIAM, D	<b>Time</b> 20:04:09	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 403207507
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 11/26/2027

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.044
Minimum Sample Volume Check: OK		0.05 g/210L Test # (Repeat) 1	0.049
Barometric Pressure Sensor Check: OK		0.08 g/210L Test # 1	0.079
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # 1	0.199
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.045
0.05 g/210L Test: OK	Yes	0.05 g/210L Test # (Repeat) 2	0.049
0.08 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.20 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.045
		0.05 g/210L Test # (Repeat) 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

**Remarks:**

05: Control Outside Tolerance.REPEATED OK

**Data Download Date/Time** 02/28/2026 20:07

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001112

<b>Agency</b> FL HIGHWAY PATROL	<b>Date</b> 03/31/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> HITCHCOCK, WILLIAM, D	<b>Time</b> 17:26:48	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 403207507
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 11/26/2027

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 04/01/2026 00:15

**Printed: 28 April 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001114

<b>Agency</b> DELAND PD	<b>Date</b> 01/30/2026	<b>Type</b> Agency Inspection	
<b>Inspector</b> WALTER,CRAIG,K	<b>Time</b> 18:34:40	<b>Software</b> 8100.27	
<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202510E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 402872586
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/28/2027	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 10/12/2026
<b>Number of Simulators</b> 5		<b>Compliance</b> Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

**Remarks:**

MONTHLY INSPECTION

**Data Download Date/Time** 01/30/2026 18:32

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001114

<b>Agency</b> DELAND PD	<b>Date</b> 02/27/2026	<b>Type</b> Agency Inspection	
<b>Inspector</b> WALTER,CRAIG,K	<b>Time</b> 16:12:24	<b>Software</b> 8100.27	
<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202510E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 402872586
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/28/2027	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 10/12/2026
<b>Number of Simulators</b> 5		<b>Compliance</b> Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

**Remarks:**

MONTHLY INSPECTION

**Data Download Date/Time** 02/27/2026 16:07

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001114

<b>Agency</b> DELAND PD	<b>Date</b> 03/31/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> WALTER,CRAIG,K	<b>Time</b> 16:19:52	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202510E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 402872586
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/28/2027	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 10/12/2026

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

**Remarks:**

MONTHLY INSPECTION

**Data Download Date/Time** 03/31/2026 16:18

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001115

<b>Agency</b> FHP	<b>Date</b> 01/12/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> MOSCHETTO,ANTHONY,C	<b>Time</b> 08:58:05	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202402E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 10124080A3
<b>Expiration</b> 02/27/2026	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 06/05/2026

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.052
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.082
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.203
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.052
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.052
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

**Remarks:**

NONE

**Data Download Date/Time** 01/12/2026 10:18

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001115

Agency FHP Date 01/20/2026 Type Department Inspection  
 Inspector PIERSON, WEN-CHI, K Time 12:33:06 Software 8100.27

0.05 Lot # 202406K 0.08 Lot # 202406L 0.20 Lot # 202406N 0.08 Gas Lot # AG510701  
 Expiration 06/19/2026 Expiration 06/19/2026 Expiration 06/20/2026 Expiration 04/17/2027

Number of Simulators 5 Average Standard Deviation 0.0002 Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.050
		0.08 g/210L Test # 4	0.080
		0.20 g/210L Test # 4	0.198
		0.08 g/210L Dry Gas Std Test # 4	0.081
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.050
		0.08 g/210L Test # 5	0.080
		0.20 g/210L Test # 5	0.198
		0.08 g/210L Dry Gas Std Test # 5	0.081
		Interferent Detect Test # 5	INT

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001115

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.050
0.08 g/210L Test # 6	0.080
0.20 g/210L Test # 6	0.199
0.08 g/210L Dry Gas Std Test # 6	0.081
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.080
0.20 g/210L Test # 7	0.198
0.08 g/210L Dry Gas Std Test # 7	0.081
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.050
0.08 g/210L Test # 8	0.080
0.20 g/210L Test # 8	0.199
0.08 g/210L Dry Gas Std Test # 8	0.081
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.051
0.08 g/210L Test # 9	0.080
0.20 g/210L Test # 9	0.199
0.08 g/210L Dry Gas Std Test # 9	0.081
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.080
0.20 g/210L Test # 10	0.199
0.08 g/210L Dry Gas Std Test # 10	0.081
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0000
0.20 g/210L Test (Std Dev)	0.0005
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0000

**Remarks:**

---

Data Download Date/Time 01/20/2026 12:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001115

<b>Agency</b> FHP	<b>Date</b> 02/02/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> MOSCHETTO,ANTHONY,C	<b>Time</b> 09:20:22	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202402E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 10124080A3
<b>Expiration</b> 02/27/2026	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 06/05/2026

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

**Remarks:**

NONE

**Data Download Date/Time** 02/03/2026 12:22

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001115

<b>Agency</b> FHP	<b>Date</b> 03/10/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> MOSCHETTO,ANTHONY,C	<b>Time</b> 07:14:40	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202510E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 10124080A3
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/28/2027	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 06/05/2026

**Number of Simulators** 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

**Remarks:**

NONE

**Data Download Date/Time** 03/10/2026 07:38

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001115

<b>Agency</b> FHP	<b>Date</b> 04/01/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> MOSCHETTO,ANTHONY,C	<b>Time</b> 09:11:58	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202510E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 10124080A3
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/28/2027	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 06/05/2026

**Number of Simulators** 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

**Remarks:**

NONE

**Data Download Date/Time** 04/01/2026 10:07

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001116

<b>Agency</b> EGLIN AFB	<b>Date</b> 01/13/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> TUSKES,ARTHUR,A	<b>Time</b> 08:50:23	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 403374313
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 06/25/2028

  

<b>Number of Simulators</b> 5	<b>Compliance</b> Yes
-------------------------------	-----------------------

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 01/13/2026 09:51

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001116

<b>Agency</b> EGLIN AFB	<b>Date</b> 02/24/2026	<b>Type</b> Agency Inspection	
<b>Inspector</b> FEENEY, ERIC, K	<b>Time</b> 08:41:36	<b>Software</b> 8100.27	
<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 403374313
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 06/25/2028
<b>Number of Simulators</b> 5		<b>Compliance</b> Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 02/25/2026 09:31

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001116

<b>Agency</b> EGLIN AFB	<b>Date</b> 03/15/2026	<b>Type</b> Agency Inspection	
<b>Inspector</b> BURNS,ALVIN,G	<b>Time</b> 12:59:04	<b>Software</b> 8100.27	
<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 403374313
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 06/25/2028
<b>Number of Simulators</b> 5		<b>Compliance</b> Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 03/16/2026 08:10

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001116

<b>Agency</b> EGLIN AFB	<b>Date</b> 04/13/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> MONARCH,BRIAN,L	<b>Time</b> 02:11:38	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 4033743138
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 06/25/2028

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 04/13/2026 03:13

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001118

<b>Agency</b> FL HIGHWAY PATROL	<b>Date</b> 01/19/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> RUBEL,NEIL,F	<b>Time</b> 15:09:01	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202510E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 403308953
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/28/2027	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 04/01/2028

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.195
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

**Remarks:**

N/A

**Data Download Date/Time** 01/19/2026 16:10

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001118

<b>Agency</b> FL HIGHWAY PATROL	<b>Date</b> 02/06/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> RUBEL,NEIL,F	<b>Time</b> 08:47:24	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202510E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 403308953
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/28/2027	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 04/01/2028

**Number of Simulators** 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

**Remarks:**

N/A

**Data Download Date/Time** 02/06/2026 09:53

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001118

<b>Agency</b> FL HIGHWAY PATROL	<b>Date</b> 02/26/2026	<b>Type</b> Department Inspection	
<b>Inspector</b> PIERSON, WEN-CHI, K	<b>Time</b> 14:10:26	<b>Software</b> 8100.27	
<b>0.05 Lot #</b> 202406K	<b>0.08 Lot #</b> 202406L	<b>0.20 Lot #</b> 202406N	<b>0.08 Gas Lot #</b> 28424080A3
<b>Expiration</b> 06/19/2026	<b>Expiration</b> 06/19/2026	<b>Expiration</b> 06/20/2026	<b>Expiration</b> 11/05/2026
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b> 0.0003	<b>Compliance</b> Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.195
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.049
		0.08 g/210L Test # 4	0.079
		0.20 g/210L Test # 4	0.195
		0.08 g/210L Dry Gas Std Test # 4	0.080
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.050
		0.08 g/210L Test # 5	0.079
		0.20 g/210L Test # 5	0.195
		0.08 g/210L Dry Gas Std Test # 5	0.080
		Interferent Detect Test # 5	INT

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001118

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.049
0.08 g/210L Test # 6	0.080
0.20 g/210L Test # 6	0.194
0.08 g/210L Dry Gas Std Test # 6	0.080
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.049
0.08 g/210L Test # 7	0.079
0.20 g/210L Test # 7	0.195
0.08 g/210L Dry Gas Std Test # 7	0.080
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.050
0.08 g/210L Test # 8	0.079
0.20 g/210L Test # 8	0.195
0.08 g/210L Dry Gas Std Test # 8	0.080
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.049
0.08 g/210L Test # 9	0.079
0.20 g/210L Test # 9	0.195
0.08 g/210L Dry Gas Std Test # 9	0.080
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.080
0.20 g/210L Test # 10	0.195
0.08 g/210L Dry Gas Std Test # 10	0.080
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0004
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0000

#### Remarks:

Data Download Date/Time 02/26/2026 15:23

Printed: 28 April 2026

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001118

<b>Agency</b> FL HIGHWAY PATROL	<b>Date</b> 03/11/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> RUBEL,NEIL,F	<b>Time</b> 13:17:25	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202510E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 403308953
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/28/2027	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 04/01/2028

**Number of Simulators** 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.195
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

**Remarks:**

N/A

**Data Download Date/Time** 03/11/2026 15:48

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001118

<b>Agency</b> FL HIGHWAY PATROL	<b>Date</b> 04/04/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> MCDOWELL, MICHAEL, J	<b>Time</b> 15:51:40	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202510E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 403308953
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/28/2027	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 04/01/2028

  

**Number of Simulators** 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

**Remarks:**

N/A

**Data Download Date/Time** 04/04/2026 16:48

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001119

<b>Agency</b> FL HIGHWAY PATROL	<b>Date</b> 01/13/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> HIGGINBOTHAM, LEANDRA,	<b>Time</b> 13:25:05	<b>Software</b> 8100.27

<b>0.05 Lot #</b>	<b>0.08 Lot #</b>	<b>0.20 Lot #</b>	<b>0.08 Gas Lot #</b>
<b>Expiration</b>	<b>Expiration</b>	<b>Expiration</b>	<b>Expiration</b>

**Number of Simulators** **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

**Remarks:**

BYPASS AI FOR OPERATION, COMPLIANCE UNDETERMINED

**Data Download Date/Time** 01/20/2026 09:42

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001119

**Agency** FL HIGHWAY PATROL      **Date** 01/27/2026      **Type** Department Inspection  
**Inspector** HIGGINBOTHAM, LEANDRA,      **Time** 15:02:52      **Software** 8100.27

**0.05 Lot #** 202406K      **0.08 Lot #** 202406L      **0.20 Lot #** 202406N      **0.08 Gas Lot #** AG510701  
**Expiration** 06/19/2026      **Expiration** 06/19/2026      **Expiration** 06/20/2026      **Expiration** 04/17/2027

**Number of Simulators** 5      **Average Standard Deviation** 0.0004      **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.049
		0.08 g/210L Test # 4	0.078
		0.20 g/210L Test # 4	0.196
		0.08 g/210L Dry Gas Std Test # 4	0.079
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.049
		0.08 g/210L Test # 5	0.078
		0.20 g/210L Test # 5	0.195
		0.08 g/210L Dry Gas Std Test # 5	0.079
		Interferent Detect Test # 5	INT

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Inspection Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001119

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.048
0.08 g/210L Test # 6	0.078
0.20 g/210L Test # 6	0.195
0.08 g/210L Dry Gas Std Test # 6	0.079
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.048
0.08 g/210L Test # 7	0.079
0.20 g/210L Test # 7	0.195
0.08 g/210L Dry Gas Std Test # 7	0.079
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.049
0.08 g/210L Test # 8	0.078
0.20 g/210L Test # 8	0.195
0.08 g/210L Dry Gas Std Test # 8	0.078
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.049
0.08 g/210L Test # 9	0.078
0.20 g/210L Test # 9	0.196
0.08 g/210L Dry Gas Std Test # 9	0.079
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.048
0.08 g/210L Test # 10	0.078
0.20 g/210L Test # 10	0.195
0.08 g/210L Dry Gas Std Test # 10	0.079
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0005
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

**Remarks:**

---

**Data Download Date/Time** 01/27/2026 15:03

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001119

<b>Agency</b> FL HIGHWAY PATROL	<b>Date</b> 02/25/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> BUDAY, JOSHUA, M	<b>Time</b> 11:41:07	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 402872586
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 10/12/2026

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

**Remarks:**

UNIT CLEAN/FUNCTIONAL AT TIME OF INSPECTION.

**Data Download Date/Time** 02/25/2026 11:41

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001120

<b>Agency</b> FL HIGHWAY PATROL	<b>Date</b> 02/27/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> GUTSCHOW,TAYLOR,D	<b>Time</b> 15:49:46	<b>Software</b> 8100.27

<b>0.05 Lot #</b>	<b>0.08 Lot #</b>	<b>0.20 Lot #</b>	<b>0.08 Gas Lot #</b>
<b>Expiration</b>	<b>Expiration</b>	<b>Expiration</b>	<b>Expiration</b>

**Number of Simulators** **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

**Remarks:**

AI NOT CONDUCTED. COMPLIANCE NOT DETERMINED.

**Data Download Date/Time** 02/27/2026 15:47

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001120

**Agency** FL HIGHWAY PATROL      **Date** 03/12/2026      **Type** Department Inspection  
**Inspector** GUTSCHOW,TAYLOR,D      **Time** 11:34:55      **Software** 8100.27

**0.05 Lot #** 202406K      **0.08 Lot #** 202406L      **0.20 Lot #** 202406N      **0.08 Gas Lot #** AG429602  
**Expiration** 06/19/2026      **Expiration** 06/19/2026      **Expiration** 06/20/2026      **Expiration** 10/22/2026

**Number of Simulators** 5      **Average Standard Deviation** 0.0001      **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.049
		0.08 g/210L Test # 4	0.079
		0.20 g/210L Test # 4	0.198
		0.08 g/210L Dry Gas Std Test # 4	0.078
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.049
		0.08 g/210L Test # 5	0.079
		0.20 g/210L Test # 5	0.198
		0.08 g/210L Dry Gas Std Test # 5	0.078
		Interferent Detect Test # 5	INT

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Inspection Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001120

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.049
0.08 g/210L Test # 6	0.079
0.20 g/210L Test # 6	0.198
0.08 g/210L Dry Gas Std Test # 6	0.078
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.049
0.08 g/210L Test # 7	0.079
0.20 g/210L Test # 7	0.198
0.08 g/210L Dry Gas Std Test # 7	0.079
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.049
0.08 g/210L Test # 8	0.079
0.20 g/210L Test # 8	0.198
0.08 g/210L Dry Gas Std Test # 8	0.079
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.049
0.08 g/210L Test # 9	0.079
0.20 g/210L Test # 9	0.198
0.08 g/210L Dry Gas Std Test # 9	0.079
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.049
0.08 g/210L Test # 10	0.079
0.20 g/210L Test # 10	0.198
0.08 g/210L Dry Gas Std Test # 10	0.079
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0000
0.08 g/210L Test (Std Dev)	0.0000
0.20 g/210L Test (Std Dev)	0.0000
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

**Remarks:**

**Data Download Date/Time** 03/12/2026 11:43

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001121

<b>Agency</b> FL HIGHWAY PATROL	<b>Date</b> 01/30/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> KEMP,TREVOR,D	<b>Time</b> 18:04:24	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202510E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 403135737
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/28/2027	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 08/30/2027

**Number of Simulators** 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

**Remarks:**

N/A

**Data Download Date/Time** 01/30/2026 18:03

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001121

<b>Agency</b> FL HIGHWAY PATROL	<b>Date</b> 02/22/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> KEMP,TREVOR,D	<b>Time</b> 17:18:16	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202510E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 403135737
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/28/2027	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 08/30/2027

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

**Remarks:**

N/A

**Data Download Date/Time** 02/22/2026 17:31

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001121

<b>Agency</b> FL HIGHWAY PATROL	<b>Date</b> 03/20/2026	<b>Type</b> Agency Inspection	
<b>Inspector</b> KEMP,TREVOR,D	<b>Time</b> 13:29:30	<b>Software</b> 8100.27	
<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202510E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 403268835
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/28/2027	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 02/13/2028
<b>Number of Simulators</b> 5		<b>Compliance</b> Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

**Remarks:**

N/A

**Data Download Date/Time** 03/23/2026 16:24

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001121

<b>Agency</b> FL HIGHWAY PATROL	<b>Date</b> 04/26/2026	<b>Type</b> Agency Inspection	
<b>Inspector</b> KEMP,TREVOR,D	<b>Time</b> 19:46:24	<b>Software</b> 8100.27	
<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202510E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 403268835
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/28/2027	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 02/13/2028
<b>Number of Simulators</b> 5		<b>Compliance</b> Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

**Remarks:**

N/A

**Data Download Date/Time** 04/26/2026 19:53

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001130

**Agency** VOLUSIA COUNTY S.O.      **Date** 01/30/2026      **Type** Agency Inspection  
**Inspector** MALETTO,NICHOLAS,R      **Time** 16:15:55      **Software** 8100.27

**0.05 Lot #** 202506E      **0.08 Lot #** 202510E      **0.20 Lot #** 202410F      **0.08 Gas Lot #** 402872586  
**Expiration** 06/24/2027      **Expiration** 10/28/2027      **Expiration** 10/30/2026      **Expiration** 10/12/2026

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 01/30/2026 16:30

**Printed: 28 April 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001130

**Agency** VOLUSIA COUNTY S.O.      **Date** 02/27/2026      **Type** Agency Inspection  
**Inspector** MALETTO,NICHOLAS,R      **Time** 16:18:38      **Software** 8100.27

**0.05 Lot #** 202506E      **0.08 Lot #** 202510E      **0.20 Lot #** 202510F      **0.08 Gas Lot #** 402872586  
**Expiration** 06/24/2027      **Expiration** 10/28/2027      **Expiration** 10/29/2027      **Expiration** 10/12/2026

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 02/27/2026 16:17

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001130

<b>Agency</b> VOLUSIA COUNTY S.O.	<b>Date</b> 03/29/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> MALETTO, NICHOLAS, R	<b>Time</b> 16:36:41	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202510E	<b>0.20 Lot #</b> 202510F	<b>0.08 Gas Lot #</b> 402872586
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/28/2027	<b>Expiration</b> 10/29/2027	<b>Expiration</b> 10/12/2026

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 03/29/2026 16:43

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001132

<b>Agency</b> VOLUSIA COUNTY SO	<b>Date</b> 03/08/2026	<b>Type</b> Department Inspection	
<b>Inspector</b> HIGGINBOTHAM, LEANDRA,	<b>Time</b> 14:45:21	<b>Software</b> 8100.27	
<b>0.05 Lot #</b> 202406K	<b>0.08 Lot #</b> 202406L	<b>0.20 Lot #</b> 202406N	<b>0.08 Gas Lot #</b> 28424080A3
<b>Expiration</b> 06/19/2026	<b>Expiration</b> 06/19/2026	<b>Expiration</b> 06/20/2026	<b>Expiration</b> 11/05/2026
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b> 0.0004	<b>Compliance</b> Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.046
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.046
		0.08 g/210L Test # 4	0.079
		0.20 g/210L Test # 4	0.199
		0.08 g/210L Dry Gas Std Test # 4	0.079
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.045
		0.08 g/210L Test # 5	0.079
		0.20 g/210L Test # 5	0.200
		0.08 g/210L Dry Gas Std Test # 5	0.078
		Interferent Detect Test # 5	INT

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Inspection Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001132

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.046
0.08 g/210L Test # 6	0.079
0.20 g/210L Test # 6	0.199
0.08 g/210L Dry Gas Std Test # 6	0.078
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.046
0.08 g/210L Test # 7	0.079
0.20 g/210L Test # 7	0.200
0.08 g/210L Dry Gas Std Test # 7	0.079
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.046
0.08 g/210L Test # 8	0.079
0.20 g/210L Test # 8	0.199
0.08 g/210L Dry Gas Std Test # 8	0.078
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.046
0.08 g/210L Test # 9	0.079
0.20 g/210L Test # 9	0.199
0.08 g/210L Dry Gas Std Test # 9	0.078
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.047
0.08 g/210L Test # 10	0.079
0.20 g/210L Test # 10	0.199
0.08 g/210L Dry Gas Std Test # 10	0.078
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0006
0.20 g/210L Test (Std Dev)	0.0004
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

**Remarks:**

**Data Download Date/Time** 03/08/2026 14:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001132

<b>Agency</b> VOLUSIA COUNTY SO	<b>Date</b> 03/29/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> MALETTO,NICHOLAS,R	<b>Time</b> 16:41:44	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202510E	<b>0.20 Lot #</b> 202510F	<b>0.08 Gas Lot #</b> 403342588
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/28/2027	<b>Expiration</b> 10/29/2027	<b>Expiration</b> 05/09/2028

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	Yes	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

**Remarks:**

Time-Date changed.

**Data Download Date/Time** 03/29/2026 16:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

<b>Agency</b> ST JOHNS COUNTY SO	<b>Date</b> 01/29/2026	<b>Type</b> Agency Inspection	
<b>Inspector</b> MILLS,S,T	<b>Time</b> 19:17:42	<b>Software</b> 8100.27	

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 403374313
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 06/25/2028

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	Yes	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

**Remarks:**

Time-Date changed.

**Data Download Date/Time** 02/02/2026 13:10

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

<b>Agency</b> ST JOHNS COUNTY SO	<b>Date</b> 02/27/2026	<b>Type</b> Agency Inspection	
<b>Inspector</b> MILLS,SCOTT,T	<b>Time</b> 17:31:00	<b>Software</b> 8100.27	

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202510E	<b>0.20 Lot #</b> 202510F	<b>0.08 Gas Lot #</b> 403374313
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/28/2027	<b>Expiration</b> 10/29/2027	<b>Expiration</b> 06/25/2028

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 02/27/2026 17:33

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

<b>Agency</b> ST JOHNS COUNTY SO	<b>Date</b> 03/25/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> MILLS,SCOTT,T	<b>Time</b> 20:29:27	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202510F	<b>0.20 Lot #</b> 202510F
<b>0.08 Gas Lot #</b> 403374313		
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/29/2027	<b>Expiration</b> 10/29/2027
		<b>Expiration</b> 06/25/2028
<b>Number of Simulators</b> 5		<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.08 g/210L Test # (Repeat) 1	0.079
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # 1	0.200
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.049
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.000
0.08 g/210L Test: OK	Yes	0.08 g/210L Test # (Repeat) 2	0.079
0.20 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.000
		0.08 g/210L Test # (Repeat) 3	0.080
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

**Remarks:**

08: Control Outside Tolerance.

**Data Download Date/Time** 03/25/2026 21:15

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

<b>Agency</b> ST JOHNS COUNTY SO	<b>Date</b> 03/25/2026	<b>Type</b> Agency Inspection	
<b>Inspector</b> MILLS,SCOTT,T	<b>Time</b> 21:13:59	<b>Software</b> 8100.27	

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202510E	<b>0.20 Lot #</b> 202510F	<b>0.08 Gas Lot #</b> 403374313
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/28/2027	<b>Expiration</b> 10/29/2027	<b>Expiration</b> 06/25/2028

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 03/25/2026 21:15

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

<b>Agency</b> ST JOHNS COUNTY SO	<b>Date</b> 01/29/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> MILLS,S,T	<b>Time</b> 19:30:21	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 403374313
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 06/25/2028

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	Yes	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

**Remarks:**

Time-Date changed.

**Data Download Date/Time** 02/02/2026 13:17

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

<b>Agency</b> ST JOHNS COUNTY SO	<b>Date</b> 02/27/2026	<b>Type</b> Agency Inspection	
<b>Inspector</b> MILLS,SCOTT,T	<b>Time</b> 17:41:58	<b>Software</b> 8100.27	

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202510E	<b>0.20 Lot #</b> 202510F	<b>0.08 Gas Lot #</b> 403525131
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/28/1027	<b>Expiration</b> 10/29/2027	<b>Expiration</b> 01/02/2029

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 02/27/2026 17:42

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

<b>Agency</b> ST JOHNS COUNTY SO	<b>Date</b> 03/25/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> MILLS,SCOTT,T	<b>Time</b> 21:35:04	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202510E	<b>0.20 Lot #</b> 202510F	<b>0.08 Gas Lot #</b> 403525131
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/28/2027	<b>Expiration</b> 10/29/2027	<b>Expiration</b> 01/02/2029

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 03/26/2026 19:00

**Printed: 28 April 2026**

COBRA ©1997-2014 CMI, Inc.