

Alcohol Testing Program - Instrument Processing Sheet

Agency: Nassau County SO

Instrument Serial Number: 80-001281

Date In: 6/5/2026

DI Completion Date: 6/25/2026

☐ Ship ☒ P/U ☐ H/D ☐ CMI ☐ EE

Intake By: <u>IAS</u> Date: <u>6/5/2026</u>		Quality Checks By: <u>SLH</u> Date: <u>6/12/2026</u>		Flow Adjustment By: _____																															
☒ Annual ☐ Dropped Off ☐ Registration ☐ Return from CMI / EE ☐ Training Instrument Visual Inspection ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/Accessories ☐ Power Cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes:		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value: <u>154</u> ☒ Flow Verification (L/s) Flow Column #: <u>ATP 102</u> 32 mm <u>0.140</u> (.139-.169) 36 mm <u>0.156</u> (.156-.190) 53 mm <u>0.230</u> (.228-.278) 103 mm <u>0.492</u> (.447-.547) ☒ Barometric Pressure Check Gauge ID #: <u>28427</u> Gauge: <u>1008</u> Instrument: <u>1009</u> ☒ Stability Checks		Flow Column #: _____ ☐ 5L/min – 17mm ☐ 15L/min – 53mm ☐ 30L/min – 103mm ☐ R-Value: _____ ☐ Post Adjustment Verification (L/S) Flow Column #: _____ 32 mm _____ (.139-.169) 36 mm _____ (.156-.190) 53 mm _____ (.228-.278) 103 mm _____ (.447-.547)																															
						<table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th style="width: 15%;">Simulator</th> <th style="width: 25%;">Serial #</th> <th style="width: 20%;">Lot#/Exp</th> <th style="width: 40%;">Maintenance By: <u>SLH</u> Date: <u>6/12/2026</u></th> </tr> </thead> <tbody> <tr> <td rowspan="2">0.050</td> <td rowspan="2">MP5088</td> <td>202603A</td> <td>☐ Battery Replacement</td> </tr> <tr> <td>03/03/2028</td> <td>☐ Dry Gas Regulator Replacement</td> </tr> <tr> <td rowspan="2">0.080</td> <td rowspan="2">MP5089</td> <td>202603B</td> <td>☐ Tank Sensor Tare</td> </tr> <tr> <td>03/03/2028</td> <td>☐ Breath Tube Replacement</td> </tr> <tr> <td rowspan="2">0.200</td> <td rowspan="2">MP5090</td> <td>202603D</td> <td>☒ Other: CMI remotored into instrument to clear error.</td> </tr> <tr> <td>03/04/2028</td> <td></td> </tr> <tr> <td rowspan="2">0.080 DGS</td> <td rowspan="2">N/A</td> <td>23825080A1</td> <td></td> </tr> <tr> <td>10/5/2027</td> <td></td> </tr> </tbody> </table>		Simulator	Serial #	Lot#/Exp	Maintenance By: <u>SLH</u> Date: <u>6/12/2026</u>	0.050	MP5088	202603A	☐ Battery Replacement	03/03/2028	☐ Dry Gas Regulator Replacement	0.080	MP5089	202603B	☐ Tank Sensor Tare	03/03/2028	☐ Breath Tube Replacement	0.200	MP5090	202603D	☒ Other: CMI remotored into instrument to clear error.	03/04/2028		0.080 DGS	N/A	23825080A1		10/5/2027	
						Simulator	Serial #	Lot#/Exp	Maintenance By: <u>SLH</u> Date: <u>6/12/2026</u>																										
						0.050	MP5088	202603A	☐ Battery Replacement																										
								03/03/2028	☐ Dry Gas Regulator Replacement																										
0.080	MP5089	202603B	☐ Tank Sensor Tare																																
		03/03/2028	☐ Breath Tube Replacement																																
0.200	MP5090	202603D	☒ Other: CMI remotored into instrument to clear error.																																
		03/04/2028																																	
0.080 DGS	N/A	23825080A1																																	
		10/5/2027																																	
<table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th style="width: 15%;">Simulator</th> <th style="width: 25%;">Serial #</th> <th style="width: 20%;">Lot #</th> <th style="width: 40%;">Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>MP6294</td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td>MP6295</td> <td>25090</td> <td>3/11/27</td> </tr> <tr> <td>0.100</td> <td>MP6296</td> <td>24390</td> <td>10/29/26</td> </tr> <tr> <td>0.200</td> <td>MP6297</td> <td>25020</td> <td>1/14/27</td> </tr> <tr> <td>0.300</td> <td>MP6298</td> <td>24430</td> <td>12/10/26</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG510701</td> <td>4/17/27</td> </tr> </tbody> </table>		Simulator	Serial #	Lot #	Expiration	0.000	MP6294	N/A	N/A	0.040	MP6295	25090	3/11/27	0.100	MP6296	24390	10/29/26	0.200	MP6297	25020	1/14/27	0.300	MP6298	24430	12/10/26	0.080 DGS	N/A	AG510701	4/17/27						
Simulator	Serial #	Lot #	Expiration																																
0.000	MP6294	N/A	N/A																																
0.040	MP6295	25090	3/11/27																																
0.100	MP6296	24390	10/29/26																																
0.200	MP6297	25020	1/14/27																																
0.300	MP6298	24430	12/10/26																																
0.080 DGS	N/A	AG510701	4/17/27																																
<table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th style="width: 15%;">Simulator</th> <th style="width: 25%;">Serial #</th> <th style="width: 20%;">Lot #</th> <th style="width: 40%;">Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>MP5088</td> <td>202603A</td> <td>3/3/28</td> </tr> <tr> <td>0.080</td> <td>MP5089</td> <td>202603B</td> <td>3/3/28</td> </tr> <tr> <td>0.200</td> <td>MP5090</td> <td>202603D</td> <td>3/4/28</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>23825080A1</td> <td>10/5/27</td> </tr> </tbody> </table>		Simulator	Serial #	Lot #	Expiration	0.050	MP5088	202603A	3/3/28	0.080	MP5089	202603B	3/3/28	0.200	MP5090	202603D	3/4/28	0.080 DGS	N/A	23825080A1	10/5/27														
Simulator	Serial #	Lot #	Expiration																																
0.050	MP5088	202603A	3/3/28																																
0.080	MP5089	202603B	3/3/28																																
0.200	MP5090	202603D	3/4/28																																
0.080 DGS	N/A	23825080A1	10/5/27																																
Gauge ID #: <u>28427</u> Gauge: <u>1008</u> Instrument: <u>1011</u>		Gauge: <u>1014</u> Instrument: <u>1014</u> Mouth Alcohol Solution Lot #: 2025D 2025-D Exp: <u>09/25/2027</u> Acetone Stock Solution Lot #: 2025B 2025-B Exp: <u>09/23/2027</u>																																	
Notes/Suggested Service: Admin Review - corrected acetone expiration date and added dashes to the mouth and acetone solution lot #. SLH 6/25/2026		☒ Post Optical Bench Adjustment Stability Checks ☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☒ Optical Bench Adjustment																																	
		☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use																																	
Taylor Gutschow Digitally signed by Taylor Gutschow Date: 2026.06.25 12:34:58 -04'00'		Wen-Chi Pierson Digitally signed by Wen-Chi Pierson Date: 2026.06.25 13:18:25 -04'00'																																	
				Tech Review Admin Review																															

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: NASSAU COUNTY SO
Time of Inspection: 17:07

Date of Inspection: 06/12/2026

Serial Number: 80-001281
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:


Number of Simulators Used: _____

Remarks:

BYPASS AI FOR OPERATION, COMPLIANCE UNDETERMINED

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.


 LEANDRA HIGGINBOTHAM
 Signature and Printed Name

06/12/2026
Date

optical bench adjustment will be performed
Stability Checks

80-001281

0.050 g/210L 0.047 to 0.053 g/210L	0.080 g/210L 0.077 to 0.083 g/210L	0.200 g/210L 0.194 to 0.206 g/210L	DGS 0.080 g/210L 0.077 to 0.083 g/210L ≤0.003 g/210L of Wet																																																																																																																																																
NASSAU COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001281 06/12/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>17:35</td></tr> <tr><td>Control Test</td><td>0.047</td><td>17:36</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>17:36</td></tr> <tr><td>Control Test</td><td>0.046</td><td>17:37</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>17:38</td></tr> <tr><td>Control Test</td><td>0.046</td><td>17:38</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>17:39</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0463</td><td></td></tr> <tr><td>Std Dev</td><td>0.0005</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.2461</td><td></td></tr> </tbody> </table> Operator's Signature: <i>[Signature]</i>	Test	g/210L	Time	Air Blank	0.000	17:35	Control Test	0.047	17:36	Air Blank	0.000	17:36	Control Test	0.046	17:37	Air Blank	0.000	17:38	Control Test	0.046	17:38	Air Blank	0.000	17:39	Control Test Stats			Average	0.0463		Std Dev	0.0005		Rel Std Dev(%)	1.2461		NASSAU COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001281 06/12/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>17:41</td></tr> <tr><td>Control Test</td><td>0.076</td><td>17:41</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>17:42</td></tr> <tr><td>Control Test</td><td>0.076</td><td>17:43</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>17:43</td></tr> <tr><td>Control Test</td><td>0.076</td><td>17:44</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>17:44</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0760</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> Operator's Signature: <i>[Signature]</i>	Test	g/210L	Time	Air Blank	0.000	17:41	Control Test	0.076	17:41	Air Blank	0.000	17:42	Control Test	0.076	17:43	Air Blank	0.000	17:43	Control Test	0.076	17:44	Air Blank	0.000	17:44	Control Test Stats			Average	0.0760		Std Dev	0.0000		Rel Std Dev(%)	0.0000		NASSAU COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001281 06/12/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>17:28</td></tr> <tr><td>Control Test</td><td>0.193</td><td>17:28</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>17:29</td></tr> <tr><td>Control Test</td><td>0.194</td><td>17:30</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>17:30</td></tr> <tr><td>Control Test</td><td>0.193</td><td>17:31</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>17:31</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1933</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.2986</td><td></td></tr> </tbody> </table> Operator's Signature: <i>[Signature]</i>	Test	g/210L	Time	Air Blank	0.000	17:28	Control Test	0.193	17:28	Air Blank	0.000	17:29	Control Test	0.194	17:30	Air Blank	0.000	17:30	Control Test	0.193	17:31	Air Blank	0.000	17:31	Control Test Stats			Average	0.1933		Std Dev	0.0006		Rel Std Dev(%)	0.2986		NASSAU COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001281 05/12/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>17:48</td></tr> <tr><td>Control Test</td><td>0.079</td><td>17:48</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>17:48</td></tr> <tr><td>Control Test</td><td>0.080</td><td>17:49</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>17:49</td></tr> <tr><td>Control Test</td><td>0.080</td><td>17:50</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>17:50</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0797</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7247</td><td></td></tr> </tbody> </table> Operator's Signature: <i>[Signature]</i>	Test	g/210L	Time	Air Blank	0.000	17:48	Control Test	0.079	17:48	Air Blank	0.000	17:48	Control Test	0.080	17:49	Air Blank	0.000	17:49	Control Test	0.080	17:50	Air Blank	0.000	17:50	Control Test Stats			Average	0.0797		Std Dev	0.0006		Rel Std Dev(%)	0.7247	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	17:35																																																																																																																																																	
Control Test	0.047	17:36																																																																																																																																																	
Air Blank	0.000	17:36																																																																																																																																																	
Control Test	0.046	17:37																																																																																																																																																	
Air Blank	0.000	17:38																																																																																																																																																	
Control Test	0.046	17:38																																																																																																																																																	
Air Blank	0.000	17:39																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0463																																																																																																																																																		
Std Dev	0.0005																																																																																																																																																		
Rel Std Dev(%)	1.2461																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	17:41																																																																																																																																																	
Control Test	0.076	17:41																																																																																																																																																	
Air Blank	0.000	17:42																																																																																																																																																	
Control Test	0.076	17:43																																																																																																																																																	
Air Blank	0.000	17:43																																																																																																																																																	
Control Test	0.076	17:44																																																																																																																																																	
Air Blank	0.000	17:44																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0760																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	17:28																																																																																																																																																	
Control Test	0.193	17:28																																																																																																																																																	
Air Blank	0.000	17:29																																																																																																																																																	
Control Test	0.194	17:30																																																																																																																																																	
Air Blank	0.000	17:30																																																																																																																																																	
Control Test	0.193	17:31																																																																																																																																																	
Air Blank	0.000	17:31																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1933																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.2986																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	17:48																																																																																																																																																	
Control Test	0.079	17:48																																																																																																																																																	
Air Blank	0.000	17:48																																																																																																																																																	
Control Test	0.080	17:49																																																																																																																																																	
Air Blank	0.000	17:49																																																																																																																																																	
Control Test	0.080	17:50																																																																																																																																																	
Air Blank	0.000	17:50																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0797																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7247																																																																																																																																																		

Verified instrument was
able to read gas pressure. ☒

NASSAU COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001281
06/16/2026 09:27:59

Auto Calibration
Max Power Res Value = 35
Auto Range Res Value = 22

Sol Value = 0.000 g/210L ***
Fit value = 0.0000 mg/l %%%
Samples Taken = 4, Discarded = 1
3um Io = 12637, 9um Io = 13536

Channel 1 Data:
Sample % Abs (% Abs Ref)
Sample #1 = 0.0750 (-0.0120)
Sample #2 = 0.1430 (-0.0490)
Sample #3 = 0.1500 (-0.0400)
Sample #4 = 0.1620 (-0.0280)
Avg % Abs = 0.1517 (-0.0390)
STD DEV = 0.0096 (0.0105)
REL STD DEV = 6.336 (27.014)

Channel 2 Data:
Sample % Abs (% Abs Ref)
Sample #1 = 0.1120 (0.0000)
Sample #2 = 0.1710 (-0.0310)
Sample #3 = 0.1670 (-0.0450)
Sample #4 = 0.1710 (-0.0390)
Avg % Abs = 0.1697 (-0.0383)
STD DEV = 0.0023 (0.0070)
REL STD DEV = 1.361 (18.323)

Sol Value = 0.040 g/210L ***
Fit value = -0.1905 mg/l %%%
Samples Taken = 4, Discarded = 1
3um Io = 12629, 9um Io = 13535
Channel 1 Data:
Sample % Abs (% Abs Ref)
Sample #1 = 0.8900 (-0.0270)
Sample #2 = 0.8620 (-0.0020)
Sample #3 = 0.8620 (0.0340)
Sample #4 = 0.8710 (0.0200)
Avg % Abs = 0.8650 (0.0173)
STD DEV = 0.0052 (0.0181)
REL STD DEV = 0.601 (104.697)

Channel 2 Data:
Sample % Abs (% Abs Ref)
Sample #1 = 1.5370 (-0.0040)
Sample #2 = 1.5070 (0.0070)
Sample #3 = 1.5180 (0.0110)
Sample #4 = 1.5010 (0.0200)
Avg % Abs = 1.5087 (0.0127)
STD DEV = 0.0086 (0.0067)
REL STD DEV = 0.571 (52.566)

Sol Value = 0.100 g/210L ***
Fit value = 0.4762 mg/l %%%
Samples Taken = 4, Discarded = 1
3um Io = 12620, 9um Io = 13528
Channel 1 Data:
Sample % Abs (% Abs Ref)
Sample #1 = 1.9700 (-0.0130)
Sample #2 = 1.9630 (-0.0010)
Sample #3 = 1.9590 (0.0260)
Sample #4 = 1.9530 (0.0430)
Avg % Abs = 1.9583 (0.0227)
STD DEV = 0.0050 (0.0222)
REL STD DEV = 0.257 (97.891)

Channel 2 Data:
Sample % Abs (% Abs Ref)
Sample #1 = 3.5080 (-0.0190)
Sample #2 = 3.4770 (0.0040)
Sample #3 = 3.4750 (0.0180)
Sample #4 = 3.4780 (0.0050)
Avg % Abs = 3.4767 (0.0090)
STD DEV = 0.0015 (0.0078)
REL STD DEV = 0.044 (86.781)

Sol Value = 0.200 g/210L ***
Fit value = 0.9524 mg/l %%%
Samples Taken = 4, Discarded = 1
3um Io = 12611, 9um Io = 13524
Channel 1 Data:
Sample % Abs (% Abs Ref)
Sample #1 = 3.7000 (-0.0210)
Sample #2 = 3.7060 (-0.0460)
Sample #3 = 3.6730 (-0.0180)
Sample #4 = 3.6340 (0.0390)
Avg % Abs = 3.6710 (-0.0083)
STD DEV = 0.0360 (0.0433)
REL STD DEV = 0.982 (519.800)

Channel 2 Data:
Sample % Abs (% Abs Ref)
Sample #1 = 6.6170 (-0.0200)
Sample #2 = 6.6290 (-0.0190)
Sample #3 = 6.6050 (0.0140)
Sample #4 = 6.5470 (0.0380)
Avg % Abs = 6.5937 (0.0110)
STD DEV = 0.0422 (0.0286)
REL STD DEV = 0.639 (260.165)

Sol Value = 0.300 g/210L ***
Fit value = 1.4286 mg/l %%%
Samples Taken = 4, Discarded = 1
3um Io = 12574, 9um Io = 13508
Channel 1 Data:
Sample % Abs (% Abs Ref)
Sample #1 = 5.3690 (-0.0250)
Sample #2 = 5.3610 (-0.0070)
Sample #3 = 5.3490 (0.0350)
Sample #4 = 5.3630 (0.0520)
Avg % Abs = 5.3577 (0.0267)
STD DEV = 0.0076 (0.0304)
REL STD DEV = 0.141 (113.887)

Channel 2 Data:
Sample % Abs (% Abs Ref)
Sample #1 = 9.5830 (0.0080)
Sample #2 = 9.5800 (0.0290)
Sample #3 = 9.5790 (0.0560)
Sample #4 = 9.5640 (0.0740)
Avg % Abs = 9.5743 (0.0530)
STD DEV = 0.0090 (0.0226)
REL STD DEV = 0.094 (42.735)

AUTO CAL DATA
Channel 1 Data:
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.152
Std Dev = 0.01 Rel Std Dev = 6.34
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 0.865
Std Dev = 0.01 Rel Std Dev = 0.60
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 1.958
Std Dev = 0.01 Rel Std Dev = 0.26
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 3.671
Std Dev = 0.04 Rel Std Dev = 0.98
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 5.358
Std Dev = 0.01 Rel Std Dev = 0.14
Zero Order Coef = -386.37
First Order Coef = 2593.50
Second Order Coef = 27.28
Standard Deviation = 25.313330

Channel 2 Data:
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.170
Std Dev = 0.00 Rel Std Dev = 1.36
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 1.509
Std Dev = 0.01 Rel Std Dev = 0.57
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 3.477
Std Dev = 0.00 Rel Std Dev = 0.04
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 6.594
Std Dev = 0.04 Rel Std Dev = 0.64
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 9.574
Std Dev = 0.01 Rel Std Dev = 0.09
Zero Order Coef = -236.01
First Order Coef = 1396.14
Second Order Coef = 12.61
Standard Deviation = 5.956115

Solution Stats Quadratic Fit Chan 2
Act Fit Residual
g/210L g/210L g/210L
0.000 0.000 -0.0000
0.040 0.040 0.0001
0.100 0.100 -0.0002
0.200 0.200 0.0001
0.300 0.300 -0.0000

Sol Value = 0.080 g/210L ***
Fit value = 0.3810 mg/l %%%
Samples Taken = 4, Discarded = 1
Channel 1 Data:
Sample #1 = 2887.00
Sample #2 = 2942.00
Sample #3 = 2904.00
Sample #4 = 2867.00
Average Result = 2904.3333
STD DEV = 37.5011
REL STD DEV = 1.291

Channel 2 Data:
Sample #1 = 3251.00
Sample #2 = 3338.00
Sample #3 = 3317.00
Sample #4 = 3306.00
Average Result = 3320.3333
STD DEV = 16.2583
REL STD DEV = 0.490

Dry Gas H2O Adjust Results
Barometric Pressure = 1011
3 um H2O Adjust (mg/l*10,000) = 905
9 um H2O Adjust (mg/l*10,000) = 489
**** AUTO CAL PASS

Solution Stats Quadratic Fit Chan 1
Act Fit Residual
g/210L g/210L g/210L
0.000 0.000 -0.0002
0.040 0.039 0.0006
0.100 0.101 -0.0007
0.200 0.200 0.0005
0.300 0.300 -0.0001

Post - Optical Adj.

Stability Checks

0.050 g/210L 0.047 to 0.053 g/210L	0.080 g/210L 0.077 to 0.083 g/210L	0.200 g/210L 0.194 to 0.206 g/210L	DGS 0.080 g/210L 0.077 to 0.083 g/210L ≤0.003 g/210L of Wet																																																																																																																																																
NASSAU COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001281 06/16/2026 Software: 8100.27 <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>12:21</td></tr><tr><td>Control Test</td><td>0.050</td><td>12:22</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:22</td></tr><tr><td>Control Test</td><td>0.050</td><td>12:23</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:23</td></tr><tr><td>Control Test</td><td>0.050</td><td>12:24</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:25</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0500</td><td></td></tr><tr><td>Std Dev</td><td>0.0000</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr></tbody></table>  Operator's Signature	Test	g/210L	Time	Air Blank	0.000	12:21	Control Test	0.050	12:22	Air Blank	0.000	12:22	Control Test	0.050	12:23	Air Blank	0.000	12:23	Control Test	0.050	12:24	Air Blank	0.000	12:25	Control Test Stats			Average	0.0500		Std Dev	0.0000		Rel Std Dev(%)	0.0000		NASSAU COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001281 06/16/2026 Software: 8100.27 <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>12:28</td></tr><tr><td>Control Test</td><td>0.079</td><td>12:29</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:29</td></tr><tr><td>Control Test</td><td>0.079</td><td>12:30</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:30</td></tr><tr><td>Control Test</td><td>0.078</td><td>12:31</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:32</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0787</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.7339</td><td></td></tr></tbody></table>  Operator's Signature	Test	g/210L	Time	Air Blank	0.000	12:28	Control Test	0.079	12:29	Air Blank	0.000	12:29	Control Test	0.079	12:30	Air Blank	0.000	12:30	Control Test	0.078	12:31	Air Blank	0.000	12:32	Control Test Stats			Average	0.0787		Std Dev	0.0006		Rel Std Dev(%)	0.7339		NASSAU COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001281 06/16/2026 Software: 8100.27 <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>12:32</td></tr><tr><td>Control Test</td><td>0.198</td><td>12:33</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:34</td></tr><tr><td>Control Test</td><td>0.198</td><td>12:34</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:35</td></tr><tr><td>Control Test</td><td>0.198</td><td>12:36</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:36</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.1980</td><td></td></tr><tr><td>Std Dev</td><td>0.0000</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr></tbody></table>  Operator's Signature	Test	g/210L	Time	Air Blank	0.000	12:32	Control Test	0.198	12:33	Air Blank	0.000	12:34	Control Test	0.198	12:34	Air Blank	0.000	12:35	Control Test	0.198	12:36	Air Blank	0.000	12:36	Control Test Stats			Average	0.1980		Std Dev	0.0000		Rel Std Dev(%)	0.0000		NASSAU COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001281 06/16/2026 Software: 8100.27 <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>12:16</td></tr><tr><td>Control Test</td><td>0.081</td><td>12:17</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:17</td></tr><tr><td>Control Test</td><td>0.081</td><td>12:18</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:18</td></tr><tr><td>Control Test</td><td>0.081</td><td>12:18</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:19</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0810</td><td></td></tr><tr><td>Std Dev</td><td>0.0000</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr></tbody></table>  Operator's Signature	Test	g/210L	Time	Air Blank	0.000	12:16	Control Test	0.081	12:17	Air Blank	0.000	12:17	Control Test	0.081	12:18	Air Blank	0.000	12:18	Control Test	0.081	12:18	Air Blank	0.000	12:19	Control Test Stats			Average	0.0810		Std Dev	0.0000		Rel Std Dev(%)	0.0000	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	12:21																																																																																																																																																	
Control Test	0.050	12:22																																																																																																																																																	
Air Blank	0.000	12:22																																																																																																																																																	
Control Test	0.050	12:23																																																																																																																																																	
Air Blank	0.000	12:23																																																																																																																																																	
Control Test	0.050	12:24																																																																																																																																																	
Air Blank	0.000	12:25																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0500																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	12:28																																																																																																																																																	
Control Test	0.079	12:29																																																																																																																																																	
Air Blank	0.000	12:29																																																																																																																																																	
Control Test	0.079	12:30																																																																																																																																																	
Air Blank	0.000	12:30																																																																																																																																																	
Control Test	0.078	12:31																																																																																																																																																	
Air Blank	0.000	12:32																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0787																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7339																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	12:32																																																																																																																																																	
Control Test	0.198	12:33																																																																																																																																																	
Air Blank	0.000	12:34																																																																																																																																																	
Control Test	0.198	12:34																																																																																																																																																	
Air Blank	0.000	12:35																																																																																																																																																	
Control Test	0.198	12:36																																																																																																																																																	
Air Blank	0.000	12:36																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1980																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	12:16																																																																																																																																																	
Control Test	0.081	12:17																																																																																																																																																	
Air Blank	0.000	12:17																																																																																																																																																	
Control Test	0.081	12:18																																																																																																																																																	
Air Blank	0.000	12:18																																																																																																																																																	
Control Test	0.081	12:18																																																																																																																																																	
Air Blank	0.000	12:19																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0810																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		

Florida Department of Law Enforcement

Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: NASSAU COUNTY SO
Time of Inspection: 10:47

Date of Inspection: 06/25/2026

Serial Number: 80-001281
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202603A Exp: 03/03/2028	0.08g/210L Test (g/210L) Lot#:202603B Exp: 03/03/2028	0.20g/210L Test (g/210L) Lot#:202603D Exp: 03/04/2028	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:23825080A1 Exp: 10/05/2027
0.000	0.049	0.079	0.199	0.079
0.000	0.049	0.079	0.199	0.078
0.000	0.049	0.080	0.199	0.079
0.000	0.049	0.080	0.199	0.078
0.000	0.049	0.080	0.199	0.078
0.000	0.049	0.080	0.200	0.079
0.000	0.050	0.080	0.200	0.078
0.000	0.050	0.081	0.199	0.078
0.000	0.050	0.081	0.199	0.078
0.000	0.050	0.080	0.200	0.078


Standard Deviations	0.0005	0.0006	0.0004	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5


Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.


 LEANDRA HIGGINBOTHAM
 Signature and Printed Name

06/25/2026
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001281, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001281</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>NASSAU COUNTY SO</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>06/25/2026</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>10:47</u>	0.200 g/ 210 L	0.008
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

06/25/2026

Date


LEANDRA HIGGINBOTHAM,
Department Inspector

FDLE/ATP Form 69 January 2026
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Page 1 of 1