

Alcohol Testing Program - Instrument Processing Sheet

Agency: PUTNAM COUNTY SO

Instrument Serial Number: 80-001244

Date In: 6/24/2026

DI Completion Date: 7/9/2026

☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake By: <u>WKP</u> Date: <u>6/24/2026</u>		Quality Checks By: <u>SLH</u> Date: <u>6/30/2025</u>		Flow Adjustment By: _____																															
☒ Annual ☐ Dropped Off ☐ Registration ☐ Return from CMI / EE ☐ Training Instrument Visual Inspection ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/Accessories ☐ Power Cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes:		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value: <u>203</u> ☒ Flow Verification (L/s) Flow Column #: <u>ATP 105</u> 32 mm <u>0.152</u> (.139-.169) 36 mm <u>0.171</u> (.156-.190) 53 mm <u>0.246</u> (.228-.278) 103 mm <u>0.519</u> (.447-.547) ☒ Barometric Pressure Check Gauge ID #: <u>28427</u> Gauge: <u>1011</u> Instrument: <u>1006</u> ☒ Stability Checks		Flow Column #: _____ ☐ 5L/min – 17mm ☐ 15L/min – 53mm ☐ 30L/min – 103mm ☐ R-Value: _____ ☐ Post Adjustment Verification (L/S) Flow Column #: _____ 32 mm (.139-.169) 36 mm (.156-.190) 53 mm (.228-.278) 103 mm (.447-.547)																															
						<table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th style="width: 15%;">Simulator</th> <th style="width: 20%;">Serial #</th> <th style="width: 25%;">Lot#/Exp</th> <th style="width: 40%;">Maintenance By: _____ Date: _____</th> </tr> </thead> <tbody> <tr> <td rowspan="2">0.050</td> <td rowspan="2">MP5088</td> <td>202603A</td> <td>☐ Battery Replacement</td> </tr> <tr> <td>03/03/2028</td> <td>☐ Dry Gas Regulator Replacement</td> </tr> <tr> <td rowspan="2">0.080</td> <td rowspan="2">MP5089</td> <td>202603B</td> <td>☐ Tank Sensor Tare</td> </tr> <tr> <td>03/03/2028</td> <td>☐ Breath Tube Replacement</td> </tr> <tr> <td rowspan="2">0.200</td> <td rowspan="2">MP5090</td> <td>202603D</td> <td>☐ Other:</td> </tr> <tr> <td>03/04/2028</td> <td></td> </tr> <tr> <td rowspan="2">0.080 DGS</td> <td rowspan="2">N/A</td> <td>23825080A1</td> <td></td> </tr> <tr> <td>10/5/2027</td> <td></td> </tr> </tbody> </table>		Simulator	Serial #	Lot#/Exp	Maintenance By: _____ Date: _____	0.050	MP5088	202603A	☐ Battery Replacement	03/03/2028	☐ Dry Gas Regulator Replacement	0.080	MP5089	202603B	☐ Tank Sensor Tare	03/03/2028	☐ Breath Tube Replacement	0.200	MP5090	202603D	☐ Other:	03/04/2028		0.080 DGS	N/A	23825080A1		10/5/2027	
						Simulator	Serial #	Lot#/Exp	Maintenance By: _____ Date: _____																										
						0.050	MP5088	202603A	☐ Battery Replacement																										
								03/03/2028	☐ Dry Gas Regulator Replacement																										
0.080	MP5089	202603B	☐ Tank Sensor Tare																																
		03/03/2028	☐ Breath Tube Replacement																																
0.200	MP5090	202603D	☐ Other:																																
		03/04/2028																																	
0.080 DGS	N/A	23825080A1																																	
		10/5/2027																																	
Optical Bench Adjustment By: <u>SLH</u>		Department Inspection By: <u>SLH</u>																																	
Barometric Pressure Gauge: <u>1014</u> ID#: <u>28662</u>		Barometric Pressure ID#: <u>34419</u>																																	
Simulator	Serial #	Lot #	Expiration	Gauge: <u>1016</u> Instrument: <u>1015</u>																															
0.000	MP6294	N/A	N/A	Mouth Alcohol Solution Lot #: <u>2025-D</u> Exp: <u>09/25/2027</u>																															
0.040	MP6295	25090	03/11/2027	Acetone Stock Solution Lot #: <u>2025-B</u> Exp: <u>09/22/2027</u>																															
0.100	MP6296	24390	10/29/2026	Simulator	Serial Number																														
0.200	MP6297	25020	1/14/2027	0.000	MP5086																														
0.300	MP6298	25150	5/6/2027	Interferent	MP5087																														
0.080 DGS	N/A	AG526603	09/23/2027	0.050	MP6291																														
☒ Post Optical Bench Adjustment Stability Checks				0.080	MP5089																														
Simulator	Serial #	Lot #	Expiration	0.200	MP5090																														
0.050	MP5088	202603A	03/03/2028	Attachments ☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☒ Optical Bench Adjustment ☒ Post-Stability Checks ☐ Flow Adjustment ☒ Form 40 ☐ Other:																															
0.080	MP5089	202603B	03/03/2028																																
0.200	MP5090	202603D	03/04/2028																																
0.080 DGS	N/A	23825080A1	10/5/2027																																
Gauge ID #: <u>34419</u> Gauge: <u>1014</u> Instrument: <u>1014</u>																																			
Notes/Suggested Service: Stabilities performed 7/1/2026. OBA=optical bench adjustment. SLH 7/1/2026 Post stabilities performed 2 separate days. SLH 7/7/2026				☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use																															
				Taylor Gutschow Digitally signed by Taylor Gutschow Date: 2026.07.20 10:40:23 -04'00'																															
				Kaitlyn Spearin Digitally signed by Kaitlyn Spearin Date: 2026.07.21 10:07:24 -04'00'																															
				Tech Review	Admin Review																														

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: PUTNAM COUNTY SO
Time of Inspection: 09:25

Date of Inspection: 07/01/2026

Serial Number: 80-001244
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:


Number of Simulators Used: _____

Remarks:

BYPASS AI FOR OPERATION, COMPLIANCE UNDETERMINED

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



LEANDRA HIGGINBOTHAM

Signature and Printed Name

07/01/2026
Date

Stability Checks

80-00 1244

0.050 g/210L 0.047 to 0.053 g/210L	0.080 g/210L 0.077 to 0.083 g/210L	0.200 g/210L 0.194 to 0.206 g/210L	DGS 0.080 g/210L 0.077 to 0.083 g/210L 50.003 g/210L of Wet																																																																																																																																				
#2 PUTNAM COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 07/01/2026 Software: 8100.27 SN 80-001244 Test g/210L Time <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>09:35</td></tr> <tr><td>Control Test</td><td>0.048</td><td>09:35</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:36</td></tr> <tr><td>Control Test</td><td>0.049</td><td>09:36</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:37</td></tr> <tr><td>Control Test</td><td>0.048</td><td>09:38</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:38</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0483</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.1945</td><td></td></tr> </table> Magee Operator's Signature	Air Blank	0.000	09:35	Control Test	0.048	09:35	Air Blank	0.000	09:36	Control Test	0.049	09:36	Air Blank	0.000	09:37	Control Test	0.048	09:38	Air Blank	0.000	09:38	Control Test Stats			Average	0.0483		Std Dev	0.0006		Rel Std Dev(%)	1.1945		PUTNAM COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 07/01/2026 Software: 8100.27 SN 80-001244 Test g/210L Time <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>09:45</td></tr> <tr><td>Control Test</td><td>0.077</td><td>09:45</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:46</td></tr> <tr><td>Control Test</td><td>0.078</td><td>09:46</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:47</td></tr> <tr><td>Control Test</td><td>0.076</td><td>09:48</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:48</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0770</td><td></td></tr> <tr><td>Std Dev</td><td>0.0010</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.2987</td><td></td></tr> </table> Magee Operator's Signature	Air Blank	0.000	09:45	Control Test	0.077	09:45	Air Blank	0.000	09:46	Control Test	0.078	09:46	Air Blank	0.000	09:47	Control Test	0.076	09:48	Air Blank	0.000	09:48	Control Test Stats			Average	0.0770		Std Dev	0.0010		Rel Std Dev(%)	1.2987		PUTNAM COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 07/01/2026 Software: 8100.27 SN 80-001244 Test g/210L Time <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>10:13</td></tr> <tr><td>Control Test</td><td>0.199</td><td>10:14</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:14</td></tr> <tr><td>Control Test</td><td>0.196</td><td>10:15</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:16</td></tr> <tr><td>Control Test</td><td>0.195</td><td>10:16</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:17</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1967</td><td></td></tr> <tr><td>Std Dev</td><td>0.0021</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.0585</td><td></td></tr> </table> Magee Operator's Signature	Air Blank	0.000	10:13	Control Test	0.199	10:14	Air Blank	0.000	10:14	Control Test	0.196	10:15	Air Blank	0.000	10:16	Control Test	0.195	10:16	Air Blank	0.000	10:17	Control Test Stats			Average	0.1967		Std Dev	0.0021		Rel Std Dev(%)	1.0585		DGS PUTNAM COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 07/01/2026 Software: 8100.27 SN 80-001244 Test g/210L Time <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>09:57</td></tr> <tr><td>Control Test</td><td>0.081</td><td>09:57</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:58</td></tr> <tr><td>Control Test</td><td>0.081</td><td>09:58</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:59</td></tr> <tr><td>Control Test</td><td>0.082</td><td>09:59</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:59</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0813</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7099</td><td></td></tr> </table> Magee Operator's Signature	Air Blank	0.000	09:57	Control Test	0.081	09:57	Air Blank	0.000	09:58	Control Test	0.081	09:58	Air Blank	0.000	09:59	Control Test	0.082	09:59	Air Blank	0.000	09:59	Control Test Stats			Average	0.0813		Std Dev	0.0006		Rel Std Dev(%)	0.7099	
Air Blank	0.000	09:35																																																																																																																																					
Control Test	0.048	09:35																																																																																																																																					
Air Blank	0.000	09:36																																																																																																																																					
Control Test	0.049	09:36																																																																																																																																					
Air Blank	0.000	09:37																																																																																																																																					
Control Test	0.048	09:38																																																																																																																																					
Air Blank	0.000	09:38																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.0483																																																																																																																																						
Std Dev	0.0006																																																																																																																																						
Rel Std Dev(%)	1.1945																																																																																																																																						
Air Blank	0.000	09:45																																																																																																																																					
Control Test	0.077	09:45																																																																																																																																					
Air Blank	0.000	09:46																																																																																																																																					
Control Test	0.078	09:46																																																																																																																																					
Air Blank	0.000	09:47																																																																																																																																					
Control Test	0.076	09:48																																																																																																																																					
Air Blank	0.000	09:48																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.0770																																																																																																																																						
Std Dev	0.0010																																																																																																																																						
Rel Std Dev(%)	1.2987																																																																																																																																						
Air Blank	0.000	10:13																																																																																																																																					
Control Test	0.199	10:14																																																																																																																																					
Air Blank	0.000	10:14																																																																																																																																					
Control Test	0.196	10:15																																																																																																																																					
Air Blank	0.000	10:16																																																																																																																																					
Control Test	0.195	10:16																																																																																																																																					
Air Blank	0.000	10:17																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.1967																																																																																																																																						
Std Dev	0.0021																																																																																																																																						
Rel Std Dev(%)	1.0585																																																																																																																																						
Air Blank	0.000	09:57																																																																																																																																					
Control Test	0.081	09:57																																																																																																																																					
Air Blank	0.000	09:58																																																																																																																																					
Control Test	0.081	09:58																																																																																																																																					
Air Blank	0.000	09:59																																																																																																																																					
Control Test	0.082	09:59																																																																																																																																					
Air Blank	0.000	09:59																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.0813																																																																																																																																						
Std Dev	0.0006																																																																																																																																						
Rel Std Dev(%)	0.7099																																																																																																																																						


Verified instrument was able to read gas pressure.

needs OBA sub

PUTNAM COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001244
07/01/2026
Software: 8100.27

#1

Test	g/210L	Time
Air Blank	0.000	09:29
Control Test	0.049	09:30
Air Blank	0.000	09:30
Control Test	0.048	09:31
Air Blank	0.000	09:31
Control Test	0.048	09:32
Air Blank	0.000	09:33
Control Test Stats		
Average	0.0483	
Std Dev	0.0006	
Rel Std Dev(%)	1.1945	


Operator's Signature

Did not know external printer was connected to instrument and printed the 0.05 g/210L stability. I reran believing it was an internal printer issue (see #2).

SLH

PUTNAM COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001244
07/07/2026 13:44:07

SW

Auto Calibration

pg 1 of 2

<<<<< 3um >>>>>			<<<<< 9um >>>>>		

Solution = 0.000 g/210L or 0.0000 mg/l, Samples = 4, Discarded = 1					
Sample	% Abs	(% Abs Ref)	% Abs	(% Abs Ref)	
Sample #1	0.1670	(-0.0290)	0.1350	(-0.0140)	
Sample #2	0.1240	(0.0170)	0.1380	(0.0050)	
Sample #3	0.0700	(0.0490)	0.0740	(0.0280)	
Sample #4	0.0880	(0.0230)	0.1090	(0.0000)	
Avg % Abs	0.0940	(0.0297)	0.1070	(0.0110)	
STD DEV	0.0275	(0.0170)	0.0320	(0.0149)	
REL STD DEV	29.250	(57.336)	29.950	(135.756)	

Solution = 0.040 g/210L or 0.1905 mg/l, Samples = 4, Discarded = 1					
Sample	% Abs	(% Abs Ref)	% Abs	(% Abs Ref)	
Sample #1	0.8800	(-0.0340)	1.5860	(-0.0270)	
Sample #2	0.8580	(-0.0350)	1.5640	(-0.0310)	
Sample #3	0.8910	(-0.0270)	1.5730	(-0.0210)	
Sample #4	0.8370	(0.0220)	1.5390	(-0.0200)	
Avg % Abs	0.8620	(-0.0133)	1.5587	(-0.0240)	
STD DEV	0.0272	(0.0309)	0.0176	(0.0061)	
REL STD DEV	3.158	(231.449)	1.130	(25.345)	

Solution = 0.100 g/210L or 0.4762 mg/l, Samples = 4, Discarded = 1					
Sample	% Abs	(% Abs Ref)	% Abs	(% Abs Ref)	
Sample #1	2.0170	(-0.0180)	3.7250	(-0.0210)	
Sample #2	1.9550	(0.0380)	3.6730	(0.0530)	
Sample #3	1.9760	(0.0330)	3.6440	(0.0550)	
Sample #4	1.9970	(0.0430)	3.6510	(0.0600)	
Avg % Abs	1.9760	(0.0380)	3.6560	(0.0560)	
STD DEV	0.0210	(0.0050)	0.0151	(0.0036)	
REL STD DEV	1.063	(13.158)	0.414	(6.438)	

Solution = 0.200 g/210L or 0.9524 mg/l, Samples = 4, Discarded = 1					
Sample	% Abs	(% Abs Ref)	% Abs	(% Abs Ref)	
Sample #1	3.8880	(-0.0290)	7.0960	(-0.0260)	
Sample #2	3.8510	(0.0310)	7.0320	(0.0570)	
Sample #3	3.8310	(0.0480)	7.0060	(0.0640)	
Sample #4	3.8370	(0.0550)	6.9950	(0.0700)	
Avg % Abs	3.8397	(0.0447)	7.0110	(0.0637)	
STD DEV	0.0103	(0.0123)	0.0190	(0.0065)	
REL STD DEV	0.267	(27.632)	0.271	(10.219)	

Solution = 0.300 g/210L or 1.4286 mg/l, Samples = 4, Discarded = 1					
Sample	% Abs	(% Abs Ref)	% Abs	(% Abs Ref)	
Sample #1	5.6580	(-0.0070)	10.2780	(-0.0180)	
Sample #2	5.6020	(0.0390)	10.2150	(0.0700)	
Sample #3	5.5770	(0.0820)	10.1550	(0.1140)	
Sample #4	5.5920	(0.0590)	10.1960	(0.0820)	
Avg % Abs	5.5903	(0.0600)	10.1887	(0.0887)	
STD DEV	0.0126	(0.0215)	0.0307	(0.0227)	
REL STD DEV	0.225	(35.862)	0.301	(25.652)	

PUTNAM COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001244
07/07/2026 13:44:07

Auto Calibration

pg 2 of 2

<<<< 3um >>>>

Zero Order Coef -221.09
First Order Coef 2450.96
Second Order Coef 25.32

Act (g/210L)	Fit (g/210L)	Residual (g/210L)
0.000	0.000	-0.0002
0.040	0.040	-0.0001
0.100	0.099	0.0009
0.200	0.201	-0.0008
0.300	0.300	0.0003

<<<< 9um >>>>

Zero Order Coef -140.56
First Order Coef 1297.00
Second Order Coef 11.66

Act (g/210L)	Fit (g/210L)	Residual (g/210L)
0.000	-0.000	0.0000
0.040	0.040	-0.0001
0.100	0.100	0.0001
0.200	0.200	-0.0000
0.300	0.300	0.0000

<<<< 3um >>>>

Solution = 0.080 g/210L or 0.3810 mg/l, Samples = 4, Discarded = 1
Sample

Sample #1	3131.00	3450.00
Sample #2	3156.00	3430.00
Sample #3	3247.00	3483.00
Sample #4	3152.00	3450.00
Avg	3185.0000	3454.3333
STD DEV	53.7308	26.7644
REL STD DEV	1.687	0.775
H2O adjust (mg/l*10k)	624	355

Barometric Pressure = 1013

*****CALIBRATION SUCCESSFUL*****

SW

86-001244
Sub

Post Optical Bench Adjustment Stability Checks

0.050 g/210L	0.080 g/210L	0.200 g/210L	DGS 0.080 g/210L																																																																																																																																				
0.047 to 0.053 g/210L ✓	0.077 to 0.083 g/210L ✓	0.194 to 0.206 g/210L ✓	0.077 to 0.083 g/210L 50.003 g/210L of Wet ✓																																																																																																																																				
PUTNAM COUNTY SO Intoxilizer - Alcohol Analyzer Model 8000 SN 80-001244 07/09/2026 Software: 8100.27 Test g/210L Time <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>09:27</td></tr> <tr><td>Control Test</td><td>0.050</td><td>09:28</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:29</td></tr> <tr><td>Control Test</td><td>0.051</td><td>09:29</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:30</td></tr> <tr><td>Control Test</td><td>0.050</td><td>09:31</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:31</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0503</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.1471</td><td></td></tr> </table>	Air Blank	0.000	09:27	Control Test	0.050	09:28	Air Blank	0.000	09:29	Control Test	0.051	09:29	Air Blank	0.000	09:30	Control Test	0.050	09:31	Air Blank	0.000	09:31	Control Test Stats			Average	0.0503		Std Dev	0.0006		Rel Std Dev(%)	1.1471		PUTNAM COUNTY SO Intoxilizer - Alcohol Analyzer Model 8000 SN 80-001244 07/07/2026 Software: 8100.27 Test g/210L Time <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>14:48</td></tr> <tr><td>Control Test</td><td>0.078</td><td>14:49</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:49</td></tr> <tr><td>Control Test</td><td>0.079</td><td>14:50</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:50</td></tr> <tr><td>Control Test</td><td>0.079</td><td>14:51</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:52</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0787</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7335</td><td></td></tr> </table>	Air Blank	0.000	14:48	Control Test	0.078	14:49	Air Blank	0.000	14:49	Control Test	0.079	14:50	Air Blank	0.000	14:50	Control Test	0.079	14:51	Air Blank	0.000	14:52	Control Test Stats			Average	0.0787		Std Dev	0.0006		Rel Std Dev(%)	0.7335		PUTNAM COUNTY SO Intoxilizer - Alcohol Analyzer Model 8000 SN 80-001244 17/09/2026 Software: 8100.27 Test g/210L Time <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>09:18</td></tr> <tr><td>Control Test</td><td>0.200</td><td>09:18</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:19</td></tr> <tr><td>Control Test</td><td>0.199</td><td>09:20</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:20</td></tr> <tr><td>Control Test</td><td>0.199</td><td>09:21</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:21</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1993</td><td></td></tr> <tr><td>Std Dev</td><td>0.0005</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.2895</td><td></td></tr> </table>	Air Blank	0.000	09:18	Control Test	0.200	09:18	Air Blank	0.000	09:19	Control Test	0.199	09:20	Air Blank	0.000	09:20	Control Test	0.199	09:21	Air Blank	0.000	09:21	Control Test Stats			Average	0.1993		Std Dev	0.0005		Rel Std Dev(%)	0.2895		PUTNAM COUNTY SO Intoxilizer - Alcohol Analyzer Model 8000 SN 80-001244 07/07/2026 Software: 8100.27 Test g/210L Time <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>14:54</td></tr> <tr><td>Control Test</td><td>0.080</td><td>14:54</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:55</td></tr> <tr><td>Control Test</td><td>0.079</td><td>14:55</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:56</td></tr> <tr><td>Control Test</td><td>0.081</td><td>14:56</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:56</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0800</td><td></td></tr> <tr><td>Std Dev</td><td>0.0010</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.2500</td><td></td></tr> </table>	Air Blank	0.000	14:54	Control Test	0.080	14:54	Air Blank	0.000	14:55	Control Test	0.079	14:55	Air Blank	0.000	14:56	Control Test	0.081	14:56	Air Blank	0.000	14:56	Control Test Stats			Average	0.0800		Std Dev	0.0010		Rel Std Dev(%)	1.2500	
Air Blank	0.000	09:27																																																																																																																																					
Control Test	0.050	09:28																																																																																																																																					
Air Blank	0.000	09:29																																																																																																																																					
Control Test	0.051	09:29																																																																																																																																					
Air Blank	0.000	09:30																																																																																																																																					
Control Test	0.050	09:31																																																																																																																																					
Air Blank	0.000	09:31																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.0503																																																																																																																																						
Std Dev	0.0006																																																																																																																																						
Rel Std Dev(%)	1.1471																																																																																																																																						
Air Blank	0.000	14:48																																																																																																																																					
Control Test	0.078	14:49																																																																																																																																					
Air Blank	0.000	14:49																																																																																																																																					
Control Test	0.079	14:50																																																																																																																																					
Air Blank	0.000	14:50																																																																																																																																					
Control Test	0.079	14:51																																																																																																																																					
Air Blank	0.000	14:52																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.0787																																																																																																																																						
Std Dev	0.0006																																																																																																																																						
Rel Std Dev(%)	0.7335																																																																																																																																						
Air Blank	0.000	09:18																																																																																																																																					
Control Test	0.200	09:18																																																																																																																																					
Air Blank	0.000	09:19																																																																																																																																					
Control Test	0.199	09:20																																																																																																																																					
Air Blank	0.000	09:20																																																																																																																																					
Control Test	0.199	09:21																																																																																																																																					
Air Blank	0.000	09:21																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.1993																																																																																																																																						
Std Dev	0.0005																																																																																																																																						
Rel Std Dev(%)	0.2895																																																																																																																																						
Air Blank	0.000	14:54																																																																																																																																					
Control Test	0.080	14:54																																																																																																																																					
Air Blank	0.000	14:55																																																																																																																																					
Control Test	0.079	14:55																																																																																																																																					
Air Blank	0.000	14:56																																																																																																																																					
Control Test	0.081	14:56																																																																																																																																					
Air Blank	0.000	14:56																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.0800																																																																																																																																						
Std Dev	0.0010																																																																																																																																						
Rel Std Dev(%)	1.2500																																																																																																																																						
Miguel Operator's Signature	Miguel Operator's Signature	Miguel Operator's Signature	Miguel Operator's Signature																																																																																																																																				

Florida Department of Law Enforcement

Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: PUTNAM COUNTY SO
Time of Inspection: 12:40

Date of Inspection: 07/09/2026

Serial Number: 80-001244
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202603A Exp: 03/03/2028	0.08g/210L Test (g/210L) Lot#:202603B Exp: 03/03/2028	0.20g/210L Test (g/210L) Lot#:202603D Exp: 03/04/2028	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:23825080A1 Exp: 10/05/2027
0.000	0.051	0.079	0.199	0.079
0.000	0.050	0.079	0.199	0.080
0.000	0.050	0.079	0.199	0.079
0.000	0.050	0.079	0.199	0.079
0.000	0.051	0.079	0.199	0.079
0.000	0.050	0.080	0.200	0.079
0.000	0.051	0.079	0.199	0.079
0.000	0.050	0.080	0.199	0.079
0.000	0.051	0.080	0.201	0.079
0.000	0.051	0.080	0.200	0.078


Standard Deviations	0.0005	0.0005	0.0006	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0005 Number of Simulators Used: 5


Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.


LEANDRA HIGGINBOTHAM
Signature and Printed Name

07/09/2026
Date



ANAB

ANSI National Accreditation Board
ACCREDITED
ISO/IEC 17025
FORENSIC CALIBRATION
LABORATORY

Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001244, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001244</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>PUTNAM COUNTY SO</u>		0.050 g/ 210 L 0.004
Calibration Date:	<u>07/09/2026</u>		0.080 g/ 210 L 0.004
Calibration Time:	<u>12:40</u>		0.200 g/ 210 L 0.008
			0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

FDLE/ATP Form 69 January 2026
Issuing Authority: Alcohol Testing Program

07/09/2026 Date
LEANDRA HIGGINBOTHAM,
Department Inspector

Service • Integrity • Respect • Quality