

Alcohol Testing Program - Instrument Processing Sheet

Agency: Coconut Creek Police Department

Instrument Serial Number: 80-001044

Date In: 4/30/2026

DI Completion Date: 6/30/2026

☒ Ship

☐ P/U

☐ H/D

☐ CMI

☐ EE

Intake By: <u>TDG</u> Date: <u>5/18/2026</u>		Quality Checks By: <u>TDG</u> Date: <u>6/8/2026</u>		Flow Adjustment By: _____																																							
☒ Annual ☐ Dropped Off ☐ Registration ☐ Return from CMI / EE ☐ Training Instrument Visual Inspection ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/Accessories ☐ Power Cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes:		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value: <u>259</u> ☒ Flow Verification (L/s) Flow Column #: <u>ATP106</u> 32 mm 0.156 (.139-.169) 36 mm 0.175 (.156-.190) 53 mm 0.234 (.228-.278) 103 mm 0.496 (.447-.547) ☒ Barometric Pressure Check Gauge ID #: <u>33363</u> Gauge: <u>1018</u> Instrument: <u>1017</u> ☒ Stability Checks		Flow Column #: _____ ☐ 5L/min – 17mm ☐ 15L/min – 53mm ☐ 30L/min – 103mm ☐ R-Value: _____ ☐ Post Adjustment Verification (L/S) Flow Column #: _____ 32 mm (.139-.169) 36 mm (.156-.190) 53 mm (.228-.278) 103 mm (.447-.547)																																							
						<table border="1" style="width: 100%; border-collapse: collapse;"> <tr> <th style="width: 15%;">Simulator</th> <th style="width: 20%;">Serial #</th> <th style="width: 25%;">Lot#/Exp</th> <th style="width: 40%;">Maintenance By: _____ Date: _____</th> </tr> <tr> <td>0.050</td> <td>MP6286</td> <td>202406K 6/19/2026</td> <td rowspan="5"> ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Tank Sensor Tare ☐ Breath Tube Replacement ☐ Other: </td> </tr> <tr> <td>0.080</td> <td>MP6287</td> <td>202603B 3/3/2028</td> </tr> <tr> <td>0.200</td> <td>MP6288</td> <td>202406N 6/20/2026</td> </tr> <tr> <td rowspan="2">0.080 DGS</td> <td rowspan="2">N/A</td> <td>AG526603</td> </tr> <tr> <td>9/23/2027</td> </tr> </table>		Simulator	Serial #	Lot#/Exp	Maintenance By: _____ Date: _____	0.050	MP6286	202406K 6/19/2026	☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Tank Sensor Tare ☐ Breath Tube Replacement ☐ Other:	0.080	MP6287	202603B 3/3/2028	0.200	MP6288	202406N 6/20/2026	0.080 DGS	N/A	AG526603	9/23/2027																		
						Simulator	Serial #	Lot#/Exp	Maintenance By: _____ Date: _____																																		
						0.050	MP6286	202406K 6/19/2026	☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Tank Sensor Tare ☐ Breath Tube Replacement ☐ Other:																																		
						0.080	MP6287	202603B 3/3/2028																																			
0.200	MP6288	202406N 6/20/2026																																									
0.080 DGS	N/A	AG526603																																									
		9/23/2027																																									
<table border="1" style="width: 100%; border-collapse: collapse;"> <tr> <th style="width: 25%;">Simulator</th> <th style="width: 20%;">Serial #</th> <th style="width: 20%;">Lot #</th> <th style="width: 35%;">Expiration</th> </tr> <tr> <td>0.000</td> <td>MP5097</td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td>MP5098</td> <td>25090</td> <td>3/11/2027</td> </tr> <tr> <td>0.100</td> <td>MP5099</td> <td>24390</td> <td>10/29/2026</td> </tr> <tr> <td>0.200</td> <td>MP5100</td> <td>25020</td> <td>1/14/2027</td> </tr> <tr> <td>0.300</td> <td>MP5101</td> <td>24430</td> <td>12/10/2026</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>28424080A3</td> <td>11/5/2026</td> </tr> </table>		Simulator	Serial #	Lot #	Expiration	0.000	MP5097	N/A	N/A	0.040	MP5098	25090	3/11/2027	0.100	MP5099	24390	10/29/2026	0.200	MP5100	25020	1/14/2027	0.300	MP5101	24430	12/10/2026	0.080 DGS	N/A	28424080A3	11/5/2026	<table border="1" style="width: 100%; border-collapse: collapse;"> <tr> <th style="width: 50%;">Gauge: <u>1017</u> Instrument: <u>1018</u></th> <th style="width: 50%;">Serial Number</th> </tr> <tr> <td>Mouth Alcohol Solution Lot #: <u>2025-D</u> Exp: <u>9/25/2027</u></td> <td>MP6284</td> </tr> <tr> <td>Acetone Stock Solution Lot #: <u>2025-B</u> Exp: <u>9/22/2027</u></td> <td>MP6285</td> </tr> <tr> <td>0.000</td> <td>MP6286</td> </tr> <tr> <td>0.050</td> <td>MP6287</td> </tr> <tr> <td>0.080</td> <td>MP6288</td> </tr> </table>		Gauge: <u>1017</u> Instrument: <u>1018</u>	Serial Number	Mouth Alcohol Solution Lot #: <u>2025-D</u> Exp: <u>9/25/2027</u>	MP6284	Acetone Stock Solution Lot #: <u>2025-B</u> Exp: <u>9/22/2027</u>	MP6285	0.000	MP6286	0.050	MP6287	0.080	MP6288
Simulator	Serial #	Lot #	Expiration																																								
0.000	MP5097	N/A	N/A																																								
0.040	MP5098	25090	3/11/2027																																								
0.100	MP5099	24390	10/29/2026																																								
0.200	MP5100	25020	1/14/2027																																								
0.300	MP5101	24430	12/10/2026																																								
0.080 DGS	N/A	28424080A3	11/5/2026																																								
Gauge: <u>1017</u> Instrument: <u>1018</u>	Serial Number																																										
Mouth Alcohol Solution Lot #: <u>2025-D</u> Exp: <u>9/25/2027</u>	MP6284																																										
Acetone Stock Solution Lot #: <u>2025-B</u> Exp: <u>9/22/2027</u>	MP6285																																										
0.000	MP6286																																										
0.050	MP6287																																										
0.080	MP6288																																										
☒ Post Optical Bench Adjustment Stability Checks		☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☒ Optical Bench Adjustment		☒ Post-Stability Checks ☐ Flow Adjustment ☒ Form 40 ☐ Other:																																							
Gauge ID #: <u>33363</u> Gauge: <u>1017</u> Instrument: <u>1018</u>		☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use		LeAndra Higginbotham Digitally signed by LeAndra Higginbotham Date: 2026.07.09 17:09:08 -04'00'																																							
Notes/Suggested Service:		Kaitlyn Spearin Digitally signed by Kaitlyn Spearin Date: 2026.07.14 07:41:08 -04'00'		Tech Review Admin Review																																							

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: COCONUT CREEK PD
Time of Inspection: 11:54

Date of Inspection: 06/08/2026

Serial Number: 80-001044
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:

Number of Simulators Used: _____

Remarks:

AI NOT CONDUCTED. COMPLIANCE NOT DETERMINED.

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

06/08/2026
Date

Stability Checks

0.05g/210L 0.047 to 0.053	0.08g/210L 0.077 to 0.083	0.20g/210L 0.194 to 0.206	DGS 0.08g/210L 0.077 to 0.083 ≤0.003 of Wet																																																																																																																																																
COCONUT CREEK PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001044 06/08/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>12:18</td></tr> <tr><td>Control Test</td><td>0.050</td><td>12:19</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:19</td></tr> <tr><td>Control Test</td><td>0.050</td><td>12:20</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:21</td></tr> <tr><td>Control Test</td><td>0.050</td><td>12:21</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:22</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0500</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> Operator's Signature 	Test	g/210L	Time	Air Blank	0.000	12:18	Control Test	0.050	12:19	Air Blank	0.000	12:19	Control Test	0.050	12:20	Air Blank	0.000	12:21	Control Test	0.050	12:21	Air Blank	0.000	12:22	Control Test Stats			Average	0.0500		Std Dev	0.0000		Rel Std Dev(%)	0.0000		COCONUT CREEK PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001044 06/08/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>12:26</td></tr> <tr><td>Control Test</td><td>0.079</td><td>12:27</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:27</td></tr> <tr><td>Control Test</td><td>0.079</td><td>12:28</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:28</td></tr> <tr><td>Control Test</td><td>0.079</td><td>12:29</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:30</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0790</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> Operator's Signature 	Test	g/210L	Time	Air Blank	0.000	12:26	Control Test	0.079	12:27	Air Blank	0.000	12:27	Control Test	0.079	12:28	Air Blank	0.000	12:28	Control Test	0.079	12:29	Air Blank	0.000	12:30	Control Test Stats			Average	0.0790		Std Dev	0.0000		Rel Std Dev(%)	0.0000		COCONUT CREEK PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001044 06/08/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>12:59</td></tr> <tr><td>Control Test</td><td>0.210</td><td>13:00</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:00</td></tr> <tr><td>Control Test</td><td>0.199</td><td>13:01</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:02</td></tr> <tr><td>Control Test</td><td>0.199</td><td>13:02</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:03</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.2027</td><td></td></tr> <tr><td>Std Dev</td><td>0.0064</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>3.1336</td><td></td></tr> </tbody> </table> Operator's Signature 	Test	g/210L	Time	Air Blank	0.000	12:59	Control Test	0.210	13:00	Air Blank	0.000	13:00	Control Test	0.199	13:01	Air Blank	0.000	13:02	Control Test	0.199	13:02	Air Blank	0.000	13:03	Control Test Stats			Average	0.2027		Std Dev	0.0064		Rel Std Dev(%)	3.1336		COCONUT CREEK PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001044 06/08/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>13:04</td></tr> <tr><td>Control Test</td><td>0.078</td><td>13:05</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:05</td></tr> <tr><td>Control Test</td><td>0.079</td><td>13:05</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:06</td></tr> <tr><td>Control Test</td><td>0.079</td><td>13:06</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:07</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0787</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7339</td><td></td></tr> </tbody> </table> Operator's Signature 	Test	g/210L	Time	Air Blank	0.000	13:04	Control Test	0.078	13:05	Air Blank	0.000	13:05	Control Test	0.079	13:05	Air Blank	0.000	13:06	Control Test	0.079	13:06	Air Blank	0.000	13:07	Control Test Stats			Average	0.0787		Std Dev	0.0006		Rel Std Dev(%)	0.7339	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	12:18																																																																																																																																																	
Control Test	0.050	12:19																																																																																																																																																	
Air Blank	0.000	12:19																																																																																																																																																	
Control Test	0.050	12:20																																																																																																																																																	
Air Blank	0.000	12:21																																																																																																																																																	
Control Test	0.050	12:21																																																																																																																																																	
Air Blank	0.000	12:22																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0500																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	12:26																																																																																																																																																	
Control Test	0.079	12:27																																																																																																																																																	
Air Blank	0.000	12:27																																																																																																																																																	
Control Test	0.079	12:28																																																																																																																																																	
Air Blank	0.000	12:28																																																																																																																																																	
Control Test	0.079	12:29																																																																																																																																																	
Air Blank	0.000	12:30																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0790																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	12:59																																																																																																																																																	
Control Test	0.210	13:00																																																																																																																																																	
Air Blank	0.000	13:00																																																																																																																																																	
Control Test	0.199	13:01																																																																																																																																																	
Air Blank	0.000	13:02																																																																																																																																																	
Control Test	0.199	13:02																																																																																																																																																	
Air Blank	0.000	13:03																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.2027																																																																																																																																																		
Std Dev	0.0064																																																																																																																																																		
Rel Std Dev(%)	3.1336																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	13:04																																																																																																																																																	
Control Test	0.078	13:05																																																																																																																																																	
Air Blank	0.000	13:05																																																																																																																																																	
Control Test	0.079	13:05																																																																																																																																																	
Air Blank	0.000	13:06																																																																																																																																																	
Control Test	0.079	13:06																																																																																																																																																	
Air Blank	0.000	13:07																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0787																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7339																																																																																																																																																		

COCONUT CREEK PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001044
06/29/2026 13:38:47

Auto Calibration
Max Power Res Value = 85
Auto Range Res Value = 49

Sol Value = 0.000 g/210L ***
Fit value = 0.0000 mg/l %%%
Samples Taken = 4, Discarded = 1
3um Io = 12355, 9um Io = 13761

***** CHANNEL 1 *****
Sample % Abs (% Abs Ref)
Sample #1 = 0.0890 (-0.0280)
Sample #2 = 0.1160 (0.0020)
Sample #3 = 0.0940 (0.0110)
Sample #4 = 0.1160 (0.0390)
Avg % Abs = 0.1087 (0.0173)
STD DEV = 0.0127 (0.0193)
REL STD DEV = 11.669 (111.323)

***** CHANNEL 2 *****
Sample % Abs (% Abs Ref)
Sample #1 = 0.1260 (-0.0160)
Sample #2 = 0.1060 (0.0100)
Sample #3 = 0.1140 (0.0130)
Sample #4 = 0.1170 (0.0210)
Avg % Abs = 0.1123 (0.0147)
STD DEV = 0.0057 (0.0057)
REL STD DEV = 5.062 (38.770)

Sol Value = 0.040 g/210L ***
Fit value = 0.1905 mg/l %%%
Samples Taken = 4, Discarded = 1
3um Io = 12344, 9um Io = 13755

***** CHANNEL 1 *****
Sample % Abs (% Abs Ref)
Sample #1 = 0.8260 (-0.0140)
Sample #2 = 0.8510 (0.0040)
Sample #3 = 0.8210 (0.0150)
Sample #4 = 0.8760 (0.0230)
Avg % Abs = 0.8493 (0.0140)
STD DEV = 0.0275 (0.0095)
REL STD DEV = 3.242 (68.139)

***** CHANNEL 2 *****
Sample % Abs (% Abs Ref)
Sample #1 = 1.5330 (0.0000)
Sample #2 = 1.5220 (0.0260)
Sample #3 = 1.5340 (0.0210)
Sample #4 = 1.5210 (0.0320)
Avg % Abs = 1.5257 (0.0263)
STD DEV = 0.0072 (0.0055)
REL STD DEV = 0.474 (20.915)

Sol Value = 0.100 g/210L ***
Fit value = 0.4762 mg/l %%%
Samples Taken = 4, Discarded = 1
3um Io = 12328, 9um Io = 13742

***** CHANNEL 1 *****
Sample % Abs (% Abs Ref)
Sample #1 = 1.9620 (-0.0160)
Sample #2 = 1.9300 (0.0050)
Sample #3 = 1.9540 (0.0170)
Sample #4 = 1.9630 (0.0050)
Avg % Abs = 1.9490 (0.0090)
STD DEV = 0.0171 (0.0069)
REL STD DEV = 0.675 (76.980)

***** CHANNEL 2 *****
Sample % Abs (% Abs Ref)
Sample #1 = 3.6530 (-0.0080)
Sample #2 = 3.6260 (-0.0030)
Sample #3 = 3.6240 (0.0070)
Sample #4 = 3.6280 (0.0070)
Avg % Abs = 3.6260 (0.0037)
STD DEV = 0.0020 (0.0058)
REL STD DEV = 0.055 (157.459)

Sol Value = 0.200 g/210L ***
Fit value = 0.9524 mg/l %%%
Samples Taken = 4, Discarded = 1
3um Io = 12323, 9um Io = 13736

***** CHANNEL 1 *****
Sample % Abs (% Abs Ref)
Sample #1 = 3.7630 (-0.0230)
Sample #2 = 3.7380 (0.0430)
Sample #3 = 3.7160 (0.0360)
Sample #4 = 3.7120 (0.0710)
Avg % Abs = 3.7220 (0.0507)
STD DEV = 0.0140 (0.0178)
REL STD DEV = 0.376 (35.103)

***** CHANNEL 2 *****
Sample % Abs (% Abs Ref)
Sample #1 = 6.9710 (-0.0100)
Sample #2 = 6.9390 (0.0420)
Sample #3 = 6.9300 (0.0290)
Sample #4 = 6.9250 (0.0510)
Avg % Abs = 6.9313 (0.0407)
STD DEV = 0.0071 (0.0111)
REL STD DEV = 0.102 (27.198)

Sol Value = 0.300 g/210L ***
Fit value = 1.4286 mg/l %%%
Samples Taken = 4, Discarded = 1
3um Io = 12315, 9um Io = 13733

***** CHANNEL 1 *****
Sample % Abs (% Abs Ref)
Sample #1 = 5.5220 (-0.0310)
Sample #2 = 5.4500 (0.0550)
Sample #3 = 5.4610 (0.0380)
Sample #4 = 5.4590 (0.0380)
Avg % Abs = 5.4567 (0.0437)
STD DEV = 0.0059 (0.0098)
REL STD DEV = 0.107 (22.477)

***** CHANNEL 2 *****
Sample % Abs (% Abs Ref)
Sample #1 = 10.1750 (-0.0030)
Sample #2 = 10.0960 (0.0960)
Sample #3 = 10.0900 (0.1130)
Sample #4 = 10.0900 (0.1040)
Avg % Abs = 10.0920 (0.1050)
STD DEV = 0.0035 (0.0075)
REL STD DEV = 0.034 (7.190)

***** AUTO CAL DATA *****
***** CHANNEL 1 *****
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.109
Std Dev = 0.01 Rel Std Dev = 11.69
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 0.849
Std Dev = 0.03 Rel Std Dev = 3.24
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 1.949
Std Dev = 0.02 Rel Std Dev = 0.88
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 3.722
Std Dev = 0.01 Rel Std Dev = 0.38
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 5.457
Std Dev = 0.01 Rel Std Dev = 0.11
Zero Order Coef = -278.63
First Order Coef = 2548.40
Second Order Coef = 22.22
Standard Deviation = 7.485148

***** CHANNEL 2 *****
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.112
Std Dev = 0.01 Rel Std Dev = 5.06
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 1.526
Std Dev = 0.01 Rel Std Dev = 0.47
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 3.626
Std Dev = 0.00 Rel Std Dev = 0.06
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 6.931
Std Dev = 0.01 Rel Std Dev = 0.10
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 10.092
Std Dev = 0.00 Rel Std Dev = 0.03
Zero Order Coef = -142.04
First Order Coef = 1314.68
Second Order Coef = 11.40
Standard Deviation = 10.551074

Solution Stats Quadratic Fit Chan 1		
Act	Fit	Residual
g/210L	g/210L	g/210L
0.000	-0.000	0.0000
0.040	0.040	0.0001
0.100	0.100	-0.0002
0.200	0.200	0.0002
0.300	0.300	-0.0001

Solution Stats Quadratic Fit Chan 2		
Act	Fit	Residual
g/210L	g/210L	g/210L
0.000	0.000	-0.0001
0.040	0.040	0.0003
0.100	0.100	-0.0003
0.200	0.200	0.0001
0.300	0.300	-0.0000

Sol Value = 0.080 g/210L ***
Fit value = 0.3810 mg/l %%%
Samples Taken = 4, Discarded = 1
***** CHANNEL 1 *****
Sample #1 = 2070.00
Sample #2 = 2968.00
Sample #3 = 3084.00
Sample #4 = 2949.00
Average Result = 3000.3333
STD DEV = 73.0776
REL STD DEV = 2.436

***** CHANNEL 2 *****
Sample #1 = 3366.00
Sample #2 = 3432.00
Sample #3 = 3424.00
Sample #4 = 3417.00
Average Result = 3424.3333
STD DEV = 7.5056
REL STD DEV = 0.219

***** DRY GAS H2O ADJUST RESULTS *****
Barometric Pressure = 1018
3 um H2O Adjust (mg/l*10,000) = 809
9 um H2O Adjust (mg/l*10,000) = 385
***** AUTO CAL PASS *****

Optical Calibration Adjustment

By: TDG

Post-Cal Stability Checks

0.05g/210L 0.047 to 0.053	0.08g/210L 0.077 to 0.083	0.20g/210L 0.194 to 0.206	DGS 0.08g/210L 0.077 to 0.083 ≤0.003 of Wet																																																																																																																																																
COCONUT CREEK PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001044 06/29/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>15:33</td></tr> <tr><td>Control Test</td><td>0.050</td><td>15:34</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:35</td></tr> <tr><td>Control Test</td><td>0.049</td><td>15:35</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:36</td></tr> <tr><td>Control Test</td><td>0.049</td><td>15:37</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:37</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0493</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.1703</td><td></td></tr> </tbody> </table> Operator's Signature 	Test	g/210L	Time	Air Blank	0.000	15:33	Control Test	0.050	15:34	Air Blank	0.000	15:35	Control Test	0.049	15:35	Air Blank	0.000	15:36	Control Test	0.049	15:37	Air Blank	0.000	15:37	Control Test Stats			Average	0.0493		Std Dev	0.0006		Rel Std Dev(%)	1.1703		COCONUT CREEK PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001044 06/29/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>15:40</td></tr> <tr><td>Control Test</td><td>0.079</td><td>15:41</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:42</td></tr> <tr><td>Control Test</td><td>0.079</td><td>15:42</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:43</td></tr> <tr><td>Control Test</td><td>0.078</td><td>15:43</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:44</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0787</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7339</td><td></td></tr> </tbody> </table> Operator's Signature 	Test	g/210L	Time	Air Blank	0.000	15:40	Control Test	0.079	15:41	Air Blank	0.000	15:42	Control Test	0.079	15:42	Air Blank	0.000	15:43	Control Test	0.078	15:43	Air Blank	0.000	15:44	Control Test Stats			Average	0.0787		Std Dev	0.0006		Rel Std Dev(%)	0.7339		COCONUT CREEK PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001044 06/29/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>15:47</td></tr> <tr><td>Control Test</td><td>0.200</td><td>15:48</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:48</td></tr> <tr><td>Control Test</td><td>0.198</td><td>15:49</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:50</td></tr> <tr><td>Control Test</td><td>0.198</td><td>15:50</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:51</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1987</td><td></td></tr> <tr><td>Std Dev</td><td>0.0012</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.5812</td><td></td></tr> </tbody> </table> Operator's Signature 	Test	g/210L	Time	Air Blank	0.000	15:47	Control Test	0.200	15:48	Air Blank	0.000	15:48	Control Test	0.198	15:49	Air Blank	0.000	15:50	Control Test	0.198	15:50	Air Blank	0.000	15:51	Control Test Stats			Average	0.1987		Std Dev	0.0012		Rel Std Dev(%)	0.5812		COCONUT CREEK PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001044 06/29/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>15:23</td></tr> <tr><td>Control Test</td><td>0.079</td><td>15:23</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:24</td></tr> <tr><td>Control Test</td><td>0.078</td><td>15:24</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:25</td></tr> <tr><td>Control Test</td><td>0.080</td><td>15:25</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:25</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0790</td><td></td></tr> <tr><td>Std Dev</td><td>0.0010</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.2658</td><td></td></tr> </tbody> </table> Operator's Signature 	Test	g/210L	Time	Air Blank	0.000	15:23	Control Test	0.079	15:23	Air Blank	0.000	15:24	Control Test	0.078	15:24	Air Blank	0.000	15:25	Control Test	0.080	15:25	Air Blank	0.000	15:25	Control Test Stats			Average	0.0790		Std Dev	0.0010		Rel Std Dev(%)	1.2658	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	15:33																																																																																																																																																	
Control Test	0.050	15:34																																																																																																																																																	
Air Blank	0.000	15:35																																																																																																																																																	
Control Test	0.049	15:35																																																																																																																																																	
Air Blank	0.000	15:36																																																																																																																																																	
Control Test	0.049	15:37																																																																																																																																																	
Air Blank	0.000	15:37																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0493																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	1.1703																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	15:40																																																																																																																																																	
Control Test	0.079	15:41																																																																																																																																																	
Air Blank	0.000	15:42																																																																																																																																																	
Control Test	0.079	15:42																																																																																																																																																	
Air Blank	0.000	15:43																																																																																																																																																	
Control Test	0.078	15:43																																																																																																																																																	
Air Blank	0.000	15:44																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0787																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7339																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	15:47																																																																																																																																																	
Control Test	0.200	15:48																																																																																																																																																	
Air Blank	0.000	15:48																																																																																																																																																	
Control Test	0.198	15:49																																																																																																																																																	
Air Blank	0.000	15:50																																																																																																																																																	
Control Test	0.198	15:50																																																																																																																																																	
Air Blank	0.000	15:51																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1987																																																																																																																																																		
Std Dev	0.0012																																																																																																																																																		
Rel Std Dev(%)	0.5812																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	15:23																																																																																																																																																	
Control Test	0.079	15:23																																																																																																																																																	
Air Blank	0.000	15:24																																																																																																																																																	
Control Test	0.078	15:24																																																																																																																																																	
Air Blank	0.000	15:25																																																																																																																																																	
Control Test	0.080	15:25																																																																																																																																																	
Air Blank	0.000	15:25																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0790																																																																																																																																																		
Std Dev	0.0010																																																																																																																																																		
Rel Std Dev(%)	1.2658																																																																																																																																																		

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: COCONUT CREEK PD
Time of Inspection: 13:45

Date of Inspection: 06/30/2026

Serial Number: 80-001044
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202603A Exp: 03/03/2028	0.08g/210L Test (g/210L) Lot#:202603B Exp: 03/03/2028	0.20g/210L Test (g/210L) Lot#:202603D Exp: 03/04/2028	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG526603 Exp: 09/23/2027
0.000	0.050	0.077	0.198	0.078
0.000	0.049	0.077	0.198	0.078
0.000	0.050	0.078	0.198	0.078
0.000	0.049	0.077	0.199	0.078
0.000	0.049	0.077	0.199	0.079
0.000	0.049	0.077	0.199	0.078
0.000	0.050	0.077	0.198	0.079
0.000	0.050	0.077	0.199	0.078
0.000	0.050	0.077	0.198	0.078
0.000	0.050	0.077	0.199	0.079


Standard Deviations	0.0005	0.0003	0.0005	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

06/30/2026
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001044, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001044</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>COCONUT CREEK PD</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>06/30/2026</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>13:45</u>	0.200 g/ 210 L	0.008
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

06/30/2026

Date

**Taylor
Gutschow**

Digitally signed by Taylor Gutschow
Date: 2026.06.30 14:35:28 -04'00'

TAYLOR D GUTSCHOW,
Department Inspector

FDLE/ATP Form 69 January 2026
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Page 1 of 1