

Alcohol Testing Program - Instrument Processing Sheet

Agency: Fort Myers Police Department

Instrument Serial Number: 80-000914

Date In: 6/23/2026

DI Completion Date: 6/30/2026

☐ Ship ☒ P/U ☐ H/D ☐ CMI ☐ EE

Intake By: <u>TDG</u> Date: <u>6/23/2026</u>		Quality Checks By: <u>TDG</u> Date: <u>6/29/2026</u>		Flow Adjustment By: _____																																							
☒ Annual ☐ Dropped Off ☐ Registration ☒ Return from CMI / EE ☐ Training Instrument Visual Inspection ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/Accessories ☐ Power Cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes:		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value: <u>230</u> ☒ Flow Verification (L/s) Flow Column #: <u>ATP106</u> 32 mm 0.152 (.139-.169) 36 mm 0.167 (.156-.190) 53 mm 0.238 (.228-.278) 103 mm 0.500 (.447-.547) ☒ Barometric Pressure Check Gauge ID #: <u>33363</u> Gauge: <u>1019</u> Instrument: <u>1020</u> ☒ Stability Checks		Flow Column #: _____ ☐ 5L/min – 17mm ☐ 15L/min – 53mm ☐ 30L/min – 103mm ☐ R-Value: _____ ☐ Post Adjustment Verification (L/S) Flow Column #: _____ 32 mm (.139-.169) 36 mm (.156-.190) 53 mm (.228-.278) 103 mm (.447-.547)																																							
						<table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th style="width: 15%;">Simulator</th> <th style="width: 20%;">Serial #</th> <th style="width: 25%;">Lot#/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>MP6286</td> <td>202603A 3/3/2028</td> </tr> <tr> <td>0.080</td> <td>MP6287</td> <td>202603B 3/3/2028</td> </tr> <tr> <td>0.200</td> <td>MP6288</td> <td>202603D 3/4/2028</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG526603 9/23/2027</td> </tr> </tbody> </table>		Simulator	Serial #	Lot#/Exp	0.050	MP6286	202603A 3/3/2028	0.080	MP6287	202603B 3/3/2028	0.200	MP6288	202603D 3/4/2028	0.080 DGS	N/A	AG526603 9/23/2027	Maintenance By: _____ Date: _____ ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Tank Sensor Tare ☐ Breath Tube Replacement ☐ Other:																				
						Simulator	Serial #	Lot#/Exp																																			
						0.050	MP6286	202603A 3/3/2028																																			
						0.080	MP6287	202603B 3/3/2028																																			
0.200	MP6288	202603D 3/4/2028																																									
0.080 DGS	N/A	AG526603 9/23/2027																																									
Optical Bench Adjustment By: <u>TDG</u> Department Inspection By: <u>TDG</u> Barometric Pressure Gauge: <u>1019</u> ID#: <u>33364</u> Barometric Pressure ID#: <u>33363</u> <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th style="width: 15%;">Simulator</th> <th style="width: 20%;">Serial #</th> <th style="width: 25%;">Lot #</th> <th style="width: 20%;">Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>MP5097</td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td>MP5098</td> <td>25090</td> <td>3/11/2027</td> </tr> <tr> <td>0.100</td> <td>MP5099</td> <td>24390</td> <td>10/29/2026</td> </tr> <tr> <td>0.200</td> <td>MP5100</td> <td>25020</td> <td>1/14/2027</td> </tr> <tr> <td>0.300</td> <td>MP5101</td> <td>24430</td> <td>12/10/2026</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>28424080A3</td> <td>11/5/2026</td> </tr> </tbody> </table>		Simulator	Serial #	Lot #	Expiration	0.000	MP5097	N/A	N/A	0.040	MP5098	25090	3/11/2027	0.100	MP5099	24390	10/29/2026	0.200	MP5100	25020	1/14/2027	0.300	MP5101	24430	12/10/2026	0.080 DGS	N/A	28424080A3	11/5/2026	Gauge: <u>1017</u> Instrument: <u>1017</u> Mouth Alcohol Solution Lot #: <u>2025-D</u> Exp: <u>9/25/2027</u> Acetone Stock Solution Lot #: <u>2025-B</u> Exp: <u>9/22/2027</u> <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th style="width: 50%;">Simulator</th> <th style="width: 50%;">Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>MP6284</td> </tr> <tr> <td>Interferent</td> <td>MP6285</td> </tr> <tr> <td>0.050</td> <td>MP6286</td> </tr> <tr> <td>0.080</td> <td>MP6287</td> </tr> <tr> <td>0.200</td> <td>MP6288</td> </tr> </tbody> </table>		Simulator	Serial Number	0.000	MP6284	Interferent	MP6285	0.050	MP6286	0.080	MP6287	0.200	MP6288
Simulator	Serial #	Lot #	Expiration																																								
0.000	MP5097	N/A	N/A																																								
0.040	MP5098	25090	3/11/2027																																								
0.100	MP5099	24390	10/29/2026																																								
0.200	MP5100	25020	1/14/2027																																								
0.300	MP5101	24430	12/10/2026																																								
0.080 DGS	N/A	28424080A3	11/5/2026																																								
Simulator	Serial Number																																										
0.000	MP6284																																										
Interferent	MP6285																																										
0.050	MP6286																																										
0.080	MP6287																																										
0.200	MP6288																																										
☒ Post Optical Bench Adjustment Stability Checks <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th style="width: 15%;">Simulator</th> <th style="width: 20%;">Serial #</th> <th style="width: 25%;">Lot #</th> <th style="width: 20%;">Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>MP6286</td> <td>202603A</td> <td>3/3/2028</td> </tr> <tr> <td>0.080</td> <td>MP6287</td> <td>202603B</td> <td>3/3/2028</td> </tr> <tr> <td>0.200</td> <td>MP6288</td> <td>202603D</td> <td>3/4/2028</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG526603</td> <td>9/23/2027</td> </tr> </tbody> </table>		Simulator	Serial #	Lot #	Expiration	0.050	MP6286	202603A	3/3/2028	0.080	MP6287	202603B	3/3/2028	0.200	MP6288	202603D	3/4/2028	0.080 DGS	N/A	AG526603	9/23/2027	☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☒ Optical Bench Adjustment																					
Simulator	Serial #	Lot #	Expiration																																								
0.050	MP6286	202603A	3/3/2028																																								
0.080	MP6287	202603B	3/3/2028																																								
0.200	MP6288	202603D	3/4/2028																																								
0.080 DGS	N/A	AG526603	9/23/2027																																								
Gauge ID #: <u>33363</u> Gauge: <u>1017</u> Instrument: <u>1017</u>		☒ Post-Stability Checks ☐ Flow Adjustment ☐ Form 40 ☐ Other:																																									
Notes/Suggested Service:		☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC																																									
		☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use																																									
		☒ Conduct an Agency Inspection Before Evidentiary Use																																									
LeAndra Higginbotham Digitally signed by LeAndra Higginbotham Date: 2026.07.09 17:03:25 -04'00'		Wen-Chi Pierson Digitally signed by Wen-Chi Pierson Date: 2026.07.10 08:22:00 -04'00'																																									
Tech Review		Admin Review																																									

Stability Checks

0.05g/210L 0.047 to 0.053	0.08g/210L 0.077 to 0.083	0.20g/210L 0.194 to 0.206	DGS 0.08g/210L 0.077 to 0.083 ≤ 0.003 of Wet																																																																																																																																																
FORT MYERS PD Intoxilyzer - Alcohol Analyzer Model: 8000 SN 80-000914 06/29/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>10:27</td></tr> <tr><td>Control Test</td><td>0.048</td><td>10:28</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:29</td></tr> <tr><td>Control Test</td><td>0.047</td><td>10:29</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:30</td></tr> <tr><td>Control Test</td><td>0.048</td><td>10:30</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:31</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0477</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.2112</td><td></td></tr> </tbody> </table> Operator's Signature <i>MC</i>	Test	g/210L	Time	Air Blank	0.000	10:27	Control Test	0.048	10:28	Air Blank	0.000	10:29	Control Test	0.047	10:29	Air Blank	0.000	10:30	Control Test	0.048	10:30	Air Blank	0.000	10:31	Control Test Stats			Average	0.0477		Std Dev	0.0006		Rel Std Dev(%)	1.2112		FORT MYERS PD Intoxilyzer - Alcohol Analyzer Model: 8000 SN 80-000914 06/29/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>12:58</td></tr> <tr><td>Control Test</td><td>0.076</td><td>12:59</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:00</td></tr> <tr><td>Control Test</td><td>0.075</td><td>13:00</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:01</td></tr> <tr><td>Control Test</td><td>0.075</td><td>13:02</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:02</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0757</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7630</td><td></td></tr> </tbody> </table> Operator's Signature <i>MC</i>	Test	g/210L	Time	Air Blank	0.000	12:58	Control Test	0.076	12:59	Air Blank	0.000	13:00	Control Test	0.075	13:00	Air Blank	0.000	13:01	Control Test	0.075	13:02	Air Blank	0.000	13:02	Control Test Stats			Average	0.0757		Std Dev	0.0006		Rel Std Dev(%)	0.7630		FORT MYERS PD Intoxilyzer - Alcohol Analyzer Model: 8000 SN 80-000914 06/29/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>13:05</td></tr> <tr><td>Control Test</td><td>0.197</td><td>13:06</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:06</td></tr> <tr><td>Control Test</td><td>0.196</td><td>13:07</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:08</td></tr> <tr><td>Control Test</td><td>0.196</td><td>13:08</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:09</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1963</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.2941</td><td></td></tr> </tbody> </table> Operator's Signature <i>MC</i>	Test	g/210L	Time	Air Blank	0.000	13:05	Control Test	0.197	13:06	Air Blank	0.000	13:06	Control Test	0.196	13:07	Air Blank	0.000	13:08	Control Test	0.196	13:08	Air Blank	0.000	13:09	Control Test Stats			Average	0.1963		Std Dev	0.0006		Rel Std Dev(%)	0.2941		FORT MYERS PD Intoxilyzer - Alcohol Analyzer Model: 8000 SN 80-000914 06/29/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>13:10</td></tr> <tr><td>Control Test</td><td>0.079</td><td>13:10</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:10</td></tr> <tr><td>Control Test</td><td>0.079</td><td>13:11</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:11</td></tr> <tr><td>Control Test</td><td>0.078</td><td>13:12</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:12</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0787</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7339</td><td></td></tr> </tbody> </table> Operator's Signature <i>MC</i>	Test	g/210L	Time	Air Blank	0.000	13:10	Control Test	0.079	13:10	Air Blank	0.000	13:10	Control Test	0.079	13:11	Air Blank	0.000	13:11	Control Test	0.078	13:12	Air Blank	0.000	13:12	Control Test Stats			Average	0.0787		Std Dev	0.0006		Rel Std Dev(%)	0.7339	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	10:27																																																																																																																																																	
Control Test	0.048	10:28																																																																																																																																																	
Air Blank	0.000	10:29																																																																																																																																																	
Control Test	0.047	10:29																																																																																																																																																	
Air Blank	0.000	10:30																																																																																																																																																	
Control Test	0.048	10:30																																																																																																																																																	
Air Blank	0.000	10:31																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0477																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	1.2112																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	12:58																																																																																																																																																	
Control Test	0.076	12:59																																																																																																																																																	
Air Blank	0.000	13:00																																																																																																																																																	
Control Test	0.075	13:00																																																																																																																																																	
Air Blank	0.000	13:01																																																																																																																																																	
Control Test	0.075	13:02																																																																																																																																																	
Air Blank	0.000	13:02																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0757																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7630																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	13:05																																																																																																																																																	
Control Test	0.197	13:06																																																																																																																																																	
Air Blank	0.000	13:06																																																																																																																																																	
Control Test	0.196	13:07																																																																																																																																																	
Air Blank	0.000	13:08																																																																																																																																																	
Control Test	0.196	13:08																																																																																																																																																	
Air Blank	0.000	13:09																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1963																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.2941																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	13:10																																																																																																																																																	
Control Test	0.079	13:10																																																																																																																																																	
Air Blank	0.000	13:10																																																																																																																																																	
Control Test	0.079	13:11																																																																																																																																																	
Air Blank	0.000	13:11																																																																																																																																																	
Control Test	0.078	13:12																																																																																																																																																	
Air Blank	0.000	13:12																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0787																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7339																																																																																																																																																		

FORT MYERS PD
Intoxilyzer - Alcohol Analyzer
Model 6000 SN 60-000914
06/29/2026 13:16:04

Auto Calibration
Max Power Res Value = 25
Auto Range Res Value = 14

Sol Value = 0.000 g/210L ***
Fit Value = 0.0000 mg/l %
Samples Taken = 4, Discarded = 1
Sum Io = 12517, Sum Io = 13887

Channel 1
Sample % Abs (% Abs Ref)
Sample #1 = 0.0940 (-0.0260)
Sample #2 = 0.0860 (-0.0070)
Sample #3 = 0.0270 (0.0220)
Sample #4 = 0.0870 (0.0130)
Avg % Abs = 0.0667 (0.0093)
STD DEV = 0.0344 (0.0148)
REL STD DEV = 51.534 (159.039)

Channel 2
Sample % Abs (% Abs Ref)
Sample #1 = 0.1150 (0.0070)
Sample #2 = 0.1150 (0.0070)
Sample #3 = 0.0690 (0.0210)
Sample #4 = 0.1140 (-0.0060)
Avg % Abs = 0.0993 (0.0073)
STD DEV = 0.0263 (0.0135)
REL STD DEV = 26.451 (184.133)

Sol Value = 0.040 g/210L ***
Fit Value = 0.1905 mg/l %
Samples Taken = 4, Discarded = 1
Sum Io = 12519, Sum Io = 13884

Channel 1
Sample % Abs (% Abs Ref)
Sample #1 = 0.7550 (-0.0060)
Sample #2 = 0.7770 (-0.0030)
Sample #3 = 0.7420 (0.0160)
Sample #4 = 0.7620 (0.0040)
Avg % Abs = 0.7603 (0.0057)
STD DEV = 0.0176 (0.0096)
REL STD DEV = 2.309 (169.571)

Channel 2
Sample % Abs (% Abs Ref)
Sample #1 = 1.4880 (0.0040)
Sample #2 = 1.5160 (0.0030)
Sample #3 = 1.4830 (0.0120)
Sample #4 = 1.5030 (0.0040)
Avg % Abs = 1.5013 (0.0063)
STD DEV = 0.0176 (0.0049)
REL STD DEV = 1.170 (77.888)

Sol Value = 0.100 g/210L ***
Fit Value = 0.4762 mg/l %
Samples Taken = 4, Discarded = 1
Sum Io = 12517, Sum Io = 13881

Channel 1
Sample % Abs (% Abs Ref)
Sample #1 = 1.7840 (-0.0060)
Sample #2 = 1.8230 (-0.0100)
Sample #3 = 1.7810 (0.0080)
Sample #4 = 1.7520 (0.0210)
Avg % Abs = 1.7853 (0.0063)
STD DEV = 0.0357 (0.0156)
REL STD DEV = 2.000 (245.796)

Channel 2
Sample % Abs (% Abs Ref)
Sample #1 = 3.5490 (-0.0150)
Sample #2 = 3.5590 (-0.0010)
Sample #3 = 3.5420 (0.0020)
Sample #4 = 3.5210 (-0.0060)
Avg % Abs = 3.5407 (-0.0017)
STD DEV = 0.0190 (0.0040)
REL STD DEV = 0.538 (242.487)

Sol Value = 0.200 g/210L ***
Fit Value = 0.9524 mg/l %
Samples Taken = 4, Discarded = 1
Sum Io = 12517, Sum Io = 13880

Channel 1
Sample % Abs (% Abs Ref)
Sample #1 = 3.4970 (-0.0120)
Sample #2 = 3.4460 (0.0130)
Sample #3 = 3.4150 (0.0440)
Sample #4 = 3.4730 (0.0180)
Avg % Abs = 3.4447 (0.0250)
STD DEV = 0.0290 (0.0166)
REL STD DEV = 0.843 (66.573)

Channel 2
Sample % Abs (% Abs Ref)
Sample #1 = 6.8090 (-0.0160)
Sample #2 = 6.7460 (0.0190)
Sample #3 = 6.7380 (0.0140)
Sample #4 = 6.8050 (0.0070)
Avg % Abs = 6.7630 (0.0133)
STD DEV = 0.0366 (0.0060)
REL STD DEV = 0.541 (45.208)

Sol Value = 0.300 g/210L ***
Fit Value = 1.4286 mg/l %
Samples Taken = 4, Discarded = 1
Sum Io = 12515, Sum Io = 13880

Channel 1
Sample % Abs (% Abs Ref)
Sample #1 = 5.0850 (-0.0040)
Sample #2 = 5.0820 (-0.0090)
Sample #3 = 5.0470 (0.0260)
Sample #4 = 5.0560 (0.0380)
Avg % Abs = 5.0617 (0.0183)
STD DEV = 0.0182 (0.0244)
REL STD DEV = 0.359 (133.200)

Channel 2
Sample % Abs (% Abs Ref)
Sample #1 = 9.8790 (0.0050)
Sample #2 = 9.8300 (0.0040)
Sample #3 = 9.8090 (0.0510)
Sample #4 = 9.8280 (0.0580)
Avg % Abs = 9.8223 (0.0377)
STD DEV = 0.0116 (0.0294)
REL STD DEV = 0.118 (77.962)

AUTO CAL DATA

Channel 1
Sol Val = 0.000 mg/l or 0.000 g/210L
% Abs = 0.067
Std Dev = 0.03 Rel Std Dev = 51.53
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 0.760
Std Dev = 0.02 Rel Std Dev = 2.31
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 1.785
Std Dev = 0.04 Rel Std Dev = 2.00
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 3.445
Std Dev = 0.03 Rel Std Dev = 0.84
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 5.062
Std Dev = 0.02 Rel Std Dev = 0.36
Zero Order Coef = -183.39
First Order Coef = 2726.08
Second Order Coef = 26.24
Standard Deviation = 3.955993

Channel 2
Sol Val = 0.000 mg/l or 0.000 g/210L
% Abs = 0.099
Std Dev = 0.03 Rel Std Dev = 26.45
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 1.501
Std Dev = 0.02 Rel Std Dev = 1.17
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 3.541
Std Dev = 0.02 Rel Std Dev = 0.54
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 6.763
Std Dev = 0.04 Rel Std Dev = 0.54
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 9.822
Std Dev = 0.01 Rel Std Dev = 0.12
Zero Order Coef = -133.69
First Order Coef = 1336.81
Second Order Coef = 13.37
Standard Deviation = 3.765532

Solution Stats Quadratic Fit Chan 1
Act Fit Residual
g/210L g/210L g/210L
0.000 -0.000 0.0000
0.040 0.040 0.0000
0.100 0.100 -0.0001
0.200 0.200 0.0001
0.300 0.300 -0.0000

Solution Stats Quadratic Fit Chan 2
Act Fit Residual
g/210L g/210L g/210L
0.000 -0.000 0.0000
0.040 0.040 0.0000
0.100 0.100 -0.0001
0.200 0.200 0.0001
0.300 0.300 -0.0000

Sol Value = 0.080 g/210L ***
Fit Value = 0.3810 mg/l %
Samples Taken = 4, Discarded = 1
Channel 1
Sample #1 = 3259.00
Sample #2 = 3284.00
Sample #3 = 3369.00
Sample #4 = 3308.00
Average Result = 3320.3333
STD DEV = 43.6216
REL STD DEV = 1.320

Channel 2
Sample #1 = 3464.00
Sample #2 = 3408.00
Sample #3 = 3483.00
Sample #4 = 3426.00
Average Result = 3439.0000
STD DEV = 39.1535
REL STD DEV = 1.139

Dry Gas H2O Adjust Results
Barometric Pressure = 1018
3 um H2O Adjust (mg/l*10,000) = 489
9 um H2O Adjust (mg/l*10,000) = 370
**** AUTO CAL PASS

Optical Calibration Adjustment

By: TDG

Post-Cal Stability Checks

0.05g/210L 0.047 to 0.053	0.08g/210L 0.077 to 0.083	0.20g/210L 0.194 to 0.206	DGS 0.08g/210L 0.077 to 0.083 ≤ 0.003 of Wet																																																																																																																																																
FORT MYERS PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000914 06/29/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>15:16</td></tr> <tr><td>Control Test</td><td>0.049</td><td>15:17</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:17</td></tr> <tr><td>Control Test</td><td>0.050</td><td>15:18</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:19</td></tr> <tr><td>Control Test</td><td>0.051</td><td>15:19</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:20</td></tr> <tr><td>Control Test</td><td>0.050</td><td>15:20</td></tr> <tr><td>Average</td><td>0.0500</td><td></td></tr> <tr><td>Std Dev</td><td>0.0010</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>2.0000</td><td></td></tr> </tbody> </table> Operator's Signature <i>ML</i>	Test	g/210L	Time	Air Blank	0.000	15:16	Control Test	0.049	15:17	Air Blank	0.000	15:17	Control Test	0.050	15:18	Air Blank	0.000	15:19	Control Test	0.051	15:19	Air Blank	0.000	15:20	Control Test	0.050	15:20	Average	0.0500		Std Dev	0.0010		Rel Std Dev(%)	2.0000		FORT MYERS PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000914 06/29/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>15:25</td></tr> <tr><td>Control Test</td><td>0.078</td><td>15:26</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:27</td></tr> <tr><td>Control Test</td><td>0.078</td><td>15:27</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:28</td></tr> <tr><td>Control Test</td><td>0.078</td><td>15:29</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:29</td></tr> <tr><td>Control Test</td><td>0.078</td><td>15:29</td></tr> <tr><td>Average</td><td>0.0780</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> Operator's Signature <i>ML</i>	Test	g/210L	Time	Air Blank	0.000	15:25	Control Test	0.078	15:26	Air Blank	0.000	15:27	Control Test	0.078	15:27	Air Blank	0.000	15:28	Control Test	0.078	15:29	Air Blank	0.000	15:29	Control Test	0.078	15:29	Average	0.0780		Std Dev	0.0000		Rel Std Dev(%)	0.0000		FORT MYERS PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000914 06/29/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>15:31</td></tr> <tr><td>Control Test</td><td>0.198</td><td>15:32</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:33</td></tr> <tr><td>Control Test</td><td>0.197</td><td>15:33</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:34</td></tr> <tr><td>Control Test</td><td>0.198</td><td>15:35</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:35</td></tr> <tr><td>Control Test</td><td>0.197</td><td>15:35</td></tr> <tr><td>Average</td><td>0.1977</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.2921</td><td></td></tr> </tbody> </table> Operator's Signature <i>ML</i>	Test	g/210L	Time	Air Blank	0.000	15:31	Control Test	0.198	15:32	Air Blank	0.000	15:33	Control Test	0.197	15:33	Air Blank	0.000	15:34	Control Test	0.198	15:35	Air Blank	0.000	15:35	Control Test	0.197	15:35	Average	0.1977		Std Dev	0.0006		Rel Std Dev(%)	0.2921		FORT MYERS PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000914 06/29/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>15:03</td></tr> <tr><td>Control Test</td><td>0.079</td><td>15:03</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:03</td></tr> <tr><td>Control Test</td><td>0.077</td><td>15:04</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:04</td></tr> <tr><td>Control Test</td><td>0.078</td><td>15:05</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:05</td></tr> <tr><td>Control Test</td><td>0.078</td><td>15:05</td></tr> <tr><td>Average</td><td>0.0780</td><td></td></tr> <tr><td>Std Dev</td><td>0.0010</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.2821</td><td></td></tr> </tbody> </table> Operator's Signature <i>ML</i>	Test	g/210L	Time	Air Blank	0.000	15:03	Control Test	0.079	15:03	Air Blank	0.000	15:03	Control Test	0.077	15:04	Air Blank	0.000	15:04	Control Test	0.078	15:05	Air Blank	0.000	15:05	Control Test	0.078	15:05	Average	0.0780		Std Dev	0.0010		Rel Std Dev(%)	1.2821	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	15:16																																																																																																																																																	
Control Test	0.049	15:17																																																																																																																																																	
Air Blank	0.000	15:17																																																																																																																																																	
Control Test	0.050	15:18																																																																																																																																																	
Air Blank	0.000	15:19																																																																																																																																																	
Control Test	0.051	15:19																																																																																																																																																	
Air Blank	0.000	15:20																																																																																																																																																	
Control Test	0.050	15:20																																																																																																																																																	
Average	0.0500																																																																																																																																																		
Std Dev	0.0010																																																																																																																																																		
Rel Std Dev(%)	2.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	15:25																																																																																																																																																	
Control Test	0.078	15:26																																																																																																																																																	
Air Blank	0.000	15:27																																																																																																																																																	
Control Test	0.078	15:27																																																																																																																																																	
Air Blank	0.000	15:28																																																																																																																																																	
Control Test	0.078	15:29																																																																																																																																																	
Air Blank	0.000	15:29																																																																																																																																																	
Control Test	0.078	15:29																																																																																																																																																	
Average	0.0780																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	15:31																																																																																																																																																	
Control Test	0.198	15:32																																																																																																																																																	
Air Blank	0.000	15:33																																																																																																																																																	
Control Test	0.197	15:33																																																																																																																																																	
Air Blank	0.000	15:34																																																																																																																																																	
Control Test	0.198	15:35																																																																																																																																																	
Air Blank	0.000	15:35																																																																																																																																																	
Control Test	0.197	15:35																																																																																																																																																	
Average	0.1977																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.2921																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	15:03																																																																																																																																																	
Control Test	0.079	15:03																																																																																																																																																	
Air Blank	0.000	15:03																																																																																																																																																	
Control Test	0.077	15:04																																																																																																																																																	
Air Blank	0.000	15:04																																																																																																																																																	
Control Test	0.078	15:05																																																																																																																																																	
Air Blank	0.000	15:05																																																																																																																																																	
Control Test	0.078	15:05																																																																																																																																																	
Average	0.0780																																																																																																																																																		
Std Dev	0.0010																																																																																																																																																		
Rel Std Dev(%)	1.2821																																																																																																																																																		

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: FORT MYERS PD

Time of Inspection: 13:03

Date of Inspection: 06/30/2026

Serial Number: 80-000914

Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202603A Exp: 03/03/2028	0.08g/210L Test (g/210L) Lot#:202603B Exp: 03/03/2028	0.20g/210L Test (g/210L) Lot#:202603D Exp: 03/04/2028	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG526603 Exp: 09/23/2027
0.000	0.050	0.078	0.198	0.079
0.000	0.050	0.078	0.197	0.078
0.000	0.049	0.078	0.198	0.079
0.000	0.050	0.078	0.198	0.078
0.000	0.050	0.078	0.198	0.079
0.000	0.050	0.079	0.198	0.078
0.000	0.050	0.078	0.198	0.079
0.000	0.050	0.078	0.198	0.079
0.000	0.050	0.078	0.198	0.079
0.000	0.050	0.078	0.198	0.079
0.000	0.050	0.078	0.197	0.079


Standard Deviations	0.0003	0.0003	0.0004	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

06/30/2026
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000914, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-000914</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>FORT MYERS PD</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>06/30/2026</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>13:03</u>	0.200 g/ 210 L	0.008
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

06/30/2026

Date

**Taylor
Gutschow**

Digitally signed by Taylor
Gutschow
Date: 2026.06.30 14:35:18
+04'00'

**TAYLOR D GUTSCHOW,
Department Inspector**