

Alcohol Testing Program - Instrument Processing Sheet

Agency: NEW SMYRNA BEACH PD Instrument Serial Number: 80-001152
 Date In: 1/6/2026 DI Completion Date: 1/27/2026 Ship P/U H/D CMI EE

Intake By: <u>KTS</u> Date: <u>1/6/2026</u>	Quality Checks By: <u>SLH</u> Date: <u>1/13/2026</u>	Flow Adjustment By: _____																			
<input checked="" type="checkbox"/> Annual <input type="checkbox"/> Dropped Off <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE <input type="checkbox"/> Training Instrument Visual Inspection <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/Accessories <input checked="" type="checkbox"/> Power Cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: AI noted instrument testing outside of tolerance for 0.080g/210L. KTS 1/6/26	<input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value: <u>145</u> * <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column #: <u>ATP 103</u> * 32 mm <u>0.152</u> (.139-.169) 36 mm <u>0.167</u> (.156-.190) 53 mm <u>0.234</u> (.228-.278) 103 mm <u>0.492</u> (.447-.547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID #: <u>28427</u> Gauge: <u>1018</u> Instrument: <u>1017</u> <input checked="" type="checkbox"/> Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot#/Exp</th> </tr> </thead> <tbody> <tr> <td rowspan="2">0.050</td> <td rowspan="2">MP5088</td> <td>202406K</td> </tr> <tr> <td>06/19/2026</td> </tr> <tr> <td rowspan="2">0.080</td> <td rowspan="2">MP5089</td> <td>202406L</td> </tr> <tr> <td>06/19/2026</td> </tr> <tr> <td rowspan="2">0.200</td> <td rowspan="2">MP5090</td> <td>202406N</td> </tr> <tr> <td>06/20/2026</td> </tr> <tr> <td rowspan="2">0.080 DGS</td> <td rowspan="2">N/A</td> <td>AG510701</td> </tr> <tr> <td>04/17/2027</td> </tr> </tbody> </table>	Simulator	Serial #	Lot#/Exp	0.050	MP5088	202406K	06/19/2026	0.080	MP5089	202406L	06/19/2026	0.200	MP5090	202406N	06/20/2026	0.080 DGS	N/A	AG510701	04/17/2027	Flow Column #: _____ <input type="checkbox"/> 5L/min – 17mm <input type="checkbox"/> 15L/min – 53mm <input type="checkbox"/> 30L/min – 103mm <input type="checkbox"/> R-Value: _____ <input type="checkbox"/> Post Adjustment Verification (L/S) Flow Column #: _____ 32 mm _____ (.139-.169) 36 mm _____ (.156-.190) 53 mm _____ (.228-.278) 103 mm _____ (.447-.547)
Simulator	Serial #	Lot#/Exp																			
0.050	MP5088	202406K																			
		06/19/2026																			
0.080	MP5089	202406L																			
		06/19/2026																			
0.200	MP5090	202406N																			
		06/20/2026																			
0.080 DGS	N/A	AG510701																			
		04/17/2027																			
Optical Bench Adjustment By: <u>WKP</u> Department Inspection By: <u>SLH</u> Barometric Pressure Gauge: <u>1028</u> ID#: <u>34419</u> Barometric Pressure ID#: <u>28662</u>		Maintenance By: _____ Date: _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Tank Sensor Tare <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other: _____																			

Optical Bench Adjustment By: <u>WKP</u> Barometric Pressure Gauge: <u>1028</u> ID#: <u>34419</u>	Department Inspection By: <u>SLH</u> Barometric Pressure ID#: <u>28662</u> Gauge: <u>1024</u> Instrument: <u>1025</u> Mouth Alcohol Solution Lot #: <u>2025-D</u> Exp: <u>09/25/2027</u> Acetone Stock Solution Lot #: <u>2025-B</u> Exp: <u>09/22/2027</u>																																								
<table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>MP6294</td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td>MP6295</td> <td>25090</td> <td>03/11/2027</td> </tr> <tr> <td>0.100</td> <td>MP6296</td> <td>25282</td> <td>08/12/2027</td> </tr> <tr> <td>0.200</td> <td>MP6297</td> <td>25020</td> <td>01/14/2027</td> </tr> <tr> <td>0.300</td> <td>MP6298</td> <td>25150</td> <td>05/06/27</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>28424080A3</td> <td>11/05/2026</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #	Expiration	0.000	MP6294	N/A	N/A	0.040	MP6295	25090	03/11/2027	0.100	MP6296	25282	08/12/2027	0.200	MP6297	25020	01/14/2027	0.300	MP6298	25150	05/06/27	0.080 DGS	N/A	28424080A3	11/05/2026	<table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>MP5086</td> </tr> <tr> <td>Interferent</td> <td>MP5087</td> </tr> <tr> <td>0.050</td> <td>MP5088</td> </tr> <tr> <td>0.080</td> <td>MP5089</td> </tr> <tr> <td>0.200</td> <td>MP5090</td> </tr> </tbody> </table>	Simulator	Serial Number	0.000	MP5086	Interferent	MP5087	0.050	MP5088	0.080	MP5089	0.200	MP5090
Simulator	Serial #	Lot #	Expiration																																						
0.000	MP6294	N/A	N/A																																						
0.040	MP6295	25090	03/11/2027																																						
0.100	MP6296	25282	08/12/2027																																						
0.200	MP6297	25020	01/14/2027																																						
0.300	MP6298	25150	05/06/27																																						
0.080 DGS	N/A	28424080A3	11/05/2026																																						
Simulator	Serial Number																																								
0.000	MP5086																																								
Interferent	MP5087																																								
0.050	MP5088																																								
0.080	MP5089																																								
0.200	MP5090																																								
<input checked="" type="checkbox"/> Post Optical Bench Adjustment Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>MP6291</td> <td>202406K</td> <td>06/19/2026</td> </tr> <tr> <td>0.080</td> <td>MP6292</td> <td>202406L</td> <td>06/19/2026</td> </tr> <tr> <td>0.200</td> <td>MP6293</td> <td>202406N</td> <td>06/20/2026</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG510701</td> <td>04/17/2027</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #	Expiration	0.050	MP6291	202406K	06/19/2026	0.080	MP6292	202406L	06/19/2026	0.200	MP6293	202406N	06/20/2026	0.080 DGS	N/A	AG510701	04/17/2027	Attachments <input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Stability Checks <input checked="" type="checkbox"/> Calibration Certificate <input checked="" type="checkbox"/> Optical Bench Adjustment <input checked="" type="checkbox"/> Post-Stability Checks <input type="checkbox"/> Flow Adjustment <input type="checkbox"/> Form 40 <input checked="" type="checkbox"/> Other: AI Note																				
Simulator	Serial #	Lot #	Expiration																																						
0.050	MP6291	202406K	06/19/2026																																						
0.080	MP6292	202406L	06/19/2026																																						
0.200	MP6293	202406N	06/20/2026																																						
0.080 DGS	N/A	AG510701	04/17/2027																																						
Gauge ID #: <u>34420</u> Gauge: <u>1026</u> Instrument: <u>1025</u>	<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use																																								

Notes/Suggested Service: * performed flow portion of quality checks on 1/20/2026. SLH Additional 0.080g/210L wet stability was performed 1/27/2026. SLH 1/28/2026 Tech Correction: Added mark to box on Post Optical Bench Adjustment Stability Checks. WKP 1/29/2026	Digitally signed by Taylor Gutschow Date: 2026.02.03 12:26:03 -05'00' Digitally signed by Kaitlyn Spearin Date: 2026.02.03 13:20:31 -05'00'
Taylor Gutschow	Kaitlyn Spearin
Tech Review	Admin Review

Helms, William

From: Helms, William
Sent: Tuesday, December 30, 2025 10:29 AM
To: leandrahigginbotham@fdle.state.fl.us
Subject: Instrument 80-001152

I just wanted to let you know that instrument 80-001152 is being sent to you, due to the 0.08 testing outside of tolerance on multiple tests.

Thank you,

Sergeant William Helms #1572

Special Operations Supervisor
New Smyrna Beach Police Department
246 Industrial Park Avenue
New Smyrna Beach, FL 32168
Office: (386) 424-2297
Whelms@cityofnsb.com
www.cityofnsb.com



Stability Checks

0.050 g/210L 0.047 to 0.053 g/210L	0.080 g/210L 0.077 to 0.083 g/210L	0.200 g/210L 0.194 to 0.206 g/210L	DGS 0.080 g/210L 0.077 to 0.083 g/210L 50.003 g/210L of Wet																																																																																																																																				
<p>NEW SWYRNA BEACH PD Intoxilyzer - Alcohol Analyzer Model 8000 01/13/2026 SN 80-001152 Software: 8100.27</p> <p>Test g/210L Time</p> <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>12:25</td></tr> <tr><td>Control Test</td><td>0.051</td><td>12:26</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:26</td></tr> <tr><td>Control Test</td><td>1.052</td><td>12:27</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:28</td></tr> <tr><td>Control Test</td><td>0.152</td><td>12:28</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:29</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0517</td><td></td></tr> <tr><td>Std Dev</td><td>0.0016</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.1175</td><td></td></tr> </table> <p>Operator's Signature: <i>Shigeloh</i></p>	Air Blank	0.000	12:25	Control Test	0.051	12:26	Air Blank	0.000	12:26	Control Test	1.052	12:27	Air Blank	0.000	12:28	Control Test	0.152	12:28	Air Blank	0.000	12:29	Control Test Stats			Average	0.0517		Std Dev	0.0016		Rel Std Dev(%)	1.1175		<p>NEW SWYRNA BEACH PD Intoxilyzer - Alcohol Analyzer Model 8000 01/13/2026 SN 80-001152 Software: 8100.27</p> <p>Test g/210L Time</p> <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>13:10</td></tr> <tr><td>Control Test</td><td>0.082</td><td>13:10</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:11</td></tr> <tr><td>Control Test</td><td>0.082</td><td>13:12</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:12</td></tr> <tr><td>Control Test</td><td>0.082</td><td>13:13</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:13</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0820</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </table> <p>Operator's Signature: <i>Shigeloh</i></p>	Air Blank	0.000	13:10	Control Test	0.082	13:10	Air Blank	0.000	13:11	Control Test	0.082	13:12	Air Blank	0.000	13:12	Control Test	0.082	13:13	Air Blank	0.000	13:13	Control Test Stats			Average	0.0820		Std Dev	0.0000		Rel Std Dev(%)	0.0000		<p>NEW SWYRNA BEACH PD Intoxilyzer - Alcohol Analyzer Model 8000 01/13/2026 SN 80-001152 Software: 8100.27</p> <p>Test g/210L Time</p> <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>12:50</td></tr> <tr><td>Control Test</td><td>0.198</td><td>12:51</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:52</td></tr> <tr><td>Control Test</td><td>0.198</td><td>12:52</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:53</td></tr> <tr><td>Control Test</td><td>0.198</td><td>12:53</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:54</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1980</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </table> <p>Operator's Signature: <i>Shigeloh</i></p>	Air Blank	0.000	12:50	Control Test	0.198	12:51	Air Blank	0.000	12:52	Control Test	0.198	12:52	Air Blank	0.000	12:53	Control Test	0.198	12:53	Air Blank	0.000	12:54	Control Test Stats			Average	0.1980		Std Dev	0.0000		Rel Std Dev(%)	0.0000		<p>DGS</p> <p>NEW SWYRNA BEACH PD Intoxilyzer - Alcohol Analyzer Model 8000 01/13/2026 SN 80-001152 Software: 8100.27</p> <p>Test g/210L Time</p> <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>12:30</td></tr> <tr><td>Control Test</td><td>0.081</td><td>12:31</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:31</td></tr> <tr><td>Control Test</td><td>0.081</td><td>12:31</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:32</td></tr> <tr><td>Control Test</td><td>0.082</td><td>12:32</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:33</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0813</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7199</td><td></td></tr> </table> <p>Operator's Signature: <i>Shigeloh</i></p>	Air Blank	0.000	12:30	Control Test	0.081	12:31	Air Blank	0.000	12:31	Control Test	0.081	12:31	Air Blank	0.000	12:32	Control Test	0.082	12:32	Air Blank	0.000	12:33	Control Test Stats			Average	0.0813		Std Dev	0.0006		Rel Std Dev(%)	0.7199	
Air Blank	0.000	12:25																																																																																																																																					
Control Test	0.051	12:26																																																																																																																																					
Air Blank	0.000	12:26																																																																																																																																					
Control Test	1.052	12:27																																																																																																																																					
Air Blank	0.000	12:28																																																																																																																																					
Control Test	0.152	12:28																																																																																																																																					
Air Blank	0.000	12:29																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.0517																																																																																																																																						
Std Dev	0.0016																																																																																																																																						
Rel Std Dev(%)	1.1175																																																																																																																																						
Air Blank	0.000	13:10																																																																																																																																					
Control Test	0.082	13:10																																																																																																																																					
Air Blank	0.000	13:11																																																																																																																																					
Control Test	0.082	13:12																																																																																																																																					
Air Blank	0.000	13:12																																																																																																																																					
Control Test	0.082	13:13																																																																																																																																					
Air Blank	0.000	13:13																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.0820																																																																																																																																						
Std Dev	0.0000																																																																																																																																						
Rel Std Dev(%)	0.0000																																																																																																																																						
Air Blank	0.000	12:50																																																																																																																																					
Control Test	0.198	12:51																																																																																																																																					
Air Blank	0.000	12:52																																																																																																																																					
Control Test	0.198	12:52																																																																																																																																					
Air Blank	0.000	12:53																																																																																																																																					
Control Test	0.198	12:53																																																																																																																																					
Air Blank	0.000	12:54																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.1980																																																																																																																																						
Std Dev	0.0000																																																																																																																																						
Rel Std Dev(%)	0.0000																																																																																																																																						
Air Blank	0.000	12:30																																																																																																																																					
Control Test	0.081	12:31																																																																																																																																					
Air Blank	0.000	12:31																																																																																																																																					
Control Test	0.081	12:31																																																																																																																																					
Air Blank	0.000	12:32																																																																																																																																					
Control Test	0.082	12:32																																																																																																																																					
Air Blank	0.000	12:33																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.0813																																																																																																																																						
Std Dev	0.0006																																																																																																																																						
Rel Std Dev(%)	0.7199																																																																																																																																						

NEW SMYRNA BEACH PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001152
01/13/2026
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:38
Control Test	0.084	12:39
Air Blank	0.000	12:40
Control Test	0.083	12:40
Air Blank	0.000	12:41
Control Test	0.082	12:42
Air Blank	0.000	12:42
Control Test Stats		
Average	0.0830	
Std Dev	0.0010	
Rel Std Dev(%)	1.2048	

re-fitted SIM / tightened and
re-ran
sub

80-001152

1/13/2026



Operator's Signature

NEW SMYRNA BEACH PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001152
01/27/2026
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:05
Control Test	0.084 *	10:06
Air Blank	0.000	10:07
Control Test	0.083	10:07
Air Blank	0.000	10:08
Control Test	0.083	10:08
Air Blank	0.000	10:09
Control Test	0.083	10:10
Air Blank	0.000	10:10
Control Test	0.083	10:11
Air Blank	0.000	10:11
Control Test	0.083	10:12
Air Blank	0.000	10:13
Control Test	0.083	10:13
Air Blank	0.000	10:14
Control Test	0.083	10:15
Air Blank	0.000	10:15
Control Test	0.083	10:16
Air Blank	0.000	10:16
Control Test	0.083	10:17
Air Blank	0.000	10:18
Control Test Stats		
Average	0.0831	
Std Dev	0.0003	
Rel Std Dev(%)	0.3805	


Operator's Signature

1/27/26 SUT
an additional wet
0.08 g/210L ~~to~~ ^{to} SUT
stability performed
to verify accuracy
again due to the
agency inspector
note. Based on the
10 replicates - an
optical bench
adjustment will
be performed.
SUT

Post-Cal Stability Checks

#75

0.05g/210L 0.047 to 0.053	0.08g/210L 0.077 to 0.083	0.20g/210L 0.194 to 0.206	DGS 0.08g/210L 0.077 to 0.083																																																																																																																																																
<p>NEW SMYRNA BEACH PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001152 01/27/2026 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>14:04</td></tr> <tr><td>Control Test</td><td>0.050</td><td>14:05</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:05</td></tr> <tr><td>Control Test</td><td>0.050</td><td>14:06</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:06</td></tr> <tr><td>Control Test</td><td>0.049</td><td>14:07</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:08</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0497</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.1625</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	14:04	Control Test	0.050	14:05	Air Blank	0.000	14:05	Control Test	0.050	14:06	Air Blank	0.000	14:06	Control Test	0.049	14:07	Air Blank	0.000	14:08	Control Test Stats			Average	0.0497		Std Dev	0.0006		Rel Std Dev(%)	1.1625		<p>NEW SMYRNA BEACH PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001152 01/27/2026 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>14:10</td></tr> <tr><td>Control Test</td><td>0.080</td><td>14:11</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:11</td></tr> <tr><td>Control Test</td><td>0.080</td><td>14:12</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:13</td></tr> <tr><td>Control Test</td><td>0.080</td><td>14:13</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:14</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0803</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7187</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	14:10	Control Test	0.080	14:11	Air Blank	0.000	14:11	Control Test	0.080	14:12	Air Blank	0.000	14:13	Control Test	0.080	14:13	Air Blank	0.000	14:14	Control Test Stats			Average	0.0803		Std Dev	0.0006		Rel Std Dev(%)	0.7187		<p>NEW SMYRNA BEACH PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001152 01/27/2026 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>14:15</td></tr> <tr><td>Control Test</td><td>0.200</td><td>14:16</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:17</td></tr> <tr><td>Control Test</td><td>0.200</td><td>14:17</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:18</td></tr> <tr><td>Control Test</td><td>0.200</td><td>14:18</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:19</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.2003</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.2882</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	14:15	Control Test	0.200	14:16	Air Blank	0.000	14:17	Control Test	0.200	14:17	Air Blank	0.000	14:18	Control Test	0.200	14:18	Air Blank	0.000	14:19	Control Test Stats			Average	0.2003		Std Dev	0.0006		Rel Std Dev(%)	0.2882		<p>NEW SMYRNA BEACH PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001152 01/27/2026 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>14:20</td></tr> <tr><td>Control Test</td><td>0.080</td><td>14:21</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:21</td></tr> <tr><td>Control Test</td><td>0.080</td><td>14:22</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:22</td></tr> <tr><td>Control Test</td><td>0.079</td><td>14:22</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:23</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0797</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7247</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	14:20	Control Test	0.080	14:21	Air Blank	0.000	14:21	Control Test	0.080	14:22	Air Blank	0.000	14:22	Control Test	0.079	14:22	Air Blank	0.000	14:23	Control Test Stats			Average	0.0797		Std Dev	0.0006		Rel Std Dev(%)	0.7247	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	14:04																																																																																																																																																	
Control Test	0.050	14:05																																																																																																																																																	
Air Blank	0.000	14:05																																																																																																																																																	
Control Test	0.050	14:06																																																																																																																																																	
Air Blank	0.000	14:06																																																																																																																																																	
Control Test	0.049	14:07																																																																																																																																																	
Air Blank	0.000	14:08																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0497																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	1.1625																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	14:10																																																																																																																																																	
Control Test	0.080	14:11																																																																																																																																																	
Air Blank	0.000	14:11																																																																																																																																																	
Control Test	0.080	14:12																																																																																																																																																	
Air Blank	0.000	14:13																																																																																																																																																	
Control Test	0.080	14:13																																																																																																																																																	
Air Blank	0.000	14:14																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0803																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7187																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	14:15																																																																																																																																																	
Control Test	0.200	14:16																																																																																																																																																	
Air Blank	0.000	14:17																																																																																																																																																	
Control Test	0.200	14:17																																																																																																																																																	
Air Blank	0.000	14:18																																																																																																																																																	
Control Test	0.200	14:18																																																																																																																																																	
Air Blank	0.000	14:19																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.2003																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.2882																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	14:20																																																																																																																																																	
Control Test	0.080	14:21																																																																																																																																																	
Air Blank	0.000	14:21																																																																																																																																																	
Control Test	0.080	14:22																																																																																																																																																	
Air Blank	0.000	14:22																																																																																																																																																	
Control Test	0.079	14:22																																																																																																																																																	
Air Blank	0.000	14:23																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0797																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7247																																																																																																																																																		
<p>Operator's Signature: <i>menpe</i></p>	<p>Operator's Signature: <i>menpe</i></p>	<p>Operator's Signature: <i>menpe</i></p>	<p>Operator's Signature: <i>menpe</i></p>																																																																																																																																																

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: NEW SMYRNA BEACH PD
Time of Inspection: 17:16

Date of Inspection: 01/27/2026

Serial Number: 80-001152
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202406K Exp: 06/19/2026	0.08g/210L Test (g/210L) Lot#:202406L Exp: 06/19/2026	0.20g/210L Test (g/210L) Lot#:202406N Exp: 06/20/2026	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG510701 Exp: 04/17/2027
0.000	0.048	0.077	0.195	0.079
0.000	0.047	0.077	0.194	0.080
0.000	0.047	0.077	0.195	0.080
0.000	0.046	0.077	0.194	0.080
0.000	0.046	0.076	0.195	0.080
0.000	0.046	0.076	0.195	0.080
0.000	0.045	0.077	0.194	0.080
0.000	0.046	0.076	0.194	0.080
0.000	0.046	0.076	0.194	0.080
0.000	0.046	0.076	0.194	0.080
0.000	0.046	0.076	0.195	0.080


Standard Deviations	0.0008	0.0005	0.0005	0.0003
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0005 Number of Simulators Used: 5


Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.


LEANDRA HIGGINBOTHAM
Signature and Printed Name

01/27/2026
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001152, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001152</u>	UNCERTAINTY* ±	
Owning Agency:	<u>NEW SMYRNA BEACH PD</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>01/27/2026</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>17:16</u>	0.200 g/ 210 L	0.008
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.


LEANDRA HIGGINBOTHAM,
Date 01/27/2026

FDLE/ATP Form 69 January 2026
Issuing Authority: Alcohol Testing Program

Department Inspector

Service • Integrity • Respect • Quality