

Alcohol Testing Program - Instrument Processing Sheet

Agency: City of Miramar Police Department Instrument Serial Number: 80-001067

Date In: 6/16/2026 DI Completion Date: 6/24/2026 ☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake By: <u>TDG</u> Date: <u>6/23/2026</u>		Quality Checks By: <u>TDG</u> Date: <u>6/24/2026</u>		Flow Adjustment By: _____																																							
☒ Annual ☐ Dropped Off ☐ Registration ☐ Return from CMI / EE ☐ Training Instrument Visual Inspection ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/Accessories ☒ Power Cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes: Mini keyboard.		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value: <u>207</u> ☒ Flow Verification (L/s) Flow Column #: <u>ATP106</u> 32 mm <u>0.156</u> (.139-.169) 36 mm <u>0.175</u> (.156-.190) 53 mm <u>0.242</u> (.228-.278) 103 mm <u>0.496</u> (.447-.547) ☒ Barometric Pressure Check Gauge ID #: <u>33363</u> Gauge: <u>1019</u> Instrument: <u>1019</u> ☒ Stability Checks		Flow Column #: _____ ☐ 5L/min – 17mm ☐ 15L/min – 53mm ☐ 30L/min – 103mm ☐ R-Value: _____ ☐ Post Adjustment Verification (L/S) Flow Column #: _____ 32 mm _____ (.139-.169) 36 mm _____ (.156-.190) 53 mm _____ (.228-.278) 103 mm _____ (.447-.547)																																							
						<table border="1" style="width:100%; border-collapse: collapse;"> <tr> <th style="width:15%;">Simulator</th> <th style="width:20%;">Serial #</th> <th style="width:20%;">Lot#/Exp</th> </tr> <tr> <td>0.050</td> <td>MP6286</td> <td>202603A 3/3/2028</td> </tr> <tr> <td>0.080</td> <td>MP6287</td> <td>202603B 3/3/2028</td> </tr> <tr> <td>0.200</td> <td>MP6288</td> <td>202603D 3/4/2028</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG526603 9/23/2027</td> </tr> </table>		Simulator	Serial #	Lot#/Exp	0.050	MP6286	202603A 3/3/2028	0.080	MP6287	202603B 3/3/2028	0.200	MP6288	202603D 3/4/2028	0.080 DGS	N/A	AG526603 9/23/2027	Maintenance By: <u>TDG</u> Date: <u>6/23/2026</u> ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Tank Sensor Tare ☐ Breath Tube Replacement ☒ Other: Added screw to printer port.																				
						Simulator	Serial #	Lot#/Exp																																			
						0.050	MP6286	202603A 3/3/2028																																			
						0.080	MP6287	202603B 3/3/2028																																			
0.200	MP6288	202603D 3/4/2028																																									
0.080 DGS	N/A	AG526603 9/23/2027																																									
Optical Bench Adjustment By: _____ Department Inspection By: <u>TDG</u> Barometric Pressure Gauge: _____ ID#: _____ Barometric Pressure ID#: <u>33363</u> Gauge: <u>1018</u> Instrument: <u>1018</u> Mouth Alcohol Solution Lot #: <u>2025-D</u> Exp: <u>9/25/2027</u> Acetone Stock Solution Lot #: <u>2024-B</u> Exp: <u>7/19/2026</u>																																											
<table border="1" style="width:100%; border-collapse: collapse;"> <tr> <th style="width:15%;">Simulator</th> <th style="width:20%;">Serial #</th> <th style="width:20%;">Lot #</th> <th style="width:20%;">Expiration</th> </tr> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </table>		Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			<table border="1" style="width:100%; border-collapse: collapse;"> <tr> <th style="width:50%;">Simulator</th> <th style="width:50%;">Serial Number</th> </tr> <tr> <td>0.000</td> <td>MP6284</td> </tr> <tr> <td>Interferent</td> <td>MP6285</td> </tr> <tr> <td>0.050</td> <td>MP6286</td> </tr> <tr> <td>0.080</td> <td>MP6287</td> </tr> <tr> <td>0.200</td> <td>MP6288</td> </tr> </table>		Simulator	Serial Number	0.000	MP6284	Interferent	MP6285	0.050	MP6286	0.080	MP6287	0.200	MP6288
Simulator	Serial #	Lot #	Expiration																																								
0.000		N/A	N/A																																								
0.040																																											
0.100																																											
0.200																																											
0.300																																											
0.080 DGS	N/A																																										
Simulator	Serial Number																																										
0.000	MP6284																																										
Interferent	MP6285																																										
0.050	MP6286																																										
0.080	MP6287																																										
0.200	MP6288																																										
☐ Post Optical Bench Adjustment Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <tr> <th style="width:15%;">Simulator</th> <th style="width:20%;">Serial #</th> <th style="width:20%;">Lot #</th> <th style="width:20%;">Expiration</th> </tr> <tr> <td>0.050</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </table> Gauge ID #: _____ Gauge: _____ Instrument: _____		Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			Attachments ☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☐ Optical Bench Adjustment ☐ Post-Stability Checks ☐ Flow Adjustment ☐ Form 40 ☐ Other:																					
		Simulator	Serial #	Lot #	Expiration																																						
		0.050																																									
		0.080																																									
0.200																																											
0.080 DGS	N/A																																										
Notes/Suggested Service:																																											
		☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use																																									
				<table border="0" style="width:100%;"> <tr> <td style="width:50%; vertical-align: top;"> LeAndra Higginbotham Digitally signed by LeAndra Higginbotham Date: 2026.07.07 08:44:40 -04'00' </td> <td style="width:50%; vertical-align: top;"> Wen-Chi Pierson Digitally signed by Wen-Chi Pierson Date: 2026.07.09 15:18:51 -04'00' </td> </tr> </table>		LeAndra Higginbotham Digitally signed by LeAndra Higginbotham Date: 2026.07.07 08:44:40 -04'00'	Wen-Chi Pierson Digitally signed by Wen-Chi Pierson Date: 2026.07.09 15:18:51 -04'00'																																				
LeAndra Higginbotham Digitally signed by LeAndra Higginbotham Date: 2026.07.07 08:44:40 -04'00'	Wen-Chi Pierson Digitally signed by Wen-Chi Pierson Date: 2026.07.09 15:18:51 -04'00'																																										
Tech Review		Admin Review																																									

Stability Checks

0.05g/210L 0.047 to 0.053	0.08g/210L 0.077 to 0.083	0.20g/210L 0.194 to 0.206	DGS 0.08g/210L 0.077 to 0.083 ☒ ≤ 0.003 of Wet ☒																																																																																																																																																
CITY OF MIAMI POL Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001067 06/24/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>12:05</td></tr> <tr><td>Control Test</td><td>0.050</td><td>12:05</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:06</td></tr> <tr><td>Control Test</td><td>0.049</td><td>12:07</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:07</td></tr> <tr><td>Control Test</td><td>0.049</td><td>12:08</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:08</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0493</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.1703</td><td></td></tr> </tbody> </table> Operator's Signature 	Test	g/210L	Time	Air Blank	0.000	12:05	Control Test	0.050	12:05	Air Blank	0.000	12:06	Control Test	0.049	12:07	Air Blank	0.000	12:07	Control Test	0.049	12:08	Air Blank	0.000	12:08	Control Test Stats			Average	0.0493		Std Dev	0.0006		Rel Std Dev(%)	1.1703		CITY OF MIAMI POL Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001067 06/24/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>12:13</td></tr> <tr><td>Control Test</td><td>0.079</td><td>12:14</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:14</td></tr> <tr><td>Control Test</td><td>0.078</td><td>12:15</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:15</td></tr> <tr><td>Control Test</td><td>0.078</td><td>12:16</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:16</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0783</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7370</td><td></td></tr> </tbody> </table> Operator's Signature 	Test	g/210L	Time	Air Blank	0.000	12:13	Control Test	0.079	12:14	Air Blank	0.000	12:14	Control Test	0.078	12:15	Air Blank	0.000	12:15	Control Test	0.078	12:16	Air Blank	0.000	12:16	Control Test Stats			Average	0.0783		Std Dev	0.0006		Rel Std Dev(%)	0.7370		CITY OF MIAMI POL Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001067 06/24/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>11:57</td></tr> <tr><td>Control Test</td><td>0.199</td><td>11:58</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:58</td></tr> <tr><td>Control Test</td><td>0.196</td><td>11:59</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:59</td></tr> <tr><td>Control Test</td><td>0.194</td><td>12:00</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:01</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1963</td><td></td></tr> <tr><td>Std Dev</td><td>0.0025</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.2818</td><td></td></tr> </tbody> </table> Operator's Signature 	Test	g/210L	Time	Air Blank	0.000	11:57	Control Test	0.199	11:58	Air Blank	0.000	11:58	Control Test	0.196	11:59	Air Blank	0.000	11:59	Control Test	0.194	12:00	Air Blank	0.000	12:01	Control Test Stats			Average	0.1963		Std Dev	0.0025		Rel Std Dev(%)	1.2818		CITY OF MIAMI POL Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001067 06/24/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>11:47</td></tr> <tr><td>Control Test</td><td>0.079</td><td>11:47</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:48</td></tr> <tr><td>Control Test</td><td>0.078</td><td>11:48</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:49</td></tr> <tr><td>Control Test</td><td>0.078</td><td>11:49</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:49</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0783</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7370</td><td></td></tr> </tbody> </table> Operator's Signature 	Test	g/210L	Time	Air Blank	0.000	11:47	Control Test	0.079	11:47	Air Blank	0.000	11:48	Control Test	0.078	11:48	Air Blank	0.000	11:49	Control Test	0.078	11:49	Air Blank	0.000	11:49	Control Test Stats			Average	0.0783		Std Dev	0.0006		Rel Std Dev(%)	0.7370	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	12:05																																																																																																																																																	
Control Test	0.050	12:05																																																																																																																																																	
Air Blank	0.000	12:06																																																																																																																																																	
Control Test	0.049	12:07																																																																																																																																																	
Air Blank	0.000	12:07																																																																																																																																																	
Control Test	0.049	12:08																																																																																																																																																	
Air Blank	0.000	12:08																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0493																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	1.1703																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	12:13																																																																																																																																																	
Control Test	0.079	12:14																																																																																																																																																	
Air Blank	0.000	12:14																																																																																																																																																	
Control Test	0.078	12:15																																																																																																																																																	
Air Blank	0.000	12:15																																																																																																																																																	
Control Test	0.078	12:16																																																																																																																																																	
Air Blank	0.000	12:16																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0783																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7370																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	11:57																																																																																																																																																	
Control Test	0.199	11:58																																																																																																																																																	
Air Blank	0.000	11:58																																																																																																																																																	
Control Test	0.196	11:59																																																																																																																																																	
Air Blank	0.000	11:59																																																																																																																																																	
Control Test	0.194	12:00																																																																																																																																																	
Air Blank	0.000	12:01																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1963																																																																																																																																																		
Std Dev	0.0025																																																																																																																																																		
Rel Std Dev(%)	1.2818																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	11:47																																																																																																																																																	
Control Test	0.079	11:47																																																																																																																																																	
Air Blank	0.000	11:48																																																																																																																																																	
Control Test	0.078	11:48																																																																																																																																																	
Air Blank	0.000	11:49																																																																																																																																																	
Control Test	0.078	11:49																																																																																																																																																	
Air Blank	0.000	11:49																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0783																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7370																																																																																																																																																		

Florida Department of Law Enforcement

Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: CITY OF MIRAMAR POL
Time of Inspection: 14:40

Date of Inspection: 06/24/2026

Serial Number: 80-001067
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202603A Exp: 03/03/2028	0.08g/210L Test (g/210L) Lot#:202603B Exp: 03/03/2028	0.20g/210L Test (g/210L) Lot#:202603D Exp: 03/04/2028	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG526603 Exp: 09/23/2027
0.000	0.050	0.078	0.198	0.078
0.000	0.050	0.078	0.198	0.077
0.000	0.049	0.078	0.198	0.077
0.000	0.049	0.078	0.198	0.077
0.000	0.049	0.078	0.198	0.077
0.000	0.049	0.078	0.198	0.077
0.000	0.049	0.078	0.198	0.077
0.000	0.049	0.077	0.198	0.077
0.000	0.049	0.078	0.198	0.077
0.000	0.049	0.078	0.198	0.077
Standard Deviations	0.0004	0.0003	0.0000	0.0003

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0002 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Taylor D Gutschow TAYLOR D GUTSCHOW
Signature and Printed Name

06/24/2026
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001067, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001067</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>CITY OF MIRAMAR POL</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>06/24/2026</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>14:40</u>	0.200 g/ 210 L	0.008
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

06/24/2026

Date

Taylor
Gutschow

Digitally signed by Taylor
Gutschow
Date: 2026.06.24 15:39:09
-04'00'

TAYLOR D GUTSCHOW,
Department Inspector