

Alcohol Testing Program - Instrument Processing Sheet

Agency: CLAY COUNTY SO

Instrument Serial Number: 80-001071

Date In: 7/28/2026

DI Completion Date: 7/31/2026

☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake By: <u>WKP</u> Date: <u>7/28/2026</u>		Quality Checks By: <u>SLH</u> Date: <u>7/28/2026</u>		Flow Adjustment By: _____																															
☒ Annual ☐ Dropped Off ☐ Registration ☒ Return from CMI / EE ☐ Training Instrument Visual Inspection ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/Accessories ☐ Power Cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes:		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value: <u>198</u> ☒ Flow Verification (L/s) Flow Column #: <u>ATP 102</u> 32 mm <u>0.144</u> (.139-.169) 36 mm <u>0.160</u> (.156-.190) 53 mm <u>0.234</u> (.228-.278) 103 mm <u>0.484</u> (.447-.547) ☒ Barometric Pressure Check Gauge ID #: <u>28427</u> Gauge: <u>1006</u> Instrument: <u>1008</u> ☒ Stability Checks		Flow Column #: _____ ☐ 5L/min – 17mm ☐ 15L/min – 53mm ☐ 30L/min – 103mm ☐ R-Value: _____ ☐ Post Adjustment Verification (L/S) Flow Column #: _____ 32 mm (.139-.169) 36 mm (.156-.190) 53 mm (.228-.278) 103 mm (.447-.547)																															
						<table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th style="width: 15%;">Simulator</th> <th style="width: 20%;">Serial #</th> <th style="width: 25%;">Lot#/Exp</th> <th style="width: 40%;">Maintenance By: _____ Date: _____</th> </tr> </thead> <tbody> <tr> <td rowspan="2">0.050</td> <td rowspan="2">MP6291</td> <td>202603A</td> <td>☐ Battery Replacement</td> </tr> <tr> <td>3/3/2028</td> <td>☐ Dry Gas Regulator Replacement</td> </tr> <tr> <td rowspan="2">0.080</td> <td rowspan="2">MP6292</td> <td>202603B</td> <td>☐ Tank Sensor Tare</td> </tr> <tr> <td>3/3/2028</td> <td>☐ Breath Tube Replacement</td> </tr> <tr> <td rowspan="2">0.200</td> <td rowspan="2">MP6293</td> <td>202603D</td> <td>☐ Other:</td> </tr> <tr> <td>3/4/2028</td> <td></td> </tr> <tr> <td rowspan="2">0.080 DGS</td> <td rowspan="2">N/A</td> <td>23825080A1</td> <td></td> </tr> <tr> <td>10/5/2027</td> <td></td> </tr> </tbody> </table>		Simulator	Serial #	Lot#/Exp	Maintenance By: _____ Date: _____	0.050	MP6291	202603A	☐ Battery Replacement	3/3/2028	☐ Dry Gas Regulator Replacement	0.080	MP6292	202603B	☐ Tank Sensor Tare	3/3/2028	☐ Breath Tube Replacement	0.200	MP6293	202603D	☐ Other:	3/4/2028		0.080 DGS	N/A	23825080A1		10/5/2027	
						Simulator	Serial #	Lot#/Exp	Maintenance By: _____ Date: _____																										
						0.050	MP6291	202603A	☐ Battery Replacement																										
								3/3/2028	☐ Dry Gas Regulator Replacement																										
0.080	MP6292	202603B	☐ Tank Sensor Tare																																
		3/3/2028	☐ Breath Tube Replacement																																
0.200	MP6293	202603D	☐ Other:																																
		3/4/2028																																	
0.080 DGS	N/A	23825080A1																																	
		10/5/2027																																	
Optical Bench Adjustment By: <u>SLH</u>		Department Inspection By: <u>SLH</u>																																	
Barometric Pressure Gauge: <u>1007</u> ID#: <u>34419</u>		Barometric Pressure ID#: <u>28427</u>																																	
Simulator	Serial #	Lot #	Expiration	Gauge: <u>1010</u> Instrument: <u>1012</u>	Mouth Alcohol Solution Lot #: <u>2025-D</u> Exp: <u>09/25/2027</u>																														
0.000	MP6294	N/A	N/A		Acetone Stock Solution Lot #: <u>2025-B</u> Exp: <u>09/22/2027</u>																														
0.040	MP6295	25090	3/11/2027																																
0.100	MP6296	24390	10/29/2026	Simulator	Serial Number																														
0.200	MP6297	25020	1/14/2027	0.000	MP5086																														
0.300	MP6298	25150	5/6/2027	Interferent	MP5087																														
0.080 DGS	N/A	AG526603	09/23/2027	0.050	MP5088																														
☒ Post Optical Bench Adjustment Stability Checks				0.080	MP5089																														
Simulator	Serial #	Lot #	Expiration	0.200	MP5090																														
0.050	MP6291	202603A	03/03/2028	Attachments ☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☒ Optical Bench Adjustment ☒ Post-Stability Checks ☐ Flow Adjustment ☐ Form 40 ☐ Other:																															
0.080	MP6292	202603B	03/03/2028																																
0.200	MP6293	202603D	03/04/2028																																
0.080 DGS	N/A	23825080A1	10/5/2027																																
Gauge ID #: <u>28427</u>																																			
Gauge: <u>1007</u> Instrument: <u>1008</u>																																			
Notes/Suggested Service:				☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC																															
				☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use																															
				☒ Conduct an Agency Inspection Before Evidentiary Use																															
				Digitally signed by Wen-Chi Pierson Date: 2026.08.11 14:36:50 -04'00' Kaitlyn Spearin Digitally signed by Kaitlyn Spearin Date: 2026.08.12 07:57:53 -04'00'																															
Tech Review		Admin Review																																	

Stability Checks

80-00 1071

0.050 g/210L 0.047 to 0.053 g/210L	0.080 g/210L 0.077 to 0.083 g/210L	0.200 g/210L 0.194 to 0.206 g/210L	DGS 0.080 g/210L 0.077 to 0.083 g/210L ≤0.003 g/210L of Wet																																																																																																																																																
CLAY COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001071 07/28/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>14:36</td></tr> <tr><td>Control Test</td><td>0.049</td><td>14:37</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:38</td></tr> <tr><td>Control Test</td><td>0.050</td><td>14:38</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:39</td></tr> <tr><td>Control Test</td><td>0.049</td><td>14:39</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:40</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0493</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.1703</td><td></td></tr> </tbody> </table> Operator's Signature: <i>Shigle</i>	Test	g/210L	Time	Air Blank	0.000	14:36	Control Test	0.049	14:37	Air Blank	0.000	14:38	Control Test	0.050	14:38	Air Blank	0.000	14:39	Control Test	0.049	14:39	Air Blank	0.000	14:40	Control Test Stats			Average	0.0493		Std Dev	0.0006		Rel Std Dev(%)	1.1703		CLAY COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001071 07/28/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>14:41</td></tr> <tr><td>Control Test</td><td>0.079</td><td>14:42</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:43</td></tr> <tr><td>Control Test</td><td>0.078</td><td>14:43</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:44</td></tr> <tr><td>Control Test</td><td>0.078</td><td>14:44</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:45</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0783</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7370</td><td></td></tr> </tbody> </table> Operator's Signature: <i>Shigle</i>	Test	g/210L	Time	Air Blank	0.000	14:41	Control Test	0.079	14:42	Air Blank	0.000	14:43	Control Test	0.078	14:43	Air Blank	0.000	14:44	Control Test	0.078	14:44	Air Blank	0.000	14:45	Control Test Stats			Average	0.0783		Std Dev	0.0006		Rel Std Dev(%)	0.7370		CLAY COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001071 07/28/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>14:48</td></tr> <tr><td>Control Test</td><td>0.200</td><td>14:49</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:50</td></tr> <tr><td>Control Test</td><td>0.199</td><td>14:50</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:51</td></tr> <tr><td>Control Test</td><td>0.199</td><td>14:51</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:52</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1993</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.2896</td><td></td></tr> </tbody> </table> Operator's Signature: <i>Shigle</i>	Test	g/210L	Time	Air Blank	0.000	14:48	Control Test	0.200	14:49	Air Blank	0.000	14:50	Control Test	0.199	14:50	Air Blank	0.000	14:51	Control Test	0.199	14:51	Air Blank	0.000	14:52	Control Test Stats			Average	0.1993		Std Dev	0.0006		Rel Std Dev(%)	0.2896		CLAY COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001071 07/28/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>14:54</td></tr> <tr><td>Control Test</td><td>0.082</td><td>14:54</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:55</td></tr> <tr><td>Control Test</td><td>0.082</td><td>14:55</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:56</td></tr> <tr><td>Control Test</td><td>0.082</td><td>14:56</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:56</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0820</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> Operator's Signature: <i>Shigle</i>	Test	g/210L	Time	Air Blank	0.000	14:54	Control Test	0.082	14:54	Air Blank	0.000	14:55	Control Test	0.082	14:55	Air Blank	0.000	14:56	Control Test	0.082	14:56	Air Blank	0.000	14:56	Control Test Stats			Average	0.0820		Std Dev	0.0000		Rel Std Dev(%)	0.0000	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	14:36																																																																																																																																																	
Control Test	0.049	14:37																																																																																																																																																	
Air Blank	0.000	14:38																																																																																																																																																	
Control Test	0.050	14:38																																																																																																																																																	
Air Blank	0.000	14:39																																																																																																																																																	
Control Test	0.049	14:39																																																																																																																																																	
Air Blank	0.000	14:40																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0493																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	1.1703																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	14:41																																																																																																																																																	
Control Test	0.079	14:42																																																																																																																																																	
Air Blank	0.000	14:43																																																																																																																																																	
Control Test	0.078	14:43																																																																																																																																																	
Air Blank	0.000	14:44																																																																																																																																																	
Control Test	0.078	14:44																																																																																																																																																	
Air Blank	0.000	14:45																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0783																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7370																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	14:48																																																																																																																																																	
Control Test	0.200	14:49																																																																																																																																																	
Air Blank	0.000	14:50																																																																																																																																																	
Control Test	0.199	14:50																																																																																																																																																	
Air Blank	0.000	14:51																																																																																																																																																	
Control Test	0.199	14:51																																																																																																																																																	
Air Blank	0.000	14:52																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1993																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.2896																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	14:54																																																																																																																																																	
Control Test	0.082	14:54																																																																																																																																																	
Air Blank	0.000	14:55																																																																																																																																																	
Control Test	0.082	14:55																																																																																																																																																	
Air Blank	0.000	14:56																																																																																																																																																	
Control Test	0.082	14:56																																																																																																																																																	
Air Blank	0.000	14:56																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0820																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		

need an optical bench
adjustment set

Verified instrument was
able to read gas pressure. ☒

CLAY COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000
07/28/2026

SN 80-001071 ✓
15:14:20

Sub

Auto Calibration

pg 1 of 2

<<<<< 3um >>>>> <<<<< 9um >>>>>

Solution = 0.000 g/210L or 0.0000 mg/l, Samples = 4, Discarded = 1

Sample	% Abs	(% Abs Ref)	% Abs	(% Abs Ref)
Sample #1	0.0890	(-0.0180)	0.1930	(-0.0260)
Sample #2	0.0660	(0.0370)	0.2450	(-0.0180)
Sample #3	0.0640	(0.0390)	0.2340	(-0.0230)
Sample #4	0.0450	(0.0680)	0.1980	(0.0020)
Avg % Abs	0.0583	(0.0480)	0.2257	(-0.0130)
STD DEV	0.0116	(0.0173)	0.0246	(0.0132)
REL STD DEV	19.869	(36.144)	10.894	(101.760)

Solution = 0.040 g/210L or 0.1905 mg/l, Samples = 4, Discarded = 1

Sample	% Abs	(% Abs Ref)	% Abs	(% Abs Ref)
Sample #1	0.7510	(-0.0160)	1.5340	(-0.0290)
Sample #2	0.7440	(-0.0090)	1.4980	(-0.0010)
Sample #3	0.7660	(0.0010)	1.5170	(-0.0150)
Sample #4	0.7420	(0.0190)	1.5170	(-0.0150)
Avg % Abs	0.7507	(0.0037)	1.5107	(-0.0103)
STD DEV	0.0133	(0.0142)	0.0110	(0.0081)
REL STD DEV	1.774	(386.978)	0.726	(78.222)

Solution = 0.100 g/210L or 0.4762 mg/l, Samples = 4, Discarded = 1

Sample	% Abs	(% Abs Ref)	% Abs	(% Abs Ref)
Sample #1	1.7540	(-0.0100)	3.4600	(-0.0050)
Sample #2	1.7550	(0.0130)	3.4450	(0.0040)
Sample #3	1.7700	(0.0140)	3.4460	(0.0120)
Sample #4	1.7880	(-0.0070)	3.4740	(0.0000)
Avg % Abs	1.7710	(0.0067)	3.4550	(0.0053)
STD DEV	0.0165	(0.0118)	0.0165	(0.0061)
REL STD DEV	0.933	(177.694)	0.476	(114.564)

Solution = 0.200 g/210L or 0.9524 mg/l, Samples = 4, Discarded = 1

Sample	% Abs	(% Abs Ref)	% Abs	(% Abs Ref)
Sample #1	3.4440	(-0.0040)	6.5820	(-0.0030)
Sample #2	3.4240	(0.0320)	6.5710	(0.0190)
Sample #3	3.4290	(0.0080)	6.5440	(0.0140)
Sample #4	3.4360	(0.0360)	6.5540	(0.0200)
Avg % Abs	3.4297	(0.0253)	6.5563	(0.0177)
STD DEV	0.0060	(0.0151)	0.0137	(0.0032)
REL STD DEV	0.176	(59.778)	0.208	(18.196)

Solution = 0.300 g/210L or 1.4286 mg/l, Samples = 4, Discarded = 1

Sample	% Abs	(% Abs Ref)	% Abs	(% Abs Ref)
Sample #1	5.0460	(-0.0220)	9.5290	(-0.0140)
Sample #2	5.0370	(-0.0260)	9.5120	(-0.0130)
Sample #3	5.0230	(0.0040)	9.4960	(0.0130)
Sample #4	5.0260	(0.0060)	9.4930	(0.0030)
Avg % Abs	5.0287	(-0.0053)	9.5003	(0.0010)
STD DEV	0.0074	(0.0179)	0.0102	(0.0131)
REL STD DEV	0.147	(336.108)	0.108	(1311.488)

CLAY COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001071
07/28/2026 15:14:20

Auto Calibration

pg 2 of 2

<<<< 3um >>>>

Zero Order Coef -156.62
First Order Coef 2722.11
Second Order Coef 29.73

Act (g/210L)	Fit (g/210L)	Residual (g/210L)
0.000	0.000	-0.0000
0.040	0.040	0.0000
0.100	0.100	0.0001
0.200	0.200	-0.0001
0.300	0.300	0.0000

<<<< 9um >>>>

Zero Order Coef -303.48
First Order Coef 1426.42
Second Order Coef 11.42

Act (g/210L)	Fit (g/210L)	Residual (g/210L)
0.000	0.000	-0.0004
0.040	0.039	0.0006
0.100	0.100	0.0000
0.200	0.200	-0.0003
0.300	0.300	0.0001

<<<< 3um >>>> <<<< 9um >>>>

Solution = 0.080 g/210L or 0.3810 mg/l, Samples = 4, Discarded = 1
Sample
Sample #1 3320.00 3178.00
Sample #2 3320.00 3185.00
Sample #3 3250.00 3171.00
Sample #4 3294.00 3210.00
Avg 3288.0000 3188.6667
STD DEV 35.3836 19.7569
REL STD DEV 1.076 0.620
H2O adjust (mg/l*10k) 521 621

Barometric Pressure = 1006

*****CALIBRATION SUCCESSFUL*****

✓
SW

Post Optical Bench Adjustment Stability Checks

80-001071
Sut

0.050 g/210L 0.047 to 0.053 g/210L	0.080 g/210L 0.077 to 0.083 g/210L	0.200 g/210L 0.194 to 0.206 g/210L	DGS 0.080 g/210L 0.077 to 0.083 g/210L ≤0.003 g/210L of Wet																																																																																																												
CLAY COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001071 07/28/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>16:07</td></tr> <tr><td>Control Test</td><td>0.049</td><td>16:07</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>16:08</td></tr> <tr><td>Control Test</td><td>0.050</td><td>16:08</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>16:09</td></tr> <tr><td>Control Test</td><td>0.050</td><td>16:10</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>16:10</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0497</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.1625</td><td></td></tr> </tbody> </table> <i>Miejske</i> Operator's Signature	Test	g/210L	Time	Air Blank	0.000	16:07	Control Test	0.049	16:07	Air Blank	0.000	16:08	Control Test	0.050	16:08	Air Blank	0.000	16:09	Control Test	0.050	16:10	Air Blank	0.000	16:10	Control Test Stats			Average	0.0497		Std Dev	0.0006		Rel Std Dev(%)	1.1625		see next page Sut	CLAY COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001071 07/28/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>16:02</td></tr> <tr><td>Control Test</td><td>0.198</td><td>16:03</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>16:03</td></tr> <tr><td>Control Test</td><td>0.198</td><td>16:04</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>16:04</td></tr> <tr><td>Control Test</td><td>0.199</td><td>16:05</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>16:05</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1983</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.2911</td><td></td></tr> </tbody> </table> <i>Miejske</i> Operator's Signature	Test	g/210L	Time	Air Blank	0.000	16:02	Control Test	0.198	16:03	Air Blank	0.000	16:03	Control Test	0.198	16:04	Air Blank	0.000	16:04	Control Test	0.199	16:05	Air Blank	0.000	16:05	Control Test Stats			Average	0.1983		Std Dev	0.0006		Rel Std Dev(%)	0.2911		CLAY COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001071 07/28/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>16:14</td></tr> <tr><td>Control Test</td><td>0.080</td><td>16:14</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>16:15</td></tr> <tr><td>Control Test</td><td>0.080</td><td>16:15</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>16:16</td></tr> <tr><td>Control Test</td><td>0.081</td><td>16:16</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>16:16</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0803</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7187</td><td></td></tr> </tbody> </table> <i>Miejske</i> Operator's Signature	Test	g/210L	Time	Air Blank	0.000	16:14	Control Test	0.080	16:14	Air Blank	0.000	16:15	Control Test	0.080	16:15	Air Blank	0.000	16:16	Control Test	0.081	16:16	Air Blank	0.000	16:16	Control Test Stats			Average	0.0803		Std Dev	0.0006		Rel Std Dev(%)	0.7187	
Test	g/210L	Time																																																																																																													
Air Blank	0.000	16:07																																																																																																													
Control Test	0.049	16:07																																																																																																													
Air Blank	0.000	16:08																																																																																																													
Control Test	0.050	16:08																																																																																																													
Air Blank	0.000	16:09																																																																																																													
Control Test	0.050	16:10																																																																																																													
Air Blank	0.000	16:10																																																																																																													
Control Test Stats																																																																																																															
Average	0.0497																																																																																																														
Std Dev	0.0006																																																																																																														
Rel Std Dev(%)	1.1625																																																																																																														
Test	g/210L	Time																																																																																																													
Air Blank	0.000	16:02																																																																																																													
Control Test	0.198	16:03																																																																																																													
Air Blank	0.000	16:03																																																																																																													
Control Test	0.198	16:04																																																																																																													
Air Blank	0.000	16:04																																																																																																													
Control Test	0.199	16:05																																																																																																													
Air Blank	0.000	16:05																																																																																																													
Control Test Stats																																																																																																															
Average	0.1983																																																																																																														
Std Dev	0.0006																																																																																																														
Rel Std Dev(%)	0.2911																																																																																																														
Test	g/210L	Time																																																																																																													
Air Blank	0.000	16:14																																																																																																													
Control Test	0.080	16:14																																																																																																													
Air Blank	0.000	16:15																																																																																																													
Control Test	0.080	16:15																																																																																																													
Air Blank	0.000	16:16																																																																																																													
Control Test	0.081	16:16																																																																																																													
Air Blank	0.000	16:16																																																																																																													
Control Test Stats																																																																																																															
Average	0.0803																																																																																																														
Std Dev	0.0006																																																																																																														
Rel Std Dev(%)	0.7187																																																																																																														

CLAY COUNTY SO

Intoxilyzer - Alcohol Analyzer

Model 8000

SN 80-001071 ✓

07/28/2026

Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	15:57
Control Test	0.080	15:58
Air Blank	0.000	15:58
Control Test	0.079	15:59
Air Blank	0.000	16:00
Control Test	0.079	16:00
Air Blank	0.000	16:01
Control Test Stats		
Average	0.0793	
Std Dev	0.0006	
Rel Std Dev(%)	0.7277	


Operator's Signature

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: CLAY COUNTY SO
Time of Inspection: 15:52

Date of Inspection: 07/31/2026

Serial Number: 80-001071
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202603A Exp: 03/03/2028	0.08g/210L Test (g/210L) Lot#:202603B Exp: 03/03/2028	0.20g/210L Test (g/210L) Lot#:202603D Exp: 03/04/2028	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:23825080A1 Exp: 10/05/2027
0.000	0.049	0.078	0.199	0.080
0.000	0.049	0.079	0.199	0.080
0.000	0.049	0.079	0.199	0.080
0.000	0.049	0.078	0.198	0.080
0.000	0.049	0.078	0.198	0.080
0.000	0.049	0.079	0.199	0.080
0.000	0.049	0.078	0.199	0.079
0.000	0.049	0.078	0.199	0.079
0.000	0.050	0.078	0.199	0.079
0.000	0.049	0.079	0.198	0.079


Standard Deviations	0.0003	0.0005	0.0004	0.0005
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:


The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



LEANDRA HIGGINBOTHAM
Signature and Printed Name

07/31/2026
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001071, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001071</u>	UNCERTAINTY * $\pm$
Owning Agency:	<u>CLAY COUNTY SO</u>	0.050 g/210 L 0.004
Calibration Date:	<u>07/31/2026</u>	0.080 g/210 L 0.004
Calibration Time:	<u>15:52</u>	0.200 g/210 L 0.008
		0.080 g/210 L Dry Gas Control 0.005

All results are reported in g/210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration. *Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/IEC 17025 standards. This document shall not be reproduced except in full without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

07/31/2026
Date

LEANDRA MIGINBOTHAM,
Department Inspector