

Alcohol Testing Program - Instrument Processing Sheet

Agency: St. Johns County SO

Instrument Serial Number: 80-007515

Date In: 7/29/2026

DI Completion Date: 8/4/2026

☐ Ship ☒ P/U ☐ H/D ☐ CMI ☐ EE

Intake By: <u>SLH</u> Date: <u>7/30/2026</u>		Quality Checks By: <u>SLH</u> Date: <u>7/30/2026</u>		Flow Adjustment By: _____																																	
☒ Annual ☒ Dropped Off ☐ Registration ☐ Return from CMI / EE ☐ Training Instrument Visual Inspection ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/Accessories ☐ Power Cord ☐ Printer Cable ☐ Static Bag ☐ 12V DC Cable Notes:		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value: <u>0.234-234</u> ☒ Flow Verification (L/s) Flow Column #: <u>ATP 102</u> 32 mm <u>0.164</u> (.139-.169) 36 mm <u>0.179</u> (.156-.190) 53 mm <u>0.253</u> (.228-.278) 103 mm <u>0.519</u> (.447-.547) ☒ Barometric Pressure Check Gauge ID #: <u>28427</u> Gauge: <u>1008</u> Instrument: <u>1008</u> ☒ Stability Checks		Flow Column #: _____ ☐ 5L/min – 17mm ☐ 15L/min – 53mm ☐ 30L/min – 103mm ☐ R-Value: _____ ☐ Post Adjustment Verification (L/S) Flow Column #: _____ 32 mm _____ (.139-.169) 36 mm _____ (.156-.190) 53 mm _____ (.228-.278) 103 mm _____ (.447-.547)																																	
		<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot#/Exp</th></tr></thead><tbody><tr><td rowspan="2">0.050</td><td rowspan="2">MP5088</td><td>202603A</td></tr><tr><td>03/03/2028</td></tr><tr><td rowspan="2">0.080</td><td rowspan="2">MP5089</td><td>202603B</td></tr><tr><td>03/03/2028</td></tr><tr><td rowspan="2">0.200</td><td rowspan="2">MP5090</td><td>202603D</td></tr><tr><td>03/04/2028</td></tr><tr><td rowspan="2">0.080 DGS</td><td rowspan="2">N/A</td><td>23825080A1</td></tr><tr><td>10/5/2027</td></tr></tbody></table>		Simulator	Serial #	Lot#/Exp	0.050	MP5088	202603A	03/03/2028	0.080	MP5089	202603B	03/03/2028	0.200	MP5090	202603D	03/04/2028	0.080 DGS	N/A	23825080A1	10/5/2027	Maintenance By: <u>SLH</u> Date: <u>7/30/2026</u> ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Tank Sensor Tare ☐ Breath Tube Replacement ☒ Other: Added internal printer paper.														
		Simulator	Serial #	Lot#/Exp																																	
		0.050	MP5088	202603A																																	
				03/03/2028																																	
0.080	MP5089	202603B																																			
		03/03/2028																																			
0.200	MP5090	202603D																																			
		03/04/2028																																			
0.080 DGS	N/A	23825080A1																																			
		10/5/2027																																			
Optical Bench Adjustment By: <u>SLH</u> Department Inspection By: <u>SLH</u>																																					
Barometric Pressure Gauge: <u>1008</u> ID#: <u>34419</u> Barometric Pressure ID#: <u>28427</u>																																					
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td>MP6294</td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td>MP6295</td><td>25090</td><td>03/11/2027</td></tr><tr><td>0.100</td><td>MP6296</td><td>24390</td><td>10/29/2026</td></tr><tr><td>0.200</td><td>MP6297</td><td>25020</td><td>01/14/2027</td></tr><tr><td>0.300</td><td>MP6298</td><td>25150</td><td>05/06/2027</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG526603</td><td>09/23/2027</td></tr></tbody></table>		Simulator	Serial #	Lot #	Expiration	0.000	MP6294	N/A	N/A	0.040	MP6295	25090	03/11/2027	0.100	MP6296	24390	10/29/2026	0.200	MP6297	25020	01/14/2027	0.300	MP6298	25150	05/06/2027	0.080 DGS	N/A	AG526603	09/23/2027	<table border="1"><thead><tr><th>Gauge: <u>1009</u> Instrument: <u>1011</u></th><th></th></tr></thead><tbody><tr><td>Mouth Alcohol Solution Lot #: <u>2025-D</u></td><td>Exp: <u>09/25/2027</u></td></tr><tr><td>Acetone Stock Solution Lot #: <u>2025-B</u></td><td>Exp: <u>09/22/2027</u></td></tr></tbody></table>		Gauge: <u>1009</u> Instrument: <u>1011</u>		Mouth Alcohol Solution Lot #: <u>2025-D</u>	Exp: <u>09/25/2027</u>	Acetone Stock Solution Lot #: <u>2025-B</u>	Exp: <u>09/22/2027</u>
Simulator	Serial #	Lot #	Expiration																																		
0.000	MP6294	N/A	N/A																																		
0.040	MP6295	25090	03/11/2027																																		
0.100	MP6296	24390	10/29/2026																																		
0.200	MP6297	25020	01/14/2027																																		
0.300	MP6298	25150	05/06/2027																																		
0.080 DGS	N/A	AG526603	09/23/2027																																		
Gauge: <u>1009</u> Instrument: <u>1011</u>																																					
Mouth Alcohol Solution Lot #: <u>2025-D</u>	Exp: <u>09/25/2027</u>																																				
Acetone Stock Solution Lot #: <u>2025-B</u>	Exp: <u>09/22/2027</u>																																				
☒ Post Optical Bench Adjustment Stability Checks		<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td>MP5088</td><td>202603A</td><td>03/03/2028</td></tr><tr><td>0.080</td><td>MP5089</td><td>202603B</td><td>03/03/2028</td></tr><tr><td>0.200</td><td>MP5090</td><td>202603D</td><td>03/04/2028</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>23825080A1</td><td>10/5/2027</td></tr></tbody></table>		Simulator	Serial #	Lot #	Expiration	0.050	MP5088	202603A	03/03/2028	0.080	MP5089	202603B	03/03/2028	0.200	MP5090	202603D	03/04/2028	0.080 DGS	N/A	23825080A1	10/5/2027	<table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>MP5086</td></tr><tr><td>Interferent</td><td>MP5087</td></tr><tr><td>0.050</td><td>MP5088</td></tr><tr><td>0.080</td><td>MP5089</td></tr><tr><td>0.200</td><td>MP5090</td></tr></tbody></table>		Simulator	Serial Number	0.000	MP5086	Interferent	MP5087	0.050	MP5088	0.080	MP5089	0.200	MP5090
Simulator	Serial #	Lot #	Expiration																																		
0.050	MP5088	202603A	03/03/2028																																		
0.080	MP5089	202603B	03/03/2028																																		
0.200	MP5090	202603D	03/04/2028																																		
0.080 DGS	N/A	23825080A1	10/5/2027																																		
Simulator	Serial Number																																				
0.000	MP5086																																				
Interferent	MP5087																																				
0.050	MP5088																																				
0.080	MP5089																																				
0.200	MP5090																																				
Gauge ID #: <u>28427</u> Gauge: <u>1009</u> Instrument: <u>1011</u>		Attachments ☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☒ Optical Bench Adjustment ☒ Post-Stability Checks ☐ Flow Adjustment ☐ Form 40 ☒ Other: Calculation form																																			
Notes/Suggested Service: Admin Review- Corrected R value to a whole number. SLH 8/6/2026		☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use																																			
		Wen-Chi Pierson Digitally signed by Wen-Chi Pierson Date: 2026.08.06 08:11:31 -04'00'																																			
		Kaitlyn Spearin Digitally signed by Kaitlyn Spearin Date: 2026.08.06 10:44:26 -04'00'																																			
Tech Review		Admin Review																																			

Stability Checks

80-00 **7515**

0.050 g/210L 0.047 to 0.053 g/210L	0.080 g/210L 0.077 to 0.083 g/210L	0.200 g/210L 0.194 to 0.206 g/210L	DGS 0.080 g/210L 0.077 to 0.083 g/210L 50.003 g/210L of Wet																																																																																																																																																
ST. JOHNS COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-007515 07/30/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>13:47</td></tr> <tr><td>Control Test</td><td>0.048</td><td>13:47</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:48</td></tr> <tr><td>Control Test</td><td>0.048</td><td>13:49</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:49</td></tr> <tr><td>Control Test</td><td>0.049</td><td>13:50</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:50</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0483</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.1945</td><td></td></tr> </tbody> </table> <i>Miguel</i> Operator's Signature	Test	g/210L	Time	Air Blank	0.000	13:47	Control Test	0.048	13:47	Air Blank	0.000	13:48	Control Test	0.048	13:49	Air Blank	0.000	13:49	Control Test	0.049	13:50	Air Blank	0.000	13:50	Control Test Stats			Average	0.0483		Std Dev	0.0006		Rel Std Dev(%)	1.1945		ST. JOHNS COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-007515 07/30/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>14:00</td></tr> <tr><td>Control Test</td><td>0.076</td><td>14:01</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:01</td></tr> <tr><td>Control Test</td><td>0.078</td><td>14:02</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:02</td></tr> <tr><td>Control Test</td><td>0.078</td><td>14:03</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:04</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0773</td><td></td></tr> <tr><td>Std Dev</td><td>0.0012</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.4932</td><td></td></tr> </tbody> </table> <i>Miguel</i> Operator's Signature	Test	g/210L	Time	Air Blank	0.000	14:00	Control Test	0.076	14:01	Air Blank	0.000	14:01	Control Test	0.078	14:02	Air Blank	0.000	14:02	Control Test	0.078	14:03	Air Blank	0.000	14:04	Control Test Stats			Average	0.0773		Std Dev	0.0012		Rel Std Dev(%)	1.4932		ST. JOHNS COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-007515 07/30/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>11:15</td></tr> <tr><td>Control Test</td><td>0.197</td><td>11:16</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:17</td></tr> <tr><td>Control Test</td><td>0.196</td><td>11:17</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:18</td></tr> <tr><td>Control Test</td><td>0.196</td><td>11:19</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:19</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1963</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.2941</td><td></td></tr> </tbody> </table> <i>Miguel</i> Operator's Signature	Test	g/210L	Time	Air Blank	0.000	11:15	Control Test	0.197	11:16	Air Blank	0.000	11:17	Control Test	0.196	11:17	Air Blank	0.000	11:18	Control Test	0.196	11:19	Air Blank	0.000	11:19	Control Test Stats			Average	0.1963		Std Dev	0.0006		Rel Std Dev(%)	0.2941		DGS ST. JOHNS COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-007515 07/30/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>12:39</td></tr> <tr><td>Control Test</td><td>0.080</td><td>12:39</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:40</td></tr> <tr><td>Control Test</td><td>0.080</td><td>12:40</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:41</td></tr> <tr><td>Control Test</td><td>0.080</td><td>12:41</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:42</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0800</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> <i>Miguel</i> Operator's Signature	Test	g/210L	Time	Air Blank	0.000	12:39	Control Test	0.080	12:39	Air Blank	0.000	12:40	Control Test	0.080	12:40	Air Blank	0.000	12:41	Control Test	0.080	12:41	Air Blank	0.000	12:42	Control Test Stats			Average	0.0800		Std Dev	0.0000		Rel Std Dev(%)	0.0000	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	13:47																																																																																																																																																	
Control Test	0.048	13:47																																																																																																																																																	
Air Blank	0.000	13:48																																																																																																																																																	
Control Test	0.048	13:49																																																																																																																																																	
Air Blank	0.000	13:49																																																																																																																																																	
Control Test	0.049	13:50																																																																																																																																																	
Air Blank	0.000	13:50																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0483																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	1.1945																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	14:00																																																																																																																																																	
Control Test	0.076	14:01																																																																																																																																																	
Air Blank	0.000	14:01																																																																																																																																																	
Control Test	0.078	14:02																																																																																																																																																	
Air Blank	0.000	14:02																																																																																																																																																	
Control Test	0.078	14:03																																																																																																																																																	
Air Blank	0.000	14:04																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0773																																																																																																																																																		
Std Dev	0.0012																																																																																																																																																		
Rel Std Dev(%)	1.4932																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	11:15																																																																																																																																																	
Control Test	0.197	11:16																																																																																																																																																	
Air Blank	0.000	11:17																																																																																																																																																	
Control Test	0.196	11:17																																																																																																																																																	
Air Blank	0.000	11:18																																																																																																																																																	
Control Test	0.196	11:19																																																																																																																																																	
Air Blank	0.000	11:19																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1963																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.2941																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	12:39																																																																																																																																																	
Control Test	0.080	12:39																																																																																																																																																	
Air Blank	0.000	12:40																																																																																																																																																	
Control Test	0.080	12:40																																																																																																																																																	
Air Blank	0.000	12:41																																																																																																																																																	
Control Test	0.080	12:41																																																																																																																																																	
Air Blank	0.000	12:42																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0800																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		

requires optical bench
adjustment. sus

Verified instrument was
able to read gas pressure. ☒

ST JOHNS COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007515
08/03/2026 08:54:58

SUB

Auto Calibration

pg 1 of 2

<<<< 3um >>>> <<<< 9um >>>>

Solution = 0.000 g/210L or 0.0000 mg/l, Samples = 4, Discarded = 1

Sample	% Abs	(% Abs Ref)	% Abs	(% Abs Ref)
Sample #1	0.0100	(-0.0030)	0.0680	(-0.0180)
Sample #2	0.0610	(-0.0140)	0.0940	(-0.0180)
Sample #3	0.0570	(-0.0080)	0.0810	(-0.0130)
Sample #4	0.0650	(-0.0020)	0.0990	(-0.0290)
Avg % Abs	0.0610	(-0.0080)	0.0913	(-0.0200)
STD DEV	0.0040	(0.0060)	0.0093	(0.0082)
REL STD DEV	6.557	(75.000)	10.173	(40.927)

Solution = 0.040 g/210L or 0.1905 mg/l, Samples = 4, Discarded = 1

Sample	% Abs	(% Abs Ref)	% Abs	(% Abs Ref)
Sample #1	0.7460	(-0.0040)	1.5180	(-0.0030)
Sample #2	0.7600	(-0.0100)	1.5260	(-0.0110)
Sample #3	0.7630	(0.0000)	1.5010	(-0.0070)
Sample #4	0.8130	(-0.0150)	1.5250	(-0.0150)
Avg % Abs	0.7787	(-0.0083)	1.5173	(-0.0110)
STD DEV	0.0298	(0.0076)	0.0142	(0.0040)
REL STD DEV	3.823	(91.652)	0.933	(36.364)

Solution = 0.100 g/210L or 0.4762 mg/l, Samples = 4, Discarded = 1

Sample	% Abs	(% Abs Ref)	% Abs	(% Abs Ref)
Sample #1	1.8560	(-0.0050)	3.6150	(-0.0170)
Sample #2	1.8330	(0.0080)	3.6200	(-0.0220)
Sample #3	1.8350	(0.0160)	3.6210	(-0.0230)
Sample #4	1.8790	(0.0120)	3.6140	(-0.0220)
Avg % Abs	1.8490	(0.0120)	3.6183	(-0.0223)
STD DEV	0.0260	(0.0040)	0.0038	(0.0006)
REL STD DEV	1.406	(33.333)	0.105	(2.585)

Solution = 0.200 g/210L or 0.9524 mg/l, Samples = 4, Discarded = 1

Sample	% Abs	(% Abs Ref)	% Abs	(% Abs Ref)
Sample #1	3.5950	(-0.0110)	6.9430	(-0.0030)
Sample #2	3.6120	(-0.0070)	6.9710	(-0.0130)
Sample #3	3.6340	(-0.0180)	6.9790	(-0.0290)
Sample #4	3.6070	(0.0110)	6.9530	(-0.0080)
Avg % Abs	3.6177	(-0.0047)	6.9677	(-0.0167)
STD DEV	0.0144	(0.0146)	0.0133	(0.0110)
REL STD DEV	0.397	(313.717)	0.191	(65.818)

Solution = 0.300 g/210L or 1.4286 mg/l, Samples = 4, Discarded = 1

Sample	% Abs	(% Abs Ref)	% Abs	(% Abs Ref)
Sample #1	5.3130	(-0.0080)	10.0870	(0.0120)
Sample #2	5.3400	(-0.0030)	10.1030	(0.0220)
Sample #3	5.3250	(0.0200)	10.1000	(0.0270)
Sample #4	5.3010	(0.0360)	10.1040	(0.0130)
Avg % Abs	5.3220	(0.0177)	10.1023	(0.0207)
STD DEV	0.0197	(0.0196)	0.0021	(0.0071)
REL STD DEV	0.370	(110.968)	0.021	(34.329)

ST JOHNS COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007515
08/03/2026 08:54:58

Auto Calibration

pg 2 of 2

<<<< 3um >>>>			<<<< 9um >>>>		
-----			-----		
Zero Order Coef	-148.76			-106.33	
First Order Coef	2613.00			1297.47	
Second Order Coef	18.37			12.51	
-----			-----		
Act	Fit	Residual	Act	Fit	Residual
(g/210L)	(g/210L)	(g/210L)	(g/210L)	(g/210L)	(g/210L)
0.000	0.000	-0.0002	0.000	0.000	-0.0003
0.040	0.040	0.0002	0.040	0.040	0.0003
0.100	0.100	0.0003	0.100	0.100	0.0002
0.200	0.200	-0.0004	0.200	0.200	-0.0004
0.300	0.300	0.0002	0.300	0.300	0.0002
-----			-----		

<<<< 3um >>>>		<<<< 9um >>>>	
-----		-----	
Solution = 0.080 g/210L or 0.3810 mg/l, Samples = 4, Discarded = 1			
Sample			
Sample #1	3399.00		3468.00
Sample #2	3436.00		3476.00
Sample #3	3333.00		3481.00
Sample #4	3404.00		3474.00
Avg	3391.0000		3477.0000
STD DEV	52.7162		3.6056
REL STD DEV	1.555		0.104
H2O adjust (mg/l*10k)	418		332

Barometric Pressure = 1009

*****CALIBRATION SUCCESSFUL*****

SWA

Post Optical Bench Adjustment Stability Checks 80-007515

* 0.050 g/210L 0.047 to 0.053 g/210L	0.080 g/210L 0.077 to 0.083 g/210L	* 0.200 g/210L 0.194 to 0.206 g/210L	DGS 0.080 g/210L 0.077 to 0.083 g/210L ≤0.003 g/210L of Wet																																																																																																																																																
ST JOHNS COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-007515 08/04/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>14:07</td></tr> <tr><td>Control Test</td><td>0.049</td><td>14:08</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:08</td></tr> <tr><td>Control Test</td><td>0.049</td><td>14:09</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:09</td></tr> <tr><td>Control Test</td><td>0.049</td><td>14:10</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:10</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0490</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> Operator's Signature: <i>Miguel</i>	Test	g/210L	Time	Air Blank	0.000	14:07	Control Test	0.049	14:08	Air Blank	0.000	14:08	Control Test	0.049	14:09	Air Blank	0.000	14:09	Control Test	0.049	14:10	Air Blank	0.000	14:10	Control Test Stats			Average	0.0490		Std Dev	0.0000		Rel Std Dev(%)	0.0000		ST JOHNS COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-007515 08/13/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>13:01</td></tr> <tr><td>Control Test</td><td>0.078</td><td>13:02</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:02</td></tr> <tr><td>Control Test</td><td>0.079</td><td>13:03</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:04</td></tr> <tr><td>Control Test</td><td>0.079</td><td>13:04</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:05</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0787</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7339</td><td></td></tr> </tbody> </table> Operator's Signature: <i>Miguel</i>	Test	g/210L	Time	Air Blank	0.000	13:01	Control Test	0.078	13:02	Air Blank	0.000	13:02	Control Test	0.079	13:03	Air Blank	0.000	13:04	Control Test	0.079	13:04	Air Blank	0.000	13:05	Control Test Stats			Average	0.0787		Std Dev	0.0006		Rel Std Dev(%)	0.7339		ST JOHNS COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-007515 08/04/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>14:12</td></tr> <tr><td>Control Test</td><td>0.197</td><td>14:12</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:13</td></tr> <tr><td>Control Test</td><td>0.197</td><td>14:14</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:14</td></tr> <tr><td>Control Test</td><td>0.197</td><td>14:15</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:15</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1970</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> Operator's Signature: <i>Miguel</i>	Test	g/210L	Time	Air Blank	0.000	14:12	Control Test	0.197	14:12	Air Blank	0.000	14:13	Control Test	0.197	14:14	Air Blank	0.000	14:14	Control Test	0.197	14:15	Air Blank	0.000	14:15	Control Test Stats			Average	0.1970		Std Dev	0.0000		Rel Std Dev(%)	0.0000		ST JOHNS COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-007515 08/03/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>13:08</td></tr> <tr><td>Control Test</td><td>0.079</td><td>13:08</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:09</td></tr> <tr><td>Control Test</td><td>0.079</td><td>13:09</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:10</td></tr> <tr><td>Control Test</td><td>0.079</td><td>13:10</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:11</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0790</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> Operator's Signature: <i>Miguel</i>	Test	g/210L	Time	Air Blank	0.000	13:08	Control Test	0.079	13:08	Air Blank	0.000	13:09	Control Test	0.079	13:09	Air Blank	0.000	13:10	Control Test	0.079	13:10	Air Blank	0.000	13:11	Control Test Stats			Average	0.0790		Std Dev	0.0000		Rel Std Dev(%)	0.0000	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	14:07																																																																																																																																																	
Control Test	0.049	14:08																																																																																																																																																	
Air Blank	0.000	14:08																																																																																																																																																	
Control Test	0.049	14:09																																																																																																																																																	
Air Blank	0.000	14:09																																																																																																																																																	
Control Test	0.049	14:10																																																																																																																																																	
Air Blank	0.000	14:10																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0490																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	13:01																																																																																																																																																	
Control Test	0.078	13:02																																																																																																																																																	
Air Blank	0.000	13:02																																																																																																																																																	
Control Test	0.079	13:03																																																																																																																																																	
Air Blank	0.000	13:04																																																																																																																																																	
Control Test	0.079	13:04																																																																																																																																																	
Air Blank	0.000	13:05																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0787																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7339																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	14:12																																																																																																																																																	
Control Test	0.197	14:12																																																																																																																																																	
Air Blank	0.000	14:13																																																																																																																																																	
Control Test	0.197	14:14																																																																																																																																																	
Air Blank	0.000	14:14																																																																																																																																																	
Control Test	0.197	14:15																																																																																																																																																	
Air Blank	0.000	14:15																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1970																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	13:08																																																																																																																																																	
Control Test	0.079	13:08																																																																																																																																																	
Air Blank	0.000	13:09																																																																																																																																																	
Control Test	0.079	13:09																																																																																																																																																	
Air Blank	0.000	13:10																																																																																																																																																	
Control Test	0.079	13:10																																																																																																																																																	
Air Blank	0.000	13:11																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0790																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		

* performed 8/4/2026 sub

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: ST JOHNS COUNTY SO
Time of Inspection: 17:33

Date of Inspection: 08/04/2026

Serial Number: 80-007515
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202603A Exp: 03/03/2028	0.08g/210L Test (g/210L) Lot#:202603B Exp: 03/03/2028	0.20g/210L Test (g/210L) Lot#:202603D Exp: 03/04/2028	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:23825080A1 Exp: 10/05/2027
0.000	0.049	0.078	0.197	0.080
0.000	0.049	0.078	0.197	0.080
0.000	0.050	0.078	0.197	0.080
0.000	0.049	0.078	0.197	0.080
0.000	0.049	0.078	0.197	0.079
0.000	0.049	0.078	0.197	0.079
0.000	0.050	0.078	0.197	0.080
0.000	0.050	0.078	0.197	0.080
0.000	0.050	0.078	0.197	0.079
0.000	0.050	0.078	0.197	0.079


Standard Deviations	0.0005	0.0000	0.0000	0.0005
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0002 Number of Simulators Used: 5

Remarks:

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



LEANDRA HIGGINBOTHAM

Signature and Printed Name

08/04/2026
Date

Florida Department of Law Enforcement
Alcohol Testing Program - Average Calculations for Calibration Certificate

Calculations By: SLH

Serial Number: 80-007515

Date: 8/4/2026

Solution:

0.05 g/210 L Test: 0.495 / 10 = 0.050 g/210L
(SUM) (AVERAGE)

0.08 g/210L Test: 0.780 / 10 = 0.078 g/210L
(SUM) (AVERAGE)

0.20 g/210L Test: 1.970 / 10 = 0.197 g/210L
(SUM) (AVERAGE)

0.08 g/210L Dry Gas Std Test: 0.796 / 10 = 0.080 g/210L
(SUM) (AVERAGE)

Comments:

Average = (Sum of observations) / (Total number of observations)

All results are reported in g/210L



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-007515, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-007515</u>	Average of $y \pm$ Uncertainty*
Owning Agency:	<u>ST JOHNS COUNTY SO</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>08/04/2026</u>	0.078 g/ 210 L 0.004
Calibration Time:	<u>17:33</u>	0.197 g/ 210 L 0.008
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

08/04/2026

Date

LeAndra
Higginbotham

Digitally signed by LeAndra
Higginbotham
Date: 2026.08.05 14:28:36 -04'00'

LEANDRA HIGGINBOTHAM,
Department Inspector

FDLE/ATP Form 69 August 2026

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Page 1 of 1