

Alcohol Testing Program - Instrument Processing Sheet

Agency: Florida Highway Patrol

Instrument Serial Number: 80-000937

Date In: 7/28/2026

DI Completion Date: 8/27/2026

☒ Ship

☐ P/U

☐ H/D

☐ CMI

☐ EE

Intake By: <u>TDG</u> Date: <u>8/3/2026</u>		Quality Checks By: <u>TDG</u> Date: <u>8/27/2026</u>		Flow Adjustment By: _____																																							
☒ Annual ☐ Dropped Off ☐ Registration ☐ Return from CMI / EE ☐ Training Instrument Visual Inspection ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/Accessories ☐ Power Cord ☐ Printer Cable ☐ Static Bag ☐ 12V DC Cable Notes:		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value: <u>170</u> ☒ Flow Verification (L/s) Flow Column #: <u>ATP106</u> 32 mm <u>0.148</u> (.139-.169) 36 mm <u>0.164</u> (.156-.190) 53 mm <u>0.238</u> (.228-.278) 103 mm <u>0.503</u> (.447-.547) ☒ Barometric Pressure Check Gauge ID #: <u>33363</u> Gauge: <u>1018</u> Instrument: <u>1015</u> ☒ Stability Checks		Flow Column #: _____ ☐ 5L/min – 17mm ☐ 15L/min – 53mm ☐ 30L/min – 103mm ☐ R-Value: _____ ☐ Post Adjustment Verification (L/S) Flow Column #: _____ 32 mm (.139-.169) 36 mm (.156-.190) 53 mm (.228-.278) 103 mm (.447-.547)																																							
						<table border="1" style="width:100%; border-collapse: collapse;"> <tr> <th style="width:15%;">Simulator</th> <th style="width:20%;">Serial #</th> <th style="width:20%;">Lot#/Exp</th> <th style="width:45%;">Maintenance By: _____ Date: _____</th> </tr> <tr> <td rowspan="2">0.050</td> <td rowspan="2">MP5099</td> <td>202603A</td> <td>☐ Battery Replacement</td> </tr> <tr> <td>3/3/2028</td> <td>☐ Dry Gas Regulator Replacement</td> </tr> <tr> <td rowspan="2">0.080</td> <td rowspan="2">MP5100</td> <td>202603B</td> <td>☐ Tank Sensor Tare</td> </tr> <tr> <td>3/3/2028</td> <td>☐ Breath Tube Replacement</td> </tr> <tr> <td rowspan="2">0.200</td> <td rowspan="2">MP5101</td> <td>202603D</td> <td>☐ Other:</td> </tr> <tr> <td>3/4/2028</td> <td></td> </tr> <tr> <td rowspan="2">0.080 DGS</td> <td rowspan="2">N/A</td> <td>28424080A3</td> <td></td> </tr> <tr> <td>11/5/2026</td> <td></td> </tr> </table>		Simulator	Serial #	Lot#/Exp	Maintenance By: _____ Date: _____	0.050	MP5099	202603A	☐ Battery Replacement	3/3/2028	☐ Dry Gas Regulator Replacement	0.080	MP5100	202603B	☐ Tank Sensor Tare	3/3/2028	☐ Breath Tube Replacement	0.200	MP5101	202603D	☐ Other:	3/4/2028		0.080 DGS	N/A	28424080A3		11/5/2026									
						Simulator	Serial #	Lot#/Exp	Maintenance By: _____ Date: _____																																		
						0.050	MP5099	202603A	☐ Battery Replacement																																		
								3/3/2028	☐ Dry Gas Regulator Replacement																																		
0.080	MP5100	202603B	☐ Tank Sensor Tare																																								
		3/3/2028	☐ Breath Tube Replacement																																								
0.200	MP5101	202603D	☐ Other:																																								
		3/4/2028																																									
0.080 DGS	N/A	28424080A3																																									
		11/5/2026																																									
<table border="1" style="width:100%; border-collapse: collapse;"> <tr> <th style="width:15%;">Simulator</th> <th style="width:20%;">Serial #</th> <th style="width:20%;">Lot #</th> <th style="width:45%;">Expiration</th> </tr> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </table>		Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			<table border="1" style="width:100%; border-collapse: collapse;"> <tr> <th style="width:50%;">Simulator</th> <th style="width:50%;">Serial Number</th> </tr> <tr> <td>0.000</td> <td>MP5097</td> </tr> <tr> <td>Interferent</td> <td>MP5098</td> </tr> <tr> <td>0.050</td> <td>MP5099</td> </tr> <tr> <td>0.080</td> <td>MP5100</td> </tr> <tr> <td>0.200</td> <td>MP5101</td> </tr> </table>		Simulator	Serial Number	0.000	MP5097	Interferent	MP5098	0.050	MP5099	0.080	MP5100	0.200	MP5101
Simulator	Serial #	Lot #	Expiration																																								
0.000		N/A	N/A																																								
0.040																																											
0.100																																											
0.200																																											
0.300																																											
0.080 DGS	N/A																																										
Simulator	Serial Number																																										
0.000	MP5097																																										
Interferent	MP5098																																										
0.050	MP5099																																										
0.080	MP5100																																										
0.200	MP5101																																										
<table border="1" style="width:100%; border-collapse: collapse;"> <tr> <th style="width:15%;">Simulator</th> <th style="width:20%;">Serial #</th> <th style="width:20%;">Lot #</th> <th style="width:45%;">Expiration</th> </tr> <tr> <td>0.050</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </table>		Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			<table border="1" style="width:100%; border-collapse: collapse;"> <tr> <th colspan="2" style="background-color: #d3d3d3;">Attachments</th> </tr> <tr> <td style="width:50%; vertical-align: top;"> ☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☐ Optical Bench Adjustment </td> <td style="width:50%; vertical-align: top;"> ☐ Post-Stability Checks ☐ Flow Adjustment ☒ Form 40 ☒ Other: Calculation Sheet </td> </tr> </table>		Attachments		☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☐ Optical Bench Adjustment	☐ Post-Stability Checks ☐ Flow Adjustment ☒ Form 40 ☒ Other: Calculation Sheet																
Simulator	Serial #	Lot #	Expiration																																								
0.050																																											
0.080																																											
0.200																																											
0.080 DGS	N/A																																										
Attachments																																											
☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☐ Optical Bench Adjustment	☐ Post-Stability Checks ☐ Flow Adjustment ☒ Form 40 ☒ Other: Calculation Sheet																																										
Gauge ID #: _____ Gauge: _____ Instrument: _____		Gauge: <u>1019</u> Instrument: <u>1017</u> Mouth Alcohol Solution Lot #: <u>2025-D</u> Exp: <u>9/25/2027</u> Acetone Stock Solution Lot #: <u>2025-B</u> Exp: <u>9/22/2027</u>																																									
Notes/Suggested Service:		☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC																																									
		☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use																																									
		☒ Conduct an Agency Inspection Before Evidentiary Use																																									
Digitally signed by Wen-Chi Pierson Date: 2026.08.31 11:49:58 -04'00'		Digitally signed by Kaitlyn Spearin Date: 2026.08.31 13:23:04 -04'00'																																									
Tech Review		Admin Review																																									

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FL HIGHWAY PATROL
Time of Inspection: 10:50

Date of Inspection: 08/27/2026

Serial Number: 80-000937
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:


Number of Simulators Used: _____

Remarks:

AI NOT CONDUCTED. COMPLIANCE NOT DETERMINED.

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

08/27/2026
Date

Stability Checks

0.05g/210L 0.047 to 0.053	0.08g/210L 0.077 to 0.083	0.20g/210L 0.194 to 0.206	DGS 0.08g/210L 0.077 to 0.083 ☒ ≤ 0.003 of Wet ☒																								
FL HIGHWAY PATROL Intoxilyzer - Alcohol Analyzer Model 8000 08/27/2026 Software: 8100.27 #1 Test g/210L Time <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>11:12</td></tr> <tr><td>Control Test</td><td>0.000</td><td>11:13</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:13</td></tr> <tr><td>Control Test</td><td>0.079</td><td>11:13</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:14</td></tr> <tr><td>Control Test</td><td>0.079</td><td>11:14</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:15</td></tr> <tr><td>Control Test</td><td>0.000</td><td>11:15</td></tr> </table> Average 0.0527 Std Dev 0.0456 Rel. Std Dev(%) 86.6025 Operator's Signature 				Air Blank	0.000	11:12	Control Test	0.000	11:13	Air Blank	0.000	11:13	Control Test	0.079	11:13	Air Blank	0.000	11:14	Control Test	0.079	11:14	Air Blank	0.000	11:15	Control Test	0.000	11:15
Air Blank	0.000	11:12																									
Control Test	0.000	11:13																									
Air Blank	0.000	11:13																									
Control Test	0.079	11:13																									
Air Blank	0.000	11:14																									
Control Test	0.079	11:14																									
Air Blank	0.000	11:15																									
Control Test	0.000	11:15																									
After sample one, I checked the electrical connection at the back of the instrument and noticed it was loose. I secured it and heard a click. Subsequent samples read normally. Will repeat. MC 8/27/26																											

Stability Checks

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																																																																																																																																				
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083																																																																																																																																				
✓	✓	✓	✓																																																																																																																																				
FL HIGHWAY PATROL Intoxilyzer - Alconol Analyzer Model 8000 08/27/2026 Software: 8100.27 Test g/210L Time <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>11:22</td></tr> <tr><td>Control Test</td><td>0.050</td><td>11:22</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:23</td></tr> <tr><td>Control Test</td><td>0.049</td><td>11:24</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:24</td></tr> <tr><td>Control Test</td><td>0.049</td><td>11:25</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:25</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0493</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel. Std Dev(%)</td><td>1.1703</td><td></td></tr> </table>	Air Blank	0.000	11:22	Control Test	0.050	11:22	Air Blank	0.000	11:23	Control Test	0.049	11:24	Air Blank	0.000	11:24	Control Test	0.049	11:25	Air Blank	0.000	11:25	Control Test Stats			Average	0.0493		Std Dev	0.0006		Rel. Std Dev(%)	1.1703		FL HIGHWAY PATROL Intoxilyzer - Alconol Analyzer Model 8000 08/27/2026 Software: 8100.27 Test g/210L Time <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>11:07</td></tr> <tr><td>Control Test</td><td>0.078</td><td>11:08</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:08</td></tr> <tr><td>Control Test</td><td>0.078</td><td>11:09</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:09</td></tr> <tr><td>Control Test</td><td>0.078</td><td>11:10</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:11</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0780</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel. Std Dev(%)</td><td>0.0000</td><td></td></tr> </table>	Air Blank	0.000	11:07	Control Test	0.078	11:08	Air Blank	0.000	11:08	Control Test	0.078	11:09	Air Blank	0.000	11:09	Control Test	0.078	11:10	Air Blank	0.000	11:11	Control Test Stats			Average	0.0780		Std Dev	0.0000		Rel. Std Dev(%)	0.0000		FL HIGHWAY PATROL Intoxilyzer - Alconol Analyzer Model 8000 08/27/2026 Software: 8100.27 Test g/210L Time <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>11:18</td></tr> <tr><td>Control Test</td><td>0.079</td><td>11:19</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:19</td></tr> <tr><td>Control Test</td><td>0.079</td><td>11:19</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:20</td></tr> <tr><td>Control Test</td><td>0.079</td><td>11:20</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:21</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0790</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel. Std Dev(%)</td><td>0.0000</td><td></td></tr> </table>	Air Blank	0.000	11:18	Control Test	0.079	11:19	Air Blank	0.000	11:19	Control Test	0.079	11:19	Air Blank	0.000	11:20	Control Test	0.079	11:20	Air Blank	0.000	11:21	Control Test Stats			Average	0.0790		Std Dev	0.0000		Rel. Std Dev(%)	0.0000		#2 FL HIGHWAY PATROL Intoxilyzer - Alconol Analyzer Model 8000 08/27/2026 Software: 8100.27 Test g/210L Time <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>11:18</td></tr> <tr><td>Control Test</td><td>0.079</td><td>11:19</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:19</td></tr> <tr><td>Control Test</td><td>0.079</td><td>11:19</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:20</td></tr> <tr><td>Control Test</td><td>0.079</td><td>11:20</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:21</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0790</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel. Std Dev(%)</td><td>0.0000</td><td></td></tr> </table>	Air Blank	0.000	11:18	Control Test	0.079	11:19	Air Blank	0.000	11:19	Control Test	0.079	11:19	Air Blank	0.000	11:20	Control Test	0.079	11:20	Air Blank	0.000	11:21	Control Test Stats			Average	0.0790		Std Dev	0.0000		Rel. Std Dev(%)	0.0000	
Air Blank	0.000	11:22																																																																																																																																					
Control Test	0.050	11:22																																																																																																																																					
Air Blank	0.000	11:23																																																																																																																																					
Control Test	0.049	11:24																																																																																																																																					
Air Blank	0.000	11:24																																																																																																																																					
Control Test	0.049	11:25																																																																																																																																					
Air Blank	0.000	11:25																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.0493																																																																																																																																						
Std Dev	0.0006																																																																																																																																						
Rel. Std Dev(%)	1.1703																																																																																																																																						
Air Blank	0.000	11:07																																																																																																																																					
Control Test	0.078	11:08																																																																																																																																					
Air Blank	0.000	11:08																																																																																																																																					
Control Test	0.078	11:09																																																																																																																																					
Air Blank	0.000	11:09																																																																																																																																					
Control Test	0.078	11:10																																																																																																																																					
Air Blank	0.000	11:11																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.0780																																																																																																																																						
Std Dev	0.0000																																																																																																																																						
Rel. Std Dev(%)	0.0000																																																																																																																																						
Air Blank	0.000	11:18																																																																																																																																					
Control Test	0.079	11:19																																																																																																																																					
Air Blank	0.000	11:19																																																																																																																																					
Control Test	0.079	11:19																																																																																																																																					
Air Blank	0.000	11:20																																																																																																																																					
Control Test	0.079	11:20																																																																																																																																					
Air Blank	0.000	11:21																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.0790																																																																																																																																						
Std Dev	0.0000																																																																																																																																						
Rel. Std Dev(%)	0.0000																																																																																																																																						
Air Blank	0.000	11:18																																																																																																																																					
Control Test	0.079	11:19																																																																																																																																					
Air Blank	0.000	11:19																																																																																																																																					
Control Test	0.079	11:19																																																																																																																																					
Air Blank	0.000	11:20																																																																																																																																					
Control Test	0.079	11:20																																																																																																																																					
Air Blank	0.000	11:21																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.0790																																																																																																																																						
Std Dev	0.0000																																																																																																																																						
Rel. Std Dev(%)	0.0000																																																																																																																																						
Operator's Signature 	Operator's Signature 	Operator's Signature 	Operator's Signature 																																																																																																																																				

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: FL HIGHWAY PATROL
Time of Inspection: 14:07

Date of Inspection: 08/27/2026

Serial Number: 80-000937
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202603A Exp: 03/03/2028	0.08g/210L Test (g/210L) Lot#:202603B Exp: 03/03/2028	0.20g/210L Test (g/210L) Lot#:202603D Exp: 03/04/2028	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:28424080A3 Exp: 11/05/2026
0.000	0.050	0.078	0.198	0.078
0.000	0.050	0.078	0.199	0.078
0.000	0.049	0.078	0.199	0.078
0.000	0.050	0.078	0.199	0.078
0.000	0.050	0.078	0.199	0.078
0.000	0.050	0.078	0.199	0.079
0.000	0.050	0.078	0.199	0.078
0.000	0.051	0.078	0.198	0.078
0.000	0.051	0.078	0.199	0.078
0.000	0.050	0.079	0.198	0.077


Standard Deviations	0.0005	0.0003	0.0004	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

08/27/2026
Date

Florida Department of Law Enforcement
Alcohol Testing Program - Average Calculations for Calibration Certificate

Calculations By: TDG

Serial Number: 80-000937

Date: 8/31/2026

Solution:

0.05 g/210 L Test: 0.501 / 10 = 0.050 g/210L
(SUM) (AVERAGE)

0.08 g/210L Test: 0.781 / 10 = 0.078 g/210L
(SUM) (AVERAGE)

0.20 g/210L Test: 1.987 / 10 = 0.199 g/210L
(SUM) (AVERAGE)

0.08 g/210L Dry Gas Std Test: 0.780 / 10 = 0.078 g/210L
(SUM) (AVERAGE)

Comments:

Average = (Sum of observations) / (Total number of observations)

All results are reported in g/210L



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000937, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-000937</u>	Average of y ± Uncertainty*
Owning Agency:	<u>FL HIGHWAY PATROL</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>08/27/2026</u>	0.078 g/210 L 0.004
Calibration Time:	<u>14:07</u>	0.199 g/ 210 L 0.008
		0.078 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Taylor
Gutschow
Digitally signed by Taylor
Gutschow
Date: 2026.08.31 10:42:09
-04'00'

08/27/2026

Date

TAYLOR D GUTSCHOW,
Department Inspector

FDLE/ATP Form 69 August 2026
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality