

Alcohol Testing Program - Instrument Processing Sheet

Agency: LEVY COUNTY SO

Instrument Serial Number: 80-001292

Date In: 6/19/2026

DI Completion Date: 6/26/2026

☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake By: <u>WKP</u> Date: <u>6/19/2026</u>		Quality Checks By: <u>WKP</u> Date: <u>6/26/2026</u>		Flow Adjustment By: _____																
☒ Annual ☐ Dropped Off ☐ Registration ☒ Return from CMI / EE ☐ Training Instrument Visual Inspection ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/Accessories ☐ Power Cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes:		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value: <u>197</u> ☒ Flow Verification (L/s) Flow Column #: <u>ATP 105</u> 32 mm <u>0.156</u> (.139-.169) 36 mm <u>0.171</u> (.156-.190) 53 mm <u>0.246</u> (.228-.278) 103 mm <u>0.507</u> (.447-.547) ☒ Barometric Pressure Check Gauge ID #: <u>28427</u> Gauge: <u>1016</u> Instrument: <u>1018</u> ☒ Stability Checks		Flow Column #: _____ ☐ 5L/min – 17mm ☐ 15L/min – 53mm ☐ 30L/min – 103mm ☐ R-Value: _____ ☐ Post Adjustment Verification (L/S) Flow Column #: _____ 32 mm _____ (.139-.169) 36 mm _____ (.156-.190) 53 mm _____ (.228-.278) 103 mm _____ (.447-.547)																
		<table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot#/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>MP5088</td> <td>202603A 3/3/2028</td> </tr> <tr> <td>0.080</td> <td>MP5089</td> <td>202603B 3/3/2028</td> </tr> <tr> <td>0.200</td> <td>MP5090</td> <td>202603D 3/4/2028</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>23825080A1 10/5/2027</td> </tr> </tbody> </table>		Simulator	Serial #	Lot#/Exp	0.050	MP5088	202603A 3/3/2028	0.080	MP5089	202603B 3/3/2028	0.200	MP5090	202603D 3/4/2028	0.080 DGS	N/A	23825080A1 10/5/2027	Maintenance By: _____ Date: _____ ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Tank Sensor Tare ☐ Breath Tube Replacement ☐ Other:	
Simulator	Serial #	Lot#/Exp																		
0.050	MP5088	202603A 3/3/2028																		
0.080	MP5089	202603B 3/3/2028																		
0.200	MP5090	202603D 3/4/2028																		
0.080 DGS	N/A	23825080A1 10/5/2027																		
Optical Bench Adjustment By: _____		Department Inspection By: <u>WKP</u>																		
Barometric Pressure Gauge: _____ ID#: _____		Barometric Pressure ID#: <u>34419</u>																		
Simulator	Serial #	Lot #	Expiration	Gauge: <u>1018</u> Instrument: <u>1018</u>																
0.000		N/A	N/A	Mouth Alcohol Solution Lot #: <u>2025-D</u> Exp: <u>9/25/2027</u>																
0.040				Acetone Stock Solution Lot #: <u>2025-B</u> Exp: <u>9/22/2027</u>																
0.100				Simulator	Serial Number															
0.200				0.000	MP5086															
0.300				Interferent	MP5087															
0.080 DGS	N/A			0.050	MP5088															
☐ Post Optical Bench Adjustment Stability Checks				0.080	MP5089															
Simulator	Serial #	Lot #	Expiration	0.200	MP5090															
0.050				Attachments																
0.080				☒ Form 41 ☐ Post-Stability Checks ☒ Stability Checks ☐ Flow Adjustment ☒ Calibration Certificate ☐ Form 40 ☐ Optical Bench Adjustment ☐ Other:																
0.200																				
0.080 DGS	N/A																			
Gauge ID #: _____																				
Gauge: _____ Instrument: _____																				
Notes/Suggested Service:				☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use																
				Shayla Platt Digitally signed by Shayla Platt Date: 2026.06.30 12:44:12 -04'00'																
				LeAndra Higginbotham Digitally signed by LeAndra Higginbotham Date: 2026.07.01 15:39:50 -04'00'																
				Tech Review Admin Review																

Stability Checks

0.05g/210L 0.047 to 0.053	0.08g/210L 0.077 to 0.083	0.20g/210L 0.194 to 0.206	DGS 0.08g/210L 0.077 to 0.083 ≤0.003 of Wet																																																																																																																																																												
LEVY COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001292 05/26/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>10:13</td></tr> <tr><td>Control Test</td><td>0.051</td><td>10:13</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:14</td></tr> <tr><td>Control Test</td><td>0.050</td><td>10:15</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:15</td></tr> <tr><td>Control Test</td><td>0.049</td><td>10:16</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:16</td></tr> <tr><td>Control Test</td><td>0.000</td><td>10:16</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0500</td><td></td></tr> <tr><td>Std Dev</td><td>0.0010</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>2.0000</td><td></td></tr> </tbody> </table> Operator's Signature: <i>menpi</i>	Test	g/210L	Time	Air Blank	0.000	10:13	Control Test	0.051	10:13	Air Blank	0.000	10:14	Control Test	0.050	10:15	Air Blank	0.000	10:15	Control Test	0.049	10:16	Air Blank	0.000	10:16	Control Test	0.000	10:16	Control Test Stats			Average	0.0500		Std Dev	0.0010		Rel Std Dev(%)	2.0000		LEVY COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001292 05/26/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>10:17</td></tr> <tr><td>Control Test</td><td>0.078</td><td>10:18</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:19</td></tr> <tr><td>Control Test</td><td>0.078</td><td>10:19</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:20</td></tr> <tr><td>Control Test</td><td>0.077</td><td>10:21</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:21</td></tr> <tr><td>Control Test</td><td>0.000</td><td>10:21</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0777</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7434</td><td></td></tr> </tbody> </table> Operator's Signature: <i>menpi</i>	Test	g/210L	Time	Air Blank	0.000	10:17	Control Test	0.078	10:18	Air Blank	0.000	10:19	Control Test	0.078	10:19	Air Blank	0.000	10:20	Control Test	0.077	10:21	Air Blank	0.000	10:21	Control Test	0.000	10:21	Control Test Stats			Average	0.0777		Std Dev	0.0006		Rel Std Dev(%)	0.7434		LEVY COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001292 05/26/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>10:22</td></tr> <tr><td>Control Test</td><td>0.200</td><td>10:23</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:23</td></tr> <tr><td>Control Test</td><td>0.200</td><td>10:24</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:24</td></tr> <tr><td>Control Test</td><td>0.199</td><td>10:25</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:26</td></tr> <tr><td>Control Test</td><td>0.000</td><td>10:26</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1997</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.2892</td><td></td></tr> </tbody> </table> Operator's Signature: <i>menpi</i>	Test	g/210L	Time	Air Blank	0.000	10:22	Control Test	0.200	10:23	Air Blank	0.000	10:23	Control Test	0.200	10:24	Air Blank	0.000	10:24	Control Test	0.199	10:25	Air Blank	0.000	10:26	Control Test	0.000	10:26	Control Test Stats			Average	0.1997		Std Dev	0.0006		Rel Std Dev(%)	0.2892		LEVY COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001292 05/26/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>10:27</td></tr> <tr><td>Control Test</td><td>0.080</td><td>10:28</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:28</td></tr> <tr><td>Control Test</td><td>0.080</td><td>10:28</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:29</td></tr> <tr><td>Control Test</td><td>0.081</td><td>10:29</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:30</td></tr> <tr><td>Control Test</td><td>0.000</td><td>10:30</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0803</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7187</td><td></td></tr> </tbody> </table> Operator's Signature: <i>menpi</i>	Test	g/210L	Time	Air Blank	0.000	10:27	Control Test	0.080	10:28	Air Blank	0.000	10:28	Control Test	0.080	10:28	Air Blank	0.000	10:29	Control Test	0.081	10:29	Air Blank	0.000	10:30	Control Test	0.000	10:30	Control Test Stats			Average	0.0803		Std Dev	0.0006		Rel Std Dev(%)	0.7187	
Test	g/210L	Time																																																																																																																																																													
Air Blank	0.000	10:13																																																																																																																																																													
Control Test	0.051	10:13																																																																																																																																																													
Air Blank	0.000	10:14																																																																																																																																																													
Control Test	0.050	10:15																																																																																																																																																													
Air Blank	0.000	10:15																																																																																																																																																													
Control Test	0.049	10:16																																																																																																																																																													
Air Blank	0.000	10:16																																																																																																																																																													
Control Test	0.000	10:16																																																																																																																																																													
Control Test Stats																																																																																																																																																															
Average	0.0500																																																																																																																																																														
Std Dev	0.0010																																																																																																																																																														
Rel Std Dev(%)	2.0000																																																																																																																																																														
Test	g/210L	Time																																																																																																																																																													
Air Blank	0.000	10:17																																																																																																																																																													
Control Test	0.078	10:18																																																																																																																																																													
Air Blank	0.000	10:19																																																																																																																																																													
Control Test	0.078	10:19																																																																																																																																																													
Air Blank	0.000	10:20																																																																																																																																																													
Control Test	0.077	10:21																																																																																																																																																													
Air Blank	0.000	10:21																																																																																																																																																													
Control Test	0.000	10:21																																																																																																																																																													
Control Test Stats																																																																																																																																																															
Average	0.0777																																																																																																																																																														
Std Dev	0.0006																																																																																																																																																														
Rel Std Dev(%)	0.7434																																																																																																																																																														
Test	g/210L	Time																																																																																																																																																													
Air Blank	0.000	10:22																																																																																																																																																													
Control Test	0.200	10:23																																																																																																																																																													
Air Blank	0.000	10:23																																																																																																																																																													
Control Test	0.200	10:24																																																																																																																																																													
Air Blank	0.000	10:24																																																																																																																																																													
Control Test	0.199	10:25																																																																																																																																																													
Air Blank	0.000	10:26																																																																																																																																																													
Control Test	0.000	10:26																																																																																																																																																													
Control Test Stats																																																																																																																																																															
Average	0.1997																																																																																																																																																														
Std Dev	0.0006																																																																																																																																																														
Rel Std Dev(%)	0.2892																																																																																																																																																														
Test	g/210L	Time																																																																																																																																																													
Air Blank	0.000	10:27																																																																																																																																																													
Control Test	0.080	10:28																																																																																																																																																													
Air Blank	0.000	10:28																																																																																																																																																													
Control Test	0.080	10:28																																																																																																																																																													
Air Blank	0.000	10:29																																																																																																																																																													
Control Test	0.081	10:29																																																																																																																																																													
Air Blank	0.000	10:30																																																																																																																																																													
Control Test	0.000	10:30																																																																																																																																																													
Control Test Stats																																																																																																																																																															
Average	0.0803																																																																																																																																																														
Std Dev	0.0006																																																																																																																																																														
Rel Std Dev(%)	0.7187																																																																																																																																																														

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: LEVY COUNTY SO
Time of Inspection: 14:54

Date of Inspection: 06/26/2026

Serial Number: 80-001292
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202603A Exp: 03/03/2028	0.08g/210L Test (g/210L) Lot#:202603B Exp: 03/03/2028	0.20g/210L Test (g/210L) Lot#:202603D Exp: 03/04/2028	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:23825080A1 Exp: 10/05/2027
0.000	0.048	0.077	0.199	0.080
0.000	0.048	0.077	0.198	0.081
0.000	0.048	0.077	0.198	0.081
0.000	0.048	0.077	0.197	0.082
0.000	0.048	0.076	0.197	0.082
0.000	0.048	0.077	0.197	0.082
0.000	0.048	0.077	0.197	0.083
0.000	0.048	0.076	0.197	0.083
0.000	0.048	0.076	0.197	0.083
0.000	0.048	0.076	0.197	0.083
0.000	0.048	0.077	0.196	0.083

Standard Deviations	0.0000	0.0004	0.0008	0.0010
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0005 Number of Simulators Used: 5

Remarks:

A F / M A: No Sample Provided. *Timed out, repeated. WEP 6/26/2026*

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Wen-Chi K Pierson WEN-CHI K PIERSON
Signature and Printed Name

06/26/2026
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001292, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001292</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>LEVY COUNTY SO</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>06/26/2026</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>14:54</u>	0.200 g/ 210 L	0.008
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

06/26/2026

Date

Wen-Chi
Pierson

Digitally signed by Wen-Chi
Pierson
Date: 2026.06.30 11:35:15
-04'00'

WEN-CHI K PIERSON,
Department Inspector