Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

Agency SARASOTA COUNTY SO **Date** 01/20/2025 **Type** Agency Inspection
Inspector SANDERS,JOSHUA,P **Time** 08:03:37 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 10124080A2
Expiration 03/28/2025 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 06/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	Yes
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.000
0.05 g/210L Test # (Repeat) 1	0.050
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.000
0.05 g/210L Test # (Repeat) 2	0.049
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.000
0.05 g/210L Test # (Repeat) 3	0.050
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

Time changed. LID WAS NOT SECURE - NO SEAL. PASSED INSPECT

Data Download Date/Time 01/20/2025 08:03

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

Agency SARASOTA COUNTY SO **Date** 02/21/2025 **Type** Agency Inspection
Inspector FELTOVIC,DEP MIKE, **Time** 07:37:55 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 10124080A2
Expiration 03/28/2025 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 06/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

INSPECTED AT JAIL

Data Download Date/Time 02/21/2025 07:39

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

Agency SARASOTA COUNTY SO **Date** 03/05/2025 **Type** Department Inspection
Inspector GUTSCHOW,TAYLOR,D **Time** 14:06:15 **Software** 8100.27

0.05 Lot # 202303K **0.08 Lot #** 202303L **0.20 Lot #** 202304C **0.08 Gas Lot #** AG429602
Expiration 03/29/2025 **Expiration** 03/29/2025 **Expiration** 04/05/2025 **Expiration** 10/22/2026

Number of Simulators 5 **Average Standard Deviation** 0.0002 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.196
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.050
0.08 g/210L Test # 4	0.079
0.20 g/210L Test # 4	0.197
0.08 g/210L Dry Gas Std Test # 4	0.080
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.050
0.08 g/210L Test # 5	0.079
0.20 g/210L Test # 5	0.197
0.08 g/210L Dry Gas Std Test # 5	0.080
Interferent Detect Test # 5	INT

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.050
0.08 g/210L Test # 6	0.079
0.20 g/210L Test # 6	0.197
0.08 g/210L Dry Gas Std Test # 6	0.079
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.079
0.20 g/210L Test # 7	0.197
0.08 g/210L Dry Gas Std Test # 7	0.080
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.050
0.08 g/210L Test # 8	0.079
0.20 g/210L Test # 8	0.197
0.08 g/210L Dry Gas Std Test # 8	0.079
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.050
0.08 g/210L Test # 9	0.079
0.20 g/210L Test # 9	0.197
0.08 g/210L Dry Gas Std Test # 9	0.079
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.079
0.20 g/210L Test # 10	0.197
0.08 g/210L Dry Gas Std Test # 10	0.080
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0000
0.08 g/210L Test (Std Dev)	0.0000
0.20 g/210L Test (Std Dev)	0.0003
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

Data Download Date/Time 03/05/2025 14:07

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

Agency SARASOTA COUNTY SO **Date** 04/04/2025 **Type** Agency Inspection
Inspector FELTOVIC,DEP MIKE, **Time** 08:51:56 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 10124080A2
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 06/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

INSPECTED AT JAIL

Data Download Date/Time 04/04/2025 08:52

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

Agency SARASOTA COUNTY SO **Date** 05/22/2025 **Type** Agency Inspection
Inspector SANDERS,JOSHUA,P **Time** 07:39:05 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 10124080A2
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 06/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

PASSED INSPECTION AT JAIL

Data Download Date/Time 05/22/2025 07:37

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

Agency SARASOTA COUNTY SO **Date** 06/24/2025 **Type** Agency Inspection
Inspector SANDERS,JOSHUA,P **Time** 07:57:02 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 10124080A2
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 06/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.200
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.201
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

PASSED INSPECTION AT JAIL

Data Download Date/Time 06/24/2025 07:56

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

Agency SARASOTA COUNTY SO **Date** 07/16/2025 **Type** Agency Inspection
Inspector SANDERS,JOSHUA,P **Time** 07:39:18 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 10124080A2
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 06/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	Yes
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.200
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.200
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.200
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

PASSED INSPECTION AT JAIL

Data Download Date/Time 07/16/2025 07:46

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

Agency SARASOTA COUNTY SO **Date** 08/28/2025 **Type** Agency Inspection
Inspector SANDERS,JOSHUA,P **Time** 09:22:54 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 15025080A1
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 07/05/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.201
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.201
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.201
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

PASSED INSPECTION AT JAIL

Data Download Date/Time 08/28/2025 09:21

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

Agency SARASOTA COUNTY SO **Date** 09/29/2025 **Type** Agency Inspection
Inspector SANDERS,JOSHUA,P **Time** 17:39:10 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 15025080A1
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 07/05/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

PASSED INSPECTION AT JAIL

Data Download Date/Time 09/29/2025 17:38

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

Agency SARASOTA COUNTY SO **Date** 10/31/2025 **Type** Agency Inspection
Inspector FELTOVIC,DEP MIKE, **Time** 14:43:17 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 15025080A1
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 07/05/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

INSPECTED AT JAIL

Data Download Date/Time 10/31/2025 14:44

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

Agency SARASOTA COUNTY SO **Date** 11/14/2025 **Type** Agency Inspection
Inspector SANDERS,JOSHUA,P **Time** 12:38:11 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 15025080A1
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 07/05/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.200
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.200
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.081
0.20 g/210L Test # 3	0.200
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

PASSED INSPECTION AT JAIL

Data Download Date/Time 11/14/2025 12:38

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

Agency SARASOTA COUNTY SO **Date** 12/12/2025 **Type** Agency Inspection
Inspector SANDERS,JOSHUA,P **Time** 09:17:36 **Software** 8100.27

0.05 Lot # 202506E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 15025080A1
Expiration 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 07/05/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.200
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.200
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.081
0.20 g/210L Test # 3	0.200
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

PASSED INSPECTION AT JAIL

Data Download Date/Time 12/12/2025 09:16

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001349

Agency	DAYTONA COLLEGE CJTC	**Date**	02/12/2025	**Type**	Agency Inspection
Inspector	WAGNER,ROBERT,F	**Time**	15:12:21	**Software**	8100.27

0.05 Lot #	**0.08 Lot #**	**0.20 Lot #**	**0.08 Gas Lot #**
Expiration	**Expiration**	**Expiration**	**Expiration**

Number of Simulators **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	Yes		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Time-Date changed.

Data Download Date/Time 01/16/2026 13:40

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001349

Agency	DAYTONA COLLEGE CJTC	**Date**	03/10/2025	**Type**	Agency Inspection
Inspector	WAGNER,ROBERT,F	**Time**	09:07:54	**Software**	8100.27

0.05 Lot #	**0.08 Lot #**	**0.20 Lot #**	**0.08 Gas Lot #**
Expiration	**Expiration**	**Expiration**	**Expiration**

Number of Simulators **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Data Download Date/Time 01/16/2026 13:40

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001349

Agency DAYTONA COLLEGE CJTC **Date** 03/12/2025 **Type** Agency Inspection
Inspector BROCK,NATHAN,A **Time** 11:23:20 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 16813080A2
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 06/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.196
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.081
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/16/2026 13:40

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001349

Agency	DAYTONA COLLEGE CJTC	**Date**	06/04/2025	**Type**	Agency Inspection
Inspector	WAGNER,ROBERT,F	**Time**	08:18:04	**Software**	8100.27

0.05 Lot #	**0.08 Lot #**	**0.20 Lot #**	**0.08 Gas Lot #**
Expiration	**Expiration**	**Expiration**	**Expiration**

Number of Simulators **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Data Download Date/Time 01/16/2026 13:40

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001349

Agency DAYTONA COLLEGE CJTC **Date** 06/05/2025 **Type** Agency Inspection
Inspector GASTON,SCOTT,B **Time** 11:38:17 **Software** 8100.27

0.05 Lot # 202101C **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 06419080A1
Expiration 01/12/2026 **Expiration** 06/14/2026 **Expiration** 06/15/2026 **Expiration** 06/12/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/16/2026 13:40

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001349

Agency	DAYTONA COLLEGE CJTC	**Date**	06/25/2025	**Type**	Agency Inspection
Inspector	WAGNER,ROBERT,F	**Time**	13:43:23	**Software**	8100.27

0.05 Lot #	**0.08 Lot #**	**0.20 Lot #**	**0.08 Gas Lot #**
Expiration	**Expiration**	**Expiration**	**Expiration**

Number of Simulators **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Data Download Date/Time 01/16/2026 13:40

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001349

Agency	DAYTONA COLLEGE CJTC	**Date**	07/07/2025	**Type**	Agency Inspection
Inspector	WAGNER,ROBERT,F	**Time**	14:20:23	**Software**	8100.27

0.05 Lot #	**0.08 Lot #**	**0.20 Lot #**	**0.08 Gas Lot #**
Expiration	**Expiration**	**Expiration**	**Expiration**

Number of Simulators **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Data Download Date/Time 01/16/2026 13:40

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001349

Agency DAYTONA COLLEGE CJTC **Date** 07/07/2025 **Type** Agency Inspection
Inspector FEELEY,JAMES,E **Time** 15:34:51 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202206B **0.20 Lot #** 202005A **0.08 Gas Lot #** 1589901
Expiration 02/27/2026 **Expiration** 06/14/2026 **Expiration** 05/12/2026 **Expiration** 10/30/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/16/2026 13:40

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001349

Agency	DAYTONA COLLEGE CJTC	**Date**	07/07/2025	**Type**	Agency Inspection
Inspector	BLASSINGAME,PAUL,X	**Time**	15:42:31	**Software**	8100.27

0.05 Lot #	**0.08 Lot #**	**0.20 Lot #**	**0.08 Gas Lot #**
Expiration	**Expiration**	**Expiration**	**Expiration**

Number of Simulators 5 **Compliance** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	Yes		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A: Sequence AbortedRange Exceeded. Non-compliance: .

Data Download Date/Time 01/16/2026 13:40

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001349

Agency DAYTONA COLLEGE CJTC **Date** 07/07/2025 **Type** Agency Inspection
Inspector BLASSINGAME,PAUL,X **Time** 16:14:34 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202206B **0.20 Lot #** 202005A **0.08 Gas Lot #** 1589901
Expiration 02/27/2026 **Expiration** 06/14/2026 **Expiration** 05/12/2026 **Expiration** 10/30/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/16/2026 13:40

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001349

Agency DAYTONA COLLEGE CJTC **Date** 07/08/2025 **Type** Agency Inspection
Inspector CHRISTENSEN,IAN,J **Time** 10:17:27 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202007A **0.20 Lot #** 201802G **0.08 Gas Lot #** 1584901
Expiration 02/27/2027 **Expiration** 07/21/2027 **Expiration** 02/22/2027 **Expiration** 10/30/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/16/2026 13:40

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001349

Agency DAYTONA COLLEGE CJTC **Date** 07/08/2025 **Type** Agency Inspection
Inspector LOGAN,TYRONE,P **Time** 11:00:04 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202007A **0.20 Lot #** 201802G **0.08 Gas Lot #** 1539801
Expiration 12/12/2025 **Expiration** 12/12/2025 **Expiration** 12/12/2025 **Expiration** 12/12/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.196
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/16/2026 13:40

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001349

Agency DAYTONA COLLEGE CJTC **Date** 07/08/2025 **Type** Agency Inspection
Inspector GROHMAN,JOHN,L **Time** 13:21:23 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202007A **0.20 Lot #** 201802G **0.08 Gas Lot #** 1589901
Expiration 02/27/2026 **Expiration** 07/21/2025 **Expiration** 02/22/2026 **Expiration** 12/25/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.196
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

A F / M A: .

Data Download Date/Time 01/16/2026 13:40

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001349

Agency	DAYTONA COLLEGE CJTC	**Date**	07/08/2025	**Type**	Agency Inspection
Inspector	BLASSINGAME,PAUL,X	**Time**	13:31:31	**Software**	8100.27

0.05 Lot #	**0.08 Lot #**	**0.20 Lot #**	**0.08 Gas Lot #**
Expiration	**Expiration**	**Expiration**	**Expiration**

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

N.

Data Download Date/Time 01/16/2026 13:40

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001349

Agency	DAYTONA COLLEGE CJTC	**Date**	07/08/2025	**Type**	Agency Inspection
Inspector	BLASSINGAME,PAUL,X	**Time**	13:36:30	**Software**	8100.27

0.05 Lot #	**0.08 Lot #**	**0.20 Lot #**	**0.08 Gas Lot #**
Expiration	**Expiration**	**Expiration**	**Expiration**

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

N.

Data Download Date/Time 01/16/2026 13:40

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001349

Agency DAYTONA COLLEGE CJTC **Date** 07/08/2025 **Type** Agency Inspection
Inspector BLASSINGAME,PAUL,X **Time** 14:13:54 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202007A **0.20 Lot #** 201802G **0.08 Gas Lot #** 1589901
Expiration 02/27/2026 **Expiration** 07/21/2025 **Expiration** 02/22/2026 **Expiration** 10/30/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

A F / M A: Range Exceeded.

Data Download Date/Time 01/16/2026 13:40

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001349

Agency DAYTONA COLLEGE CJTC **Date** 07/09/2025 **Type** Agency Inspection
Inspector RODRIGUEZCAMACH,ANTHONY **Time** 09:03:07 **Software** 8100.27

0.05 Lot #	202402E	**0.08 Lot #**	202007A	**0.20 Lot #**	202005A	**0.08 Gas Lot #**	1589901
Expiration	02/27/2026	**Expiration**	02/27/2026	**Expiration**	02/27/2026	**Expiration**	02/27/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/16/2026 13:40

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001349

Agency DAYTONA COLLEGE CJTC **Date** 07/09/2025 **Type** Agency Inspection
Inspector GASTON,SCOTT,B **Time** 09:58:12 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202007A **0.20 Lot #** 202005A **0.08 Gas Lot #** 1345765
Expiration 02/27/2026 **Expiration** 07/21/2025 **Expiration** 05/12/2026 **Expiration** 07/21/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/16/2026 13:41

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001349

Agency	DAYTONA COLLEGE CJTC	**Date**	08/26/2025	**Type**	Agency Inspection
Inspector	WAGNER,ROBERT,F	**Time**	14:42:11	**Software**	8100.27

0.05 Lot #	**0.08 Lot #**	**0.20 Lot #**	**0.08 Gas Lot #**
Expiration	**Expiration**	**Expiration**	**Expiration**

Number of Simulators **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	Yes		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Time-Date changed.

Data Download Date/Time 01/16/2026 13:41

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001349

Agency	DAYTONA COLLEGE CJTC	**Date**	11/10/2025	**Type**	Agency Inspection
Inspector	WAGNER,ROBERT,F	**Time**	10:31:32	**Software**	8100.27

0.05 Lot #	**0.08 Lot #**	**0.20 Lot #**	**0.08 Gas Lot #**
Expiration	**Expiration**	**Expiration**	**Expiration**

Number of Simulators **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Data Download Date/Time 01/16/2026 13:41

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001349

Agency	DAYTONA COLLEGE CJTC	**Date**	12/10/2025	**Type**	Agency Inspection
Inspector	WAGNER,ROBERT,F	**Time**	13:01:23	**Software**	8100.27

0.05 Lot #	**0.08 Lot #**	**0.20 Lot #**	**0.08 Gas Lot #**
Expiration	**Expiration**	**Expiration**	**Expiration**

Number of Simulators **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Data Download Date/Time 01/16/2026 13:41

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001350

Agency BARTOW PD **Date** 01/12/2025 **Type** Agency Inspection
Inspector SILLS,MARK,A **Time** 08:49:48 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202307C **0.20 Lot #** 202307C **0.08 Gas Lot #** 231603
Expiration 02/27/2026 **Expiration** 07/12/2025 **Expiration** 07/12/2025 **Expiration** 07/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.195
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.196
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.196
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/12/2025 08:48

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001350

Agency BARTOW PD **Date** 01/12/2025 **Type** Agency Inspection
Inspector SILLS,MARK,A **Time** 09:29:17 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 231603
Expiration 02/27/2026 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 07/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # (Repeat) 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # (Repeat) 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # (Repeat) 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

00: Ambient Fail.

Data Download Date/Time 01/12/2025 09:27

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001350

Agency	BARTOW PD	**Date**	02/09/2025	**Type**	Agency Inspection
Inspector	SILLS,MARK,A	**Time**	08:37:42	**Software**	8100.27

0.05 Lot #	202402E	**0.08 Lot #**	202307A	**0.20 Lot #**	202307C	**0.08 Gas Lot #**	231603
Expiration	02/27/2026	**Expiration**	07/11/2025	**Expiration**	07/12/2025	**Expiration**	07/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.195
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/09/2025 08:35

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001350

Agency BARTOW PD **Date** 03/05/2025 **Type** Department Inspection
Inspector GUTSCHOW,TAYLOR,D **Time** 13:51:54 **Software** 8100.27

0.05 Lot # 202303K **Expiration** 03/29/2025
0.08 Lot # 202303L **Expiration** 03/29/2025
0.20 Lot # 202304C **Expiration** 04/05/2025
0.08 Gas Lot # AG429602 **Expiration** 10/22/2026

Number of Simulators 5 **Average Standard Deviation** 0.0002 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.195
0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.195
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.195
0.08 g/210L Dry Gas Std Test # 3	0.078
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.049
0.08 g/210L Test # 4	0.079
0.20 g/210L Test # 4	0.195
0.08 g/210L Dry Gas Std Test # 4	0.077
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.050
0.08 g/210L Test # 5	0.079
0.20 g/210L Test # 5	0.195
0.08 g/210L Dry Gas Std Test # 5	0.078
Interferent Detect Test # 5	INT

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001350

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.049
0.08 g/210L Test # 6	0.079
0.20 g/210L Test # 6	0.195
0.08 g/210L Dry Gas Std Test # 6	0.078
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.049
0.08 g/210L Test # 7	0.079
0.20 g/210L Test # 7	0.195
0.08 g/210L Dry Gas Std Test # 7	0.077
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.050
0.08 g/210L Test # 8	0.079
0.20 g/210L Test # 8	0.195
0.08 g/210L Dry Gas Std Test # 8	0.077
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.050
0.08 g/210L Test # 9	0.079
0.20 g/210L Test # 9	0.195
0.08 g/210L Dry Gas Std Test # 9	0.077
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.079
0.20 g/210L Test # 10	0.195
0.08 g/210L Dry Gas Std Test # 10	0.078
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0005
0.08 g/210L Test (Std Dev)	0.0000
0.20 g/210L Test (Std Dev)	0.0000
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

Data Download Date/Time 03/05/2025 13:50

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001350

Agency BARTOW PD **Date** 03/15/2025 **Type** Agency Inspection
Inspector SILLS,MARK,A **Time** 11:11:58 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202307C **0.08 Gas Lot #** 231603
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 07/12/2025 **Expiration** 07/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.195
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.195
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.196
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 03/15/2025 11:09

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001350

Agency BARTOW PD **Date** 04/29/2025 **Type** Agency Inspection
Inspector SILLS,MARK,A **Time** 12:57:39 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202307C **0.08 Gas Lot #** 231603
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 07/12/2025 **Expiration** 07/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.194
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.196
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.196
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 04/29/2025 12:54

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001350

Agency BARTOW PD **Date** 05/14/2025 **Type** Agency Inspection
Inspector SILLS,MARK,A **Time** 13:03:02 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202307C **0.08 Gas Lot #** 231603
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 07/12/2025 **Expiration** 07/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.195
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.196
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.196
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 05/14/2025 13:00

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001350

Agency	BARTOW PD	**Date**	06/15/2025	**Type**	Agency Inspection
Inspector	SILLS,MARK,A	**Time**	08:32:14	**Software**	8100.27

0.05 Lot #	202402E	**0.08 Lot #**	202410E	**0.20 Lot #**	202307C	**0.08 Gas Lot #**	231603
Expiration	02/27/2026	**Expiration**	10/22/2026	**Expiration**	07/12/2025	**Expiration**	07/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.194
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.195
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.196
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 06/15/2025 08:28

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001350

Agency BARTOW PD **Date** 07/07/2025 **Type** Agency Inspection
Inspector SILLS,MARK,A **Time** 09:34:24 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202307C **0.08 Gas Lot #** 231603
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 07/12/2025 **Expiration** 07/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.193
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.194
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.194
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 07/07/2025 09:30

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001350

Agency BARTOW PD **Date** 08/05/2025 **Type** Agency Inspection
Inspector SILLS,MARK,A **Time** 20:35:25 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 231603
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 07/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 08/05/2025 20:33

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001350

Agency BARTOW PD **Date** 09/25/2025 **Type** Agency Inspection
Inspector SILLS,MARK,A **Time** 17:32:37 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 231603
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 07/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.195
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 09/25/2025 17:31

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001350

Agency BARTOW PD **Date** 10/19/2025 **Type** Agency Inspection
Inspector SILLS,MARK,A **Time** 09:41:08 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 231603
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 07/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.196
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.196
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.196
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 10/19/2025 09:39

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001350

Agency BARTOW PD **Date** 11/16/2025 **Type** Agency Inspection
Inspector SILLS,MARK,A **Time** 08:35:37 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 231603
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 07/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.081
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 11/16/2025 08:33

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001350

Agency BARTOW PD **Date** 12/13/2025 **Type** Agency Inspection
Inspector SILLS,MARK,A **Time** 08:11:18 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 231603
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 07/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.196
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 12/14/2025 08:10

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001350

Agency BARTOW PD **Date** 12/28/2025 **Type** Agency Inspection
Inspector SILLS,MARK,A **Time** 08:23:11 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 231603
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 07/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.196
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.196
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 12/28/2025 08:26

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001360

Agency EVERGLADES NAT PARK **Date** 08/13/2025 **Type** Department Inspection
Inspector GUTSCHOW,TAYLOR,D **Time** 12:53:41 **Software** 8100.27

0.05 Lot # 202406K **0.08 Lot #** 202406L **0.20 Lot #** 202406N **0.08 Gas Lot #** AG429602
Expiration 06/19/2026 **Expiration** 06/19/2026 **Expiration** 06/20/2026 **Expiration** 10/22/2026

Number of Simulators 5 **Average Standard Deviation** 0.0000 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.201
0.08 g/210L Dry Gas Std Test # 1	0.077
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.201
0.08 g/210L Dry Gas Std Test # 2	0.077
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.201
0.08 g/210L Dry Gas Std Test # 3	0.077
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.049
0.08 g/210L Test # 4	0.079
0.20 g/210L Test # 4	0.201
0.08 g/210L Dry Gas Std Test # 4	0.077
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.049
0.08 g/210L Test # 5	0.079
0.20 g/210L Test # 5	0.201
0.08 g/210L Dry Gas Std Test # 5	0.077
Interferent Detect Test # 5	INT

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001360

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.049
0.08 g/210L Test # 6	0.079
0.20 g/210L Test # 6	0.201
0.08 g/210L Dry Gas Std Test # 6	0.077
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.049
0.08 g/210L Test # 7	0.079
0.20 g/210L Test # 7	0.201
0.08 g/210L Dry Gas Std Test # 7	0.077
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.049
0.08 g/210L Test # 8	0.079
0.20 g/210L Test # 8	0.201
0.08 g/210L Dry Gas Std Test # 8	0.077
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.049
0.08 g/210L Test # 9	0.079
0.20 g/210L Test # 9	0.201
0.08 g/210L Dry Gas Std Test # 9	0.077
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.049
0.08 g/210L Test # 10	0.079
0.20 g/210L Test # 10	0.201
0.08 g/210L Dry Gas Std Test # 10	0.078
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0000
0.08 g/210L Test (Std Dev)	0.0000
0.20 g/210L Test (Std Dev)	0.0000
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

Data Download Date/Time 08/13/2025 12:52

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001360

Agency EVERGLADES NAT PARK **Date** 10/24/2025 **Type** Agency Inspection
Inspector SAAVEDRA,OSCAR, **Time** 12:02:59 **Software** 8100.27

0.05 Lot # 202506E **0.08 Lot #** 202410E **0.20 Lot #** 202410E **0.08 Gas Lot #** 28423080A2
Expiration 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 11/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.203
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.204
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.203
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT

Remarks:

A F / M A: Range ExceededRFI Detect.

Data Download Date/Time 10/24/2025 12:06

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001360

Agency EVERGLADES NAT PARK **Date** 11/25/2025 **Type** Agency Inspection
Inspector SAAVEDRA,OSCAR, **Time** 12:25:54 **Software** 8100.27

0.05 Lot # 202506E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 403056709
Expiration 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 05/28/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.203
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 11/25/2025 12:35

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001360

Agency EVERGLADES NAT PARK **Date** 12/15/2025 **Type** Agency Inspection
Inspector SAAVEDRA,OSCAR,- **Time** 14:57:53 **Software** 8100.27

0.05 Lot # 202506E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 403056709
Expiration 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 05/28/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.200
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.200
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.200
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 12/15/2025 14:56

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001361

Agency FL HIGHWAY PATROL **Date** 02/21/2025 **Type** Department Inspection
Inspector GUTSCHOW,TAYLOR,D **Time** 13:46:50 **Software** 8100.27

0.05 Lot # 202303K **0.08 Lot #** 202303L **0.20 Lot #** 202304C **0.08 Gas Lot #** AG429602
Expiration 03/29/2025 **Expiration** 03/29/2025 **Expiration** 04/05/2025 **Expiration** 10/22/2026

Number of Simulators 5 **Average Standard Deviation** 0.0001 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.077
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.077
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.077
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.049
0.08 g/210L Test # 4	0.079
0.20 g/210L Test # 4	0.198
0.08 g/210L Dry Gas Std Test # 4	0.077
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.049
0.08 g/210L Test # 5	0.079
0.20 g/210L Test # 5	0.198
0.08 g/210L Dry Gas Std Test # 5	0.077
Interferent Detect Test # 5	INT

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001361

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.049
0.08 g/210L Test # 6	0.079
0.20 g/210L Test # 6	0.198
0.08 g/210L Dry Gas Std Test # 6	0.077
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.049
0.08 g/210L Test # 7	0.079
0.20 g/210L Test # 7	0.198
0.08 g/210L Dry Gas Std Test # 7	0.077
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.049
0.08 g/210L Test # 8	0.079
0.20 g/210L Test # 8	0.198
0.08 g/210L Dry Gas Std Test # 8	0.076
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.049
0.08 g/210L Test # 9	0.079
0.20 g/210L Test # 9	0.198
0.08 g/210L Dry Gas Std Test # 9	0.077
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.049
0.08 g/210L Test # 10	0.079
0.20 g/210L Test # 10	0.198
0.08 g/210L Dry Gas Std Test # 10	0.076
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0003
0.08 g/210L Test (Std Dev)	0.0000
0.20 g/210L Test (Std Dev)	0.0000
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Data Download Date/Time 02/21/2025 13:45

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001361

Agency FL HIGHWAY PATROL **Date** 02/28/2025 **Type** Agency Inspection
Inspector HITCHCOCK,WILLIAM,D **Time** 14:08:32 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 403207507
Expiration 02/27/2027 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 11/26/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.077
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.076
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/28/2025 14:07

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001361

Agency FL HIGHWAY PATROL **Date** 03/31/2025 **Type** Agency Inspection
Inspector HITCHCOCK,WILLIAM,D **Time** 11:16:01 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 403207507
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 11/26/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.077
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.077
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 04/04/2025 11:47

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001361

Agency FL HIGHWAY PATROL **Date** 04/30/2025 **Type** Agency Inspection
Inspector HITCHCOCK,WILLIAM,D **Time** 14:24:06 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 403207507
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 11/26/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.077
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.077
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 05/30/2025 09:38

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001361

Agency FL HIGHWAY PATROL **Date** 05/30/2025 **Type** Agency Inspection
Inspector HITCHCOCK,WILLIAM,D **Time** 09:29:52 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 403207507
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 11/26/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.200
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 05/30/2025 09:39

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001361

Agency FL HIGHWAY PATROL **Date** 06/30/2025 **Type** Agency Inspection
Inspector HITCHCOCK,WILLIAM,D **Time** 13:40:20 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 403207507
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 11/26/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.196
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.077
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 06/30/2025 13:38

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001361

Agency FL HIGHWAY PATROL **Date** 07/31/2025 **Type** Agency Inspection
Inspector HITCHCOCK,WILLIAM,D **Time** 01:01:06 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 403207507
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 11/26/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.196
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.077
0.20 g/210L Test # 2	0.196
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.077
0.20 g/210L Test # 3	0.196
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 07/31/2025 00:59

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001361

Agency FL HIGHWAY PATROL **Date** 08/29/2025 **Type** Agency Inspection
Inspector HITCHCOCK,WILLIAM,D **Time** 10:35:59 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 403402507
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 11/26/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.077
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.077
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.078
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 08/29/2025 11:01

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001361

Agency FL HIGHWAY PATROL **Date** 09/30/2025 **Type** Agency Inspection
Inspector HITCHCOCK,WILLIAM,D **Time** 09:38:04 **Software** 8100.27

0.05 Lot # 202506E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 403207507
Expiration 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 11/26/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.194
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.077
0.20 g/210L Test # 2	0.195
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.077
0.20 g/210L Test # 3	0.195
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 09/30/2025 09:42

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001361

Agency FL HIGHWAY PATROL **Date** 10/31/2025 **Type** Agency Inspection
Inspector HITCHCOCK,WILLIAM,D **Time** 21:56:00 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 403207507
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 11/26/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.078
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 11/07/2025 08:48

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001361

Agency FL HIGHWAY PATROL **Date** 11/24/2025 **Type** Agency Inspection
Inspector HITCHCOCK,WILLIAM,D **Time** 13:03:19 **Software** 8100.27

0.05 Lot # 202506E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 403207507
Expiration 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 11/26/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 11/24/2025 13:03

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001361

Agency FL HIGHWAY PATROL **Date** 12/31/2025 **Type** Agency Inspection
Inspector HITCHCOCK,WILLIAM,D **Time** 11:44:03 **Software** 8100.27

0.05 Lot # 202506E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 403207507
Expiration 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 11/26/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.082
0.20 g/210L Test # 1	0.195
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.195
0.08 g/210L Dry Gas Std Test # 2	0.082
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.196
0.08 g/210L Dry Gas Std Test # 3	0.082
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 12/31/2025 11:43

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

Agency CHARLOTTE COUNTY SO **Date** 01/12/2025 **Type** Agency Inspection
Inspector CHANDLER,VANESSA,N **Time** 20:51:02 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 06723080A5
Expiration 02/27/2026 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 04/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.194
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.193
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.193
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

CLEANED AND INSPECTED

Data Download Date/Time 01/12/2025 20:48

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

Agency CHARLOTTE COUNTY SO **Date** 02/04/2025 **Type** Agency Inspection
Inspector LEONARD,ALEYA,D **Time** 22:19:29 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 06723080A5
Expiration 02/27/2026 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 04/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.196
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.077
0.20 g/210L Test # 2	0.196
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.077
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

CLEANED AND INSPECTED

Data Download Date/Time 02/04/2025 22:16

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

Agency CHARLOTTE COUNTY SO **Date** 03/09/2025 **Type** Agency Inspection
Inspector DEMPSTER,PAUL,R **Time** 20:11:44 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 06723080A5
Expiration 02/27/2026 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 04/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.196
0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.078
Interferent Detect Test # 3	INT

Remarks:

CLEANED AND INSPECTED

Data Download Date/Time 03/09/2025 20:08

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

Agency CHARLOTTE COUNTY SO **Date** 04/24/2025 **Type** Agency Inspection
Inspector BUTZER,PAIGE, **Time** 07:13:23 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 28423080A4
Expiration 02/27/2026 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 11/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.075
		0.20 g/210L Test # 3	0.192
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

CLEANED AND INSPECTED

Data Download Date/Time 04/24/2025 07:10

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

Agency CHARLOTTE COUNTY SO **Date** 05/25/2025 **Type** Agency Inspection
Inspector LEONARD,ALEYA,D **Time** 21:45:40 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 28423080A4
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 11/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

CLEANED AND INSPECTED

Data Download Date/Time 05/25/2025 21:44

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

Agency CHARLOTTE COUNTY SO **Date** 06/24/2025 **Type** Agency Inspection
Inspector CHANDLER,VANESSA,N **Time** 20:51:04 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 28423080A4
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 11/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

A F / M A: Improper Sample.MOUTH ALCOHOL PROVIDED TOO SOON.

Data Download Date/Time 06/24/2025 20:50

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

Agency CHARLOTTE COUNTY SO **Date** 07/15/2025 **Type** Agency Inspection
Inspector LEONARD,ALEYA,D **Time** 23:24:13 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 28423080A4
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 11/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.200
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

CLEANED AND INSPECTED.

Data Download Date/Time 07/15/2025 23:20

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

Agency	CHARLOTTE COUNTY SO	**Date**	08/22/2025	**Type**	Agency Inspection
Inspector	CHANDLER,VANESSA,N	**Time**	20:36:14	**Software**	8100.27

0.05 Lot #	202402E	**0.08 Lot #**	202410E	**0.20 Lot #**	202410F	**0.08 Gas Lot #**	28423080A4
Expiration	02/27/2026	**Expiration**	10/22/2026	**Expiration**	10/30/2026	**Expiration**	11/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

CLEANED AND INSPECTED

Data Download Date/Time 08/22/2025 20:35

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

Agency CHARLOTTE COUNTY SO **Date** 09/17/2025 **Type** Agency Inspection
Inspector LEONARD,ALEYA,D **Time** 21:55:58 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 28423080A4
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 11/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

CLEANED AND INSPECTED

Data Download Date/Time 09/17/2025 21:55

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

Agency	CHARLOTTE COUNTY SO	**Date**	12/09/2025	**Type**	Agency Inspection
Inspector	GUTSCHOW,TAYLOR,D	**Time**	13:06:10	**Software**	8100.27

0.05 Lot #	**0.08 Lot #**	**0.20 Lot #**	**0.08 Gas Lot #**
Expiration	**Expiration**	**Expiration**	**Expiration**

Number of Simulators **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

AI NOT CONDUCTED. COMPLIANCE NOT DETERMINED.

Data Download Date/Time 12/10/2025 13:01

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

Agency CHARLOTTE COUNTY SO **Date** 12/10/2025 **Type** Department Inspection
Inspector GUTSCHOW,TAYLOR,D **Time** 13:04:40 **Software** 8100.27

0.05 Lot # 202406K **0.08 Lot #** 202406L **0.20 Lot #** 202406N **0.08 Gas Lot #** AG526603
Expiration 06/19/2026 **Expiration** 06/19/2026 **Expiration** 06/20/2026 **Expiration** 09/23/2027

Number of Simulators 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.078
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.048
0.08 g/210L Test # 4	0.078
0.20 g/210L Test # 4	0.199
0.08 g/210L Dry Gas Std Test # 4	0.079
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.048
0.08 g/210L Test # 5	0.078
0.20 g/210L Test # 5	0.197
0.08 g/210L Dry Gas Std Test # 5	0.079
Interferent Detect Test # 5	INT

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.048
0.08 g/210L Test # 6	0.078
0.20 g/210L Test # 6	0.199
0.08 g/210L Dry Gas Std Test # 6	0.078
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.048
0.08 g/210L Test # 7	0.078
0.20 g/210L Test # 7	0.199
0.08 g/210L Dry Gas Std Test # 7	0.078
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.048
0.08 g/210L Test # 8	0.078
0.20 g/210L Test # 8	0.198
0.08 g/210L Dry Gas Std Test # 8	0.078
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.048
0.08 g/210L Test # 9	0.079
0.20 g/210L Test # 9	0.198
0.08 g/210L Dry Gas Std Test # 9	0.078
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.048
0.08 g/210L Test # 10	0.078
0.20 g/210L Test # 10	0.198
0.08 g/210L Dry Gas Std Test # 10	0.078
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0000
0.08 g/210L Test (Std Dev)	0.0003
0.20 g/210L Test (Std Dev)	0.0006
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Data Download Date/Time 12/10/2025 13:01

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

Agency CHARLOTTE COUNTY SO **Date** 12/27/2025 **Type** Agency Inspection
Inspector CHANDLER,VANESSA,N **Time** 23:33:41 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 15025080A1
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 07/05/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.200
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.200
0.08 g/210L Dry Gas Std Test # 3	0.078
Interferent Detect Test # 3	INT

Remarks:

CLEANED AND INSPECTED

Data Download Date/Time 12/27/2025 23:32

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

Agency	GAINESVILLE PD	**Date**	01/27/2025	**Type**	Agency Inspection
Inspector	STEMEN,AARON,M	**Time**	20:11:01	**Software**	8100.27

0.05 Lot #	202402E	**0.08 Lot #**	202307A	**0.20 Lot #**	202307C	**0.08 Gas Lot #**	24923080A3
Expiration	02/27/2026	**Expiration**	07/11/2025	**Expiration**	07/12/2025	**Expiration**	10/05/2025

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

IN COMPLIANCE

Data Download Date/Time 01/27/2025 20:14

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

Agency	GAINESVILLE PD	**Date**	02/26/2025	**Type**	Agency Inspection
Inspector	STEMEN,AARON,M	**Time**	19:03:14	**Software**	8100.27

0.05 Lot #	202402E	**0.08 Lot #**	202410E	**0.20 Lot #**	202410F	**0.08 Gas Lot #**	24923080A3
Expiration	02/27/2026	**Expiration**	10/22/2026	**Expiration**	10/30/2026	**Expiration**	10/05/2025

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

IN COMPLIANCE

Data Download Date/Time 02/26/2025 19:06

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

Agency GAINESVILLE PD **Date** 03/25/2025 **Type** Agency Inspection
Inspector STEMEN,AARON,M **Time** 19:50:14 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 24923080A3
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 10/05/2025

Number of Simulators 4 **Compliance** No

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	No
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

Non-compliance: DIAGNOSTIC CHECK FAIL.

Data Download Date/Time 03/25/2025 20:12

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

Agency GAINESVILLE PD **Date** 09/24/2025 **Type** Department Inspection
Inspector PLATT,SHAYLA,D **Time** 13:12:25 **Software** 8100.27

0.05 Lot # 202406K **0.08 Lot #** 202406L **0.20 Lot #** 202406N **0.08 Gas Lot #** AG510701
Expiration 06/19/2026 **Expiration** 06/19/2026 **Expiration** 06/20/2026 **Expiration** 04/17/2027

Number of Simulators 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.081
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.051
0.08 g/210L Test # 2	0.081
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.078
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.051
0.08 g/210L Test # 4	0.080
0.20 g/210L Test # 4	0.199
0.08 g/210L Dry Gas Std Test # 4	0.078
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.050
0.08 g/210L Test # 5	0.081
0.20 g/210L Test # 5	0.199
0.08 g/210L Dry Gas Std Test # 5	0.078
Interferent Detect Test # 5	INT

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.050
0.08 g/210L Test # 6	0.081
0.20 g/210L Test # 6	0.199
0.08 g/210L Dry Gas Std Test # 6	0.078
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.051
0.08 g/210L Test # 7	0.080
0.20 g/210L Test # 7	0.199
0.08 g/210L Dry Gas Std Test # 7	0.079
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.051
0.08 g/210L Test # 8	0.081
0.20 g/210L Test # 8	0.199
0.08 g/210L Dry Gas Std Test # 8	0.078
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.051
0.08 g/210L Test # 9	0.081
0.20 g/210L Test # 9	0.199
0.08 g/210L Dry Gas Std Test # 9	0.078
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.051
0.08 g/210L Test # 10	0.081
0.20 g/210L Test # 10	0.199
0.08 g/210L Dry Gas Std Test # 10	0.078
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0005
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0003
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

Data Download Date/Time 09/24/2025 13:10

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

Agency GAINESVILLE PD **Date** 10/30/2025 **Type** Agency Inspection
Inspector STEMEN,AARON,M **Time** 02:22:32 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 19825080A1
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 09/05/2027

Number of Simulators 4 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.051
0.08 g/210L Test # 1	0.081
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.081
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.051
0.08 g/210L Test # 3	0.081
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.078
Interferent Detect Test # 3	INT

Remarks:

IN COMPLIANCE

Data Download Date/Time 10/30/2025 18:35

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

Agency GAINESVILLE PD **Date** 11/26/2025 **Type** Agency Inspection
Inspector STEMEN,AARON,M **Time** 19:03:08 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 19825080A1
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 09/05/2027

Number of Simulators 4 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.196
0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.078
Interferent Detect Test # 3	INT

Remarks:

IN COMPLIANCE

Data Download Date/Time 11/26/2025 19:02

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

Agency GAINESVILLE PD **Date** 12/29/2025 **Type** Agency Inspection
Inspector STEMEN,AARON,M **Time** 19:39:02 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 19825080A1
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 09/05/2027

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

IN COMPLIANCE

Data Download Date/Time 12/29/2025 19:37

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001365

Agency LEVY COUNTY SO **Date** 01/04/2025 **Type** Agency Inspection
Inspector LEVY,JAYSON,I **Time** 21:28:47 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 01524080A4
Expiration 03/28/2025 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 03/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.194
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

A F / M A:USED TOO MUCH MA.

Data Download Date/Time 01/10/2025 10:46

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001365

Agency LEVY COUNTY SO **Date** 01/29/2025 **Type** Agency Inspection
Inspector LEVY,JAYSON,I **Time** 14:57:53 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 10124080A2
Expiration 03/28/2025 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 06/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.191
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.192
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/29/2025 17:48

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001365

Agency LEVY COUNTY SO **Date** 02/02/2025 **Type** Agency Inspection
Inspector LEVY,JAYSON,I **Time** 10:58:34 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 10124080A2
Expiration 03/28/2025 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 06/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.202
0.08 g/210L Dry Gas Std Test # 1	0.082
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.202
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.201
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/02/2025 10:58

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001365

Agency LEVY COUNTY SO **Date** 03/30/2025 **Type** Agency Inspection
Inspector LEVY,JAYSON,I **Time** 20:16:41 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 10124080A2
Expiration 02/27/2026 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 06/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.194
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 04/01/2025 08:00

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001365

Agency LEVY COUNTY SO **Date** 04/01/2025 **Type** Agency Inspection
Inspector LEVY,JAYSON,I **Time** 07:50:17 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 10124080A2
Expiration 02/27/2026 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 06/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 04/01/2025 08:00

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001365

Agency LEVY COUNTY SO **Date** 04/11/2025 **Type** Agency Inspection
Inspector LEVY,JAYSON,I **Time** 14:17:43 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 10124080A2
Expiration 02/27/2026 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 06/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.195
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 04/16/2025 19:52

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001365

Agency LEVY COUNTY SO **Date** 05/11/2025 **Type** Agency Inspection
Inspector LEVY,JAYSON,I **Time** 10:27:01 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #**
Expiration 02/27/2026 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration**

Number of Simulators 5 **Compliance** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.235
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # (Repeat) 1	0.189
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	No	0.05 g/210L Test # 2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.076
0.05 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Test: OK	Yes	Alcohol Free Test # 3	0.000
0.20 g/210L Test: OK	No	0.05 g/210L Test # 3	0.046
0.08 g/210L Dry Gas Std Test: OK	No	0.08 g/210L Test # 3	0.076
		Interferent Detect Test # 3	INT

Remarks:

20: Purge Fail SIM ERROR, Improper Sample. Non-compliance:FAIL .20 INSP.

Data Download Date/Time 05/11/2025 10:50

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001365

Agency	LEVY COUNTY SO	**Date**	09/08/2025	**Type**	Agency Inspection
Inspector	HIGGINBOTHAM,LEANDRA,	**Time**	13:14:41	**Software**	8100.27

0.05 Lot #	**0.08 Lot #**	**0.20 Lot #**	**0.08 Gas Lot #**
Expiration	**Expiration**	**Expiration**	**Expiration**

Number of Simulators **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

BYPASS AI FOR OPERATION; COMPLIANCE UNDETERMINED

Data Download Date/Time 09/08/2025 13:43

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

Agency PINELLAS COUNTY SO **Date** 03/05/2025 **Type** Department Inspection
Inspector GUTSCHOW,TAYLOR,D **Time** 13:26:06 **Software** 8100.27

0.05 Lot # 202303K **0.08 Lot #** 202303L **0.20 Lot #** 202304C **0.08 Gas Lot #** AG429602
Expiration 03/29/2025 **Expiration** 03/29/2025 **Expiration** 04/05/2025 **Expiration** 10/22/2026

Number of Simulators 5 **Average Standard Deviation** 0.0002 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.201
0.08 g/210L Dry Gas Std Test # 1	0.075
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.200
0.08 g/210L Dry Gas Std Test # 2	0.076
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.201
0.08 g/210L Dry Gas Std Test # 3	0.076
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.050
0.08 g/210L Test # 4	0.080
0.20 g/210L Test # 4	0.201
0.08 g/210L Dry Gas Std Test # 4	0.075
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.050
0.08 g/210L Test # 5	0.080
0.20 g/210L Test # 5	0.201
0.08 g/210L Dry Gas Std Test # 5	0.076
Interferent Detect Test # 5	INT

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.051
0.08 g/210L Test # 6	0.080
0.20 g/210L Test # 6	0.201
0.08 g/210L Dry Gas Std Test # 6	0.076
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.080
0.20 g/210L Test # 7	0.201
0.08 g/210L Dry Gas Std Test # 7	0.076
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.050
0.08 g/210L Test # 8	0.080
0.20 g/210L Test # 8	0.201
0.08 g/210L Dry Gas Std Test # 8	0.076
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.051
0.08 g/210L Test # 9	0.080
0.20 g/210L Test # 9	0.201
0.08 g/210L Dry Gas Std Test # 9	0.075
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.080
0.20 g/210L Test # 10	0.201
0.08 g/210L Dry Gas Std Test # 10	0.076
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0000
0.20 g/210L Test (Std Dev)	0.0003
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Data Download Date/Time 03/05/2025 13:24

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

Agency PINELLAS COUNTY SO **Date** 03/19/2025 **Type** Department Inspection
Inspector GUTSCHOW,TAYLOR,D **Time** 13:12:50 **Software** 8100.27

0.05 Lot # 202303K **0.08 Lot #** 202303L **0.20 Lot #** 202304C **0.08 Gas Lot #** AG429602
Expiration 03/29/2025 **Expiration** 03/29/2025 **Expiration** 04/05/2025 **Expiration** 10/22/2026

Number of Simulators 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.051
0.08 g/210L Test # 1	0.082
0.20 g/210L Test # 1	0.204
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.051
0.08 g/210L Test # 2	0.082
0.20 g/210L Test # 2	0.205
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.051
0.08 g/210L Test # 3	0.082
0.20 g/210L Test # 3	0.205
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.052
0.08 g/210L Test # 4	0.082
0.20 g/210L Test # 4	0.205
0.08 g/210L Dry Gas Std Test # 4	0.080
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.052
0.08 g/210L Test # 5	0.082
0.20 g/210L Test # 5	0.205
0.08 g/210L Dry Gas Std Test # 5	0.080
Interferent Detect Test # 5	INT

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.051
0.08 g/210L Test # 6	0.082
0.20 g/210L Test # 6	0.205
0.08 g/210L Dry Gas Std Test # 6	0.080
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.051
0.08 g/210L Test # 7	0.082
0.20 g/210L Test # 7	0.205
0.08 g/210L Dry Gas Std Test # 7	0.081
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.051
0.08 g/210L Test # 8	0.082
0.20 g/210L Test # 8	0.205
0.08 g/210L Dry Gas Std Test # 8	0.080
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.052
0.08 g/210L Test # 9	0.082
0.20 g/210L Test # 9	0.204
0.08 g/210L Dry Gas Std Test # 9	0.081
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.052
0.08 g/210L Test # 10	0.082
0.20 g/210L Test # 10	0.205
0.08 g/210L Dry Gas Std Test # 10	0.080
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0005
0.08 g/210L Test (Std Dev)	0.0000
0.20 g/210L Test (Std Dev)	0.0004
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Data Download Date/Time 03/19/2025 13:11

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

Agency PINELLAS COUNTY SO **Date** 03/26/2025 **Type** Agency Inspection
Inspector DEANE,M, **Time** 22:45:08 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** AG500903
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 01/09/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.200
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.081
0.20 g/210L Test # 2	0.204
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.051
0.08 g/210L Test # 3	0.082
0.20 g/210L Test # 3	0.206
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 03/26/2025 22:43

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

Agency PINELLAS COUNTY SO **Date** 04/23/2025 **Type** Agency Inspection
Inspector DEANE,M, **Time** 15:24:10 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** AG500903
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 01/09/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.081
0.20 g/210L Test # 1	0.203
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.051
0.08 g/210L Test # 2	0.082
0.20 g/210L Test # 2	0.205
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.051
0.08 g/210L Test # 3	0.083
0.20 g/210L Test # 3	0.205
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 04/23/2025 15:21

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

Agency PINELLAS COUNTY SO **Date** 05/22/2025 **Type** Agency Inspection
Inspector DEANE,M, **Time** 22:26:03 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** AG500903
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 01/09/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.200
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.081
0.20 g/210L Test # 2	0.204
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.051
0.08 g/210L Test # 3	0.081
0.20 g/210L Test # 3	0.206
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 05/22/2025 22:23

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

Agency PINELLAS COUNTY SO **Date** 06/18/2025 **Type** Agency Inspection
Inspector DEANE,M, **Time** 15:02:22 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** AG500903
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 01/09/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.081
0.20 g/210L Test # 1	0.202
0.08 g/210L Dry Gas Std Test # 1	0.082
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.051
0.08 g/210L Test # 2	0.082
0.20 g/210L Test # 2	0.205
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.051
0.08 g/210L Test # 3	0.082
0.20 g/210L Test # 3	0.205
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 06/18/2025 14:59

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

Agency PINELLAS COUNTY SO **Date** 07/09/2025 **Type** Agency Inspection
Inspector DEANE,M, **Time** 10:54:55 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** AG500903
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 01/09/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.204
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.081
0.20 g/210L Test # 2	0.205
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.081
0.20 g/210L Test # 3	0.205
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 07/09/2025 10:52

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

Agency PINELLAS COUNTY SO **Date** 08/13/2025 **Type** Agency Inspection
Inspector DEANE,M, **Time** 11:38:32 **Software** 8100.27

0.05 Lot # 202506E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** AG500903
Expiration 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 01/09/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.202
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.051
0.08 g/210L Test # 2	0.082
0.20 g/210L Test # 2	0.205
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.082
0.20 g/210L Test # 3	0.205
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 08/13/2025 11:37

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

Agency PINELLAS COUNTY SO **Date** 09/17/2025 **Type** Agency Inspection
Inspector DEANE,M, **Time** 12:07:20 **Software** 8100.27

0.05 Lot # 202506E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** AG500903
Expiration 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 01/09/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.202
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.082
0.20 g/210L Test # 2	0.205
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.051
0.08 g/210L Test # 3	0.081
0.20 g/210L Test # 3	0.205
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 09/17/2025 12:04

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

Agency PINELLAS COUNTY SO **Date** 10/15/2025 **Type** Agency Inspection
Inspector DEANE,M, **Time** 16:30:53 **Software** 8100.27

0.05 Lot # 202506E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** AG500903
Expiration 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 01/09/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.081
0.20 g/210L Test # 1	0.201
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.051
0.08 g/210L Test # 2	0.082
0.20 g/210L Test # 2	0.205
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.082
0.20 g/210L Test # 3	0.206
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 10/15/2025 16:26

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

Agency PINELLAS COUNTY SO **Date** 11/12/2025 **Type** Agency Inspection
Inspector DEANE,M, **Time** 10:46:48 **Software** 8100.27

0.05 Lot # 202506E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** AG500903
Expiration 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 01/09/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.051
0.08 g/210L Test # 1	0.081
0.20 g/210L Test # 1	0.202
0.08 g/210L Dry Gas Std Test # 1	0.082
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.051
0.08 g/210L Test # 2	0.082
0.20 g/210L Test # 2	0.205
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.051
0.08 g/210L Test # 3	0.082
0.20 g/210L Test # 3	0.205
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 11/12/2025 10:45

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

Agency PINELLAS COUNTY SO **Date** 12/17/2025 **Type** Agency Inspection
Inspector DEANE,M, **Time** 11:45:05 **Software** 8100.27

0.05 Lot # 202506E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** AG500903
Expiration 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 01/09/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.051
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.203
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.051
0.08 g/210L Test # 2	0.081
0.20 g/210L Test # 2	0.205
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.051
0.08 g/210L Test # 3	0.082
0.20 g/210L Test # 3	0.204
0.08 g/210L Dry Gas Std Test # 3	0.082
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 12/17/2025 11:43

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001368

Agency NAS MAYPORT **Date** 01/14/2025 **Type** Agency Inspection
Inspector POTTER,MICHAEL,R **Time** 14:19:08 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 15924080A2
Expiration 03/28/2025 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 07/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.196
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

IN SERV. NS MAYPORT

Data Download Date/Time 01/15/2025 09:09

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001368

Agency NAS MAYPORT **Date** 02/25/2025 **Type** Agency Inspection
Inspector LIGHTSEY,ROBERT,D **Time** 09:45:50 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 15924080A2
Expiration 03/28/2025 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 07/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

EVED USE AT M/P N/S

Data Download Date/Time 02/25/2025 09:47

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001368

Agency NAS MAYPORT **Date** 03/25/2025 **Type** Agency Inspection
Inspector POTTER,MICHAEL,R **Time** 14:49:54 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 15924080A2
Expiration 03/28/2025 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 07/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.082
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.082
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.082
Interferent Detect Test # 3	INT

Remarks:

IN SERV NS MAYPORT

Data Download Date/Time 03/25/2025 14:46

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001368

Agency NAS MAYPORT **Date** 04/29/2025 **Type** Agency Inspection
Inspector LIGHTSEY,ROBERT,D **Time** 18:01:11 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 15924080A2
Expiration 02/27/2026 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 07/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

EVED USE AT NS/MP

Data Download Date/Time 04/29/2025 17:57

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001368

Agency NAS MAYPORT **Date** 05/28/2025 **Type** Agency Inspection
Inspector POTTER,MICHAEL,R **Time** 14:14:41 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 15924080A2
Expiration 02/27/2026 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 07/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.051
0.08 g/210L Test # 1	0.081
0.20 g/210L Test # 1	0.200
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.051
0.08 g/210L Test # 2	0.081
0.20 g/210L Test # 2	0.200
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.051
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

REMOVED FROM SERV NS MAYPORT FOR DEPT INSP @ FDLE 5/28/25

Data Download Date/Time 05/28/2025 14:10

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001368

Agency NAS MAYPORT **Date** 06/13/2025 **Type** Department Inspection
Inspector HIGGINBOTHAM,LEANDRA, **Time** 15:16:21 **Software** 8100.27

0.05 Lot # 202406K **0.08 Lot #** 202406L **0.20 Lot #** 202406N **0.08 Gas Lot #** AG429602
Expiration 06/19/2026 **Expiration** 06/19/2026 **Expiration** 06/20/2026 **Expiration** 10/22/2026

Number of Simulators 5 **Average Standard Deviation** 0.0006 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.045
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.201
0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.200
0.08 g/210L Dry Gas Std Test # 2	0.077
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.046
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.201
0.08 g/210L Dry Gas Std Test # 3	0.078
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.048
0.08 g/210L Test # 4	0.080
0.20 g/210L Test # 4	0.201
0.08 g/210L Dry Gas Std Test # 4	0.078
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.046
0.08 g/210L Test # 5	0.080
0.20 g/210L Test # 5	0.202
0.08 g/210L Dry Gas Std Test # 5	0.078
Interferent Detect Test # 5	INT

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001368

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.048
0.08 g/210L Test # 6	0.080
0.20 g/210L Test # 6	0.201
0.08 g/210L Dry Gas Std Test # 6	0.078
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.048
0.08 g/210L Test # 7	0.079
0.20 g/210L Test # 7	0.201
0.08 g/210L Dry Gas Std Test # 7	0.078
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.047
0.08 g/210L Test # 8	0.080
0.20 g/210L Test # 8	0.201
0.08 g/210L Dry Gas Std Test # 8	0.077
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.047
0.08 g/210L Test # 9	0.080
0.20 g/210L Test # 9	0.201
0.08 g/210L Dry Gas Std Test # 9	0.078
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.048
0.08 g/210L Test # 10	0.080
0.20 g/210L Test # 10	0.202
0.08 g/210L Dry Gas Std Test # 10	0.078
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0011
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0005
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Data Download Date/Time 06/13/2025 16:13

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001368

Agency NAS MAYPORT **Date** 07/09/2025 **Type** Agency Inspection
Inspector POTTER,MICHAEL,R **Time** 12:23:32 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 13225080A2
Expiration 02/27/2026 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 06/05/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

IN SERV @ NS MAYPORT PD

Data Download Date/Time 07/09/2025 12:19

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001368

Agency NAS MAYPORT **Date** 08/28/2025 **Type** Agency Inspection
Inspector POTTER,MICHAEL,R **Time** 13:05:46 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 13225080A2
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 06/05/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.051
0.08 g/210L Test # 1	0.081
0.20 g/210L Test # 1	0.202
0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.051
0.08 g/210L Test # 2	0.082
0.20 g/210L Test # 2	0.202
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.051
0.08 g/210L Test # 3	0.082
0.20 g/210L Test # 3	0.201
0.08 g/210L Dry Gas Std Test # 3	0.078
Interferent Detect Test # 3	INT

Remarks:

IN SERV NS MAYPORT PD

Data Download Date/Time 08/28/2025 13:00

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001368

Agency NAS MAYPORT **Date** 09/26/2025 **Type** Agency Inspection
Inspector POTTER,MICHAEL,R **Time** 09:50:50 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 13225080A2
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 06/05/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.081
0.20 g/210L Test # 1	0.203
0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.200
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

IN SERV # NS MAYPORT

Data Download Date/Time 09/26/2025 09:45

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001368

Agency NAS MAYPORT **Date** 10/31/2025 **Type** Agency Inspection
Inspector LIGHTSEY,ROBERT,D **Time** 11:35:29 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 13225080A2
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 06/05/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.196
0.08 g/210L Dry Gas Std Test # 1	0.077
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.077
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.193
0.08 g/210L Dry Gas Std Test # 3	0.077
Interferent Detect Test # 3	INT

Remarks:

EVED USE @ MAY NS

Data Download Date/Time 10/31/2025 14:31

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001368

Agency NAS MAYPORT **Date** 11/25/2025 **Type** Agency Inspection
Inspector POTTER,MICHAEL,R **Time** 10:56:29 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 13225080A2
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 06/05/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.081
0.20 g/210L Test # 1	0.202
0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.051
0.08 g/210L Test # 2	0.082
0.20 g/210L Test # 2	0.201
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.051
0.08 g/210L Test # 3	0.082
0.20 g/210L Test # 3	0.201
0.08 g/210L Dry Gas Std Test # 3	0.078
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 11/25/2025 11:30

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001368

Agency NAS MAYPORT **Date** 12/16/2025 **Type** Agency Inspection
Inspector POTTER,MICHAEL,R **Time** 09:50:47 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 13225080A2
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 06/05/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.081
0.20 g/210L Test # 1	0.204
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.081
0.20 g/210L Test # 2	0.201
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.081
0.20 g/210L Test # 3	0.202
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

IN SERV @ NS MAYPORT PD

Data Download Date/Time 12/16/2025 09:57

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001415

Agency	PINELLAS COUNTY SO	**Date**	01/16/2025	**Type**	Agency Inspection
Inspector	GUTSCHOW,TAYLOR,D	**Time**	11:58:43	**Software**	8100.27

0.05 Lot #	**0.08 Lot #**	**0.20 Lot #**	**0.08 Gas Lot #**
Expiration	**Expiration**	**Expiration**	**Expiration**

Number of Simulators **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

AI NOT CONDUCTED. BYPASSED TO BRING OUT OF DISABLED MODE.

Data Download Date/Time 01/24/2025 10:03

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001415

Agency PINELLAS COUNTY SO **Date** 01/17/2025 **Type** Department Inspection
Inspector GUTSCHOW,TAYLOR,D **Time** 13:36:28 **Software** 8100.27

0.05 Lot # 202303K **0.08 Lot #** 202303L **0.20 Lot #** 202304C **0.08 Gas Lot #** AG429602
Expiration 03/29/2025 **Expiration** 03/29/2025 **Expiration** 04/05/2025 **Expiration** 10/22/2026

Number of Simulators 5 **Average Standard Deviation** 0.0002 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.078
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.050
0.08 g/210L Test # 4	0.079
0.20 g/210L Test # 4	0.199
0.08 g/210L Dry Gas Std Test # 4	0.078
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.050
0.08 g/210L Test # 5	0.079
0.20 g/210L Test # 5	0.199
0.08 g/210L Dry Gas Std Test # 5	0.078
Interferent Detect Test # 5	INT

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001415

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.050
0.08 g/210L Test # 6	0.079
0.20 g/210L Test # 6	0.199
0.08 g/210L Dry Gas Std Test # 6	0.078
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.079
0.20 g/210L Test # 7	0.199
0.08 g/210L Dry Gas Std Test # 7	0.078
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.050
0.08 g/210L Test # 8	0.079
0.20 g/210L Test # 8	0.199
0.08 g/210L Dry Gas Std Test # 8	0.078
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.050
0.08 g/210L Test # 9	0.080
0.20 g/210L Test # 9	0.199
0.08 g/210L Dry Gas Std Test # 9	0.078
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.051
0.08 g/210L Test # 10	0.079
0.20 g/210L Test # 10	0.199
0.08 g/210L Dry Gas Std Test # 10	0.078
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0003
0.20 g/210L Test (Std Dev)	0.0000
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

Data Download Date/Time 01/24/2025 10:03

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001415

Agency	PINELLAS COUNTY SO	**Date**	02/19/2025	**Type**	Agency Inspection
Inspector	DEANE,M,	**Time**	15:16:17	**Software**	8100.27

0.05 Lot #	202402E	**0.08 Lot #**	202307A	**0.20 Lot #**	202307C	**0.08 Gas Lot #**	AG409204
Expiration	02/27/2026	**Expiration**	07/11/2025	**Expiration**	07/12/2025	**Expiration**	04/01/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.047
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.195
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/19/2025 15:16

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001415

Agency PINELLAS COUNTY SO **Date** 03/19/2025 **Type** Agency Inspection
Inspector DEANE,M, **Time** 16:06:07 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** AG409204
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 04/01/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.200
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.201
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 03/19/2025 16:06

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001415

Agency PINELLAS COUNTY SO **Date** 04/24/2025 **Type** Agency Inspection
Inspector DEANE,M, **Time** 22:42:17 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** AG409204
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 04/01/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.047
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.196
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 04/24/2025 22:43

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

Agency ORANGE COUNTY SO **Date** 01/14/2025 **Type** Agency Inspection
Inspector MELVILLE,KELLY,M **Time** 11:37:52 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202307C **0.08 Gas Lot #** 28423080A4
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 07/12/2025 **Expiration** 11/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.078
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/14/2025 13:17

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

Agency ORANGE COUNTY SO **Date** 02/05/2025 **Type** Department Inspection
Inspector ARMSTRONG,DESTINEE,N **Time** 13:12:31 **Software** 8100.27

0.05 Lot # 202406K **0.08 Lot #** 202406L **0.20 Lot #** 202406N **0.08 Gas Lot #** AG429602
Expiration 06/19/2026 **Expiration** 06/19/2026 **Expiration** 06/20/2026 **Expiration** 10/22/2026

Number of Simulators 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.202
0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.202
0.08 g/210L Dry Gas Std Test # 2	0.077
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.202
0.08 g/210L Dry Gas Std Test # 3	0.077
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.049
0.08 g/210L Test # 4	0.079
0.20 g/210L Test # 4	0.201
0.08 g/210L Dry Gas Std Test # 4	0.078
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.049
0.08 g/210L Test # 5	0.080
0.20 g/210L Test # 5	0.202
0.08 g/210L Dry Gas Std Test # 5	0.078
Interferent Detect Test # 5	INT

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.049
0.08 g/210L Test # 6	0.080
0.20 g/210L Test # 6	0.202
0.08 g/210L Dry Gas Std Test # 6	0.077
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.049
0.08 g/210L Test # 7	0.079
0.20 g/210L Test # 7	0.201
0.08 g/210L Dry Gas Std Test # 7	0.077
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.049
0.08 g/210L Test # 8	0.080
0.20 g/210L Test # 8	0.201
0.08 g/210L Dry Gas Std Test # 8	0.078
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.049
0.08 g/210L Test # 9	0.080
0.20 g/210L Test # 9	0.202
0.08 g/210L Dry Gas Std Test # 9	0.078
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.049
0.08 g/210L Test # 10	0.080
0.20 g/210L Test # 10	0.202
0.08 g/210L Dry Gas Std Test # 10	0.077
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0005
0.20 g/210L Test (Std Dev)	0.0004
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

Data Download Date/Time 02/05/2025 13:11

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

Agency ORANGE COUNTY SO **Date** 02/18/2025 **Type** Agency Inspection
Inspector MELVILLE,KELLY,M **Time** 10:16:07 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202307C **0.08 Gas Lot #** 15924080A2
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 07/12/2025 **Expiration** 07/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.081
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.081
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/18/2025 10:17

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

Agency ORANGE COUNTY SO **Date** 03/20/2025 **Type** Agency Inspection
Inspector MELVILLE,KELLY,M **Time** 10:04:25 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202307C **0.08 Gas Lot #** 15924080A2
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 07/12/2025 **Expiration** 07/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.081
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 03/20/2025 10:19

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

Agency ORANGE COUNTY SO **Date** 04/22/2025 **Type** Agency Inspection
Inspector MELVILLE,KELLY,M **Time** 10:42:10 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202307C **0.08 Gas Lot #** 15924080A2
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 07/12/2025 **Expiration** 07/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.200
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 04/22/2025 10:56

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

Agency ORANGE COUNTY SO **Date** 05/14/2025 **Type** Agency Inspection
Inspector MELVILLE,KELLY,M **Time** 10:12:18 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202307C **0.08 Gas Lot #** 15924080A2
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 07/12/2025 **Expiration** 07/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 05/14/2025 11:34

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

Agency ORANGE COUNTY SO **Date** 06/25/2025 **Type** Agency Inspection
Inspector MELVILLE,KELLY,M **Time** 10:27:45 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202307C **0.08 Gas Lot #** 15924080A2
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 07/12/2025 **Expiration** 07/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.200
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.200
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 06/25/2025 11:02

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

Agency ORANGE COUNTY SO **Date** 07/16/2025 **Type** Agency Inspection
Inspector MELVILLE,KELLY,M **Time** 12:27:44 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 15924080A2
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 07/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.201
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.203
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.202
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 07/16/2025 12:39

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

Agency ORANGE COUNTY SO **Date** 08/13/2025 **Type** Agency Inspection
Inspector MELVILLE,KELLY,M **Time** 09:59:30 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 15924080A2
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 07/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.200
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.201
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.202
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 08/13/2025 10:00

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

Agency ORANGE COUNTY SO **Date** 09/01/2025 **Type** Agency Inspection
Inspector MELVILLE,KELLY,M **Time** 10:16:21 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 15924080A2
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 07/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.201
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.202
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.081
0.20 g/210L Test # 3	0.202
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 09/01/2025 10:30

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

Agency ORANGE COUNTY SO **Date** 10/16/2025 **Type** Agency Inspection
Inspector MELVILLE,KELLY,M **Time** 12:42:38 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 15924080A2
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 07/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.201
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.202
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.202
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 10/16/2025 12:40

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

Agency ORANGE COUNTY SO **Date** 11/11/2025 **Type** Agency Inspection
Inspector MELVILLE,KELLY,M **Time** 09:59:28 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 15924080A2
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 07/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.081
0.20 g/210L Test # 1	0.201
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.081
0.20 g/210L Test # 2	0.202
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.081
0.20 g/210L Test # 3	0.203
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 11/11/2025 10:01

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

Agency ORANGE COUNTY SO **Date** 12/10/2025 **Type** Agency Inspection
Inspector MELVILLE,KELLY,M **Time** 12:54:40 **Software** 8100.27

0.05 Lot # 202506E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 17725080A1
Expiration 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 08/05/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.200
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.202
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.203
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 12/10/2025 12:58

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

Agency ORANGE COUNTY SO **Date** 01/14/2025 **Type** Agency Inspection
Inspector MELVILLE,KELLY,M **Time** 11:42:13 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202307C **0.08 Gas Lot #** 24923080A2
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 07/12/2025 **Expiration** 10/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.083
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.081
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.083
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.081
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.082
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/14/2025 13:20

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

Agency ORANGE COUNTY SO **Date** 02/18/2025 **Type** Agency Inspection
Inspector MELVILLE,KELLY,M **Time** 10:11:00 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202307C **0.08 Gas Lot #** 24923080A2
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 07/12/2025 **Expiration** 10/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.195
0.08 g/210L Dry Gas Std Test # 1	0.083
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.195
0.08 g/210L Dry Gas Std Test # 2	0.083
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.196
0.08 g/210L Dry Gas Std Test # 3	0.083
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/18/2025 10:11

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

Agency ORANGE COUNTY SO **Date** 03/20/2025 **Type** Agency Inspection
Inspector MELVILLE,KELLY,M **Time** 12:16:27 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202307C **0.08 Gas Lot #** 15924080A2
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 07/12/2025 **Expiration** 07/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.196
0.08 g/210L Dry Gas Std Test # 1	0.084
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.083
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.083
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 03/20/2025 12:19

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

Agency ORANGE COUNTY SO **Date** 04/22/2025 **Type** Agency Inspection
Inspector MELVILLE,KELLY,M **Time** 13:25:43 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202307C **0.08 Gas Lot #** 15924080A2
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 07/12/2025 **Expiration** 07/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.195
0.08 g/210L Dry Gas Std Test # 1	0.083
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.196
0.08 g/210L Dry Gas Std Test # 2	0.082
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.082
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 04/22/2025 15:05

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

Agency ORANGE COUNTY SO **Date** 05/12/2025 **Type** Department Inspection
Inspector PLATT,SHAYLA,D **Time** 11:59:00 **Software** 8100.27

0.05 Lot # 202406K **0.08 Lot #** 202406L **0.20 Lot #** 202406N **0.08 Gas Lot #** AG429602
Expiration 06/19/2026 **Expiration** 06/19/2026 **Expiration** 06/20/2026 **Expiration** 10/22/2026

Number of Simulators 5 **Average Standard Deviation** 0.0005 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.200
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.049
0.08 g/210L Test # 4	0.080
0.20 g/210L Test # 4	0.199
0.08 g/210L Dry Gas Std Test # 4	0.081
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.049
0.08 g/210L Test # 5	0.080
0.20 g/210L Test # 5	0.200
0.08 g/210L Dry Gas Std Test # 5	0.080
Interferent Detect Test # 5	INT

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.049
0.08 g/210L Test # 6	0.080
0.20 g/210L Test # 6	0.200
0.08 g/210L Dry Gas Std Test # 6	0.080
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.049
0.08 g/210L Test # 7	0.081
0.20 g/210L Test # 7	0.200
0.08 g/210L Dry Gas Std Test # 7	0.081
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.050
0.08 g/210L Test # 8	0.081
0.20 g/210L Test # 8	0.200
0.08 g/210L Dry Gas Std Test # 8	0.081
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.050
0.08 g/210L Test # 9	0.081
0.20 g/210L Test # 9	0.200
0.08 g/210L Dry Gas Std Test # 9	0.081
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.081
0.20 g/210L Test # 10	0.201
0.08 g/210L Dry Gas Std Test # 10	0.080
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0006
0.20 g/210L Test (Std Dev)	0.0006
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

Data Download Date/Time 05/12/2025 11:56

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

Agency ORANGE COUNTY SO **Date** 06/25/2025 **Type** Agency Inspection
Inspector MELVILLE,KELLY,M **Time** 11:59:45 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202307C **0.08 Gas Lot #** 00925080A3
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 07/12/2025 **Expiration** 02/05/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.082
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.082
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.081
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.082
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 06/25/2025 12:03

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

Agency ORANGE COUNTY SO **Date** 07/16/2025 **Type** Agency Inspection
Inspector MELVILLE,KELLY,M **Time** 11:36:57 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 00925080A3
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 02/05/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.082
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.051
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 07/16/2025 14:09

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

Agency ORANGE COUNTY SO **Date** 08/13/2025 **Type** Agency Inspection
Inspector MELVILLE,KELLY,M **Time** 11:49:05 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 00925080A3
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 02/05/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.200
0.08 g/210L Dry Gas Std Test # 1	0.082
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.201
0.08 g/210L Dry Gas Std Test # 2	0.082
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.201
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 08/13/2025 12:38

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

Agency ORANGE COUNTY SO **Date** 09/01/2025 **Type** Agency Inspection
Inspector MELVILLE,KELLY,M **Time** 10:11:17 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 00925080A3
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 02/05/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.083
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.082
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.083
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 09/01/2025 10:36

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

Agency ORANGE COUNTY SO **Date** 10/16/2025 **Type** Agency Inspection
Inspector MELVILLE,KELLY,M **Time** 10:24:00 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 00925080A3
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 02/05/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.083
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.083
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.083
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 10/16/2025 10:26

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

Agency ORANGE COUNTY SO **Date** 11/11/2025 **Type** Agency Inspection
Inspector MELVILLE,KELLY,M **Time** 09:57:00 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 17725080A1
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 08/05/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.084
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.084
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.051
0.08 g/210L Test # 3	0.081
0.20 g/210L Test # 3	0.200
0.08 g/210L Dry Gas Std Test # 3	0.084
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 11/11/2025 09:56

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

Agency ORANGE COUNTY SO **Date** 12/10/2025 **Type** Agency Inspection
Inspector MELVILLE,KELLY,M **Time** 10:34:20 **Software** 8100.27

0.05 Lot # 202506E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 17725080A1
Expiration 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 08/05/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.196
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 12/10/2025 10:50

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

Agency ORANGE COUNTY SO **Date** 12/30/2025 **Type** Department Inspection
Inspector HIGGINBOTHAM,LEANDRA, **Time** 13:53:13 **Software** 8100.27

0.05 Lot # 202406K **Expiration** 06/19/2026
0.08 Lot # 202406L **Expiration** 06/19/2026
0.20 Lot # 202406N **Expiration** 06/20/2026
0.08 Gas Lot # AG510701 **Expiration** 04/17/2027

Number of Simulators 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.081
0.20 g/210L Test # 1	0.202
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.202
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.081
0.20 g/210L Test # 3	0.201
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.050
0.08 g/210L Test # 4	0.080
0.20 g/210L Test # 4	0.201
0.08 g/210L Dry Gas Std Test # 4	0.081
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.050
0.08 g/210L Test # 5	0.080
0.20 g/210L Test # 5	0.200
0.08 g/210L Dry Gas Std Test # 5	0.081
Interferent Detect Test # 5	INT

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.050
0.08 g/210L Test # 6	0.080
0.20 g/210L Test # 6	0.201
0.08 g/210L Dry Gas Std Test # 6	0.082
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.080
0.20 g/210L Test # 7	0.200
0.08 g/210L Dry Gas Std Test # 7	0.082
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.050
0.08 g/210L Test # 8	0.081
0.20 g/210L Test # 8	0.201
0.08 g/210L Dry Gas Std Test # 8	0.082
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.050
0.08 g/210L Test # 9	0.081
0.20 g/210L Test # 9	0.200
0.08 g/210L Dry Gas Std Test # 9	0.082
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.081
0.20 g/210L Test # 10	0.201
0.08 g/210L Dry Gas Std Test # 10	0.082
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0000
0.08 g/210L Test (Std Dev)	0.0005
0.20 g/210L Test (Std Dev)	0.0007
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

Data Download Date/Time 12/30/2025 13:50

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

Agency ORANGE COUNTY SO **Date** 01/14/2025 **Type** Agency Inspection
Inspector MELVILLE,KELLY,M **Time** 10:21:20 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202307C **0.08 Gas Lot #** 15924080A2
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 07/12/2025 **Expiration** 07/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.196
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.196
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.196
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/14/2025 10:41

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

Agency ORANGE COUNTY SO **Date** 02/18/2025 **Type** Agency Inspection
Inspector MELVILLE,KELLY,M **Time** 11:28:42 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202307C **0.08 Gas Lot #** 15924080A2
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 07/12/2025 **Expiration** 07/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.195
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.195
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.195
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/18/2025 11:31

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

Agency ORANGE COUNTY SO **Date** 03/20/2025 **Type** Agency Inspection
Inspector MELVILLE,KELLY,M **Time** 11:13:05 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202307C **0.08 Gas Lot #** 15924080A2
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 07/12/2025 **Expiration** 07/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.193
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.195
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.195
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 03/20/2025 12:26

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

Agency ORANGE COUNTY SO **Date** 04/22/2025 **Type** Agency Inspection
Inspector MELVILLE,KELLY,M **Time** 12:25:36 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202307C **0.08 Gas Lot #** 15924080A2
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 07/12/2025 **Expiration** 07/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.194
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.194
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.194
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 04/22/2025 12:27

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

Agency ORANGE COUNTY SO **Date** 05/14/2025 **Type** Agency Inspection
Inspector MELVILLE,KELLY,M **Time** 11:54:30 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202307C **0.08 Gas Lot #** 15924080A2
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 07/12/2025 **Expiration** 07/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.195
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.195
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.195
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 05/14/2025 12:39

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

Agency ORANGE COUNTY SO **Date** 06/03/2025 **Type** Department Inspection
Inspector HIGGINBOTHAM,LEANDRA, **Time** 14:41:04 **Software** 8100.27

0.05 Lot # 202406K **0.08 Lot #** 202406L **0.20 Lot #** 202406N **0.08 Gas Lot #** AG429602
Expiration 06/19/2026 **Expiration** 06/19/2026 **Expiration** 06/20/2026 **Expiration** 10/22/2026

Number of Simulators 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.196
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.196
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.049
0.08 g/210L Test # 4	0.079
0.20 g/210L Test # 4	0.197
0.08 g/210L Dry Gas Std Test # 4	0.080
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.049
0.08 g/210L Test # 5	0.079
0.20 g/210L Test # 5	0.196
0.08 g/210L Dry Gas Std Test # 5	0.080
Interferent Detect Test # 5	INT

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.049
0.08 g/210L Test # 6	0.079
0.20 g/210L Test # 6	0.198
0.08 g/210L Dry Gas Std Test # 6	0.080
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.049
0.08 g/210L Test # 7	0.079
0.20 g/210L Test # 7	0.197
0.08 g/210L Dry Gas Std Test # 7	0.080
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.049
0.08 g/210L Test # 8	0.079
0.20 g/210L Test # 8	0.197
0.08 g/210L Dry Gas Std Test # 8	0.080
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.049
0.08 g/210L Test # 9	0.079
0.20 g/210L Test # 9	0.197
0.08 g/210L Dry Gas Std Test # 9	0.080
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.049
0.08 g/210L Test # 10	0.079
0.20 g/210L Test # 10	0.197
0.08 g/210L Dry Gas Std Test # 10	0.080
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0003
0.08 g/210L Test (Std Dev)	0.0003
0.20 g/210L Test (Std Dev)	0.0006
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

Data Download Date/Time 06/03/2025 14:45

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

Agency ORANGE COUNTY SO **Date** 06/25/2025 **Type** Agency Inspection
Inspector MELVILLE,KELLY,M **Time** 11:54:38 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202307C **0.08 Gas Lot #** 00925080A3
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 07/12/2025 **Expiration** 02/05/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.195
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.196
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.196
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 06/25/2025 11:54

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

Agency ORANGE COUNTY SO **Date** 07/16/2025 **Type** Agency Inspection
Inspector MELVILLE,KELLY,M **Time** 11:34:02 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 00925080A3
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 02/05/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.195
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.196
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.196
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 07/16/2025 13:46

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

Agency ORANGE COUNTY SO **Date** 08/13/2025 **Type** Agency Inspection
Inspector MELVILLE,KELLY,M **Time** 11:40:44 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 00925080A3
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 02/05/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 08/13/2025 12:21

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

Agency ORANGE COUNTY SO **Date** 09/01/2025 **Type** Agency Inspection
Inspector MELVILLE,KELLY,M **Time** 10:05:22 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 00925080A3
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 02/05/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.195
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.195
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.196
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 09/01/2025 10:41

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

Agency	ORANGE COUNTY SO	**Date**	10/16/2025	**Type**	Agency Inspection
Inspector	MELVILLE,KELLY,M	**Time**	10:31:35	**Software**	8100.27

0.05 Lot #	202402E	**0.08 Lot #**	202410E	**0.20 Lot #**	202410F	**0.08 Gas Lot #**	00925080A3
Expiration	02/27/2026	**Expiration**	10/22/2026	**Expiration**	10/30/2026	**Expiration**	02/05/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.196
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 10/16/2025 10:29

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

Agency	ORANGE COUNTY SO	**Date**	11/11/2025	**Type**	Agency Inspection
Inspector	MELVILLE,KELLY,M	**Time**	10:05:39	**Software**	8100.27

0.05 Lot #	202402E	**0.08 Lot #**	202410E	**0.20 Lot #**	202410F	**0.08 Gas Lot #**	00925080A3
Expiration	02/27/2026	**Expiration**	10/22/2026	**Expiration**	10/30/2026	**Expiration**	02/05/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.082
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 11/11/2025 10:05

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

Agency	ORANGE COUNTY SO	**Date**	12/10/2025	**Type**	Agency Inspection
Inspector	MELVILLE,KELLY,M	**Time**	12:00:15	**Software**	8100.27

0.05 Lot #	202506E	**0.08 Lot #**	202410E	**0.20 Lot #**	202410F	**0.08 Gas Lot #**	00925080A3
Expiration	06/24/2027	**Expiration**	10/22/2026	**Expiration**	10/30/2026	**Expiration**	02/05/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.196
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.196
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 12/10/2025 11:59

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001439

Agency MARTIN COUNTY SO **Date** 01/27/2025 **Type** Agency Inspection
Inspector THOMAS,CHRISTOPHER,L **Time** 22:13:14 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202307C **0.08 Gas Lot #** 08223080A5
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 07/12/2025 **Expiration** 06/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.078
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/27/2025 22:14

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001439

Agency	MARTIN COUNTY SO	**Date**	02/02/2025	**Type**	Agency Inspection
Inspector	THOMAS,CHRISTOPHER,L	**Time**	23:03:13	**Software**	8100.27

0.05 Lot #	202402E	**0.08 Lot #**	202410E	**0.20 Lot #**	202307C	**0.08 Gas Lot #**	08223080A5
Expiration	02/27/2026	**Expiration**	10/22/2026	**Expiration**	07/12/2025	**Expiration**	06/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/02/2025 23:02

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001439

Agency	MARTIN COUNTY SO	**Date**	03/10/2025	**Type**	Agency Inspection
Inspector	THOMAS,CHRISTOPHER,L	**Time**	21:58:55	**Software**	8100.27

0.05 Lot #	202402E	**0.08 Lot #**	202410E	**0.20 Lot #**	202307C	**0.08 Gas Lot #**	08223080A5
Expiration	02/27/2026	**Expiration**	10/22/2026	**Expiration**	07/12/2025	**Expiration**	06/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.078
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 03/10/2025 21:59

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001439

Agency	MARTIN COUNTY SO	**Date**	04/22/2025	**Type**	Agency Inspection
Inspector	THOMAS,CHRISTOPHER,L	**Time**	21:21:20	**Software**	8100.27

0.05 Lot #	202402E	**0.08 Lot #**	202410E	**0.20 Lot #**	202307C	**0.08 Gas Lot #**	08223080A5
Expiration	02/27/2026	**Expiration**	10/22/2026	**Expiration**	07/12/2025	**Expiration**	06/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 04/22/2025 21:21

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001439

Agency MARTIN COUNTY SO **Date** 05/15/2025 **Type** Agency Inspection
Inspector THOMAS,CHRISTOPHER,L **Time** 21:11:36 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 31124080A2
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 12/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.078
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 05/15/2025 21:20

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001439

Agency MARTIN COUNTY SO **Date** 06/22/2025 **Type** Agency Inspection
Inspector THOMAS,CHRISTOPHER,L **Time** 19:00:57 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 31124080A2
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 12/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.200
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.200
0.08 g/210L Dry Gas Std Test # 3	0.078
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 06/22/2025 19:00

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001439

Agency MARTIN COUNTY SO **Date** 07/24/2025 **Type** Agency Inspection
Inspector THOMAS,CHRISTOPHER,L **Time** 23:37:36 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 31124080A2
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 12/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.078
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 07/24/2025 23:37

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001439

Agency MARTIN COUNTY SO **Date** 08/26/2025 **Type** Department Inspection
Inspector GUTSCHOW,TAYLOR,D **Time** 12:50:20 **Software** 8100.27

0.05 Lot # 202406K **0.08 Lot #** 202406L **0.20 Lot #** 202406N **0.08 Gas Lot #** AG429602
Expiration 06/19/2026 **Expiration** 06/19/2026 **Expiration** 06/20/2026 **Expiration** 10/22/2026

Number of Simulators 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.077
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.077
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.077
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.048
0.08 g/210L Test # 4	0.078
0.20 g/210L Test # 4	0.198
0.08 g/210L Dry Gas Std Test # 4	0.077
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.048
0.08 g/210L Test # 5	0.078
0.20 g/210L Test # 5	0.198
0.08 g/210L Dry Gas Std Test # 5	0.077
Interferent Detect Test # 5	INT

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001439

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.048
0.08 g/210L Test # 6	0.078
0.20 g/210L Test # 6	0.199
0.08 g/210L Dry Gas Std Test # 6	0.077
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.049
0.08 g/210L Test # 7	0.078
0.20 g/210L Test # 7	0.199
0.08 g/210L Dry Gas Std Test # 7	0.078
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.048
0.08 g/210L Test # 8	0.078
0.20 g/210L Test # 8	0.199
0.08 g/210L Dry Gas Std Test # 8	0.077
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.048
0.08 g/210L Test # 9	0.078
0.20 g/210L Test # 9	0.198
0.08 g/210L Dry Gas Std Test # 9	0.077
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.048
0.08 g/210L Test # 10	0.078
0.20 g/210L Test # 10	0.199
0.08 g/210L Dry Gas Std Test # 10	0.078
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0003
0.20 g/210L Test (Std Dev)	0.0005
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Data Download Date/Time 08/26/2025 12:59

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001439

Agency MARTIN COUNTY SO **Date** 09/18/2025 **Type** Agency Inspection
Inspector THOMAS,CHRISTOPHER,L **Time** 00:21:19 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 31124080A2
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 12/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.078
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 09/18/2025 00:19

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001439

Agency MARTIN COUNTY SO **Date** 10/15/2025 **Type** Agency Inspection
Inspector THOMAS,CHRISTOPHER,L **Time** 22:48:01 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 31124080A2
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 12/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.078
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 10/15/2025 22:48

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001439

Agency MARTIN COUNTY SO **Date** 11/12/2025 **Type** Agency Inspection
Inspector THOMAS,CHRISTOPHER,L **Time** 21:36:14 **Software** 8100.27

0.05 Lot # 202506E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 31124080A2
Expiration 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 12/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # (Repeat) 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test # 1	INT
Alcohol Free Test # (Repeat) 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # (Repeat) 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.078
Interferent Detect Test # 3	INT

Remarks:

00: Ambient Fail/ LET ROOM AIR OUT .

Data Download Date/Time 11/12/2025 21:34

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001439

Agency MARTIN COUNTY SO **Date** 12/10/2025 **Type** Agency Inspection
Inspector THOMAS,CHRISTOPHER,L **Time** 22:01:37 **Software** 8100.27

0.05 Lot # 202506E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 31124080A2
Expiration 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 12/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.200
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.200
0.08 g/210L Dry Gas Std Test # 3	0.078
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 12/10/2025 22:02

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001442

Agency BISCAYNE NAT PARK **Date** 01/24/2025 **Type** Agency Inspection
Inspector SAAVEDRA,OSCAR, **Time** 12:43:48 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 402987895
Expiration 02/27/2026 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 03/05/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.047
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.047
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/24/2025 12:43

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001442

Agency BISCAYNE NAT PARK **Date** 02/26/2025 **Type** Agency Inspection
Inspector GORNEY,WILLIAM,M **Time** 10:58:05 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 402987895
Expiration 02/27/2026 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 03/05/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.047
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.078
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/26/2025 10:59

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001442

Agency BISCAYNE NAT PARK **Date** 03/18/2025 **Type** Agency Inspection
Inspector GORNEY,WILLIAM,M **Time** 11:48:41 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 402987895
Expiration 02/27/2026 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 03/05/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.047
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.047
0.08 g/210L Test # 2	0.077
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.077
0.20 g/210L Test # 3	0.200
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 03/18/2025 11:50

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001442

Agency BISCAYNE NAT PARK **Date** 04/08/2025 **Type** Agency Inspection
Inspector GORNEY,WILLIAM,M **Time** 11:18:39 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 402987895
Expiration 02/27/2026 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 03/05/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.046
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.047
0.08 g/210L Test # 2	0.077
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.047
0.08 g/210L Test # 3	0.077
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 04/08/2025 11:20

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001442

Agency BISCAYNE NAT PARK **Date** 05/06/2025 **Type** Agency Inspection
Inspector SAAVEDRA,OSCAR, **Time** 13:14:29 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 402987895
Expiration 02/27/2026 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 03/05/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

A F / M A: No Sample Provided.

Data Download Date/Time 05/06/2025 13:15

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001442

Agency BISCAYNE NAT PARK **Date** 06/27/2025 **Type** Agency Inspection
Inspector GORNEY,WILLIAM,M **Time** 13:54:55 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 402987895
Expiration 02/27/2026 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 03/05/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.047
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.077
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 06/27/2025 13:57

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001442

Agency	BISCAYNE NAT PARK	**Date**	12/22/2025	**Type**	Agency Inspection
Inspector	PIERSON,WEN-CHI,K	**Time**	14:33:20	**Software**	8100.27

0.05 Lot #	**0.08 Lot #**	**0.20 Lot #**	**0.08 Gas Lot #**
Expiration	**Expiration**	**Expiration**	**Expiration**

Number of Simulators **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

BYPASS AI TO OPERATE INSTRUMENT. COMPLIANCE NOT DETERMINED.

Data Download Date/Time 12/23/2025 16:08

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001442

Agency BISCAYNE NAT PARK **Date** 12/23/2025 **Type** Department Inspection
Inspector HIGGINBOTHAM,LEANDRA, **Time** 16:12:11 **Software** 8100.27

0.05 Lot # 202406K **0.08 Lot #** 202406L **0.20 Lot #** 202406N **0.08 Gas Lot #** AG510701
Expiration 06/19/2026 **Expiration** 06/19/2026 **Expiration** 06/20/2026 **Expiration** 04/17/2027

Number of Simulators 5 **Average Standard Deviation** 0.0005 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
Alcohol Free Test # (Repeat) 1	0.000
0.05 g/210L Test # 1	0.045
0.08 g/210L Test # 1	0.075
0.20 g/210L Test # 1	0.195
0.08 g/210L Dry Gas Std Test # 1	0.079
0.08 g/210L Dry Gas Std Test # (Repeat) 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
Alcohol Free Test # (Repeat) 2	0.000
0.05 g/210L Test # 2	0.046
0.08 g/210L Test # 2	0.075
0.20 g/210L Test # 2	0.194
0.08 g/210L Dry Gas Std Test # 2	0.079
0.08 g/210L Dry Gas Std Test # (Repeat) 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
Alcohol Free Test # (Repeat) 3	0.000
0.05 g/210L Test # 3	0.045
0.08 g/210L Test # 3	0.075
0.20 g/210L Test # 3	0.194
0.08 g/210L Dry Gas Std Test # 3	0.079
0.08 g/210L Dry Gas Std Test # (Repeat) 3	0.082
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
Alcohol Free Test # (Repeat) 4	0.000
0.05 g/210L Test # 4	0.046
0.08 g/210L Test # 4	0.075
0.20 g/210L Test # 4	0.194
0.08 g/210L Dry Gas Std Test # 4	0.079

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001442

0.08 g/210L Dry Gas Std Test # (Repeat) 4	0.081
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
Alcohol Free Test # (Repeat) 5	0.000
0.05 g/210L Test # 5	0.046
0.08 g/210L Test # 5	0.075
0.20 g/210L Test # 5	0.195
0.08 g/210L Dry Gas Std Test # 5	0.079
0.08 g/210L Dry Gas Std Test # (Repeat) 5	0.081
Interferent Detect Test # 5	INT
Alcohol Free Test # 6	0.000
Alcohol Free Test # (Repeat) 6	0.000
0.05 g/210L Test # 6	0.046
0.08 g/210L Test # 6	0.075
0.20 g/210L Test # 6	0.194
0.08 g/210L Dry Gas Std Test # 6	0.079
0.08 g/210L Dry Gas Std Test # (Repeat) 6	0.080
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
Alcohol Free Test # (Repeat) 7	0.000
0.05 g/210L Test # 7	0.046
0.08 g/210L Test # 7	0.075
0.20 g/210L Test # 7	0.195
0.08 g/210L Dry Gas Std Test # 7	0.078
0.08 g/210L Dry Gas Std Test # (Repeat) 7	0.079
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
Alcohol Free Test # (Repeat) 8	0.000
0.05 g/210L Test # 8	0.046
0.08 g/210L Test # 8	0.075
0.20 g/210L Test # 8	0.194
0.08 g/210L Dry Gas Std Test # 8	0.078
0.08 g/210L Dry Gas Std Test # (Repeat) 8	0.079
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	INT
Alcohol Free Test # (Repeat) 9	0.000
0.05 g/210L Test # 9	0.046
0.08 g/210L Test # 9	0.075
0.20 g/210L Test # 9	0.194
0.08 g/210L Dry Gas Std Test # 9	INT

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001442

0.08 g/210L Dry Gas Std Test # (Repeat) 9	0.078
Interferent Detect Test # 9	INT
Alcohol Free Test # (Repeat) 10	0.000
0.05 g/210L Test # 10	0.046
0.08 g/210L Test # 10	0.075
0.20 g/210L Test # 10	0.195
0.08 g/210L Dry Gas Std Test # (Repeat) 10	0.078
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0000
0.20 g/210L Test (Std Dev)	0.0005
0.08 g/210L Dry Gas Std Test (Std Dev) (Repeat)	0.0013

Remarks:

00: Interferent Detect; TURNED FAN ON. 08: Interferent Detect; APPLIED FAN TOWARD INSTRUMENT.

Data Download Date/Time 12/23/2025 16:09

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001442

Agency BISCAYNE NAT PARK **Date** 12/30/2025 **Type** Department Inspection
Inspector HIGGINBOTHAM,LEANDRA, **Time** 14:20:28 **Software** 8100.27

0.05 Lot # 202406K **0.08 Lot #** 202406L **0.20 Lot #** 202406N **0.08 Gas Lot #** AG510701
Expiration 06/19/2026 **Expiration** 06/19/2026 **Expiration** 06/20/2026 **Expiration** 04/17/2027

Number of Simulators 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.081
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.051
0.08 g/210L Test # 2	0.082
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.051
0.08 g/210L Test # 3	0.082
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.051
0.08 g/210L Test # 4	0.082
0.20 g/210L Test # 4	0.199
0.08 g/210L Dry Gas Std Test # 4	0.079
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.051
0.08 g/210L Test # 5	0.082
0.20 g/210L Test # 5	0.198
0.08 g/210L Dry Gas Std Test # 5	0.079
Interferent Detect Test # 5	INT

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001442

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.052
0.08 g/210L Test # 6	0.082
0.20 g/210L Test # 6	0.198
0.08 g/210L Dry Gas Std Test # 6	0.079
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.051
0.08 g/210L Test # 7	0.082
0.20 g/210L Test # 7	0.198
0.08 g/210L Dry Gas Std Test # 7	0.078
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.052
0.08 g/210L Test # 8	0.082
0.20 g/210L Test # 8	0.197
0.08 g/210L Dry Gas Std Test # 8	0.079
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.052
0.08 g/210L Test # 9	0.082
0.20 g/210L Test # 9	0.198
0.08 g/210L Dry Gas Std Test # 9	0.079
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.051
0.08 g/210L Test # 10	0.082
0.20 g/210L Test # 10	0.197
0.08 g/210L Dry Gas Std Test # 10	0.079
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0006
0.08 g/210L Test (Std Dev)	0.0003
0.20 g/210L Test (Std Dev)	0.0007
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

Data Download Date/Time 12/30/2025 14:17

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001471

Agency	SUMTER COUNTY SO	**Date**	01/28/2025	**Type**	Agency Inspection
Inspector	HAYES,CLINTON,R	**Time**	03:07:45	**Software**	8100.27

0.05 Lot #	202402E	**0.08 Lot #**	202410E	**0.20 Lot #**	202307C	**0.08 Gas Lot #**	231603
Expiration	02/27/2026	**Expiration**	10/22/2026	**Expiration**	07/12/2025	**Expiration**	06/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.047
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.193
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.047
0.08 g/210L Test # 2	0.077
0.20 g/210L Test # 2	0.194
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.195
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/28/2025 03:11

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001471

Agency	SUMTER COUNTY SO	**Date**	02/27/2025	**Type**	Agency Inspection
Inspector	HOLLIS,TERENCE,L	**Time**	19:09:19	**Software**	8100.27

0.05 Lot #	**0.08 Lot #**	**0.20 Lot #**	**0.08 Gas Lot #**
Expiration	**Expiration**	**Expiration**	**Expiration**

Number of Simulators 5 **Compliance** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	Interferent Detect Test # 1	0.000
Minimum Sample Volume Check: OK		Interferent Detect Test # (Repeat) 1	0.000
Barometric Pressure Sensor Check: OK		Alcohol Free Test # 2	0.000
Alcohol Free Subject Test: 0.000	Yes	Interferent Detect Test # 2	0.000
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # (Repeat) 2	0.000
Interferent Detect Test: Interferent Detect	No	Alcohol Free Test # 3	0.000
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test # 3	0.000
Alcohol Free Test: OK	Yes	Interferent Detect Test # (Repeat) 3	0.000
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Int Det:RADIO TO CLOSE TO DUI ROOM,. Non-compliance:INT NOT DETECTED RESTART INSPECTION.

Data Download Date/Time 02/27/2025 19:52

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001471

Agency SUMTER COUNTY SO **Date** 02/27/2025 **Type** Agency Inspection
Inspector HOLLIS,TERENCE,L **Time** 19:52:31 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202307C **0.08 Gas Lot #** 231603
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 07/12/2025 **Expiration** 06/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.047
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.194
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.047
0.08 g/210L Test # 2	0.077
0.20 g/210L Test # 2	0.194
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.077
0.20 g/210L Test # 3	0.193
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/27/2025 19:52

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001471

Agency SUMTER COUNTY SO **Date** 03/30/2025 **Type** Agency Inspection
Inspector HAYES,CLINTON,R **Time** 02:21:19 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202307C **0.08 Gas Lot #** 231603
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 07/12/2025 **Expiration** 06/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.046
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.195
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.047
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.195
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.047
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.196
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 03/30/2025 02:49

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001471

Agency SUMTER COUNTY SO **Date** 04/30/2025 **Type** Agency Inspection
Inspector CAUDLE,PAUL,A **Time** 11:05:53 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202307C **0.08 Gas Lot #** 231603
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 07/12/2025 **Expiration** 06/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.047
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.195
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	0.000
Interferent Detect Test # (Repeat) 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.047
0.08 g/210L Test # 2	0.077
0.20 g/210L Test # 2	0.196
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	0.000
Interferent Detect Test # (Repeat) 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.047
0.08 g/210L Test # 3	0.077
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	0.000
Interferent Detect Test # (Repeat) 3	INT

Remarks:

INT DET CHANGED SOLUTION, CHECKED SEALS, RETESTED

Data Download Date/Time 04/30/2025 11:08

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001471

Agency SUMTER COUNTY SO **Date** 05/30/2025 **Type** Agency Inspection
Inspector HOLLIS,TERENCE,L **Time** 17:09:06 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202307C **0.08 Gas Lot #** 231603
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 07/12/2025 **Expiration** 06/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.047
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.196
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.047
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.196
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 05/30/2025 17:13

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001471

Agency	SUMTER COUNTY SO	**Date**	06/22/2025	**Type**	Agency Inspection
Inspector	HAYES,CLINTON,R	**Time**	07:58:58	**Software**	8100.27

0.05 Lot #	202402E	**0.08 Lot #**	202410E	**0.20 Lot #**	202307C	**0.08 Gas Lot #**	231603
Expiration	02/27/2026	**Expiration**	10/22/2026	**Expiration**	07/12/2025	**Expiration**	06/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.046
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.195
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.047
0.08 g/210L Test # 2	0.077
0.20 g/210L Test # 2	0.195
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.047
0.08 g/210L Test # 3	0.077
0.20 g/210L Test # 3	0.196
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 06/22/2025 08:01

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001471

Agency SUMTER COUNTY SO **Date** 07/01/2025 **Type** Agency Inspection
Inspector CAUDLE,PAUL,A **Time** 10:16:00 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202307C **0.08 Gas Lot #** 231603
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 07/12/2025 **Expiration** 06/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.045
0.08 g/210L Test # 1	0.076
0.20 g/210L Test # 1	0.193
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.046
0.08 g/210L Test # 2	0.077
0.20 g/210L Test # 2	0.194
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.046
0.08 g/210L Test # 3	0.076
0.20 g/210L Test # 3	0.195
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 07/01/2025 10:17

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001471

Agency SUMTER COUNTY SO **Date** 07/11/2025 **Type** Department Inspection
Inspector PLATT,SHAYLA,D **Time** 11:33:51 **Software** 8100.27

0.05 Lot # 202406K **0.08 Lot #** 202406L **0.20 Lot #** 202406N **0.08 Gas Lot #** AG429602
Expiration 06/19/2026 **Expiration** 06/19/2026 **Expiration** 06/20/2026 **Expiration** 10/22/2026

Number of Simulators 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.081
0.20 g/210L Test # 1	0.200
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.081
0.20 g/210L Test # 2	0.200
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.201
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.050
0.08 g/210L Test # 4	0.081
0.20 g/210L Test # 4	0.200
0.08 g/210L Dry Gas Std Test # 4	0.079
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.050
0.08 g/210L Test # 5	0.080
0.20 g/210L Test # 5	0.200
0.08 g/210L Dry Gas Std Test # 5	0.079
Interferent Detect Test # 5	INT

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001471

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.050
0.08 g/210L Test # 6	0.080
0.20 g/210L Test # 6	0.200
0.08 g/210L Dry Gas Std Test # 6	0.079
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.080
0.20 g/210L Test # 7	0.200
0.08 g/210L Dry Gas Std Test # 7	0.079
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.049
0.08 g/210L Test # 8	0.080
0.20 g/210L Test # 8	0.200
0.08 g/210L Dry Gas Std Test # 8	0.080
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.050
0.08 g/210L Test # 9	0.080
0.20 g/210L Test # 9	0.201
0.08 g/210L Dry Gas Std Test # 9	0.079
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.080
0.20 g/210L Test # 10	0.200
0.08 g/210L Dry Gas Std Test # 10	0.079
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0004
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Data Download Date/Time 07/11/2025 11:32

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001471

Agency SUMTER COUNTY SO **Date** 08/04/2025 **Type** Agency Inspection
Inspector CAUDLE,PAUL,A **Time** 11:03:15 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 231603
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 06/05/2026

Number of Simulators 4 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.200
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 08/04/2025 11:09

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001471

Agency SUMTER COUNTY SO **Date** 09/15/2025 **Type** Agency Inspection
Inspector HOLLIS,TERENCE,L **Time** 21:56:36 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 231603
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 06/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.047
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.077
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 09/15/2025 22:03

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001471

Agency SUMTER COUNTY SO **Date** 10/30/2025 **Type** Agency Inspection
Inspector HAYES,CLINTON,R **Time** 11:55:02 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 231603
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 06/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.075
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.046
		0.08 g/210L Test # 3	0.076
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 11/03/2025 12:00

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001471

Agency SUMTER COUNTY SO **Date** 11/29/2025 **Type** Agency Inspection
Inspector HOLLIS,TERENCE,L **Time** 07:49:00 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 231603
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 06/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	No
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.046
0.08 g/210L Test # 1	0.076
0.20 g/210L Test # 1	0.194
0.08 g/210L Dry Gas Std Test # 1	0.082
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.046
0.08 g/210L Test # 2	0.076
0.20 g/210L Test # 2	0.195
0.08 g/210L Dry Gas Std Test # 2	0.083
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.046
0.08 g/210L Test # 3	0.076
0.20 g/210L Test # 3	0.195
0.08 g/210L Dry Gas Std Test # 3	0.082
Interferent Detect Test # 3	INT

Remarks:

DIANOSTIC FAILED

Data Download Date/Time 11/29/2025 07:46

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001475

Agency	DAYTONA STATE COLL	**Date**	02/13/2025	**Type**	Agency Inspection
Inspector	WAGNER,ROBERT,F	**Time**	07:54:52	**Software**	8100.27

0.05 Lot #	**0.08 Lot #**	**0.20 Lot #**	**0.08 Gas Lot #**
Expiration	**Expiration**	**Expiration**	**Expiration**

Number of Simulators **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Data Download Date/Time 02/04/2026 16:14

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001475

Agency	DAYTONA STATE COLL	**Date**	03/10/2025	**Type**	Agency Inspection
Inspector	WQAGNER,ROBERT,F	**Time**	09:00:11	**Software**	8100.27

0.05 Lot #	**0.08 Lot #**	**0.20 Lot #**	**0.08 Gas Lot #**
Expiration	**Expiration**	**Expiration**	**Expiration**

Number of Simulators **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	Yes		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Time-Date changed.

Data Download Date/Time 02/04/2026 16:14

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001475

Agency	DAYTONA STATE COLL	**Date**	03/25/2025	**Type**	Agency Inspection
Inspector	ARMSTRONG,DESTINEE,N	**Time**	14:27:34	**Software**	8100.27

0.05 Lot #	**0.08 Lot #**	**0.20 Lot #**	**0.08 Gas Lot #**
Expiration	**Expiration**	**Expiration**	**Expiration**

Number of Simulators **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

BYPASS AI TO OPERATE INSTRUMENT COMPLIANCE NOT DETERMINED

Data Download Date/Time 02/04/2026 16:14

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001475

Agency	DAYTONA STATE COLL	**Date**	06/25/2025	**Type**	Agency Inspection
Inspector	WAGNER,ROBERT,F	**Time**	10:48:48	**Software**	8100.27

0.05 Lot #	**0.08 Lot #**	**0.20 Lot #**	**0.08 Gas Lot #**
Expiration	**Expiration**	**Expiration**	**Expiration**

Number of Simulators **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Data Download Date/Time 02/04/2026 16:14

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001475

Agency	DAYTONA STATE COLL	**Date**	07/07/2025	**Type**	Agency Inspection
Inspector	WAGNER,ROBERT,F	**Time**	14:04:24	**Software**	8100.27

0.05 Lot #	**0.08 Lot #**	**0.20 Lot #**	**0.08 Gas Lot #**
Expiration	**Expiration**	**Expiration**	**Expiration**

Number of Simulators **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Data Download Date/Time 02/04/2026 16:14

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001475

Agency DAYTONA STATE COLL **Date** 07/07/2025 **Type** Agency Inspection
Inspector IZQUIERDO,JEREMY,M **Time** 15:34:10 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202007A **0.20 Lot #** 202005A **0.08 Gas Lot #** 0641908A1
Expiration 02/27/2026 **Expiration** 07/21/2026 **Expiration** 05/12/2026 **Expiration** 05/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

TESTING

Data Download Date/Time 02/04/2026 16:14

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001475

Agency DAYTONA STATE COLL **Date** 07/07/2025 **Type** Agency Inspection
Inspector SEARS,COLBY,D **Time** 16:06:09 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202007A **0.20 Lot #** 202005A **0.08 Gas Lot #** 06419080A1
Expiration 02/27/2026 **Expiration** 07/07/2026 **Expiration** 07/07/2026 **Expiration** 07/07/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.194
0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.047
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.195
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.196
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/04/2026 16:14

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001475

Agency DAYTONA STATE COLL **Date** 07/07/2025 **Type** Agency Inspection
Inspector YEATES,SCOTT,W **Time** 16:42:27 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202007A **0.20 Lot #** 202005A **0.08 Gas Lot #** 06419080A1
Expiration 02/27/2026 **Expiration** 12/25/2025 **Expiration** 12/25/2025 **Expiration** 12/25/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.194
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.081
0.20 g/210L Test # 2	0.196
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.081
0.20 g/210L Test # 3	0.195
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/04/2026 16:14

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001475

Agency DAYTONA STATE COLL **Date** 07/08/2025 **Type** Agency Inspection
Inspector SEARS,COLBY,D **Time** 09:57:20 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202007A **0.20 Lot #** 202005A **0.08 Gas Lot #** 06419080A1
Expiration 07/07/2026 **Expiration** 07/07/2026 **Expiration** 07/07/2026 **Expiration** 07/07/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.044
0.05 g/210L Test # (Repeat) 1	0.047
0.08 g/210L Test # 1	0.076
0.20 g/210L Test # 1	0.191
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.046
0.05 g/210L Test # (Repeat) 2	0.047
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.194
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.046
0.05 g/210L Test # (Repeat) 3	0.048
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.194
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

05: Control Outside Tolerance.

Data Download Date/Time 02/04/2026 16:14

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001475

Agency DAYTONA STATE COLL **Date** 07/08/2025 **Type** Agency Inspection
Inspector IZQUIERDO,JEREMY,M **Time** 10:29:39 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202007A **0.20 Lot #** 202005A **0.08 Gas Lot #** 064190080A
Expiration 02/27/2026 **Expiration** 07/21/2026 **Expiration** 05/12/2026 **Expiration** 05/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.047
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.194
0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.047
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.196
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/04/2026 16:14

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001475

Agency DAYTONA STATE COLL **Date** 07/08/2025 **Type** Agency Inspection
Inspector GASTON,SCOTT,B **Time** 11:10:18 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202007A **0.20 Lot #** 202005A **0.08 Gas Lot #** 0641080A1
Expiration 02/27/2026 **Expiration** 07/21/2026 **Expiration** 05/12/2026 **Expiration** 05/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.194
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.196
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/04/2026 16:15

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001475

Agency DAYTONA STATE COLL **Date** 07/08/2025 **Type** Agency Inspection
Inspector YEATES,SCOTT,W **Time** 12:45:55 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202007A **0.20 Lot #** 202005A **0.08 Gas Lot #**
Expiration 02/27/2026 **Expiration** 07/21/2026 **Expiration** 05/12/2026 **Expiration**

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	Interferent Detect Test # 1	INT
Mouth Alcohol Test: Slope Not Met	Yes	Alcohol Free Test # 2	0.000
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test # 2	0.049
Diagnostic Check (Post-Inspection): OK	No	0.08 g/210L Test # 2	0.081
Alcohol Free Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.05 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Test: OK	Yes	Alcohol Free Test # 3	0.000
0.20 g/210L Test: OK	Yes	0.05 g/210L Test # 3	0.048
0.08 g/210L Dry Gas Std Test: OK	No	0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.197
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/04/2026 16:15

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001475

Agency DAYTONA STATE COLL **Date** 07/08/2025 **Type** Agency Inspection
Inspector YEATES,SCOTT,W **Time** 13:20:06 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202007A **0.20 Lot #** 202005A **0.08 Gas Lot #** 06419080A1
Expiration 02/27/2026 **Expiration** 07/21/2026 **Expiration** 05/12/2026 **Expiration** 05/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.047
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.078
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/04/2026 16:15

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001475

Agency DAYTONA STATE COLL **Date** 07/08/2025 **Type** Agency Inspection
Inspector LEME,SANTIAGO, **Time** 14:04:24 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202007A **0.20 Lot #** 202005A **0.08 Gas Lot #**
Expiration 12/12/2025 **Expiration** 12/12/2025 **Expiration** 12/12/2025 **Expiration**

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	Interferent Detect Test # 1	INT
Mouth Alcohol Test: Slope Not Met	Yes	Alcohol Free Test # 2	0.000
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test # 2	0.049
Diagnostic Check (Post-Inspection): OK	No	0.08 g/210L Test # 2	0.080
Alcohol Free Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.05 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Test: OK	Yes	Alcohol Free Test # 3	0.000
0.20 g/210L Test: OK	Yes	0.05 g/210L Test # 3	0.049
0.08 g/210L Dry Gas Std Test: OK	No	0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.200
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/04/2026 16:15

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001475

Agency	DAYTONA STATE COLL	**Date**	07/08/2025	**Type**	Agency Inspection
Inspector	HARTLEY,BRANDEN,C	**Time**	14:09:25	**Software**	8100.27

0.05 Lot #	**0.08 Lot #**	**0.20 Lot #**	**0.08 Gas Lot #**
Expiration	**Expiration**	**Expiration**	**Expiration**

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	Yes		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A: .

Data Download Date/Time 02/04/2026 16:15

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001475

Agency DAYTONA STATE COLL **Date** 07/08/2025 **Type** Agency Inspection
Inspector HARTLEY,BRANDEN,C **Time** 14:38:18 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202007A **0.20 Lot #** 202005A **0.08 Gas Lot #**
Expiration 02/27/2026 **Expiration** 07/21/2026 **Expiration** 05/12/2026 **Expiration**

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	Interferent Detect Test # 1	INT
Mouth Alcohol Test: Slope Not Met	Yes	Alcohol Free Test # 2	0.000
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test # 2	0.049
Diagnostic Check (Post-Inspection): OK	No	0.08 g/210L Test # 2	0.080
Alcohol Free Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.05 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Test: OK	Yes	Alcohol Free Test # 3	0.000
0.20 g/210L Test: OK	Yes	0.05 g/210L Test # 3	0.048
0.08 g/210L Dry Gas Std Test: OK	No	0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.198
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/04/2026 16:15

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001475

Agency DAYTONA STATE COLL **Date** 07/08/2025 **Type** Agency Inspection
Inspector YEATES,SCOTT,W **Time** 16:19:47 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202007A **0.20 Lot #** 202005A **0.08 Gas Lot #** 06419080A1
Expiration 02/27/2026 **Expiration** 07/21/2026 **Expiration** 05/12/2026 **Expiration** 05/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.000
Minimum Sample Volume Check: OK		0.05 g/210L Test # (Repeat) 1	0.047
Barometric Pressure Sensor Check: OK		0.08 g/210L Test # 1	0.080
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # 1	0.196
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.000
0.05 g/210L Test: OK	Yes	0.05 g/210L Test # (Repeat) 2	0.049
0.08 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.20 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.000
		0.05 g/210L Test # (Repeat) 3	0.048
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

05: Control Outside Tolerance.

Data Download Date/Time 02/04/2026 16:15

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001475

Agency DAYTONA STATE COLL **Date** 07/09/2025 **Type** Agency Inspection
Inspector YEATES,SCOTT,W **Time** 08:59:41 **Software** 8100.27

0.05 Lot # 202101C **0.08 Lot #** 202007A **0.20 Lot #** 202206C **0.08 Gas Lot #** 06419080A1
Expiration 01/12/2026 **Expiration** 07/21/2026 **Expiration** 06/15/2026 **Expiration** 05/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.044
0.05 g/210L Test # (Repeat) 1	0.046
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.000
0.20 g/210L Test # (Repeat) 1	0.195
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.045
0.05 g/210L Test # (Repeat) 2	0.047
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.188
0.20 g/210L Test # (Repeat) 2	0.196
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.046
0.05 g/210L Test # (Repeat) 3	0.046
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.194
0.20 g/210L Test # (Repeat) 3	0.196
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

05: Control Outside Tolerance. 20: Control Outside Tolerance.

Data Download Date/Time 02/04/2026 16:15

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001475

Agency	DAYTONA STATE COLL	**Date**	07/09/2025	**Type**	Agency Inspection
Inspector	BLASSINGAME,PAUL,X	**Time**	09:05:37	**Software**	8100.27

0.05 Lot #	**0.08 Lot #**	**0.20 Lot #**	**0.08 Gas Lot #**
Expiration	**Expiration**	**Expiration**	**Expiration**

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	Yes		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A: .

Data Download Date/Time 02/04/2026 16:15

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001475

Agency	DAYTONA STATE COLL	**Date**	07/09/2025	**Type**	Agency Inspection
Inspector	BLASSINGAME,PAUL,X	**Time**	09:08:37	**Software**	8100.27

0.05 Lot #	**0.08 Lot #**	**0.20 Lot #**	**0.08 Gas Lot #**
Expiration	**Expiration**	**Expiration**	**Expiration**

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A: Subject Test Refused.

Data Download Date/Time 02/04/2026 16:15

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001475

Agency DAYTONA STATE COLL **Date** 07/09/2025 **Type** Agency Inspection
Inspector BLASSINGAME,PAUL,X **Time** 09:45:54 **Software** 8100.27

0.05 Lot # 202101C **0.08 Lot #** 202007A **0.20 Lot #** 202206C **0.08 Gas Lot #** 06419080A1
Expiration 01/12/2026 **Expiration** 07/21/2026 **Expiration** 06/15/2026 **Expiration** 05/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/04/2026 16:15

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001475

Agency	DAYTONA STATE COLL	**Date**	08/26/2025	**Type**	Agency Inspection
Inspector	WAGNER,ROBERT,F	**Time**	14:41:36	**Software**	8100.27

0.05 Lot #	**0.08 Lot #**	**0.20 Lot #**	**0.08 Gas Lot #**
Expiration	**Expiration**	**Expiration**	**Expiration**

Number of Simulators **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Data Download Date/Time 02/04/2026 16:15

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001475

Agency	DAYTONA STATE COLL	**Date**	11/10/2025	**Type**	Agency Inspection
Inspector	WAGNER,ROBERT,F	**Time**	10:21:04	**Software**	8100.27

0.05 Lot #	**0.08 Lot #**	**0.20 Lot #**	**0.08 Gas Lot #**
Expiration	**Expiration**	**Expiration**	**Expiration**

Number of Simulators **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Data Download Date/Time 02/04/2026 16:15

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001475

Agency	DAYTONA STATE COLL	**Date**	12/10/2025	**Type**	Agency Inspection
Inspector	WAGNER,ROBERT,F	**Time**	10:29:49	**Software**	8100.27

0.05 Lot #	**0.08 Lot #**	**0.20 Lot #**	**0.08 Gas Lot #**
Expiration	**Expiration**	**Expiration**	**Expiration**

Number of Simulators **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Data Download Date/Time 02/04/2026 16:15

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001646

Agency TYNDALL AFB **Date** 01/06/2025 **Type** Agency Inspection
Inspector LORD,LANNY,A **Time** 14:08:39 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 402477284
Expiration 02/27/2026 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 06/27/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.194
0.08 g/210L Dry Gas Std Test # 1	0.082
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.196
0.08 g/210L Dry Gas Std Test # 2	0.082
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.196
0.08 g/210L Dry Gas Std Test # 3	0.082
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/08/2025 09:49

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001646

Agency TYNDALL AFB **Date** 02/04/2025 **Type** Agency Inspection
Inspector LORD,LANNY,A **Time** 13:27:24 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 402477284
Expiration 02/27/2026 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 06/27/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.196
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/04/2025 16:08

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001646

Agency TYNDALL AFB **Date** 03/04/2025 **Type** Agency Inspection
Inspector LORD,LANNY,A **Time** 12:36:25 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 402477284
Expiration 02/27/2026 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 06/27/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.196
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 03/04/2025 13:34

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001646

Agency TYNDALL AFB **Date** 04/02/2025 **Type** Agency Inspection
Inspector LORD,LANNY,A **Time** 10:17:10 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 403268835
Expiration 02/27/2026 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 02/13/2028

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.076
0.20 g/210L Test # 1	0.194
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.077
0.20 g/210L Test # 2	0.195
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.196
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 04/02/2025 11:21

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001646

Agency TYNDALL AFB **Date** 05/07/2025 **Type** Agency Inspection
Inspector LORD,LANNY,A **Time** 10:28:07 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 403268835
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 02/13/2028

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.195
0.08 g/210L Dry Gas Std Test # 1	0.082
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.196
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.082
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 05/07/2025 14:32

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001646

Agency TYNDALL AFB **Date** 06/02/2025 **Type** Agency Inspection
Inspector LORD,LANNY,A **Time** 11:05:19 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 403268835
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 02/13/2028

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.082
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.082
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 06/02/2025 12:08

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

Agency CAPE CORAL PD **Date** 01/14/2025 **Type** Agency Inspection
Inspector BARKLEY,JEFFREY,G **Time** 23:02:37 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 28423080A4
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 11/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.195
0.08 g/210L Dry Gas Std Test # 1	0.082
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.194
0.08 g/210L Dry Gas Std Test # 2	0.082
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.195
0.08 g/210L Dry Gas Std Test # 3	0.082
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/18/2025 03:41

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

Agency CAPE CORAL PD **Date** 02/06/2025 **Type** Agency Inspection
Inspector BARKLEY,JEFFREY,G **Time** 03:02:53 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 28423080A4
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 11/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.192
0.08 g/210L Dry Gas Std Test # 1	0.082
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.194
0.08 g/210L Dry Gas Std Test # 2	0.082
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.195
0.08 g/210L Dry Gas Std Test # 3	0.082
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/06/2025 03:13

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

Agency CAPE CORAL PD **Date** 03/06/2025 **Type** Agency Inspection
Inspector BARKLEY,JEFFREY,G **Time** 01:44:18 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 28423080A4
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 11/05/2025

Number of Simulators 5 **Compliance** No

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	No
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.195
0.08 g/210L Dry Gas Std Test # 1	0.082
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.195
0.08 g/210L Dry Gas Std Test # 2	0.083
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.082
Interferent Detect Test # 3	INT

Remarks:

Non-compliance:ALCOHOL FREE SUBJECT OUT OF COMPLIANCE .

Data Download Date/Time 03/06/2025 01:53

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

Agency CAPE CORAL PD **Date** 10/30/2025 **Type** Department Inspection
Inspector GUTSCHOW,TAYLOR,D **Time** 13:49:55 **Software** 8100.27

0.05 Lot # 202406K **0.08 Lot #** 202406L **0.20 Lot #** 202406N **0.08 Gas Lot #** AG429602
Expiration 06/19/2026 **Expiration** 06/19/2026 **Expiration** 06/20/2026 **Expiration** 10/22/2026

Number of Simulators 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.078
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.049
0.08 g/210L Test # 4	0.080
0.20 g/210L Test # 4	0.198
0.08 g/210L Dry Gas Std Test # 4	0.079
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.050
0.08 g/210L Test # 5	0.080
0.20 g/210L Test # 5	0.197
0.08 g/210L Dry Gas Std Test # 5	0.078
Interferent Detect Test # 5	INT

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.050
0.08 g/210L Test # 6	0.080
0.20 g/210L Test # 6	0.197
0.08 g/210L Dry Gas Std Test # 6	0.078
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.080
0.20 g/210L Test # 7	0.198
0.08 g/210L Dry Gas Std Test # 7	0.077
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.049
0.08 g/210L Test # 8	0.080
0.20 g/210L Test # 8	0.198
0.08 g/210L Dry Gas Std Test # 8	0.078
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.050
0.08 g/210L Test # 9	0.080
0.20 g/210L Test # 9	0.198
0.08 g/210L Dry Gas Std Test # 9	0.077
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.080
0.20 g/210L Test # 10	0.197
0.08 g/210L Dry Gas Std Test # 10	0.078
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0005
0.08 g/210L Test (Std Dev)	0.0000
0.20 g/210L Test (Std Dev)	0.0005
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0006

Remarks:

Data Download Date/Time 10/30/2025 13:48

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

Agency CAPE CORAL PD **Date** 11/24/2025 **Type** Agency Inspection
Inspector BARKLEY,JEFFREY,G **Time** 20:37:09 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 26824080A1
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 11/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.195
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 11/24/2025 20:43

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

Agency CAPE CORAL PD **Date** 12/12/2025 **Type** Agency Inspection
Inspector BARKLEY,JEFFREY,G **Time** 20:17:51 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202510E **0.20 Lot #** 202410F **0.08 Gas Lot #** 26824080A1
Expiration 02/27/2026 **Expiration** 10/28/2027 **Expiration** 10/30/2026 **Expiration** 11/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.195
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.196
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 12/12/2025 20:24

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001648

Agency	PUTNAM COUNTY SO	**Date**	06/26/2025	**Type**	Agency Inspection
Inspector	PIERSON,WEN-CHI,K	**Time**	08:57:12	**Software**	8100.27

0.05 Lot #	**0.08 Lot #**	**0.20 Lot #**	**0.08 Gas Lot #**
Expiration	**Expiration**	**Expiration**	**Expiration**

Number of Simulators **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

BYPASS AI TO OPERATE INSTRUMENT COMPLIANCE NOT DETERMINED

Data Download Date/Time 06/30/2025 17:12

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001648

Agency PUTNAM COUNTY SO **Date** 06/30/2025 **Type** Department Inspection
Inspector HIGGINBOTHAM,LEANDRA, **Time** 16:57:47 **Software** 8100.27

0.05 Lot # 202406K **0.08 Lot #** 202406L **0.20 Lot #** 202406N **0.08 Gas Lot #** AG429602
Expiration 06/19/2026 **Expiration** 06/19/2026 **Expiration** 06/20/2026 **Expiration** 10/22/2026

Number of Simulators 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.194
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.047
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.195
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.047
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.196
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.047
0.08 g/210L Test # 4	0.079
0.20 g/210L Test # 4	0.196
0.08 g/210L Dry Gas Std Test # 4	0.079
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.047
0.08 g/210L Test # 5	0.079
0.20 g/210L Test # 5	0.195
0.08 g/210L Dry Gas Std Test # 5	0.079
Interferent Detect Test # 5	INT

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001648

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.048
0.08 g/210L Test # 6	0.079
0.20 g/210L Test # 6	0.195
0.08 g/210L Dry Gas Std Test # 6	0.079
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.047
0.08 g/210L Test # 7	0.079
0.20 g/210L Test # 7	0.195
0.08 g/210L Dry Gas Std Test # 7	0.079
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.047
0.08 g/210L Test # 8	0.079
0.20 g/210L Test # 8	0.195
0.08 g/210L Dry Gas Std Test # 8	0.079
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.047
0.08 g/210L Test # 9	0.079
0.20 g/210L Test # 9	0.195
0.08 g/210L Dry Gas Std Test # 9	0.079
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.047
0.08 g/210L Test # 10	0.080
0.20 g/210L Test # 10	0.196
0.08 g/210L Dry Gas Std Test # 10	0.079
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0006
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0000

Remarks:

Data Download Date/Time 06/30/2025 17:12

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001648

Agency PUTNAM COUNTY SO **Date** 07/16/2025 **Type** Agency Inspection
Inspector BISHOP,TREVOR,C **Time** 12:04:05 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 28423080A2
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 11/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.195
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.200
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.201
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 07/16/2025 12:03

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001648

Agency PUTNAM COUNTY SO **Date** 08/29/2025 **Type** Agency Inspection
Inspector BISHOP,TREVOR,C **Time** 15:34:17 **Software** 8100.27

0.05 Lot # 202506E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 28423080A2
Expiration 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 11/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.047
0.08 g/210L Test # 1	0.076
0.20 g/210L Test # 1	0.186
0.20 g/210L Test # (Repeat) 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.195
0.20 g/210L Test # (Repeat) 2	0.200
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.197
0.20 g/210L Test # (Repeat) 3	0.200
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

20: Control Outside Tolerance.

Data Download Date/Time 08/30/2025 20:21

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001648

Agency PUTNAM COUNTY SO **Date** 08/30/2025 **Type** Agency Inspection
Inspector BISHOP,TREVOR,C **Time** 20:17:18 **Software** 8100.27

0.05 Lot # 202506E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 28423080A2
Expiration 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 11/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.045
0.08 g/210L Test # 1	0.075
0.20 g/210L Test # 1	0.195
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.047
0.08 g/210L Test # 2	0.077
0.20 g/210L Test # 2	0.196
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.196
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 08/30/2025 20:21

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001648

Agency PUTNAM COUNTY SO **Date** 09/26/2025 **Type** Agency Inspection
Inspector BISHOP,TREVOR,C **Time** 20:07:17 **Software** 8100.27

0.05 Lot # 202506E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 28423080A2
Expiration 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 11/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.047
0.08 g/210L Test # 2	0.077
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 09/26/2025 20:07

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001648

Agency PUTNAM COUNTY SO **Date** 10/11/2025 **Type** Agency Inspection
Inspector BISHOP,TREVOR,C **Time** 19:29:26 **Software** 8100.27

0.05 Lot # 202506E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 28423080A2
Expiration 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 11/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 10/11/2025 19:32

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001648

Agency PUTNAM COUNTY SO **Date** 11/23/2025 **Type** Agency Inspection
Inspector BISHOP,TREVOR,C **Time** 18:41:59 **Software** 8100.27

0.05 Lot # 202506E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 251683
Expiration 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 11/05/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 12/03/2025 14:58

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001648

Agency PUTNAM COUNTY SO **Date** 12/23/2025 **Type** Agency Inspection
Inspector BISHOP,TREVOR,C **Time** 18:02:25 **Software** 8100.27

0.05 Lot # 202506E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 251683
Expiration 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 11/05/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.194
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 12/23/2025 18:45

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001651

Agency COLLIER COUNTY SO **Date** 01/16/2025 **Type** Agency Inspection
Inspector LANDAS,CARTER, **Time** 15:58:00 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 24923080A2
Expiration 03/28/2025 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 10/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

TESTED AT NJC

Data Download Date/Time 01/16/2025 16:58

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001651

Agency COLLIER COUNTY SO **Date** 02/05/2025 **Type** Agency Inspection
Inspector GONZALES,YVETTE,R **Time** 12:17:12 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 24923080A2
Expiration 03/28/2025 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 10/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/05/2025 13:21

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001651

Agency COLLIER COUNTY SO **Date** 03/18/2025 **Type** Agency Inspection
Inspector GONZALES,YVETTE,R **Time** 12:54:21 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 24923080A2
Expiration 03/28/2025 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 10/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.200
0.08 g/210L Dry Gas Std Test # 1	0.082
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.081
0.20 g/210L Test # 2	0.200
0.08 g/210L Dry Gas Std Test # 2	0.082
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.051
0.08 g/210L Test # 3	0.081
0.20 g/210L Test # 3	0.200
0.08 g/210L Dry Gas Std Test # 3	0.082
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 03/18/2025 14:33

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001651

Agency COLLIER COUNTY SO **Date** 04/09/2025 **Type** Agency Inspection
Inspector GONZALES,YVETTE,R **Time** 12:46:47 **Software** 8100.27

0.05 Lot # 202302E **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 24923080A2
Expiration 02/27/2026 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 10/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 04/09/2025 13:14

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001651

Agency COLLIER COUNTY SO **Date** 05/13/2025 **Type** Agency Inspection
Inspector GONZALES,YVETTE,R **Time** 09:57:48 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 24923080A2
Expiration 02/27/2026 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 10/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.051
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.051
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.081
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 05/13/2025 12:27

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001651

Agency COLLIER COUNTY SO **Date** 05/29/2025 **Type** Agency Inspection
Inspector GONZALES,YVETTE,R **Time** 13:44:13 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 24923080A2
Expiration 02/27/2026 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 10/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 05/29/2025 13:51

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001651

Agency COLLIER COUNTY SO **Date** 06/13/2025 **Type** Department Inspection
Inspector GUTSCHOW,TAYLOR,D **Time** 14:17:29 **Software** 8100.27

0.05 Lot # 202406K **0.08 Lot #** 202406L **0.20 Lot #** 202406N **0.08 Gas Lot #** AG429602
Expiration 06/19/2026 **Expiration** 06/19/2026 **Expiration** 06/20/2026 **Expiration** 10/22/2026

Number of Simulators 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.200
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.049
0.08 g/210L Test # 4	0.079
0.20 g/210L Test # 4	0.200
0.08 g/210L Dry Gas Std Test # 4	0.079
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.049
0.08 g/210L Test # 5	0.079
0.20 g/210L Test # 5	0.200
0.08 g/210L Dry Gas Std Test # 5	0.079
Interferent Detect Test # 5	INT

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001651

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.050
0.08 g/210L Test # 6	0.079
0.20 g/210L Test # 6	0.200
0.08 g/210L Dry Gas Std Test # 6	0.079
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.049
0.08 g/210L Test # 7	0.080
0.20 g/210L Test # 7	0.200
0.08 g/210L Dry Gas Std Test # 7	0.080
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.049
0.08 g/210L Test # 8	0.080
0.20 g/210L Test # 8	0.199
0.08 g/210L Dry Gas Std Test # 8	0.080
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.049
0.08 g/210L Test # 9	0.080
0.20 g/210L Test # 9	0.200
0.08 g/210L Dry Gas Std Test # 9	0.079
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.049
0.08 g/210L Test # 10	0.080
0.20 g/210L Test # 10	0.200
0.08 g/210L Dry Gas Std Test # 10	0.079
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0003
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0004
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Data Download Date/Time 06/13/2025 14:16

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001651

Agency COLLIER COUNTY SO **Date** 06/24/2025 **Type** Agency Inspection
Inspector GONZALES,YVETTE,R **Time** 10:08:14 **Software** 8100.27

0.05 Lot # 202302E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 24923080A2
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 10/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.081
0.20 g/210L Test # 1	0.200
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.081
0.20 g/210L Test # 2	0.200
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.082
0.20 g/210L Test # 3	0.200
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 06/24/2025 10:43

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001651

Agency COLLIER COUNTY SO **Date** 07/19/2025 **Type** Agency Inspection
Inspector GONZALES,YVETTE,R **Time** 11:02:36 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 24923080A2
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 10/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.200
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.200
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 07/19/2025 11:27

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001651

Agency COLLIER COUNTY SO **Date** 08/02/2025 **Type** Agency Inspection
Inspector GONZALES,YVETTE,R **Time** 12:48:05 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 24923080A2
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 10/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.081
0.20 g/210L Test # 2	0.200
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.081
0.20 g/210L Test # 3	0.200
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 08/02/2025 13:02

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001651

Agency COLLIER COUNTY SO **Date** 09/28/2025 **Type** Agency Inspection
Inspector LANDAS,CARTER, **Time** 14:32:50 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 24923080A2
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 10/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

TESTED AT NJC

Data Download Date/Time 09/28/2025 15:09

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001651

Agency COLLIER COUNTY SO **Date** 10/25/2025 **Type** Agency Inspection
Inspector LANDAS,CARTER, **Time** 08:41:27 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 04125080A5
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 03/05/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

TESTED AT NJC

Data Download Date/Time 10/25/2025 10:14

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001651

Agency COLLIER COUNTY SO **Date** 11/30/2025 **Type** Agency Inspection
Inspector LANDAS,CARTER, **Time** 11:55:59 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 04125080A5
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 03/05/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.081
0.20 g/210L Test # 1	0.200
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.081
0.20 g/210L Test # 2	0.200
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.081
0.20 g/210L Test # 3	0.201
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

M.A. TOO MUCH SOLUTION USED Sample Prov . M.A. TOO MUCH SOLUTION USED

Data Download Date/Time 12/01/2025 10:09

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001651

Agency COLLIER COUNTY SO **Date** 12/17/2025 **Type** Agency Inspection
Inspector LANDAS,CARTER,- **Time** 06:55:01 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 04125080A5
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 03/05/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

TESTED AT NJC

Data Download Date/Time 12/17/2025 08:44

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

Agency ST PETERSBURG PD **Date** 01/15/2025 **Type** Agency Inspection
Inspector HALL,DUSTIN,L **Time** 22:27:17 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 15924080A2
Expiration 02/27/2026 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 07/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.052
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.052
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.082
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.051
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.082
Interferent Detect Test # 3	INT

Remarks:

IN COMPLIANCE

Data Download Date/Time 01/15/2025 22:50

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

Agency ST PETERSBURG PD **Date** 02/11/2025 **Type** Agency Inspection
Inspector HALL,DUSTIN,L **Time** 22:20:42 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202307A **0.20 Lot #** 202410F **0.08 Gas Lot #** 31124080A1
Expiration 02/27/2026 **Expiration** 07/11/2025 **Expiration** 10/30/2026 **Expiration** 12/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.200
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

IN COMPLIANCE

Data Download Date/Time 02/13/2025 09:34

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

Agency ST PETERSBURG PD **Date** 03/12/2025 **Type** Agency Inspection
Inspector HALL,DUSTIN,L **Time** 01:45:40 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202307A **0.20 Lot #** 202410F **0.08 Gas Lot #** 31124080A1
Expiration 02/27/2026 **Expiration** 07/11/2025 **Expiration** 10/30/2026 **Expiration** 12/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.081
0.20 g/210L Test # 1	0.201
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.200
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

IN COMPLIANCE

Data Download Date/Time 03/12/2025 03:25

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

Agency	ST PETERSBURG PD	**Date**	08/13/2025	**Type**	Agency Inspection
Inspector	GUTSCHOW,TAYLOR,D	**Time**	10:30:12	**Software**	8100.27

0.05 Lot #	**0.08 Lot #**	**0.20 Lot #**	**0.08 Gas Lot #**
Expiration	**Expiration**	**Expiration**	**Expiration**

Number of Simulators **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

AI NOT CONDUCTED. COMPLIANCE NOT DETERMINED.

Data Download Date/Time 08/13/2025 12:01

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

Agency	ST PETERSBURG PD	**Date**	09/30/2025	**Type**	Agency Inspection
Inspector	GUTSCHOW,TAYLOR,D	**Time**	12:43:58	**Software**	8100.27

0.05 Lot #	**0.08 Lot #**	**0.20 Lot #**	**0.08 Gas Lot #**
Expiration	**Expiration**	**Expiration**	**Expiration**

Number of Simulators **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

AI NOT CONDUCTED. COMPLIANCE NOT DETERMINED.

Data Download Date/Time 10/06/2025 11:48

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

Agency ST PETERSBURG PD **Date** 10/06/2025 **Type** Department Inspection
Inspector GUTSCHOW,TAYLOR,D **Time** 16:01:37 **Software** 8100.27

0.05 Lot # 202406K **0.08 Lot #** 202406L **0.20 Lot #** 202406N **0.08 Gas Lot #** AG526603
Expiration 06/19/2026 **Expiration** 06/19/2026 **Expiration** 06/20/2026 **Expiration** 09/23/2027

Number of Simulators 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.048
0.08 g/210L Test # 4	0.078
0.20 g/210L Test # 4	0.198
0.08 g/210L Dry Gas Std Test # 4	0.078
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.048
0.08 g/210L Test # 5	0.079
0.20 g/210L Test # 5	0.197
0.08 g/210L Dry Gas Std Test # 5	0.078
Interferent Detect Test # 5	INT

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.048
0.08 g/210L Test # 6	0.078
0.20 g/210L Test # 6	0.197
0.08 g/210L Dry Gas Std Test # 6	0.078
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.048
0.08 g/210L Test # 7	0.078
0.20 g/210L Test # 7	0.196
0.08 g/210L Dry Gas Std Test # 7	0.079
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.048
0.08 g/210L Test # 8	0.078
0.20 g/210L Test # 8	0.197
0.08 g/210L Dry Gas Std Test # 8	0.079
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.048
0.08 g/210L Test # 9	0.079
0.20 g/210L Test # 9	0.197
0.08 g/210L Dry Gas Std Test # 9	0.079
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.048
0.08 g/210L Test # 10	0.078
0.20 g/210L Test # 10	0.197
0.08 g/210L Dry Gas Std Test # 10	0.079
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0000
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0004
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Data Download Date/Time 10/06/2025 15:58

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

Agency ST PETERSBURG PD **Date** 10/24/2025 **Type** Agency Inspection
Inspector HALL,DUSTIN,L **Time** 22:51:09 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 04125080A5
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 03/05/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

IN COMPLIANCE

Data Download Date/Time 10/24/2025 23:38

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001654

Agency MIAMI PD **Date** 01/30/2025 **Type** Agency Inspection
Inspector FERNANDEZ,ANGEL, **Time** 19:57:48 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 402367487
Expiration 02/27/2026 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 04/28/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.047
0.08 g/210L Test # 1	0.076
0.20 g/210L Test # 1	0.191
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.047
0.08 g/210L Test # 2	0.076
0.20 g/210L Test # 2	0.191
0.08 g/210L Dry Gas Std Test # 2	0.082
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.046
0.08 g/210L Test # 3	0.076
0.20 g/210L Test # 3	0.192
0.08 g/210L Dry Gas Std Test # 3	0.082
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/30/2025 20:03

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001654

Agency MIAMI PD **Date** 02/22/2025 **Type** Agency Inspection
Inspector FERNANDEZ,ANGEL, **Time** 00:21:17 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 402367487
Expiration 02/27/2026 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 04/28/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.194
0.08 g/210L Dry Gas Std Test # 3	0.082
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/22/2025 00:22

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001654

Agency MIAMI PD **Date** 03/26/2025 **Type** Agency Inspection
Inspector FERNANDEZ,ANGEL, **Time** 21:41:24 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202402E **0.20 Lot #** 202410F **0.08 Gas Lot #** 402367487
Expiration 02/27/2026 **Expiration** 02/27/2026 **Expiration** 10/30/2026 **Expiration** 04/28/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.077
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 03/26/2025 21:47

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001654

Agency MIAMI PD **Date** 04/30/2025 **Type** Agency Inspection
Inspector FERNANDEZ,ANGEL, **Time** 23:11:08 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 402732434
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 04/28/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.082
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.082
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.082
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 04/30/2025 23:10

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001654

Agency MIAMI PD **Date** 05/01/2025 **Type** Agency Inspection
Inspector FERNANDEZ,ANGEL, **Time** 01:09:50 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 402732434
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 04/28/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.082
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.082
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 05/01/2025 01:10

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001654

Agency MIAMI PD **Date** 05/31/2025 **Type** Agency Inspection
Inspector FERNANDEZ,ANGEL, **Time** 01:12:30 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 402732434
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 04/28/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 05/31/2025 01:11

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001654

Agency MIAMI PD **Date** 06/27/2025 **Type** Agency Inspection
Inspector FERNANDEZ,ANGEL, **Time** 11:14:20 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 402732434
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 04/28/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.046
0.08 g/210L Test # 1	0.076
0.20 g/210L Test # 1	0.196
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.046
0.08 g/210L Test # 2	0.076
0.20 g/210L Test # 2	0.196
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.046
0.08 g/210L Test # 3	0.077
0.20 g/210L Test # 3	0.196
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 06/27/2025 11:15

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001654

Agency	MIAMI PD	**Date**	09/03/2025	**Type**	Agency Inspection
Inspector	GUTSCHOW,TAYLOR,D	**Time**	09:35:18	**Software**	8100.27

0.05 Lot #	**0.08 Lot #**	**0.20 Lot #**	**0.08 Gas Lot #**
Expiration	**Expiration**	**Expiration**	**Expiration**

Number of Simulators **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

AI NOT CONDUCTED. COMPLIANCE NOT DETERMINED.

Data Download Date/Time 09/03/2025 14:06

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001654

Agency MIAMI PD **Date** 09/03/2025 **Type** Department Inspection
Inspector GUTSCHOW,TAYLOR,D **Time** 13:29:03 **Software** 8100.27

0.05 Lot # 202406K **0.08 Lot #** 202406L **0.20 Lot #** 202406N **0.08 Gas Lot #** AG429602
Expiration 06/19/2026 **Expiration** 06/19/2026 **Expiration** 06/20/2026 **Expiration** 10/22/2026

Number of Simulators 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.049
0.08 g/210L Test # 4	0.079
0.20 g/210L Test # 4	0.199
0.08 g/210L Dry Gas Std Test # 4	0.079
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.049
0.08 g/210L Test # 5	0.079
0.20 g/210L Test # 5	0.199
0.08 g/210L Dry Gas Std Test # 5	0.078
Interferent Detect Test # 5	INT

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001654

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.048
0.08 g/210L Test # 6	0.079
0.20 g/210L Test # 6	0.198
0.08 g/210L Dry Gas Std Test # 6	0.078
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.048
0.08 g/210L Test # 7	0.079
0.20 g/210L Test # 7	0.199
0.08 g/210L Dry Gas Std Test # 7	0.078
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.049
0.08 g/210L Test # 8	0.079
0.20 g/210L Test # 8	0.199
0.08 g/210L Dry Gas Std Test # 8	0.079
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.049
0.08 g/210L Test # 9	0.079
0.20 g/210L Test # 9	0.199
0.08 g/210L Dry Gas Std Test # 9	0.079
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.048
0.08 g/210L Test # 10	0.079
0.20 g/210L Test # 10	0.199
0.08 g/210L Dry Gas Std Test # 10	0.078
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0000
0.20 g/210L Test (Std Dev)	0.0003
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

Data Download Date/Time 09/03/2025 14:06

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001654

Agency	MIAMI PD	**Date**	12/09/2025	**Type**	Agency Inspection
Inspector	GUTSCHOW,TAYLOR,D	**Time**	12:20:23	**Software**	8100.27

0.05 Lot #	**0.08 Lot #**	**0.20 Lot #**	**0.08 Gas Lot #**
Expiration	**Expiration**	**Expiration**	**Expiration**

Number of Simulators **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

AI NOT CONDUCTED. COMPLIANCE NOT DETERMINED.

Data Download Date/Time 12/09/2025 12:47

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

Agency ST JOHNS COUNTY SO **Date** 01/31/2025 **Type** Agency Inspection
Inspector ELLIS,ALEXANDER, **Time** 19:52:19 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 402816921
Expiration 02/27/2026 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 08/09/2026

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.083
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.083
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.195
		0.08 g/210L Dry Gas Std Test # 3	0.083
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/31/2025 20:21

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

Agency ST JOHNS COUNTY SO **Date** 02/26/2025 **Type** Agency Inspection
Inspector MILLS,SCOTT,T **Time** 21:04:20 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 402816921
Expiration 02/27/2026 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 08/09/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.082
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.083
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.083
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/26/2025 21:08

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

Agency ST JOHNS COUNTY SO **Date** 03/28/2025 **Type** Agency Inspection
Inspector ELLIS,ALEXANDER, **Time** 21:21:38 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 402816921
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 08/09/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.082
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.083
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.083
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 03/28/2025 21:24

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

Agency	ST JOHNS COUNTY SO	**Date**	04/24/2025	**Type**	Agency Inspection
Inspector	MILLS,SCOTT,T	**Time**	21:41:32	**Software**	8100.27

0.05 Lot #	202402E	**0.08 Lot #**	202410E	**0.20 Lot #**	202410F	**0.08 Gas Lot #**	402816921
Expiration	02/27/2026	**Expiration**	10/22/2026	**Expiration**	10/30/2026	**Expiration**	08/09/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

NO COMMENT

Data Download Date/Time 04/24/2025 21:43

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

Agency ST JOHNS COUNTY SO **Date** 05/30/2025 **Type** Agency Inspection
Inspector MILLS,SCOTT,T **Time** 19:07:07 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 402816921
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 08/09/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.051
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.200
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.051
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.200
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.051
0.08 g/210L Test # 3	0.081
0.20 g/210L Test # 3	0.201
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 05/30/2025 19:11

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

Agency ST JOHNS COUNTY SO **Date** 06/26/2025 **Type** Agency Inspection
Inspector MILLS,SCOTT,T **Time** 20:48:32 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 402816921
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 08/09/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.082
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.200
0.08 g/210L Dry Gas Std Test # 2	0.082
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.051
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.200
0.08 g/210L Dry Gas Std Test # 3	0.082
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 06/26/2025 20:55

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

Agency ST JOHNS COUNTY SO **Date** 07/30/2025 **Type** Agency Inspection
Inspector MILLS,SCOTT,T **Time** 21:29:52 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 403374313
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 06/25/2028

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.051
0.08 g/210L Test # 1	0.081
0.20 g/210L Test # 1	0.200
0.08 g/210L Dry Gas Std Test # 1	0.083
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.051
0.08 g/210L Test # 2	0.081
0.20 g/210L Test # 2	0.201
0.08 g/210L Dry Gas Std Test # 2	0.083
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.051
0.08 g/210L Test # 3	0.081
0.20 g/210L Test # 3	0.201
0.08 g/210L Dry Gas Std Test # 3	0.082
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 07/30/2025 21:34

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

Agency	ST JOHNS COUNTY SO	**Date**	08/28/2025	**Type**	Agency Inspection
Inspector	MILLS,SCOTT,T	**Time**	21:58:49	**Software**	8100.27

0.05 Lot #	**0.08 Lot #**	**0.20 Lot #**	**0.08 Gas Lot #**
Expiration	**Expiration**	**Expiration**	**Expiration**

Number of Simulators **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Data Download Date/Time 08/28/2025 22:41

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

Agency ST JOHNS COUNTY SO **Date** 08/28/2025 **Type** Agency Inspection
Inspector MILLS,SCOTT,T **Time** 22:38:52 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 403374313
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 06/25/2028

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	No
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.200
0.08 g/210L Dry Gas Std Test # 1	0.078
0.08 g/210L Dry Gas Std Test # (Repeat) 1	0.083
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.200
0.08 g/210L Dry Gas Std Test # 2	0.000
0.08 g/210L Dry Gas Std Test # (Repeat) 2	0.083
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.051
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.200
0.08 g/210L Dry Gas Std Test # 3	0.000
0.08 g/210L Dry Gas Std Test # (Repeat) 3	0.083
Interferent Detect Test # 3	INT

Remarks:

08: Control Outside Tolerance.

Data Download Date/Time 08/28/2025 22:41

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

Agency ST JOHNS COUNTY SO **Date** 09/30/2025 **Type** Agency Inspection
Inspector MILLS,SCOTT,T **Time** 20:04:11 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 403374313
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 06/25/2028

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.051
0.08 g/210L Test # 1	0.081
0.20 g/210L Test # 1	0.200
0.08 g/210L Dry Gas Std Test # 1	0.082
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.051
0.08 g/210L Test # 2	0.081
0.20 g/210L Test # 2	0.201
0.08 g/210L Dry Gas Std Test # 2	0.082
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.051
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.201
0.08 g/210L Dry Gas Std Test # 3	0.082
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 09/30/2025 20:05

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

Agency ST JOHNS COUNTY SO **Date** 10/14/2025 **Type** Department Inspection
Inspector HIGGINBOTHAM,LEANDRA, **Time** 13:26:24 **Software** 8100.27

0.05 Lot # 202406K **0.08 Lot #** 202406K **0.20 Lot #** 202406N **0.08 Gas Lot #** AG510701
Expiration 06/19/2026 **Expiration** 06/19/2026 **Expiration** 06/20/2026 **Expiration** 04/17/2027

Number of Simulators 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.200
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.081
0.20 g/210L Test # 2	0.200
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.200
0.08 g/210L Dry Gas Std Test # 3	0.082
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.050
0.08 g/210L Test # 4	0.080
0.20 g/210L Test # 4	0.199
0.08 g/210L Dry Gas Std Test # 4	0.082
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.050
0.08 g/210L Test # 5	0.081
0.20 g/210L Test # 5	0.200
0.08 g/210L Dry Gas Std Test # 5	0.081
Interferent Detect Test # 5	INT

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.050
0.08 g/210L Test # 6	0.081
0.20 g/210L Test # 6	0.200
0.08 g/210L Dry Gas Std Test # 6	0.081
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.051
0.08 g/210L Test # 7	0.081
0.20 g/210L Test # 7	0.200
0.08 g/210L Dry Gas Std Test # 7	0.082
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.051
0.08 g/210L Test # 8	0.080
0.20 g/210L Test # 8	0.200
0.08 g/210L Dry Gas Std Test # 8	0.082
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.050
0.08 g/210L Test # 9	0.080
0.20 g/210L Test # 9	0.201
0.08 g/210L Dry Gas Std Test # 9	0.082
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.081
0.20 g/210L Test # 10	0.200
0.08 g/210L Dry Gas Std Test # 10	0.081
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0005
0.20 g/210L Test (Std Dev)	0.0004
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

Data Download Date/Time 10/14/2025 13:26

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

Agency ST JOHNS COUNTY SO **Date** 10/29/2025 **Type** Agency Inspection
Inspector MILLS,SCOTT,T **Time** 18:30:43 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 403374313
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 06/25/2028

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.081
0.20 g/210L Test # 1	0.202
0.08 g/210L Dry Gas Std Test # 1	0.082
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.051
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.202
0.08 g/210L Dry Gas Std Test # 2	0.083
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.202
0.08 g/210L Dry Gas Std Test # 3	0.083
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 10/29/2025 18:32

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

Agency ST JOHNS COUNTY SO **Date** 11/27/2025 **Type** Agency Inspection
Inspector MILLS,SCOTT,T **Time** 20:43:47 **Software** 8100.27

0.05 Lot # 202506E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 403374313
Expiration 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 06/25/2028

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.052
0.08 g/210L Test # 1	0.082
0.20 g/210L Test # 1	0.201
0.08 g/210L Dry Gas Std Test # 1	0.083
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.052
0.08 g/210L Test # 2	0.083
0.20 g/210L Test # 2	0.202
0.08 g/210L Dry Gas Std Test # 2	0.083
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.053
0.08 g/210L Test # 3	0.082
0.20 g/210L Test # 3	0.202
0.08 g/210L Dry Gas Std Test # 3	0.084
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 11/27/2025 20:43

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

Agency ST JOHNS COUNTY SO **Date** 12/16/2025 **Type** Agency Inspection
Inspector MILLS,SCOTT,T **Time** 18:54:44 **Software** 8100.27

0.05 Lot # 202506E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 403374313
Expiration 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 06/25/2028

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.051
0.08 g/210L Test # 1	0.081
0.20 g/210L Test # 1	0.201
0.08 g/210L Dry Gas Std Test # 1	0.083
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.052
0.08 g/210L Test # 2	0.081
0.20 g/210L Test # 2	0.201
0.08 g/210L Dry Gas Std Test # 2	0.082
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.052
0.08 g/210L Test # 3	0.081
0.20 g/210L Test # 3	0.201
0.08 g/210L Dry Gas Std Test # 3	0.082
Interferent Detect Test # 3	INT

Remarks:

N/A

Data Download Date/Time 12/16/2025 19:13

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001656

Agency NORTH MIAMI PD **Date** 01/20/2025 **Type** Agency Inspection
Inspector DE JESUS,RAY, **Time** 14:04:59 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 26824080A1
Expiration 03/28/2025 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 11/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.046
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.196
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.046
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.196
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.046
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.196
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/20/2025 14:06

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001656

Agency NORTH MIAMI PD **Date** 02/05/2025 **Type** Department Inspection
Inspector GUTSCHOW,TAYLOR,D **Time** 13:50:05 **Software** 8100.27

0.05 Lot # 202303K **0.08 Lot #** 202303L **0.20 Lot #** 202304C **0.08 Gas Lot #** AG429602
Expiration 03/29/2025 **Expiration** 03/29/2025 **Expiration** 04/05/2025 **Expiration** 10/22/2026

Number of Simulators 5 **Average Standard Deviation** 0.0002 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.078
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.050
0.08 g/210L Test # 4	0.080
0.20 g/210L Test # 4	0.199
0.08 g/210L Dry Gas Std Test # 4	0.078
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.050
0.08 g/210L Test # 5	0.080
0.20 g/210L Test # 5	0.199
0.08 g/210L Dry Gas Std Test # 5	0.078
Interferent Detect Test # 5	INT

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001656

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.050
0.08 g/210L Test # 6	0.080
0.20 g/210L Test # 6	0.199
0.08 g/210L Dry Gas Std Test # 6	0.078
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.080
0.20 g/210L Test # 7	0.198
0.08 g/210L Dry Gas Std Test # 7	0.078
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.050
0.08 g/210L Test # 8	0.080
0.20 g/210L Test # 8	0.198
0.08 g/210L Dry Gas Std Test # 8	0.078
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.050
0.08 g/210L Test # 9	0.080
0.20 g/210L Test # 9	0.199
0.08 g/210L Dry Gas Std Test # 9	0.079
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.080
0.20 g/210L Test # 10	0.199
0.08 g/210L Dry Gas Std Test # 10	0.078
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0000
0.08 g/210L Test (Std Dev)	0.0003
0.20 g/210L Test (Std Dev)	0.0004
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

Data Download Date/Time 02/05/2025 13:53

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001656

Agency NORTH MIAMI PD **Date** 02/23/2025 **Type** Agency Inspection
Inspector DE JESUS,RAY, **Time** 18:52:01 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 26824080A1
Expiration 03/28/2025 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 11/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/23/2025 18:54

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001656

Agency NORTH MIAMI PD **Date** 03/24/2025 **Type** Agency Inspection
Inspector DE JESUS,RAY, **Time** 12:18:25 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 26824080A1
Expiration 03/28/2025 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 11/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.047
0.08 g/210L Test # 1	0.075
0.20 g/210L Test # 1	0.193
0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.047
0.08 g/210L Test # 2	0.076
0.20 g/210L Test # 2	0.193
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.047
0.08 g/210L Test # 3	0.075
0.20 g/210L Test # 3	0.194
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 03/24/2025 12:20

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001656

Agency NORTH MIAMI PD **Date** 04/23/2025 **Type** Agency Inspection
Inspector DE JESUS,RAY, **Time** 18:30:29 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 26824080A1
Expiration 02/27/2026 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 11/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.078
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 04/23/2025 18:31

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001656

Agency NORTH MIAMI PD **Date** 05/26/2025 **Type** Agency Inspection
Inspector DE JESUS,RAY, **Time** 22:47:46 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 26824080A1
Expiration 02/27/2026 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 11/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # (Repeat) 1	0.000
0.05 g/210L Test # 1	0.046
0.08 g/210L Test # 1	0.076
0.20 g/210L Test # 1	0.196
0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test # 1	INT
Alcohol Free Test # (Repeat) 2	0.000
0.05 g/210L Test # 2	0.046
0.08 g/210L Test # 2	0.076
0.20 g/210L Test # 2	0.196
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # (Repeat) 3	0.000
0.05 g/210L Test # 3	0.046
0.08 g/210L Test # 3	0.075
0.20 g/210L Test # 3	0.195
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

00: Ambient Fail-MOUTH ALCOHOL AIR.

Data Download Date/Time 05/26/2025 22:48

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001656

Agency NORTH MIAMI PD **Date** 06/17/2025 **Type** Agency Inspection
Inspector DE JESUS,RAY, **Time** 21:38:15 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 26824080A1
Expiration 02/27/2026 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 11/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 06/17/2025 21:39

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001656

Agency NORTH MIAMI PD **Date** 07/20/2025 **Type** Agency Inspection
Inspector DE JESUS,RAY, **Time** 17:14:00 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 26824080A1
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 11/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.047
0.08 g/210L Test # 1	0.075
0.20 g/210L Test # 1	0.196
0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.047
0.08 g/210L Test # 2	0.076
0.20 g/210L Test # 2	0.196
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.047
0.08 g/210L Test # 3	0.076
0.20 g/210L Test # 3	0.196
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

VERIFYING

Data Download Date/Time 07/20/2025 17:15

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001656

Agency NORTH MIAMI PD **Date** 08/11/2025 **Type** Agency Inspection
Inspector DE JESUS,RAY, **Time** 22:23:36 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 26824080A1
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 11/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.078
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 08/11/2025 22:24

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001656

Agency NORTH MIAMI PD **Date** 09/01/2025 **Type** Agency Inspection
Inspector DE JESUS,RAY, **Time** 17:58:21 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 26824080A1
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 11/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.077
0.20 g/210L Test # 2	0.195
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.077
0.20 g/210L Test # 3	0.193
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 09/01/2025 18:01

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001656

Agency NORTH MIAMI PD **Date** 10/15/2025 **Type** Agency Inspection
Inspector DE JESUS,RAY, **Time** 14:21:06 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 26824080A1
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 11/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.184
0.20 g/210L Test # (Repeat) 1	0.195
0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.077
0.20 g/210L Test # 2	0.190
0.20 g/210L Test # (Repeat) 2	0.194
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.077
0.20 g/210L Test # 3	0.191
0.20 g/210L Test # (Repeat) 3	0.195
0.08 g/210L Dry Gas Std Test # 3	0.078
Interferent Detect Test # 3	INT

Remarks:

20: Control Outside Tolerance-SIMULATOR WARMING UP.

Data Download Date/Time 10/15/2025 14:23

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001656

Agency NORTH MIAMI PD **Date** 11/09/2025 **Type** Agency Inspection
Inspector DE JESUS,RAY, **Time** 18:59:29 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 26824080A1
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 11/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.192
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.077
0.20 g/210L Test # 2	0.193
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.077
0.20 g/210L Test # 3	0.192
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 11/09/2025 19:01

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001656

Agency NORTH MIAMI PD **Date** 12/23/2025 **Type** Agency Inspection
Inspector DE JESUS,RAY, **Time** 16:05:52 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 26824080A1
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 11/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.076
0.08 g/210L Test # (Repeat) 1	0.075
0.20 g/210L Test # 1	0.193
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.076
0.08 g/210L Test # (Repeat) 2	0.077
0.20 g/210L Test # 2	0.193
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.074
0.08 g/210L Test # (Repeat) 3	0.076
0.20 g/210L Test # 3	0.194
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

08: Control Outside Tolerance-COLD SIMULATOR.

Data Download Date/Time 12/23/2025 16:08

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

Agency OSCEOLA COUNTY SO **Date** 01/20/2025 **Type** Agency Inspection
Inspector LYONS,DANIEL,E **Time** 17:04:47 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 403207507
Expiration 03/28/2025 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 11/26/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.051
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.200
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.082
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.082
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.051
0.08 g/210L Test # 3	0.082
0.20 g/210L Test # 3	0.202
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/24/2025 14:40

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

Agency OSCEOLA COUNTY SO **Date** 02/04/2025 **Type** Department Inspection
Inspector ARMSTRONG,DESTINEE,N **Time** 09:52:18 **Software** 8100.27

0.05 Lot # 202406K **0.08 Lot #** 202406L **0.20 Lot #** 202406N **0.08 Gas Lot #** AG429602
Expiration 06/19/2026 **Expiration** 06/19/2026 **Expiration** 06/20/2026 **Expiration** 10/22/2026

Number of Simulators 5 **Average Standard Deviation** 0.0008 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.081
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.200
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.196
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.050
0.08 g/210L Test # 4	0.082
0.20 g/210L Test # 4	0.197
0.08 g/210L Dry Gas Std Test # 4	0.079
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.051
0.08 g/210L Test # 5	0.081
0.20 g/210L Test # 5	0.197
0.08 g/210L Dry Gas Std Test # 5	0.079
Interferent Detect Test # 5	INT

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.050
0.08 g/210L Test # 6	0.079
0.20 g/210L Test # 6	0.197
0.08 g/210L Dry Gas Std Test # 6	0.079
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.051
0.08 g/210L Test # 7	0.080
0.20 g/210L Test # 7	0.199
0.08 g/210L Dry Gas Std Test # 7	0.080
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.050
0.08 g/210L Test # 8	0.082
0.20 g/210L Test # 8	0.198
0.08 g/210L Dry Gas Std Test # 8	0.079
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.051
0.08 g/210L Test # 9	0.079
0.20 g/210L Test # 9	0.198
0.08 g/210L Dry Gas Std Test # 9	0.079
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.080
0.20 g/210L Test # 10	0.198
0.08 g/210L Dry Gas Std Test # 10	0.079
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0006
0.08 g/210L Test (Std Dev)	0.0012
0.20 g/210L Test (Std Dev)	0.0011
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Data Download Date/Time 02/04/2025 09:54

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

Agency OSCEOLA COUNTY SO **Date** 02/26/2025 **Type** Agency Inspection
Inspector LYONS,DANIEL,E **Time** 15:44:28 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 403207507
Expiration 02/27/2026 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 11/26/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.082
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.052
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.083
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/26/2025 16:06

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

Agency OSCEOLA COUNTY SO **Date** 03/25/2025 **Type** Agency Inspection
Inspector LYONS/BLACK,DANIEL/TIMOTHY,E **Time** 11:40:32 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 403207507
Expiration 02/27/2026 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 11/26/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.195
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.195
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 03/25/2025 12:01

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

Agency OSCEOLA COUNTY SO **Date** 04/24/2025 **Type** Agency Inspection
Inspector LYONS/BLACK,DANIEL/TIMOTHY,E **Time** 13:58:10 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 403207507
Expiration 02/27/2026 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 11/26/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.081
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.081
0.20 g/210L Test # 2	0.200
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.081
0.20 g/210L Test # 3	0.195
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 04/24/2025 13:57

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

Agency OSCEOLA COUNTY SO **Date** 05/14/2025 **Type** Agency Inspection
Inspector LYONS,DANIEL',E **Time** 13:15:40 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 403207507
Expiration 02/27/2026 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 11/26/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.052
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.052
0.08 g/210L Test # 2	0.081
0.20 g/210L Test # 2	0.200
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.053
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.196
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 05/14/2025 13:17

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

Agency OSCEOLA COUNTY SO **Date** 06/23/2025 **Type** Agency Inspection
Inspector LYONS/AULTMAN,DANIEL/SHAWN,E **Time** 14:13:21 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 403207507
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 11/26/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.051
0.08 g/210L Test # 1	0.081
0.20 g/210L Test # 1	0.000
0.20 g/210L Test # (Repeat) 1	0.202
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.081
0.20 g/210L Test # 2	0.000
0.20 g/210L Test # (Repeat) 2	0.205
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.051
0.08 g/210L Test # 3	0.081
0.20 g/210L Test # 3	0.000
0.20 g/210L Test # (Repeat) 3	0.202
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

20:Control Outside Tol DID NOT CONNECT SIM

Data Download Date/Time 06/23/2025 14:16

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

Agency OSCEOLA COUNTY SO **Date** 07/24/2025 **Type** Agency Inspection
Inspector LYONS,DANIEL,E **Time** 13:32:57 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 403207507
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 11/26/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.052
0.08 g/210L Test # 1	0.082
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.051
0.08 g/210L Test # 2	0.083
0.20 g/210L Test # 2	0.202
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.051
0.08 g/210L Test # 3	0.083
0.20 g/210L Test # 3	0.201
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 07/24/2025 13:53

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

Agency OSCEOLA COUNTY SO **Date** 08/21/2025 **Type** Agency Inspection
Inspector LYONS,DANIEL,E **Time** 13:29:18 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 403207507
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 11/26/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.052
0.08 g/210L Test # 1	0.081
0.20 g/210L Test # 1	0.201
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.084
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.052
0.08 g/210L Test # 3	0.084
0.20 g/210L Test # 3	0.203
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 08/21/2025 13:28

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

Agency OSCEOLA COUNTY SO **Date** 09/16/2025 **Type** Agency Inspection
Inspector AULTMAN,SHAWN,A **Time** 12:31:05 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 403207507
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 11/26/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.081
0.20 g/210L Test # 1	0.202
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	0.028
Interferent Detect Test # (Repeat) 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.081
0.20 g/210L Test # 2	0.201
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	0.028
Interferent Detect Test # (Repeat) 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.082
0.20 g/210L Test # 3	0.200
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	0.029
Interferent Detect Test # (Repeat) 3	INT

Remarks:

Int Det:M/A USED IN ERROR SIM RESET W/ACETONE. RETESTED.

Data Download Date/Time 09/16/2025 12:31

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

Agency OSCEOLA COUNTY SO **Date** 10/29/2025 **Type** Agency Inspection
Inspector LYONS/AULTMAN,DANIEL/SHAWN,A **Time** 12:54:04 **Software** 8100.27

0.05 Lot # 202402E **Expiration** 02/27/2026
0.08 Lot # 202410E **Expiration** 10/22/2026
0.20 Lot # 202410F **Expiration** 10/30/2026
0.08 Gas Lot # 403207507 **Expiration** 11/26/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.201
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 10/29/2025 13:18

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

Agency OSCEOLA COUNTY SO **Date** 11/21/2025 **Type** Agency Inspection
Inspector LYONS DAN,AULTMAN SHAWN, **Time** 12:05:35 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 403207507
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 11/26/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.082
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.081
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 11/21/2025 12:04

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

Agency	OSCEOLA COUNTY SO	**Date**	12/24/2025	**Type**	Agency Inspection
Inspector	LYONS/DANIEL,DANIEL/SHAWN,E	**Time**	09:28:08	**Software**	8100.27

0.05 Lot #	**0.08 Lot #**	**0.20 Lot #**	**0.08 Gas Lot #**
Expiration	**Expiration**	**Expiration**	**Expiration**

Number of Simulators **Compliance** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

DSP FAILNon-compliance:DIAGNOSTIC CHECK SHOWED DSP FAIL.

Data Download Date/Time 12/24/2025 09:28

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001720

Agency ALACHUA PD **Date** 01/29/2025 **Type** Agency Inspection
Inspector LEVY,JAYSON,I **Time** 15:01:04 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 10124080A2
Expiration 03/28/2025 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 06/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.047
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.196
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.047
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.196
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.196
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/29/2025 17:40

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001720

Agency ALACHUA PD **Date** 02/07/2025 **Type** Agency Inspection
Inspector LEVY,JAYSON,I **Time** 16:08:11 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 10124080A2
Expiration 03/28/2025 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 06/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.200
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/07/2025 16:49

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001720

Agency ALACHUA PD **Date** 03/31/2025 **Type** Agency Inspection
Inspector LEVY,JAYSON,I **Time** 16:24:41 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 10124080A2
Expiration 02/27/2026 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 06/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 03/31/2025 17:15

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001720

Agency ALACHUA PD **Date** 04/27/2025 **Type** Agency Inspection
Inspector LEVY,JAYSON,I **Time** 17:50:16 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 10124080A2
Expiration 02/27/2026 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 06/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.196
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.196
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.196
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 04/27/2025 18:40

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001720

Agency ALACHUA PD **Date** 05/30/2025 **Type** Agency Inspection
Inspector LEVY,JAYSON,I **Time** 20:33:20 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 10124080A2
Expiration 02/27/2026 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 06/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # (Repeat) 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # (Repeat) 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.195
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # (Repeat) 3	INT

Remarks:

Int Det: Ambient Fail UNK WHY.

Data Download Date/Time 05/30/2025 21:00

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001720

Agency ALACHUA PD **Date** 06/01/2025 **Type** Agency Inspection
Inspector LEVY,JAYSON,I **Time** 10:50:57 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 10124080A2
Expiration 02/27/2026 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 06/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.207
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.081
0.20 g/210L Test # 2	0.208
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.051
0.08 g/210L Test # 3	0.081
0.20 g/210L Test # 3	0.210
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 06/01/2025 11:23

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001720

Agency ALACHUA PD **Date** 07/08/2025 **Type** Agency Inspection
Inspector LEVY,JAYSON,I **Time** 09:53:06 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 10124080A2
Expiration 02/27/2026 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 06/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 07/08/2025 09:56

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001720

Agency ALACHUA PD **Date** 08/05/2025 **Type** Agency Inspection
Inspector LEVY,JAYSON,I **Time** 16:20:07 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202406N **0.08 Gas Lot #** 10124080A2
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 06/20/2026 **Expiration** 06/05/2026

Number of Simulators 5 **Compliance** No

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	No
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	No

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.209
0.08 g/210L Dry Gas Std Test # 1	0.099
0.08 g/210L Dry Gas Std Test # (Repeat) 1	0.093
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.208
0.08 g/210L Dry Gas Std Test # 2	0.000
0.08 g/210L Dry Gas Std Test # (Repeat) 2	0.087
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.081
0.20 g/210L Test # 3	0.205
0.08 g/210L Dry Gas Std Test # 3	0.094
0.08 g/210L Dry Gas Std Test # (Repeat) 3	0.085
Interferent Detect Test # 3	INT

Remarks:

08: Control Outside Tolerance FLUID FROM .20 HOSE, Control . Non-compliance:08 DRY GAS.

Data Download Date/Time 08/06/2025 15:57

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001720

Agency	ALACHUA PD	**Date**	08/06/2025	**Type**	Agency Inspection
Inspector	LEVY,JAYSON,I	**Time**	15:50:13	**Software**	8100.27

0.05 Lot #	202402E	**0.08 Lot #**	202410E	**0.20 Lot #**	202406N	**0.08 Gas Lot #**	01524080A4
Expiration	02/27/2026	**Expiration**	10/22/2026	**Expiration**	06/20/2026	**Expiration**	03/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 08/06/2025 15:57

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001720

Agency	ALACHUA PD	**Date**	09/14/2025	**Type**	Agency Inspection
Inspector	LEVY,JAYSON,I	**Time**	13:18:16	**Software**	8100.27

0.05 Lot #	202402E	**0.08 Lot #**	202410E	**0.20 Lot #**	202406N	**0.08 Gas Lot #**	01524080A4
Expiration	02/27/2026	**Expiration**	10/22/2026	**Expiration**	06/20/2026	**Expiration**	03/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.047
0.08 g/210L Test # 1	0.076
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.076
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.047
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.077
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.047
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.077
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 09/14/2025 13:20

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001720

Agency ALACHUA PD **Date** 10/31/2025 **Type** Agency Inspection
Inspector LEVY,JAYSON,I **Time** 09:34:22 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** **0.08 Gas Lot #**
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** **Expiration**

Number of Simulators 5 **Compliance** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.093
Barometric Pressure Sensor Check: OK		0.08 g/210L Test # (Repeat) 1	0.091
Alcohol Free Subject Test: 0.000	Yes	Interferent Detect Test # 1	INT
Mouth Alcohol Test: Slope Not Met	Yes	Alcohol Free Test # 2	0.000
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test # 2	0.050
Diagnostic Check (Post-Inspection): OK	No	0.08 g/210L Test # 2	0.094
Alcohol Free Test: OK	Yes	0.08 g/210L Test # (Repeat) 2	0.199
0.05 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Test: OK	No	Alcohol Free Test # 3	0.000
0.20 g/210L Test: OK	No	0.05 g/210L Test # 3	0.050
0.08 g/210L Dry Gas Std Test: OK	No	0.08 g/210L Test # 3	0.093
		0.08 g/210L Test # (Repeat) 3	0.199
		Interferent Detect Test # 3	INT

Remarks:

08: Control Outside Tolerance SIM ISSUE, Control Outside To. Non-compliance:SIM ISSUE.

Data Download Date/Time 10/31/2025 10:35

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001720

Agency	ALACHUA PD	**Date**	10/31/2025	**Type**	Agency Inspection
Inspector	LEVY,JAYSON,I	**Time**	10:34:38	**Software**	8100.27

0.05 Lot #	202402E	**0.08 Lot #**	202410E	**0.20 Lot #**	202406N	**0.08 Gas Lot #**	01524080A4
Expiration	02/27/2026	**Expiration**	10/22/2026	**Expiration**	06/20/2026	**Expiration**	03/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.052
0.08 g/210L Test # 1	0.082
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.051
0.08 g/210L Test # 2	0.081
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.052
0.08 g/210L Test # 3	0.081
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 10/31/2025 10:35

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001720

Agency ALACHUA PD **Date** 11/26/2025 **Type** Agency Inspection
Inspector LEVY,JAYSON,I **Time** 07:46:49 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202406N **0.08 Gas Lot #** 01524080A4
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 06/20/2026 **Expiration** 03/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.076
0.20 g/210L Test # 1	0.195
0.08 g/210L Dry Gas Std Test # 1	0.076
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.047
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.195
0.08 g/210L Dry Gas Std Test # 2	0.077
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.194
0.08 g/210L Dry Gas Std Test # 3	0.076
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 11/26/2025 07:47

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001720

Agency	ALACHUA PD	**Date**	12/12/2025	**Type**	Agency Inspection
Inspector	LEVY,JAYSON,I	**Time**	18:15:23	**Software**	8100.27

0.05 Lot #	202402E	**0.08 Lot #**	202410E	**0.20 Lot #**	202406N	**0.08 Gas Lot #**	01524080A4
Expiration	02/27/2026	**Expiration**	10/22/2026	**Expiration**	06/20/2026	**Expiration**	03/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.200
0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.200
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.081
0.20 g/210L Test # 3	0.201
0.08 g/210L Dry Gas Std Test # 3	0.078
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 12/22/2025 14:28

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001720

Agency ALACHUA PD **Date** 12/31/2025 **Type** Department Inspection
Inspector PIERSON,WEN-CHI,K **Time** 13:47:00 **Software** 8100.27

0.05 Lot # 202406K **0.08 Lot #** 202406L **0.20 Lot #** 202406N **0.08 Gas Lot #** AG510701
Expiration 06/19/2026 **Expiration** 06/19/2026 **Expiration** 06/20/2026 **Expiration** 04/17/2027

Number of Simulators 5 **Average Standard Deviation** 0.0007 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.053
0.08 g/210L Test # 1	0.081
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.053
0.08 g/210L Test # 2	0.081
0.20 g/210L Test # 2	0.202
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.053
0.08 g/210L Test # 3	0.081
0.20 g/210L Test # 3	0.203
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.053
0.08 g/210L Test # 4	0.082
0.20 g/210L Test # 4	0.203
0.08 g/210L Dry Gas Std Test # 4	0.080
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.053
0.08 g/210L Test # 5	0.081
0.20 g/210L Test # 5	0.202
0.08 g/210L Dry Gas Std Test # 5	0.081
Interferent Detect Test # 5	INT

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001720

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.053
0.08 g/210L Test # 6	0.082
0.20 g/210L Test # 6	0.204
0.08 g/210L Dry Gas Std Test # 6	0.080
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.054
0.08 g/210L Test # 7	0.082
0.20 g/210L Test # 7	0.202
0.08 g/210L Dry Gas Std Test # 7	0.081
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.053
0.08 g/210L Test # 8	0.082
0.20 g/210L Test # 8	0.202
0.08 g/210L Dry Gas Std Test # 8	0.081
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.054
0.08 g/210L Test # 9	0.081
0.20 g/210L Test # 9	0.204
0.08 g/210L Dry Gas Std Test # 9	0.080
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.053
0.08 g/210L Test # 10	0.081
0.20 g/210L Test # 10	0.204
0.08 g/210L Dry Gas Std Test # 10	0.081
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0005
0.20 g/210L Test (Std Dev)	0.0017
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

Data Download Date/Time 12/31/2025 13:51

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

Agency PALM BAY PD **Date** 01/28/2025 **Type** Agency Inspection
Inspector LUTZ,EDWIN,J **Time** 09:43:36 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 24923080A4
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 10/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.047
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.191
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.192
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.194
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/28/2025 09:44

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

Agency PALM BAY PD **Date** 02/26/2025 **Type** Agency Inspection
Inspector HODGDON,TODD,S **Time** 11:58:19 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 24923080A4
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 10/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.192
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.194
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.194
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/26/2025 11:58

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

Agency PALM BAY PD **Date** 03/27/2025 **Type** Agency Inspection
Inspector FERINO,WESLEY,G **Time** 12:41:14 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 24923080A4
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 10/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.081
0.20 g/210L Test # 1	0.194
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.081
0.20 g/210L Test # 2	0.195
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.081
0.20 g/210L Test # 3	0.195
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

AI TODD HODGDON AND EDWIN LUTZ PRESENT AND OBSERVED

Data Download Date/Time 03/27/2025 12:44

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

Agency PALM BAY PD **Date** 04/30/2025 **Type** Agency Inspection
Inspector FERINO,WESLEY,G **Time** 10:46:21 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 24923080A4
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 10/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.000
0.20 g/210L Test # (Repeat) 1	0.193
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.000
0.20 g/210L Test # (Repeat) 2	0.194
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.000
0.20 g/210L Test # (Repeat) 3	0.195
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

RANGE EXCEEDED REDONE, CONTROL DUE TO LEAK,AI HODGEDON PRESR HAD A BREAK BY SEAL,REDID.

Data Download Date/Time 04/30/2025 10:54

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

Agency PALM BAY PD **Date** 05/23/2025 **Type** Agency Inspection
Inspector LUTZ,EDWIN,J **Time** 10:36:05 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 24923080A4
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 10/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.076
0.20 g/210L Test # 1	0.194
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.077
0.20 g/210L Test # 2	0.195
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.077
0.20 g/210L Test # 3	0.195
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 05/23/2025 10:37

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

Agency PALM BAY PD **Date** 06/26/2025 **Type** Agency Inspection
Inspector HODGDON,TODD,S **Time** 15:00:41 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 24923080A4
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 10/05/2025

Number of Simulators 5 **Compliance** No

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	No
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	No

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.046
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.196
0.08 g/210L Dry Gas Std Test # 1	0.095
0.08 g/210L Dry Gas Std Test # (Repeat) 1	0.091
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.077
0.20 g/210L Test # 2	0.196
0.08 g/210L Dry Gas Std Test # 2	0.092
0.08 g/210L Dry Gas Std Test # (Repeat) 2	0.089
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.196
0.08 g/210L Dry Gas Std Test # 3	0.091
0.08 g/210L Dry Gas Std Test # (Repeat) 3	0.088
Interferent Detect Test # 3	INT

Remarks:

08: Control Outside Tolerance, Control Outside Tolerance. Non-compliance:ISSUE WITH ROOM.

Data Download Date/Time 07/10/2025 10:07

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

Agency PALM BAY PD **Date** 07/29/2025 **Type** Department Inspection
Inspector HIGGINBOTHAM,LEANDRA, **Time** 17:47:28 **Software** 8100.27

0.05 Lot # 202406K **0.08 Lot #** 202406L **0.20 Lot #** 202406N **0.08 Gas Lot #** AG429602
Expiration 06/19/2026 **Expiration** 06/19/2026 **Expiration** 06/20/2026 **Expiration** 10/22/2026

Number of Simulators 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.195
0.08 g/210L Dry Gas Std Test # 1	0.073
0.08 g/210L Dry Gas Std Test # (Repeat) 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.196
0.08 g/210L Dry Gas Std Test # 2	0.077
0.08 g/210L Dry Gas Std Test # (Repeat) 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.196
0.08 g/210L Dry Gas Std Test # 3	0.079
0.08 g/210L Dry Gas Std Test # (Repeat) 3	0.081
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.048
0.08 g/210L Test # 4	0.078
0.20 g/210L Test # 4	0.196
0.08 g/210L Dry Gas Std Test # 4	0.079
0.08 g/210L Dry Gas Std Test # (Repeat) 4	0.081
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.048

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

0.08 g/210L Test # 5	0.078
0.20 g/210L Test # 5	0.197
0.08 g/210L Dry Gas Std Test # 5	0.080
0.08 g/210L Dry Gas Std Test # (Repeat) 5	0.081
Interferent Detect Test # 5	INT
Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.048
0.08 g/210L Test # 6	0.079
0.20 g/210L Test # 6	0.196
0.08 g/210L Dry Gas Std Test # 6	0.080
0.08 g/210L Dry Gas Std Test # (Repeat) 6	0.081
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.049
0.08 g/210L Test # 7	0.079
0.20 g/210L Test # 7	0.196
0.08 g/210L Dry Gas Std Test # 7	0.081
0.08 g/210L Dry Gas Std Test # (Repeat) 7	0.081
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.049
0.08 g/210L Test # 8	0.079
0.20 g/210L Test # 8	0.196
0.08 g/210L Dry Gas Std Test # 8	0.080
0.08 g/210L Dry Gas Std Test # (Repeat) 8	0.081
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.049
0.08 g/210L Test # 9	0.079
0.20 g/210L Test # 9	0.197
0.08 g/210L Dry Gas Std Test # 9	0.081
0.08 g/210L Dry Gas Std Test # (Repeat) 9	0.081
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.048
0.08 g/210L Test # 10	0.079
0.20 g/210L Test # 10	0.196
0.08 g/210L Dry Gas Std Test # 10	0.081
0.08 g/210L Dry Gas Std Test # (Repeat) 10	0.081
Interferent Detect Test # 10	INT

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0005
0.20 g/210L Test (Std Dev)	0.0005
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0024
0.08 g/210L Dry Gas Std Test (Std Dev) (Repeat)	0.0003

Remarks:

08: Control Outside Tolerance.TIGHTENED CYLINDER, RERAN

Data Download Date/Time 07/29/2025 17:46

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

Agency PALM BAY PD **Date** 08/25/2025 **Type** Agency Inspection
Inspector LUTZ,EDWIN,J **Time** 04:50:48 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 24923080A4
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 10/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.074
0.08 g/210L Dry Gas Std Test # (Repeat) 1	0.078
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.077
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.077
0.08 g/210L Dry Gas Std Test # (Repeat) 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.078
0.08 g/210L Dry Gas Std Test # (Repeat) 3	0.080
Interferent Detect Test # 3	INT

Remarks:

08: Control Outside Tolerance.08DG NOT SEATED

Data Download Date/Time 08/25/2025 04:49

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

Agency PALM BAY PD **Date** 09/30/2025 **Type** Agency Inspection
Inspector HODGDON,TODD,S **Time** 17:07:08 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 24923080A4
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 10/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.047
0.08 g/210L Test # 1	0.075
0.20 g/210L Test # 1	0.192
0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.047
0.08 g/210L Test # 2	0.076
0.20 g/210L Test # 2	0.193
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.047
0.08 g/210L Test # 3	0.076
0.20 g/210L Test # 3	0.194
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 09/30/2025 17:16

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

Agency PALM BAY PD **Date** 10/23/2025 **Type** Agency Inspection
Inspector HODGDON,TODD,S **Time** 19:55:11 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 32324080A2
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2024 **Expiration** 12/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.195
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.077
0.20 g/210L Test # 2	0.196
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.077
0.20 g/210L Test # 3	0.196
0.08 g/210L Dry Gas Std Test # 3	0.082
Interferent Detect Test # 3	INT

Remarks:

PASS

Data Download Date/Time 10/23/2025 19:56

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

Agency	PALM BAY PD	**Date**	11/27/2025	**Type**	Agency Inspection
Inspector	BARNETT,JAMES,C	**Time**	19:43:03	**Software**	8100.27

0.05 Lot #	**0.08 Lot #**	**0.20 Lot #**	**0.08 Gas Lot #**
Expiration	**Expiration**	**Expiration**	**Expiration**

Number of Simulators **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Data Download Date/Time 11/27/2025 20:32

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

Agency PALM BAY PD **Date** 11/27/2025 **Type** Agency Inspection
Inspector BARNETT,JAMES,C **Time** 20:25:23 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 32324080A2
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 12/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.047
0.08 g/210L Test # 1	0.076
0.20 g/210L Test # 1	0.194
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.047
0.08 g/210L Test # 2	0.076
0.20 g/210L Test # 2	0.195
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.076
0.20 g/210L Test # 3	0.196
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

REVIEW

Data Download Date/Time 11/27/2025 20:32

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

Agency PALM BAY PD **Date** 12/26/2025 **Type** Agency Inspection
Inspector LUTZ,EDWIN,J **Time** 19:13:25 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 32324080A2
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 12/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.046
0.08 g/210L Test # 1	0.076
0.20 g/210L Test # 1	0.194
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.046
0.08 g/210L Test # 2	0.076
0.20 g/210L Test # 2	0.194
0.08 g/210L Dry Gas Std Test # 2	0.082
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.047
0.08 g/210L Test # 3	0.077
0.20 g/210L Test # 3	0.194
0.08 g/210L Dry Gas Std Test # 3	0.082
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 12/26/2025 19:11

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

Agency LEE COUNTY SO **Date** 01/22/2025 **Type** Agency Inspection
Inspector DUNCAN,NICHOLAS, **Time** 11:55:36 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202307C **0.08 Gas Lot #** 31124080A2
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 07/12/2025 **Expiration** 12/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.046
0.08 g/210L Test # 1	0.075
0.20 g/210L Test # 1	0.191
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.047
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.195
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.047
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.195
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/22/2025 11:56

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

Agency LEE COUNTY SO **Date** 02/19/2025 **Type** Agency Inspection
Inspector DUNCAN,NICHOLAS, **Time** 12:49:59 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202307C **0.08 Gas Lot #** 31124080A2
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 07/12/2025 **Expiration** 12/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.047
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.190
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.194
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.194
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/19/2025 12:54

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

Agency LEE COUNTY SO **Date** 03/25/2025 **Type** Agency Inspection
Inspector DUNCAN,NICHOLAS, **Time** 11:18:14 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202307C **0.08 Gas Lot #** 01524080A2
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 07/12/2025 **Expiration** 03/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.047
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.190
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.193
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.193
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 03/25/2025 11:22

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

Agency LEE COUNTY SO **Date** 04/01/2025 **Type** Department Inspection
Inspector GUTSCHOW,TAYLOR,D **Time** 16:25:36 **Software** 8100.27

0.05 Lot # 202406K **0.08 Lot #** 202406L **0.20 Lot #** 202304C **0.08 Gas Lot #** AG429602
Expiration 06/19/2026 **Expiration** 06/19/2026 **Expiration** 04/05/2025 **Expiration** 10/22/2026

Number of Simulators 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.196
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.196
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.196
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.048
0.08 g/210L Test # 4	0.079
0.20 g/210L Test # 4	0.196
0.08 g/210L Dry Gas Std Test # 4	0.080
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.048
0.08 g/210L Test # 5	0.079
0.20 g/210L Test # 5	0.197
0.08 g/210L Dry Gas Std Test # 5	0.080
Interferent Detect Test # 5	INT

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.048
0.08 g/210L Test # 6	0.080
0.20 g/210L Test # 6	0.196
0.08 g/210L Dry Gas Std Test # 6	0.080
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.049
0.08 g/210L Test # 7	0.079
0.20 g/210L Test # 7	0.196
0.08 g/210L Dry Gas Std Test # 7	0.080
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.049
0.08 g/210L Test # 8	0.079
0.20 g/210L Test # 8	0.196
0.08 g/210L Dry Gas Std Test # 8	0.080
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.049
0.08 g/210L Test # 9	0.079
0.20 g/210L Test # 9	0.197
0.08 g/210L Dry Gas Std Test # 9	0.080
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.049
0.08 g/210L Test # 10	0.079
0.20 g/210L Test # 10	0.196
0.08 g/210L Dry Gas Std Test # 10	0.080
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0005
0.08 g/210L Test (Std Dev)	0.0003
0.20 g/210L Test (Std Dev)	0.0004
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0000

Remarks:

Data Download Date/Time 04/01/2025 16:26

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

Agency LEE COUNTY SO **Date** 04/23/2025 **Type** Agency Inspection
Inspector DUNCAN,NICHOLAS, **Time** 09:11:16 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202307C **0.08 Gas Lot #** 24923080A3
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 07/12/2025 **Expiration** 10/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.047
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.191
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.047
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.193
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.047
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.194
0.08 g/210L Dry Gas Std Test # 3	0.082
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 04/23/2025 09:12

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

Agency LEE COUNTY SO **Date** 05/08/2025 **Type** Agency Inspection
Inspector DUNCAN,NICHOLAS, **Time** 09:28:47 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202307C **0.08 Gas Lot #** 24923080A3
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 07/12/2025 **Expiration** 10/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # (Repeat) 1	0.000
0.05 g/210L Test # 1	0.047
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.192
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # (Repeat) 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.194
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # (Repeat) 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.195
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

00: Ambient Fail-CLEAR AREA- RETEST.

Data Download Date/Time 05/08/2025 09:28

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

Agency LEE COUNTY SO **Date** 06/09/2025 **Type** Agency Inspection
Inspector DUNCAN,NICHOLAS, **Time** 11:33:33 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 24923080A3
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 10/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # (Repeat) 1	0.000
0.05 g/210L Test # 1	0.047
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.192
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # (Repeat) 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.194
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # (Repeat) 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.196
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

00: Ambient Fail-CLEAR AREA-RETEST.

Data Download Date/Time 06/09/2025 11:36

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

Agency LEE COUNTY SO **Date** 07/22/2025 **Type** Agency Inspection
Inspector DUNCAN,NICHOLAS, **Time** 08:40:07 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 24923080A3
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 10/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # (Repeat) 1	0.000
0.05 g/210L Test # 1	0.046
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.193
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # (Repeat) 2	0.000
0.05 g/210L Test # 2	0.047
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.196
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # (Repeat) 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.082
Interferent Detect Test # 3	INT

Remarks:

00: Ambient Fail-CLEAR AREA-RETEST.

Data Download Date/Time 07/22/2025 08:40

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

Agency LEE COUNTY SO **Date** 08/27/2025 **Type** Agency Inspection
Inspector DUNCAN,NICHOLAS, **Time** 09:40:50 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 24923080A3
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 10/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.046
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.193
0.08 g/210L Dry Gas Std Test # 1	0.082
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.047
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.196
0.08 g/210L Dry Gas Std Test # 2	0.082
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.195
0.08 g/210L Dry Gas Std Test # 3	0.082
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 08/27/2025 09:43

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

Agency LEE COUNTY SO **Date** 09/22/2025 **Type** Agency Inspection
Inspector DUNCAN,NICHOLAS, **Time** 10:12:21 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 31124080A2
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 12/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.046
0.08 g/210L Test # 1	0.076
0.20 g/210L Test # 1	0.192
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.047
0.08 g/210L Test # 2	0.077
0.20 g/210L Test # 2	0.195
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.195
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 09/22/2025 10:10

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

Agency LEE COUNTY SO **Date** 10/21/2025 **Type** Agency Inspection
Inspector DUNCAN,NICHOLAS, **Time** 11:12:37 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 31124080A2
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 12/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.046
0.08 g/210L Test # 1	0.076
0.20 g/210L Test # 1	0.193
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.195
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.196
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

A F / M A:SLOPE NOT MET- RETEST.

Data Download Date/Time 10/21/2025 11:11

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

Agency LEE COUNTY SO **Date** 11/25/2025 **Type** Agency Inspection
Inspector DUNCAN,NICHOLAS, **Time** 10:30:09 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 13225080A3
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 06/05/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.046
0.08 g/210L Test # 1	0.075
0.20 g/210L Test # 1	0.191
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.047
0.08 g/210L Test # 2	0.077
0.20 g/210L Test # 2	0.195
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.047
0.08 g/210L Test # 3	0.077
0.20 g/210L Test # 3	0.195
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/22/2026 14:26

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

Agency LEE COUNTY SO **Date** 12/16/2025 **Type** Agency Inspection
Inspector DUNCAN,NICHOLAS, **Time** 10:03:01 **Software** 8100.27

0.05 Lot # 202506E **0.08 Lot #** 202510E **0.20 Lot #** 202410F **0.08 Gas Lot #** 31124080A2
Expiration 06/24/2027 **Expiration** 10/28/2027 **Expiration** 10/30/2026 **Expiration** 12/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.047
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.191
0.08 g/210L Dry Gas Std Test # 1	0.082
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.193
0.08 g/210L Dry Gas Std Test # 2	0.082
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.195
0.08 g/210L Dry Gas Std Test # 3	0.082
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/22/2026 14:26

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001723

Agency VENICE PD **Date** 08/30/2025 **Type** Agency Inspection
Inspector NELSON,ANDREW,G **Time** 12:48:41 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 23324080A2
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 10/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.200
0.08 g/210L Dry Gas Std Test # 1	0.059
0.08 g/210L Dry Gas Std Test # (Repeat) 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.200
0.08 g/210L Dry Gas Std Test # 2	0.081
0.08 g/210L Dry Gas Std Test # (Repeat) 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.200
0.08 g/210L Dry Gas Std Test # 3	0.081
0.08 g/210L Dry Gas Std Test # (Repeat) 3	0.081
Interferent Detect Test # 3	INT

Remarks:

08: Control Outside Tolerance.

Data Download Date/Time 08/30/2025 12:45

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

Agency PINELLAS PARK PD **Date** 01/31/2025 **Type** Agency Inspection
Inspector ADAMS,STERLING,M **Time** 02:36:48 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 28423080A4
Expiration 02/27/2026 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 11/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

N/A

Data Download Date/Time 01/31/2025 02:53

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

Agency PINELLAS PARK PD **Date** 02/01/2025 **Type** Agency Inspection
Inspector ADAMS,STERLING,M **Time** 19:36:30 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 28423080A4
Expiration 02/27/2026 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 11/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.077
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.078
Interferent Detect Test # 3	INT

Remarks:

N/A

Data Download Date/Time 02/01/2025 19:52

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

Agency PINELLAS PARK PD **Date** 03/30/2025 **Type** Agency Inspection
Inspector ADAMS,STERLING,M **Time** 05:10:27 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 28423080A4
Expiration 02/27/2026 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 11/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.047
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.196
0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.077
0.20 g/210L Test # 2	0.196
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.047
0.08 g/210L Test # 3	0.077
0.20 g/210L Test # 3	0.196
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

N/A

Data Download Date/Time 04/02/2025 20:43

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

Agency PINELLAS PARK PD **Date** 04/02/2025 **Type** Agency Inspection
Inspector ADAMS,STERLING,M **Time** 20:36:45 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 28423080A4
Expiration 02/27/2026 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 11/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # (Repeat) 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test # 1	INT
Alcohol Free Test # (Repeat) 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # (Repeat) 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.196
0.08 g/210L Dry Gas Std Test # 3	0.078
Interferent Detect Test # 3	INT

Remarks:

00: Ambient Fail / CLEARED ROOM AND AREA AROUND INSTRUMENT . N/A

Data Download Date/Time 04/02/2025 20:43

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

Agency PINELLAS PARK PD **Date** 05/31/2025 **Type** Agency Inspection
Inspector ADAMS,STERLING,M **Time** 12:35:38 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 28423080A4
Expiration 02/27/2026 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 11/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.047
0.08 g/210L Test # 1	0.076
0.20 g/210L Test # 1	0.193
0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.076
0.20 g/210L Test # 2	0.194
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.077
0.20 g/210L Test # 3	0.194
0.08 g/210L Dry Gas Std Test # 3	0.078
Interferent Detect Test # 3	INT

Remarks:

N/A

Data Download Date/Time 06/04/2025 17:50

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

Agency PINELLAS PARK PD **Date** 06/04/2025 **Type** Agency Inspection
Inspector ADAMS,STERLING,M **Time** 17:34:16 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202307C **0.08 Gas Lot #** 28423080A4
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 07/12/2025 **Expiration** 11/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # (Repeat) 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.195
0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test # 1	INT
Alcohol Free Test # (Repeat) 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.077
0.20 g/210L Test # 2	0.195
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # (Repeat) 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.077
0.20 g/210L Test # 3	0.195
0.08 g/210L Dry Gas Std Test # 3	0.078
Interferent Detect Test # 3	INT

Remarks:

00: Ambient Fail. CLEARED ROOM AND VACATED AREA AROUND INST

Data Download Date/Time 06/04/2025 17:50

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

Agency PINELLAS PARK PD **Date** 07/31/2025 **Type** Agency Inspection
Inspector ADAMS,STERLING,M **Time** 18:59:07 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 28423080A4
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 11/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.076
0.20 g/210L Test # 1	0.195
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.077
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

N/A

Data Download Date/Time 08/01/2025 07:09

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

Agency PINELLAS PARK PD **Date** 08/01/2025 **Type** Agency Inspection
Inspector ADAMS,STERLING,M **Time** 06:59:44 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 28423080A4
Expiration 02/27/2026 **Expiration** 10/22/2025 **Expiration** 10/30/2026 **Expiration** 11/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.047
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.195
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.047
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.196
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.077
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

N/A

Data Download Date/Time 08/01/2025 07:09

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

Agency PINELLAS PARK PD **Date** 09/02/2025 **Type** Agency Inspection
Inspector ADAMS,STERLING,M **Time** 22:38:50 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 28423080A4
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 11/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.047
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.194
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.196
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

OUT OF SERVICE

Data Download Date/Time 09/02/2025 22:54

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

Agency PINELLAS PARK PD **Date** 10/06/2025 **Type** Department Inspection
Inspector GUTSCHOW,TAYLOR,D **Time** 15:15:03 **Software** 8100.27

0.05 Lot # 202406K **0.08 Lot #** 202406L **0.20 Lot #** 202406N **0.08 Gas Lot #** AG526603
Expiration 06/19/2026 **Expiration** 06/19/2026 **Expiration** 06/20/2026 **Expiration** 09/23/2027

Number of Simulators 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.047
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.195
0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.077
0.20 g/210L Test # 2	0.195
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.047
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.196
0.08 g/210L Dry Gas Std Test # 3	0.078
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.048
0.08 g/210L Test # 4	0.078
0.20 g/210L Test # 4	0.195
0.08 g/210L Dry Gas Std Test # 4	0.078
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.048
0.08 g/210L Test # 5	0.077
0.20 g/210L Test # 5	0.195
0.08 g/210L Dry Gas Std Test # 5	0.078
Interferent Detect Test # 5	INT

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.048
0.08 g/210L Test # 6	0.078
0.20 g/210L Test # 6	0.195
0.08 g/210L Dry Gas Std Test # 6	0.078
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.048
0.08 g/210L Test # 7	0.078
0.20 g/210L Test # 7	0.196
0.08 g/210L Dry Gas Std Test # 7	0.078
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.048
0.08 g/210L Test # 8	0.078
0.20 g/210L Test # 8	0.195
0.08 g/210L Dry Gas Std Test # 8	0.078
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.048
0.08 g/210L Test # 9	0.079
0.20 g/210L Test # 9	0.196
0.08 g/210L Dry Gas Std Test # 9	0.078
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.048
0.08 g/210L Test # 10	0.078
0.20 g/210L Test # 10	0.195
0.08 g/210L Dry Gas Std Test # 10	0.078
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0006
0.20 g/210L Test (Std Dev)	0.0004
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0000

Remarks:

Data Download Date/Time 10/06/2025 15:15

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

Agency PINELLAS PARK PD **Date** 11/30/2025 **Type** Agency Inspection
Inspector ADAMS,STERLING,M **Time** 19:24:49 **Software** 8100.27

0.05 Lot # 202002E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 19825080A1
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 09/05/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
Alcohol Free Test # (Repeat) 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
Alcohol Free Test # (Repeat) 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # (Repeat) 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.201
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

00: RFI Detect / RM CLRD.INSERVICE

Data Download Date/Time 12/04/2025 00:15

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

Agency PINELLAS PARK PD **Date** 12/01/2025 **Type** Agency Inspection
Inspector ADAMS,STERLING,M **Time** 04:36:05 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 19825080A1
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 09/05/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.196
0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.200
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.051
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.200
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

N/A

Data Download Date/Time 12/04/2025 00:15

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001737

Agency	SOUTH MIAMI PD	**Date**	02/05/2025	**Type**	Agency Inspection
Inspector	GUTSCHOW,TAYLOR,D	**Time**	11:13:20	**Software**	8100.27

0.05 Lot #	**0.08 Lot #**	**0.20 Lot #**	**0.08 Gas Lot #**
Expiration	**Expiration**	**Expiration**	**Expiration**

Number of Simulators **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

AI NOT CONDUCTED. BYPASSED TO BRING OUT OF DISABLED MODE.

Data Download Date/Time 02/13/2025 09:22

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001737

Agency SOUTH MIAMI PD **Date** 02/14/2025 **Type** Department Inspection
Inspector GUTSCHOW,TAYLOR,D **Time** 11:57:25 **Software** 8100.27

0.05 Lot # 202303K **0.08 Lot #** 202303L **0.20 Lot #** 202304C **0.08 Gas Lot #** AG429602
Expiration 03/29/2025 **Expiration** 03/29/2025 **Expiration** 04/05/2025 **Expiration** 10/22/2026

Number of Simulators 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.050
0.08 g/210L Test # 4	0.080
0.20 g/210L Test # 4	0.198
0.08 g/210L Dry Gas Std Test # 4	0.080
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.050
0.08 g/210L Test # 5	0.080
0.20 g/210L Test # 5	0.197
0.08 g/210L Dry Gas Std Test # 5	0.080
Interferent Detect Test # 5	INT

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001737

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.050
0.08 g/210L Test # 6	0.079
0.20 g/210L Test # 6	0.197
0.08 g/210L Dry Gas Std Test # 6	0.080
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.080
0.20 g/210L Test # 7	0.197
0.08 g/210L Dry Gas Std Test # 7	0.081
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.051
0.08 g/210L Test # 8	0.080
0.20 g/210L Test # 8	0.198
0.08 g/210L Dry Gas Std Test # 8	0.081
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.051
0.08 g/210L Test # 9	0.080
0.20 g/210L Test # 9	0.198
0.08 g/210L Dry Gas Std Test # 9	0.080
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.051
0.08 g/210L Test # 10	0.080
0.20 g/210L Test # 10	0.197
0.08 g/210L Dry Gas Std Test # 10	0.080
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0005
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Data Download Date/Time 02/14/2025 11:55

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001737

Agency SOUTH MIAMI PD **Date** 04/24/2025 **Type** Agency Inspection
Inspector MUNOZ,ROBERT,A **Time** 16:30:58 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** AG326805
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 09/25/2025

Number of Simulators 5 **Compliance** No

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	No
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	No

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.000
0.08 g/210L Dry Gas Std Test # (Repeat) 1	0.000
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.200
0.08 g/210L Dry Gas Std Test # 2	0.000
0.08 g/210L Dry Gas Std Test # (Repeat) 2	0.000
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.000
0.08 g/210L Dry Gas Std Test # (Repeat) 3	0.000
Interferent Detect Test # 3	INT

Remarks:

08: GAS CYLINDER OUT OF GAS, Control Outside Tolerance. Non-compliance:NO GAS IN CYLINDER.

Data Download Date/Time 04/24/2025 17:14

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001737

Agency SOUTH MIAMI PD **Date** 04/24/2025 **Type** Agency Inspection
Inspector MUNOZ,ROBERT,A **Time** 17:15:23 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 403207507
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 11/26/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.046
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.200
0.08 g/210L Dry Gas Std Test # 1	0.000
0.08 g/210L Dry Gas Std Test # (Repeat) 1	0.082
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.082
0.08 g/210L Dry Gas Std Test # (Repeat) 2	0.083
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.200
0.08 g/210L Dry Gas Std Test # 3	0.082
0.08 g/210L Dry Gas Std Test # (Repeat) 3	0.082
Interferent Detect Test # 3	INT

Remarks:

08: GAS CYLINDER NOT CONNECTED.

Data Download Date/Time 04/24/2025 17:14

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001737

Agency SOUTH MIAMI PD **Date** 08/14/2025 **Type** Agency Inspection
Inspector VEGA,OBEL, **Time** 19:53:17 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #**
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration**

Number of Simulators 5 **Compliance** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # (Repeat) 1	0.201
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	No	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # (Repeat) 2	0.200
0.08 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.20 g/210L Test: OK	No	Alcohol Free Test # 3	0.000
0.08 g/210L Dry Gas Std Test: OK	No	0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # (Repeat) 3	
		Interferent Detect Test # 3	INT

Remarks:

20: RFI Detect, RFI Detect. RFI DETCTED DURING .20G/210L CONTROL TEST..

Data Download Date/Time 08/14/2025 19:51

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001737

Agency	SOUTH MIAMI PD	**Date**	10/31/2025	**Type**	Agency Inspection
Inspector	MUNOZ,ROBERT,A	**Time**	11:33:22	**Software**	8100.27

0.05 Lot #	202402E	**0.08 Lot #**	202410E	**0.20 Lot #**	202410F	**0.08 Gas Lot #**	15025080A2
Expiration	02/27/2026	**Expiration**	10/22/2026	**Expiration**	10/30/2026	**Expiration**	07/05/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	Yes	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Time-Date changed.

Data Download Date/Time 10/31/2025 11:32

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

Agency	CHARLOTTE COUNTY SO	**Date**	01/12/2025	**Type**	Agency Inspection
Inspector	CHANDLER,VANESSA,N	**Time**	19:52:42	**Software**	8100.27

0.05 Lot #	202402E	**0.08 Lot #**	202307A	**0.20 Lot #**	202307C	**0.08 Gas Lot #**	06723080A5
Expiration	02/27/2026	**Expiration**	07/11/2025	**Expiration**	07/12/2025	**Expiration**	04/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.076
		0.20 g/210L Test # 3	0.195
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

CLEANED AND INSPECTED

Data Download Date/Time 01/12/2025 19:50

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

Agency CHARLOTTE COUNTY SO **Date** 02/04/2025 **Type** Agency Inspection
Inspector LEONARD,ALEYA,D **Time** 21:12:34 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 06723080A5
Expiration 02/27/2026 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 04/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.075
0.20 g/210L Test # 1	0.191
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.076
0.20 g/210L Test # 2	0.193
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.075
0.20 g/210L Test # 3	0.193
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

CLEANED AND INSPECTED

Data Download Date/Time 02/04/2025 21:15

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

Agency CHARLOTTE COUNTY SO **Date** 03/09/2025 **Type** Agency Inspection
Inspector DEMPSTER,PAUL,R **Time** 21:31:58 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 06723080A5
Expiration 02/27/2026 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 04/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.196
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

CLEANED AND INSPECTED

Data Download Date/Time 03/09/2025 21:34

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

Agency CHARLOTTE COUNTY SO **Date** 04/24/2025 **Type** Agency Inspection
Inspector BUTZER,PAIGE, **Time** 06:56:38 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #**
Expiration 02/27/2026 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration**

Number of Simulators 5 **Compliance** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.174
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # (Repeat) 1	0.194
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	No	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.171
0.08 g/210L Test: OK	Yes	0.20 g/210L Test # (Repeat) 2	0.191
0.20 g/210L Test: OK	No	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	No	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.076
		0.20 g/210L Test # 3	0.173
		0.20 g/210L Test # (Repeat) 3	0.189
		Interferent Detect Test # 3	INT

Remarks:

20: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: .

Data Download Date/Time 04/24/2025 06:58

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

Agency CHARLOTTE COUNTY SO **Date** 05/25/2025 **Type** Agency Inspection
Inspector LEONARD,ALEYA,D **Time** 21:49:53 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 28423080A4
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 11/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.200
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

CLEANED AND INSPECTED

Data Download Date/Time 05/25/2025 21:53

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

Agency	CHARLOTTE COUNTY SO	**Date**	07/18/2025	**Type**	Agency Inspection
Inspector	GUTSCHOW,TAYLOR,D	**Time**	12:17:33	**Software**	8100.27

0.05 Lot #	**0.08 Lot #**	**0.20 Lot #**	**0.08 Gas Lot #**
Expiration	**Expiration**	**Expiration**	**Expiration**

Number of Simulators **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

BYPASSED AI TO OPERATE INSTRUMENT. COMPLIANCE UNDETERMINED.

Data Download Date/Time 07/18/2025 15:05

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

Agency CHARLOTTE COUNTY SO **Date** 07/18/2025 **Type** Department Inspection
Inspector GUTSCHOW,TAYLOR,D **Time** 15:07:47 **Software** 8100.27

0.05 Lot # 202406K **0.08 Lot #** 202406L **0.20 Lot #** 202406N **0.08 Gas Lot #** AG429602
Expiration 06/19/2026 **Expiration** 06/19/2026 **Expiration** 06/20/2026 **Expiration** 10/22/2026

Number of Simulators 5 **Average Standard Deviation** 0.0002 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.077
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.077
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.078
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.049
0.08 g/210L Test # 4	0.079
0.20 g/210L Test # 4	0.197
0.08 g/210L Dry Gas Std Test # 4	0.079
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.049
0.08 g/210L Test # 5	0.079
0.20 g/210L Test # 5	0.197
0.08 g/210L Dry Gas Std Test # 5	0.078
Interferent Detect Test # 5	INT

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.049
0.08 g/210L Test # 6	0.079
0.20 g/210L Test # 6	0.197
0.08 g/210L Dry Gas Std Test # 6	0.079
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.049
0.08 g/210L Test # 7	0.079
0.20 g/210L Test # 7	0.197
0.08 g/210L Dry Gas Std Test # 7	0.078
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.050
0.08 g/210L Test # 8	0.079
0.20 g/210L Test # 8	0.197
0.08 g/210L Dry Gas Std Test # 8	0.078
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.049
0.08 g/210L Test # 9	0.079
0.20 g/210L Test # 9	0.197
0.08 g/210L Dry Gas Std Test # 9	0.078
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.049
0.08 g/210L Test # 10	0.079
0.20 g/210L Test # 10	0.197
0.08 g/210L Dry Gas Std Test # 10	0.078
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0003
0.08 g/210L Test (Std Dev)	0.0000
0.20 g/210L Test (Std Dev)	0.0000
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0006

Remarks:

Data Download Date/Time 07/18/2025 15:05

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

Agency CHARLOTTE COUNTY SO **Date** 08/22/2025 **Type** Agency Inspection
Inspector CHANDLER,VANESSA,N **Time** 19:50:59 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 28423080A4
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 11/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.201
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.081
0.20 g/210L Test # 2	0.200
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.081
0.20 g/210L Test # 3	0.201
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

CLEANED AND INSPECTED

Data Download Date/Time 08/22/2025 19:50

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

Agency CHARLOTTE COUNTY SO **Date** 09/17/2025 **Type** Agency Inspection
Inspector LEONARD,ALEYA,D **Time** 22:41:32 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 28423080A4
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 11/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.200
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

CLEANED AND INSPECTED

Data Download Date/Time 09/17/2025 22:41

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

Agency CHARLOTTE COUNTY SO **Date** 10/21/2025 **Type** Agency Inspection
Inspector LEONARD,ALEYA,D **Time** 21:48:20 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 15025080A1
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 07/05/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

CLEANED AND INSPECTED

Data Download Date/Time 10/21/2025 21:47

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

Agency CHARLOTTE COUNTY SO **Date** 11/26/2025 **Type** Agency Inspection
Inspector LEONARD,ALEYA,D **Time** 21:39:30 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 15025080A1
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 07/05/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

CLEANED AND INSPECTED

Data Download Date/Time 11/26/2025 21:39

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

Agency CHARLOTTE COUNTY SO **Date** 12/27/2025 **Type** Agency Inspection
Inspector CHANDLER,VANESSA,N **Time** 20:34:05 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 15025080A1
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 07/05/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

CLEANED AND INSPECTED

Data Download Date/Time 12/27/2025 20:35

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

Agency CLEARWATER PD **Date** 01/30/2025 **Type** Agency Inspection
Inspector HURT,JONATHAN,R **Time** 19:20:23 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 14323080A3
Expiration 02/27/2025 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 09/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.202
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.051
0.08 g/210L Test # 2	0.081
0.20 g/210L Test # 2	0.203
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.052
0.08 g/210L Test # 3	0.081
0.20 g/210L Test # 3	0.202
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

ACETONE LOT #2023-B, M/A LOT

Data Download Date/Time 02/03/2025 23:06

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

Agency	CLEARWATER PD	**Date**	03/02/2025	**Type**	Agency Inspection
Inspector	NUGENT,DAVID,A	**Time**	02:16:28	**Software**	8100.27

0.05 Lot #	**0.08 Lot #**	**0.20 Lot #**	**0.08 Gas Lot #**
Expiration	**Expiration**	**Expiration**	**Expiration**

Number of Simulators 5 **Compliance** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	Alcohol Free Test # 2	0.000
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	Yes		
Mouth Alcohol Test: Slope Not Met	Yes		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A: RFI Detect. 00: Sequence Aborted. Non-compliance:ABORTED.

Data Download Date/Time 03/02/2025 02:52

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

Agency CLEARWATER PD **Date** 03/02/2025 **Type** Agency Inspection
Inspector NUGENT,DAVID,A **Time** 02:53:07 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 14323080A3
Expiration 02/27/2026 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 09/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.047
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	0.000
Interferent Detect Test # (Repeat) 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.196
0.08 g/210L Dry Gas Std Test # 2	0.084
Interferent Detect Test # 2	0.000
Interferent Detect Test # (Repeat) 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.076
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.082
Interferent Detect Test # 3	0.000
Interferent Detect Test # (Repeat) 3	INT

Remarks:

Int Det:NOT DETECTED.ACT2023-B;M/A2024-A

Data Download Date/Time 03/02/2025 02:52

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

Agency	CLEARWATER PD	**Date**	04/29/2025	**Type**	Agency Inspection
Inspector	HURT,JONATHAN,R	**Time**	20:14:33	**Software**	8100.27

0.05 Lot #	202402E	**0.08 Lot #**	202307A	**0.20 Lot #**	202307C	**0.08 Gas Lot #**	01524080A2
Expiration	02/27/2026	**Expiration**	07/11/2025	**Expiration**	07/12/2025	**Expiration**	03/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.193
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.046
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.047
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.196
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

M/A LOT 2024-A, ACETONE LOT 2023-B

Data Download Date/Time 05/03/2025 00:11

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

Agency CLEARWATER PD **Date** 05/27/2025 **Type** Agency Inspection
Inspector HURT,JONATHAN,R **Time** 23:07:01 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 01524080A2
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 03/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.047
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.051
0.08 g/210L Test # 2	0.081
0.20 g/210L Test # 2	0.200
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.051
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.200
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

M/A LOT # 2024-A, ACETONE LOT#2023-B

Data Download Date/Time 05/31/2025 03:51

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

Agency CLEARWATER PD **Date** 06/30/2025 **Type** Agency Inspection
Inspector HURT,JONATHAN,R **Time** 06:30:11 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 31124080A2
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 12/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.200
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.200
0.08 g/210L Dry Gas Std Test # 2	0.082
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.046
0.08 g/210L Test # 3	0.077
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

M-A LOT 2024-B, ACETONE LOT 2023-B

Data Download Date/Time 07/04/2025 00:02

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

Agency	CLEARWATER PD	**Date**	07/17/2025	**Type**	Agency Inspection
Inspector	HURT,JONATHAN,R	**Time**	18:31:16	**Software**	8100.27

0.05 Lot #	**0.08 Lot #**	**0.20 Lot #**	**0.08 Gas Lot #**
Expiration	**Expiration**	**Expiration**	**Expiration**

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	Interferent Detect Test # 1	0.000
Minimum Sample Volume Check: OK		Interferent Detect Test # (Repeat) 1	0.000
Barometric Pressure Sensor Check: OK		Alcohol Free Test # 2	0.000
Alcohol Free Subject Test: 0.000	Yes	Interferent Detect Test # 2	INT
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # (Repeat) 2	0.000
Interferent Detect Test: Interferent Detect	No	Alcohol Free Test # 3	0.000
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test # 3	INT
Alcohol Free Test: OK	Yes	Interferent Detect Test # (Repeat) 3	0.000
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

INT NOT DET FIRST, .

Data Download Date/Time 07/17/2025 19:50

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

Agency CLEARWATER PD **Date** 07/17/2025 **Type** Agency Inspection
Inspector HURT,JONATHAN,R **Time** 19:09:12 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 31124080A2
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 12/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.051
0.08 g/210L Test # 1	0.075
0.20 g/210L Test # 1	0.202
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.053
0.08 g/210L Test # 2	0.075
0.20 g/210L Test # 2	0.205
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.076
0.20 g/210L Test # 3	0.206
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 07/17/2025 19:50

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

Agency	CLEARWATER PD	**Date**	09/09/2025	**Type**	Agency Inspection
Inspector	HIGGINBOTHAM,LEANDRA,	**Time**	09:54:11	**Software**	8100.27

0.05 Lot #	**0.08 Lot #**	**0.20 Lot #**	**0.08 Gas Lot #**
Expiration	**Expiration**	**Expiration**	**Expiration**

Number of Simulators **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

BYPASS AI FOR OPERATION, COMPLIANCE UNDETERMINED

Data Download Date/Time 09/10/2025 11:59

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

Agency CLEARWATER PD **Date** 09/10/2025 **Type** Department Inspection
Inspector HIGGINBOTHAM,LEANDRA, **Time** 12:01:22 **Software** 8100.27

0.05 Lot # 202406K **0.08 Lot #** 202406L **0.20 Lot #** 202406N **0.08 Gas Lot #** AG429602
Expiration 06/19/2026 **Expiration** 06/19/2026 **Expiration** 06/20/2026 **Expiration** 10/22/2026

Number of Simulators 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	INT
Alcohol Free Test # (Repeat) 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.077
Interferent Detect Test # 1	INT
Alcohol Free Test # (Repeat) 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.077
Interferent Detect Test # 2	INT
Alcohol Free Test # (Repeat) 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.077
Interferent Detect Test # 3	INT
Alcohol Free Test # (Repeat) 4	0.000
0.05 g/210L Test # 4	0.048
0.08 g/210L Test # 4	0.079
0.20 g/210L Test # 4	0.199
0.08 g/210L Dry Gas Std Test # 4	0.077
Interferent Detect Test # 4	INT
Alcohol Free Test # (Repeat) 5	0.000
0.05 g/210L Test # 5	0.048
0.08 g/210L Test # 5	0.079
0.20 g/210L Test # 5	0.198
0.08 g/210L Dry Gas Std Test # 5	0.077

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

Interferent Detect Test # 5	INT
Alcohol Free Test # (Repeat) 6	0.000
0.05 g/210L Test # 6	0.049
0.08 g/210L Test # 6	0.079
0.20 g/210L Test # 6	0.198
0.08 g/210L Dry Gas Std Test # 6	0.077
Interferent Detect Test # 6	INT
Alcohol Free Test # (Repeat) 7	0.000
0.05 g/210L Test # 7	0.049
0.08 g/210L Test # 7	0.079
0.20 g/210L Test # 7	0.199
0.08 g/210L Dry Gas Std Test # 7	0.077
Interferent Detect Test # 7	INT
Alcohol Free Test # (Repeat) 8	0.000
0.05 g/210L Test # 8	0.049
0.08 g/210L Test # 8	0.079
0.20 g/210L Test # 8	0.199
0.08 g/210L Dry Gas Std Test # 8	0.076
Interferent Detect Test # 8	INT
Alcohol Free Test # (Repeat) 9	0.000
0.05 g/210L Test # 9	0.049
0.08 g/210L Test # 9	0.079
0.20 g/210L Test # 9	0.198
0.08 g/210L Dry Gas Std Test # 9	0.077
Interferent Detect Test # 9	INT
Alcohol Free Test # (Repeat) 10	0.000
0.05 g/210L Test # 10	0.049
0.08 g/210L Test # 10	0.079
0.20 g/210L Test # 10	0.198
0.08 g/210L Dry Gas Std Test # 10	0.077
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0005
0.08 g/210L Test (Std Dev)	0.0003
0.20 g/210L Test (Std Dev)	0.0007
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

00: Interferent Detect.

Data Download Date/Time 09/10/2025 11:59

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

Agency	CLEARWATER PD	**Date**	09/26/2025	**Type**	Agency Inspection
Inspector	YEATES,SCOTT,W	**Time**	18:37:11	**Software**	8100.27

0.05 Lot #	202402E	**0.08 Lot #**	202410E	**0.20 Lot #**	202410F	**0.08 Gas Lot #**	31124080A2
Expiration	02/27/2026	**Expiration**	10/22/2026	**Expiration**	10/30/2026	**Expiration**	12/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.196
0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.078
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 09/26/2025 21:53

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

Agency CLEARWATER PD **Date** 10/22/2025 **Type** Agency Inspection
Inspector YEATES,SCOTT,W **Time** 17:00:59 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 31124080A2
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 12/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.037
0.05 g/210L Test # (Repeat) 1	0.046
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.196
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.037
0.05 g/210L Test # (Repeat) 2	0.047
0.08 g/210L Test # 2	0.077
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.047
0.05 g/210L Test # (Repeat) 3	0.048
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.078
Interferent Detect Test # 3	INT

Remarks:

A F / M A: RFI Detect. 05: Control Outside Tolerance.

Data Download Date/Time 10/23/2025 18:39

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

Agency	CLEARWATER PD	**Date**	11/22/2025	**Type**	Agency Inspection
Inspector	YEATES,SCOTT,W	**Time**	22:00:29	**Software**	8100.27

0.05 Lot #	202402E	**0.08 Lot #**	202410E	**0.20 Lot #**	202410F	**0.08 Gas Lot #**	31124080A2
Expiration	02/27/2026	**Expiration**	10/22/2026	**Expiration**	10/30/2026	**Expiration**	12/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 11/22/2025 22:51

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

Agency	CLEARWATER PD	**Date**	12/17/2025	**Type**	Agency Inspection
Inspector	YEATES,SCOTT,W	**Time**	13:55:26	**Software**	8100.27

0.05 Lot #	202506E	**0.08 Lot #**	202410E	**0.20 Lot #**	202410F	**0.08 Gas Lot #**	31124080A2
Expiration	06/24/2027	**Expiration**	10/22/2026	**Expiration**	10/30/2026	**Expiration**	12/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	No	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 12/17/2025 14:57

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

Agency PALM BEACH CSO **Date** 01/17/2025 **Type** Agency Inspection
Inspector KARLECKE,JASON,C **Time** 14:24:51 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 14323080A3
Expiration 03/28/2025 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 09/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.194
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.195
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.195
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/17/2025 14:25

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

Agency PALM BEACH COUNTY SO **Date** 02/14/2025 **Type** Department Inspection
Inspector GUTSCHOW,TAYLOR,D **Time** 12:30:27 **Software** 8100.27

0.05 Lot # 202303K **0.08 Lot #** 202303L **0.20 Lot #** 202304C **0.08 Gas Lot #** AG429602
Expiration 03/29/2025 **Expiration** 03/29/2025 **Expiration** 04/05/2025 **Expiration** 10/22/2026

Number of Simulators 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.051
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.051
0.08 g/210L Test # 4	0.079
0.20 g/210L Test # 4	0.196
0.08 g/210L Dry Gas Std Test # 4	0.080
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.051
0.08 g/210L Test # 5	0.079
0.20 g/210L Test # 5	0.197
0.08 g/210L Dry Gas Std Test # 5	0.079
Interferent Detect Test # 5	INT

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.052
0.08 g/210L Test # 6	0.079
0.20 g/210L Test # 6	0.197
0.08 g/210L Dry Gas Std Test # 6	0.080
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.051
0.08 g/210L Test # 7	0.079
0.20 g/210L Test # 7	0.197
0.08 g/210L Dry Gas Std Test # 7	0.080
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.051
0.08 g/210L Test # 8	0.080
0.20 g/210L Test # 8	0.197
0.08 g/210L Dry Gas Std Test # 8	0.079
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.051
0.08 g/210L Test # 9	0.079
0.20 g/210L Test # 9	0.198
0.08 g/210L Dry Gas Std Test # 9	0.079
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.051
0.08 g/210L Test # 10	0.079
0.20 g/210L Test # 10	0.197
0.08 g/210L Dry Gas Std Test # 10	0.080
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0007
0.08 g/210L Test (Std Dev)	0.0003
0.20 g/210L Test (Std Dev)	0.0004
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

Data Download Date/Time 02/14/2025 12:30

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

Agency PALM BEACH COUNTY SO **Date** 03/14/2025 **Type** Agency Inspection
Inspector KARLECKE,JASON,C **Time** 12:32:22 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 10124080A2
Expiration 03/28/2025 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 06/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 03/14/2025 12:33

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

Agency PALM BEACH COUNTY SO **Date** 04/25/2025 **Type** Agency Inspection
Inspector KARLECKE,JASON,C **Time** 13:09:26 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 10124080A2
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 06/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 04/25/2025 13:09

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

Agency PALM BEACH COUNTY SO **Date** 05/09/2025 **Type** Agency Inspection
Inspector KARLECKE,JASON,C **Time** 14:02:00 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 32324080A2
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 12/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.200
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 05/09/2025 14:03

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

Agency PALM BEACH COUNTY SO **Date** 06/20/2025 **Type** Agency Inspection
Inspector KARLECKE,JASON,C **Time** 16:31:31 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 32324080A2
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 12/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.081
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.081
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.082
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 06/20/2025 16:32

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

Agency PALM BEACH COUNTY SO **Date** 07/11/2025 **Type** Agency Inspection
Inspector KARLECKE,JASON,C **Time** 16:11:58 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 14323080A3
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 09/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.081
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 07/11/2025 16:11

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

Agency PALM BEACH COUNTY SO **Date** 08/15/2025 **Type** Agency Inspection
Inspector KARLECKE,JASON,C **Time** 15:46:01 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 14323080A3
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 09/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 08/15/2025 15:45

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

Agency PALM BEACH COUNTY SO **Date** 09/12/2025 **Type** Agency Inspection
Inspector KARLECKE,JASON,C **Time** 13:19:48 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 10124080A2
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 06/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.051
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.051
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.051
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.200
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 09/12/2025 13:20

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

Agency PALM BEACH COUNTY SO **Date** 10/17/2025 **Type** Agency Inspection
Inspector KARLECKE,JASON,C **Time** 15:17:14 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 10124080A2
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 06/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.082
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 10/17/2025 15:19

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

Agency PALM BEACH COUNTY SO **Date** 11/14/2025 **Type** Agency Inspection
Inspector KARLECKE,JASON,C **Time** 13:54:20 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 19825080A2
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 09/05/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.081
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 11/14/2025 14:09

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

Agency PALM BEACH COUNTY SO **Date** 12/12/2025 **Type** Agency Inspection
Inspector KARLECKE,JASON,C **Time** 14:00:59 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 19825080A2
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 09/05/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 12/12/2025 14:01

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

Agency MANATEE COUNTY SO **Date** 01/17/2025 **Type** Agency Inspection
Inspector COLEMAN,WILLIAM, **Time** 14:20:15 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 403141809
Expiration 02/27/2026 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 09/10/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.195
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.195
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.077
0.20 g/210L Test # 3	0.195
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/17/2025 14:22

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

Agency MANATEE COUNTY SO **Date** 02/07/2025 **Type** Agency Inspection
Inspector NOVAK,CHRISTOPHER, **Time** 12:44:01 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 403141809
Expiration 02/27/2026 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 09/10/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.195
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.195
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.195
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/07/2025 12:45

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

Agency MANATEE COUNTY SO **Date** 03/07/2025 **Type** Agency Inspection
Inspector COLEMAN,WILLIAM, **Time** 09:32:14 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 403141809
Expiration 03/28/2025 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 09/10/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.196
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 03/07/2025 09:33

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

Agency MANATEE COUNTY SO **Date** 03/25/2025 **Type** Department Inspection
Inspector GUTSCHOW,TAYLOR,D **Time** 15:54:00 **Software** 8100.27

0.05 Lot # 202303K **0.08 Lot #** 202303L **0.20 Lot #** 202304C **0.08 Gas Lot #** AG429602
Expiration 03/29/2025 **Expiration** 03/29/2025 **Expiration** 04/05/2025 **Expiration** 10/22/2026

Number of Simulators 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.194
0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.193
0.08 g/210L Dry Gas Std Test # 2	0.077
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.194
0.08 g/210L Dry Gas Std Test # 3	0.077
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.049
0.08 g/210L Test # 4	0.078
0.20 g/210L Test # 4	0.194
0.08 g/210L Dry Gas Std Test # 4	0.078
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.049
0.08 g/210L Test # 5	0.079
0.20 g/210L Test # 5	0.194
0.08 g/210L Dry Gas Std Test # 5	0.077
Interferent Detect Test # 5	INT

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.049
0.08 g/210L Test # 6	0.078
0.20 g/210L Test # 6	0.194
0.08 g/210L Dry Gas Std Test # 6	0.078
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.049
0.08 g/210L Test # 7	0.079
0.20 g/210L Test # 7	0.194
0.08 g/210L Dry Gas Std Test # 7	0.078
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.049
0.08 g/210L Test # 8	0.079
0.20 g/210L Test # 8	0.194
0.08 g/210L Dry Gas Std Test # 8	0.078
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.049
0.08 g/210L Test # 9	0.079
0.20 g/210L Test # 9	0.194
0.08 g/210L Dry Gas Std Test # 9	0.078
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.049
0.08 g/210L Test # 10	0.079
0.20 g/210L Test # 10	0.194
0.08 g/210L Dry Gas Std Test # 10	0.078
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0000
0.08 g/210L Test (Std Dev)	0.0005
0.20 g/210L Test (Std Dev)	0.0003
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Data Download Date/Time 03/25/2025 15:56

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

Agency MANATEE COUNTY SO **Date** 05/16/2025 **Type** Agency Inspection
Inspector NOVAK,CHRISTOPHER, **Time** 10:03:01 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 402987895
Expiration 02/27/2026 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 03/05/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 05/16/2025 10:03

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

Agency MANATEE COUNTY SO **Date** 06/20/2025 **Type** Agency Inspection
Inspector COLEMAN,WILLIAM, **Time** 10:35:11 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202307C **0.08 Gas Lot #** 402987895
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 07/12/2025 **Expiration** 03/05/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.196
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.196
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.196
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 06/20/2025 10:36

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

Agency MANATEE COUNTY SO **Date** 07/15/2025 **Type** Agency Inspection
Inspector COLEMAN,WILLIAM, **Time** 12:53:07 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 403141809
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 09/10/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	Yes
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

Time-Date changed.

Data Download Date/Time 07/15/2025 12:51

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

Agency MANATEE COUNTY SO **Date** 08/14/2025 **Type** Agency Inspection
Inspector COLEMAN,WILLIAM, **Time** 09:32:48 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 403141809
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 09/10/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.200
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 08/14/2025 09:29

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

Agency MANATEE COUNTY SO **Date** 09/15/2025 **Type** Agency Inspection
Inspector COLEMAN,WILLIAM, **Time** 09:19:08 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 403411042
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 08/02/2028

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.051
0.08 g/210L Test # 3	0.081
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 09/15/2025 09:59

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

Agency MANATEE COUNTY SO **Date** 10/17/2025 **Type** Agency Inspection
Inspector COLEMAN,WILLIAM, **Time** 13:57:20 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 403411042
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 08/02/2028

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.046
0.08 g/210L Test # 1	0.075
0.20 g/210L Test # 1	0.195
0.08 g/210L Dry Gas Std Test # 1	0.077
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.046
0.08 g/210L Test # 2	0.076
0.20 g/210L Test # 2	0.194
0.08 g/210L Dry Gas Std Test # 2	0.077
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.046
0.08 g/210L Test # 3	0.076
0.20 g/210L Test # 3	0.194
0.08 g/210L Dry Gas Std Test # 3	0.078
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 10/17/2025 13:55

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

Agency MANATEE COUNTY SO **Date** 11/10/2025 **Type** Agency Inspection
Inspector COLEMAN,WILLIAM, **Time** 09:29:19 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 403411042
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 08/02/2028

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 11/10/2025 09:27

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

Agency MANATEE COUNTY SO **Date** 12/22/2025 **Type** Agency Inspection
Inspector COLEMAN,WILLIAM, **Time** 14:28:29 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 403419188
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 08/13/2028

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.082
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.082
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.082
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 12/22/2025 14:31

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

Agency JACKSONVILLE SO **Date** 01/09/2025 **Type** Agency Inspection
Inspector LIGHTSEY,ROBERT,D **Time** 09:39:47 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 15924080A2
Expiration 03/28/2025 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 07/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.083
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.083
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.083
		Interferent Detect Test # 3	INT

Remarks:

IN SERV @ PTDF

Data Download Date/Time 01/09/2025 09:45

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

Agency JACKSONVILLE SO **Date** 02/24/2025 **Type** Agency Inspection
Inspector LIGHTSEY,ROBERT,D **Time** 13:54:29 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 28424080A3
Expiration 03/28/2025 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 11/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.082
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.082
Interferent Detect Test # 3	INT

Remarks:

EVED USE AT PTDF

Data Download Date/Time 02/24/2025 13:57

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

Agency JACKSONVILLE SO **Date** 03/13/2025 **Type** Agency Inspection
Inspector POTTER,MICHAEL,R **Time** 14:47:55 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 28424080A3
Expiration 03/28/2025 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 11/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

IN SERV @ PTDF

Data Download Date/Time 03/13/2025 14:59

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

Agency JACKSONVILLE SO **Date** 04/22/2025 **Type** Agency Inspection
Inspector LIGHTSEY,ROBERT,D **Time** 11:16:44 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 28424080A3
Expiration 02/27/2026 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 11/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.196
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.196
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

EVED USE AT PTDF

Data Download Date/Time 04/22/2025 11:39

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

Agency JACKSONVILLE SO **Date** 05/13/2025 **Type** Agency Inspection
Inspector LIGHTSEY,ROBERT,D **Time** 12:27:17 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 28424080A3
Expiration 02/27/2026 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 11/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.076
0.20 g/210L Test # 1	0.195
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.077
0.20 g/210L Test # 2	0.195
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.077
0.20 g/210L Test # 3	0.196
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

IN USE @ PTDF 05/13/2025

Data Download Date/Time 05/13/2025 12:33

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

Agency JACKSONVILLE SO **Date** 06/11/2025 **Type** Agency Inspection
Inspector LIGHTSEY,ROBERT,D **Time** 13:30:53 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 28424080A3
Expiration 02/27/2026 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 11/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.200
0.08 g/210L Dry Gas Std Test # 1	0.082
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.082
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.082
Interferent Detect Test # 3	INT

Remarks:

IN SERV @ PTDF

Data Download Date/Time 06/11/2025 13:37

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

Agency JACKSONVILLE SO **Date** 07/10/2025 **Type** Agency Inspection
Inspector LIGHTSEY,ROBERT,D **Time** 13:11:47 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 28424080A3
Expiration 02/27/2026 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 11/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.076
0.20 g/210L Test # 1	0.194
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.077
0.20 g/210L Test # 2	0.195
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.077
0.20 g/210L Test # 3	0.195
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

IN SERV @ PTDF

Data Download Date/Time 07/10/2025 13:20

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

Agency JACKSONVILLE SO **Date** 07/24/2025 **Type** Agency Inspection
Inspector LIGHTSEY,ROBERT,D **Time** 11:07:07 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 28424080A3
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 11/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.196
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.077
0.20 g/210L Test # 2	0.195
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.077
0.20 g/210L Test # 3	0.194
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 07/24/2025 11:15

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

Agency JACKSONVILLE SO **Date** 08/20/2025 **Type** Agency Inspection
Inspector LIGHTSEY,ROBERT,D **Time** 15:08:19 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 28424080A3
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 11/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.196
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.200
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

REMOVED FROM SERV FOR FDLE DEPT INSP

Data Download Date/Time 08/20/2025 15:13

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

Agency JACKSONVILLE SO **Date** 09/16/2025 **Type** Department Inspection
Inspector HIGGINBOTHAM,LEANDRA, **Time** 17:15:34 **Software** 8100.27

0.05 Lot # 202406K **0.08 Lot #** 202406L **0.20 Lot #** 202406N **0.08 Gas Lot #** AG429602
Expiration 06/19/2026 **Expiration** 06/19/2026 **Expiration** 06/20/2026 **Expiration** 10/22/2026

Number of Simulators 5 **Average Standard Deviation** 0.0002 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.200
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.200
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.200
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.049
0.08 g/210L Test # 4	0.080
0.20 g/210L Test # 4	0.200
0.08 g/210L Dry Gas Std Test # 4	0.080
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.049
0.08 g/210L Test # 5	0.080
0.20 g/210L Test # 5	0.200
0.08 g/210L Dry Gas Std Test # 5	0.080
Interferent Detect Test # 5	INT

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.049
0.08 g/210L Test # 6	0.080
0.20 g/210L Test # 6	0.200
0.08 g/210L Dry Gas Std Test # 6	0.080
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.049
0.08 g/210L Test # 7	0.081
0.20 g/210L Test # 7	0.201
0.08 g/210L Dry Gas Std Test # 7	0.080
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.049
0.08 g/210L Test # 8	0.081
0.20 g/210L Test # 8	0.200
0.08 g/210L Dry Gas Std Test # 8	0.081
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.049
0.08 g/210L Test # 9	0.080
0.20 g/210L Test # 9	0.200
0.08 g/210L Dry Gas Std Test # 9	0.080
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.049
0.08 g/210L Test # 10	0.081
0.20 g/210L Test # 10	0.200
0.08 g/210L Dry Gas Std Test # 10	0.080
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0000
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0003
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Data Download Date/Time 09/16/2025 17:22

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

Agency JACKSONVILLE SO **Date** 10/29/2025 **Type** Agency Inspection
Inspector LIGHTSEY,ROBERT,D **Time** 11:38:51 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 28424080A3
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 11/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.076
0.20 g/210L Test # 1	0.193
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.077
0.20 g/210L Test # 2	0.194
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.076
0.20 g/210L Test # 3	0.194
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

IN SERV @ PTDF

Data Download Date/Time 10/29/2025 11:41

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

Agency JACKSONVILLE SO **Date** 11/19/2025 **Type** Agency Inspection
Inspector LIGHTSEY,ROBERT,D **Time** 13:05:59 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 13225080A2
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 06/05/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.195
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.195
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

IN SERV. 11.19.25

Data Download Date/Time 11/19/2025 13:11

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

Agency	JACKSONVILLE SO	**Date**	12/17/2025	**Type**	Agency Inspection
Inspector	LIGHTSEY,ROBERT,D	**Time**	12:44:45	**Software**	8100.27

0.05 Lot #	202402E	**0.08 Lot #**	202410E	**0.20 Lot #**	202410F	**0.08 Gas Lot #**	23725080A1
Expiration	02/27/2026	**Expiration**	10/22/2026	**Expiration**	10/30/2026	**Expiration**	10/05/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.205
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.203
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

EVID USE AT PTDF

Data Download Date/Time 12/17/2025 12:55

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001866

Agency	LAKE HAMILTON PD	**Date**	12/30/2025	**Type**	Agency Inspection
Inspector	PIERSON,WEN-CHI,K	**Time**	14:59:15	**Software**	8100.27

0.05 Lot #	**0.08 Lot #**	**0.20 Lot #**	**0.08 Gas Lot #**
Expiration	**Expiration**	**Expiration**	**Expiration**

Number of Simulators **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

BYPASS AI TO OPERATE INSTRUMENT. COMPLIANCE NOT DETERMINED.

Data Download Date/Time 12/31/2025 12:24

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001866

Agency LAKE HAMILTON PD **Date** 12/31/2025 **Type** Department Inspection
Inspector PIERSON,WEN-CHI,K **Time** 12:20:59 **Software** 8100.27

0.05 Lot # 202406K **0.08 Lot #** 202406L **0.20 Lot #** 202406N **0.08 Gas Lot #** AG510701
Expiration 06/19/2026 **Expiration** 06/19/2026 **Expiration** 06/20/2026 **Expiration** 04/17/2027

Number of Simulators 5 **Average Standard Deviation** 0.0006 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.051
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.201
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.052
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.201
0.08 g/210L Dry Gas Std Test # 2	0.082
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.051
0.08 g/210L Test # 3	0.081
0.20 g/210L Test # 3	0.202
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.051
0.08 g/210L Test # 4	0.081
0.20 g/210L Test # 4	0.201
0.08 g/210L Dry Gas Std Test # 4	0.082
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.052
0.08 g/210L Test # 5	0.081
0.20 g/210L Test # 5	0.201
0.08 g/210L Dry Gas Std Test # 5	0.081
Interferent Detect Test # 5	INT

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001866

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.051
0.08 g/210L Test # 6	0.081
0.20 g/210L Test # 6	0.202
0.08 g/210L Dry Gas Std Test # 6	0.081
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.052
0.08 g/210L Test # 7	0.082
0.20 g/210L Test # 7	0.202
0.08 g/210L Dry Gas Std Test # 7	0.081
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.052
0.08 g/210L Test # 8	0.081
0.20 g/210L Test # 8	0.201
0.08 g/210L Dry Gas Std Test # 8	0.082
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.052
0.08 g/210L Test # 9	0.081
0.20 g/210L Test # 9	0.201
0.08 g/210L Dry Gas Std Test # 9	0.082
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.054
0.08 g/210L Test # 10	0.082
0.20 g/210L Test # 10	0.202
0.08 g/210L Dry Gas Std Test # 10	0.081
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0009
0.08 g/210L Test (Std Dev)	0.0006
0.20 g/210L Test (Std Dev)	0.0005
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

Data Download Date/Time 12/31/2025 12:24

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001962

Agency PORT ST LUCIE PD **Date** 02/18/2025 **Type** Agency Inspection
Inspector APPELBAUM,KEITH,A **Time** 11:19:55 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202410E **0.20 Lot #** 202307C **0.08 Gas Lot #** 403008479
Expiration 03/28/2025 **Expiration** 10/22/2026 **Expiration** 07/12/2025 **Expiration** 03/28/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.051
0.08 g/210L Test # 1	0.081
0.20 g/210L Test # 1	0.196
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.082
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.051
0.08 g/210L Test # 3	0.082
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/18/2025 11:28

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001962

Agency PORT ST LUCIE PD **Date** 03/28/2025 **Type** Agency Inspection
Inspector THOMPSON,BRIAN,J **Time** 13:19:03 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202401E **0.20 Lot #** 202410F **0.08 Gas Lot #** 403008479
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 03/28/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.046
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.082
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.047
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.082
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.047
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.082
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 03/28/2025 13:24

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001962

Agency PORT ST LUCIE PD **Date** 04/27/2025 **Type** Agency Inspection
Inspector APPELBAUM,KEITH,A **Time** 11:00:52 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 403008479
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 03/28/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.081
0.20 g/210L Test # 1	0.200
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.200
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.081
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 04/27/2025 11:30

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001962

Agency PORT ST LUCIE PD **Date** 05/30/2025 **Type** Agency Inspection
Inspector THOMPSON,BRIAN,J **Time** 12:19:28 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 403008479
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 03/28/2027

Number of Simulators 4 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.082
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.077
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 05/30/2025 12:19

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001962

Agency	PORT ST LUCIE PD	**Date**	07/11/2025	**Type**	Agency Inspection
Inspector	GUTSCHOW,TAYLOR,D	**Time**	11:40:53	**Software**	8100.27

0.05 Lot #	**0.08 Lot #**	**0.20 Lot #**	**0.08 Gas Lot #**
Expiration	**Expiration**	**Expiration**	**Expiration**

Number of Simulators **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

AI BYPASSED TO OPERATE INSTRUMENT. COMPLIANCE UNDETERMINED.

Data Download Date/Time 07/11/2025 15:28

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001962

Agency PORT ST LUCIE PD **Date** 07/11/2025 **Type** Department Inspection
Inspector GUTSCHOW,TAYLOR,D **Time** 15:17:44 **Software** 8100.27

0.05 Lot # 202406K **0.08 Lot #** 202406L **0.20 Lot #** 202406N **0.08 Gas Lot #** AG429602
Expiration 06/19/2026 **Expiration** 06/19/2026 **Expiration** 06/20/2026 **Expiration** 10/22/2026

Number of Simulators 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
Alcohol Free Test # (Repeat) 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
Alcohol Free Test # (Repeat) 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	INT
Alcohol Free Test # (Repeat) 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT
Alcohol Free Test # (Repeat) 4	0.000
0.05 g/210L Test # 4	0.050
0.08 g/210L Test # 4	0.080
0.20 g/210L Test # 4	0.198
0.08 g/210L Dry Gas Std Test # 4	0.081
Interferent Detect Test # 4	INT
Alcohol Free Test # (Repeat) 5	0.000
0.05 g/210L Test # 5	0.050
0.08 g/210L Test # 5	0.080

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001962

0.20 g/210L Test # 5	0.198
0.08 g/210L Dry Gas Std Test # 5	0.081
Interferent Detect Test # 5	INT
Alcohol Free Test # (Repeat) 6	0.000
0.05 g/210L Test # 6	0.051
0.08 g/210L Test # 6	0.080
0.20 g/210L Test # 6	0.199
0.08 g/210L Dry Gas Std Test # 6	0.081
Interferent Detect Test # 6	INT
Alcohol Free Test # (Repeat) 7	0.000
0.05 g/210L Test # 7	0.051
0.08 g/210L Test # 7	0.080
0.20 g/210L Test # 7	0.199
0.08 g/210L Dry Gas Std Test # 7	0.082
Interferent Detect Test # 7	INT
Alcohol Free Test # (Repeat) 8	0.000
0.05 g/210L Test # 8	0.051
0.08 g/210L Test # 8	0.081
0.20 g/210L Test # 8	0.199
0.08 g/210L Dry Gas Std Test # 8	0.081
Interferent Detect Test # 8	INT
Alcohol Free Test # (Repeat) 9	0.000
0.05 g/210L Test # 9	0.051
0.08 g/210L Test # 9	0.080
0.20 g/210L Test # 9	0.199
0.08 g/210L Dry Gas Std Test # 9	0.081
Interferent Detect Test # 9	INT
Alcohol Free Test # (Repeat) 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.080
0.20 g/210L Test # 10	0.199
0.08 g/210L Dry Gas Std Test # 10	0.081
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0005
0.08 g/210L Test (Std Dev)	0.0003
0.20 g/210L Test (Std Dev)	0.0004
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

00: Interferent Detect. AIRED OUT ROOM AND REPEATED.

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001962

Data Download Date/Time 07/11/2025 15:29

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001962

Agency PORT ST LUCIE PD **Date** 07/23/2025 **Type** Agency Inspection
Inspector APPELBAUM,KEITH,A **Time** 16:16:54 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 403008479
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 03/28/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.051
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.051
0.08 g/210L Test # 2	0.081
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.051
0.08 g/210L Test # 3	0.081
0.20 g/210L Test # 3	0.201
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 07/23/2025 16:15

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001962

Agency PORT ST LUCIE PD **Date** 08/22/2025 **Type** Agency Inspection
Inspector THOMPSON,BRIAN,J **Time** 10:26:01 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 403008479
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 03/28/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.081
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.051
0.08 g/210L Test # 2	0.081
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.051
0.08 g/210L Test # 3	0.082
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

N/A

Data Download Date/Time 08/22/2025 10:33

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001962

Agency	PORT ST LUCIE PD	**Date**	09/29/2025	**Type**	Agency Inspection
Inspector	THOMPSON,BRIAN,J	**Time**	12:27:52	**Software**	8100.27

0.05 Lot #	**0.08 Lot #**	**0.20 Lot #**	**0.08 Gas Lot #**
Expiration	**Expiration**	**Expiration**	**Expiration**

Number of Simulators 5 **Compliance** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A: . SWITCH PERSON FOR INTER.

Data Download Date/Time 09/29/2025 13:14

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001962

Agency PORT ST LUCIE PD **Date** 09/29/2025 **Type** Agency Inspection
Inspector THOMPSON,BRIAN,J **Time** 13:05:31 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 403008479
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 03/28/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.051
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.081
0.20 g/210L Test # 3	0.200
0.08 g/210L Dry Gas Std Test # 3	0.078
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 09/29/2025 13:14

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001962

Agency PORT ST LUCIE PD **Date** 10/28/2025 **Type** Agency Inspection
Inspector APPELBAUM,KEITH,A **Time** 15:46:26 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 403008479
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 02/28/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test # 1	0.000
Interferent Detect Test # (Repeat) 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.081
0.20 g/210L Test # 2	0.200
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	0.000
Interferent Detect Test # (Repeat) 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.081
0.20 g/210L Test # 3	0.200
0.08 g/210L Dry Gas Std Test # 3	0.078
Interferent Detect Test # 3	0.000
Interferent Detect Test # (Repeat) 3	INT

Remarks:

Int Det:ACE ADDED.

Data Download Date/Time 10/28/2025 15:45

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001962

Agency PORT ST LUCIE PD **Date** 11/24/2025 **Type** Agency Inspection
Inspector APPELBAUM,KEITH,A **Time** 10:37:17 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 403008479
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 02/28/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 11/24/2025 10:40

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002097

Agency NAS JACKSONVILLE PD **Date** 01/15/2025 **Type** Agency Inspection
Inspector POTTER,MICHAEL,R **Time** 15:07:41 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 01923080A4
Expiration 03/28/2025 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 02/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.206
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.206
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.204
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

IN SERV. NAS JAX

Data Download Date/Time 01/15/2025 15:08

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002097

Agency NAS JACKSONVILLE PD **Date** 02/25/2025 **Type** Agency Inspection
Inspector POTTER,MICHAEL,R **Time** 14:36:50 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 28424080A3
Expiration 03/28/2025 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 11/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.203
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.203
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

REMOVED FROM SERV FDLE D/I

Data Download Date/Time 02/25/2025 14:37

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002097

Agency NAS JACKSONVILLE PD **Date** 10/28/2025 **Type** Department Inspection
Inspector PLATT,SHAYLA,D **Time** 15:06:15 **Software** 8100.27

0.05 Lot # 202406K **0.08 Lot #** 202406L **0.20 Lot #** 202406N **0.08 Gas Lot #** AG510701
Expiration 06/19/2026 **Expiration** 06/19/2026 **Expiration** 06/20/2026 **Expiration** 04/17/2027

Number of Simulators 5 **Average Standard Deviation** 0.0001 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.047
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.047
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.047
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.078
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.047
0.08 g/210L Test # 4	0.078
0.20 g/210L Test # 4	0.199
0.08 g/210L Dry Gas Std Test # 4	0.078
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.047
0.08 g/210L Test # 5	0.078
0.20 g/210L Test # 5	0.199
0.08 g/210L Dry Gas Std Test # 5	0.078
Interferent Detect Test # 5	INT

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002097

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.047
0.08 g/210L Test # 6	0.078
0.20 g/210L Test # 6	0.199
0.08 g/210L Dry Gas Std Test # 6	0.078
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.048
0.08 g/210L Test # 7	0.078
0.20 g/210L Test # 7	0.199
0.08 g/210L Dry Gas Std Test # 7	0.078
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.048
0.08 g/210L Test # 8	0.078
0.20 g/210L Test # 8	0.199
0.08 g/210L Dry Gas Std Test # 8	0.078
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.048
0.08 g/210L Test # 9	0.078
0.20 g/210L Test # 9	0.199
0.08 g/210L Dry Gas Std Test # 9	0.078
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.048
0.08 g/210L Test # 10	0.078
0.20 g/210L Test # 10	0.199
0.08 g/210L Dry Gas Std Test # 10	0.078
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0005
0.08 g/210L Test (Std Dev)	0.0000
0.20 g/210L Test (Std Dev)	0.0000
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0000

Remarks:

Data Download Date/Time 10/28/2025 15:06

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002097

Agency NAS JACKSONVILLE PD **Date** 11/20/2025 **Type** Agency Inspection
Inspector LIGHTSEY,ROBERT,D **Time** 18:11:30 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 13225080A2
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 06/05/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.045
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.045
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

EVID USE AT NAS JAX

Data Download Date/Time 11/20/2025 18:10

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002097

Agency NAS JACKSONVILLE PD **Date** 12/19/2025 **Type** Agency Inspection
Inspector POTTER,MICHAEL,R **Time** 14:00:31 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 13225080A2
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 06/05/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.047
0.08 g/210L Test # 1	0.076
0.20 g/210L Test # 1	0.196
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.047
0.08 g/210L Test # 2	0.076
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.047
0.08 g/210L Test # 3	0.077
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

IN SERV @ NAS JAX PD

Data Download Date/Time 12/19/2025 14:00

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002170

Agency NAS KEY WEST **Date** 02/14/2025 **Type** Department Inspection
Inspector GUTSCHOW,TAYLOR,D **Time** 13:31:17 **Software** 8100.27

0.05 Lot # 202303K **0.08 Lot #** 202303L **0.20 Lot #** 202304C **0.08 Gas Lot #** AG429602
Expiration 03/29/2025 **Expiration** 03/29/2025 **Expiration** 04/05/2025 **Expiration** 10/22/2026

Number of Simulators 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.050
0.08 g/210L Test # 4	0.079
0.20 g/210L Test # 4	0.197
0.08 g/210L Dry Gas Std Test # 4	0.080
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.050
0.08 g/210L Test # 5	0.079
0.20 g/210L Test # 5	0.197
0.08 g/210L Dry Gas Std Test # 5	0.079
Interferent Detect Test # 5	INT

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002170

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.049
0.08 g/210L Test # 6	0.079
0.20 g/210L Test # 6	0.197
0.08 g/210L Dry Gas Std Test # 6	0.080
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.080
0.20 g/210L Test # 7	0.196
0.08 g/210L Dry Gas Std Test # 7	0.079
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.050
0.08 g/210L Test # 8	0.080
0.20 g/210L Test # 8	0.196
0.08 g/210L Dry Gas Std Test # 8	0.079
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.050
0.08 g/210L Test # 9	0.080
0.20 g/210L Test # 9	0.197
0.08 g/210L Dry Gas Std Test # 9	0.080
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.079
0.20 g/210L Test # 10	0.196
0.08 g/210L Dry Gas Std Test # 10	0.080
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0005
0.20 g/210L Test (Std Dev)	0.0004
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Data Download Date/Time 02/14/2025 13:29

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002170

Agency NAS KEY WEST **Date** 09/03/2025 **Type** Agency Inspection
Inspector ANDERSON,LEE,B **Time** 14:46:46 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 403411042
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 08/02/2028

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

INSTRUMENT CLEANED AND CERTIFIED

Data Download Date/Time 09/03/2025 14:51

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002170

Agency NAS KEY WEST **Date** 11/06/2025 **Type** Agency Inspection
Inspector ANDERSON,LEE,B **Time** 13:48:15 **Software** 8100.27

0.05 Lot # 202506E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 403411042
Expiration 06/24/2027 **Expiration** 10/22/2028 **Expiration** 10/30/2026 **Expiration** 08/02/2028

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.076
0.20 g/210L Test # 1	0.192
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.077
0.20 g/210L Test # 2	0.194
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.077
0.20 g/210L Test # 3	0.195
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

INSTRUMENT CLEANED AND INSPECTED

Data Download Date/Time 11/06/2025 13:52

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002231

Agency NEW SMYRNA BEACH PD **Date** 01/31/2025 **Type** Agency Inspection
Inspector HUNNEFELD,RALPH,A **Time** 13:35:46 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202307C **0.08 Gas Lot #** 24923080A4
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 07/12/2025 **Expiration** 10/05/2025

Number of Simulators 4 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.047
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.196
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.196
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.196
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/31/2025 14:10

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002231

Agency NEW SMYRNA BEACH PD **Date** 02/27/2025 **Type** Agency Inspection
Inspector HELMS,WILLIAM,A **Time** 17:09:32 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202307C **0.08 Gas Lot #** 24923080A4
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 07/12/2025 **Expiration** 10/05/2025

Number of Simulators 4 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.195
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.195
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.195
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/28/2025 08:24

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002231

Agency NEW SMYRNA BEACH PD **Date** 03/06/2025 **Type** Agency Inspection
Inspector HUNNEFELD,RALPH,A **Time** 09:49:09 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202307C **0.08 Gas Lot #** 24923080A4
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 07/12/2025 **Expiration** 10/05/2025

Number of Simulators 4 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.081
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.081
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.081
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 03/06/2025 13:35

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002231

Agency NEW SMYRNA BEACH PD **Date** 04/28/2025 **Type** Agency Inspection
Inspector HELMS,WILLIAM,A **Time** 16:29:18 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202307C **0.08 Gas Lot #** 24923080A4
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 07/12/2025 **Expiration** 10/05/2025

Number of Simulators 4 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.047
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.193
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.195
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.196
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 04/28/2025 16:32

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002231

Agency NEW SMYRNA BEACH PD **Date** 05/26/2025 **Type** Agency Inspection
Inspector HUNNEFELD,RALPH,A **Time** 08:01:16 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202307C **0.08 Gas Lot #** 24923080A4
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 07/12/2025 **Expiration** 10/05/2025

Number of Simulators 4 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.046
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.191
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.047
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.193
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.047
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.194
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 05/27/2025 08:32

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002231

Agency NEW SMYRNA BEACH PD **Date** 06/23/2025 **Type** Agency Inspection
Inspector HELMS,WILLIAM,A **Time** 16:43:58 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202307C **0.08 Gas Lot #** 24923080A4
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 07/12/2025 **Expiration** 10/05/2025

Number of Simulators 4 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.046
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.189
0.20 g/210L Test # (Repeat) 1	0.193
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.047
0.08 g/210L Test # 2	0.077
0.20 g/210L Test # 2	0.191
0.20 g/210L Test # (Repeat) 2	0.194
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.193
0.20 g/210L Test # (Repeat) 3	0.194
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

20: Control Outside Tolerance.

Data Download Date/Time 06/23/2025 16:48

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002231

Agency NEW SMYRNA BEACH PD **Date** 07/04/2025 **Type** Agency Inspection
Inspector HUNNEFELD,RALPH,A **Time** 11:09:34 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202307C **0.08 Gas Lot #** 24923080A4
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 07/12/2025 **Expiration** 10/05/2025

Number of Simulators 4 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # (Repeat) 1	0.000
0.05 g/210L Test # 1	0.045
0.08 g/210L Test # 1	0.076
0.20 g/210L Test # 1	0.188
0.20 g/210L Test # (Repeat) 1	0.193
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # (Repeat) 2	0.000
0.05 g/210L Test # 2	0.046
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.192
0.20 g/210L Test # (Repeat) 2	0.193
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # (Repeat) 3	0.000
0.05 g/210L Test # 3	0.047
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.193
0.20 g/210L Test # (Repeat) 3	0.194
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

A F / M A: RESIDUAL MOUTH ALC. 00: Ambient Fail. 20: Control Outside Tolerance.

Data Download Date/Time 07/04/2025 11:07

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002231

Agency NEW SMYRNA BEACH PD **Date** 08/14/2025 **Type** Agency Inspection
Inspector HELMS,WILLIAM,A **Time** 16:58:51 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 24923080A4
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 10/05/2025

Number of Simulators 4 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.047
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.191
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.193
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.195
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 08/14/2025 17:00

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002231

Agency NEW SMYRNA BEACH PD **Date** 08/26/2025 **Type** Department Inspection
Inspector HIGGINBOTHAM,LEANDRA, **Time** 13:40:24 **Software** 8100.27

0.05 Lot # 202406K **0.08 Lot #** 202406L **0.20 Lot #** 202406N **0.08 Gas Lot #** AG429602
Expiration 06/19/2026 **Expiration** 06/19/2026 **Expiration** 06/20/2026 **Expiration** 10/22/2026

Number of Simulators 5 **Average Standard Deviation** 0.0002 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.047
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.048
0.08 g/210L Test # 4	0.078
0.20 g/210L Test # 4	0.197
0.08 g/210L Dry Gas Std Test # 4	0.080
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.048
0.08 g/210L Test # 5	0.078
0.20 g/210L Test # 5	0.197
0.08 g/210L Dry Gas Std Test # 5	0.079
Interferent Detect Test # 5	INT

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002231

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.048
0.08 g/210L Test # 6	0.078
0.20 g/210L Test # 6	0.197
0.08 g/210L Dry Gas Std Test # 6	0.080
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.048
0.08 g/210L Test # 7	0.078
0.20 g/210L Test # 7	0.197
0.08 g/210L Dry Gas Std Test # 7	0.079
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.048
0.08 g/210L Test # 8	0.077
0.20 g/210L Test # 8	0.197
0.08 g/210L Dry Gas Std Test # 8	0.080
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.048
0.08 g/210L Test # 9	0.078
0.20 g/210L Test # 9	0.197
0.08 g/210L Dry Gas Std Test # 9	0.080
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.048
0.08 g/210L Test # 10	0.078
0.20 g/210L Test # 10	0.197
0.08 g/210L Dry Gas Std Test # 10	0.079
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0003
0.08 g/210L Test (Std Dev)	0.0003
0.20 g/210L Test (Std Dev)	0.0000
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

Data Download Date/Time 08/26/2025 13:34

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002231

Agency NEW SMYRNA BEACH PD **Date** 09/23/2025 **Type** Agency Inspection
Inspector HUNNEFELD,RALPH,A **Time** 08:41:44 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 24923080A4
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 10/05/2025

Number of Simulators 4 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.195
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.196
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 09/24/2025 06:58

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002231

Agency NEW SMYRNA BEACH PD **Date** 10/27/2025 **Type** Agency Inspection
Inspector HELMS,WILLIAM,A **Time** 17:00:02 **Software** 8100.27

0.05 Lot # 202506E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 23725080A1
Expiration 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 10/05/2027

Number of Simulators 4 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	Yes
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.047
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.193
0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.196
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.196
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

Time-Date changed.

Data Download Date/Time 10/27/2025 17:14

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002231

Agency NEW SMYRNA BEACH PD **Date** 11/26/2025 **Type** Agency Inspection
Inspector HUNNEFELD,RALPH,A **Time** 08:21:19 **Software** 8100.27

0.05 Lot # 202506E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 23725080A1
Expiration 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 10/05/2027

Number of Simulators 4 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.193
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.047
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.195
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.196
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 11/26/2025 10:22

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002231

Agency NEW SMYRNA BEACH PD **Date** 12/22/2025 **Type** Agency Inspection
Inspector HELMS,WILLIAM,A **Time** 14:53:44 **Software** 8100.27

0.05 Lot # 202506E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 23725080A1
Expiration 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 10/05/2027

Number of Simulators 4 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	Yes
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.194
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

Time-Date changed.

Data Download Date/Time 12/23/2025 00:59

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

Agency MIAMI BEACH PD **Date** 01/16/2025 **Type** Agency Inspection
Inspector SOCARRAS,PEDRO,A **Time** 11:26:16 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 402477284
Expiration 03/28/2025 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 06/27/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.047
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.196
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.047
0.08 g/210L Test # 2	0.077
0.20 g/210L Test # 2	0.196
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.047
0.08 g/210L Test # 3	0.077
0.20 g/210L Test # 3	0.196
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

AI PRICE #958 PRESENT

Data Download Date/Time 01/16/2025 13:24

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

Agency MIAMI BEACH PD **Date** 02/20/2025 **Type** Agency Inspection
Inspector SOCARRAS,PEDRO,A **Time** 10:22:16 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 402477284
Expiration 03/28/2025 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 06/27/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.047
0.08 g/210L Test # 1	0.076
0.20 g/210L Test # 1	0.196
0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.047
0.08 g/210L Test # 2	0.077
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.076
0.20 g/210L Test # 3	0.196
0.08 g/210L Dry Gas Std Test # 3	0.077
Interferent Detect Test # 3	INT

Remarks:

AI MILLAN #141 PRESENT

Data Download Date/Time 02/20/2025 11:16

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

Agency	MIAMI BEACH PD	**Date**	11/13/2025	**Type**	Agency Inspection
Inspector	GUTSCHOW,TAYLOR,D	**Time**	09:48:37	**Software**	8100.27

0.05 Lot #	**0.08 Lot #**	**0.20 Lot #**	**0.08 Gas Lot #**
Expiration	**Expiration**	**Expiration**	**Expiration**

Number of Simulators **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Data Download Date/Time 11/13/2025 09:51

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

Agency MIAMI BEACH PD **Date** 11/14/2025 **Type** Department Inspection
Inspector GUTSCHOW,TAYLOR,D **Time** 13:32:44 **Software** 8100.27

0.05 Lot # 202406K **0.08 Lot #** 202406L **0.20 Lot #** 202406N **0.08 Gas Lot #** AG429602
Expiration 06/19/2026 **Expiration** 06/19/2026 **Expiration** 06/20/2026 **Expiration** 10/22/2026

Number of Simulators 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.051
0.08 g/210L Test # 1	0.082
0.20 g/210L Test # 1	0.201
0.08 g/210L Dry Gas Std Test # 1	0.076
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.051
0.08 g/210L Test # 2	0.082
0.20 g/210L Test # 2	0.201
0.08 g/210L Dry Gas Std Test # 2	0.076
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.051
0.08 g/210L Test # 3	0.082
0.20 g/210L Test # 3	0.201
0.08 g/210L Dry Gas Std Test # 3	0.075
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.051
0.08 g/210L Test # 4	0.081
0.20 g/210L Test # 4	0.202
0.08 g/210L Dry Gas Std Test # 4	0.076
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.051
0.08 g/210L Test # 5	0.081
0.20 g/210L Test # 5	0.202
0.08 g/210L Dry Gas Std Test # 5	0.075
Interferent Detect Test # 5	INT

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.051
0.08 g/210L Test # 6	0.081
0.20 g/210L Test # 6	0.201
0.08 g/210L Dry Gas Std Test # 6	0.076
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.051
0.08 g/210L Test # 7	0.081
0.20 g/210L Test # 7	0.201
0.08 g/210L Dry Gas Std Test # 7	0.076
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.051
0.08 g/210L Test # 8	0.081
0.20 g/210L Test # 8	0.201
0.08 g/210L Dry Gas Std Test # 8	0.075
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.051
0.08 g/210L Test # 9	0.081
0.20 g/210L Test # 9	0.202
0.08 g/210L Dry Gas Std Test # 9	0.076
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.051
0.08 g/210L Test # 10	0.081
0.20 g/210L Test # 10	0.201
0.08 g/210L Dry Gas Std Test # 10	0.075
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0000
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0004
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

Data Download Date/Time 11/14/2025 13:37

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

Agency MIAMI BEACH PD **Date** 12/11/2025 **Type** Agency Inspection
Inspector SOCARRAS,PEDRO,A **Time** 10:55:32 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 402872586
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 10/12/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.051
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.201
0.08 g/210L Dry Gas Std Test # 1	0.076
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.051
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.201
0.08 g/210L Dry Gas Std Test # 2	0.075
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.052
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.201
0.08 g/210L Dry Gas Std Test # 3	0.075
Interferent Detect Test # 3	INT

Remarks:

AI PRICE #958 PRESENT

Data Download Date/Time 12/11/2025 11:39

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002418

Agency	BREVARD STATE COLL	**Date**	01/08/2025	**Type**	Agency Inspection
Inspector	WAGNER,ROBERT,F	**Time**	10:12:34	**Software**	8100.27

0.05 Lot #	**0.08 Lot #**	**0.20 Lot #**	**0.08 Gas Lot #**
Expiration	**Expiration**	**Expiration**	**Expiration**

Number of Simulators **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Data Download Date/Time 11/19/2025 13:25

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002418

Agency	BREVARD STATE COLL	**Date**	02/25/2025	**Type**	Agency Inspection
Inspector	WAGNER,ROBERT,F	**Time**	11:24:04	**Software**	8100.27

0.05 Lot #	**0.08 Lot #**	**0.20 Lot #**	**0.08 Gas Lot #**
Expiration	**Expiration**	**Expiration**	**Expiration**

Number of Simulators **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Data Download Date/Time 11/19/2025 13:25

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002418

Agency BREVARD STATE COLL **Date** 02/25/2025 **Type** Agency Inspection
Inspector WAGNER,ROBERT,F **Time** 14:59:29 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 30510080A1
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 12/25/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.047
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.190
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.047
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.193
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.194
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 11/19/2025 13:25

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002418

Agency BREVARD STATE COLL **Date** 02/25/2025 **Type** Agency Inspection
Inspector BLACK,TIMOTHY,J **Time** 16:09:40 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 30510080A1
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 12/01/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.194
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.196
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.195
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 11/19/2025 13:25

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002418

Agency BREVARD STATE COLL **Date** 02/26/2025 **Type** Agency Inspection
Inspector AULTMAN,SHAWN,A **Time** 10:14:19 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 30510080A1
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 12/01/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.194
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.196
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.196
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 11/19/2025 13:25

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002418

Agency BREVARD STATE COLL **Date** 02/26/2025 **Type** Agency Inspection
Inspector BLACK,TIMOTHY,J **Time** 10:56:31 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 20241OE **0.20 Lot #** 202410F **0.08 Gas Lot #** 30510080A1
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 12/01/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.047
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.196
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.196
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 11/19/2025 13:25

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002418

Agency BREVARD STATE COLL **Date** 02/26/2025 **Type** Agency Inspection
Inspector AULTMAN,SHAWN,A **Time** 14:38:05 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 30510080A1
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 12/01/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.200
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.200
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.200
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 11/19/2025 13:25

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002418

Agency BREVARD STATE COLL **Date** 02/26/2025 **Type** Agency Inspection
Inspector RENTAS,N,M **Time** 15:15:51 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 30510080A1
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 12/01/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.195
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.196
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.196
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 11/19/2025 13:25

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002418

Agency BREVARD STATE COLL **Date** 02/27/2025 **Type** Agency Inspection
Inspector ETTMAN,KRISTOFER,C **Time** 10:37:47 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 30510080A1
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 12/01/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.047
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.193
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.195
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.195
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 11/19/2025 13:25

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002418

Agency BREVARD STATE COLL **Date** 02/27/2025 **Type** Agency Inspection
Inspector RENTAS,N, **Time** 11:14:45 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 30510080A1
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 12/01/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.195
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.196
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 11/19/2025 13:25

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002418

Agency BREVARD STATE COLL **Date** 02/27/2025 **Type** Agency Inspection
Inspector BLACK,TIMOTHY,J **Time** 13:23:39 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 30510080A1
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 12/01/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.195
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.196
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 11/19/2025 13:25

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002418

Agency	BREVARD STATE COLL	**Date**	03/25/2025	**Type**	Agency Inspection
Inspector	MORGAN,DARIN,	**Time**	13:05:14	**Software**	8100.27

0.05 Lot #	**0.08 Lot #**	**0.20 Lot #**	**0.08 Gas Lot #**
Expiration	**Expiration**	**Expiration**	**Expiration**

Number of Simulators **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Data Download Date/Time 11/19/2025 13:25

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002418

Agency	BREVARD STATE COLL	**Date**	04/02/2025	**Type**	Agency Inspection
Inspector	MORGAN,D,	**Time**	11:01:02	**Software**	8100.27

0.05 Lot #	**0.08 Lot #**	**0.20 Lot #**	**0.08 Gas Lot #**
Expiration	**Expiration**	**Expiration**	**Expiration**

Number of Simulators **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Data Download Date/Time 11/19/2025 13:25

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002418

Agency	BREVARD STATE COLL	**Date**	05/21/2025	**Type**	Agency Inspection
Inspector	WAGNER,ROBERT,F	**Time**	09:51:12	**Software**	8100.27

0.05 Lot #	**0.08 Lot #**	**0.20 Lot #**	**0.08 Gas Lot #**
Expiration	**Expiration**	**Expiration**	**Expiration**

Number of Simulators **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Data Download Date/Time 11/19/2025 13:25

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002418

Agency	BREVARD STATE COLL	**Date**	07/16/2025	**Type**	Agency Inspection
Inspector	MORGAN,D,	**Time**	10:53:13	**Software**	8100.27

0.05 Lot #	**0.08 Lot #**	**0.20 Lot #**	**0.08 Gas Lot #**
Expiration	**Expiration**	**Expiration**	**Expiration**

Number of Simulators **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Data Download Date/Time 11/19/2025 13:25

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002418

Agency	BREVARD STATE COLL	**Date**	10/13/2025	**Type**	Agency Inspection
Inspector	WAGNER,ROBERT,F	**Time**	08:58:31	**Software**	8100.27

0.05 Lot #	**0.08 Lot #**	**0.20 Lot #**	**0.08 Gas Lot #**
Expiration	**Expiration**	**Expiration**	**Expiration**

Number of Simulators **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Data Download Date/Time 11/19/2025 13:25

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002418

Agency	BREVARD STATE COLL	**Date**	11/17/2025	**Type**	Agency Inspection
Inspector	MORGAN,D,	**Time**	13:45:24	**Software**	8100.27

0.05 Lot #	**0.08 Lot #**	**0.20 Lot #**	**0.08 Gas Lot #**
Expiration	**Expiration**	**Expiration**	**Expiration**

Number of Simulators **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Data Download Date/Time 11/19/2025 13:25

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002462

Agency HIALEAH PD **Date** 01/22/2025 **Type** Agency Inspection
Inspector DELGADO,ARMANDO, **Time** 10:51:23 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 402370890
Expiration 02/27/2026 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 05/09/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.195
0.08 g/210L Dry Gas Std Test # 1	0.076
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.196
0.08 g/210L Dry Gas Std Test # 2	0.077
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.196
0.08 g/210L Dry Gas Std Test # 3	0.077
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/22/2025 12:17

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002462

Agency HIALEAH PD **Date** 02/10/2025 **Type** Department Inspection
Inspector GUTSCHOW,TAYLOR,D **Time** 11:46:38 **Software** 8100.27

0.05 Lot # 202303K **0.08 Lot #** 202303L **0.20 Lot #** 202304C **0.08 Gas Lot #** AG429602
Expiration 03/29/2025 **Expiration** 03/29/2025 **Expiration** 04/05/2025 **Expiration** 10/22/2026

Number of Simulators 5 **Average Standard Deviation** 0.0009 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.191
0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.077
0.20 g/210L Test # 2	0.196
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.196
0.08 g/210L Dry Gas Std Test # 3	0.078
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.048
0.08 g/210L Test # 4	0.077
0.20 g/210L Test # 4	0.196
0.08 g/210L Dry Gas Std Test # 4	0.079
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.049
0.08 g/210L Test # 5	0.079
0.20 g/210L Test # 5	0.196
0.08 g/210L Dry Gas Std Test # 5	0.078
Interferent Detect Test # 5	INT

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002462

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.049
0.08 g/210L Test # 6	0.079
0.20 g/210L Test # 6	0.196
0.08 g/210L Dry Gas Std Test # 6	0.078
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.049
0.08 g/210L Test # 7	0.079
0.20 g/210L Test # 7	0.196
0.08 g/210L Dry Gas Std Test # 7	0.078
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.049
0.08 g/210L Test # 8	0.079
0.20 g/210L Test # 8	0.192
0.08 g/210L Dry Gas Std Test # 8	0.079
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.049
0.08 g/210L Test # 9	0.078
0.20 g/210L Test # 9	0.196
0.08 g/210L Dry Gas Std Test # 9	0.079
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.049
0.08 g/210L Test # 10	0.079
0.20 g/210L Test # 10	0.196
0.08 g/210L Dry Gas Std Test # 10	0.079
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0008
0.20 g/210L Test (Std Dev)	0.0019
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

Data Download Date/Time 02/10/2025 12:31

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002462

Agency HIALEAH PD **Date** 02/14/2025 **Type** Department Inspection
Inspector GUTSCHOW,TAYLOR,D **Time** 11:53:35 **Software** 8100.27

0.05 Lot # 202303K **0.08 Lot #** 202303L **0.20 Lot #** 202304C **0.08 Gas Lot #** AG429602
Expiration 03/29/2025 **Expiration** 03/29/2025 **Expiration** 04/05/2025 **Expiration** 10/22/2026

Number of Simulators 5 **Average Standard Deviation** 0.0005 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.051
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.200
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.051
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.200
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.051
0.08 g/210L Test # 4	0.080
0.20 g/210L Test # 4	0.196
0.08 g/210L Dry Gas Std Test # 4	0.080
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.051
0.08 g/210L Test # 5	0.080
0.20 g/210L Test # 5	0.198
0.08 g/210L Dry Gas Std Test # 5	0.080
Interferent Detect Test # 5	INT

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002462

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.051
0.08 g/210L Test # 6	0.080
0.20 g/210L Test # 6	0.196
0.08 g/210L Dry Gas Std Test # 6	0.080
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.051
0.08 g/210L Test # 7	0.081
0.20 g/210L Test # 7	0.197
0.08 g/210L Dry Gas Std Test # 7	0.080
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.051
0.08 g/210L Test # 8	0.080
0.20 g/210L Test # 8	0.197
0.08 g/210L Dry Gas Std Test # 8	0.080
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.051
0.08 g/210L Test # 9	0.080
0.20 g/210L Test # 9	0.198
0.08 g/210L Dry Gas Std Test # 9	0.080
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.051
0.08 g/210L Test # 10	0.081
0.20 g/210L Test # 10	0.198
0.08 g/210L Dry Gas Std Test # 10	0.080
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0003
0.08 g/210L Test (Std Dev)	0.0005
0.20 g/210L Test (Std Dev)	0.0014
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0000

Remarks:

Data Download Date/Time 02/14/2025 11:54

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002462

Agency	HIALEAH PD	**Date**	09/30/2025	**Type**	Agency Inspection
Inspector	GUTSCHOW,TAYLOR,D	**Time**	12:45:24	**Software**	8100.27

0.05 Lot #	**0.08 Lot #**	**0.20 Lot #**	**0.08 Gas Lot #**
Expiration	**Expiration**	**Expiration**	**Expiration**

Number of Simulators **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

AI NOT CONDUCTED. COMPLIANCE NOT DETERMINED.

Data Download Date/Time 10/01/2025 11:00

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002463

Agency HIALEAH PD **Date** 01/22/2025 **Type** Agency Inspection
Inspector PEREZ,VINCENTE, **Time** 10:55:38 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 402370890
Expiration 02/27/2026 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 05/09/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/22/2025 11:56

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002463

Agency HIALEAH PD **Date** 02/06/2025 **Type** Agency Inspection
Inspector MONTERO,JOSE, **Time** 09:38:05 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 402370890
Expiration 02/27/2026 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 05/09/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.196
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.196
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.196
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/06/2025 09:45

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002463

Agency HIALEAH PD **Date** 03/12/2025 **Type** Agency Inspection
Inspector MONTERO,JOSE,. **Time** 09:41:38 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202307C **0.08 Gas Lot #** 402370890
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 07/12/2025 **Expiration** 05/09/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.078
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 03/12/2025 09:46

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002463

Agency HIALEAH PD **Date** 04/10/2025 **Type** Agency Inspection
Inspector MONTERO,JOSE, **Time** 10:16:03 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 402370890
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 05/09/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 04/10/2025 10:27

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002463

Agency HIALEAH PD **Date** 05/08/2025 **Type** Agency Inspection
Inspector MONTERO,JOSE, **Time** 09:58:18 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 403255643
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 01/29/2028

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 05/08/2025 10:02

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002463

Agency HIALEAH PD **Date** 06/04/2025 **Type** Agency Inspection
Inspector MONTERO,JOSE, **Time** 09:49:13 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 403255643
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 01/29/2028

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 06/04/2025 09:55

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002463

Agency HIALEAH PD **Date** 07/11/2025 **Type** Department Inspection
Inspector GUTSCHOW,TAYLOR,D **Time** 16:24:59 **Software** 8100.27

0.05 Lot # 202406K **0.08 Lot #** 202406L **0.20 Lot #** 202406N **0.08 Gas Lot #** AG429602
Expiration 06/19/2026 **Expiration** 06/19/2026 **Expiration** 06/20/2026 **Expiration** 10/22/2026

Number of Simulators 5 **Average Standard Deviation** 0.0002 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.077
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.048
0.08 g/210L Test # 4	0.079
0.20 g/210L Test # 4	0.198
0.08 g/210L Dry Gas Std Test # 4	0.077
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.048
0.08 g/210L Test # 5	0.079
0.20 g/210L Test # 5	0.198
0.08 g/210L Dry Gas Std Test # 5	0.077
Interferent Detect Test # 5	INT

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002463

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.048
0.08 g/210L Test # 6	0.079
0.20 g/210L Test # 6	0.198
0.08 g/210L Dry Gas Std Test # 6	0.078
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.048
0.08 g/210L Test # 7	0.079
0.20 g/210L Test # 7	0.198
0.08 g/210L Dry Gas Std Test # 7	0.077
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.048
0.08 g/210L Test # 8	0.079
0.20 g/210L Test # 8	0.198
0.08 g/210L Dry Gas Std Test # 8	0.078
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.048
0.08 g/210L Test # 9	0.079
0.20 g/210L Test # 9	0.198
0.08 g/210L Dry Gas Std Test # 9	0.078
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.048
0.08 g/210L Test # 10	0.078
0.20 g/210L Test # 10	0.198
0.08 g/210L Dry Gas Std Test # 10	0.077
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0000
0.08 g/210L Test (Std Dev)	0.0003
0.20 g/210L Test (Std Dev)	0.0000
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

Data Download Date/Time 07/11/2025 16:26

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002463

Agency HIALEAH PD **Date** 08/06/2025 **Type** Agency Inspection
Inspector MONTERO,JOSE, **Time** 10:26:48 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 403255643
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 01/29/2028

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.078
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 08/06/2025 10:32

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002463

Agency HIALEAH PD **Date** 09/04/2025 **Type** Agency Inspection
Inspector MONTERO,JOSE, **Time** 10:27:13 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 403255643
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 01/29/2028

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	0.000
Interferent Detect Test # (Repeat) 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	0.000
Interferent Detect Test # (Repeat) 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	0.000
Interferent Detect Test # (Repeat) 3	INT

Remarks:

Int Det:RE-DID SOLUTION,FDLE TAYLOR ADVISED. .

Data Download Date/Time 09/04/2025 10:31

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002463

Agency HIALEAH PD **Date** 10/09/2025 **Type** Agency Inspection
Inspector MONTERO,JOSE, **Time** 10:25:55 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 403255643
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 01/29/2028

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.078
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 10/09/2025 10:30

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002463

Agency HIALEAH PD **Date** 11/05/2025 **Type** Agency Inspection
Inspector MONTERO,JOSE, **Time** 11:50:21 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 403255643
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 01/29/2028

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.078
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 11/05/2025 11:52

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002463

Agency HIALEAH PD **Date** 12/03/2025 **Type** Agency Inspection
Inspector MONTERO,JOSE, **Time** 09:31:53 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 403255643
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 01/29/2028

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.078
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 12/03/2025 09:31

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002563

Agency	DAYTONA COLLEGE CJTC	**Date**	02/12/2025	**Type**	Agency Inspection
Inspector	WAGNER,ROBERT,F	**Time**	15:45:22	**Software**	8100.27

0.05 Lot #	**0.08 Lot #**	**0.20 Lot #**	**0.08 Gas Lot #**
Expiration	**Expiration**	**Expiration**	**Expiration**

Number of Simulators **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Data Download Date/Time 01/16/2026 13:54

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002563

Agency	DAYTONA COLLEGE CJTC	**Date**	03/12/2025	**Type**	Agency Inspection
Inspector	WAGNER,ROBERT,F	**Time**	09:44:02	**Software**	8100.27

0.05 Lot #	**0.08 Lot #**	**0.20 Lot #**	**0.08 Gas Lot #**
Expiration	**Expiration**	**Expiration**	**Expiration**

Number of Simulators **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Data Download Date/Time 01/16/2026 13:54

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002563

Agency DAYTONA COLLEGE CJTC **Date** 03/12/2025 **Type** Agency Inspection
Inspector ELLIS,ALEXANDER, **Time** 11:18:36 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 166440
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 10/30/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.078
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/16/2026 13:54

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002563

Agency	DAYTONA COLLEGE CJTC	**Date**	06/04/2025	**Type**	Agency Inspection
Inspector	WAGNER,ROBERT,F	**Time**	08:09:47	**Software**	8100.27

0.05 Lot #	**0.08 Lot #**	**0.20 Lot #**	**0.08 Gas Lot #**
Expiration	**Expiration**	**Expiration**	**Expiration**

Number of Simulators **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Data Download Date/Time 01/16/2026 13:54

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002563

Agency DAYTONA COLLEGE CJTC **Date** 06/05/2025 **Type** Agency Inspection
Inspector BARNETT,JAMES,C **Time** 11:07:02 **Software** 8100.27

0.05 Lot # **0.08 Lot #** **0.20 Lot #** **0.08 Gas Lot #**
Expiration **Expiration** **Expiration** **Expiration**

Number of Simulators 5 **Compliance** No

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	No
Diagnostic Check (Post-Inspection): OK	No
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	No
0.08 g/210L Test: OK	No
0.20 g/210L Test: OK	No
0.08 g/210L Dry Gas Std Test: OK	No

Test	Result
Alcohol Free Test # 1	0.000
Interferent Detect Test # 1	0.000
Alcohol Free Test # 2	0.000
Interferent Detect Test # 2	0.000
Alcohol Free Test # 3	0.000
Interferent Detect Test # 3	0.000

Remarks:

Int Det: . Non-compliance: .

Data Download Date/Time 01/16/2026 13:54

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002563

Agency DAYTONA COLLEGE CJTC **Date** 06/05/2025 **Type** Agency Inspection
Inspector BARNETT,JAMES,C **Time** 11:46:24 **Software** 8100.27

0.05 Lot # 202101C **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 1187042
Expiration 01/12/2026 **Expiration** 06/14/2026 **Expiration** 06/15/2026 **Expiration** 07/22/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.195
0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.196
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.196
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

Y

Data Download Date/Time 01/16/2026 13:54

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002563

Agency	DAYTONA COLLEGE CJTC	**Date**	06/25/2025	**Type**	Agency Inspection
Inspector	WAGNER,ROBERT,F	**Time**	01:51:33	**Software**	8100.27

0.05 Lot #	**0.08 Lot #**	**0.20 Lot #**	**0.08 Gas Lot #**
Expiration	**Expiration**	**Expiration**	**Expiration**

Number of Simulators **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Data Download Date/Time 01/16/2026 13:54

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002563

Agency	DAYTONA COLLEGE CJTC	**Date**	11/12/2025	**Type**	Agency Inspection
Inspector	WAGNER,ROBERT,F	**Time**	14:35:55	**Software**	8100.27

0.05 Lot #	**0.08 Lot #**	**0.20 Lot #**	**0.08 Gas Lot #**
Expiration	**Expiration**	**Expiration**	**Expiration**

Number of Simulators **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Data Download Date/Time 01/16/2026 13:54

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002563

Agency	DAYTONA COLLEGE CJTC	**Date**	12/10/2025	**Type**	Agency Inspection
Inspector	WAGNER,ROBERT,F	**Time**	13:07:44	**Software**	8100.27

0.05 Lot #	**0.08 Lot #**	**0.20 Lot #**	**0.08 Gas Lot #**
Expiration	**Expiration**	**Expiration**	**Expiration**

Number of Simulators **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Data Download Date/Time 01/16/2026 13:54

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002987

Agency MIAMI GARDENS PD **Date** 01/06/2025 **Type** Agency Inspection
Inspector OMALLEY,BRIAN,P **Time** 11:12:34 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 403005538
Expiration 03/28/2025 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 03/25/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.046
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.196
0.08 g/210L Dry Gas Std Test # 1	0.082
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.047
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.082
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.047
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.082
Interferent Detect Test # 3	INT

Remarks:

IN COMPLIANCE

Data Download Date/Time 01/06/2025 11:15

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002987

Agency MIAMI GARDENS PD **Date** 02/26/2025 **Type** Agency Inspection
Inspector OMALLEY,BRIAN,P **Time** 13:58:52 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 403005538
Expiration 03/28/2025 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 03/25/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.045
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.045
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

IN COMPLIANCE

Data Download Date/Time 02/26/2025 14:06

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002987

Agency	MIAMI GARDENS PD	**Date**	05/19/2025	**Type**	Agency Inspection
Inspector	OMALLEY,BRIAN,P	**Time**	16:11:34	**Software**	8100.27

0.05 Lot #	**0.08 Lot #**	**0.20 Lot #**	**0.08 Gas Lot #**
Expiration	**Expiration**	**Expiration**	**Expiration**

Number of Simulators 5 **Compliance** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	Interferent Detect Test # 1	0.000
Minimum Sample Volume Check: OK		Interferent Detect Test # (Repeat) 1	0.000
Barometric Pressure Sensor Check: OK		Alcohol Free Test # 2	0.000
Alcohol Free Subject Test: 0.000	Yes	Interferent Detect Test # 2	INT
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # (Repeat) 2	INT
Interferent Detect Test: Interferent Detect	No	Alcohol Free Test # 3	0.000
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test # 3	INT
Alcohol Free Test: OK	Yes	Interferent Detect Test # (Repeat) 3	INT
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Int Det: , . Non-compliance: .

Data Download Date/Time 06/23/2025 12:55

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002987

Agency	MIAMI GARDENS PD	**Date**	05/20/2025	**Type**	Agency Inspection
Inspector	OMALLEY,BRIAN,P	**Time**	11:26:18	**Software**	8100.27

0.05 Lot #	**0.08 Lot #**	**0.20 Lot #**	**0.08 Gas Lot #**
Expiration	**Expiration**	**Expiration**	**Expiration**

Number of Simulators 5 **Compliance** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	Interferent Detect Test # 1	0.000
Minimum Sample Volume Check: OK		Interferent Detect Test # (Repeat) 1	0.000
Barometric Pressure Sensor Check: OK		Alcohol Free Test # 2	0.000
Alcohol Free Subject Test: 0.000	Yes	Interferent Detect Test # 2	0.000
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # (Repeat) 2	0.000
Interferent Detect Test: Interferent Detect	No	Alcohol Free Test # 3	0.000
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test # 3	0.000
Alcohol Free Test: OK	Yes	Interferent Detect Test # (Repeat) 3	0.000
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Int Det: , . Non-compliance: .

Data Download Date/Time 06/23/2025 12:55

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002987

Agency MIAMI GARDENS PD **Date** 06/23/2025 **Type** Agency Inspection
Inspector OMALLEY,BRIAN,P **Time** 12:55:00 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 403005538
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 03/25/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.196
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.082
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.082
Interferent Detect Test # 3	INT

Remarks:

IN COMPLIANCE

Data Download Date/Time 06/23/2025 12:56

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002987

Agency MIAMI GARDENS PD **Date** 07/28/2025 **Type** Agency Inspection
Inspector OMALLEY,BRIAN,P **Time** 15:57:09 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 403005538
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 03/25/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.047
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.196
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.046
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.046
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

IN COMPLIANCE

Data Download Date/Time 07/28/2025 15:57

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002987

Agency MIAMI GARDENS PD **Date** 08/13/2025 **Type** Department Inspection
Inspector GUTSCHOW,TAYLOR,D **Time** 13:24:31 **Software** 8100.27

0.05 Lot # 202406K **0.08 Lot #** 202406L **0.20 Lot #** 202406N **0.08 Gas Lot #** AG429602
Expiration 06/19/2026 **Expiration** 06/19/2026 **Expiration** 06/20/2026 **Expiration** 10/22/2026

Number of Simulators 5 **Average Standard Deviation** 0.0002 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.047
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.048
0.08 g/210L Test # 4	0.078
0.20 g/210L Test # 4	0.199
0.08 g/210L Dry Gas Std Test # 4	0.079
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.048
0.08 g/210L Test # 5	0.078
0.20 g/210L Test # 5	0.199
0.08 g/210L Dry Gas Std Test # 5	0.079
Interferent Detect Test # 5	INT

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002987

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.048
0.08 g/210L Test # 6	0.078
0.20 g/210L Test # 6	0.199
0.08 g/210L Dry Gas Std Test # 6	0.079
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.047
0.08 g/210L Test # 7	0.078
0.20 g/210L Test # 7	0.199
0.08 g/210L Dry Gas Std Test # 7	0.079
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.048
0.08 g/210L Test # 8	0.078
0.20 g/210L Test # 8	0.198
0.08 g/210L Dry Gas Std Test # 8	0.080
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.048
0.08 g/210L Test # 9	0.078
0.20 g/210L Test # 9	0.199
0.08 g/210L Dry Gas Std Test # 9	0.079
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.047
0.08 g/210L Test # 10	0.078
0.20 g/210L Test # 10	0.199
0.08 g/210L Dry Gas Std Test # 10	0.079
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0000
0.20 g/210L Test (Std Dev)	0.0004
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

Data Download Date/Time 08/13/2025 13:23

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002987

Agency MIAMI GARDENS PD **Date** 08/18/2025 **Type** Agency Inspection
Inspector OMALLEY,BRIAN,P **Time** 16:03:39 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 403005538
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 03/25/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.047
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.196
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.047
0.08 g/210L Test # 2	0.077
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.047
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

IN COMPLIANCE

Data Download Date/Time 08/18/2025 16:02

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002987

Agency MIAMI GARDENS PD **Date** 09/30/2025 **Type** Agency Inspection
Inspector OMALLEY,BRIAN,P **Time** 12:44:57 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 403005538
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 03/25/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.045
0.08 g/210L Test # 1	0.075
0.20 g/210L Test # 1	0.193
0.08 g/210L Dry Gas Std Test # 1	0.082
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.046
0.08 g/210L Test # 2	0.075
0.20 g/210L Test # 2	0.194
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.045
0.08 g/210L Test # 3	0.075
0.20 g/210L Test # 3	0.194
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

IN COMPLIANCE

Data Download Date/Time 09/30/2025 12:45

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002987

Agency MIAMI GARDENS PD **Date** 10/28/2025 **Type** Agency Inspection
Inspector OMALLEY,BRIAN,P **Time** 11:56:39 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 403005538
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 03/25/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

IN COMPLIANCE

Data Download Date/Time 10/28/2025 11:56

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002987

Agency MIAMI GARDENS PD **Date** 11/20/2025 **Type** Agency Inspection
Inspector OMALLEY,BRIAN,P **Time** 15:53:47 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 403005538
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 03/25/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

IN COMPLIANCE

Data Download Date/Time 11/20/2025 16:06

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002987

Agency MIAMI GARDENS PD **Date** 12/31/2025 **Type** Agency Inspection
Inspector OMALLEY,BRIAN,P **Time** 10:19:16 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 403005538
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 03/25/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.082
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.082
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.082
Interferent Detect Test # 3	INT

Remarks:

IN COMPLIANCE

Data Download Date/Time 12/31/2025 10:17

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.