

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: OKALOOSA COUNTY SO
Time of Inspection: 10:05

Date of Inspection: 06/29/2022

Serial Number: 80-000763
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202101C Exp: 01/12/2023	0.08g/210L Test (g/210L) Lot#:202108D Exp: 08/17/2023	0.20g/210L Test (g/210L) Lot#:202106B Exp: 06/22/2023	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:1349789 Exp: 07/15/2023
0.000	0.051	0.081	0.200	0.081
0.000	0.050	0.080	0.200	0.081
0.000	0.049	0.081	0.201	0.081


Number of Simulators Used: 5

Remarks:

A F / M A: RFI Detect No Sample Provided ISRAEL SOTO CONT - AF/MA RFI Detected and MA and test reinitiated. Passed 2nd test (human error). Israel Soto contacted and made aware.

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

 JONATHAN D SCHLAGER
Signature and Printed Name

06/29/2022
Date