

Instrument Processing Sheet

Agency: FL HIGHWAY PATROL Instrument Serial Number: 80-000784
 Date In: 12/10/2025 DI Completion Date: 12/11/2025 Ship P/U H/D CMI EE

Intake By: <u>WKP</u> Date: <u>12/10/25</u>	Quality Checks By: <u>KTS</u> Date: <u>12/11/25</u>	Flow Adjustment By: _____ Date: _____															
<input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input checked="" type="checkbox"/> Return from CMI / EE <input type="checkbox"/> Return unworked <input type="checkbox"/> Training Visual Inspection <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/Accessories <input type="checkbox"/> Power Cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes:	<input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value: <u>235</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column #: <u>ATP103</u> 32 mm <u>0.160</u> (.139-.169) 36 mm <u>0.179</u> (.156-.190) 53 mm <u>0.246</u> (.228-.278) 103 mm <u>0.500</u> (.447-.547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID #: <u>28427</u> Gauge: <u>1012</u> Instrument: <u>1012</u> <input checked="" type="checkbox"/> Stability Checks	Flow Column #: _____ <input type="checkbox"/> 5L/min – 17mm <input type="checkbox"/> 15L/min – 53mm <input type="checkbox"/> 30L/min – 103mm <input type="checkbox"/> R-Value: _____ <input type="checkbox"/> Post Adjustment Verification (L/S) Flow Column #: _____ 32 mm _____ (.139-.169) 36 mm _____ (.156-.190) 53 mm _____ (.228-.278) 103 mm _____ (.447-.547)															
<table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot#/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>MP5088</td> <td>202406K 6/19/2026</td> </tr> <tr> <td>0.080</td> <td>MP5089</td> <td>202406L 6/19/2026</td> </tr> <tr> <td>0.200</td> <td>MP5090</td> <td>202406N 6/20/2026</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG510701 4/17/2027</td> </tr> </tbody> </table>		Simulator	Serial #	Lot#/Exp	0.050	MP5088	202406K 6/19/2026	0.080	MP5089	202406L 6/19/2026	0.200	MP5090	202406N 6/20/2026	0.080 DGS	N/A	AG510701 4/17/2027	Maintenance By: _____ Date: _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement and Tank Sensor Tare <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other:
Simulator	Serial #	Lot#/Exp															
0.050	MP5088	202406K 6/19/2026															
0.080	MP5089	202406L 6/19/2026															
0.200	MP5090	202406N 6/20/2026															
0.080 DGS	N/A	AG510701 4/17/2027															

Optical Bench Adjustment By: _____	Department Inspection By: <u>KTS</u>																																								
Barometric Pressure Gauge: _____ ID#: _____	Barometric Pressure ID#: <u>28427</u>																																								
<table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>	Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Gauge: <u>1012</u> Instrument: <u>1013</u> Mouth Alcohol Solution Lot #: <u>2025-D</u> Exp: <u>9/25/2027</u> Acetone Stock Solution Lot #: <u>2025-B</u> Exp: <u>9/22/2027</u> <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>MP5086</td> </tr> <tr> <td>Interferent</td> <td>MP5087</td> </tr> <tr> <td>0.050</td> <td>MP5088</td> </tr> <tr> <td>0.080</td> <td>MP5089</td> </tr> <tr> <td>0.200</td> <td>MP5090</td> </tr> </tbody> </table>	Simulator	Serial Number	0.000	MP5086	Interferent	MP5087	0.050	MP5088	0.080	MP5089	0.200	MP5090
Simulator	Serial #	Lot #	Expiration																																						
0.000		N/A	N/A																																						
0.040																																									
0.100																																									
0.200																																									
0.300																																									
0.080 DGS	N/A																																								
Simulator	Serial Number																																								
0.000	MP5086																																								
Interferent	MP5087																																								
0.050	MP5088																																								
0.080	MP5089																																								
0.200	MP5090																																								
<input type="checkbox"/> Post Optical Bench Adjustment Stability Checks <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>	Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			Attachments <input checked="" type="checkbox"/> Form 41 <input type="checkbox"/> Post-Stability Checks <input checked="" type="checkbox"/> Stability Checks <input type="checkbox"/> Flow Adjustment <input checked="" type="checkbox"/> Calibration Certificate <input type="checkbox"/> Form 40 <input type="checkbox"/> Optical Bench Adjustment <input type="checkbox"/> Other:																				
Simulator	Serial #	Lot #	Expiration																																						
0.050																																									
0.080																																									
0.200																																									
0.080 DGS	N/A																																								
Barometric Pressure Gauge: _____ ID#: _____																																									

Notes/Suggested Service: Tech Review: Added check mark to return method. KTS 12/12/25 Admin Review: Added check mark to static bag for intake. KTS 12/16/25	<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use				
	<table style="width: 100%;"> <tr> <td style="width: 50%; text-align: center;"> Digitally signed by Taylor Gutschow <small>Date: 2025.12.12 11:01:52 -05'00'</small> </td> <td style="width: 50%; text-align: center;"> Digitally signed by LeAndra Higginbotham <small>Date: 2025.12.17 10:07:01 -05'00'</small> </td> </tr> <tr> <td style="text-align: center;">Taylor Gutschow</td> <td style="text-align: center;">LeAndra Higginbotham</td> </tr> </table>	Digitally signed by Taylor Gutschow <small>Date: 2025.12.12 11:01:52 -05'00'</small>	Digitally signed by LeAndra Higginbotham <small>Date: 2025.12.17 10:07:01 -05'00'</small>	Taylor Gutschow	LeAndra Higginbotham
Digitally signed by Taylor Gutschow <small>Date: 2025.12.12 11:01:52 -05'00'</small>	Digitally signed by LeAndra Higginbotham <small>Date: 2025.12.17 10:07:01 -05'00'</small>				
Taylor Gutschow	LeAndra Higginbotham				
	<table style="width: 100%;"> <tr> <td style="width: 50%; text-align: center;">Tech Review</td> <td style="width: 50%; text-align: center;">Admin Review</td> </tr> </table>	Tech Review	Admin Review		
Tech Review	Admin Review				

Stability Checks

0.050 g/210L	0.080 g/210L	0.200 g/210L	DGS 0.080 g/210L																																																																																																																																																
0.047 to 0.053 g/210L <input checked="" type="checkbox"/>	0.077 to 0.083 g/210L <input checked="" type="checkbox"/>	0.194 to 0.206 g/210L <input checked="" type="checkbox"/>	0.077 to 0.083 g/210L <input checked="" type="checkbox"/> ≤0.003 g/210L of Wet <input checked="" type="checkbox"/>																																																																																																																																																
Performed Root Case Analysis	Performed Root Case Analysis	Performed Root Case Analysis	Performed Root Case Analysis																																																																																																																																																
<p>FL HIGHWAY PATROL Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000784 12/11/2025 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>06:58</td></tr> <tr><td>Control Test</td><td>0.051</td><td>06:59</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>06:59</td></tr> <tr><td>Control Test</td><td>0.050</td><td>07:00</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>07:01</td></tr> <tr><td>Control Test</td><td>0.050</td><td>07:01</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>07:02</td></tr> <tr><td colspan="3">Control Test Stats</td></tr> <tr><td>Average</td><td>0.0503</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.1471</td><td></td></tr> </tbody> </table> <p> Operator's Signature</p>	Test	g/210L	Time	Air Blank	0.000	06:58	Control Test	0.051	06:59	Air Blank	0.000	06:59	Control Test	0.050	07:00	Air Blank	0.000	07:01	Control Test	0.050	07:01	Air Blank	0.000	07:02	Control Test Stats			Average	0.0503		Std Dev	0.0006		Rel Std Dev(%)	1.1471		<p>FL HIGHWAY PATROL Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000784 12/11/2025 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>07:03</td></tr> <tr><td>Control Test</td><td>0.081</td><td>07:04</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>07:04</td></tr> <tr><td>Control Test</td><td>0.080</td><td>07:05</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>07:06</td></tr> <tr><td>Control Test</td><td>0.080</td><td>07:06</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>07:07</td></tr> <tr><td colspan="3">Control Test Stats</td></tr> <tr><td>Average</td><td>0.0803</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7187</td><td></td></tr> </tbody> </table> <p> Operator's Signature</p>	Test	g/210L	Time	Air Blank	0.000	07:03	Control Test	0.081	07:04	Air Blank	0.000	07:04	Control Test	0.080	07:05	Air Blank	0.000	07:06	Control Test	0.080	07:06	Air Blank	0.000	07:07	Control Test Stats			Average	0.0803		Std Dev	0.0006		Rel Std Dev(%)	0.7187		<p>FL HIGHWAY PATROL Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000784 12/11/2025 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>07:09</td></tr> <tr><td>Control Test</td><td>0.205</td><td>07:09</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>07:10</td></tr> <tr><td>Control Test</td><td>0.203</td><td>07:11</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>07:11</td></tr> <tr><td>Control Test</td><td>0.204</td><td>07:12</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>07:13</td></tr> <tr><td colspan="3">Control Test Stats</td></tr> <tr><td>Average</td><td>0.2040</td><td></td></tr> <tr><td>Std Dev</td><td>0.0010</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.4902</td><td></td></tr> </tbody> </table> <p> Operator's Signature</p>	Test	g/210L	Time	Air Blank	0.000	07:09	Control Test	0.205	07:09	Air Blank	0.000	07:10	Control Test	0.203	07:11	Air Blank	0.000	07:11	Control Test	0.204	07:12	Air Blank	0.000	07:13	Control Test Stats			Average	0.2040		Std Dev	0.0010		Rel Std Dev(%)	0.4902		<p>FL HIGHWAY PATROL Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000784 12/11/2025 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>06:53</td></tr> <tr><td>Control Test</td><td>0.080</td><td>06:53</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>06:53</td></tr> <tr><td>Control Test</td><td>0.079</td><td>06:54</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>06:54</td></tr> <tr><td>Control Test</td><td>0.079</td><td>06:55</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>06:55</td></tr> <tr><td colspan="3">Control Test Stats</td></tr> <tr><td>Average</td><td>0.0793</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7277</td><td></td></tr> </tbody> </table> <p> Operator's Signature</p>	Test	g/210L	Time	Air Blank	0.000	06:53	Control Test	0.080	06:53	Air Blank	0.000	06:53	Control Test	0.079	06:54	Air Blank	0.000	06:54	Control Test	0.079	06:55	Air Blank	0.000	06:55	Control Test Stats			Average	0.0793		Std Dev	0.0006		Rel Std Dev(%)	0.7277	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	06:58																																																																																																																																																	
Control Test	0.051	06:59																																																																																																																																																	
Air Blank	0.000	06:59																																																																																																																																																	
Control Test	0.050	07:00																																																																																																																																																	
Air Blank	0.000	07:01																																																																																																																																																	
Control Test	0.050	07:01																																																																																																																																																	
Air Blank	0.000	07:02																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0503																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	1.1471																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	07:03																																																																																																																																																	
Control Test	0.081	07:04																																																																																																																																																	
Air Blank	0.000	07:04																																																																																																																																																	
Control Test	0.080	07:05																																																																																																																																																	
Air Blank	0.000	07:06																																																																																																																																																	
Control Test	0.080	07:06																																																																																																																																																	
Air Blank	0.000	07:07																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0803																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7187																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	07:09																																																																																																																																																	
Control Test	0.205	07:09																																																																																																																																																	
Air Blank	0.000	07:10																																																																																																																																																	
Control Test	0.203	07:11																																																																																																																																																	
Air Blank	0.000	07:11																																																																																																																																																	
Control Test	0.204	07:12																																																																																																																																																	
Air Blank	0.000	07:13																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.2040																																																																																																																																																		
Std Dev	0.0010																																																																																																																																																		
Rel Std Dev(%)	0.4902																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	06:53																																																																																																																																																	
Control Test	0.080	06:53																																																																																																																																																	
Air Blank	0.000	06:53																																																																																																																																																	
Control Test	0.079	06:54																																																																																																																																																	
Air Blank	0.000	06:54																																																																																																																																																	
Control Test	0.079	06:55																																																																																																																																																	
Air Blank	0.000	06:55																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0793																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7277																																																																																																																																																		

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: FL HIGHWAY PATROL
Time of Inspection: 08:50

Date of Inspection: 12/11/2025

Serial Number: 80-000784
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202406K Exp: 06/19/2026	0.08g/210L Test (g/210L) Lot#:202406L Exp: 06/19/2026	0.20g/210L Test (g/210L) Lot#:202406N Exp: 06/20/2026	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG510701 Exp: 04/17/2027
0.000	0.050	0.080	0.201	0.079
0.000	0.050	0.080	0.202	0.078
0.000	0.049	0.080	0.202	0.078
0.000	0.050	0.080	0.203	0.077
0.000	0.050	0.081	0.202	0.077
0.000	0.050	0.080	0.203	0.077
0.000	0.050	0.081	0.203	0.078
0.000	0.050	0.080	0.202	0.077
0.000	0.050	0.080	0.203	0.077
0.000	0.049	0.080	0.202	0.077
Standard Deviations	0.0004	0.0004	0.0006	0.0007

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0005 Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



KATIE T SPEARIN

Signature and Printed Name

12/11/2025
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000784 , manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-000784</u>	UNCERTAINTY* ±	
Owning Agency:	<u>FL HIGHWAY PATROL</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>12/11/2025</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>08:50</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

12/11/2025

Date

Kaitlyn Spearin

KATIE T SPEARIN,
Department Inspector

Digitally signed by Kaitlyn Spearin
Date: 2025.12.11 10:05:00 -05'00'

Alcohol Testing Program - Instrument Processing Sheet

Agency: FL HIGHWAY PATROL Instrument Serial Number: 80-000784
 Date In: 12/30/2025 DI Completion Date: _____ Ship P/U H/D CMI EE

Intake By: <u>WKP</u> Date: <u>12/30/2025</u> <input type="checkbox"/> Annual <input checked="" type="checkbox"/> Dropped Off <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE <input type="checkbox"/> Training Instrument Visual Inspection <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/Accessories <input type="checkbox"/> Power Cord <input type="checkbox"/> Printer Cable <input type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes:	Quality Checks By: <u>WKP</u> Date: <u>12/30/2025</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value: <u>230</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column #: _____ 32 mm <u>0.160</u> (.139-.169) 36 mm <u>0.179</u> (.156-.190) 53 mm <u>0.242</u> (.228-.278) 103 mm <u>0.496</u> (.447-.547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID #: <u>28427</u> Gauge: <u>1019</u> Instrument: <u>1018</u> <input checked="" type="checkbox"/> Stability Checks	Flow Adjustment By: _____ Flow Column #: _____ <input type="checkbox"/> 5L/min – 17mm <input type="checkbox"/> 15L/min – 53mm <input type="checkbox"/> 30L/min – 103mm <input type="checkbox"/> R-Value: _____ <input type="checkbox"/> Post Adjustment Verification (L/S) Flow Column #: _____ 32 mm _____ (.139-.169) 36 mm _____ (.156-.190) 53 mm _____ (.228-.278) 103 mm _____ (.447-.547)
---	---	---

Simulator	Serial #	Lot#/Exp
0.050	MP5088	202406K
		6/19/2026
0.080	MP5089	202406L
		6/19/2026
0.200	MP5090	202406N
		6/20/2026
0.080 DGS	N/A	AG510701
		4/17/2027

Maintenance By: _____	Date: _____
<input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Tank Sensor Tare <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other:	

Optical Bench Adjustment	By: _____
Barometric Pressure Gauge: _____	ID#: _____
Department Inspection	By: _____
Barometric Pressure ID#: _____	
Gauge: _____	Instrument: _____
Mouth Alcohol Solution Lot #: _____	Exp: _____
Acetone Stock Solution Lot #: _____	Exp: _____
Simulator	Serial Number
0.000	
0.040	
0.100	
0.200	
0.300	
0.080 DGS	N/A
<input type="checkbox"/> Post Optical Bench Adjustment Stability Checks	
Simulator	Serial #
0.050	
0.080	
0.200	
0.080 DGS	N/A
Gauge ID #: _____	Instrument: _____
Gauge: _____ Instrument: _____	

Attachments	
<input type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Stability Checks <input type="checkbox"/> Calibration Certificate <input type="checkbox"/> Optical Bench Adjustment	<input type="checkbox"/> Post-Stability Checks <input type="checkbox"/> Flow Adjustment <input type="checkbox"/> Form 40 <input type="checkbox"/> Other:

Notes/Suggested Service: Instrument dropped off by the agency believing it had returned from repair, however, it had already been previously worked. Quality checks were completed and no further processing was performed since the 2025 department inspection was already completed. WKP 12/30/25

*Tech Correction: Marked boxes at the bottom of the IPS relating to compliance, evidentiary use, and agency inspection. WKP 1/5/26

* **Instrument Complies with Chapter 11D-8, FAC**
 Instrument Does Not Comply with Chapter 11D-8, FAC

* **Return to/Place into Evidentiary Use**
 Remain Out of Evidentiary Use

* **Conduct an Agency Inspection Before Evidentiary Use**

Shayla Platt Digitally signed by Shayla Platt Date: 2026.01.12 10:11:26 -05'00'	Shayla Platt Digitally signed by Shayla Platt Date: 2026.01.12 10:11:41 -05'00'
Tech Review	Admin Review

Stability Checks

0.05g/210L 0.047 to 0.053 ✓	0.08g/210L 0.077 to 0.083 ✓	0.20g/210L 0.194 to 0.206 ✓	DGS 0.08g/210L 0.077 to 0.083 ✓ ≤0.003 of Wet ✓																																																																																																																																																
<p>FL HIGHWAY PATROL Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000784 12/30/2025 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>09:50</td></tr> <tr><td>Control Test</td><td>0.052</td><td>09:50</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:51</td></tr> <tr><td>Control Test</td><td>0.052</td><td>09:52</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:52</td></tr> <tr><td>Control Test</td><td>0.051</td><td>09:53</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:54</td></tr> <tr><td colspan="3">Control Test Stats</td></tr> <tr><td>Average</td><td>0.0517</td><td></td></tr> <tr><td>Std Dev</td><td>0.0016</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.1175</td><td></td></tr> </tbody> </table> <p>Operator's Signature: <i>menpi</i></p>	Test	g/210L	Time	Air Blank	0.000	09:50	Control Test	0.052	09:50	Air Blank	0.000	09:51	Control Test	0.052	09:52	Air Blank	0.000	09:52	Control Test	0.051	09:53	Air Blank	0.000	09:54	Control Test Stats			Average	0.0517		Std Dev	0.0016		Rel Std Dev(%)	1.1175		<p>FL HIGHWAY PATROL Intoxilyzer - Alcohol Analyzer Model: 8000 SN 80-000784 12/30/2025 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>09:55</td></tr> <tr><td>Control Test</td><td>0.082</td><td>09:55</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:56</td></tr> <tr><td>Control Test</td><td>0.082</td><td>09:56</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:57</td></tr> <tr><td>Control Test</td><td>0.082</td><td>09:58</td></tr> <tr><td>Air Blank</td><td>0.001</td><td>09:58</td></tr> <tr><td colspan="3">Control Test Stats</td></tr> <tr><td>Average</td><td>0.0820</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> <p>Operator's Signature: <i>menpi</i></p>	Test	g/210L	Time	Air Blank	0.000	09:55	Control Test	0.082	09:55	Air Blank	0.000	09:56	Control Test	0.082	09:56	Air Blank	0.000	09:57	Control Test	0.082	09:58	Air Blank	0.001	09:58	Control Test Stats			Average	0.0820		Std Dev	0.0000		Rel Std Dev(%)	0.0000		<p>FL HIGHWAY PATROL Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000784 12/30/2025 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>10:00</td></tr> <tr><td>Control Test</td><td>0.205</td><td>10:01</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:01</td></tr> <tr><td>Control Test</td><td>0.205</td><td>10:02</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:03</td></tr> <tr><td>Control Test</td><td>0.204</td><td>10:03</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:04</td></tr> <tr><td colspan="3">Control Test Stats</td></tr> <tr><td>Average</td><td>0.2047</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.2821</td><td></td></tr> </tbody> </table> <p>Operator's Signature: <i>menpi</i></p>	Test	g/210L	Time	Air Blank	0.000	10:00	Control Test	0.205	10:01	Air Blank	0.000	10:01	Control Test	0.205	10:02	Air Blank	0.000	10:03	Control Test	0.204	10:03	Air Blank	0.000	10:04	Control Test Stats			Average	0.2047		Std Dev	0.0006		Rel Std Dev(%)	0.2821		<p>FL HIGHWAY PATROL Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000784 12/30/2025 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>10:07</td></tr> <tr><td>Control Test</td><td>0.080</td><td>10:08</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:08</td></tr> <tr><td>Control Test</td><td>0.081</td><td>10:09</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:09</td></tr> <tr><td>Control Test</td><td>0.081</td><td>10:09</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:10</td></tr> <tr><td colspan="3">Control Test Stats</td></tr> <tr><td>Average</td><td>0.0807</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7157</td><td></td></tr> </tbody> </table> <p>Operator's Signature: <i>menpi</i></p>	Test	g/210L	Time	Air Blank	0.000	10:07	Control Test	0.080	10:08	Air Blank	0.000	10:08	Control Test	0.081	10:09	Air Blank	0.000	10:09	Control Test	0.081	10:09	Air Blank	0.000	10:10	Control Test Stats			Average	0.0807		Std Dev	0.0006		Rel Std Dev(%)	0.7157	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	09:50																																																																																																																																																	
Control Test	0.052	09:50																																																																																																																																																	
Air Blank	0.000	09:51																																																																																																																																																	
Control Test	0.052	09:52																																																																																																																																																	
Air Blank	0.000	09:52																																																																																																																																																	
Control Test	0.051	09:53																																																																																																																																																	
Air Blank	0.000	09:54																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0517																																																																																																																																																		
Std Dev	0.0016																																																																																																																																																		
Rel Std Dev(%)	1.1175																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	09:55																																																																																																																																																	
Control Test	0.082	09:55																																																																																																																																																	
Air Blank	0.000	09:56																																																																																																																																																	
Control Test	0.082	09:56																																																																																																																																																	
Air Blank	0.000	09:57																																																																																																																																																	
Control Test	0.082	09:58																																																																																																																																																	
Air Blank	0.001	09:58																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0820																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	10:00																																																																																																																																																	
Control Test	0.205	10:01																																																																																																																																																	
Air Blank	0.000	10:01																																																																																																																																																	
Control Test	0.205	10:02																																																																																																																																																	
Air Blank	0.000	10:03																																																																																																																																																	
Control Test	0.204	10:03																																																																																																																																																	
Air Blank	0.000	10:04																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.2047																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.2821																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	10:07																																																																																																																																																	
Control Test	0.080	10:08																																																																																																																																																	
Air Blank	0.000	10:08																																																																																																																																																	
Control Test	0.081	10:09																																																																																																																																																	
Air Blank	0.000	10:09																																																																																																																																																	
Control Test	0.081	10:09																																																																																																																																																	
Air Blank	0.000	10:10																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0807																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7157																																																																																																																																																		

Flow Calibration Adjustment(s)

Performed by TDG

MIAMI DADE SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007084
10/01/2025
Software: 8100.27

Flow Rate Calibration*****

1: Rate (Liters/min) = 5

SQRT(Diff) = 6.926

2: Rate (Liters/min) = 15

SQRT(Diff) = 11.957

3: Rate (Liters/min) = 30

SQRT(Diff) = 21.094

Dependent Data Scale Factor = 100000 L/min

Independent Data Scale Factor = 256

Rounded Slope = 683

Rounded Intercept = -664845

Correlation = 0.99874

Stability Checks

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																																																																																																																																				
<p>MIAMI DRAE SO Intoxilyzer - Alcohol Analyzer Model: 8000 10/01/2025 Software: 8100.27</p> <p>SN 80-007084</p> <p>Test g/210L Time</p> <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>15:30</td></tr> <tr><td>Control Test</td><td>0.049</td><td>15:31</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:32</td></tr> <tr><td>Control Test</td><td>0.049</td><td>15:32</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:33</td></tr> <tr><td>Control Test</td><td>0.049</td><td>15:33</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:34</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0490</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </table> <p>Operator's Signature </p>	Air Blank	0.000	15:30	Control Test	0.049	15:31	Air Blank	0.000	15:32	Control Test	0.049	15:32	Air Blank	0.000	15:33	Control Test	0.049	15:33	Air Blank	0.000	15:34	Control Test Stats			Average	0.0490		Std Dev	0.0000		Rel Std Dev(%)	0.0000		<p>MIAMI DRAE SC Intoxilyzer - Alcohol Analyzer Model: 8000 10/01/2025 Software: 8100.27</p> <p>SN 80-007084</p> <p>Test g/210L Time</p> <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>15:37</td></tr> <tr><td>Control Test</td><td>0.080</td><td>15:38</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:38</td></tr> <tr><td>Control Test</td><td>0.078</td><td>15:39</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:40</td></tr> <tr><td>Control Test</td><td>0.079</td><td>15:40</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:41</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0790</td><td></td></tr> <tr><td>Std Dev</td><td>0.0010</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.2658</td><td></td></tr> </table> <p>Operator's Signature </p>	Air Blank	0.000	15:37	Control Test	0.080	15:38	Air Blank	0.000	15:38	Control Test	0.078	15:39	Air Blank	0.000	15:40	Control Test	0.079	15:40	Air Blank	0.000	15:41	Control Test Stats			Average	0.0790		Std Dev	0.0010		Rel Std Dev(%)	1.2658		<p>MIAMI DRAE SO Intoxilyzer - Alcohol Analyzer Model: 8000 10/01/2025 Software: 8100.27</p> <p>SN 80-007084</p> <p>Test g/210L Time</p> <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>15:21</td></tr> <tr><td>Control Test</td><td>0.201</td><td>15:21</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:22</td></tr> <tr><td>Control Test</td><td>0.199</td><td>15:23</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:23</td></tr> <tr><td>Control Test</td><td>0.199</td><td>15:24</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:25</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1997</td><td></td></tr> <tr><td>Std Dev</td><td>0.0012</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.5783</td><td></td></tr> </table> <p>Operator's Signature </p>	Air Blank	0.000	15:21	Control Test	0.201	15:21	Air Blank	0.000	15:22	Control Test	0.199	15:23	Air Blank	0.000	15:23	Control Test	0.199	15:24	Air Blank	0.000	15:25	Control Test Stats			Average	0.1997		Std Dev	0.0012		Rel Std Dev(%)	0.5783		<p>MIAMI DRAE SO Intoxilyzer - Alcohol Analyzer Model: 8000 10/01/2025 Software: 8100.27</p> <p>SN 80-007084</p> <p>Test g/210L Time</p> <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>15:14</td></tr> <tr><td>Control Test</td><td>0.079</td><td>15:14</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:15</td></tr> <tr><td>Control Test</td><td>0.080</td><td>15:15</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:16</td></tr> <tr><td>Control Test</td><td>0.078</td><td>15:16</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:16</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0790</td><td></td></tr> <tr><td>Std Dev</td><td>0.0010</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.2658</td><td></td></tr> </table> <p>Operator's Signature </p>	Air Blank	0.000	15:14	Control Test	0.079	15:14	Air Blank	0.000	15:15	Control Test	0.080	15:15	Air Blank	0.000	15:16	Control Test	0.078	15:16	Air Blank	0.000	15:16	Control Test Stats			Average	0.0790		Std Dev	0.0010		Rel Std Dev(%)	1.2658	
Air Blank	0.000	15:30																																																																																																																																					
Control Test	0.049	15:31																																																																																																																																					
Air Blank	0.000	15:32																																																																																																																																					
Control Test	0.049	15:32																																																																																																																																					
Air Blank	0.000	15:33																																																																																																																																					
Control Test	0.049	15:33																																																																																																																																					
Air Blank	0.000	15:34																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.0490																																																																																																																																						
Std Dev	0.0000																																																																																																																																						
Rel Std Dev(%)	0.0000																																																																																																																																						
Air Blank	0.000	15:37																																																																																																																																					
Control Test	0.080	15:38																																																																																																																																					
Air Blank	0.000	15:38																																																																																																																																					
Control Test	0.078	15:39																																																																																																																																					
Air Blank	0.000	15:40																																																																																																																																					
Control Test	0.079	15:40																																																																																																																																					
Air Blank	0.000	15:41																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.0790																																																																																																																																						
Std Dev	0.0010																																																																																																																																						
Rel Std Dev(%)	1.2658																																																																																																																																						
Air Blank	0.000	15:21																																																																																																																																					
Control Test	0.201	15:21																																																																																																																																					
Air Blank	0.000	15:22																																																																																																																																					
Control Test	0.199	15:23																																																																																																																																					
Air Blank	0.000	15:23																																																																																																																																					
Control Test	0.199	15:24																																																																																																																																					
Air Blank	0.000	15:25																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.1997																																																																																																																																						
Std Dev	0.0012																																																																																																																																						
Rel Std Dev(%)	0.5783																																																																																																																																						
Air Blank	0.000	15:14																																																																																																																																					
Control Test	0.079	15:14																																																																																																																																					
Air Blank	0.000	15:15																																																																																																																																					
Control Test	0.080	15:15																																																																																																																																					
Air Blank	0.000	15:16																																																																																																																																					
Control Test	0.078	15:16																																																																																																																																					
Air Blank	0.000	15:16																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.0790																																																																																																																																						
Std Dev	0.0010																																																																																																																																						
Rel Std Dev(%)	1.2658																																																																																																																																						
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083																																																																																																																																				
✓	✓	✓	✓																																																																																																																																				
✓	✓	✓	✓																																																																																																																																				

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: MIAMI DADE SO
Time of Inspection: 14:42

Date of Inspection: 10/06/2025

Serial Number: 80-007084
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202406K Exp: 06/19/2026	0.08g/210L Test (g/210L) Lot#:202406L Exp: 06/19/2026	0.20g/210L Test (g/210L) Lot#:202406N Exp: 06/20/2026	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG526603 Exp: 09/23/2027
0.000	0.048	0.079	0.200	0.081
0.000	0.049	0.079	0.200	0.081
0.000	0.048	0.079	0.200	0.080
0.000	0.049	0.079	0.200	0.080
0.000	0.048	0.079	0.200	0.081
0.000	0.049	0.078	0.199	0.080
0.000	0.048	0.079	0.200	0.081
0.000	0.049	0.079	0.199	0.080
0.000	0.049	0.079	0.200	0.081
0.000	0.049	0.080	0.200	0.080

Standard Deviations	0.0005	0.0004	0.0004	0.0005
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

10/06/2025
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-007084, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-007084</u>	UNCERTAINTY* \pm	
Owning Agency:	<u>MIAMI DADE SO</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>10/06/2025</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>14:42</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.
Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).
The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.
This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Taylor
Gutschow
Digitally signed by Taylor
Gutschow
Date: 2025.10.07 11:34:55
-04'00'

10/06/2025 Date
TAYLOR D GUTSCHOW,
Department Inspector

FDLE/ATP Form 69 December 2021
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality



INSTRUMENT PROCESSING SHEET

Agency Miami Dade PDS/N 80-007084Florida Department of
Law EnforcementDate In 01/06/2025 DI Completion Date 01/17/2025 Ship P/U H/D CMI EE

Intake By TDG _____ Date <u>01/14/2025</u> <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: <u>AI indicates that during the last inspection, the DGS read 0's and the instrument displayed a DSP Fail.</u>	Quality Checks By TDG _____ Date <u>01/16/2025</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>169</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP101</u> 32 mm <u>0.156</u> (.139 - .169) 36 mm <u>0.179</u> (.156 - .190) 53 mm <u>0.250</u> (.228 - .278) 103 mm <u>0.500</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>26932</u> <input checked="" type="checkbox"/> Stability Checks <table border="1" style="width:100%; border-collapse: collapse; margin-top: 5px;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>MP6286</td> <td>202303K 03/29/2025</td> </tr> <tr> <td>0.080</td> <td>MP6287</td> <td>202303L 03/29/2025</td> </tr> <tr> <td>0.200</td> <td>MP6288</td> <td>202304C 04/05/2025</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG429602 10/22/2026</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.050	MP6286	202303K 03/29/2025	0.080	MP6287	202303L 03/29/2025	0.200	MP6288	202304C 04/05/2025	0.080 DGS	N/A	AG429602 10/22/2026	Flow Calibration By _____ Date _____ Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)
Simulator	Serial #	Lot #/Exp															
0.050	MP6286	202303K 03/29/2025															
0.080	MP6287	202303L 03/29/2025															
0.200	MP6288	202304C 04/05/2025															
0.080 DGS	N/A	AG429602 10/22/2026															
		Maintenance By TDG _____ Date <u>01/16/2025</u> <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input checked="" type="checkbox"/> Other <u>Detached and reattached the gas delivery system connection prior to Quality Checks to ensure a secure connection.</u>															

Calibration Adjustment By _____ Barometric Pressure Gauge _____ ID # _____ <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> <input type="checkbox"/> Post Calibration Adjustment Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>	Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			Department Inspection By TDG _____ Barometric Pressure ID# <u>26932</u> Gauge <u>1021</u> Instrument <u>1022</u> Mouth Alcohol Solution Lot # <u>2024-A</u> Acetone Stock Solution Lot # <u>2023-B</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>MP6284</td> </tr> <tr> <td>Interferent</td> <td>MP6285</td> </tr> <tr> <td>0.050</td> <td>MP6286</td> </tr> <tr> <td>0.080</td> <td>MP6287</td> </tr> <tr> <td>0.200</td> <td>MP6288</td> </tr> </tbody> </table>	Simulator	Serial Number	0.000	MP6284	Interferent	MP6285	0.050	MP6286	0.080	MP6287	0.200	MP6288
Simulator	Serial #	Lot #	Expiration																																																										
0.000		N/A	N/A																																																										
0.040																																																													
0.100																																																													
0.200																																																													
0.300																																																													
0.080 DGS	N/A																																																												
Simulator	Serial #	Lot #	Expiration																																																										
0.050																																																													
0.080																																																													
0.200																																																													
0.080 DGS	N/A																																																												
Simulator	Serial Number																																																												
0.000	MP6284																																																												
Interferent	MP6285																																																												
0.050	MP6286																																																												
0.080	MP6287																																																												
0.200	MP6288																																																												
Attachments <input checked="" type="checkbox"/> Form 41 <input type="checkbox"/> Post-Stability Checks <input checked="" type="checkbox"/> Stability Checks <input type="checkbox"/> Flow Calibration <input checked="" type="checkbox"/> Calibration Certificate <input type="checkbox"/> Form 40 <input type="checkbox"/> Calibration Adjustment <input checked="" type="checkbox"/> Other <u>Extra Stabilities</u>																																																													

Notes/Suggested Service: <u>Could not replicate DSP Fail or 0 reading on DGS. The Agency Inspector will notify the Department Inspector if either recurs. (TDG 01/17/2025)</u>	<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use
Destinee Armstrong <small>Digitally signed by Destinee Armstrong Date: 2025.01.24 07:41:13 -0500</small>	Phil Nicodemo <small>Digitally signed by Phil Nicodemo Date: 2025.01.24 12:24:09 -05'00'</small>
Tech Review / Date _____	Admin Review / Date _____

Stability Checks

0.05g/210L 0.047 to 0.053	0.08g/210L 0.077 to 0.083	0.20g/210L 0.194 to 0.206	DGS 0.08g/210L 0.077 to 0.083																																																																																																																																																
<p>MIAMI DADE PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-007084 01/16/2025 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>13:16</td></tr> <tr><td>Control Test</td><td>0.051</td><td>13:17</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:17</td></tr> <tr><td>Control Test</td><td>0.051</td><td>13:18</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:19</td></tr> <tr><td>Control Test</td><td>0.050</td><td>13:19</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:20</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0507</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.1355</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	13:16	Control Test	0.051	13:17	Air Blank	0.000	13:17	Control Test	0.051	13:18	Air Blank	0.000	13:19	Control Test	0.050	13:19	Air Blank	0.000	13:20	Control Test Stats			Average	0.0507		Std Dev	0.0006		Rel Std Dev(%)	1.1355		<p>MIAMI DADE PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-007084 01/16/2025 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>13:09</td></tr> <tr><td>Control Test</td><td>0.081</td><td>13:10</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:10</td></tr> <tr><td>Control Test</td><td>0.080</td><td>13:11</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:12</td></tr> <tr><td>Control Test</td><td>0.080</td><td>13:12</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:13</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0803</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7187</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	13:09	Control Test	0.081	13:10	Air Blank	0.000	13:10	Control Test	0.080	13:11	Air Blank	0.000	13:12	Control Test	0.080	13:12	Air Blank	0.000	13:13	Control Test Stats			Average	0.0803		Std Dev	0.0006		Rel Std Dev(%)	0.7187		<p>MIAMI DADE PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-007084 01/16/2025 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>13:03</td></tr> <tr><td>Control Test</td><td>0.198</td><td>13:03</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:04</td></tr> <tr><td>Control Test</td><td>0.197</td><td>13:05</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:05</td></tr> <tr><td>Control Test</td><td>0.196</td><td>13:06</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:07</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1970</td><td></td></tr> <tr><td>Std Dev</td><td>0.0010</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.5076</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	13:03	Control Test	0.198	13:03	Air Blank	0.000	13:04	Control Test	0.197	13:05	Air Blank	0.000	13:05	Control Test	0.196	13:06	Air Blank	0.000	13:07	Control Test Stats			Average	0.1970		Std Dev	0.0010		Rel Std Dev(%)	0.5076		<p>MIAMI DADE PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-007084 01/16/2025 Software: 8100.27</p> <p style="text-align: right; font-size: 2em; color: blue;">DGS</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>12:57</td></tr> <tr><td>Control Test</td><td>0.081</td><td>12:57</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:58</td></tr> <tr><td>Control Test</td><td>0.080</td><td>12:58</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:59</td></tr> <tr><td>Control Test</td><td>0.080</td><td>12:59</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:59</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0803</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7187</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	12:57	Control Test	0.081	12:57	Air Blank	0.000	12:58	Control Test	0.080	12:58	Air Blank	0.000	12:59	Control Test	0.080	12:59	Air Blank	0.000	12:59	Control Test Stats			Average	0.0803		Std Dev	0.0006		Rel Std Dev(%)	0.7187	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	13:16																																																																																																																																																	
Control Test	0.051	13:17																																																																																																																																																	
Air Blank	0.000	13:17																																																																																																																																																	
Control Test	0.051	13:18																																																																																																																																																	
Air Blank	0.000	13:19																																																																																																																																																	
Control Test	0.050	13:19																																																																																																																																																	
Air Blank	0.000	13:20																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0507																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	1.1355																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	13:09																																																																																																																																																	
Control Test	0.081	13:10																																																																																																																																																	
Air Blank	0.000	13:10																																																																																																																																																	
Control Test	0.080	13:11																																																																																																																																																	
Air Blank	0.000	13:12																																																																																																																																																	
Control Test	0.080	13:12																																																																																																																																																	
Air Blank	0.000	13:13																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0803																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7187																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	13:03																																																																																																																																																	
Control Test	0.198	13:03																																																																																																																																																	
Air Blank	0.000	13:04																																																																																																																																																	
Control Test	0.197	13:05																																																																																																																																																	
Air Blank	0.000	13:05																																																																																																																																																	
Control Test	0.196	13:06																																																																																																																																																	
Air Blank	0.000	13:07																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1970																																																																																																																																																		
Std Dev	0.0010																																																																																																																																																		
Rel Std Dev(%)	0.5076																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	12:57																																																																																																																																																	
Control Test	0.081	12:57																																																																																																																																																	
Air Blank	0.000	12:58																																																																																																																																																	
Control Test	0.080	12:58																																																																																																																																																	
Air Blank	0.000	12:59																																																																																																																																																	
Control Test	0.080	12:59																																																																																																																																																	
Air Blank	0.000	12:59																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0803																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7187																																																																																																																																																		
<p style="text-align: right; font-size: 2em; color: blue;">ML</p> <p>Operator's Signature</p>	<p style="text-align: right; font-size: 2em; color: blue;">ML</p> <p>Operator's Signature</p>	<p style="text-align: right; font-size: 2em; color: blue;">ML</p> <p>Operator's Signature</p>	<p style="text-align: right; font-size: 2em; color: blue;">ML</p> <p>Operator's Signature</p>																																																																																																																																																

NLS

MIAMI DADE PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007084
01/16/2025
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:39
Control Test	0.081	13:39
Air Blank	0.000	13:40
Control Test	0.082	13:40
Air Blank	0.000	13:41
Control Test	0.080	13:41
Air Blank	0.000	13:41
Control Test	0.081	13:42
Air Blank	0.000	13:42
Control Test	0.081	13:43
Air Blank	0.000	13:43
Control Test	0.082	13:44
Air Blank	0.000	13:44
Control Test	0.081	13:44
Air Blank	0.000	13:45
Control Test	0.080	13:45
Air Blank	0.000	13:46
Control Test	0.081	13:46
Air Blank	0.000	13:47
Control Test	0.080	13:47
Air Blank	0.000	13:47
Control Test	0.081	13:48
Air Blank	0.000	13:48
Control Test	0.081	13:49
Air Blank	0.000	13:49
Control Test	0.081	13:50
Air Blank	0.000	13:50
Control Test	0.081	13:50
Air Blank	0.000	13:51
Control Test	0.081	13:51
Air Blank	0.000	13:52
Control Test	0.081	13:52
Air Blank	0.000	13:53
Control Test	0.081	13:53
Air Blank	0.000	13:53
Control Test	0.082	13:54
Air Blank	0.000	13:54
Control Test	0.081	13:55
Air Blank	0.000	13:55
Control Test	0.081	13:56
Air Blank	0.000	13:56
Control Test Stats		
Average	0.0810	
Std Dev	0.0006	
Rel Std Dev(%)	0.6938	

MG

Operator's Signature

Extra Stabilities

ML

Conducted extra NLS stability checks on different days.

ML

1/17/25

NLS

MIAMI DADE PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007084
01/17/2025
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:31
Control Test	0.081	09:31
Air Blank	0.000	09:32
Control Test	0.081	09:32
Air Blank	0.000	09:33
Control Test	0.081	09:33
Air Blank	0.000	09:34
Control Test	0.080	09:34
Air Blank	0.000	09:34
Control Test	0.081	09:35
Air Blank	0.000	09:35
Control Test	0.081	09:36
Air Blank	0.000	09:36
Control Test	0.080	09:37
Air Blank	0.000	09:37
Control Test	0.081	09:37
Air Blank	0.000	09:38
Control Test	0.081	09:38
Air Blank	0.000	09:39
Control Test	0.081	09:39
Air Blank	0.000	09:40
Control Test	0.081	09:40
Air Blank	0.000	09:41
Control Test	0.081	09:41
Air Blank	0.000	09:41
Control Test	0.082	09:42
Air Blank	0.000	09:42
Control Test	0.081	09:43
Air Blank	0.000	09:43
Control Test	0.082	09:43
Air Blank	0.000	09:44
Control Test	0.081	09:44
Air Blank	0.000	09:45
Control Test	0.081	09:45
Air Blank	0.000	09:46
Control Test	0.082	09:46
Air Blank	0.000	09:47
Control Test	0.081	09:47
Air Blank	0.000	09:47
Control Test	0.081	09:48
Air Blank	0.000	09:48
Control Test Stats		
Average	0.0810	
Std Dev	0.0005	
Rel Std Dev(%)	0.6298	

ML

Operator's Signature

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: MIAMI DADE PD
Time of Inspection: 11:33

Date of Inspection: 01/17/2025

Serial Number: 80-007084
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202303K Exp: 03/29/2025	0.08g/210L Test (g/210L) Lot#:202303L Exp: 03/29/2025	0.20g/210L Test (g/210L) Lot#:202304C Exp: 04/05/2025	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG429602 Exp: 10/22/2026
0.000	0.051	0.080	0.198	0.080
0.000	0.050	0.079	0.198	0.080
0.000	0.050	0.080	0.198	0.080
0.000	0.050	0.079	0.198	0.080
0.000	0.051	0.079	0.199	0.080
0.000	0.051	0.079	0.199	0.080
0.000	0.051	0.079	0.199	0.081
0.000	0.051	0.079	0.199	0.080
0.000	0.051	0.080	0.198	0.080
0.000	0.051	0.080	0.198	0.080

Standard Deviations	0.0004	0.0005	0.0005	0.0003
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5


Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.


 _____ TAYLOR D GUTSCHOW
 Signature and Printed Name

01/17/2025
 Date



Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

Calibration Certificate

This is to certify the calibration of Intoxilyzer 8000 serial number 80-007084, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-007084</u>	UNCERTAINTY* ±	
Owning Agency:	<u>MIAMI DADE PD</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>01/17/2025</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>11:33</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

01/17/2025

Date

TAYLOR D GUTSCHOW,
Department Inspector

FDLE/ATP Form 69 December 2021

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality