

Alcohol Testing Program - Instrument Processing Sheet

Agency: Sarasota County SO

Instrument Serial Number: 80-001348


Date In: 4/14/2026

DI Completion Date: 4/22/26

☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake By: <u>IAS</u> Date: <u>4/14/2026</u>		Quality Checks By: <u>KTS</u> Date: <u>4/16/2026</u>		Flow Adjustment By: _____																																									
☒ Annual ☐ Dropped Off ☐ Registration ☐ Return from CMI / EE ☐ Training Instrument Visual Inspection ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/Accessories ☐ Power Cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes: Missing back right foot. IAS 4/14/26		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value: <u>211</u> ☒ Flow Verification (L/s) Flow Column #: <u>ATP105</u> 32 mm <u>0.160</u> (.139-.169) 36 mm <u>0.171</u> (.156-.190) 53 mm <u>0.242</u> (.228-.278) 103 mm <u>0.539</u> (.447-.547) ☒ Barometric Pressure Check Gauge ID #: <u>28427</u> Gauge: <u>1016</u> Instrument: <u>1012</u> ☒ Stability Checks		Flow Column #: _____ ☐ 5L/min – 17mm ☐ 15L/min – 53mm ☐ 30L/min – 103mm ☐ R-Value: _____ ☐ Post Adjustment Verification (L/S) Flow Column #: _____ 32 mm _____ (.139-.169) 36 mm _____ (.156-.190) 53 mm _____ (.228-.278) 103 mm _____ (.447-.547)																																									
						<table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot#/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>MP6291</td> <td>202406K 6/19/2026</td> </tr> <tr> <td>0.080</td> <td>MP6292</td> <td>202603B 3/3/2028</td> </tr> <tr> <td>0.200</td> <td>MP6293</td> <td>202406N 6/20/2026</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>23825080A1 10/5/2027</td> </tr> </tbody> </table>		Simulator	Serial #	Lot#/Exp	0.050	MP6291	202406K 6/19/2026	0.080	MP6292	202603B 3/3/2028	0.200	MP6293	202406N 6/20/2026	0.080 DGS	N/A	23825080A1 10/5/2027																							
								Simulator	Serial #	Lot#/Exp																																			
								0.050	MP6291	202406K 6/19/2026																																			
								0.080	MP6292	202603B 3/3/2028																																			
0.200	MP6293	202406N 6/20/2026																																											
0.080 DGS	N/A	23825080A1 10/5/2027																																											
Maintenance By: KTS Date: 4/14/2026 ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Tank Sensor Tare ☐ Breath Tube Replacement ☒ Other: Replaced back right foot. KTS 4/14/26																																													
Optical Bench Adjustment By: <u>WKP</u>		Department Inspection By: <u>KTS</u>																																											
Barometric Pressure Gauge: <u>1015</u> ID#: <u>28662</u>		Barometric Pressure ID#: <u>28427</u>																																											
<table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>MP6294</td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td>MP6295</td> <td>25090</td> <td>03/11/2027</td> </tr> <tr> <td>0.100</td> <td>MP6296</td> <td>25282</td> <td>08/12/2027</td> </tr> <tr> <td>0.200</td> <td>MP6297</td> <td>25020</td> <td>1/14/2027</td> </tr> <tr> <td>0.300</td> <td>MP6298</td> <td>25150</td> <td>05/06/2027</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG510701</td> <td>04/17/2027</td> </tr> </tbody> </table>		Simulator	Serial #	Lot #	Expiration	0.000	MP6294	N/A	N/A	0.040	MP6295	25090	03/11/2027	0.100	MP6296	25282	08/12/2027	0.200	MP6297	25020	1/14/2027	0.300	MP6298	25150	05/06/2027	0.080 DGS	N/A	AG510701	04/17/2027	Gauge: <u>1018</u> Instrument: <u>1018</u> Mouth Alcohol Solution Lot #: <u>2025-D</u> Exp: <u>9/25/27</u> Acetone Stock Solution Lot #: <u>2025-B</u> Exp: <u>9/22/27</u>		<table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>MP6289</td> </tr> <tr> <td>Interferent</td> <td>MP6290</td> </tr> <tr> <td>0.050</td> <td>MP6291</td> </tr> <tr> <td>0.080</td> <td>MP6292</td> </tr> <tr> <td>0.200</td> <td>MP6293</td> </tr> </tbody> </table>		Simulator	Serial Number	0.000	MP6289	Interferent	MP6290	0.050	MP6291	0.080	MP6292	0.200	MP6293
Simulator	Serial #	Lot #	Expiration																																										
0.000	MP6294	N/A	N/A																																										
0.040	MP6295	25090	03/11/2027																																										
0.100	MP6296	25282	08/12/2027																																										
0.200	MP6297	25020	1/14/2027																																										
0.300	MP6298	25150	05/06/2027																																										
0.080 DGS	N/A	AG510701	04/17/2027																																										
Simulator	Serial Number																																												
0.000	MP6289																																												
Interferent	MP6290																																												
0.050	MP6291																																												
0.080	MP6292																																												
0.200	MP6293																																												
☒ Post Optical Bench Adjustment Stability Checks		Attachments																																											
<table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>MP6291</td> <td>202406K</td> <td>06/19/2026</td> </tr> <tr> <td>0.080</td> <td>MP6292</td> <td>202603B</td> <td>03/03/2028</td> </tr> <tr> <td>0.200</td> <td>MP6293</td> <td>202406N</td> <td>06/20/2026</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>23825080A1</td> <td>10/05/2027</td> </tr> </tbody> </table>		Simulator	Serial #	Lot #	Expiration	0.050	MP6291	202406K	06/19/2026	0.080	MP6292	202603B	03/03/2028	0.200	MP6293	202406N	06/20/2026	0.080 DGS	N/A	23825080A1	10/05/2027	☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☒ Optical Bench Adjustment		☒ Post-Stability Checks ☐ Flow Adjustment ☐ Form 40 ☐ Other:																					
Simulator	Serial #	Lot #	Expiration																																										
0.050	MP6291	202406K	06/19/2026																																										
0.080	MP6292	202603B	03/03/2028																																										
0.200	MP6293	202406N	06/20/2026																																										
0.080 DGS	N/A	23825080A1	10/05/2027																																										
Gauge ID #: <u>28427</u> Gauge: <u>1013</u> Instrument: <u>1014</u>		☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use																																											
Notes/Suggested Service:		Digitally signed by Shayla Platt Date: 2026.04.23 10:42:21 -04'00'																																											
		Digitally signed by Wen-Chi Pierson Date: 2026.04.30 13:48:58 -04'00'																																											
		Tech Review		Admin Review																																									

Stability Checks

0.050 g/210L 0.047 to 0.053 g/210L	0.080 g/210L 0.077 to 0.083 g/210L	0.200 g/210L 0.194 to 0.206 g/210L	DGS 0.080 g/210L 0.077 to 0.083 g/210L ≤0.003 g/210L of Wet
✓ SARASOTA COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001348 04/16/2026 Software: 8100.27 Test g/210L Time Air Blank 0.000 09:08 Control Test 0.049 09:09 Air Blank 0.000 09:09 Control Test 0.050 09:10 Air Blank 0.000 09:10 Control Test 0.048 09:11 Air Blank 0.000 09:12 Control Test Stats Average 0.0490 Std Dev 1.0010 Rel Std Dev(%) 2.0408 Operator's Signature 	✓ SARASOTA COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001348 04/16/2026 Software: 8100.27 Test g/210L Time Air Blank 0.000 09:15 Control Test 0.079 09:15 Air Blank 0.000 09:16 Control Test 0.078 09:16 Air Blank 0.000 09:17 Control Test 0.079 09:18 Air Blank 0.000 09:18 Control Test Stats Average 0.0787 Std Dev 0.0006 Rel Std Dev(%) 0.7339 Operator's Signature 	✓ SARASOTA COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001348 04/16/2026 Software: 8100.27 Test g/210L Time Air Blank 0.000 09:19 Control Test 0.199 09:20 Air Blank 0.000 09:21 Control Test 0.200 09:21 Air Blank 0.000 09:22 Control Test 0.199 09:22 Air Blank 0.000 09:23 Control Test Stats Average 0.1993 Std Dev 0.0006 Rel Std Dev(%) 0.2896 Operator's Signature 	✓ SARASOTA COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001348 04/16/2026 Software: 8100.27 Test g/210L Time Air Blank 0.000 09:00 Control Test 0.082 09:00 Air Blank 0.000 09:01 Control Test 0.082 09:01 Air Blank 0.000 09:01 Control Test 0.082 09:02 Air Blank 0.000 09:02 Control Test Stats Average 0.0820 Std Dev 0.0000 Rel Std Dev(%) 0.0000 Operator's Signature 

INTOXILYZER 8000
Instrument Initialization
09:26 04/17/2026

Instrument power chord inadvertently
got unplugged when moving the
instrument on lab bench to create
more space during the optical bench
calibration adjustment. #1

wkp 4/17/2026

SARASOTA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 60-001348
04/17/2026 11:42:47

Auto Calibration
Max Power Res Value = 56
Auto Range Res Value = 25

INTOXILYZER 8000
Instrument Initialization
09:26 04/17/2026

SAFESOTA COUNTY SC
Intoxilizer - Alcohol analyzer
Model 8000
11/17/2026
SN 80-001348
11:57:05
Auto Calibration
Max Power Res Value = 55
Auto Range Res Value = 25

***** AUTO CAL DATA *****
***** CHANNEL 1 *****
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.144
Std Dev = 0.06 Rel Std Dev = 41.53
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 0.839
Std Dev = 0.03 Rel Std Dev = 3.36
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 1.899
Std Dev = 0.02 Rel Std Dev = 1.14
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 3.695
Std Dev = 0.01 Rel Std Dev = 0.18
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 5.404
Std Dev = 0.01 Rel Std Dev = 0.15
Zero Order Coef = -350.35
First Order Coef = 2650.97
Second Order Coef = 9.92
Standard Deviation = 41.619289
***** CHANNEL 2 *****
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.127
Std Dev = 0.02 Rel Std Dev = 16.30
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 1.535
Std Dev = 0.03 Rel Std Dev = 1.59
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 3.617
Std Dev = 0.02 Rel Std Dev = 0.45
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 6.985
Std Dev = 0.00 Rel Std Dev = 0.04
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 10.167
Std Dev = 0.02 Rel Std Dev = 0.20
Zero Order Coef = -155.99
First Order Coef = 1320.67
Second Order Coef = 9.73
Standard Deviation = 14.655045

***** CHANNEL 2 *****
Sample % Abs (% Abs Ref)
Sample #1 = 7.0490 (-0.0030)
Sample #2 = 6.9840 (-0.0540)
Sample #3 = 6.9880 (-0.0480)
Sample #4 = 6.9820 (-0.0410)
Avg % Abs = 6.9847 (-0.0477)
STD DEV = 0.0031 (-0.0065)
REL STD DEV = 0.044 (-13.650)

Sol Value = 0.300 g/210L ***
Fit Value = 1.4286 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12507, Sum Io = 13545
***** CHANNEL 1 *****
Sample % Abs (% Abs Ref)
Sample #1 = 5.4200 (-0.0220)
Sample #2 = 5.4130 (-0.0150)
Sample #3 = 5.4010 (-0.0270)
Sample #4 = 5.3980 (-0.0220)
Avg % Abs = 5.4040 (-0.0213)
STD DEV = 0.0079 (-0.0160)
REL STD DEV = 0.147 (-28.255)

Sol Value = 0.100 g/210L ***
Fit Value = 0.4762 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12503, Sum Io = 13542
***** CHANNEL 1 *****
Sample % Abs (% Abs Ref)
Sample #1 = 1.9600 (-0.0470)
Sample #2 = 1.9130 (-0.0260)
Sample #3 = 1.9100 (-0.0300)
Sample #4 = 1.8740 (-0.0150)
Avg % Abs = 1.8990 (-0.0063)
STD DEV = 0.0217 (-0.0290)
REL STD DEV = 1.143 (-457.713)

***** CHANNEL 2 *****
Sample % Abs (% Abs Ref)
Sample #1 = 1.5880 (-0.0030)
Sample #2 = 1.5710 (-0.0300)
Sample #3 = 1.5120 (-0.0620)
Sample #4 = 1.5370 (-0.0590)
Avg % Abs = 1.5400 (-0.0503)
STD DEV = 0.0296 (-0.0177)
REL STD DEV = 1.923 (-35.112)

Sol Value = 0.040 g/210L ***
Fit Value = 0.1905 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12502, Sum Io = 13536
***** CHANNEL 1 *****
Sample % Abs (% Abs Ref)
Sample #1 = 0.8370 (-0.0010)
Sample #2 = 0.8850 (-0.0380)
Sample #3 = 0.8810 (-0.0190)
Sample #4 = 0.7870 (-0.0380)
Avg % Abs = 0.8243 (-0.0063)
STD DEV = 0.0530 (-0.0356)
REL STD DEV = 6.430 (-624.500)

Sol Value = 0.200 g/210L ***
Fit Value = 0.9524 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12499, Sum Io = 13539
***** CHANNEL 1 *****
Sample % Abs (% Abs Ref)
Sample #1 = 0.8360 (-0.0280)
Sample #2 = 0.8260 (-0.0640)
Sample #3 = 0.8190 (-0.0250)
Sample #4 = 0.6710 (-0.0610)
Avg % Abs = 0.8387 (-0.0500)
STD DEV = 0.0282 (-0.0277)
REL STD DEV = 3.365 (-43.405)

Sol Value = 0.140 g/210L ***
Fit Value = 0.1905 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12522, Sum Io = 13549
***** CHANNEL 1 *****
Sample % Abs (% Abs Ref)
Sample #1 = 0.9030 (-0.0040)
Sample #2 = 0.8770 (-0.0490)
Sample #3 = 0.8130 (-0.1050)
Sample #4 = 0.8010 (-0.1250)
Avg % Abs = 0.8303 (-0.0930)
STD DEV = 0.0409 (-0.0394)
REL STD DEV = 4.921 (-42.361)

***** CHANNEL 2 *****
Sample % Abs (% Abs Ref)
Sample #1 = 10.1710 (-0.0010)
Sample #2 = 10.1910 (-0.0120)
Sample #3 = 10.1520 (-0.0210)
Sample #4 = 10.1590 (-0.0210)
Avg % Abs = 10.1673 (-0.0180)
STD DEV = 0.0208 (-0.0052)
REL STD DEV = 0.205 (-28.868)

Sol Value = 0.200 g/210L ***
Fit Value = 0.9524 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12507, Sum Io = 13548
***** CHANNEL 1 *****
Sample % Abs (% Abs Ref)
Sample #1 = 3.7140 (-0.0050)
Sample #2 = 3.7000 (-0.0220)
Sample #3 = 3.6870 (-0.0380)
Sample #4 = 3.6970 (-0.0300)
Avg % Abs = 3.6947 (-0.0300)
STD DEV = 0.0068 (-0.0080)
REL STD DEV = 0.184 (-26.667)

***** CHANNEL 2 *****
Sample % Abs (% Abs Ref)
Sample #1 = 10.1710 (-0.0010)
Sample #2 = 10.1910 (-0.0120)
Sample #3 = 10.1520 (-0.0210)
Sample #4 = 10.1590 (-0.0210)
Avg % Abs = 10.1673 (-0.0180)
STD DEV = 0.0208 (-0.0052)
REL STD DEV = 0.205 (-28.868)

Sol Value = 0.200 g/210L ***
Fit Value = 0.9524 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12507, Sum Io = 13548
***** CHANNEL 1 *****
Sample % Abs (% Abs Ref)
Sample #1 = 3.7140 (-0.0050)
Sample #2 = 3.7000 (-0.0220)
Sample #3 = 3.6870 (-0.0380)
Sample #4 = 3.6970 (-0.0300)
Avg % Abs = 3.6947 (-0.0300)
STD DEV = 0.0068 (-0.0080)
REL STD DEV = 0.184 (-26.667)

Optical Bench
calibration
adjustment #2
CWP
4/17/2026

Solution Stats Quadratic Fit Chan 1
Act Fit Residual
g/210L g/210L g/210L
1.000 0.001 -0.0007
1.040 0.039 0.0005
1.100 0.059 0.0009
0.200 0.201 -0.0012
0.300 0.300 0.0004

optical calibration adjustment

berken #2

04/17/2020

80-001348

Solution Status Quadratic Fit Chan 2			
Act	Fit	Residual	
g/210L	g/210L	g/210L	
0.000	0.000	-0.0002	
0.140	0.040	0.0002	
0.100	0.100	0.0003	
0.200	0.200	-0.0004	
0.300	0.300	0.0002	

Sol Value = 0.080 g/210L ***
Fit Value = 0.3811 mg/l %%%
Samples Taken = 4, Discarded = 1

***** CHANNEL 1

Sample #1 = 3248.00
Sample #2 = 3213.00
Sample #3 = 3208.00
Sample #4 = 3229.00

Average Result = 3216.6667

STD DEV = 10.9697

REL STD DEV = 0.341

***** CHANNEL 2

Sample #1 = 3508.00
Sample #2 = 3535.00
Sample #3 = 3528.00
Sample #4 = 3537.00

Average Result = 3533.3333

STD DEV = 4.7258

REL STD DEV = 0.134

Dry Gas H2O Adjust Results *****

Barometric Pressure = 1013

3 um H2O Adjust (mg/l * 10,000) = 593

9 um H2O Adjust (mg/l * 10,000) = 276

***** AUTO CAL PASS

Post-Cal Stability Checks

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																																																																																																																																				
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083																																																																																																																																				
SARASOTA COUNTY SO Intoxilizer - Alconol Analyzer Model 8000 SN 80-001348 04/17/2026 Software: 8100.27 Test: g/210L Time <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>13:15</td></tr> <tr><td>Control Test</td><td>0.049</td><td>13:15</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:07</td></tr> <tr><td>Control Test</td><td>0.149</td><td>13:08</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:08</td></tr> <tr><td>Control Test</td><td>0.149</td><td>13:09</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:09</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.149</td><td></td></tr> <tr><td>Std Dev</td><td>0.000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </table> Operator's Signature: <i>menpi</i>	Air Blank	0.000	13:15	Control Test	0.049	13:15	Air Blank	0.000	13:07	Control Test	0.149	13:08	Air Blank	0.000	13:08	Control Test	0.149	13:09	Air Blank	0.000	13:09	Control Test Stats			Average	0.149		Std Dev	0.000		Rel Std Dev(%)	0.0000		SARASOTA COUNTY SO Intoxilizer - Alconol Analyzer Model 8000 SN 80-001348 04/17/2026 Software: 8100.27 Test: g/210L Time <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>13:18</td></tr> <tr><td>Control Test</td><td>0.080</td><td>13:19</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:19</td></tr> <tr><td>Control Test</td><td>0.080</td><td>13:20</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:20</td></tr> <tr><td>Control Test</td><td>0.080</td><td>13:21</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:22</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0800</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </table> Operator's Signature: <i>menpi</i>	Air Blank	0.000	13:18	Control Test	0.080	13:19	Air Blank	0.000	13:19	Control Test	0.080	13:20	Air Blank	0.000	13:20	Control Test	0.080	13:21	Air Blank	0.000	13:22	Control Test Stats			Average	0.0800		Std Dev	0.0000		Rel Std Dev(%)	0.0000		SARASOTA COUNTY SO Intoxilizer - Alconol Analyzer Model 8000 SN 80-001348 04/17/2026 Software: 8100.27 Test: g/210L Time <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>13:12</td></tr> <tr><td>Control Test</td><td>0.198</td><td>13:13</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:13</td></tr> <tr><td>Control Test</td><td>0.199</td><td>13:14</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:14</td></tr> <tr><td>Control Test</td><td>0.198</td><td>13:15</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:16</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1983</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.2911</td><td></td></tr> </table> Operator's Signature: <i>menpi</i>	Air Blank	0.000	13:12	Control Test	0.198	13:13	Air Blank	0.000	13:13	Control Test	0.199	13:14	Air Blank	0.000	13:14	Control Test	0.198	13:15	Air Blank	0.000	13:16	Control Test Stats			Average	0.1983		Std Dev	0.0006		Rel Std Dev(%)	0.2911		SARASOTA COUNTY SO Intoxilizer - Alconol Analyzer Model 8000 SN 80-001348 04/17/2026 Software: 8100.27 Test: g/210L Time <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>13:24</td></tr> <tr><td>Control Test</td><td>0.079</td><td>13:24</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:25</td></tr> <tr><td>Control Test</td><td>0.079</td><td>13:25</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:26</td></tr> <tr><td>Control Test</td><td>0.080</td><td>13:26</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:27</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0793</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7277</td><td></td></tr> </table> Operator's Signature: <i>menpi</i>	Air Blank	0.000	13:24	Control Test	0.079	13:24	Air Blank	0.000	13:25	Control Test	0.079	13:25	Air Blank	0.000	13:26	Control Test	0.080	13:26	Air Blank	0.000	13:27	Control Test Stats			Average	0.0793		Std Dev	0.0006		Rel Std Dev(%)	0.7277	
Air Blank	0.000	13:15																																																																																																																																					
Control Test	0.049	13:15																																																																																																																																					
Air Blank	0.000	13:07																																																																																																																																					
Control Test	0.149	13:08																																																																																																																																					
Air Blank	0.000	13:08																																																																																																																																					
Control Test	0.149	13:09																																																																																																																																					
Air Blank	0.000	13:09																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.149																																																																																																																																						
Std Dev	0.000																																																																																																																																						
Rel Std Dev(%)	0.0000																																																																																																																																						
Air Blank	0.000	13:18																																																																																																																																					
Control Test	0.080	13:19																																																																																																																																					
Air Blank	0.000	13:19																																																																																																																																					
Control Test	0.080	13:20																																																																																																																																					
Air Blank	0.000	13:20																																																																																																																																					
Control Test	0.080	13:21																																																																																																																																					
Air Blank	0.000	13:22																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.0800																																																																																																																																						
Std Dev	0.0000																																																																																																																																						
Rel Std Dev(%)	0.0000																																																																																																																																						
Air Blank	0.000	13:12																																																																																																																																					
Control Test	0.198	13:13																																																																																																																																					
Air Blank	0.000	13:13																																																																																																																																					
Control Test	0.199	13:14																																																																																																																																					
Air Blank	0.000	13:14																																																																																																																																					
Control Test	0.198	13:15																																																																																																																																					
Air Blank	0.000	13:16																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.1983																																																																																																																																						
Std Dev	0.0006																																																																																																																																						
Rel Std Dev(%)	0.2911																																																																																																																																						
Air Blank	0.000	13:24																																																																																																																																					
Control Test	0.079	13:24																																																																																																																																					
Air Blank	0.000	13:25																																																																																																																																					
Control Test	0.079	13:25																																																																																																																																					
Air Blank	0.000	13:26																																																																																																																																					
Control Test	0.080	13:26																																																																																																																																					
Air Blank	0.000	13:27																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.0793																																																																																																																																						
Std Dev	0.0006																																																																																																																																						
Rel Std Dev(%)	0.7277																																																																																																																																						

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: SARASOTA COUNTY SO
Time of Inspection: 15:31

Date of Inspection: 04/22/2026

Serial Number: 80-001348
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202406K Exp: 06/19/2026	0.08g/210L Test (g/210L) Lot#:202603B Exp: 03/03/2028	0.20g/210L Test (g/210L) Lot#:202406N Exp: 06/20/2026	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:23825080A1 Exp: 10/05/2027
0.000	0.050	0.080	0.201	0.080
0.000	0.050	0.081	0.201	0.080
0.000	0.050	0.081	0.201	0.079
0.000	0.050	0.080	0.201	0.079
0.000	0.050	0.080	0.201	0.079
0.000	0.050	0.080	0.201	0.080
0.000	0.050	0.080	0.201	0.079
0.000	0.050	0.080	0.202	0.080
0.000	0.050	0.080	0.202	0.079
0.000	0.050	0.080	0.201	0.079

Standard Deviations	0.0000	0.0004	0.0004	0.0005
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



KATIE T SPEARIN

Signature and Printed Name

04/22/2026
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001348, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001348</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>SARASOTA COUNTY SO</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>04/22/2026</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>15:31</u>	0.200 g/ 210 L	0.008
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

04/22/2026

Date


KATIE T SPEARIN,
Department Inspector

FDLE/ATP Form 69 January 2026
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality