



INSTRUMENT PROCESSING SHEET

Agency ORMOND BEACH PDS/N 80-001156Florida Department of
Law EnforcementDate In 9/17/2025DI Completion Date 10/2/2025☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake By <u>WKP</u> Date <u>9/17/2025</u>	Quality Checks By <u>WKP</u> Date <u>09/25/2025</u>	Flow Calibration By <u>SLH</u> Date <u>9/30/2025</u>																																															
☒ Annual ☐ Registration ☒ Return from CMI / EE Visual Inspection: <table border="0"><tr><td>☒ Case</td><td>☒ Handle</td></tr><tr><td>☒ Keyboard</td><td>☒ Dry Gas Shelf</td></tr><tr><td>☒ Feet</td><td>☒ Breath Tube</td></tr><tr><td>☒ Ports</td><td>☒ Screws Tight</td></tr></table> Other Equipment/ Accessories: <table border="0"><tr><td>☐ Power cord</td><td>☐ Printer Cable</td></tr><tr><td>☒ Static Bag</td><td>☐ 12V DC Cable</td></tr></table> Notes: _____ _____ _____ _____ _____ _____ _____ _____ _____ _____	☒ Case	☒ Handle	☒ Keyboard	☒ Dry Gas Shelf	☒ Feet	☒ Breath Tube	☒ Ports	☒ Screws Tight	☐ Power cord	☐ Printer Cable	☒ Static Bag	☐ 12V DC Cable	☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>196</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP 103</u> 32 mm <u>*0.261</u> (.139 - .169) 36 mm <u>*0.273</u> (.156 - .190) 53 mm <u>*0.328</u> (.228 - .278) 103 mm <u>0.515</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>28421</u> ☒ Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP5088</td><td>202406K 06/19/2026</td></tr><tr><td>0.080</td><td>MP5089</td><td>202406L 06/19/2026</td></tr><tr><td>0.200</td><td>MP5090</td><td>202406N 06/20/2026</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG510701 04/17/2027</td></tr></tbody></table>	Simulator	Serial #	Lot #/Exp	0.050	MP5088	202406K 06/19/2026	0.080	MP5089	202406L 06/19/2026	0.200	MP5090	202406N 06/20/2026	0.080 DGS	N/A	AG510701 04/17/2027	Flow Column # <u>ATP102</u> ☒ 5L/min – 17mm ☒ 15L/min – 53mm ☒ 30L/min – 103mm ☒ R-Value <u>196</u> ☒ Post Calibration Verification (L/s) Flow Column # <u>ATP103</u> 32 mm <u>0.156</u> (.139 - .169) 36 mm <u>0.175</u> (.156 - .190) 53 mm <u>0.242</u> (.228 - .278) 103 mm <u>0.519</u> (.447 - .547) <table border="1" style="width:100%; border-collapse: collapse;"><thead><tr><th>Maintenance By</th><th>Date</th></tr></thead><tbody><tr><td>☐ Battery Replacement</td><td></td></tr><tr><td>☐ Dry Gas Regulator Replacement</td><td></td></tr><tr><td>☐ Breath Tube Replacement</td><td></td></tr><tr><td>☐ Other _____</td><td></td></tr><tr><td> </td><td></td></tr><tr><td> </td><td></td></tr><tr><td> </td><td></td></tr><tr><td> </td><td></td></tr><tr><td> </td><td></td></tr></tbody></table>	Maintenance By	Date	☐ Battery Replacement		☐ Dry Gas Regulator Replacement		☐ Breath Tube Replacement		☐ Other _____											
☒ Case	☒ Handle																																																
☒ Keyboard	☒ Dry Gas Shelf																																																
☒ Feet	☒ Breath Tube																																																
☒ Ports	☒ Screws Tight																																																
☐ Power cord	☐ Printer Cable																																																
☒ Static Bag	☐ 12V DC Cable																																																
Simulator	Serial #	Lot #/Exp																																															
0.050	MP5088	202406K 06/19/2026																																															
0.080	MP5089	202406L 06/19/2026																																															
0.200	MP5090	202406N 06/20/2026																																															
0.080 DGS	N/A	AG510701 04/17/2027																																															
Maintenance By	Date																																																
☐ Battery Replacement																																																	
☐ Dry Gas Regulator Replacement																																																	
☐ Breath Tube Replacement																																																	
☐ Other _____																																																	
Calibration Adjustment By _____		Department Inspection By <u>SLH</u>																																															
Barometric Pressure Gauge _____ ID # _____		Barometric Pressure ID# <u>28421</u>																																															
<table border="1" style="width:100%; border-collapse: collapse;"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td></td><td></td><td></td></tr><tr><td>0.100</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.300</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>		Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Gauge <u>1018</u> Instrument <u>1015</u> Mouth Alcohol Solution Lot # <u>2025-C</u> Acetone Stock Solution Lot # <u>2024-B</u>																			
Simulator	Serial #	Lot #	Expiration																																														
0.000		N/A	N/A																																														
0.040																																																	
0.100																																																	
0.200																																																	
0.300																																																	
0.080 DGS	N/A																																																
☐ Post Calibration Adjustment Stability Checks		<table border="1" style="width:100%; border-collapse: collapse;"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>MP6289</td></tr><tr><td>Interferent</td><td>MP5087</td></tr><tr><td>0.050</td><td>MP6291</td></tr><tr><td>0.080</td><td>MP5089</td></tr><tr><td>0.200</td><td>MP5090</td></tr></tbody></table>	Simulator	Serial Number	0.000	MP6289	Interferent	MP5087	0.050	MP6291	0.080	MP5089	0.200	MP5090																																			
Simulator	Serial Number																																																
0.000	MP6289																																																
Interferent	MP5087																																																
0.050	MP6291																																																
0.080	MP5089																																																
0.200	MP5090																																																
<table border="1" style="width:100%; border-collapse: collapse;"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td></td><td></td><td></td></tr><tr><td>0.080</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>		Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			Attachments <table border="0"><tr><td>☒ Form 41</td><td>☐ Post-Stability Checks</td></tr><tr><td>☒ Stability Checks</td><td>☒ Flow Calibration</td></tr><tr><td>☒ Calibration Certificate</td><td>☐ Form 40</td></tr><tr><td>☐ Calibration Adjustment</td><td>☐ Other _____</td></tr></table>	☒ Form 41	☐ Post-Stability Checks	☒ Stability Checks	☒ Flow Calibration	☒ Calibration Certificate	☐ Form 40	☐ Calibration Adjustment	☐ Other _____																			
Simulator	Serial #	Lot #	Expiration																																														
0.050																																																	
0.080																																																	
0.200																																																	
0.080 DGS	N/A																																																
☒ Form 41	☐ Post-Stability Checks																																																
☒ Stability Checks	☒ Flow Calibration																																																
☒ Calibration Certificate	☐ Form 40																																																
☐ Calibration Adjustment	☐ Other _____																																																
Notes/Suggested Service: * Flow rates outside of nominal. No external equipment or user error identified. WKP <u>09/25/2025</u> **Tech Correction: Corrected intake selection. WKP <u>10/7/2025</u> _____ _____ _____		☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use Taylor Gutschow Tech Review / Date _____ Shayla Platt Admin Review / Date _____ Digitally signed by Taylor Gutschow Date: 2025.10.07 16:09:22 -0400 Digitally signed by Shayla Platt Date: 2025.10.09 22:39:57 -0400																																															

Stability Checks

0.05g/210L 0.047 to 0.053 ✓	0.08g/210L 0.077 to 0.083 ✓	0.20g/210L 0.194 to 0.206 ✓	DGS 0.08g/210L 0.077 to 0.083 ✓ ≤0.003 of Wet ✓																																																																																																																																																
Performed Root Case Analysis	Performed Root Case Analysis	Performed Root Case Analysis	Performed Root Case Analysis																																																																																																																																																
ORMOND BEACH PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001156 09/25/2025 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>14:39</td></tr> <tr><td>Control Test</td><td>0.048</td><td>14:39</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:40</td></tr> <tr><td>Control Test</td><td>0.048</td><td>14:41</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:41</td></tr> <tr><td>Control Test</td><td>0.048</td><td>14:42</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:42</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0480</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> <i>menpi</i> Operator's Signature	Test	g/210L	Time	Air Blank	0.000	14:39	Control Test	0.048	14:39	Air Blank	0.000	14:40	Control Test	0.048	14:41	Air Blank	0.000	14:41	Control Test	0.048	14:42	Air Blank	0.000	14:42	Control Test Stats			Average	0.0480		Std Dev	0.0000		Rel Std Dev(%)	0.0000		ORMOND BEACH PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001156 09/25/2025 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>14:50</td></tr> <tr><td>Control Test</td><td>0.079</td><td>14:51</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:51</td></tr> <tr><td>Control Test</td><td>0.079</td><td>14:52</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:52</td></tr> <tr><td>Control Test</td><td>0.078</td><td>14:53</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:54</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0787</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7339</td><td></td></tr> </tbody> </table> <i>menpi</i> Operator's Signature	Test	g/210L	Time	Air Blank	0.000	14:50	Control Test	0.079	14:51	Air Blank	0.000	14:51	Control Test	0.079	14:52	Air Blank	0.000	14:52	Control Test	0.078	14:53	Air Blank	0.000	14:54	Control Test Stats			Average	0.0787		Std Dev	0.0006		Rel Std Dev(%)	0.7339		ORMOND BEACH PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001156 09/25/2025 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>14:55</td></tr> <tr><td>Control Test</td><td>0.203</td><td>14:56</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:56</td></tr> <tr><td>Control Test</td><td>0.200</td><td>14:56</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:57</td></tr> <tr><td>Control Test</td><td>0.200</td><td>14:58</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:58</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.2010</td><td></td></tr> <tr><td>Std Dev</td><td>0.0017</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.6617</td><td></td></tr> </tbody> </table> <i>menpi</i> Operator's Signature	Test	g/210L	Time	Air Blank	0.000	14:55	Control Test	0.203	14:56	Air Blank	0.000	14:56	Control Test	0.200	14:56	Air Blank	0.000	14:57	Control Test	0.200	14:58	Air Blank	0.000	14:58	Control Test Stats			Average	0.2010		Std Dev	0.0017		Rel Std Dev(%)	0.6617		ORMOND BEACH PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001156 09/25/2025 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>14:23</td></tr> <tr><td>Control Test</td><td>0.080</td><td>14:23</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:23</td></tr> <tr><td>Control Test</td><td>0.080</td><td>14:24</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:24</td></tr> <tr><td>Control Test</td><td>0.080</td><td>14:25</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:25</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0800</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> <i>menpi</i> Operator's Signature	Test	g/210L	Time	Air Blank	0.000	14:23	Control Test	0.080	14:23	Air Blank	0.000	14:23	Control Test	0.080	14:24	Air Blank	0.000	14:24	Control Test	0.080	14:25	Air Blank	0.000	14:25	Control Test Stats			Average	0.0800		Std Dev	0.0000		Rel Std Dev(%)	0.0000	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	14:39																																																																																																																																																	
Control Test	0.048	14:39																																																																																																																																																	
Air Blank	0.000	14:40																																																																																																																																																	
Control Test	0.048	14:41																																																																																																																																																	
Air Blank	0.000	14:41																																																																																																																																																	
Control Test	0.048	14:42																																																																																																																																																	
Air Blank	0.000	14:42																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0480																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	14:50																																																																																																																																																	
Control Test	0.079	14:51																																																																																																																																																	
Air Blank	0.000	14:51																																																																																																																																																	
Control Test	0.079	14:52																																																																																																																																																	
Air Blank	0.000	14:52																																																																																																																																																	
Control Test	0.078	14:53																																																																																																																																																	
Air Blank	0.000	14:54																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0787																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7339																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	14:55																																																																																																																																																	
Control Test	0.203	14:56																																																																																																																																																	
Air Blank	0.000	14:56																																																																																																																																																	
Control Test	0.200	14:56																																																																																																																																																	
Air Blank	0.000	14:57																																																																																																																																																	
Control Test	0.200	14:58																																																																																																																																																	
Air Blank	0.000	14:58																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.2010																																																																																																																																																		
Std Dev	0.0017																																																																																																																																																		
Rel Std Dev(%)	0.6617																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	14:23																																																																																																																																																	
Control Test	0.080	14:23																																																																																																																																																	
Air Blank	0.000	14:23																																																																																																																																																	
Control Test	0.080	14:24																																																																																																																																																	
Air Blank	0.000	14:24																																																																																																																																																	
Control Test	0.080	14:25																																																																																																																																																	
Air Blank	0.000	14:25																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0800																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		

80-001156

SAT 9/30/2025

Flow Calibration

Root Cause Analysis Performed Prior



Performed by SLH

INTOXILYZER 8000
Instrument Initialization
12:05 09/30/2025

ORMOND BEACH PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001156
09/30/2025
Software: 8100.27

Flow Rate Calibration*****

1: Rate (Liters/min) = 5

SQRT(Diff)) = 6.926

2: Rate (Liters/min) = 15

SQRT(Diff)) = 11.703

3: Rate (Liters/min) = 30

SQRT(Diff)) = 20.590

Dependent Data Scale Factor = 100000 L/min

Independent Data Scale Factor = 256

Rounded Slope = 708

Rounded Intercept = -701726

Correlation = 0.99838

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: ORMOND BEACH PD
Time of Inspection: 15:22

Date of Inspection: 10/02/2025

Serial Number: 80-001156
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202406K Exp: 06/19/2026	0.08g/210L Test (g/210L) Lot#:202406L Exp: 06/19/2026	0.20g/210L Test (g/210L) Lot#:202406N Exp: 06/20/2026	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG510701 Exp: 04/17/2027
0.000	0.047	0.078	0.200	0.079
0.000	0.047	0.078	0.201	0.079
0.000	0.047	0.078	0.201	0.079
0.000	0.047	0.078	0.201	0.080
0.000	0.047	0.078	0.202	0.080
0.000	0.047	0.078	0.202	0.079
0.000	0.047	0.078	0.202	0.079
0.000	0.047	0.077	0.202	0.079
0.000	0.048	0.078	0.202	0.080
0.000	0.048	0.078	0.202	0.079


Standard Deviations	0.0004	0.0003	0.0007	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.



I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



LEANDRA HIGGINBOTHAM

Signature and Printed Name

10/02/2025
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001156, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001156</u>	UNCERTAINTY* $\pm$
Owning Agency:	<u>ORMOND BEACH PD</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>10/02/2025</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>15:22</u>	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

FDLE/ATP Form 69 October 2024
Issuing Authority: Alcohol Testing Program

10/02/2025 Date
Leandra Higginbotham
LEANDRA HIGGINBOTHAM,
Department Inspector

Service • Integrity • Respect • Quality



INSTRUMENT PROCESSING SHEET

Agency Ormond Beach PDS/N 80-001156Florida Department of
Law EnforcementDate In 01/28/2025 DI Completion Date N/A☐ Ship ☐ P/U ☐ H/D ☒ CMI ☐ EE

Intake By <u>ALL</u> Date <u>01/28/2025</u>		Quality Checks By <u>DA</u> Date <u>02/26/2025</u>		Flow Calibration By _____ Date _____																																																													
☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☐ Handle ☒ Keyboard ☐ Dry Gas Shelf ☒ Feet ☐ Breath Tube ☒ Ports ☐ Screws Tight Other Equipment/ Accessories: ☐ Power cord ☐ Printer Cable ☐ Static Bag ☐ 12V DC Cable Notes: arrived in pelican case inside a cardboard box		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>86</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP102</u> 32 mm <u>0.062</u> (.139 - .169) 36 mm <u>0.078</u> (.156 - .190) 53 mm <u>0.152</u> (.228 - .278) 103 mm <u>0.445</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>30793</u> ☒ Stability Checks		☐ Flow Column # _____ ☐ 5L/min – 17mm ☐ 15L/min – 53mm ☐ 30L/min – 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																																													
		<table border="1" style="width:100%"><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr><tr><td>0.050</td><td>MP5088</td><td>202406K 06/19/2026</td></tr><tr><td>0.080</td><td>MP5089</td><td>202406L 06/19/2026</td></tr><tr><td>0.200</td><td>MP5090</td><td>202406N 06/20/2026</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG429602 10/22/2026</td></tr></table>		Simulator	Serial #	Lot #/Exp	0.050	MP5088	202406K 06/19/2026	0.080	MP5089	202406L 06/19/2026	0.200	MP5090	202406N 06/20/2026	0.080 DGS	N/A	AG429602 10/22/2026	Maintenance By _____ Date _____ ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____																																														
Simulator	Serial #	Lot #/Exp																																																															
0.050	MP5088	202406K 06/19/2026																																																															
0.080	MP5089	202406L 06/19/2026																																																															
0.200	MP5090	202406N 06/20/2026																																																															
0.080 DGS	N/A	AG429602 10/22/2026																																																															
Calibration Adjustment By _____				Department Inspection By _____																																																													
Barometric Pressure Gauge _____ ID # _____ <table border="1" style="width:100%"><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr><tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td></td><td></td><td></td></tr><tr><td>0.100</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.300</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></table> ☐ Post Calibration Adjustment Stability Checks <table border="1" style="width:100%"><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr><tr><td>0.050</td><td></td><td></td><td></td></tr><tr><td>0.080</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></table>				Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			Barometric Pressure ID# _____ Gauge _____ Instrument _____ Mouth Alcohol Solution Lot # _____ Acetone Stock Solution Lot # _____ <table border="1" style="width:100%"><tr><th>Simulator</th><th>Serial Number</th></tr><tr><td>0.000</td><td></td></tr><tr><td>Interferent</td><td></td></tr><tr><td>0.050</td><td></td></tr><tr><td>0.080</td><td></td></tr><tr><td>0.200</td><td></td></tr></table>		Simulator	Serial Number	0.000		Interferent		0.050		0.080		0.200	
Simulator	Serial #	Lot #	Expiration																																																														
0.000		N/A	N/A																																																														
0.040																																																																	
0.100																																																																	
0.200																																																																	
0.300																																																																	
0.080 DGS	N/A																																																																
Simulator	Serial #	Lot #	Expiration																																																														
0.050																																																																	
0.080																																																																	
0.200																																																																	
0.080 DGS	N/A																																																																
Simulator	Serial Number																																																																
0.000																																																																	
Interferent																																																																	
0.050																																																																	
0.080																																																																	
0.200																																																																	
Notes/Suggested Service: R-value <100. Flow verification values outside of nominal range. I ensured all tubing was free of kinks/tears and that there was sufficient pressure to the flow column from the air compressor. No external/user error was found. The instrument will need to be sent to repair. DA 02/26/2025				Attachments ☐ Form 41 ☐ Post-Stability Checks ☒ Stability Checks ☐ Flow Calibration ☐ Calibration Certificate ☐ Form 40 ☐ Calibration Adjustment ☒ Other <u>Form 51</u>																																																													
				☐ Instrument Complies with Chapter 11D-8, FAC ☒ Instrument Does Not Comply with Chapter 11D-8, FAC ☐ Return to/Place into Evidentiary Use ☒ Remain Out of Evidentiary Use ☐ Conduct an Agency Inspection Before Evidentiary Use Phil Nicodemus Digitally signed by Phil Nicodemus Date: 2025.03.05 15:09:09 -05'00' Shayla Platt Digitally signed by Shayla Platt Date: 2025.03.05 15:12:40 -05'00' Tech Review / Date _____ Admin Review / Date _____																																																													

Stability Checks 80-001156 DA 2/26/25

ORMOND BEACH PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001156
02/26/2025
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:34
Control Test	0.050	11:35
Air Blank	0.000	11:35
Control Test	0.050	11:36
Air Blank	0.000	11:36
Control Test	0.050	11:37
Air Blank	0.000	11:38
Control Test Stats		
Average	0.0500	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	




Operator's Signature

ORMOND BEACH PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001156
02/26/2025
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:39
Control Test	0.080	11:40
Air Blank	0.000	11:41
Control Test	0.080	11:41
Air Blank	0.000	11:42
Control Test	0.079	11:43
Air Blank	0.000	11:43
Control Test Stats		
Average	0.0797	
Std Dev	0.0006	
Rel Std Dev(%)	0.7247	


Wet



Operator's Signature

ORMOND BEACH PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001156
02/26/2025
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:44
Control Test	0.201	11:45
Air Blank	0.000	11:46
Control Test	0.200	11:46
Air Blank	0.000	11:47
Control Test	0.198	11:48
Air Blank	0.000	11:48
Control Test Stats		
Average	0.1997	
Std Dev	0.0015	
Rel Std Dev(%)	0.7650	



Operator's Signature

ORMOND BEACH PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001156
02/26/2025
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:49
Control Test	0.079	11:50
Air Blank	0.000	11:50
Control Test	0.080	11:51
Air Blank	0.000	11:51
Control Test	0.080	11:51
Air Blank	0.000	11:52
Control Test Stats		
Average	0.0797	
Std Dev	0.0006	
Rel Std Dev(%)	0.7247	

DGS



Operator's Signature

Return Material Authorization

Ship to: ☒ CMI, Inc.

☐ Enforcement Electronics

Shipment to repair facility authorized by: Michael Garner on 03/04/2025

Items Returned: Instrument ☒ Supplies ☐ Other ☐ Describe: _____

Instrument Model: Intoxilyzer 8000 Serial Number: 80-001156

Bill To Address:

Ormond Beach Police Department

ATTN: Michael Garner

Ship to Address:

FDLE Off-Site Mail Facility

c/o Florida Department of Law Enforcement

Alcohol Testing Program

813 B Lake Bradford Rd

Tallahassee, FL 32304

Reason for Return:

Flow sensor R-value <100. Instrument flow verification values outside nominal range.

Please choose one of the following options:

☐ 1. I _____, authorize all repairs.

☐ 2. I _____, authorize repairs up to \$_____.

☒ 3. I require an estimate **BEFORE** any repairs will be authorized and/ or conducted.

Please contact: Name: Michael Garner

Phone #: 386-559-0622 Email: michael.garner@ormondbeach.org

ATP Contact Name: Destinee Armstrong ATP Email: destinee.armstrong@fdle.state.fl.us