

Alcohol Testing Program - Instrument Processing Sheet

Agency: Volusia CO SO

Instrument Serial Number: 80-001154

Date In: 8/3/2026

DI Completion Date: 08/13/2026

☒ Ship☐ P/U☐ H/D

☐ CMI

☐ EE

Intake By: IAS Date: 8/3/2026		Quality Checks By: SLH Date: 8/11/2026		Flow Adjustment By: _____	
☒ Annual ☐ Dropped Off ☐ Registration ☐ Return from CMI / EE ☐ Training Instrument Visual Inspection ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/Accessories ☒ Power Cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes: AC to DC converter cord in box. SLH 8/11/2026		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value: 225 ☒ Flow Verification (L/s) Flow Column #: ATP 105 32 mm 0.156 (.139-.169) 36 mm 0.164 (.156-.190) 53 mm 0.238 (.228-.278) 103 mm 0.500 (.447-.547) ☒ Barometric Pressure Check Gauge ID #: 28427 Gauge: 1012 Instrument: 1014 ☒ Stability Checks		Flow Column #: _____ ☐ 5L/min – 17mm ☐ 15L/min – 53mm ☐ 30L/min – 103mm ☐ R-Value: _____ ☐ Post Adjustment Verification (L/S) Flow Column #: _____ 32 mm (.139-.169) 36 mm (.156-.190) 53 mm (.228-.278) 103 mm (.447-.547)	
Optical Bench Adjustment By: _____		Department Inspection By: SLH			
Barometric Pressure Gauge: _____ ID#: _____		Barometric Pressure ID#: 28427			
Simulator	Serial #	Lot #	Expiration	Gauge: 1012 Instrument: 1014	
0.000		N/A	N/A	Mouth Alcohol Solution Lot #: 2025-D Exp: 09/25/2027	
0.040				Acetone Stock Solution Lot #: 2025-B Exp: 09/22/2027	
0.100				Simulator	Serial Number
0.200				0.000	MP6289
0.300				Interferent	MP6290
0.080 DGS	N/A			0.050	MP6291
☐ Post Optical Bench Adjustment Stability Checks				0.080	MP6292
Simulator	Serial #	Lot #	Expiration	0.200	MP6293
0.050				Attachments	
0.080				☒ Form 41	
0.200				☒ Stability Checks	
0.080 DGS	N/A			☒ Calibration Certificate	
Gauge ID #: _____				☐ Optical Bench Adjustment	
Gauge: _____ Instrument: _____				☐ Post-Stability Checks	
Notes/Suggested Service: Instrument power button only blinks and will not turn on. SLH 8/4/2026 The instrument required the AC to DC converter to initialize and operate as expected. SLH 8/11/2026				☐ Flow Adjustment	
				☐ Form 40	
				☒ Other: Calculation Form	
				☒ Instrument Complies with Chapter 11D-8, FAC	
				☐ Instrument Does Not Comply with Chapter 11D-8, FAC	
				☒ Return to/Place into Evidentiary Use	
				☐ Remain Out of Evidentiary Use	
				☒ Conduct an Agency Inspection Before Evidentiary Use	
Taylor Gutschow				Wen-Chi Pierson	
Digitally signed by Taylor Gutschow Date: 2026.08.30 20:53:28 -04'00'				Digitally signed by Wen-Chi Pierson Date: 2026.08.31 09:23:05 -04'00'	
Tech Review				Admin Review	

80-00 1154

0.050 g/210L

0.047 to 0.053 g/210L

JULIUSIA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000
08/12/2026
SN 80-001154
Software: 8100.27

Test	g/20L	Time
Air Blank	0.000	15:09
Control Test	0.050	15:09
Air Blank	0.000	15:10
Control Test	0.049	15:11
Air Blank	0.000	15:11
Control Test	0.049	15:12
Air Blank	0.000	15:12
Control Test Stats		
Average	0.0493	
Std Dev	0.0006	
Rel. Std. Dev.(%)	1.1703	

Operator's Signature

0.080 g/210L

0.077 to 0.083 g/210L

2021/21/21

VOLUSTA COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000
 08/12/2026
 Software: 8100.27
 SN 80-001154

Test	g/210L	Time
Air Blank	0.000	12:33
Control Test	0.079	12:34
Air Blank	0.000	12:34
Control Test	0.078	12:35
Air Blank	0.000	12:36
Control Test	0.078	12:36
Air Blank	0.000	12:37
Control Test Stats		
Average	0.0783	
Std Dev	0.0006	
Rel. Std. Dev.(%)	0.7370	

Operator's Signature

0.200 g/210L

0.194 to 0.206 g/210L

VOLUSIF COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000
08/12/2026
Software: 8100.27
SN 80-001154

Test	g/210L	Time
Air Blank	0.100	15:15
Control Test	0.197	15:15
Air Blank	0.300	15:16
Control Test	0.195	15:17
Air Blank	0.000	15:17
Control Test	0.195	15:18
Air Blank	0.000	15:18
Control Test Stats		
Average	0.1953	
Std Dev	0.3006	
Rel Std Dev[%]	0.2941	

Operator's Signature

DGS 0.080 g/210L

**0.077 to 0.083 g/210L
≤0.003 g/210L of Wet**

VOLUSIA COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000
 08/11/2026
 Software: 8100.27
 SN 80-C01154

Test	g/2.0L	Time
Air Blank	0.000	16:23
Control Test	0.080	16:24
Air Blank	0.000	16:24
Control Test	0.079	16:25
Air Blank	0.000	16:25
Control Test	0.078	16:25
Air Blank	0.000	16:26
Control Test Stats		
Average	0.0790	
Std Dev	0.0010	
Rel Std Dev(%)	1.2658	

Operator's Signature

Verified instrument was able to read gas pressure.

* performed 8/11/2026 ~~self~~

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: VOLUSIA COUNTY SO
Time of Inspection: 18:08

Date of Inspection: 08/13/2026

Serial Number: 80-001154
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 202603A Exp: 03/03/2028	0.08g/210L Test (g/210L) Lot#: 202603B Exp: 03/03/2028	0.20g/210L Test (g/210L) Lot#: 202603D Exp: 03/04/2028	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 23825080A1 Exp: 10/05/2027
0.000	0.050	0.078	0.196	0.079
0.000	0.050	0.079	0.196	0.078
0.000	0.049	0.078	0.196	0.078
0.000	0.050	0.078	0.196	0.078
0.000	0.050	0.079	0.197	0.078
0.000	0.050	0.078	0.196	0.079
0.000	0.050	0.079	0.196	0.078
0.000	0.049	0.079	0.196	0.079
0.000	0.050	0.079	0.197	0.078
0.000	0.050	0.079	0.196	0.079


Standard Deviations	0.0004	0.0005	0.0004	0.0005
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.


LEANDRA HIGGINBOTHAM
Signature and Printed Name

08/13/2026
Date

Florida Department of Law Enforcement
Alcohol Testing Program - Average Calculations for Calibration Certificate

Calculations By: SLH

Serial Number: 80-001154

Date: 8/20/2026

Solution:

0.05 g/210 L Test: 0.498 / 10 = 0.050 g/210L
(SUM) (AVERAGE)

0.08 g/210L Test: 0.786 / 10 = 0.079 g/210L
(SUM) (AVERAGE)

0.20 g/210L Test: 1.962 / 10 = 0.196 g/210L
(SUM) (AVERAGE)

0.08 g/210L Dry Gas Std Test: 0.784 / 10 = 0.078 g/210L
(SUM) (AVERAGE)

Comments:

Average = (Sum of observations) / (Total number of observations)

All results are reported in g/210L



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001154, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001154</u>	Average of y ± Uncertainty*
Owning Agency:	<u>VOLUSIA COUNTY SO</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>08/13/2026</u>	0.079 g/ 210 L 0.004
Calibration Time:	<u>18:08</u>	0.196 g/ 210 L 0.008
		0.078 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

S. LeAndra

Higginbotham

Digitally signed by S. LeAndra
Higginbotham
Date: 2026.08.20 10:12:42 -04'00'

08/13/2026

Date

LEANDRA HIGGINBOTHAM,

FDLE/ATP Form 69 August 2026

Issuing Authority: Alcohol Testing Program

Department Inspector

Service • Integrity • Respect • Quality