

Instrument Processing Sheet

Agency: CLAY COUNTY SO Instrument Serial Number: 80-001071
 Date In: 10/23/2025 DI Completion Date: 10/28/2025 Ship P/U H/D CMI EE

Intake By: <u>WKP</u> Date: <u>10/23/25</u>	Quality Checks By: <u>SLH</u> Date: <u>10/27/25</u>	Flow Adjustment By: _____ Date: _____
<input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input checked="" type="checkbox"/> Return from CMI / EE <input type="checkbox"/> Return unworked <input type="checkbox"/> Training Visual Inspection <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/Accessories <input type="checkbox"/> Power Cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable	<input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value: <u>198</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column #: <u>ATP 102</u> 32 mm <u>0.152</u> (.139-.169) 36 mm <u>0.171</u> (.156-.190) 53 mm <u>0.238</u> (.228-.278) 103 mm <u>0.488</u> (.447-.547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID #: <u>28662</u> Gauge: <u>1008</u> Instrument: <u>1007</u> <input checked="" type="checkbox"/> Stability Checks	Flow Column #: _____ <input type="checkbox"/> 5L/min – 17mm <input type="checkbox"/> 15L/min – 53mm <input type="checkbox"/> 30L/min – 103mm <input type="checkbox"/> R-Value: _____ <input type="checkbox"/> Post Adjustment Verification (L/S) Flow Column #: _____ 32 mm _____ (.139-.169) 36 mm _____ (.156-.190) 53 mm _____ (.228-.278) 103 mm _____ (.447-.547)

Notes: 	<table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot#/Exp</th> </tr> </thead> <tbody> <tr> <td rowspan="2">0.050</td> <td rowspan="2">MP5088</td> <td>202406K</td> </tr> <tr> <td>06/19/2026</td> </tr> <tr> <td rowspan="2">0.080</td> <td rowspan="2">MP5089</td> <td>202406L</td> </tr> <tr> <td>6/19/2026</td> </tr> <tr> <td rowspan="2">0.200</td> <td rowspan="2">MP5090</td> <td>202406N</td> </tr> <tr> <td>06/20/2026</td> </tr> <tr> <td rowspan="2">0.080 DGS</td> <td rowspan="2">N/A</td> <td>AG510701</td> </tr> <tr> <td>4/17/2027</td> </tr> </tbody> </table>	Simulator	Serial #	Lot#/Exp	0.050	MP5088	202406K	06/19/2026	0.080	MP5089	202406L	6/19/2026	0.200	MP5090	202406N	06/20/2026	0.080 DGS	N/A	AG510701	4/17/2027	<table border="1" style="width: 100%; border-collapse: collapse;"> <tr> <td style="width: 25%;">Maintenance</td> <td>By: _____</td> <td>Date: _____</td> </tr> <tr> <td colspan="3"> <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement and Tank Sensor Tare <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other: _____ </td> </tr> </table>	Maintenance	By: _____	Date: _____	<input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement and Tank Sensor Tare <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other: _____		
Simulator	Serial #	Lot#/Exp																									
0.050	MP5088	202406K																									
		06/19/2026																									
0.080	MP5089	202406L																									
		6/19/2026																									
0.200	MP5090	202406N																									
		06/20/2026																									
0.080 DGS	N/A	AG510701																									
		4/17/2027																									
Maintenance	By: _____	Date: _____																									
<input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement and Tank Sensor Tare <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other: _____																											

Optical Bench Adjustment By: _____	Department Inspection By: <u>SLH</u>																																								
Barometric Pressure Gauge: _____ ID#: _____	Barometric Pressure ID#: <u>28421</u>																																								
<table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>	Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Gauge: <u>1011</u> Instrument: <u>1009</u> Mouth Alcohol Solution Lot #: <u>2025-C</u> Exp: <u>09/25/2027</u> Acetone Stock Solution Lot #: <u>2025-B</u> Exp: <u>09/22/2027</u> <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>MP6289</td> </tr> <tr> <td>Interferent</td> <td>MP6290</td> </tr> <tr> <td>0.050</td> <td>MP5088</td> </tr> <tr> <td>0.080</td> <td>MP6292</td> </tr> <tr> <td>0.200</td> <td>MP5090</td> </tr> </tbody> </table>	Simulator	Serial Number	0.000	MP6289	Interferent	MP6290	0.050	MP5088	0.080	MP6292	0.200	MP5090
Simulator	Serial #	Lot #	Expiration																																						
0.000		N/A	N/A																																						
0.040																																									
0.100																																									
0.200																																									
0.300																																									
0.080 DGS	N/A																																								
Simulator	Serial Number																																								
0.000	MP6289																																								
Interferent	MP6290																																								
0.050	MP5088																																								
0.080	MP6292																																								
0.200	MP5090																																								
<input type="checkbox"/> Post Optical Bench Adjustment Stability Checks <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>	Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			<table border="1" style="width: 100%; border-collapse: collapse;"> <tr> <td colspan="2">Attachments</td> </tr> <tr> <td> <input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Stability Checks <input checked="" type="checkbox"/> Calibration Certificate <input type="checkbox"/> Optical Bench Adjustment </td> <td> <input type="checkbox"/> Post-Stability Checks <input type="checkbox"/> Flow Adjustment <input type="checkbox"/> Form 40 <input type="checkbox"/> Other: _____ </td> </tr> </table>	Attachments		<input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Stability Checks <input checked="" type="checkbox"/> Calibration Certificate <input type="checkbox"/> Optical Bench Adjustment	<input type="checkbox"/> Post-Stability Checks <input type="checkbox"/> Flow Adjustment <input type="checkbox"/> Form 40 <input type="checkbox"/> Other: _____																
Simulator	Serial #	Lot #	Expiration																																						
0.050																																									
0.080																																									
0.200																																									
0.080 DGS	N/A																																								
Attachments																																									
<input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Stability Checks <input checked="" type="checkbox"/> Calibration Certificate <input type="checkbox"/> Optical Bench Adjustment	<input type="checkbox"/> Post-Stability Checks <input type="checkbox"/> Flow Adjustment <input type="checkbox"/> Form 40 <input type="checkbox"/> Other: _____																																								
Barometric Pressure Gauge: _____ ID#: _____																																									

Notes/Suggested Service: Stabilities performed on 2 days. SLH 10/28/25	<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <hr/> <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <hr/> <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use
Taylor Gutschow <small>Digitally signed by Taylor Gutschow Date: 2025.10.29 14:33:39 -04'00'</small>	Shayla Platt <small>Digitally signed by Shayla Platt Date: 2025.11.04 16:26:01 -05'00'</small>
Tech Review	Admin Review

Stability Checks

80-001071
Sub 10/27/2025

0.050 g/210L 0.047 to 0.053 g/210L	0.080 g/210L 0.077 to 0.083 g/210L	0.200 g/210L 0.194 to 0.206 g/210L	DGS 0.080 g/210L 0.077 to 0.083 g/210L 50.003 g/210L of Wet																																																																																																																																				
<p>CLAY COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 10/28/2025 SN 80-001071 Software: 8100.27</p> <p>Test 9/211L Time</p> <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>09:39</td></tr> <tr><td>Control Test</td><td>0.048</td><td>09:39</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:40</td></tr> <tr><td>Control Test</td><td>0.048</td><td>09:41</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:41</td></tr> <tr><td>Control Test</td><td>0.048</td><td>09:42</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:42</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0480</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </table>	Air Blank	0.000	09:39	Control Test	0.048	09:39	Air Blank	0.000	09:40	Control Test	0.048	09:41	Air Blank	0.000	09:41	Control Test	0.048	09:42	Air Blank	0.000	09:42	Control Test Stats			Average	0.0480		Std Dev	0.0000		Rel Std Dev(%)	0.0000		<p>CLAY COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 10/27/2025 SN 80-001071 Software: 8100.27</p> <p>Test 9/210L Time</p> <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>12:58</td></tr> <tr><td>Control Test</td><td>0.078</td><td>12:58</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:59</td></tr> <tr><td>Control Test</td><td>0.078</td><td>12:59</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:00</td></tr> <tr><td>Control Test</td><td>0.078</td><td>13:01</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:01</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0780</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </table>	Air Blank	0.000	12:58	Control Test	0.078	12:58	Air Blank	0.000	12:59	Control Test	0.078	12:59	Air Blank	0.000	13:00	Control Test	0.078	13:01	Air Blank	0.000	13:01	Control Test Stats			Average	0.0780		Std Dev	0.0000		Rel Std Dev(%)	0.0000		<p>CLAY COUNTY SC Intoxilyzer - Alcohol Analyzer Model 8000 10/29/2025 SN 80-001071 Software: 8100.27</p> <p>Test 9/2.0L Time</p> <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>09:47</td></tr> <tr><td>Control Test</td><td>0.199</td><td>09:48</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:49</td></tr> <tr><td>Control Test</td><td>0.199</td><td>09:49</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:50</td></tr> <tr><td>Control Test</td><td>0.199</td><td>09:50</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:51</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1990</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </table>	Air Blank	0.000	09:47	Control Test	0.199	09:48	Air Blank	0.000	09:49	Control Test	0.199	09:49	Air Blank	0.000	09:50	Control Test	0.199	09:50	Air Blank	0.000	09:51	Control Test Stats			Average	0.1990		Std Dev	0.0000		Rel Std Dev(%)	0.0000		<p>CLAY COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 10/27/2025 SN 80-001071 Software: 8100.27</p> <p>Test 9/210L Time</p> <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>13:06</td></tr> <tr><td>Control Test</td><td>0.080</td><td>13:06</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:07</td></tr> <tr><td>Control Test</td><td>0.080</td><td>13:07</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:08</td></tr> <tr><td>Control Test</td><td>0.080</td><td>13:08</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:08</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0800</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </table>	Air Blank	0.000	13:06	Control Test	0.080	13:06	Air Blank	0.000	13:07	Control Test	0.080	13:07	Air Blank	0.000	13:08	Control Test	0.080	13:08	Air Blank	0.000	13:08	Control Test Stats			Average	0.0800		Std Dev	0.0000		Rel Std Dev(%)	0.0000	
Air Blank	0.000	09:39																																																																																																																																					
Control Test	0.048	09:39																																																																																																																																					
Air Blank	0.000	09:40																																																																																																																																					
Control Test	0.048	09:41																																																																																																																																					
Air Blank	0.000	09:41																																																																																																																																					
Control Test	0.048	09:42																																																																																																																																					
Air Blank	0.000	09:42																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.0480																																																																																																																																						
Std Dev	0.0000																																																																																																																																						
Rel Std Dev(%)	0.0000																																																																																																																																						
Air Blank	0.000	12:58																																																																																																																																					
Control Test	0.078	12:58																																																																																																																																					
Air Blank	0.000	12:59																																																																																																																																					
Control Test	0.078	12:59																																																																																																																																					
Air Blank	0.000	13:00																																																																																																																																					
Control Test	0.078	13:01																																																																																																																																					
Air Blank	0.000	13:01																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.0780																																																																																																																																						
Std Dev	0.0000																																																																																																																																						
Rel Std Dev(%)	0.0000																																																																																																																																						
Air Blank	0.000	09:47																																																																																																																																					
Control Test	0.199	09:48																																																																																																																																					
Air Blank	0.000	09:49																																																																																																																																					
Control Test	0.199	09:49																																																																																																																																					
Air Blank	0.000	09:50																																																																																																																																					
Control Test	0.199	09:50																																																																																																																																					
Air Blank	0.000	09:51																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.1990																																																																																																																																						
Std Dev	0.0000																																																																																																																																						
Rel Std Dev(%)	0.0000																																																																																																																																						
Air Blank	0.000	13:06																																																																																																																																					
Control Test	0.080	13:06																																																																																																																																					
Air Blank	0.000	13:07																																																																																																																																					
Control Test	0.080	13:07																																																																																																																																					
Air Blank	0.000	13:08																																																																																																																																					
Control Test	0.080	13:08																																																																																																																																					
Air Blank	0.000	13:08																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.0800																																																																																																																																						
Std Dev	0.0000																																																																																																																																						
Rel Std Dev(%)	0.0000																																																																																																																																						
<p>Performed Root Case Analysis</p> <p>Operator's Signature: <i>M. Siggel</i></p>	<p>Performed Root Case Analysis</p> <p>Operator's Signature: <i>M. Siggel</i></p>	<p>Performed Root Case Analysis</p> <p>Operator's Signature: <i>M. Siggel</i></p>	<p>Performed Root Case Analysis</p> <p>Operator's Signature: <i>M. Siggel</i></p>																																																																																																																																				

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: CLAY COUNTY SO
Time of Inspection: 14:29

Date of Inspection: 10/28/2025

Serial Number: 80-001071
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202406K Exp: 06/19/2026	0.08g/210L Test (g/210L) Lot#:202406L Exp: 06/19/2026	0.20g/210L Test (g/210L) Lot#:202406N Exp: 06/20/2026	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG510701 Exp: 04/17/2027
0.000	0.047	0.077	0.197	0.080
0.000	0.048	0.077	0.198	0.080
0.000	0.047	0.077	0.198	0.080
0.000	0.047	0.078	0.198	0.080
0.000	0.047	0.077	0.198	0.080
0.000	0.047	0.077	0.198	0.080
0.000	0.048	0.078	0.199	0.080
0.000	0.047	0.078	0.199	0.080
0.000	0.047	0.078	0.199	0.080
0.000	0.047	0.078	0.199	0.080
0.000	0.048	0.078	0.199	0.080


Standard Deviations	0.0004	0.0005	0.0006	0.0000
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5


Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.


 LEANDRA HIGGINBOTHAM
 Signature and Printed Name

10/28/2025
 Date



Florida Department of Law Enforcement
 Alcohol Testing Program
 2331 Phillips Road
 Tallahassee, FL 32308

Calibration Certificate

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001071, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001071</u>	UNCERTAINTY* ±	
Owning Agency:	<u>CLAY COUNTY SO</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>10/28/2025</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>14:29</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
 *Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.
 This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

10/28/2025

Date

LEANDRA HIGGINBOTHAM,

Department Inspector

FDLE/ATP Form 69 October 2024
 Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality