

Agency: Rockledge PD **Instrument Serial Number:** 80-001261

Date In: 7/17/2026 **DI Completion Date:** 7/24/2026 ☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

*All printed or downloaded copies are uncontrolled
Confirm version prior to use*

Issued By: Forensic Quality Manager
Issue Date: December 22, 2025

Stability Checks

80-00 1261

0.050 g/210L	0.080 g/210L	0.200 g/210L	DGS 0.080 g/210L
0.047 to 0.053 g/210L ROCKLEDGE PD Intoxilyzer - Alcohol Analyzer Model 8000 07/22/2026 Software: 8100.27 SN 80-001261 Test g/210L Time Air Blank 0.000 13:42 Control Test 0.050 13:43 Air Blank 0.000 13:44 Control Test 0.050 13:44 Air Blank 0.000 13:45 Control Test 0.050 13:46 Air Blank 0.000 13:46 Control Test Status Average 0.0500 Std Dev 0.0000 Rel Std Dev(%) 0.0000 Operator Signature	0.077 to 0.083 g/210L ROCKLEDGE PD Intoxilyzer - Alcohol Analyzer Model 8000 07/22/2026 Software: 8100.27 SN 80-001261 Test g/210L Time Air Blank 0.000 13:07 Control Test 0.079 13:08 Air Blank 0.000 13:08 Control Test 0.079 13:09 Air Blank 0.000 13:10 Control Test 0.079 13:10 Air Blank 0.000 13:11 Control Test Status Average 0.0790 Std Dev 0.0000 Rel Std Dev(%) 0.0000 Operator Signature	0.194 to 0.206 g/210L ROCKLEDGE PD Intoxilyzer - Alcohol Analyzer Model 8000 07/22/2026 Software: 8100.27 SN 80-001261 Test g/210L Time Air Blank 0.000 13:12 Control Test 0.199 13:13 Air Blank 0.000 13:14 Control Test 0.198 13:14 Air Blank 0.000 13:15 Control Test 0.199 13:15 Air Blank 0.000 13:16 Control Test Status Average 0.1987 Std Dev 0.0006 Rel Std Dev(%) 0.2906 Operator Signature	0.077 to 0.083 g/210L ROCKLEDGE PD Intoxilyzer - Alcohol Analyzer Model 8000 07/22/2026 Software: 8100.27 SN 80-001261 Test g/210L Time Air Blank 0.000 13:17 Control Test 0.079 13:17 Air Blank 0.000 13:18 Control Test 0.081 13:18 Air Blank 0.000 13:19 Control Test 0.080 13:19 Air Blank 0.000 13:19 Control Test Status Average 0.0800 Std Dev 0.0010 Rel Std Dev(%) 1.2500 Operator Signature

Verified instrument was able to read gas pressure.

Florida Department of Law Enforcement

Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: ROCKLEDGE PD
Time of Inspection: 13:16

Date of Inspection: 07/24/2026

Serial Number: 80-001261
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202603A Exp: 03/03/2028	0.08g/210L Test (g/210L) Lot#:202603B Exp: 03/03/2028	0.20g/210L Test (g/210L) Lot#:202603D Exp: 03/04/2028	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:23825080A1 Exp: 10/05/2027
0.000	0.050	0.079	0.200	0.079
0.000	0.051	0.079	0.200	0.079
0.000	0.051	0.080	0.200	0.079
0.000	0.050	0.079	0.200	0.079
0.000	0.051	0.079	0.200	0.080
0.000	0.051	0.079	0.201	0.080
0.000	0.051	0.080	0.201	0.080
0.000	0.051	0.080	0.201	0.080
0.000	0.051	0.080	0.201	0.079
0.000	0.051	0.079	0.202	0.080

Standard Deviations	0.0004	0.0005	0.0006	0.0005
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0005 Number of Simulators Used: 5


Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.


LEANDRA HIGGINBOTHAM
Signature and Printed Name

07/24/2026
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001261, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001261</u>	UNCERTAINTY* $\pm$
Owning Agency:	<u>ROCKLEDGE PD</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>07/24/2026</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>13:16</u>	0.200 g/ 210 L 0.008
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.


LEANDRA HIGGINBOTHAM,
Department Inspector

07/24/2026

Date

FDLE/ATP Form 69 January 2026
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality