



INSTRUMENT PROCESSING SHEET

Agency Hendry CSO

S/N 80-006234

Florida Department of Law Enforcement

Date In 01/16/2025 DI Completion Date 02/03/2025

Ship P/U H/D CMI EE

Intake By TDG _____ Date <u>01/27/2025</u> <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: <u>Dropped off.</u>	Quality Checks By TDG _____ Date <u>02/03/2025</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>166</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP101</u> 32 mm <u>0.152</u> (.139 - .169) 36 mm <u>0.175</u> (.156 - .190) 53 mm <u>0.250</u> (.228 - .278) 103 mm <u>0.515</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>26932</u> <input checked="" type="checkbox"/> Stability Checks	Flow Calibration By _____ Date _____ Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)															
	<table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>MP6286</td> <td>202303K 03/29/2025</td> </tr> <tr> <td>0.080</td> <td>MP6287</td> <td>202303L 03/29/2025</td> </tr> <tr> <td>0.200</td> <td>MP6288</td> <td>202304C 04/05/2025</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG429602 10/22/2026</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.050	MP6286	202303K 03/29/2025	0.080	MP6287	202303L 03/29/2025	0.200	MP6288	202304C 04/05/2025	0.080 DGS	N/A	AG429602 10/22/2026	Maintenance By TDG _____ Date <u>02/03/2025</u> <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input checked="" type="checkbox"/> Other <u>Replaced internal printer paper</u>
Simulator	Serial #	Lot #/Exp															
0.050	MP6286	202303K 03/29/2025															
0.080	MP6287	202303L 03/29/2025															
0.200	MP6288	202304C 04/05/2025															
0.080 DGS	N/A	AG429602 10/22/2026															

Calibration Adjustment By _____ Barometric Pressure Gauge _____ ID # _____ <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> <input type="checkbox"/> Post Calibration Adjustment Stability Checks <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>	Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			Department Inspection By TDG _____ Barometric Pressure ID# <u>26932</u> Gauge <u>1022</u> Instrument <u>1020</u> Mouth Alcohol Solution Lot # <u>2024-A</u> Acetone Stock Solution Lot # <u>2023-B</u> <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>MP4863</td> </tr> <tr> <td>Interferent</td> <td>MP4864</td> </tr> <tr> <td>0.050</td> <td>MP6286</td> </tr> <tr> <td>0.080</td> <td>MP6287</td> </tr> <tr> <td>0.200</td> <td>MP6288</td> </tr> </tbody> </table>	Simulator	Serial Number	0.000	MP4863	Interferent	MP4864	0.050	MP6286	0.080	MP6287	0.200	MP6288
Simulator	Serial #	Lot #	Expiration																																																										
0.000		N/A	N/A																																																										
0.040																																																													
0.100																																																													
0.200																																																													
0.300																																																													
0.080 DGS	N/A																																																												
Simulator	Serial #	Lot #	Expiration																																																										
0.050																																																													
0.080																																																													
0.200																																																													
0.080 DGS	N/A																																																												
Simulator	Serial Number																																																												
0.000	MP4863																																																												
Interferent	MP4864																																																												
0.050	MP6286																																																												
0.080	MP6287																																																												
0.200	MP6288																																																												
	Attachments <input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Stability Checks <input checked="" type="checkbox"/> Calibration Certificate <input type="checkbox"/> Calibration Adjustment <input type="checkbox"/> Post-Stability Checks <input type="checkbox"/> Flow Calibration <input checked="" type="checkbox"/> Form 40 <input type="checkbox"/> Other _____																																																												

Notes/Suggested Service: _____ _____ _____ _____ _____ _____	<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use Digitally signed by _____ Destinee Armstrong Digitally signed by Destinee Armstrong Date: 2025.02.05 11:53:54 -0500 Shayla Platt Shayla Platt Date: 2025.02.05 Tech Review / Date _____ Admin Review / Date <u>14:09:07-05'00'</u>
---	---

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HENDRY COUNTY SO
Time of Inspection: 10:13

Date of Inspection: 02/03/2025

Serial Number: 80-006234
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08g/210L Test (g/210L) Lot#: _____ Exp: _____	0.20g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: _____ Exp: _____

Number of Simulators Used: _____

Remarks:

AI NOT CONDUCTED. BYPASSED TO BRING OUT OF DISABLED MODE.

Not determined *MG*
2/3/25

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Taylor D Gutschow

TAYLOR D GUTSCHOW

Signature and Printed Name

02/03/2025
Date

Stability Checks

0.05g/210L 0.047 to 0.053	0.08g/210L 0.077 to 0.083	0.20g/210L 0.194 to 0.206	DGS 0.08g/210L 0.077 to 0.083																																																																																																															
<p>HENRY COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 02/03/2025 SN 80-006234 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>10:58</td></tr> <tr><td>Control Test</td><td>0.051</td><td>10:56</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:59</td></tr> <tr><td>Control Test</td><td>0.051</td><td>11:00</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:00</td></tr> <tr><td>Control Test</td><td>0.051</td><td>11:01</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:02</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0510</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel. Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> <p>Operator's Signature <i>MW</i></p>	Test	g/210L	Time	Air Blank	0.000	10:58	Control Test	0.051	10:56	Air Blank	0.000	10:59	Control Test	0.051	11:00	Air Blank	0.000	11:00	Control Test	0.051	11:01	Air Blank	0.000	11:02	Control Test Stats			Average	0.0510		Std Dev	0.0000		Rel. Std Dev(%)	0.0000		<p>HENRY COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 02/03/2025 SN 80-006234 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>10:42</td></tr> <tr><td>Control Test</td><td>0.082</td><td>10:42</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:43</td></tr> <tr><td>Control Test</td><td>0.081</td><td>10:44</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:44</td></tr> <tr><td>Control Test</td><td>0.080</td><td>10:45</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:46</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0810</td><td></td></tr> <tr><td>Std Dev</td><td>0.0010</td><td></td></tr> <tr><td>Rel. Std Dev(%)</td><td>1.2346</td><td></td></tr> </tbody> </table> <p>Operator's Signature <i>MW</i></p>	Test	g/210L	Time	Air Blank	0.000	10:42	Control Test	0.082	10:42	Air Blank	0.000	10:43	Control Test	0.081	10:44	Air Blank	0.000	10:44	Control Test	0.080	10:45	Air Blank	0.000	10:46	Control Test Stats			Average	0.0810		Std Dev	0.0010		Rel. Std Dev(%)	1.2346		<p>HENRY COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 02/03/2025 SN 80-006234 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>10:49</td></tr> <tr><td>Control Test</td><td>0.205</td><td>10:50</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:51</td></tr> <tr><td>Control Test</td><td>0.201</td><td>10:51</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:51</td></tr> <tr><td>Control Test</td><td>0.201</td><td>10:52</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:53</td></tr> <tr><td>Control Test</td><td>0.201</td><td>10:53</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.2023</td><td></td></tr> <tr><td>Std Dev</td><td>0.0023</td><td></td></tr> <tr><td>Rel. Std Dev(%)</td><td>1.1414</td><td></td></tr> </tbody> </table> <p>Operator's Signature <i>MW</i></p>	Test	g/210L	Time	Air Blank	0.000	10:49	Control Test	0.205	10:50	Air Blank	0.000	10:51	Control Test	0.201	10:51	Air Blank	0.000	10:51	Control Test	0.201	10:52	Air Blank	0.000	10:53	Control Test	0.201	10:53	Control Test Stats			Average	0.2023		Std Dev	0.0023		Rel. Std Dev(%)	1.1414		<p>0.077 to 0.083</p> <p>0.077 to 0.083</p> <p>0.077 to 0.083</p> <p>0.077 to 0.083</p> <p>Printed to external printer. Will attach.</p> <p>MW 2/3/25</p>
Test	g/210L	Time																																																																																																																
Air Blank	0.000	10:58																																																																																																																
Control Test	0.051	10:56																																																																																																																
Air Blank	0.000	10:59																																																																																																																
Control Test	0.051	11:00																																																																																																																
Air Blank	0.000	11:00																																																																																																																
Control Test	0.051	11:01																																																																																																																
Air Blank	0.000	11:02																																																																																																																
Control Test Stats																																																																																																																		
Average	0.0510																																																																																																																	
Std Dev	0.0000																																																																																																																	
Rel. Std Dev(%)	0.0000																																																																																																																	
Test	g/210L	Time																																																																																																																
Air Blank	0.000	10:42																																																																																																																
Control Test	0.082	10:42																																																																																																																
Air Blank	0.000	10:43																																																																																																																
Control Test	0.081	10:44																																																																																																																
Air Blank	0.000	10:44																																																																																																																
Control Test	0.080	10:45																																																																																																																
Air Blank	0.000	10:46																																																																																																																
Control Test Stats																																																																																																																		
Average	0.0810																																																																																																																	
Std Dev	0.0010																																																																																																																	
Rel. Std Dev(%)	1.2346																																																																																																																	
Test	g/210L	Time																																																																																																																
Air Blank	0.000	10:49																																																																																																																
Control Test	0.205	10:50																																																																																																																
Air Blank	0.000	10:51																																																																																																																
Control Test	0.201	10:51																																																																																																																
Air Blank	0.000	10:51																																																																																																																
Control Test	0.201	10:52																																																																																																																
Air Blank	0.000	10:53																																																																																																																
Control Test	0.201	10:53																																																																																																																
Control Test Stats																																																																																																																		
Average	0.2023																																																																																																																	
Std Dev	0.0023																																																																																																																	
Rel. Std Dev(%)	1.1414																																																																																																																	

HENDRY COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006234
02/03/2025
Software: 8100.27

ALS

Test	g/210L	Time
Air Blank	0.000	10:14
Control Test	0.080	10:14
Air Blank	0.000	10:15
Control Test	0.080	10:15
Air Blank	0.000	10:16
Control Test	0.079	10:16
Air Blank	0.000	10:17
Control Test Stats		
Average	0.0797	
Std Dev	0.0006	
Rel Std Dev(%)	0.7247	

nk

Operator's Signature

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: HENDRY COUNTY SO
Time of Inspection: 14:59

Date of Inspection: 02/03/2025

Serial Number: 80-006234
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202303K Exp: 03/29/2025	0.08g/210L Test (g/210L) Lot#:202303L Exp: 03/29/2025	0.20g/210L Test (g/210L) Lot#:202304C Exp: 04/05/2025	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG429602 Exp: 10/22/2026
0.000	0.051	0.081	0.204	0.080
0.000	0.051	0.080	0.202	0.079
0.000	0.051	0.080	0.202	0.079
0.000	0.051	0.080	0.201	0.079
0.000	0.051	0.080	0.201	0.079
0.000	0.051	0.080	0.201	0.079
0.000	0.051	0.081	0.201	0.079
0.000	0.051	0.081	0.201	0.079
0.000	0.051	0.081	0.201	0.079
0.000	0.051	0.081	0.201	0.079
0.000	0.051	0.081	0.201	0.079


Standard Deviations	0.0000	0.0005	0.0009	0.0003
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.


I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

02/03/2025
Date



Florida Department of Law Enforcement
 Alcohol Testing Program
 4700 Terminal Drive, Suite 1
 Ft. Myers, FL 33907

Calibration Certificate

This is to certify the calibration of Intoxilyzer 8000 serial number 80-006234, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-006234</u>	UNCERTAINTY* ±	
Owning Agency:	<u>HENDRY COUNTY SO</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>02/03/2025</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>14:59</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.
 Bias is limited by calibration acceptance criteria: All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
 *Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).
 The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.
 This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

02/03/2025

Date
 TAYLOR D GUTSCHOW,
 Department Inspector

FDLE/ATP Form 69 December 2021
 Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality