

Alcohol Testing Program - Instrument Processing Sheet

Agency: Hendry County Sheriff's Office Instrument Serial Number: 80-001207

Date In: 5/20/2026 DI Completion Date: 6/8/2026 ☐ Ship ☒ P/U ☐ H/D ☐ CMI ☐ EE

Intake By: <u>TDG</u> Date: <u>5/20/2026</u>		Quality Checks By: <u>TDG</u> Date: <u>6/8/2026</u>		Flow Adjustment By: <u>TDG</u>																														
☒ Annual ☒ Dropped Off ☐ Registration ☐ Return from CMI / EE ☐ Training Instrument Visual Inspection ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/Accessories ☐ Power Cord ☐ Printer Cable ☐ Static Bag ☐ 12V DC Cable Notes: No box. Uploaded records.		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value: <u>129</u> ☒ Flow Verification (L/s) Flow Column #: <u>ATP106</u> 32 mm 0.128* (.139-.169) 36 mm 0.144* (.156-.190) 53 mm 0.218* (.228-.278) 103 mm 0.496 (.447-.547) ☒ Barometric Pressure Check Gauge ID #: <u>33363</u> Gauge: <u>1018</u> Instrument: <u>1017</u> ☒ Stability Checks <table border="1" style="width: 100%; border-collapse: collapse; margin-top: 5px;"> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot#/Exp</th> </tr> <tr> <td>0.050</td> <td>MP6286</td> <td>202406K 6/19/2026</td> </tr> <tr> <td>0.080</td> <td>MP6287</td> <td>202603B 3/3/2028</td> </tr> <tr> <td>0.200</td> <td>MP6288</td> <td>202406N 6/20/2026</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG526603 9/23/2027</td> </tr> </table>		Simulator	Serial #	Lot#/Exp	0.050	MP6286	202406K 6/19/2026	0.080	MP6287	202603B 3/3/2028	0.200	MP6288	202406N 6/20/2026	0.080 DGS	N/A	AG526603 9/23/2027	Flow Column #: <u>ATP104</u> ☒ 5L/min – 17mm ☒ 15L/min – 53mm ☒ 30L/min – 103mm ☒ R-Value: <u>130</u> ☒ Post Adjustment Verification (L/S) Flow Column #: <u>ATP106</u> 32 mm 0.148 (.139-.169) 36 mm 0.160 (.156-.190) 53 mm 0.238 (.228-.278) 103 mm 0.500 (.447-.547)		<table border="1" style="width: 100%; border-collapse: collapse;"> <tr> <th style="background-color: #e0e0e0;">Maintenance By:</th> <th style="background-color: #e0e0e0;">Date:</th> </tr> <tr> <td>☐ Battery Replacement</td> <td></td> </tr> <tr> <td>☐ Dry Gas Regulator Replacement</td> <td></td> </tr> <tr> <td>☐ Tank Sensor Tare</td> <td></td> </tr> <tr> <td>☐ Breath Tube Replacement</td> <td></td> </tr> <tr> <td>☐ Other:</td> <td></td> </tr> </table>		Maintenance By:	Date:	☐ Battery Replacement		☐ Dry Gas Regulator Replacement		☐ Tank Sensor Tare		☐ Breath Tube Replacement		☐ Other:	
Simulator	Serial #	Lot#/Exp																																
0.050	MP6286	202406K 6/19/2026																																
0.080	MP6287	202603B 3/3/2028																																
0.200	MP6288	202406N 6/20/2026																																
0.080 DGS	N/A	AG526603 9/23/2027																																
Maintenance By:	Date:																																	
☐ Battery Replacement																																		
☐ Dry Gas Regulator Replacement																																		
☐ Tank Sensor Tare																																		
☐ Breath Tube Replacement																																		
☐ Other:																																		
Optical Bench Adjustment By: _____		Department Inspection By: <u>TDG</u>																																
Barometric Pressure Gauge: _____ ID#: _____		Barometric Pressure ID#: <u>33363</u>																																
Simulator	Serial #	Lot #	Expiration	Gauge: <u>1017</u> Instrument: <u>1016</u>																														
0.000		N/A	N/A	Mouth Alcohol Solution Lot #: <u>2025-D</u> Exp: <u>9/25/2027</u>																														
0.040				Acetone Stock Solution Lot #: <u>2024-B</u> Exp: <u>7/19/2026</u>																														
0.100				Simulator	Serial Number																													
0.200				0.000	MP6284																													
0.300				Interferent	MP6285																													
0.080 DGS	N/A			0.050	MP6286																													
☐ Post Optical Bench Adjustment Stability Checks				0.080	MP6287																													
Simulator	Serial #	Lot #	Expiration	0.200	MP6288																													
0.050				Attachments																														
0.080				<table style="width: 100%;"> <tr> <td>☒ Form 41</td> <td>☐ Post-Stability Checks</td> </tr> <tr> <td>☒ Stability Checks</td> <td>☒ Flow Adjustment</td> </tr> <tr> <td>☒ Calibration Certificate</td> <td>☐ Form 40</td> </tr> <tr> <td>☐ Optical Bench Adjustment</td> <td>☒ Other: Additional Stabilities</td> </tr> </table>		☒ Form 41	☐ Post-Stability Checks	☒ Stability Checks	☒ Flow Adjustment	☒ Calibration Certificate	☐ Form 40	☐ Optical Bench Adjustment	☒ Other: Additional Stabilities																					
☒ Form 41	☐ Post-Stability Checks																																	
☒ Stability Checks	☒ Flow Adjustment																																	
☒ Calibration Certificate	☐ Form 40																																	
☐ Optical Bench Adjustment	☒ Other: Additional Stabilities																																	
0.200																																		
0.080 DGS	N/A																																	
Gauge ID #: _____				☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use																														
Gauge: _____ Instrument: _____																																		
Notes/Suggested Service: *Outside nominal. (TDG 6/8/26) Ran additional stabilities after the Department Inspection due to the low DGS readings. Used the same DGS cylinder from the inspection. All results were passing. (TDG 6/8/26)																																		
Tech Review		Admin Review																																

Flow Calibration Adjustment(s)

Performed by TDG

HENDRY COUNTY SO.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001207
06/08/2026
Software: 8100.27

Flow Rate Calibration*****
1: Rate (Liters/min) = 5
SQRT(Diff) = 6.324
2: Rate (Liters/min) = 15
SQRT(Diff) = 11.223
3: Rate (Liters/min) = 30
SQRT(Diff) = 20.711
Dependent Data Scale Factor = 100000 L/min
Independent Data Scale Factor = 256
Rounded Slope = 670
Rounded Intercept = -522239
Correlation = 0.99774

Stability Checks

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																																																																																																																																																
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083																																																																																																																																																
HENDRY COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001207 06/08/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>12:58</td></tr> <tr><td>Control Test</td><td>0.048</td><td>12:59</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:59</td></tr> <tr><td>Control Test</td><td>0.048</td><td>13:00</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:01</td></tr> <tr><td>Control Test</td><td>0.049</td><td>13:01</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:02</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0483</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.1945</td><td></td></tr> </tbody> </table> Operator's Signature 	Test	g/210L	Time	Air Blank	0.000	12:58	Control Test	0.048	12:59	Air Blank	0.000	12:59	Control Test	0.048	13:00	Air Blank	0.000	13:01	Control Test	0.049	13:01	Air Blank	0.000	13:02	Control Test Stats			Average	0.0483		Std Dev	0.0006		Rel Std Dev(%)	1.1945		HENDRY COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001207 06/08/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>12:17</td></tr> <tr><td>Control Test</td><td>0.081</td><td>12:18</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:18</td></tr> <tr><td>Control Test</td><td>0.079</td><td>12:19</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:20</td></tr> <tr><td>Control Test</td><td>0.078</td><td>12:20</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:21</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0793</td><td></td></tr> <tr><td>Std Dev</td><td>0.0015</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.9255</td><td></td></tr> </tbody> </table> Operator's Signature 	Test	g/210L	Time	Air Blank	0.000	12:17	Control Test	0.081	12:18	Air Blank	0.000	12:18	Control Test	0.079	12:19	Air Blank	0.000	12:20	Control Test	0.078	12:20	Air Blank	0.000	12:21	Control Test Stats			Average	0.0793		Std Dev	0.0015		Rel Std Dev(%)	1.9255		HENDRY COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001207 06/08/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>12:25</td></tr> <tr><td>Control Test</td><td>0.201</td><td>12:26</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:26</td></tr> <tr><td>Control Test</td><td>0.200</td><td>12:27</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:28</td></tr> <tr><td>Control Test</td><td>0.200</td><td>12:28</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:29</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.2003</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.2882</td><td></td></tr> </tbody> </table> Operator's Signature 	Test	g/210L	Time	Air Blank	0.000	12:25	Control Test	0.201	12:26	Air Blank	0.000	12:26	Control Test	0.200	12:27	Air Blank	0.000	12:28	Control Test	0.200	12:28	Air Blank	0.000	12:29	Control Test Stats			Average	0.2003		Std Dev	0.0006		Rel Std Dev(%)	0.2882		HENDRY COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001207 06/08/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>13:03</td></tr> <tr><td>Control Test</td><td>0.077</td><td>13:04</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:04</td></tr> <tr><td>Control Test</td><td>0.077</td><td>13:04</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:05</td></tr> <tr><td>Control Test</td><td>0.078</td><td>13:05</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:06</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0773</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7466</td><td></td></tr> </tbody> </table> Operator's Signature 	Test	g/210L	Time	Air Blank	0.000	13:03	Control Test	0.077	13:04	Air Blank	0.000	13:04	Control Test	0.077	13:04	Air Blank	0.000	13:05	Control Test	0.078	13:05	Air Blank	0.000	13:06	Control Test Stats			Average	0.0773		Std Dev	0.0006		Rel Std Dev(%)	0.7466	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	12:58																																																																																																																																																	
Control Test	0.048	12:59																																																																																																																																																	
Air Blank	0.000	12:59																																																																																																																																																	
Control Test	0.048	13:00																																																																																																																																																	
Air Blank	0.000	13:01																																																																																																																																																	
Control Test	0.049	13:01																																																																																																																																																	
Air Blank	0.000	13:02																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0483																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	1.1945																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	12:17																																																																																																																																																	
Control Test	0.081	12:18																																																																																																																																																	
Air Blank	0.000	12:18																																																																																																																																																	
Control Test	0.079	12:19																																																																																																																																																	
Air Blank	0.000	12:20																																																																																																																																																	
Control Test	0.078	12:20																																																																																																																																																	
Air Blank	0.000	12:21																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0793																																																																																																																																																		
Std Dev	0.0015																																																																																																																																																		
Rel Std Dev(%)	1.9255																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	12:25																																																																																																																																																	
Control Test	0.201	12:26																																																																																																																																																	
Air Blank	0.000	12:26																																																																																																																																																	
Control Test	0.200	12:27																																																																																																																																																	
Air Blank	0.000	12:28																																																																																																																																																	
Control Test	0.200	12:28																																																																																																																																																	
Air Blank	0.000	12:29																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.2003																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.2882																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	13:03																																																																																																																																																	
Control Test	0.077	13:04																																																																																																																																																	
Air Blank	0.000	13:04																																																																																																																																																	
Control Test	0.077	13:04																																																																																																																																																	
Air Blank	0.000	13:05																																																																																																																																																	
Control Test	0.078	13:05																																																																																																																																																	
Air Blank	0.000	13:06																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0773																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7466																																																																																																																																																		

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: HENDRY COUNTY SO
Time of Inspection: 15:36

Date of Inspection: 06/08/2026

Serial Number: 80-001207
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202406K Exp: 06/19/2026	0.08g/210L Test (g/210L) Lot#:202603B Exp: 03/03/2028	0.20g/210L Test (g/210L) Lot#:202406N Exp: 06/20/2026	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG526603 Exp: 09/23/2027
0.000	0.049	0.078	0.200	0.076
0.000	0.049	0.078	0.200	0.076
0.000	0.048	0.078	0.200	0.076
0.000	0.049	0.078	0.200	0.076
0.000	0.049	0.078	0.200	0.075
0.000	0.049	0.078	0.199	0.075
0.000	0.048	0.078	0.200	0.075
0.000	0.049	0.078	0.200	0.075
0.000	0.048	0.078	0.199	0.075
0.000	0.048	0.078	0.199	0.075

Standard Deviations	0.0005	0.0000	0.0004	0.0005
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

06/08/2026
Date

Extra DGS Stabilities

HENDRY COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001207
06/08/2026
Software: 8100.27

TAG

Test	g/210L	Time
Air Blank	0.000	15:39
Control Test	0.078 ✓	15:39
Air Blank	0.000	15:39
Control Test	0.078 ✓	15:40
Air Blank	0.000	15:40
Control Test	0.078 ✓	15:41
Air Blank	0.000	15:41
Control Test	0.078 ✓	15:41
Air Blank	0.000	15:42
Control Test	0.078 ✓	15:42
Air Blank	0.000	15:43
Control Test	0.078 ✓	15:43
Air Blank	0.000	15:44
Control Test	0.078 ✓	15:44
Air Blank	0.000	15:45
Control Test	0.078 ✓	15:45
Air Blank	0.000	15:45
Control Test	0.079 ✓	15:46
Air Blank	0.000	15:46
Control Test	0.079 ✓	15:47
Air Blank	0.000	15:47
Control Test	0.078 ✓	15:47
Air Blank	0.000	15:48
Control Test	0.079 ✓	15:48
Air Blank	0.000	15:49
Control Test	0.078 ✓	15:49
Air Blank	0.000	15:50
Control Test	0.079 ✓	15:50
Air Blank	0.000	15:51
Control Test	0.078 ✓	15:51
Air Blank	0.000	15:51
Control Test	0.079 ✓	15:52
Air Blank	0.000	15:52
Control Test	0.079 ✓	15:53
Air Blank	0.000	15:53
Control Test	0.078 ✓	15:53
Air Blank	0.000	15:54
Control Test	0.079 ✓	15:54
Air Blank	0.000	15:55
Control Test	0.079 ✓	15:55
Air Blank	0.000	15:56
Control Test Stats		
Average	0.0784	
Std Dev	0.0005	
Rel Std Dev(%)	0.6411	



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001207, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001207</u>	UNCERTAINTY* $\pm$
Owning Agency:	<u>HENDRY COUNTY SO</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>06/08/2026</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>15:36</u>	0.200 g/ 210 L 0.008
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Taylor
Gutschow
Digitally signed by Taylor
Gutschow
Date: 2026.06.08 16:42:47
+04'00'

06/08/2026

Date

TAYLOR D GUTSCHOW,
Department Inspector

FDLE/ATP Form 69 January 2026

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality