

Alcohol Testing Program - Instrument Processing Sheet

Agency: Columbia County SO

Instrument Serial Number: 80-000777

Date In: 6/5/2026

DI Completion Date: 6/9/2026

☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake By: IAS Date: 6/5/2026		Quality Checks By: WKP Date: 6/9/2026		Flow Adjustment By: _____																																													
☒ Annual ☐ Dropped Off ☐ Registration ☐ Return from CMI / EE ☐ Training Instrument Visual Inspection ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/Accessories ☐ Power Cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes:		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value: 177 ☒ Flow Verification (L/s) Flow Column #: ATP 105 32 mm 0.144 (.139-.169) 36 mm 0.164 (.156-.190) 53 mm 0.238 (.228-.278) 103 mm 0.511 (.447-.547) ☒ Barometric Pressure Check Gauge ID #: 28662 Gauge: 1018 Instrument: 1015 ☒ Stability Checks		Flow Column #: _____ ☐ 5L/min – 17mm ☐ 15L/min – 53mm ☐ 30L/min – 103mm ☐ R-Value: _____ ☐ Post Adjustment Verification (L/S) Flow Column #: _____ 32 mm (.139-.169) 36 mm (.156-.190) 53 mm (.228-.278) 103 mm (.447-.547)																																													
						<table border="1" style="width:100%; border-collapse: collapse;"> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot#/Exp</th> </tr> <tr> <td>0.050</td> <td>MP5088</td> <td>202603A 3/3/2028</td> </tr> <tr> <td>0.080</td> <td>MP5089</td> <td>202603B 3/3/2028</td> </tr> <tr> <td>0.200</td> <td>MP5090</td> <td>202603D 3/4/2028</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>23825080A1 10/5/2027</td> </tr> </table>		Simulator	Serial #	Lot#/Exp	0.050	MP5088	202603A 3/3/2028	0.080	MP5089	202603B 3/3/2028	0.200	MP5090	202603D 3/4/2028	0.080 DGS	N/A	23825080A1 10/5/2027																											
						Simulator	Serial #	Lot#/Exp																																									
						0.050	MP5088	202603A 3/3/2028																																									
						0.080	MP5089	202603B 3/3/2028																																									
0.200	MP5090	202603D 3/4/2028																																															
0.080 DGS	N/A	23825080A1 10/5/2027																																															
Maintenance By: WKP Date: 6/9/2026 ☒ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Tank Sensor Tare ☐ Breath Tube Replacement ☐ Other:																																																	
Optical Bench Adjustment By: _____ Barometric Pressure Gauge: _____ ID#: _____ <table border="1" style="width:100%; border-collapse: collapse;"> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </table> ☐ Post Optical Bench Adjustment Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> <tr> <td>0.050</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </table> Gauge ID #: _____ Gauge: _____ Instrument: _____		Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A		
Simulator	Serial #	Lot #	Expiration																																														
0.000		N/A	N/A																																														
0.040																																																	
0.100																																																	
0.200																																																	
0.300																																																	
0.080 DGS	N/A																																																
Simulator	Serial #	Lot #	Expiration																																														
0.050																																																	
0.080																																																	
0.200																																																	
0.080 DGS	N/A																																																
Notes/Suggested Service:		Department Inspection By: WKP Barometric Pressure ID#: 34419 Gauge: 1018 Instrument: 1016 Mouth Alcohol Solution Lot #: 2025-D Exp: 9/25/2027 Acetone Stock Solution Lot #: 2025-B Exp: 9/22/2027 <table border="1" style="width:100%; border-collapse: collapse;"> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> <tr> <td>0.000</td> <td>MP5086</td> </tr> <tr> <td>Interferent</td> <td>MP5087</td> </tr> <tr> <td>0.050</td> <td>MP5088</td> </tr> <tr> <td>0.080</td> <td>MP5089</td> </tr> <tr> <td>0.200</td> <td>MP5090</td> </tr> </table>		Simulator	Serial Number	0.000	MP5086	Interferent	MP5087	0.050	MP5088	0.080	MP5089	0.200	MP5090																																		
		Simulator	Serial Number																																														
		0.000	MP5086																																														
		Interferent	MP5087																																														
		0.050	MP5088																																														
0.080	MP5089																																																
0.200	MP5090																																																
<table border="1" style="width:100%; border-collapse: collapse;"> <tr> <td style="width: 50%;"> ☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☐ Optical Bench Adjustment </td> <td style="width: 50%;"> ☐ Post-Stability Checks ☐ Flow Adjustment ☐ Form 40 ☐ Other: </td> </tr> </table>		☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☐ Optical Bench Adjustment	☐ Post-Stability Checks ☐ Flow Adjustment ☐ Form 40 ☐ Other:																																														
☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☐ Optical Bench Adjustment	☐ Post-Stability Checks ☐ Flow Adjustment ☐ Form 40 ☐ Other:																																																
☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use																																																	
<table style="width:100%;"> <tr> <td style="width: 50%;"> LeAndra Higginbotham Digitally signed by LeAndra Higginbotham Date: 2026.06.14 14:04:55 -04'00' </td> <td style="width: 50%;"> Kaitlyn Spearin Digitally signed by Kaitlyn Spearin Date: 2026.06.15 07:21:59 -04'00' </td> </tr> </table>		LeAndra Higginbotham Digitally signed by LeAndra Higginbotham Date: 2026.06.14 14:04:55 -04'00'	Kaitlyn Spearin Digitally signed by Kaitlyn Spearin Date: 2026.06.15 07:21:59 -04'00'																																														
LeAndra Higginbotham Digitally signed by LeAndra Higginbotham Date: 2026.06.14 14:04:55 -04'00'	Kaitlyn Spearin Digitally signed by Kaitlyn Spearin Date: 2026.06.15 07:21:59 -04'00'																																																
Tech Review Admin Review																																																	

Stability Checks

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083
COLUMBIA COUNTY SO Intoxilyzer - Alcotest Analyzer Model 8000 06/09/2026 Software: 8100.27 SN 80-000777 Test g/210L Time Air Blank 0.000 10:39 Control Test 0.049 10:40 Air Blank 0.000 10:40 Control Test 0.049 10:41 Air Blank 0.000 10:41 Control Test 0.048 10:42 Air Blank 0.000 10:42 Control Test 0.048 10:42 Average 0.048 Std Dev 0.000 Rel Std Dev(%) 0.000 <i>Wentz</i> Operator's Signature	COLUMBIA COUNTY SO Intoxilyzer - Alcotest Analyzer Model 8000 06/09/2026 Software: 8100.27 SN 80-000777 Test g/210L Time Air Blank 0.000 10:45 Control Test 0.077 10:46 Air Blank 0.000 10:46 Control Test 0.077 10:47 Air Blank 0.000 10:47 Control Test 0.077 10:48 Air Blank 0.000 10:49 Control Test 0.077 10:49 Average 0.077 Std Dev 0.000 Rel Std Dev(%) 0.000 <i>Wentz</i> Operator's Signature	COLUMBIA COUNTY SO Intoxilyzer - Alcotest Analyzer Model 8000 06/09/2026 Software: 8100.27 SN 81-000777 Test g/210L Time Air Blank 0.000 10:50 Control Test 0.196 10:50 Air Blank 0.000 10:51 Control Test 0.196 10:52 Air Blank 0.000 10:52 Control Test 0.196 10:53 Air Blank 0.000 10:53 Control Test 0.196 10:53 Average 0.196 Std Dev 0.000 Rel Std Dev(%) 0.000 <i>Wentz</i> Operator's Signature	COLUMBIA COUNTY SO Intoxilyzer - Alcotest Analyzer Model 8000 06/09/2026 Software: 8100.27 SN 80-000777 Test g/210L Time Air Blank 0.000 10:54 Control Test 0.079 10:54 Air Blank 0.000 10:55 Control Test 0.080 10:55 Air Blank 0.000 10:56 Control Test 0.079 10:56 Air Blank 0.000 10:57 Control Test 0.079 10:57 Average 0.079 Std Dev 0.006 Rel Std Dev(%) 0.7277 <i>Wentz</i> Operator's Signature

Florida Department of Law Enforcement

Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: COLUMBIA COUNTY SO
Time of Inspection: 15:31

Date of Inspection: 06/09/2026

Serial Number: 80-000777
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202603A Exp: 03/03/2028	0.08g/210L Test (g/210L) Lot#:202603B Exp: 03/03/2028	0.20g/210L Test (g/210L) Lot#:202603D Exp: 03/04/2028	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:23825080A1 Exp: 10/05/2027
0.000	0.048	0.076	0.195	0.078
0.000	0.048	0.076	0.195	0.078
0.000	0.049	0.076	0.195	0.079
0.000	0.049	0.076	0.195	0.078
0.000	0.048	0.077	0.195	0.078
0.000	0.049	0.076	0.194	0.079
0.000	0.048	0.076	0.194	0.079
0.000	0.049	0.076	0.195	0.079
0.000	0.049	0.076	0.195	0.079
0.000	0.049	0.076	0.194	0.079


Standard Deviations	0.0005	0.0003	0.0004	0.0005
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



WEN-CHI K. PIERSON

Signature and Printed Name

06/09/2026
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000777, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-000777</u>	UNCERTAINTY* $\pm$
Owning Agency:	<u>COLUMBIA COUNTY SO</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>06/09/2026</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>15:31</u>	0.200 g/ 210 L 0.008
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Wen-Chi

Chi Pierson
Date: 2026.06.10
09:08:31 -04'00'

Pierson

06/09/2026

Date

WEN-CHI K PIERSON,

Department Inspector

FDLE/ATP Form 69 January 2026

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality