

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: GULF COUNTY SO
Time of Inspection: 21:26

Date of Inspection: 04/28/2023

Serial Number: 80-000795
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202112A Exp: 12/07/2023	0.08g/210L Test (g/210L) Lot#:202108D Exp: 08/17/2023	0.20g/210L Test (g/210L) Lot#:202106B Exp: 06/22/2023	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:402607186 Exp: 12/02/2025
0.000 / 0.000	0.047	0.080	0.201	0.081 / 0.081
0.007 / 0.000	0.055	0.080	0.200	0.087 / 0.082
0.000 / 0.000	0.045	0.082	0.201	0.087 / 0.082

Number of Simulators Used: 5

Remarks:

00: Control Outside Tolerance. 08: Control Outside Tolerance.UNKNOWN ERROR

*.05 reference was done without issue on PSTPD intoxilyzer. Repeat yielded no errors.
Repeat of dry gas had no issue. Reason for anomaly unknown.*

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

OFC. Dalton Brammer #109

DALTON J BRAMMER

Signature and Printed Name

04/28/2023
Date