

Instrument Processing Sheet

Agency: ORANGE COUNTY SO Instrument Serial Number: 80-001418
 Date In: 12/19/2025 DI Completion Date: 12/30/2025 Ship P/U H/D CMI EE

Intake By: <u>WKP</u> Date: <u>12/19/25</u>	Quality Checks By: <u>SLH</u> Date: <u>12/23/25</u>	Flow Adjustment By: <u>SLH</u> Date: <u>12/29/2025</u>
<input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE <input type="checkbox"/> Return unworked <input type="checkbox"/> Training Visual Inspection <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/Accessories <input type="checkbox"/> Power Cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable	<input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value: <u>210</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column #: <u>ATP 103</u> 32 mm <u>0.148</u> (.139-.169) 36 mm <u>0.160</u> (.156-.190) 53 mm <u>0.226</u> (.228-.278) 103 mm <u>0.496</u> (.447-.547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID #: <u>28421</u> Gauge: <u>1022</u> Instrument: <u>1019</u> <input checked="" type="checkbox"/> Stability Checks	Flow Column #: <u>ATP 102 / ATP 102</u> <input checked="" type="checkbox"/> 5L/min – 17mm <input checked="" type="checkbox"/> 15L/min – 53mm <input checked="" type="checkbox"/> 30L/min – 103mm <input checked="" type="checkbox"/> R-Value: <u>216 / 216</u> <input checked="" type="checkbox"/> Post Adjustment Verification (L/S) Flow Column #: <u>ATP 103 / ATP 103</u> 32 mm <u>0.113 / 0.148</u> (.139-.169) 36 mm <u>0.152 / 0.160</u> (.156-.190) 53 mm <u>0.222 / 0.230</u> (.228-.278) 103 mm <u>0.527 / 0.496</u> (.447-.547)

Notes:	<table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot#/Exp</th> </tr> </thead> <tbody> <tr> <td rowspan="2">0.050</td> <td rowspan="2">MP5088</td> <td>202406K</td> </tr> <tr> <td>06/19/2026</td> </tr> <tr> <td rowspan="2">0.080</td> <td rowspan="2">MP5089</td> <td>202406L</td> </tr> <tr> <td>06/19/2026</td> </tr> <tr> <td rowspan="2">0.200</td> <td rowspan="2">MP5090</td> <td>202406N</td> </tr> <tr> <td>06/20/2026</td> </tr> <tr> <td rowspan="2">0.080 DGS</td> <td rowspan="2">N/A</td> <td>AG510701</td> </tr> <tr> <td>04/17/2027</td> </tr> </tbody> </table>	Simulator	Serial #	Lot#/Exp	0.050	MP5088	202406K	06/19/2026	0.080	MP5089	202406L	06/19/2026	0.200	MP5090	202406N	06/20/2026	0.080 DGS	N/A	AG510701	04/17/2027	Maintenance By: _____ Date: _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement and Tank Sensor Tare <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other:
Simulator	Serial #	Lot#/Exp																			
0.050	MP5088	202406K																			
		06/19/2026																			
0.080	MP5089	202406L																			
		06/19/2026																			
0.200	MP5090	202406N																			
		06/20/2026																			
0.080 DGS	N/A	AG510701																			
		04/17/2027																			

Optical Bench Adjustment By: <u>SLH</u>	Department Inspection By: <u>SLH</u>																																								
Barometric Pressure Gauge: <u>1014</u> ID#: <u>28427</u>	Barometric Pressure ID#: <u>28662</u>																																								
<table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>MP6289</td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td>MP6295</td> <td>25090</td> <td>03/11/2027</td> </tr> <tr> <td>0.100</td> <td>MP6296</td> <td>25282</td> <td>08/12/2027</td> </tr> <tr> <td>0.200</td> <td>MP6297</td> <td>24080</td> <td>02/13/2026</td> </tr> <tr> <td>0.300</td> <td>MP6298</td> <td>25150</td> <td>05/06/27</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>28424080A3</td> <td>11/05/2026</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #	Expiration	0.000	MP6289	N/A	N/A	0.040	MP6295	25090	03/11/2027	0.100	MP6296	25282	08/12/2027	0.200	MP6297	24080	02/13/2026	0.300	MP6298	25150	05/06/27	0.080 DGS	N/A	28424080A3	11/05/2026	Gauge: <u>1019</u> Instrument: <u>1019</u> Mouth Alcohol Solution Lot #: <u>2025-D</u> Exp: <u>09/25/2027</u> Acetone Stock Solution Lot #: <u>2025-B</u> Exp: <u>09/22/2027</u> <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>MP6294</td> </tr> <tr> <td>Interferent</td> <td>MP6290</td> </tr> <tr> <td>0.050</td> <td>MP5088</td> </tr> <tr> <td>0.080</td> <td>MP5089</td> </tr> <tr> <td>0.200</td> <td>MP5090</td> </tr> </tbody> </table>	Simulator	Serial Number	0.000	MP6294	Interferent	MP6290	0.050	MP5088	0.080	MP5089	0.200	MP5090
Simulator	Serial #	Lot #	Expiration																																						
0.000	MP6289	N/A	N/A																																						
0.040	MP6295	25090	03/11/2027																																						
0.100	MP6296	25282	08/12/2027																																						
0.200	MP6297	24080	02/13/2026																																						
0.300	MP6298	25150	05/06/27																																						
0.080 DGS	N/A	28424080A3	11/05/2026																																						
Simulator	Serial Number																																								
0.000	MP6294																																								
Interferent	MP6290																																								
0.050	MP5088																																								
0.080	MP5089																																								
0.200	MP5090																																								
<input checked="" type="checkbox"/> Post Optical Bench Adjustment Stability Checks																																									
<table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>MP5088</td> <td>202406K</td> <td>06/19/2026</td> </tr> <tr> <td>0.080</td> <td>MP5089</td> <td>202406L</td> <td>06/19/2026</td> </tr> <tr> <td>0.200</td> <td>MP5090</td> <td>202406N</td> <td>06/20/2026</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG510701</td> <td>04/17/2027</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #	Expiration	0.050	MP5088	202406K	06/19/2026	0.080	MP5089	202406L	06/19/2026	0.200	MP5090	202406N	06/20/2026	0.080 DGS	N/A	AG510701	04/17/2027	Attachments <input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Stability Checks <input checked="" type="checkbox"/> Calibration Certificate <input checked="" type="checkbox"/> Optical Bench Adjustment <input checked="" type="checkbox"/> Post-Stability Checks <input checked="" type="checkbox"/> Flow Adjustment (x2) <input type="checkbox"/> Form 40 <input checked="" type="checkbox"/> Other: stability to assess DGS results																				
Simulator	Serial #	Lot #	Expiration																																						
0.050	MP5088	202406K	06/19/2026																																						
0.080	MP5089	202406L	06/19/2026																																						
0.200	MP5090	202406N	06/20/2026																																						
0.080 DGS	N/A	AG510701	04/17/2027																																						
Barometric Pressure Gauge: <u>1014</u> ID#: <u>28421</u>																																									

Notes/Suggested Service: 12/23/25 performed 10 replicates for 0.080 g/210L DGS to assess if INT DETECT triggered as told by AI. SLH 12/23/25 Flow Verification and stabilities for quality checks performed 12/29/25. SLH 12/29/2025	<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use
Taylor Gutschow <small>Digitally signed by Taylor Gutschow Date: 2026.01.13 13:37:48 -05'00'</small>	Wen-Chi Pierson <small>Digitally signed by Wen-Chi Pierson Date: 2026.01.14 09:18:40 -05'00'</small>
Tech Review	Admin Review

80-001418

D9S

ORANGE COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001418
12/23/2025
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	16:21
Control Test	0.082	16:22
Air Blank	0.000	16:22
Control Test	0.083	16:22
Air Blank	0.000	16:23
Control Test	0.082	16:23
Air Blank	0.000	16:24
Control Test	0.082	16:24
Air Blank	0.000	16:24
Control Test	0.082	16:25
Air Blank	0.000	16:25
Control Test	0.083	16:26
Air Blank	0.000	16:26
Control Test	0.082	16:26
Air Blank	0.000	16:27
Control Test	0.082	16:27
Air Blank	0.000	16:28
Control Test	0.082	16:28
Air Blank	0.000	16:28
Control Test	0.083	16:29
Air Blank	0.000	16:29
Control Test Stats		
Average	0.0823	
Std Dev	0.0005	
Rel Std Dev(%)	0.5869	

10 stability checks
of 0.0809/210L D9S
to assess if
INT DETECT was
triggered as ~~note~~ ^{SUT}
agency indicated.
SUT
12/23/2025


Operator's Signature

Flow Calibration

Root Cause Analysis Performed Prior



Performed by SLH

#1

ORANGE COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001418
12/29/2025
Software: 8100.27

Flow Rate Calibration*****

1: Rate (Liters/min) = 5
SQRT(Diff)) = 7.680
2: Rate (Liters/min) = 15
SQRT(Diff)) = 11.617
3: Rate (Liters/min) = 30
SQRT(Diff)) = 20.637
Dependent Data Scale Factor = 100000 L/min
Independent Data Scale Factor = 256
Rounded Slope = 736
Rounded Intercept = -840271
Correlation = 0.99420

post verification failed -
repeated flow
calibration

Set

12/29/25

#2

ORANGE COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001418
12/29/2025
Software: 8100.27

Flow Rate Calibration*****

1: Rate (Liters/min) = 5
SQRT(Diff)) = 7.277
2: Rate (Liters/min) = 15
SQRT(Diff)) = 11.445
3: Rate (Liters/min) = 30
SQRT(Diff)) = 20.395
Dependent Data Scale Factor = 100000 L/min
Independent Data Scale Factor = 256
Rounded Slope = 730
Rounded Intercept = -770879
Correlation = 0.99573

Stability Checks

80-001418

SUB 12/29/25

perform calibration adjustment

0.050 g/210L	0.080 g/210L	0.200 g/210L	DGS 0.080 g/210L
<p>0.047 to 0.053 g/210L <input checked="" type="checkbox"/></p> <p>0.077 to 0.083 g/210L <input checked="" type="checkbox"/></p> <p>0.194 to 0.206 g/210L <input checked="" type="checkbox"/></p> <p>0.077 to 0.083 g/210L <input checked="" type="checkbox"/></p> <p>50.003 g/210L of Wet <input checked="" type="checkbox"/></p>	<p>0.077 to 0.083 g/210L <input checked="" type="checkbox"/></p> <p>0.194 to 0.206 g/210L <input checked="" type="checkbox"/></p> <p>0.077 to 0.083 g/210L <input checked="" type="checkbox"/></p> <p>50.003 g/210L of Wet <input checked="" type="checkbox"/></p>	<p>0.077 to 0.083 g/210L <input checked="" type="checkbox"/></p> <p>0.194 to 0.206 g/210L <input checked="" type="checkbox"/></p> <p>0.077 to 0.083 g/210L <input checked="" type="checkbox"/></p> <p>50.003 g/210L of Wet <input checked="" type="checkbox"/></p>	<p>0.077 to 0.083 g/210L <input checked="" type="checkbox"/></p> <p>0.194 to 0.206 g/210L <input checked="" type="checkbox"/></p> <p>0.077 to 0.083 g/210L <input checked="" type="checkbox"/></p> <p>50.003 g/210L of Wet <input checked="" type="checkbox"/></p>
<p>Performed Root Case Analysis</p> <p>ORANGE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 12/29/2025 Software: 8100.27</p>	<p>Performed Root Case Analysis</p> <p>ORANGE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 12/29/2025 Software: 8100.27</p>	<p>Performed Root Case Analysis</p> <p>ORANGE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 12/29/2025 Software: 8100.27</p>	<p>Performed Root Case Analysis</p> <p>ORANGE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 12/29/2025 Software: 8100.27</p>
<p>Test</p> <p>Air Blank 0.000 16:49</p> <p>Control Test 0.052 16:50</p> <p>Air Blank 0.000 16:50</p> <p>Control Test 0.050 16:51</p> <p>Air Blank 0.000 16:52</p> <p>Control Test 0.050 16:52</p> <p>Air Blank 0.000 16:53</p> <p>Control Test Stats</p> <p>Average 0.0507</p> <p>Std Dev 0.0112</p> <p>Rel Std Dev(%) 2.2791</p>	<p>Test</p> <p>Air Blank 0.000 16:28</p> <p>Control Test 0.080 16:29</p> <p>Air Blank 0.000 16:29</p> <p>Control Test 0.079 16:30</p> <p>Air Blank 0.000 16:30</p> <p>Control Test 0.079 16:31</p> <p>Air Blank 0.000 16:32</p> <p>Control Test Stats</p> <p>Average 0.0793</p> <p>Std Dev 0.0006</p> <p>Rel Std Dev(%) 0.7277</p>	<p>Test</p> <p>Air Blank 0.000 16:22</p> <p>Control Test 0.139 16:23</p> <p>Air Blank 0.000 16:23</p> <p>Control Test 0.138 16:24</p> <p>Air Blank 0.000 16:24</p> <p>Control Test 0.137 16:25</p> <p>Air Blank 0.000 16:26</p> <p>Control Test Stats</p> <p>Average 0.1380</p> <p>Std Dev 0.0110</p> <p>Rel Std Dev(%) 0.5351</p>	<p>Test</p> <p>Air Blank 0.000 16:33</p> <p>Control Test 0.083 16:34</p> <p>Air Blank 0.000 16:34</p> <p>Control Test 0.083 16:34</p> <p>Air Blank 0.000 16:35</p> <p>Control Test 0.084 16:35</p> <p>Air Blank 0.000 16:36</p> <p>Control Test Stats</p> <p>Average 0.0833</p> <p>Std Dev 0.0006</p> <p>Rel Std Dev(%) 0.5928</p>
<p>Operator's Signature</p> <p><i>Mijalok</i></p>	<p>Operator's Signature</p> <p><i>Mijalok</i></p>	<p>Operator's Signature</p> <p><i>Mijalok</i></p>	<p>Operator's Signature</p> <p><i>Mijalok</i></p>

D95

ORANGE COUNTY SO

Intoxilyzer - Alcohol Analyzer

Model 8000

12/29/2025

SN 80-001418

16:54:15

- *SWT*

Auto Calibration

pg 1 of 2

<<<<< 3um >>>>> <<<<< 9um >>>>>

Solution = 0.000 g/210L or 0.0000 mg/l, Samples = 4, Discarded = 1

Sample	% Abs	(% Abs Ref)	% Abs	(% Abs Ref)
Sample #1	0.1800	(-0.0190)	0.3520	(-0.0290)
Sample #2	0.1500	(0.0010)	0.3030	(-0.0120)
Sample #3	0.1550	(0.0140)	0.2970	(-0.0100)
Sample #4	0.1580	(0.0480)	0.3090	(-0.0080)
Avg % Abs	0.1543	(0.0210)	0.3030	(-0.0100)
STD DEV	0.0040	(0.0243)	0.0060	(0.0020)
REL STD DEV	2.619	(115.568)	1.980	(20.000)

Solution = 0.040 g/210L or 0.1905 mg/l, Samples = 4, Discarded = 1

Sample	% Abs	(% Abs Ref)	% Abs	(% Abs Ref)
Sample #1	0.8640	(-0.0200)	1.6910	(-0.0130)
Sample #2	0.8560	(0.0030)	1.6760	(0.0060)
Sample #3	0.8470	(0.0050)	1.6730	(0.0050)
Sample #4	0.8300	(0.0240)	1.6600	(0.0250)
Avg % Abs	0.8443	(0.0107)	1.6697	(0.0120)
STD DEV	0.0132	(0.0116)	0.0085	(0.0113)
REL STD DEV	1.564	(108.658)	0.509	(93.912)

Solution = 0.100 g/210L or 0.4762 mg/l, Samples = 4, Discarded = 1

Sample	% Abs	(% Abs Ref)	% Abs	(% Abs Ref)
Sample #1	1.8640	(-0.0010)	3.7130	(-0.0020)
Sample #2	1.8660	(-0.0030)	3.6850	(0.0170)
Sample #3	1.8660	(0.0050)	3.6850	(0.0310)
Sample #4	1.8790	(0.0290)	3.6860	(0.0490)
Avg % Abs	1.8703	(0.0103)	3.6853	(0.0323)
STD DEV	0.0075	(0.0167)	0.0006	(0.0160)
REL STD DEV	0.401	(161.161)	0.016	(49.613)

Solution = 0.200 g/210L or 0.9524 mg/l, Samples = 4, Discarded = 1

Sample	% Abs	(% Abs Ref)	% Abs	(% Abs Ref)
Sample #1	3.6040	(-0.0110)	6.9770	(-0.0090)
Sample #2	3.5550	(0.0000)	6.8790	(0.0500)
Sample #3	3.5670	(0.0260)	6.8980	(0.0580)
Sample #4	3.5600	(0.0330)	6.9050	(0.0660)
Avg % Abs	3.5607	(0.0197)	6.8940	(0.0580)
STD DEV	0.0060	(0.0174)	0.0135	(0.0080)
REL STD DEV	0.169	(88.412)	0.195	(13.793)

Solution = 0.300 g/210L or 1.4286 mg/l, Samples = 4, Discarded = 1

Sample	% Abs	(% Abs Ref)	% Abs	(% Abs Ref)
Sample #1	5.2430	(-0.0070)	10.0310	(0.0050)
Sample #2	5.1870	(0.0270)	9.9430	(0.0790)
Sample #3	5.1760	(0.0540)	9.9510	(0.0900)
Sample #4	5.1930	(0.0270)	9.9250	(0.1040)
Avg % Abs	5.1853	(0.0360)	9.9397	(0.0910)
STD DEV	0.0086	(0.0156)	0.0133	(0.0125)
REL STD DEV	0.166	(43.301)	0.134	(13.769)

ORANGE COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001418
 12/29/2025 16:54:15

Auto Calibration

pg 2 of 2

<<<<< 3um >>>>>

 Zero Order Coef -409.53
 First Order Coef 2713.60
 Second Order Coef 22.95

<<<<< 9um >>>>>

 -405.96
 1357.89
 12.06

Act (g/210L)	Fit (g/210L)	Residual (g/210L)
0.000	0.000	-0.0002
0.040	0.040	0.0001
0.100	0.100	0.0003
0.200	0.200	-0.0004
0.300	0.300	0.0002

Act (g/210L)	Fit (g/210L)	Residual (g/210L)
0.000	0.000	-0.0001
0.040	0.040	0.0002
0.100	0.100	-0.0000
0.200	0.200	-0.0001
0.300	0.300	0.0001

<<<<< 3um >>>>> <<<<< 9um >>>>>

 Solution = 0.080 g/210L or 0.3810 mg/l, Samples = 4, Discarded = 1
 Sample

Sample #1	3160.00	3104.00
Sample #2	3061.00	3136.00
Sample #3	3077.00	3104.00
Sample #4	3116.00	3104.00
Avg	3084.6667	3114.6667
STD DEV	28.2902	18.4752
REL STD DEV	0.917	0.593
H2O adjust (mg/l*10k)	725	695

Barometric Pressure = 1014

*****CALIBRATION SUCCESSFUL*****

SW

Post-Calibration Adjustment Stability Checks

80-001418
Sub 12/29/25

0.050 g/210L	0.080 g/210L	0.200 g/210L	DGS 0.080 g/210L																																																																																																																																				
<p>0.047 to 0.053 g/210L</p> <p>Performed Root Case Analysis</p>	<p>0.077 to 0.083 g/210L</p> <p>Performed Root Case Analysis</p>	<p>0.194 to 0.206 g/210L</p> <p>Performed Root Case Analysis</p>	<p>0.077 to 0.083 g/210L</p> <p>50.003 g/210L of Wet</p> <p>Performed Root Case Analysis</p>																																																																																																																																				
<p>ORANGE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001418 12/29/2025 Software: 8100.27</p> <p>Test g/210L Time</p> <table border="1"> <tr><td>Air Blank</td><td>0.010</td><td>17:58</td></tr> <tr><td>Control Test</td><td>0.051</td><td>17:59</td></tr> <tr><td>Air Blank</td><td>0.010</td><td>17:59</td></tr> <tr><td>Control Test</td><td>0.050</td><td>18:00</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>18:01</td></tr> <tr><td>Control Test</td><td>0.049</td><td>18:01</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>18:02</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0500</td><td></td></tr> <tr><td>Std Dev</td><td>0.0010</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>2.0000</td><td></td></tr> </table>	Air Blank	0.010	17:58	Control Test	0.051	17:59	Air Blank	0.010	17:59	Control Test	0.050	18:00	Air Blank	0.000	18:01	Control Test	0.049	18:01	Air Blank	0.000	18:02	Control Test Stats			Average	0.0500		Std Dev	0.0010		Rel Std Dev(%)	2.0000		<p>ORANGE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001418 12/29/2025 Software: 8100.27</p> <p>Test g/210L Time</p> <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>17:48</td></tr> <tr><td>Control Test</td><td>0.083</td><td>17:49</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>17:49</td></tr> <tr><td>Control Test</td><td>0.081</td><td>17:50</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>17:51</td></tr> <tr><td>Control Test</td><td>0.080</td><td>17:51</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>17:52</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0813</td><td></td></tr> <tr><td>Std Dev</td><td>0.0015</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.8781</td><td></td></tr> </table>	Air Blank	0.000	17:48	Control Test	0.083	17:49	Air Blank	0.000	17:49	Control Test	0.081	17:50	Air Blank	0.000	17:51	Control Test	0.080	17:51	Air Blank	0.000	17:52	Control Test Stats			Average	0.0813		Std Dev	0.0015		Rel Std Dev(%)	1.8781		<p>ORANGE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001418 12/29/2025 Software: 8100.27</p> <p>Test g/210L Time</p> <table border="1"> <tr><td>Air Blank</td><td>0.300</td><td>17:53</td></tr> <tr><td>Control Test</td><td>0.203</td><td>17:54</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>17:55</td></tr> <tr><td>Control Test</td><td>0.201</td><td>17:55</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>17:56</td></tr> <tr><td>Control Test</td><td>0.200</td><td>17:56</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>17:57</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.203</td><td></td></tr> <tr><td>Std Dev</td><td>0.005</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7587</td><td></td></tr> </table>	Air Blank	0.300	17:53	Control Test	0.203	17:54	Air Blank	0.000	17:55	Control Test	0.201	17:55	Air Blank	0.000	17:56	Control Test	0.200	17:56	Air Blank	0.000	17:57	Control Test Stats			Average	0.203		Std Dev	0.005		Rel Std Dev(%)	0.7587		<p>ORANGE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001418 12/29/2025 Software: 8100.27</p> <p>Test g/210L Time</p> <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>17:44</td></tr> <tr><td>Control Test</td><td>0.080</td><td>17:44</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>17:45</td></tr> <tr><td>Control Test</td><td>0.080</td><td>17:45</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>17:46</td></tr> <tr><td>Control Test</td><td>0.080</td><td>17:46</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>17:46</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0800</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </table>	Air Blank	0.000	17:44	Control Test	0.080	17:44	Air Blank	0.000	17:45	Control Test	0.080	17:45	Air Blank	0.000	17:46	Control Test	0.080	17:46	Air Blank	0.000	17:46	Control Test Stats			Average	0.0800		Std Dev	0.0000		Rel Std Dev(%)	0.0000	
Air Blank	0.010	17:58																																																																																																																																					
Control Test	0.051	17:59																																																																																																																																					
Air Blank	0.010	17:59																																																																																																																																					
Control Test	0.050	18:00																																																																																																																																					
Air Blank	0.000	18:01																																																																																																																																					
Control Test	0.049	18:01																																																																																																																																					
Air Blank	0.000	18:02																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.0500																																																																																																																																						
Std Dev	0.0010																																																																																																																																						
Rel Std Dev(%)	2.0000																																																																																																																																						
Air Blank	0.000	17:48																																																																																																																																					
Control Test	0.083	17:49																																																																																																																																					
Air Blank	0.000	17:49																																																																																																																																					
Control Test	0.081	17:50																																																																																																																																					
Air Blank	0.000	17:51																																																																																																																																					
Control Test	0.080	17:51																																																																																																																																					
Air Blank	0.000	17:52																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.0813																																																																																																																																						
Std Dev	0.0015																																																																																																																																						
Rel Std Dev(%)	1.8781																																																																																																																																						
Air Blank	0.300	17:53																																																																																																																																					
Control Test	0.203	17:54																																																																																																																																					
Air Blank	0.000	17:55																																																																																																																																					
Control Test	0.201	17:55																																																																																																																																					
Air Blank	0.000	17:56																																																																																																																																					
Control Test	0.200	17:56																																																																																																																																					
Air Blank	0.000	17:57																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.203																																																																																																																																						
Std Dev	0.005																																																																																																																																						
Rel Std Dev(%)	0.7587																																																																																																																																						
Air Blank	0.000	17:44																																																																																																																																					
Control Test	0.080	17:44																																																																																																																																					
Air Blank	0.000	17:45																																																																																																																																					
Control Test	0.080	17:45																																																																																																																																					
Air Blank	0.000	17:46																																																																																																																																					
Control Test	0.080	17:46																																																																																																																																					
Air Blank	0.000	17:46																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.0800																																																																																																																																						
Std Dev	0.0000																																																																																																																																						
Rel Std Dev(%)	0.0000																																																																																																																																						
<p>Operator's Signature</p> 	<p>Operator's Signature</p> 	<p>Operator's Signature</p> 	<p>Operator's Signature</p> 																																																																																																																																				

DPS

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: ORANGE COUNTY SO
Time of Inspection: 13:53

Date of Inspection: 12/30/2025

Serial Number: 80-001418
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202406K Exp: 06/19/2026	0.08g/210L Test (g/210L) Lot#:202406L Exp: 06/19/2026	0.20g/210L Test (g/210L) Lot#:202406N Exp: 06/20/2026	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG510701 Exp: 04/17/2027
0.000	0.050	0.081	0.202	0.081
0.000	0.050	0.080	0.202	0.081
0.000	0.050	0.081	0.201	0.081
0.000	0.050	0.080	0.201	0.081
0.000	0.050	0.080	0.200	0.081
0.000	0.050	0.080	0.201	0.082
0.000	0.050	0.080	0.200	0.082
0.000	0.050	0.081	0.201	0.082
0.000	0.050	0.081	0.200	0.082
0.000	0.050	0.081	0.201	0.082


Standard Deviations	0.0000	0.0005	0.0007	0.0005
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5



Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.


LEANDRA HIGGINBOTHAM
Signature and Printed Name

12/30/2025
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001418, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001418</u>	UNCERTAINTY* ±	
Owning Agency:	<u>ORANGE COUNTY SO</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>12/30/2025</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>13:53</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

M. J. G. H.
LEANDRA HIGGINBOTHAM,
Department Inspector

12/30/2025

Date

FDLE/ATP Form 69 October 2024
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Stability Checks

80-001418 *SNA*

0.050 g/210L 0.047 to 0.053 g/210L	0.080 g/210L 0.077 to 0.083 g/210L	0.200 g/210L 0.194 to 0.206 g/210L	DGS 0.080 g/210L 0.077 to 0.083 g/210L ≤ 0.003 g/210L of Wet
<p>ORANGE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001418 05/06/2025 Software: 8100.27</p> <p>Performed Root Case Analysis</p>	<p>ORANGE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001418 05/06/2025 Software: 8100.27</p> <p>Performed Root Case Analysis</p>	<p>ORANGE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001418 05/06/2025 Software: 8100.27</p> <p>Performed Root Case Analysis</p>	<p>ORANGE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001418 05/06/2025 Software: 8100.27</p> <p>Performed Root Case Analysis</p>
<p>Test g/210L Time</p> <p>Air Blank 0.000 14:18</p> <p>Control Test 0.050 14:19</p> <p>Air Blank 0.000 14:20</p> <p>Control Test 0.049 14:20</p> <p>Air Blank 0.000 14:21</p> <p>Control Test 0.049 14:21</p> <p>Air Blank 0.000 14:22</p> <p>Control Test Stats</p> <p>Average 0.0493</p> <p>Std Dev 0.0006</p> <p>Rel Std Dev(%) 1.1703</p>	<p>Test g/210L Time</p> <p>Air Blank 0.000 14:39</p> <p>Control Test 0.080 14:40</p> <p>Air Blank 0.000 14:40</p> <p>Control Test 0.080 14:41</p> <p>Air Blank 0.000 14:41</p> <p>Control Test 0.080 14:42</p> <p>Air Blank 0.000 14:43</p> <p>Control Test Stats</p> <p>Average 0.0800</p> <p>Std Dev 0.0000</p> <p>Rel Std Dev(%) 0.0000</p>	<p>Test g/210L Time</p> <p>Air Blank 0.000 14:26</p> <p>Control Test 0.200 14:27</p> <p>Air Blank 0.000 14:28</p> <p>Control Test 0.199 14:28</p> <p>Air Blank 0.000 14:29</p> <p>Control Test 0.198 14:30</p> <p>Air Blank 0.000 14:30</p> <p>Control Test Stats</p> <p>Average 0.1990</p> <p>Std Dev 0.0010</p> <p>Rel Std Dev(%) 0.5025</p>	<p>Test g/210L Time</p> <p>Air Blank 0.000 15:00</p> <p>Control Test 0.082 15:00</p> <p>Air Blank 0.000 15:01</p> <p>Control Test 0.082 15:01</p> <p>Air Blank 0.000 15:01</p> <p>Control Test 0.082 15:01</p> <p>Air Blank 0.000 15:02</p> <p>Control Test Stats</p> <p>Average 0.0820</p> <p>Std Dev 0.0000</p> <p>Rel Std Dev(%) 0.0000</p>
<p><i>M. J. J. J.</i> Operator Signature</p>			

DGS

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: ORANGE COUNTY SO
Time of Inspection: 11:59

Date of Inspection: 05/12/2025

Serial Number: 80-001418
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202406K Exp: 06/19/2026	0.08g/210L Test (g/210L) Lot#:202406L Exp: 06/19/2026	0.20g/210L Test (g/210L) Lot#:202406N Exp: 06/20/2026	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG429602 Exp: 10/22/2026
0.000	0.049	0.079	0.199	0.081
0.000	0.049	0.080	0.199	0.081
0.000	0.049	0.080	0.200	0.080
0.000	0.049	0.080	0.199	0.081
0.000	0.049	0.080	0.200	0.080
0.000	0.049	0.080	0.200	0.080
0.000	0.049	0.081	0.200	0.081
0.000	0.050	0.081	0.200	0.081
0.000	0.050	0.081	0.200	0.081
0.000	0.050	0.081	0.201	0.080

Standard Deviations	0.0004	0.0006	0.0006	0.0005
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0005 Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Shayla Platt

SHAYLA D PLATT

Signature and Printed Name

05/12/2025
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001418, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001418</u>	UNCERTAINTY* ±	
Owning Agency:	<u>ORANGE COUNTY SO</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>05/12/2025</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>11:59</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Shayla Platt
Digitally signed by
Shayla Platt
Date: 2025.05.15
21:03:30 -04'00'

05/12/2025

Date

SHAYLA D PLATT,

Department Inspector

FDLE/ATP Form 69 October 2024

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality