



AGENCY INSPECTION AND OTHER ELECTRONIC DATA REVIEW

Agency: Sumter County Sheriff's Office

Instrument Serial Number: 80-000816

AGENCY INSPECTION DATA REVIEW
Agency Inspector: Paul Caudle
Date of Inspection: 08/29/2022
Time of Inspection: 06:47:12
Agency Inspection Discrepancy: [ ] Incomplete [ ] Untimely/Not Received [ ] Erroneous Information [ ] Procedural [x] Other (Missing Required Information)
[ ] Agency Inspection Not Conducted or Records regarding Agency Inspection have not been uploaded.
[ ] Lot Number [ ] Expiration Date for g/ 210L [ ] Alcohol Reference Solution [ ] Dry Gas Standard is [ ] Incorrect [ ] Expired.
[x] FDLE/ATP Form 39 states in part, "If a test must be repeated, the REASON must be entered when prompted and recorded in the Remarks section of FDLE/ATP Form 40 Agency Inspection Report - Intoxilyzer 8000. The [x] REASON for repeating the following test(s); OR the [x] Possible Cause and Corrective Action Taken on the following test(s) was not recorded:
[ ] Alcohol Free Subject Test [ ] Mouth Alcohol Test [ ] Alcohol Free Test [ ] Interferent Detect Test
[x] 0.05 g/210L Test [ ] 0.08 g/210L Test [ ] 0.20 g/210L Test [ ] 0.08 g/210L Dry Gas Standard Test
[ ] FDLE/ATP Form 39 states in part, "If the instrument does not comply with the requirements of Chapter 11D-8, FAC, remove the instrument from service and notify the Department Inspector."
[ ] The Department Inspector was not notified. However, the issue was satisfactorily corrected and the repeated Agency Inspection complies with the requirements of Chapter 11D-8, FAC.
[ ] The Department Inspector was not notified. However, the repeated Agency Inspection does not comply with the requirements of Chapter 11D-8, FAC and the instrument was correctly removed from evidentiary use.
[ ] The Department Inspector was not notified. The repeated Agency Inspection does not comply with the requirements of Chapter 11D-8, FAC and the instrument was not removed from evidentiary use.
[ ] The Agency Inspection is noted as "Complies" when it does not comply with the requirements of Chapter 11D-8, FAC.
[ ] Other: \_\_\_\_\_

OTHER ELECTRONIC DATA REVIEW
[ ] Login Records Date: \_\_\_\_\_
[ ] Cylinder Change Records Date: \_\_\_\_\_
[ ] Control Test Records Date: \_\_\_\_\_
[ ] Diagnostic Check Records Date: \_\_\_\_\_
Comments:
The 0.05 ARS Test was repeated. The reason for repeating the test, as well as the corrective action taken prior to repeating the test, must be included in the remarks.

CORRECTIVE ACTION
[x] Record hand-written amendments on the FDLE/ATP Form 40, Agency Inspection Report, initial and date the amendments, mark the report "AMENDED", and forward a copy to the Department Inspector by 01/05/2023 (Date).
[ ] Provide a written explanation regarding the referenced item(s) to the Department Inspector by \_\_\_\_\_ (Date).
[ ] Upload the Agency Inspection(s).
[ ] Remove the instrument from evidentiary use until otherwise directed by the Department.
[ ] No action required
[ ] Other: \_\_\_\_\_

Taylor Durackow

Signature of Alcohol Testing Program Staff Member

12/5/2022
Date