

Alcohol Testing Program - Instrument Processing Sheet

Agency: PERRY PD Instrument Serial Number: 80-005337

Date In: 8/25/2026 DI Completion Date: 8/26/2026 ☐ Ship ☒ P/U ☐ H/D ☐ CMI ☐ EE

Intake By: <u>WKP</u> Date: <u>8/25/2026</u>	Quality Checks By: <u>WKP</u> Date: <u>8/25/2026</u>	Flow Adjustment By: _____																																																																																															
☒ Annual ☒ Dropped Off ☐ Registration ☐ Return from CMI / EE ☐ Training Instrument Visual Inspection ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/Accessories ☐ Power Cord ☐ Printer Cable ☐ Static Bag ☐ 12V DC Cable Notes:	☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value: <u>210</u> ☒ Flow Verification (L/s) Flow Column #: <u>ATP 102</u> 32 mm <u>0.156</u> (.139-.169) 36 mm <u>0.171</u> (.156-.190) 53 mm <u>0.238</u> (.228-.278) 103 mm <u>0.492</u> (.447-.547) ☒ Barometric Pressure Check Gauge ID #: <u>34420</u> Gauge: <u>1012</u> Instrument: <u>999</u> ☒ Stability Checks <table border="1" style="width: 100%; border-collapse: collapse; margin-top: 5px;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot#/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>MP5088</td> <td>202603A 3/3/2028</td> </tr> <tr> <td>0.080</td> <td>MP5089</td> <td>202603B 3/3/2028</td> </tr> <tr> <td>0.200</td> <td>MP5090</td> <td>202603D 3/4/2028</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>23725080A1 10/5/2027</td> </tr> </tbody> </table>	Simulator	Serial #	Lot#/Exp	0.050	MP5088	202603A 3/3/2028	0.080	MP5089	202603B 3/3/2028	0.200	MP5090	202603D 3/4/2028	0.080 DGS	N/A	23725080A1 10/5/2027	Flow Column #: _____ ☐ 5L/min – 17mm ☐ 15L/min – 53mm ☐ 30L/min – 103mm ☐ R-Value: _____ ☐ Post Adjustment Verification (L/S) Flow Column #: _____ 32 mm (.139-.169) 36 mm (.156-.190) 53 mm (.228-.278) 103 mm (.447-.547)																																																																																
Simulator	Serial #	Lot#/Exp																																																																																															
0.050	MP5088	202603A 3/3/2028																																																																																															
0.080	MP5089	202603B 3/3/2028																																																																																															
0.200	MP5090	202603D 3/4/2028																																																																																															
0.080 DGS	N/A	23725080A1 10/5/2027																																																																																															
<table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th style="width: 30%;">Optical Bench Adjustment</th> <th style="width: 20%;">By: <u>WKP</u></th> <th style="width: 50%;">Department Inspection</th> <th style="width: 20%;">By: <u>WKP</u></th> </tr> </thead> <tbody> <tr> <td colspan="2">Barometric Pressure Gauge: <u>1013</u> ID#: <u>28662</u></td> <td colspan="2">Barometric Pressure ID#: <u>34420</u></td> </tr> <tr> <td> <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr><td>0.000</td><td>MP6294</td><td>N/A</td><td>N/A</td></tr> <tr><td>0.040</td><td>MP6295</td><td>25090</td><td>3/11/2027</td></tr> <tr><td>0.100</td><td>MP6296</td><td>24390</td><td>10/29/2026</td></tr> <tr><td>0.200</td><td>MP6297</td><td>25020</td><td>1/14/2027</td></tr> <tr><td>0.300</td><td>MP6298</td><td>24430</td><td>12/10/2026</td></tr> <tr><td>0.080 DGS</td><td>N/A</td><td>AG526603</td><td>9/23/2027</td></tr> </tbody> </table> </td> <td colspan="2"> Gauge: <u>1013</u> Instrument: <u>1013</u> Mouth Alcohol Solution Lot #: <u>2025-D</u> Exp: <u>9/25/2027</u> Acetone Stock Solution Lot #: <u>2025-B</u> Exp: <u>9/22/2027</u> </td> <td> <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr><td>0.000</td><td>MP5086</td></tr> <tr><td>Interferent</td><td>MP5087</td></tr> <tr><td>0.050</td><td>MP5088</td></tr> <tr><td>0.080</td><td>MP5089</td></tr> <tr><td>0.200</td><td>MP5090</td></tr> </tbody> </table> </td> </tr> <tr> <td colspan="2"> ☒ Post Optical Bench Adjustment Stability Checks <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr><td>0.050</td><td>MP5088</td><td>202603A</td><td>3/3/2028</td></tr> <tr><td>0.080</td><td>MP5089</td><td>202603B</td><td>3/3/2028</td></tr> <tr><td>0.200</td><td>MP5090</td><td>202603D</td><td>3/4/2028</td></tr> <tr><td>0.080 DGS</td><td>N/A</td><td>23725080A1</td><td>10/5/2027</td></tr> </tbody> </table> </td> <td colspan="2"> ☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☒ Optical Bench Adjustment ☒ Average Calculations </td> <td> ☒ Post-Stability Checks ☐ Flow Adjustment ☐ Form 40 ☐ Other: </td> </tr> <tr> <td colspan="2"> Gauge ID #: <u>28427</u> Gauge: <u>1012</u> Instrument: <u>1012</u> </td> <td colspan="2"> <table border="1" style="width: 100%; border-collapse: collapse;"> <tr> <td colspan="2"> ☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC </td> </tr> <tr> <td colspan="2"> ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use </td> </tr> <tr> <td colspan="2"> ☒ Conduct an Agency Inspection Before Evidentiary Use </td> </tr> <tr> <td style="width: 50%;"> Taylor Gutschow Digitally signed by Taylor Gutschow Date: 2026.08.30 21:15:16 -04'00' </td> <td style="width: 50%;"> S. LeAndra Higginbotham Digitally signed by S. LeAndra Higginbotham Date: 2026.08.31 14:51:40 -04'00' </td> </tr> <tr> <td style="text-align: center;">Tech Review</td> <td style="text-align: center;">Admin Review</td> </tr> </table> </td> </tr> <tr> <td colspan="2"> Notes/Suggested Service: Admin Correction: Added signature to the first page of the optical bench adjustment. WKP 8/31/2026 </td> <td colspan="2"></td> </tr> </tbody> </table>			Optical Bench Adjustment	By: <u>WKP</u>	Department Inspection	By: <u>WKP</u>	Barometric Pressure Gauge: <u>1013</u> ID#: <u>28662</u>		Barometric Pressure ID#: <u>34420</u>		<table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr><td>0.000</td><td>MP6294</td><td>N/A</td><td>N/A</td></tr> <tr><td>0.040</td><td>MP6295</td><td>25090</td><td>3/11/2027</td></tr> <tr><td>0.100</td><td>MP6296</td><td>24390</td><td>10/29/2026</td></tr> <tr><td>0.200</td><td>MP6297</td><td>25020</td><td>1/14/2027</td></tr> <tr><td>0.300</td><td>MP6298</td><td>24430</td><td>12/10/2026</td></tr> <tr><td>0.080 DGS</td><td>N/A</td><td>AG526603</td><td>9/23/2027</td></tr> </tbody> </table>	Simulator	Serial #	Lot #	Expiration	0.000	MP6294	N/A	N/A	0.040	MP6295	25090	3/11/2027	0.100	MP6296	24390	10/29/2026	0.200	MP6297	25020	1/14/2027	0.300	MP6298	24430	12/10/2026	0.080 DGS	N/A	AG526603	9/23/2027	Gauge: <u>1013</u> Instrument: <u>1013</u> Mouth Alcohol Solution Lot #: <u>2025-D</u> Exp: <u>9/25/2027</u> Acetone Stock Solution Lot #: <u>2025-B</u> Exp: <u>9/22/2027</u>		<table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr><td>0.000</td><td>MP5086</td></tr> <tr><td>Interferent</td><td>MP5087</td></tr> <tr><td>0.050</td><td>MP5088</td></tr> <tr><td>0.080</td><td>MP5089</td></tr> <tr><td>0.200</td><td>MP5090</td></tr> </tbody> </table>	Simulator	Serial Number	0.000	MP5086	Interferent	MP5087	0.050	MP5088	0.080	MP5089	0.200	MP5090	☒ Post Optical Bench Adjustment Stability Checks <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr><td>0.050</td><td>MP5088</td><td>202603A</td><td>3/3/2028</td></tr> <tr><td>0.080</td><td>MP5089</td><td>202603B</td><td>3/3/2028</td></tr> <tr><td>0.200</td><td>MP5090</td><td>202603D</td><td>3/4/2028</td></tr> <tr><td>0.080 DGS</td><td>N/A</td><td>23725080A1</td><td>10/5/2027</td></tr> </tbody> </table>		Simulator	Serial #	Lot #	Expiration	0.050	MP5088	202603A	3/3/2028	0.080	MP5089	202603B	3/3/2028	0.200	MP5090	202603D	3/4/2028	0.080 DGS	N/A	23725080A1	10/5/2027	☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☒ Optical Bench Adjustment ☒ Average Calculations		☒ Post-Stability Checks ☐ Flow Adjustment ☐ Form 40 ☐ Other:	Gauge ID #: <u>28427</u> Gauge: <u>1012</u> Instrument: <u>1012</u>		<table border="1" style="width: 100%; border-collapse: collapse;"> <tr> <td colspan="2"> ☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC </td> </tr> <tr> <td colspan="2"> ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use </td> </tr> <tr> <td colspan="2"> ☒ Conduct an Agency Inspection Before Evidentiary Use </td> </tr> <tr> <td style="width: 50%;"> Taylor Gutschow Digitally signed by Taylor Gutschow Date: 2026.08.30 21:15:16 -04'00' </td> <td style="width: 50%;"> S. LeAndra Higginbotham Digitally signed by S. LeAndra Higginbotham Date: 2026.08.31 14:51:40 -04'00' </td> </tr> <tr> <td style="text-align: center;">Tech Review</td> <td style="text-align: center;">Admin Review</td> </tr> </table>		☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC		☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use		☒ Conduct an Agency Inspection Before Evidentiary Use		Taylor Gutschow Digitally signed by Taylor Gutschow Date: 2026.08.30 21:15:16 -04'00'	S. LeAndra Higginbotham Digitally signed by S. LeAndra Higginbotham Date: 2026.08.31 14:51:40 -04'00'	Tech Review	Admin Review	Notes/Suggested Service: Admin Correction: Added signature to the first page of the optical bench adjustment. WKP 8/31/2026			
Optical Bench Adjustment	By: <u>WKP</u>	Department Inspection	By: <u>WKP</u>																																																																																														
Barometric Pressure Gauge: <u>1013</u> ID#: <u>28662</u>		Barometric Pressure ID#: <u>34420</u>																																																																																															
<table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr><td>0.000</td><td>MP6294</td><td>N/A</td><td>N/A</td></tr> <tr><td>0.040</td><td>MP6295</td><td>25090</td><td>3/11/2027</td></tr> <tr><td>0.100</td><td>MP6296</td><td>24390</td><td>10/29/2026</td></tr> <tr><td>0.200</td><td>MP6297</td><td>25020</td><td>1/14/2027</td></tr> <tr><td>0.300</td><td>MP6298</td><td>24430</td><td>12/10/2026</td></tr> <tr><td>0.080 DGS</td><td>N/A</td><td>AG526603</td><td>9/23/2027</td></tr> </tbody> </table>	Simulator	Serial #	Lot #	Expiration	0.000	MP6294	N/A	N/A	0.040	MP6295	25090	3/11/2027	0.100	MP6296	24390	10/29/2026	0.200	MP6297	25020	1/14/2027	0.300	MP6298	24430	12/10/2026	0.080 DGS	N/A	AG526603	9/23/2027	Gauge: <u>1013</u> Instrument: <u>1013</u> Mouth Alcohol Solution Lot #: <u>2025-D</u> Exp: <u>9/25/2027</u> Acetone Stock Solution Lot #: <u>2025-B</u> Exp: <u>9/22/2027</u>		<table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr><td>0.000</td><td>MP5086</td></tr> <tr><td>Interferent</td><td>MP5087</td></tr> <tr><td>0.050</td><td>MP5088</td></tr> <tr><td>0.080</td><td>MP5089</td></tr> <tr><td>0.200</td><td>MP5090</td></tr> </tbody> </table>	Simulator	Serial Number	0.000	MP5086	Interferent	MP5087	0.050	MP5088	0.080	MP5089	0.200	MP5090																																																						
Simulator	Serial #	Lot #	Expiration																																																																																														
0.000	MP6294	N/A	N/A																																																																																														
0.040	MP6295	25090	3/11/2027																																																																																														
0.100	MP6296	24390	10/29/2026																																																																																														
0.200	MP6297	25020	1/14/2027																																																																																														
0.300	MP6298	24430	12/10/2026																																																																																														
0.080 DGS	N/A	AG526603	9/23/2027																																																																																														
Simulator	Serial Number																																																																																																
0.000	MP5086																																																																																																
Interferent	MP5087																																																																																																
0.050	MP5088																																																																																																
0.080	MP5089																																																																																																
0.200	MP5090																																																																																																
☒ Post Optical Bench Adjustment Stability Checks <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr><td>0.050</td><td>MP5088</td><td>202603A</td><td>3/3/2028</td></tr> <tr><td>0.080</td><td>MP5089</td><td>202603B</td><td>3/3/2028</td></tr> <tr><td>0.200</td><td>MP5090</td><td>202603D</td><td>3/4/2028</td></tr> <tr><td>0.080 DGS</td><td>N/A</td><td>23725080A1</td><td>10/5/2027</td></tr> </tbody> </table>		Simulator	Serial #	Lot #	Expiration	0.050	MP5088	202603A	3/3/2028	0.080	MP5089	202603B	3/3/2028	0.200	MP5090	202603D	3/4/2028	0.080 DGS	N/A	23725080A1	10/5/2027	☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☒ Optical Bench Adjustment ☒ Average Calculations		☒ Post-Stability Checks ☐ Flow Adjustment ☐ Form 40 ☐ Other:																																																																									
Simulator	Serial #	Lot #	Expiration																																																																																														
0.050	MP5088	202603A	3/3/2028																																																																																														
0.080	MP5089	202603B	3/3/2028																																																																																														
0.200	MP5090	202603D	3/4/2028																																																																																														
0.080 DGS	N/A	23725080A1	10/5/2027																																																																																														
Gauge ID #: <u>28427</u> Gauge: <u>1012</u> Instrument: <u>1012</u>		<table border="1" style="width: 100%; border-collapse: collapse;"> <tr> <td colspan="2"> ☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC </td> </tr> <tr> <td colspan="2"> ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use </td> </tr> <tr> <td colspan="2"> ☒ Conduct an Agency Inspection Before Evidentiary Use </td> </tr> <tr> <td style="width: 50%;"> Taylor Gutschow Digitally signed by Taylor Gutschow Date: 2026.08.30 21:15:16 -04'00' </td> <td style="width: 50%;"> S. LeAndra Higginbotham Digitally signed by S. LeAndra Higginbotham Date: 2026.08.31 14:51:40 -04'00' </td> </tr> <tr> <td style="text-align: center;">Tech Review</td> <td style="text-align: center;">Admin Review</td> </tr> </table>		☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC		☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use		☒ Conduct an Agency Inspection Before Evidentiary Use		Taylor Gutschow Digitally signed by Taylor Gutschow Date: 2026.08.30 21:15:16 -04'00'	S. LeAndra Higginbotham Digitally signed by S. LeAndra Higginbotham Date: 2026.08.31 14:51:40 -04'00'	Tech Review	Admin Review																																																																																				
☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC																																																																																																	
☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use																																																																																																	
☒ Conduct an Agency Inspection Before Evidentiary Use																																																																																																	
Taylor Gutschow Digitally signed by Taylor Gutschow Date: 2026.08.30 21:15:16 -04'00'	S. LeAndra Higginbotham Digitally signed by S. LeAndra Higginbotham Date: 2026.08.31 14:51:40 -04'00'																																																																																																
Tech Review	Admin Review																																																																																																
Notes/Suggested Service: Admin Correction: Added signature to the first page of the optical bench adjustment. WKP 8/31/2026																																																																																																	

Stability Checks

0.05g/210L 0.047 to 0.053	0.08g/210L 0.077 to 0.083	0.20g/210L 0.194 to 0.206	DGS 0.08g/210L 0.077 to 0.083																																																																																																																																										
PERK 20 Intoxilizer - Alcohol Analyzer Model 8000 SN 80-05537 08/25/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>13:51</td></tr> <tr><td>Control Test</td><td>0.050</td><td>13:52</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:53</td></tr> <tr><td>Control Test</td><td>0.049</td><td>13:53</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:54</td></tr> <tr><td>Control Test</td><td>0.050</td><td>13:55</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:55</td></tr> <tr><td>Control Test</td><td>0.050</td><td>13:55</td></tr> <tr><td>Average</td><td>0.0497</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.1625</td><td></td></tr> </tbody> </table> Operator's Signature _____	Test	g/210L	Time	Air Blank	0.000	13:51	Control Test	0.050	13:52	Air Blank	0.000	13:53	Control Test	0.049	13:53	Air Blank	0.000	13:54	Control Test	0.050	13:55	Air Blank	0.000	13:55	Control Test	0.050	13:55	Average	0.0497		Std Dev	0.0006		Rel Std Dev(%)	1.1625		PERK 20 Intoxilizer - Alcohol Analyzer Model 8000 SN 80-05537 08/25/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>13:42</td></tr> <tr><td>Control Test</td><td>0.079</td><td>13:43</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:43</td></tr> <tr><td>Control Test</td><td>0.078</td><td>13:44</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:44</td></tr> <tr><td>Control Test</td><td>0.078</td><td>13:45</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:45</td></tr> <tr><td>Control Test</td><td>0.078</td><td>13:46</td></tr> <tr><td>Average</td><td>0.0783</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7370</td><td></td></tr> </tbody> </table> Operator's Signature _____	Test	g/210L	Time	Air Blank	0.000	13:42	Control Test	0.079	13:43	Air Blank	0.000	13:43	Control Test	0.078	13:44	Air Blank	0.000	13:44	Control Test	0.078	13:45	Air Blank	0.000	13:45	Control Test	0.078	13:46	Average	0.0783		Std Dev	0.0006		Rel Std Dev(%)	0.7370		PERK 20 Intoxilizer - Alcohol Analyzer Model 8000 SN 80-05537 08/25/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>13:36</td></tr> <tr><td>Control Test</td><td>0.200</td><td>13:37</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:36</td></tr> <tr><td>Control Test</td><td>0.198</td><td>13:38</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:39</td></tr> <tr><td>Control Test</td><td>0.198</td><td>13:40</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:40</td></tr> <tr><td>Control Test</td><td>0.1987</td><td>13:41</td></tr> <tr><td>Average</td><td>0.1987</td><td></td></tr> <tr><td>Std Dev</td><td>0.0012</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.5812</td><td></td></tr> </tbody> </table> Operator's Signature _____	Test	g/210L	Time	Air Blank	0.000	13:36	Control Test	0.200	13:37	Air Blank	0.000	13:36	Control Test	0.198	13:38	Air Blank	0.000	13:39	Control Test	0.198	13:40	Air Blank	0.000	13:40	Control Test	0.1987	13:41	Average	0.1987		Std Dev	0.0012		Rel Std Dev(%)	0.5812		PERK 20 Intoxilizer - Alcohol Analyzer Model 8000 SN 80-05537 08/25/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>13:47</td></tr> <tr><td>Control Test</td><td>0.081</td><td>13:48</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:48</td></tr> <tr><td>Control Test</td><td>0.080</td><td>13:48</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:48</td></tr> <tr><td>Control Test</td><td>0.080</td><td>13:48</td></tr> <tr><td>Average</td><td>0.0803</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7187</td><td></td></tr> </tbody> </table> Operator's Signature _____	Test	g/210L	Time	Air Blank	0.000	13:47	Control Test	0.081	13:48	Air Blank	0.000	13:48	Control Test	0.080	13:48	Air Blank	0.000	13:48	Control Test	0.080	13:48	Average	0.0803		Std Dev	0.0006		Rel Std Dev(%)	0.7187	
Test	g/210L	Time																																																																																																																																											
Air Blank	0.000	13:51																																																																																																																																											
Control Test	0.050	13:52																																																																																																																																											
Air Blank	0.000	13:53																																																																																																																																											
Control Test	0.049	13:53																																																																																																																																											
Air Blank	0.000	13:54																																																																																																																																											
Control Test	0.050	13:55																																																																																																																																											
Air Blank	0.000	13:55																																																																																																																																											
Control Test	0.050	13:55																																																																																																																																											
Average	0.0497																																																																																																																																												
Std Dev	0.0006																																																																																																																																												
Rel Std Dev(%)	1.1625																																																																																																																																												
Test	g/210L	Time																																																																																																																																											
Air Blank	0.000	13:42																																																																																																																																											
Control Test	0.079	13:43																																																																																																																																											
Air Blank	0.000	13:43																																																																																																																																											
Control Test	0.078	13:44																																																																																																																																											
Air Blank	0.000	13:44																																																																																																																																											
Control Test	0.078	13:45																																																																																																																																											
Air Blank	0.000	13:45																																																																																																																																											
Control Test	0.078	13:46																																																																																																																																											
Average	0.0783																																																																																																																																												
Std Dev	0.0006																																																																																																																																												
Rel Std Dev(%)	0.7370																																																																																																																																												
Test	g/210L	Time																																																																																																																																											
Air Blank	0.000	13:36																																																																																																																																											
Control Test	0.200	13:37																																																																																																																																											
Air Blank	0.000	13:36																																																																																																																																											
Control Test	0.198	13:38																																																																																																																																											
Air Blank	0.000	13:39																																																																																																																																											
Control Test	0.198	13:40																																																																																																																																											
Air Blank	0.000	13:40																																																																																																																																											
Control Test	0.1987	13:41																																																																																																																																											
Average	0.1987																																																																																																																																												
Std Dev	0.0012																																																																																																																																												
Rel Std Dev(%)	0.5812																																																																																																																																												
Test	g/210L	Time																																																																																																																																											
Air Blank	0.000	13:47																																																																																																																																											
Control Test	0.081	13:48																																																																																																																																											
Air Blank	0.000	13:48																																																																																																																																											
Control Test	0.080	13:48																																																																																																																																											
Air Blank	0.000	13:48																																																																																																																																											
Control Test	0.080	13:48																																																																																																																																											
Average	0.0803																																																																																																																																												
Std Dev	0.0006																																																																																																																																												
Rel Std Dev(%)	0.7187																																																																																																																																												

PERRY PD
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-005337
 08/26/2026 09:16:57

Auto Calibration *Wen-Chi Pierson*

pg 1 of 2

<<<<< 3um >>>>>			<<<<< 9um >>>>>		
Solution = 0.000 g/210L or 0.0000 mg/l, Samples = 4, Discarded = 1					
Sample	% Abs	(% Abs Ref)	% Abs	(% Abs Ref)	
Sample #1	0.0600	(-0.0130)	0.1780	(-0.0120)	
Sample #2	0.0550	(0.0000)	0.1390	(0.0080)	
Sample #3	0.0750	(0.0200)	0.1620	(0.0070)	
Sample #4	0.0370	(0.0360)	0.1570	(0.0090)	
Avg % Abs	0.0557	(0.0187)	0.1527	(0.0080)	
STD DEV	0.0190	(0.0180)	0.0121	(0.0010)	
REL STD DEV	34.147	(96.627)	7.924	(12.500)	
Solution = 0.040 g/210L or 0.1905 mg/l, Samples = 4, Discarded = 1					
Sample	% Abs	(% Abs Ref)	% Abs	(% Abs Ref)	
Sample #1	0.7910	(-0.0130)	1.5670	(-0.0140)	
Sample #2	0.7690	(0.0030)	1.5350	(0.0130)	
Sample #3	0.7430	(0.0430)	1.4900	(0.0410)	
Sample #4	0.8120	(0.0180)	1.5290	(0.0240)	
Avg % Abs	0.7747	(0.0213)	1.5180	(0.0260)	
STD DEV	0.0348	(0.0202)	0.0244	(0.0141)	
REL STD DEV	4.498	(94.722)	1.610	(54.257)	
Solution = 0.100 g/210L or 0.4762 mg/l, Samples = 4, Discarded = 1					
Sample	% Abs	(% Abs Ref)	% Abs	(% Abs Ref)	
Sample #1	1.8470	(0.0000)	3.5080	(0.0070)	
Sample #2	1.8360	(-0.0020)	3.4910	(0.0320)	
Sample #3	1.8700	(0.0120)	3.5260	(0.0260)	
Sample #4	1.8270	(0.0310)	3.4900	(0.0520)	
Avg % Abs	1.8443	(0.0137)	3.5023	(0.0367)	
STD DEV	0.0227	(0.0166)	0.0205	(0.0136)	
REL STD DEV	1.230	(121.193)	0.585	(37.128)	
Solution = 0.200 g/210L or 0.9524 mg/l, Samples = 4, Discarded = 1					
Sample	% Abs	(% Abs Ref)	% Abs	(% Abs Ref)	
Sample #1	3.6480	(-0.0020)	6.7590	(0.0000)	
Sample #2	3.5750	(0.0550)	6.6700	(0.0820)	
Sample #3	3.6230	(0.0390)	6.6580	(0.0800)	
Sample #4	3.6210	(0.0410)	6.6920	(0.0800)	
Avg % Abs	3.6063	(0.0450)	6.6733	(0.0807)	
STD DEV	0.0272	(0.0087)	0.0172	(0.0012)	
REL STD DEV	0.753	(19.373)	0.258	(1.431)	
Solution = 0.300 g/210L or 1.4286 mg/l, Samples = 4, Discarded = 1					
Sample	% Abs	(% Abs Ref)	% Abs	(% Abs Ref)	
Sample #1	5.2770	(-0.0190)	9.7200	(-0.0200)	
Sample #2	5.2730	(0.0160)	9.6410	(0.0590)	
Sample #3	5.2820	(0.0230)	9.6380	(0.0840)	
Sample #4	5.2780	(0.0300)	9.6420	(0.0670)	
Avg % Abs	5.2777	(0.0230)	9.6403	(0.0700)	
STD DEV	0.0045	(0.0070)	0.0021	(0.0128)	
REL STD DEV	0.085	(30.435)	0.022	(18.239)	

PERRY PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-005337
08/26/2026 09:16:57

Auto Calibration

pg 2 of 2

<<<< 3um >>>>			<<<< 9um >>>>		
-----			-----		
Zero Order Coef	-128.77			-202.27	
First Order Coef	2590.23			1363.70	
Second Order Coef	26.30			14.37	
-----			-----		
Act	Fit	Residual	Act	Fit	Residual
(g/210L)	(g/210L)	(g/210L)	(g/210L)	(g/210L)	(g/210L)
0.000	0.000	-0.0003	0.000	0.000	-0.0001
0.040	0.040	0.0002	0.040	0.040	0.0001
0.100	0.099	0.0005	0.100	0.100	0.0002
0.200	0.201	-0.0006	0.200	0.200	-0.0003
0.300	0.300	0.0002	0.300	0.300	0.0001
-----			-----		

<<<< 3um >>>>		<<<< 9um >>>>	
-----		-----	
Solution = 0.080 g/210L or 0.3810 mg/l, Samples = 4, Discarded = 1			
Sample			
Sample #1	3397.00		3261.00
Sample #2	3453.00		3268.00
Sample #3	3427.00		3282.00
Sample #4	3454.00		3268.00
Avg	3444.6667		3272.6667
STD DEV	15.3080		8.0829
REL STD DEV	0.444		0.247
H2O adjust (mg/l*10k)	365		537

Barometric Pressure = 1012

*****CALIBRATION SUCCESSFUL*****
men
8/26/2026

Post-Cal Stability Checks

0.05g/210L 0.047 to 0.053	0.08g/210L 0.077 to 0.083	0.20g/210L 0.194 to 0.206	DGS 0.08g/210L 0.077 to 0.083 ✓ ≤0.003 of Wet ✓																																																																																																																																																
PERRY PD Intoxilyzer - Alcohol Analyzer Model 8000 09/26/2026 Software: 6100.27 SN 81-05337 <table border="1"> <thead> <tr> <th>Test</th> <th>9/21/02</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.050</td><td>11:01</td></tr> <tr><td>Control Test</td><td>0.050</td><td>11:02</td></tr> <tr><td>Air Blank</td><td>0.050</td><td>11:03</td></tr> <tr><td>Control Test</td><td>0.050</td><td>11:03</td></tr> <tr><td>Air Blank</td><td>0.050</td><td>11:04</td></tr> <tr><td>Control Test</td><td>0.050</td><td>11:04</td></tr> <tr><td>Air Blank</td><td>0.050</td><td>11:05</td></tr> <tr><td>Control Test</td><td>0.050</td><td>11:05</td></tr> <tr><td>Average</td><td>0.0500</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel. Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table>	Test	9/21/02	Time	Air Blank	0.050	11:01	Control Test	0.050	11:02	Air Blank	0.050	11:03	Control Test	0.050	11:03	Air Blank	0.050	11:04	Control Test	0.050	11:04	Air Blank	0.050	11:05	Control Test	0.050	11:05	Average	0.0500		Std Dev	0.0000		Rel. Std Dev(%)	0.0000		PERRY PD Intoxilyzer - Alcohol Analyzer Model 8000 09/26/2026 Software: 6100.27 SN 81-05337 <table border="1"> <thead> <tr> <th>Test</th> <th>9/21/02</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>11:15</td></tr> <tr><td>Control Test</td><td>0.078</td><td>11:16</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:16</td></tr> <tr><td>Control Test</td><td>0.078</td><td>11:17</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:18</td></tr> <tr><td>Control Test</td><td>0.078</td><td>11:18</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:19</td></tr> <tr><td>Control Test</td><td>0.000</td><td>11:19</td></tr> <tr><td>Average</td><td>0.0780</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel. Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table>	Test	9/21/02	Time	Air Blank	0.000	11:15	Control Test	0.078	11:16	Air Blank	0.000	11:16	Control Test	0.078	11:17	Air Blank	0.000	11:18	Control Test	0.078	11:18	Air Blank	0.000	11:19	Control Test	0.000	11:19	Average	0.0780		Std Dev	0.0000		Rel. Std Dev(%)	0.0000		PERRY PD Intoxilyzer - Alcohol Analyzer Model 8000 09/26/2026 Software: 6100.27 SN 81-05337 <table border="1"> <thead> <tr> <th>Test</th> <th>9/21/02</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>11:07</td></tr> <tr><td>Control Test</td><td>0.199</td><td>11:08</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:09</td></tr> <tr><td>Control Test</td><td>0.198</td><td>11:09</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:10</td></tr> <tr><td>Control Test</td><td>0.199</td><td>11:11</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:11</td></tr> <tr><td>Control Test</td><td>0.000</td><td>11:11</td></tr> <tr><td>Average</td><td>0.1987</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel. Std Dev(%)</td><td>0.2906</td><td></td></tr> </tbody> </table>	Test	9/21/02	Time	Air Blank	0.000	11:07	Control Test	0.199	11:08	Air Blank	0.000	11:09	Control Test	0.198	11:09	Air Blank	0.000	11:10	Control Test	0.199	11:11	Air Blank	0.000	11:11	Control Test	0.000	11:11	Average	0.1987		Std Dev	0.0006		Rel. Std Dev(%)	0.2906		PERRY PD Intoxilyzer - Alcohol Analyzer Model 8000 09/26/2026 Software: 6100.27 SN 81-05337 <table border="1"> <thead> <tr> <th>Test</th> <th>9/21/02</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>10:30</td></tr> <tr><td>Control Test</td><td>0.079</td><td>10:30</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:31</td></tr> <tr><td>Control Test</td><td>0.079</td><td>10:31</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:31</td></tr> <tr><td>Control Test</td><td>0.078</td><td>10:32</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:32</td></tr> <tr><td>Control Test</td><td>0.000</td><td>10:32</td></tr> <tr><td>Average</td><td>0.0787</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel. Std Dev(%)</td><td>0.7339</td><td></td></tr> </tbody> </table>	Test	9/21/02	Time	Air Blank	0.000	10:30	Control Test	0.079	10:30	Air Blank	0.000	10:31	Control Test	0.079	10:31	Air Blank	0.000	10:31	Control Test	0.078	10:32	Air Blank	0.000	10:32	Control Test	0.000	10:32	Average	0.0787		Std Dev	0.0006		Rel. Std Dev(%)	0.7339	
Test	9/21/02	Time																																																																																																																																																	
Air Blank	0.050	11:01																																																																																																																																																	
Control Test	0.050	11:02																																																																																																																																																	
Air Blank	0.050	11:03																																																																																																																																																	
Control Test	0.050	11:03																																																																																																																																																	
Air Blank	0.050	11:04																																																																																																																																																	
Control Test	0.050	11:04																																																																																																																																																	
Air Blank	0.050	11:05																																																																																																																																																	
Control Test	0.050	11:05																																																																																																																																																	
Average	0.0500																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel. Std Dev(%)	0.0000																																																																																																																																																		
Test	9/21/02	Time																																																																																																																																																	
Air Blank	0.000	11:15																																																																																																																																																	
Control Test	0.078	11:16																																																																																																																																																	
Air Blank	0.000	11:16																																																																																																																																																	
Control Test	0.078	11:17																																																																																																																																																	
Air Blank	0.000	11:18																																																																																																																																																	
Control Test	0.078	11:18																																																																																																																																																	
Air Blank	0.000	11:19																																																																																																																																																	
Control Test	0.000	11:19																																																																																																																																																	
Average	0.0780																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel. Std Dev(%)	0.0000																																																																																																																																																		
Test	9/21/02	Time																																																																																																																																																	
Air Blank	0.000	11:07																																																																																																																																																	
Control Test	0.199	11:08																																																																																																																																																	
Air Blank	0.000	11:09																																																																																																																																																	
Control Test	0.198	11:09																																																																																																																																																	
Air Blank	0.000	11:10																																																																																																																																																	
Control Test	0.199	11:11																																																																																																																																																	
Air Blank	0.000	11:11																																																																																																																																																	
Control Test	0.000	11:11																																																																																																																																																	
Average	0.1987																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel. Std Dev(%)	0.2906																																																																																																																																																		
Test	9/21/02	Time																																																																																																																																																	
Air Blank	0.000	10:30																																																																																																																																																	
Control Test	0.079	10:30																																																																																																																																																	
Air Blank	0.000	10:31																																																																																																																																																	
Control Test	0.079	10:31																																																																																																																																																	
Air Blank	0.000	10:31																																																																																																																																																	
Control Test	0.078	10:32																																																																																																																																																	
Air Blank	0.000	10:32																																																																																																																																																	
Control Test	0.000	10:32																																																																																																																																																	
Average	0.0787																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel. Std Dev(%)	0.7339																																																																																																																																																		
<i>menya</i> Operator's Signature	<i>menya</i> Operator's Signature	<i>menya</i> Operator's Signature	<i>menya</i> Operator's Signature																																																																																																																																																

Florida Department of Law Enforcement

Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: PERRY PD
Time of Inspection: 14:12

Date of Inspection: 08/26/2026

Serial Number: 80-005337
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202603A Exp: 03/03/2028	0.08g/210L Test (g/210L) Lot#:202603B Exp: 03/03/2028	0.20g/210L Test (g/210L) Lot#:202603D Exp: 03/04/2028	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:23725080A1 Exp: 10/05/2027
0.000	0.050	0.078	0.199	0.079
0.000	0.050	0.078	0.199	0.079
0.000	0.051	0.078	0.198	0.078
0.000	0.051	0.078	0.198	0.078
0.000	0.051	0.078	0.199	0.078
0.000	0.051	0.078	0.199	0.078
0.000	0.051	0.078	0.199	0.078
0.000	0.051	0.079	0.198	0.078
0.000	0.050	0.078	0.199	0.078
0.000	0.051	0.078	0.199	0.077

Standard Deviations	0.0004	0.0003	0.0004	0.0005
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



WEN-CHI K PIERSON

Signature and Printed Name

08/26/2026
Date

Florida Department of Law Enforcement
Alcohol Testing Program - Average Calculations for Calibration Certificate

Calculations By: WKP

Serial Number: 80-005337

Date: 8/27/2026

Solution:

0.05 g/210 L Test: 0.507 / 10 = 0.051 g/210L
(SUM) (AVERAGE)

0.08 g/210L Test: 0.781 / 10 = 0.078 g/210L
(SUM) (AVERAGE)

0.20 g/210L Test: 1.987 / 10 = 0.199 g/210L
(SUM) (AVERAGE)

0.08 g/210L Dry Gas Std Test: 0.781 / 10 = 0.078 g/210L
(SUM) (AVERAGE)

Comments:

Average = (Sum of observations) / (Total number of observations)

All results are reported in g/210L



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-005337, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-005337</u>	Average of y ± Uncertainty*
Owning Agency:	<u>PERRY PD</u>	0.051 g/ 210 L 0.004
Calibration Date:	<u>08/26/2026</u>	0.078 g/ 210 L 0.004
Calibration Time:	<u>14:12</u>	0.199 g/ 210 L 0.008
		0.078 g/ 210 L Dry Gas Control 0.005

All results are reported in g/210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Wen-Chi

Chi Pierson
Date: 2026.08.27 11:55:50
-04'00'

Pierson

08/26/2026

Date

WEN-CHI K PIERSON,

Department Inspector

FDLE/ATP Form 69 August 2026

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality