

Alcohol Testing Program - Instrument Processing Sheet

Agency: HILLSBOROUGH CO SO **Instrument Serial Number:** 80-007480

Date In: 2/24/2026 **DI Completion Date:** 2/26/2026 ☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake By: <u>WKP</u> Date: <u>2/24/2026</u>		Quality Checks By: <u>KTS</u> Date: <u>2/25/2026</u>		Flow Adjustment By: _____																
☒ Annual ☐ Dropped Off ☐ Registration ☐ Return from CMI / EE ☐ Training Instrument Visual Inspection ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/Accessories ☐ Power Cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes:		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value: <u>120</u> ☒ Flow Verification (L/s) Flow Column #: <u>ATP103</u> 32 mm <u>0.144</u> (.139-.169) 36 mm <u>0.156</u> (.156-.190) 53 mm <u>0.234</u> (.228-.278) 103 mm <u>0.484</u> (.447-.547) ☒ Barometric Pressure Check Gauge ID #: <u>28421</u> Gauge: <u>1016</u> Instrument: <u>1016</u> ☒ Stability Checks <table border="1" style="width: 100%; border-collapse: collapse; margin-top: 5px;"> <tr> <th style="width: 15%;">Simulator</th> <th style="width: 25%;">Serial #</th> <th style="width: 60%;">Lot#/Exp</th> </tr> <tr> <td>0.050</td> <td>MP5088</td> <td>202406K 6/19/2026</td> </tr> <tr> <td>0.080</td> <td>MP5089</td> <td>202406L 6/19/2026</td> </tr> <tr> <td>0.200</td> <td>MP5090</td> <td>202406N 6/20/2026</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>28424080A3 11/5/2026</td> </tr> </table>		Simulator	Serial #	Lot#/Exp	0.050	MP5088	202406K 6/19/2026	0.080	MP5089	202406L 6/19/2026	0.200	MP5090	202406N 6/20/2026	0.080 DGS	N/A	28424080A3 11/5/2026	Flow Column #: _____ ☐ 5L/min – 17mm ☐ 15L/min – 53mm ☐ 30L/min – 103mm ☐ R-Value: _____ ☐ Post Adjustment Verification (L/S) Flow Column #: _____ 32 mm (.139-.169) 36 mm (.156-.190) 53 mm (.228-.278) 103 mm (.447-.547)	
Simulator	Serial #	Lot#/Exp																		
0.050	MP5088	202406K 6/19/2026																		
0.080	MP5089	202406L 6/19/2026																		
0.200	MP5090	202406N 6/20/2026																		
0.080 DGS	N/A	28424080A3 11/5/2026																		
		Maintenance By: _____ Date: _____																		
				☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Tank Sensor Tare ☐ Breath Tube Replacement ☐ Other:																
Optical Bench Adjustment By: _____		Department Inspection By: <u>KTS</u>																		
Barometric Pressure Gauge: _____ ID#: _____		Barometric Pressure ID#: <u>28421</u>																		
Simulator	Serial #	Lot #	Expiration	Gauge: <u>1013</u> Instrument: <u>1013</u>																
0.000		N/A	N/A	Mouth Alcohol Solution Lot #: <u>2025-D</u> Exp: <u>9/25/2027</u>																
0.040				Acetone Stock Solution Lot #: <u>2025-B</u> Exp: <u>9/22/2027</u>																
0.100				Simulator	Serial Number															
0.200				0.000	MP5086															
0.300				Interferent	MP5087															
0.080 DGS	N/A			0.050	MP5088															
☐ Post Optical Bench Adjustment Stability Checks				0.080	MP5089															
Simulator	Serial #	Lot #	Expiration	0.200	MP5090															
0.050				Attachments																
0.080				☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☐ Optical Bench Adjustment ☐ Post-Stability Checks ☐ Flow Adjustment ☐ Form 40 ☐ Other:																
0.200																				
0.080 DGS	N/A																			
Gauge ID #: _____ Gauge: _____ Instrument: _____																				
Notes/Suggested Service:				☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use																
				Digitally signed by Shayla Platt Date: 2026.03.06 13:24:45 -05'00'																
				Digitally signed by Shayla Platt Date: 2026.03.06 13:25:15 -05'00'																
				Tech Review Admin Review																

Stability Checks


0.050 g/210L	0.080 g/210L	0.200 g/210L	DGS 0.080 g/210L																																																																																																																																																
0.047 to 0.053 g/210L	0.077 to 0.083 g/210L	0.194 to 0.206 g/210L	0.077 to 0.083 g/210L																																																																																																																																																
☒	☒	☒	☒																																																																																																																																																
Performed Root Case Analysis	Performed Root Case Analysis	Performed Root Case Analysis	Performed Root Case Analysis																																																																																																																																																
HILLSBOROUGH CSO Intoxilizer - Alcohol Analyzer Model 8000 02/25/2026 Software: 8100.27 SN 80-007480 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>08:20</td></tr> <tr><td>Control Test</td><td>0.049</td><td>08:21</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>08:22</td></tr> <tr><td>Control Test</td><td>0.048</td><td>08:22</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>08:23</td></tr> <tr><td>Control Test</td><td>0.049</td><td>08:23</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>08:24</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0487</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.1863</td><td></td></tr> </tbody> </table> Operator's Signature _____	Test	g/210L	Time	Air Blank	0.000	08:20	Control Test	0.049	08:21	Air Blank	0.000	08:22	Control Test	0.048	08:22	Air Blank	0.000	08:23	Control Test	0.049	08:23	Air Blank	0.000	08:24	Control Test Stats			Average	0.0487		Std Dev	0.0006		Rel Std Dev(%)	1.1863		HILLSBOROUGH CSO Intoxilizer - Alcohol Analyzer Model 8000 02/25/2026 Software: 8100.27 SN 80-007480 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>08:10</td></tr> <tr><td>Control Test</td><td>0.079</td><td>08:11</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>08:12</td></tr> <tr><td>Control Test</td><td>0.079</td><td>08:12</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>08:13</td></tr> <tr><td>Control Test</td><td>0.079</td><td>08:13</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>08:14</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0790</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> Operator's Signature _____	Test	g/210L	Time	Air Blank	0.000	08:10	Control Test	0.079	08:11	Air Blank	0.000	08:12	Control Test	0.079	08:12	Air Blank	0.000	08:13	Control Test	0.079	08:13	Air Blank	0.000	08:14	Control Test Stats			Average	0.0790		Std Dev	0.0000		Rel Std Dev(%)	0.0000		HILLSBOROUGH CSO Intoxilizer - Alcohol Analyzer Model 8000 02/25/2026 Software: 8100.27 SN 80-007480 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>08:35</td></tr> <tr><td>Control Test</td><td>0.197</td><td>08:36</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>08:36</td></tr> <tr><td>Control Test</td><td>0.198</td><td>08:37</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>08:38</td></tr> <tr><td>Control Test</td><td>0.198</td><td>08:38</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>08:39</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1977</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.2921</td><td></td></tr> </tbody> </table> Operator's Signature _____	Test	g/210L	Time	Air Blank	0.000	08:35	Control Test	0.197	08:36	Air Blank	0.000	08:36	Control Test	0.198	08:37	Air Blank	0.000	08:38	Control Test	0.198	08:38	Air Blank	0.000	08:39	Control Test Stats			Average	0.1977		Std Dev	0.0006		Rel Std Dev(%)	0.2921		HILLSBOROUGH CSO Intoxilizer - Alcohol Analyzer Model 8000 02/25/2026 Software: 8100.27 SN 80-007480 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>08:03</td></tr> <tr><td>Control Test</td><td>0.080</td><td>08:03</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>08:04</td></tr> <tr><td>Control Test</td><td>0.080</td><td>08:04</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>08:05</td></tr> <tr><td>Control Test</td><td>0.080</td><td>08:05</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>08:06</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0800</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> Operator's Signature _____	Test	g/210L	Time	Air Blank	0.000	08:03	Control Test	0.080	08:03	Air Blank	0.000	08:04	Control Test	0.080	08:04	Air Blank	0.000	08:05	Control Test	0.080	08:05	Air Blank	0.000	08:06	Control Test Stats			Average	0.0800		Std Dev	0.0000		Rel Std Dev(%)	0.0000	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	08:20																																																																																																																																																	
Control Test	0.049	08:21																																																																																																																																																	
Air Blank	0.000	08:22																																																																																																																																																	
Control Test	0.048	08:22																																																																																																																																																	
Air Blank	0.000	08:23																																																																																																																																																	
Control Test	0.049	08:23																																																																																																																																																	
Air Blank	0.000	08:24																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0487																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	1.1863																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	08:10																																																																																																																																																	
Control Test	0.079	08:11																																																																																																																																																	
Air Blank	0.000	08:12																																																																																																																																																	
Control Test	0.079	08:12																																																																																																																																																	
Air Blank	0.000	08:13																																																																																																																																																	
Control Test	0.079	08:13																																																																																																																																																	
Air Blank	0.000	08:14																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0790																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	08:35																																																																																																																																																	
Control Test	0.197	08:36																																																																																																																																																	
Air Blank	0.000	08:36																																																																																																																																																	
Control Test	0.198	08:37																																																																																																																																																	
Air Blank	0.000	08:38																																																																																																																																																	
Control Test	0.198	08:38																																																																																																																																																	
Air Blank	0.000	08:39																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1977																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.2921																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	08:03																																																																																																																																																	
Control Test	0.080	08:03																																																																																																																																																	
Air Blank	0.000	08:04																																																																																																																																																	
Control Test	0.080	08:04																																																																																																																																																	
Air Blank	0.000	08:05																																																																																																																																																	
Control Test	0.080	08:05																																																																																																																																																	
Air Blank	0.000	08:06																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0800																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		

HILLSBOROUGH CSO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007480
02/25/2026
Software: 8100.27

DRY CONTROL TEST

Test	g/210L	Time
Air Blank	ABT*	08:25
Air Blank	0.000	08:26

*Sequence Aborted



Operator's Signature

Paper jammed. Reprinted from Database.
Inadvertently hit dry control test instead
of DMV menu.

 2/26/26

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: HILLSBOROUGH CSO
Time of Inspection: 08:48

Date of Inspection: 02/26/2026

Serial Number: 80-007480
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202406K Exp: 06/19/2026	0.08g/210L Test (g/210L) Lot#:202406L Exp: 06/19/2026	0.20g/210L Test (g/210L) Lot#:202406N Exp: 06/20/2026	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:28424080A3 Exp: 11/05/2026
0.000	0.047	0.078	0.196	0.080
0.000	0.048	0.078	0.197	0.080
0.000	0.049	0.078	0.197	0.079
0.000	0.048	0.078	0.197	0.079
0.000	0.048	0.078	0.197	0.079
0.000	0.048	0.078	0.197	0.079
0.000	0.048	0.078	0.197	0.079
0.000	0.048	0.078	0.197	0.079
0.000	0.048	0.078	0.197	0.079
0.000	0.048	0.079	0.197	0.078

Standard Deviations	0.0004	0.0003	0.0003	0.0005
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

KATIE T SPEARIN

Signature and Printed Name

02/26/2026
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-007480 , manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-007480</u>	UNCERTAINTY* $\pm$
Owning Agency:	<u>HILLSBOROUGH CSO</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>02/26/2026</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>08:48</u>	0.200 g/ 210 L 0.008
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).
The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.
This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Date 02/26/2026 KATIE T SPEARIN,

Department Inspector

FDLE/ATP Form 69 January 2026
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality