

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: WALTON COUNTY SO  
Time of Inspection: 10:51

Date of Inspection: 04/26/2022

Serial Number: 80-001311  
Software: 8100.27

| Check or Test                               | YES | NO |
|---|-----|----|
| Date and/or Time Adjusted                   |     | No |
| Diagnostic Check (Pre-Inspection): OK       | Yes |    |
| Alcohol Free Subject Test: 0.000            | Yes |    |
| Mouth Alcohol Test: Slope Not Met           | Yes |    |
| Interferent Detect Test: Interferent Detect | Yes |    |
| Diagnostic Check (Post-Inspection): OK      | Yes |    |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L)<br>Lot#:202101C<br>Exp: 01/12/2023 | 0.08g/210L Test (g/210L)<br>Lot#:202108D<br>Exp: 08/17/2023 | 0.20g/210L Test (g/210L)<br>Lot#:202106B<br>Exp: 06/22/2023 | 0.08 g/210L Dry Gas Std Test (g/210L)<br>Lot#:WO32302427<br>Exp: 09/11/2024 |
|----------------------------|---|---|---|---|
| 0.000                      | 0.048   | 0.078   | 0.197   | 0.080   |
| 0.000                      | 0.048   | 0.078   | 0.197   | 0.080   |
| 0.000                      | 0.049   | 0.078   | 0.197   | 0.080   |

Number of Simulators Used: 5

Remarks: *In Compliance* <sup>WLM</sup>

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*W/5 William P May #385* WILLIAM P MAY  
\_\_\_\_\_  
Signature and Printed Name

04/26/2022  
Date

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: WALTON COUNTY SO  
Time of Inspection: 08:23

Date of Inspection: 06/24/2022

Serial Number: 80-001311  
Software: 8100.27

| Check or Test                               | YES | NO |
|---|-----|----|
| Date and/or Time Adjusted                   |     | No |
| Diagnostic Check (Pre-Inspection): OK       | Yes |    |
| Alcohol Free Subject Test: 0.000            | Yes |    |
| Mouth Alcohol Test: Slope Not Met           | Yes |    |
| Interferent Detect Test: Interferent Detect | Yes |    |
| Diagnostic Check (Post-Inspection): OK      | Yes |    |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L)<br>Lot#:202101C<br>Exp: 01/12/2023 | 0.08g/210L Test (g/210L)<br>Lot#:202108D<br>Exp: 08/17/2023 | 0.20g/210L Test (g/210L)<br>Lot#:202106B<br>Exp: 06/22/2023 | 0.08 g/210L Dry Gas Std Test (g/210L)<br>Lot#:WO32302427<br>Exp: 09/11/2024 |
|----------------------------|---|---|---|---|
| 0.000                      | 0.048   | 0.078   | 0.197   | 0.080   |
| 0.000                      | 0.049   | 0.078   | 0.197   | 0.080   |
| 0.000                      | 0.049   | 0.079   | 0.197   | 0.080   |

Number of Simulators Used: 5

Remarks:

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*William P May* #383 WILLIAM P MAY  
Signature and Printed Name

06/24/2022  
Date

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: WALTON COUNTY SO  
Time of Inspection: 08:59

Date of Inspection: 07/29/2022

Serial Number: 80-001311  
Software: 8100.27

| Check or Test                               | YES | NO |
|---|-----|----|
| Date and/or Time Adjusted                   |     | No |
| Diagnostic Check (Pre-Inspection): OK       | Yes |    |
| Alcohol Free Subject Test: 0.000            | Yes |    |
| Mouth Alcohol Test: Slope Not Met           | Yes |    |
| Interferent Detect Test: Interferent Detect | Yes |    |
| Diagnostic Check (Post-Inspection): OK      | Yes |    |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L)<br>Lot#:202101C<br>Exp: 01/12/2023 | 0.08g/210L Test (g/210L)<br>Lot#:202108D<br>Exp: 08/17/2023 | 0.20g/210L Test (g/210L)<br>Lot#:202106B<br>Exp: 06/22/2023 | 0.08 g/210L Dry Gas Std Test (g/210L)<br>Lot#:W032302427<br>Exp: 09/11/2024 |
|----------------------------|---|---|---|---|
| 0.000                      | 0.048   | 0.000 / 0.079   | 0.198   | 0.080   |
| 0.000                      | 0.049   | 0.078 / 0.079   | 0.198   | 0.079   |
| 0.000                      | 0.048   | 0.079 / 0.079   | 0.198   | 0.079   |

Number of Simulators Used: 5

**Remarks:**

08: Control Outside FORGOT TO CNCT SIM.

Reconnected Sim - Passed .08

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

W. P. May #383 WILLIAM P MAY  
Signature and Printed Name

07/29/2022  
Date

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: WALTON COUNTY SO  
Time of Inspection: 06:46

Date of Inspection: 08/31/2022

Serial Number: 80-001311  
Software: 8100.27

| Check or Test                               | YES | NO |
|---|-----|----|
| Date and/or Time Adjusted                   |     | No |
| Diagnostic Check (Pre-Inspection): OK       | Yes |    |
| Alcohol Free Subject Test: 0.000            | Yes |    |
| Mouth Alcohol Test: Slope Not Met           | Yes |    |
| Interferent Detect Test: Interferent Detect | Yes |    |
| Diagnostic Check (Post-Inspection): OK      | Yes |    |


| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L)<br>Lot#:202101C<br>Exp: 01/12/2023 | 0.08g/210L Test (g/210L)<br>Lot#:202108D<br>Exp: 08/17/2023 | 0.20g/210L Test (g/210L)<br>Lot#:202106B<br>Exp: 06/22/2023 | 0.08 g/210L Dry Gas Std Test (g/210L)<br>Lot#:W032302427<br>Exp: 09/11/2024 |
|----------------------------|---|---|---|---|
| 0.000                      | 0.049   | 0.079   | 0.197   | 0.079   |
| 0.000                      | 0.049   | 0.078   | 0.198   | 0.079   |
| 0.000                      | 0.050   | 0.078   | 0.198   | 0.080   |

Number of Simulators Used: 5

Remarks:

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



WILLIAM P MAY

Signature and Printed Name

08/31/2022  
Date

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: WALTON COUNTY SO  
Time of Inspection: 11:47

Date of Inspection: 09/30/2022

Serial Number: 80-001311  
Software: 8100.27

| Check or Test                               | YES | NO |
|---|-----|----|
| Date and/or Time Adjusted                   |     | No |
| Diagnostic Check (Pre-Inspection): OK       | Yes |    |
| Alcohol Free Subject Test: 0.000            | Yes |    |
| Mouth Alcohol Test: Slope Not Met           | Yes |    |
| Interferent Detect Test: Interferent Detect | Yes |    |
| Diagnostic Check (Post-Inspection): OK      | Yes |    |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L)<br>Lot#:202101C<br>Exp: 01/12/2023 | 0.08g/210L Test (g/210L)<br>Lot#:202108D<br>Exp: 08/17/2023 | 0.20g/210L Test (g/210L)<br>Lot#:202106B<br>Exp: 06/22/2023 | 0.08 g/210L Dry Gas Std Test (g/210L)<br>Lot#:WO32302427<br>Exp: 09/11/2024 |
|----------------------------|---|---|---|---|
| 0.000                      | 0.049   | 0.079   | 0.197   | 0.080   |
| 0.000                      | 0.049   | 0.079   | 0.198   | 0.080   |
| 0.000                      | 0.049   | 0.079   | 0.198   | 0.080   |

Number of Simulators Used: 5

Remarks:

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*William P May #383*

WILLIAM P MAY

Signature and Printed Name

09/30/2022  
Date

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: WALTON COUNTY SO  
Time of Inspection: 11:43

Date of Inspection: 10/19/2022

Serial Number: 80-001311  
Software: 8100.27

| Check or Test                               | YES | NO |
|---|-----|----|
| Date and/or Time Adjusted                   |     | No |
| Diagnostic Check (Pre-Inspection): OK       | Yes |    |
| Alcohol Free Subject Test: 0.000            | Yes |    |
| Mouth Alcohol Test: Slope Not Met           | Yes |    |
| Interferent Detect Test: Interferent Detect | Yes |    |
| Diagnostic Check (Post-Inspection): OK      | Yes |    |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L)<br>Lot#:202101C<br>Exp: 01/12/2023 | 0.08g/210L Test (g/210L)<br>Lot#:202108D<br>Exp: 08/17/2023 | 0.20g/210L Test (g/210L)<br>Lot#:202106B<br>Exp: 06/22/2023 | 0.08 g/210L Dry Gas Std Test (g/210L)<br>Lot#:W032302427<br>Exp: 09/11/2024 |
|----------------------------|---|---|---|---|
| 0.000                      | 0.050   | 0.079   | 0.197   | 0.081   |
| 0.000                      | 0.050   | 0.079   | 0.198   | 0.080   |
| 0.000                      | 0.050   | 0.079   | 0.199   | 0.081   |

Number of Simulators Used: 5

Remarks:

Taken out of Service  
sent to FDLE  
for Department Inspection - Annual

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

D/S William P. May #383 WILLIAM P MAY  
Signature and Printed Name

10/19/2022  
Date

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: WALTON COUNTY SO  
Time of Inspection: 16:45

Date of Inspection: 11/30/2022

Serial Number: 80-001311  
Software: 8100.27

| Check or Test                               | YES | NO |
|---|-----|----|
| Date and/or Time Adjusted                   |     | No |
| Diagnostic Check (Pre-Inspection): OK       | Yes |    |
| Alcohol Free Subject Test: 0.000            | Yes |    |
| Mouth Alcohol Test: Slope Not Met           | Yes |    |
| Interferent Detect Test: Interferent Detect | Yes |    |
| Diagnostic Check (Post-Inspection): OK      | Yes |    |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L)<br>Lot#:202101C<br>Exp: 01/12/2023 | 0.08g/210L Test (g/210L)<br>Lot#:202108D<br>Exp: 08/17/2023 | 0.20g/210L Test (g/210L)<br>Lot#:202106B<br>Exp: 06/22/2023 | 0.08 g/210L Dry Gas Std Test (g/210L)<br>Lot#:WO32302427<br>Exp: 09/11/2024 |
|----------------------------|---|---|---|---|
| 0.000                      | 0.049   | 0.079   | 0.198   | 0.080   |
| 0.000                      | 0.049   | 0.079   | 0.197   | 0.079   |
| 0.000                      | 0.049   | 0.079   | 0.198   | 0.080   |

Number of Simulators Used: 5

Remarks:

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

W. P. May #383 WILLIAM P MAY  
 \_\_\_\_\_  
 Signature and Printed Name

11/30/2022  
Date



# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: WALTON COUNTY SO  
Time of Inspection: 11:25

Date of Inspection: 12/29/2022

Serial Number: 80-001311  
Software: 8100.27

| Check or Test                               | YES | NO |
|---|-----|----|
| Date and/or Time Adjusted                   |     | No |
| Diagnostic Check (Pre-Inspection): OK       | Yes |    |
| Alcohol Free Subject Test: 0.000            | Yes |    |
| Mouth Alcohol Test: Slope Not Met           | Yes |    |
| Interferent Detect Test: Interferent Detect | Yes |    |
| Diagnostic Check (Post-Inspection): OK      | Yes |    |


| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L)<br>Lot#:202101C<br>Exp: 01/12/2023 | 0.08g/210L Test (g/210L)<br>Lot#:202108D<br>Exp: 08/17/2023 | 0.20g/210L Test (g/210L)<br>Lot#:202106B<br>Exp: 06/22/2023 | 0.08 g/210L Dry Gas Std Test (g/210L)<br>Lot#:WO32302427<br>Exp: 09/11/2024 |
|----------------------------|---|---|---|---|
| 0.000                      | 0.050   | 0.078   | 0.198   | 0.080   |
| 0.000                      | 0.049   | 0.079   | 0.198   | 0.080   |
| 0.000                      | 0.049   | 0.079   | 0.198   | 0.080   |

Number of Simulators Used: 5

Remarks:

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

 #383

WILLIAM P MAY

Signature and Printed Name

12/29/2022  
Date