



INSTRUMENT PROCESSING SHEET

Agency FL HIGHWAY PATROLS/N 80-003409Florida Department of
Law EnforcementDate In 9/8/2025DI Completion Date 12/30/2025 Ship P/U H/D CMI EE

Intake	By	Date	Quality Checks	By	SLH	Date	12/30/2025	Flow Calibration	By	Date																
<input type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE Visual Inspection: <input type="checkbox"/> Case <input type="checkbox"/> Handle <input type="checkbox"/> Keyboard <input type="checkbox"/> Dry Gas Shelf <input type="checkbox"/> Feet <input type="checkbox"/> Breath Tube <input type="checkbox"/> Ports <input type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: <u>Already performed Intake 9/8/25 by WKP. See attached record. SLH 12/30/2025</u>			<input type="checkbox"/> Breath Tube Screen <input type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Flow Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547) <input type="checkbox"/> Barometric Pressure Check Gauge ID # _____ <input type="checkbox"/> Stability Checks					Flow Column # _____ <input type="checkbox"/> 5L/min – 17mm <input type="checkbox"/> 15L/min – 53mm <input type="checkbox"/> 30L/min – 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																		
			<table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> </tr> </tbody> </table>			Simulator	Serial #	Lot #/Exp	0.050			0.080			0.200			0.080 DGS	N/A					Maintenance By <u>SLH/SP</u> Date <u>10/9/2025</u> <input checked="" type="checkbox"/> Battery Replacement (x2) <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other <u>Replaced battery again on 12/30/2025 due to time/date still not correct. SLH 12/30/2025</u>		
Simulator	Serial #	Lot #/Exp																								
0.050																										
0.080																										
0.200																										
0.080 DGS	N/A																									

Calibration Adjustment	Department Inspection																																																												
By _____ Barometric Pressure Gauge ID # _____ <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> <input type="checkbox"/> Post Calibration Adjustment Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>	Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			By <u>SLH</u> Barometric Pressure ID# <u>28427</u> Gauge <u>1017</u> Instrument <u>1012</u> Mouth Alcohol Solution Lot # <u>2025-D expire 9/25/27</u> Acetone Stock Solution Lot # <u>2025-B expire 9/22/27</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>MP6294</td> </tr> <tr> <td>Interferent</td> <td>MP6290</td> </tr> <tr> <td>0.050</td> <td>MP5088</td> </tr> <tr> <td>0.080</td> <td>MP5089</td> </tr> <tr> <td>0.200</td> <td>MP5090</td> </tr> </tbody> </table> Attachments <input checked="" type="checkbox"/> Form 41 <input type="checkbox"/> Stability Checks <input checked="" type="checkbox"/> Calibration Certificate <input type="checkbox"/> Calibration Adjustment <input type="checkbox"/> Post-Stability Checks <input type="checkbox"/> Flow Calibration <input checked="" type="checkbox"/> Form 40 calibration/repair packet originally submitted for review <input checked="" type="checkbox"/> Other _____	Simulator	Serial Number	0.000	MP6294	Interferent	MP6290	0.050	MP5088	0.080	MP5089	0.200	MP5090
Simulator	Serial #	Lot #	Expiration																																																										
0.000		N/A	N/A																																																										
0.040																																																													
0.100																																																													
0.200																																																													
0.300																																																													
0.080 DGS	N/A																																																												
Simulator	Serial #	Lot #	Expiration																																																										
0.050																																																													
0.080																																																													
0.200																																																													
0.080 DGS	N/A																																																												
Simulator	Serial Number																																																												
0.000	MP6294																																																												
Interferent	MP6290																																																												
0.050	MP5088																																																												
0.080	MP5089																																																												
0.200	MP5090																																																												

Notes/Suggested Service: <u>There are 2 IPS forms because the instrument was originally processed to go to repair, but the decision was made to retain until FDLE could replace batteries. The record was withdrawn from review. SLH 12/30/2025</u> <u>Quality Checks were performed 9/9/25 by WKP. See attached record. Settings were again verified after the 12/30/25 battery change. SLH 12/30/25</u>	<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use
Taylor Gutschow <small>Digitally signed by Taylor Gutschow Date: 2026.01.13 14:12:02 -0500</small>	Wen-Chi Pierson <small>Digitally signed by Wen-Chi Pierson Date: 2026.01.15 08:19:54 -0500</small>
Tech Review / Date	Admin Review / Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FL HIGHWAY PATROL
Time of Inspection: 12:37

Date of Inspection: 12/30/2025

Serial Number: 80-003409
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08g/210L Test (g/210L) Lot#: _____ Exp: _____	0.20g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: _____ Exp: _____


Number of Simulators Used: _____

Remarks:

BYPASS AI FOR OPERATION, COMPLIANCE NOT DETERMINED

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



LEANDRA HIGGINBOTHAM

Signature and Printed Name

12/30/2025
Date

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: FL HIGHWAY PATROL
Time of Inspection: 14:49

Date of Inspection: 12/30/2025

Serial Number: 80-003409
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202406K Exp: 06/19/2026	0.08g/210L Test (g/210L) Lot#:202406L Exp: 06/19/2026	0.20g/210L Test (g/210L) Lot#:202406N Exp: 06/20/2026	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG510701 Exp: 04/17/2027
0.000	0.048	0.077	0.195	0.080
0.000	0.049	0.078	0.195	0.080
0.000	0.049	0.078	0.195	0.080
0.000	0.048	0.078	0.195	0.081
0.000	0.048	0.078	0.196	0.081
0.000	0.048	0.079	0.195	0.081
0.000	0.048	0.079	0.196	0.081
0.000	0.048	0.078	0.196	0.081
0.000	0.048	0.079	0.196	0.081
0.000	0.048	0.078	0.196	0.081
0.000	0.049	0.080	0.197	0.082


Standard Deviations	0.0004	0.0008	0.0006	0.0006
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0006 Number of Simulators Used: 5



Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.


 LEANDRA HIGGINBOTHAM
 Signature and Printed Name

12/30/2025
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-003409, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-003409</u>	UNCERTAINTY* ±	
Owning Agency:	<u>FL HIGHWAY PATROL</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>12/30/2025</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>14:49</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

M. J. Latta
12/30/2025
Date

LEANDRA HIGGINBOTHAM,
Department Inspector

FDLE/ATP Form 69 October 2024
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Stability Checks

0.05g/210L 0.047 to 0.053	0.08g/210L 0.077 to 0.083	0.20g/210L 0.194 to 0.206	DGS 0.08g/210L 0.077 to 0.083 ≤0.003 of Wet																																																																																																																																																
Performed Root Case Analysis		Performed Root Case Analysis																																																																																																																																																	
<p>FL HIGHWAY PATROL Intoxilyzer - Alcohol Analyzer Model 8100 SN 80-003409 09/09/2025 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>08:52</td></tr> <tr><td>Control Test</td><td>0.048</td><td>08:53</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>08:53</td></tr> <tr><td>Control Test</td><td>0.048</td><td>08:54</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>08:55</td></tr> <tr><td>Control Test</td><td>0.048</td><td>08:55</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>08:56</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0480</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> <p><i>Wemp</i> Operator's Signature</p>	Test	g/210L	Time	Air Blank	0.000	08:52	Control Test	0.048	08:53	Air Blank	0.000	08:53	Control Test	0.048	08:54	Air Blank	0.000	08:55	Control Test	0.048	08:55	Air Blank	0.000	08:56	Control Test Stats			Average	0.0480		Std Dev	0.0000		Rel Std Dev(%)	0.0000		<p>FL HIGHWAY PATROL Intoxilyzer - Alcohol Analyzer Model 8100 SN 80-003409 09/09/2025 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>08:57</td></tr> <tr><td>Control Test</td><td>0.078</td><td>08:58</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>08:58</td></tr> <tr><td>Control Test</td><td>0.078</td><td>08:59</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>08:59</td></tr> <tr><td>Control Test</td><td>0.078</td><td>09:00</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:00</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0780</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> <p><i>Wemp</i> Operator's Signature</p>	Test	g/210L	Time	Air Blank	0.000	08:57	Control Test	0.078	08:58	Air Blank	0.000	08:58	Control Test	0.078	08:59	Air Blank	0.000	08:59	Control Test	0.078	09:00	Air Blank	0.000	09:00	Control Test Stats			Average	0.0780		Std Dev	0.0000		Rel Std Dev(%)	0.0000		<p>FL HIGHWAY PATROL Intoxilyzer - Alcohol Analyzer Model 8100 SN 80-003409 09/09/2025 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>09:04</td></tr> <tr><td>Control Test</td><td>0.198</td><td>09:05</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:06</td></tr> <tr><td>Control Test</td><td>0.198</td><td>09:06</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:07</td></tr> <tr><td>Control Test</td><td>0.197</td><td>09:07</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:08</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1977</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.2921</td><td></td></tr> </tbody> </table> <p><i>Wemp</i> Operator's Signature</p>	Test	g/210L	Time	Air Blank	0.000	09:04	Control Test	0.198	09:05	Air Blank	0.000	09:06	Control Test	0.198	09:06	Air Blank	0.000	09:07	Control Test	0.197	09:07	Air Blank	0.000	09:08	Control Test Stats			Average	0.1977		Std Dev	0.0006		Rel Std Dev(%)	0.2921		<p>FL HIGHWAY PATROL Intoxilyzer - Alcohol Analyzer Model 8100 SN 80-003409 09/09/2025 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>08:40</td></tr> <tr><td>Control Test</td><td>0.078</td><td>08:40</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>08:40</td></tr> <tr><td>Control Test</td><td>0.060</td><td>08:41</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>08:41</td></tr> <tr><td>Control Test</td><td>0.079</td><td>08:42</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>08:42</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0790</td><td></td></tr> <tr><td>Std Dev</td><td>0.0010</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.2658</td><td></td></tr> </tbody> </table> <p><i>Wemp</i> Operator's Signature</p>	Test	g/210L	Time	Air Blank	0.000	08:40	Control Test	0.078	08:40	Air Blank	0.000	08:40	Control Test	0.060	08:41	Air Blank	0.000	08:41	Control Test	0.079	08:42	Air Blank	0.000	08:42	Control Test Stats			Average	0.0790		Std Dev	0.0010		Rel Std Dev(%)	1.2658	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	08:52																																																																																																																																																	
Control Test	0.048	08:53																																																																																																																																																	
Air Blank	0.000	08:53																																																																																																																																																	
Control Test	0.048	08:54																																																																																																																																																	
Air Blank	0.000	08:55																																																																																																																																																	
Control Test	0.048	08:55																																																																																																																																																	
Air Blank	0.000	08:56																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0480																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	08:57																																																																																																																																																	
Control Test	0.078	08:58																																																																																																																																																	
Air Blank	0.000	08:58																																																																																																																																																	
Control Test	0.078	08:59																																																																																																																																																	
Air Blank	0.000	08:59																																																																																																																																																	
Control Test	0.078	09:00																																																																																																																																																	
Air Blank	0.000	09:00																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0780																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	09:04																																																																																																																																																	
Control Test	0.198	09:05																																																																																																																																																	
Air Blank	0.000	09:06																																																																																																																																																	
Control Test	0.198	09:06																																																																																																																																																	
Air Blank	0.000	09:07																																																																																																																																																	
Control Test	0.197	09:07																																																																																																																																																	
Air Blank	0.000	09:08																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1977																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.2921																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	08:40																																																																																																																																																	
Control Test	0.078	08:40																																																																																																																																																	
Air Blank	0.000	08:40																																																																																																																																																	
Control Test	0.060	08:41																																																																																																																																																	
Air Blank	0.000	08:41																																																																																																																																																	
Control Test	0.079	08:42																																																																																																																																																	
Air Blank	0.000	08:42																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0790																																																																																																																																																		
Std Dev	0.0010																																																																																																																																																		
Rel Std Dev(%)	1.2658																																																																																																																																																		

Return Material Authorization

Ship to: CMI, Inc.
 Enforcement Electronics

Shipment to repair facility authorized by: Taylor Ledford on 9/15/2025

Items Returned: Instrument Supplies Other Describe: _____
Instrument Model: Intoxilyzer 8000 Serial Number: 80-003409

Bill To Address: FHP _____ 5023 53rd Ave _____ Bradenton, FL 34203 _____ _____	Ship to Address: FDLE Tallahassee _____ _____ _____ _____
---	--

Reason for Return:
Battery replacement required

Please choose one of the following options:

- 1. I _____, authorize all repairs.
- 2. I _____, authorize repairs up to \$_____.
- 3. I require an estimate **BEFORE** any repairs will be authorized and/ or conducted.

Please contact: Name: Taylor Ledford
Phone #: 941-962-7782 Email: TaylorLedford@flhsmv.gov

ATP Contact Name: LeAndra Higginbotham ATP Email: LeAndraHigginbotham@fdle.state.fl.us