

# Instrument Processing Sheet

Agency: BISCAYNE NAT PARK

Instrument Serial Number: 80-001442

Date In: 12/19/2025

DI Completion Date: 12/30/2025 and 12/23/2025

Ship  P/U  H/D  CMI  EE

<b>Intake</b> By: <u>WKP</u> Date: <u>12/19/25</u> <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE <input type="checkbox"/> Return unworked <input type="checkbox"/> Training Visual Inspection <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/Accessories <input checked="" type="checkbox"/> Power Cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes:	<b>Quality Checks</b> By: <u>WKP</u> Date: <u>12/22/25</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value: <u>244</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column #: <u>ATP 103</u> 32 mm <u>0.156</u> (.139-.169) 36 mm <u>0.171</u> (.156-.190) 53 mm <u>0.242</u> (.228-.278) 103 mm <u>0.500</u> (.447-.547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID #: <u>28662</u> Gauge: <u>1023</u> Instrument: <u>1023</u> <input checked="" type="checkbox"/> Stability Checks	<b>Flow Adjustment</b> By: _____ Date: _____ Flow Column #: _____ <input type="checkbox"/> 5L/min – 17mm <input type="checkbox"/> 15L/min – 53mm <input type="checkbox"/> 30L/min – 103mm <input type="checkbox"/> R-Value: _____ <input type="checkbox"/> Post Adjustment Verification (L/S) Flow Column #: _____ 32 mm _____ (.139-.169) 36 mm _____ (.156-.190) 53 mm _____ (.228-.278) 103 mm _____ (.447-.547)																								
<table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot#/Exp</th> <th>Maintenance</th> </tr> </thead> <tbody> <tr> <td rowspan="2">0.050</td> <td rowspan="2">MP5088</td> <td>202406K</td> <td rowspan="2"><input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement and Tank Sensor Tare <input type="checkbox"/> Breath Tube Replacement <input checked="" type="checkbox"/> Other: replaced internal printer paper</td> </tr> <tr> <td>06/19/2026</td> </tr> <tr> <td rowspan="2">0.080</td> <td rowspan="2">MP5089</td> <td>202406L</td> <td rowspan="2"></td> </tr> <tr> <td>06/19/2026</td> </tr> <tr> <td rowspan="2">0.200</td> <td rowspan="2">MP5090</td> <td>202406N</td> <td rowspan="2"></td> </tr> <tr> <td>06/20/2026</td> </tr> <tr> <td rowspan="2">0.080 DGS</td> <td rowspan="2">N/A</td> <td>AG510701</td> <td rowspan="2"></td> </tr> <tr> <td>04/17/2027</td> </tr> </tbody> </table>			Simulator	Serial #	Lot#/Exp	Maintenance	0.050	MP5088	202406K	<input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement and Tank Sensor Tare <input type="checkbox"/> Breath Tube Replacement <input checked="" type="checkbox"/> Other: replaced internal printer paper	06/19/2026	0.080	MP5089	202406L		06/19/2026	0.200	MP5090	202406N		06/20/2026	0.080 DGS	N/A	AG510701		04/17/2027
Simulator	Serial #	Lot#/Exp	Maintenance																							
0.050	MP5088	202406K	<input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement and Tank Sensor Tare <input type="checkbox"/> Breath Tube Replacement <input checked="" type="checkbox"/> Other: replaced internal printer paper																							
		06/19/2026																								
0.080	MP5089	202406L																								
		06/19/2026																								
0.200	MP5090	202406N																								
		06/20/2026																								
0.080 DGS	N/A	AG510701																								
		04/17/2027																								

<b>Optical Bench Adjustment</b> By: <u>SLH</u> Barometric Pressure Gauge: <u>1011</u> ID#: <u>28427</u> <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>MP6289</td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td>MP6295</td> <td>25090</td> <td>03/11/27</td> </tr> <tr> <td>0.100</td> <td>MP6296</td> <td>25282</td> <td>08/12/27</td> </tr> <tr> <td>0.200</td> <td>MP6297</td> <td>24080</td> <td>02/13/26</td> </tr> <tr> <td>0.300</td> <td>MP6298</td> <td>25150</td> <td>05/06/27</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>28424080A3</td> <td>11/05/2026</td> </tr> </tbody> </table> <input checked="" type="checkbox"/> Post Optical Bench Adjustment Stability Checks <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>MP5088</td> <td>202406K</td> <td>06/19/2026</td> </tr> <tr> <td>0.080</td> <td>MP5089</td> <td>202406L</td> <td>06/19/2026</td> </tr> <tr> <td>0.200</td> <td>MP5090</td> <td>202406N</td> <td>06/20/2026</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG510701</td> <td>04/17/2027</td> </tr> </tbody> </table> Barometric Pressure Gauge: <u>1013</u> ID#: <u>28421</u>	Simulator	Serial #	Lot #	Expiration	0.000	MP6289	N/A	N/A	0.040	MP6295	25090	03/11/27	0.100	MP6296	25282	08/12/27	0.200	MP6297	24080	02/13/26	0.300	MP6298	25150	05/06/27	0.080 DGS	N/A	28424080A3	11/05/2026	Simulator	Serial #	Lot #	Expiration	0.050	MP5088	202406K	06/19/2026	0.080	MP5089	202406L	06/19/2026	0.200	MP5090	202406N	06/20/2026	0.080 DGS	N/A	AG510701	04/17/2027	<b>Department Inspection</b> By: <u>SLH</u> Barometric Pressure ID#: <u>28421 / 28421</u> Gauge: <u>1025 / 1019</u> Instrument: <u>1024 / 1018</u> Mouth Alcohol Solution Lot #: <u>2025-D / same</u> Exp: <u>09/25/2027</u> Acetone Stock Solution Lot #: <u>2025-B / same</u> Exp: <u>09/22/2027</u> <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>MP6294</td> </tr> <tr> <td>Interferent</td> <td>MP6290</td> </tr> <tr> <td>0.050</td> <td>MP5088</td> </tr> <tr> <td>0.080</td> <td>MP5089</td> </tr> <tr> <td>0.200</td> <td>MP5090</td> </tr> </tbody> </table> <b>Attachments</b> <input checked="" type="checkbox"/> Form 41 (x2) <input checked="" type="checkbox"/> Post-Stability Checks <input checked="" type="checkbox"/> Stability Checks <input type="checkbox"/> Flow Adjustment <input checked="" type="checkbox"/> Calibration Certificate <input checked="" type="checkbox"/> Form 40 <input checked="" type="checkbox"/> Optical Bench Adjustment <input type="checkbox"/> Other:	Simulator	Serial Number	0.000	MP6294	Interferent	MP6290	0.050	MP5088	0.080	MP5089	0.200	MP5090
Simulator	Serial #	Lot #	Expiration																																																										
0.000	MP6289	N/A	N/A																																																										
0.040	MP6295	25090	03/11/27																																																										
0.100	MP6296	25282	08/12/27																																																										
0.200	MP6297	24080	02/13/26																																																										
0.300	MP6298	25150	05/06/27																																																										
0.080 DGS	N/A	28424080A3	11/05/2026																																																										
Simulator	Serial #	Lot #	Expiration																																																										
0.050	MP5088	202406K	06/19/2026																																																										
0.080	MP5089	202406L	06/19/2026																																																										
0.200	MP5090	202406N	06/20/2026																																																										
0.080 DGS	N/A	AG510701	04/17/2027																																																										
Simulator	Serial Number																																																												
0.000	MP6294																																																												
Interferent	MP6290																																																												
0.050	MP5088																																																												
0.080	MP5089																																																												
0.200	MP5090																																																												

Notes/Suggested Service: Due to Interferent Detect message for two levels during the inspection and results falling at the lower acceptability range, I will perform a optical bench adjustment. It is noted the DI conducted 12/23/2025 was compliant. SLH 12/23/25 Discretionary optical bench adjustment performed 12/29/2025. SLH 12/30/2025 performed DI post optical bench adjustment. There were two compliant inspections performed, 12/23 and 12/30/2025. SIM info for 2nd DI were the same. SLH 12/30/25	<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use <div style="display: flex; justify-content: space-between; font-size: small;"> <div style="text-align: center;">                         Digitally signed by Taylor Gutschow                          Date: 2026.01.13 13:50:54 -05'00'                     </div> <div style="text-align: center;">                         Digitally signed by Wen-Chi Pierson                          Date: 2026.01.14 10:05:29 -05'00'                     </div> </div>
<b>Tech Review</b>	<b>Admin Review</b>

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: BISCAYNE NAT PARK  
Time of Inspection: 14:33

Date of Inspection: 12/22/2025

Serial Number: 80-001442  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08g/210L Test (g/210L) Lot#: _____ Exp: _____	0.20g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: _____ Exp: _____

Number of Simulators Used: \_\_\_\_\_

Remarks:  
BYPASS AI TO OPERATE INSTRUMENT. COMPLIANCE NOT DETERMINED.

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



Signature and Printed Name

WEN-CHI K PIERSON

12/22/2025  
Date

# Stability Checks

0.05g/210L 0.047 to 0.053	0.08g/210L 0.077 to 0.083	0.20g/210L 0.194 to 0.206	DGS 0.08g/210L 0.077 to 0.083	≤0.003 of Wet																																																																								
✓	✓	✓	✓	✓																																																																								
Performed Root Case Analysis	Performed Root Case Analysis	Performed Root Case Analysis	Performed Root Case Analysis	Performed Root Case Analysis																																																																								
<p>BISCAYNE NAT PARK Intoxilyzer - Alcohol Analyzer Model 8100 SN 80-101442 12/22/2025 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>15:37</td></tr> <tr><td>Control Test</td><td>0.046</td><td>15:38</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:38</td></tr> <tr><td>Control Test</td><td>0.047</td><td>15:39</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:40</td></tr> <tr><td>Control Test</td><td>0.047</td><td>15:40</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:41</td></tr> <tr><td colspan="3">Control Test Stats</td></tr> <tr><td>Average</td><td>0.0473</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.2198</td><td></td></tr> </tbody> </table> <p><i>Wen</i> Operator's Signature</p>	Test	g/210L	Time	Air Blank	0.000	15:37	Control Test	0.046	15:38	Air Blank	0.000	15:38	Control Test	0.047	15:39	Air Blank	0.000	15:40	Control Test	0.047	15:40	Air Blank	0.000	15:41	Control Test Stats			Average	0.0473		Std Dev	0.0006		Rel Std Dev(%)	1.2198		<p>View stability sheet print out</p>	<p>View stability sheet print out</p>	<p>BISCAYNE NAT PARK Intoxilyzer - Alcohol Analyzer Model 8100 SN 80-101442 12/22/2025 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>15:26</td></tr> <tr><td>Control Test</td><td>0.081</td><td>15:26</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:27</td></tr> <tr><td>Control Test</td><td>0.080</td><td>15:27</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:28</td></tr> <tr><td>Control Test</td><td>0.080</td><td>15:28</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:29</td></tr> <tr><td colspan="3">Control Test Stats</td></tr> <tr><td>Average</td><td>0.0803</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7187</td><td></td></tr> </tbody> </table> <p><i>Wen</i> Operator's Signature</p>	Test	g/210L	Time	Air Blank	0.000	15:26	Control Test	0.081	15:26	Air Blank	0.000	15:27	Control Test	0.080	15:27	Air Blank	0.000	15:28	Control Test	0.080	15:28	Air Blank	0.000	15:29	Control Test Stats			Average	0.0803		Std Dev	0.0006		Rel Std Dev(%)	0.7187		
Test	g/210L	Time																																																																										
Air Blank	0.000	15:37																																																																										
Control Test	0.046	15:38																																																																										
Air Blank	0.000	15:38																																																																										
Control Test	0.047	15:39																																																																										
Air Blank	0.000	15:40																																																																										
Control Test	0.047	15:40																																																																										
Air Blank	0.000	15:41																																																																										
Control Test Stats																																																																												
Average	0.0473																																																																											
Std Dev	0.0006																																																																											
Rel Std Dev(%)	1.2198																																																																											
Test	g/210L	Time																																																																										
Air Blank	0.000	15:26																																																																										
Control Test	0.081	15:26																																																																										
Air Blank	0.000	15:27																																																																										
Control Test	0.080	15:27																																																																										
Air Blank	0.000	15:28																																																																										
Control Test	0.080	15:28																																																																										
Air Blank	0.000	15:29																																																																										
Control Test Stats																																																																												
Average	0.0803																																																																											
Std Dev	0.0006																																																																											
Rel Std Dev(%)	0.7187																																																																											



#3

0.05g/210L net

Inadvertently

conducted

analysis MCP

12/22/25

Wenp

BISCAYNE NAT PARK  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-001442  
 12/22/2025  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	15:46
Control Test	0.047	15:47
Air Blank	0.000	15:48
Control Test	0.047	15:48
Air Blank	0.000	15:49
Control Test	0.047	15:50
Air Blank	0.000	15:50
Control Test Stats		
Average	0.0470	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	



Operator's Signature

0.08g/210L

Wet

BISCAYNE NAT PARK  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001442  
12/22/2025  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	15:51
Control Test	0.078	15:52
Air Blank	0.000	15:52
Control Test	0.077	15:53
Air Blank	0.000	15:54
Control Test	0.077	15:54
Air Blank	0.000	15:55
Control Test Stats		
Average	0.0773	
Std Dev	0.0006	
Rel Std Dev(%)	0.7466	



Operator's Signature

0.20g/210L  
wet

BISCAYNE NAT PARK  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001442  
12/22/2025  
Software: 8100.27

Test	g/210L	Time
-----	-----	-----
Air Blank	0.000	15:56
Control Test	0.199	15:57
Air Blank	0.000	15:57
Control Test	0.198	15:58
Air Blank	0.000	15:59
Control Test	0.197	15:59
Air Blank	0.000	16:00
Control Test Stats		
Average	0.1980	
Std Dev	0.0010	
Rel Std Dev(%)	0.5051	

  
Operator's Signature

# Florida Department of Law Enforcement Alcohol Testing Program

## DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: BISCAYNE NAT PARK  
Time of Inspection: 16:12

Date of Inspection: 12/23/2025

Serial Number: 80-001442  
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202406K Exp: 06/19/2026	0.08g/210L Test (g/210L) Lot#:202406L Exp: 06/19/2026	0.20g/210L Test (g/210L) Lot#:202406N Exp: 06/20/2026	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG510701 Exp: 04/17/2027
0.000 / 0.000	0.045	0.075	0.195	0.079 / 0.080
0.000 / 0.000	0.046	0.075	0.194	0.079 / 0.080
0.000 / 0.000	0.045	0.075	0.194	0.079 / 0.082
0.000 / 0.000	0.046	0.075	0.194	0.079 / 0.081
0.000 / 0.000	0.046	0.075	0.195	0.079 / 0.081
0.000 / 0.000	0.046	0.075	0.194	0.079 / 0.080
0.000 / 0.000	0.046	0.075	0.195	0.078 / 0.079
0.000 / 0.000	0.046	0.075	0.194	0.078 / 0.079
INT / 0.000	0.046	0.075	0.194	INT / 0.078
/ 0.000	0.046	0.075	0.195	/ 0.078
Standard Deviations	0.0004	0.0000	0.0005	/ 0.0013


Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0005 Number of Simulators Used: 5

**Remarks:**

00: Interferent Detect; TURNED FAN ON. 08: Interferent Detect; APPLIED FAN TOWARD INSTRUMENT.

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



LEANDRA HIGGINBOTHAM

Signature and Printed Name

12/23/2025  
Date

BISCAYNE NAT PARK  
Intoxilyzer - Alcohol Analyzer  
Model 8000  
12/29/2025

SN 80-001442 ✓  
12:49:48

Auto Calibration

pg 1 of 2

<<<<< 3um >>>>> <<<<< 9um >>>>>

-----  
Solution = 0.000 g/210L or 0.0000 mg/l, Samples = 4, Discarded = 1  
Sample % Abs (% Abs Ref) % Abs (% Abs Ref)  
Sample #1 0.1080 (0.0000) 0.1360 (-0.0040)  
Sample #2 0.1020 (-0.0020) 0.1420 (-0.0050)  
Sample #3 0.1100 (0.0000) 0.1390 (-0.0050)  
Sample #4 0.1280 (-0.0050) 0.1270 (-0.0130)  
Avg % Abs 0.1133 (-0.0023) 0.1360 (-0.0077)  
STD DEV 0.0133 (0.0025) 0.0079 (0.0046)  
REL STD DEV 11.750 (107.855) 5.836 (60.245)  
-----

-----  
Solution = 0.040 g/210L or 0.1905 mg/l, Samples = 4, Discarded = 1  
Sample % Abs (% Abs Ref) % Abs (% Abs Ref)  
Sample #1 0.7970 (0.0080) 1.5240 (-0.0030)  
Sample #2 0.8040 (0.0180) 1.5170 (0.0020)  
Sample #3 0.8000 (0.0200) 1.4810 (0.0200)  
Sample #4 0.7990 (0.0290) 1.5000 (0.0180)  
Avg % Abs 0.8010 (0.0223) 1.4993 (0.0133)  
STD DEV 0.0026 (0.0059) 0.0180 (0.0099)  
REL STD DEV 0.330 (26.236) 1.201 (73.993)  
-----

-----  
Solution = 0.100 g/210L or 0.4762 mg/l, Samples = 4, Discarded = 1  
Sample % Abs (% Abs Ref) % Abs (% Abs Ref)  
Sample #1 1.8380 (-0.0030) 3.5210 (0.0040)  
Sample #2 1.8070 (0.0300) 3.4880 (0.0370)  
Sample #3 1.8000 (0.0350) 3.4860 (0.0530)  
Sample #4 1.7920 (0.0430) 3.4740 (0.0550)  
Avg % Abs 1.7997 (0.0360) 3.4827 (0.0483)  
STD DEV 0.0075 (0.0066) 0.0076 (0.0099)  
REL STD DEV 0.417 (18.215) 0.217 (20.412)  
-----

-----  
Solution = 0.200 g/210L or 0.9524 mg/l, Samples = 4, Discarded = 1  
Sample % Abs (% Abs Ref) % Abs (% Abs Ref)  
Sample #1 3.5040 (0.0020) 6.7310 (0.0030)  
Sample #2 3.4740 (0.0450) 6.6840 (0.0800)  
Sample #3 3.4720 (0.0530) 6.6470 (0.0890)  
Sample #4 3.4560 (0.0550) 6.6270 (0.1040)  
Avg % Abs 3.4673 (0.0510) 6.6527 (0.0910)  
STD DEV 0.0099 (0.0053) 0.0289 (0.0121)  
REL STD DEV 0.285 (10.375) 0.435 (13.323)  
-----

-----  
Solution = 0.300 g/210L or 1.4286 mg/l, Samples = 4, Discarded = 1  
Sample % Abs (% Abs Ref) % Abs (% Abs Ref)  
Sample #1 5.1060 (-0.0130) 9.7560 (-0.0060)  
Sample #2 5.0850 (0.0520) 9.6840 (0.1130)  
Sample #3 5.0760 (0.0460) 9.6510 (0.1240)  
Sample #4 5.0580 (0.0580) 9.6520 (0.1200)  
Avg % Abs 5.0730 (0.0520) 9.6623 (0.1190)  
STD DEV 0.0137 (0.0060) 0.0188 (0.0056)  
REL STD DEV 0.271 (11.538) 0.194 (4.679)  
-----

SUA

BISCAYNE NAT PARK  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-001442 ✓  
 12/29/2025 12:49:48

Auto Calibration

pg 2 of 2

<<<<< 3um >>>>>			<<<<< 9um >>>>>		
Zero Order Coef	-314.48		Zero Order Coef	-186.99	
First Order Coef	2768.88		First Order Coef	1376.64	
Second Order Coef	21.28		Second Order Coef	12.53	
Act (g/210L)	Fit (g/210L)	Residual (g/210L)	Act (g/210L)	Fit (g/210L)	Residual (g/210L)
0.000	-0.000	0.0000	0.000	0.000	-0.0000
0.040	0.040	-0.0003	0.040	0.040	-0.0000
0.100	0.099	0.0005	0.100	0.100	0.0001
0.200	0.200	-0.0004	0.200	0.200	-0.0000
0.300	0.300	0.0001	0.300	0.300	0.0000

<<<<< 3um >>>>>		<<<<< 9um >>>>>	
Solution = 0.080 g/210L or 0.3810 mg/l, Samples = 4, Discarded = 1			
Sample #1	2968.00	3242.00	
Sample #2	2977.00	3303.00	
Sample #3	2944.00	3316.00	
Sample #4	2985.00	3328.00	
Avg	2968.6667	3315.6667	
STD DEV	21.7332	12.5033	
REL STD DEV	0.732	0.377	
H2O adjust (mg/l*10k)	841	494	

Barometric Pressure = 1010

\*\*\*\*\*CALIBRATION SUCCESSFUL\*\*\*\*\*

SUA

Post stability

BISCAYNE NAT PARK  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001442  
12/29/2025  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	14:40
Control Test	0.051	14:41
Air Blank	0.000	14:42
Control Test	0.050	14:42
Air Blank	0.000	14:43
Control Test	0.050	14:43
Air Blank	0.000	14:44
Control Test Stats		
Average	0.0503	
Std Dev	0.0006	
Rel Std Dev(%)	1.1471	

  
Operator's Signature

BISCAYNE NAT PARK  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001442  
12/29/2025  
Software: 8100.27

*post stability*

Test	g/210L	Time
-----	-----	-----
Air Blank	0.000	15:02
Control Test	0.080	15:03
Air Blank	0.000	15:03
Control Test	0.080	15:04
Air Blank	0.000	15:04
Control Test	0.080	15:05
Air Blank	0.000	15:06
Control Test Stats		
Average	0.0800	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

*Stiggell*

Operator's Signature

BISCAYNE NAT PARK  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001442  
12/29/2025  
Software: 8100.27

*post stability*


Test	g/210L	Time
Air Blank	0.000	15:15
Control Test	0.200	15:16
Air Blank	0.000	15:17
Control Test	0.200	15:17
Air Blank	0.000	15:18
Control Test	0.199	15:19
Air Blank	0.000	15:19
Control Test Stats		
Average	0.1997	
Std Dev	0.0006	
Rel Std Dev(%)	0.2892	

  
Operator's Signature

BISCAYNE NAT PARK  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001442  
12/29/2025  
Software: 8100.27

post stability  
DGS

Test	g/210L	Time
Air Blank	0.000	15:23
Control Test	0.079	15:23
Air Blank	0.000	15:23
Control Test	0.079	15:24
Air Blank	0.000	15:24
Control Test	0.079	15:25
Air Blank	0.000	15:25
Control Test Stats		
Average	0.0790	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

  
Operator's Signature

# Florida Department of Law Enforcement Alcohol Testing Program

## DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: BISCAYNE NAT PARK  
Time of Inspection: 14:20

Date of Inspection: 12/30/2025

Serial Number: 80-001442  
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	


Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202406K Exp: 06/19/2026	0.08g/210L Test (g/210L) Lot#:202406L Exp: 06/19/2026	0.20g/210L Test (g/210L) Lot#:202406N Exp: 06/20/2026	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG510701 Exp: 04/17/2027
0.000	0.050	0.081	0.199	0.079
0.000	0.051	0.082	0.199	0.079
0.000	0.051	0.082	0.199	0.079
0.000	0.051	0.082	0.199	0.079
0.000	0.051	0.082	0.198	0.079
0.000	0.052	0.082	0.198	0.079
0.000	0.051	0.082	0.198	0.078
0.000	0.052	0.082	0.197	0.079
0.000	0.052	0.082	0.198	0.079
0.000	0.051	0.082	0.197	0.079
Standard Deviations	0.0006	0.0003	0.0007	0.0003

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.


I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



LEANDRA HIGGINBOTHAM

Signature and Printed Name

12/30/2025  
Date



# Calibration Certificate

Florida Department of Law Enforcement  
Alcohol Testing Program  
2331 Phillips Road  
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001442, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001442</u>	UNCERTAINTY* ±	
Owning Agency:	<u>BISCAYNE NAT PARK</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>12/30/2025</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>14:20</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.  
Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.  
\*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).  
The instrument results before and after any adjustment are found in the associated pre and post stability checks

### TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.  
This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

  
Date 12/30/2025  
LEANDRA HIGGINBOTHAM,  
Department Inspector