

Alcohol Testing Program - Instrument Processing Sheet

Agency: JACKSONVILLE SO

Instrument Serial Number: 80-001045

Date In: 6/8/2026

DI Completion Date: 6/12/2026

☒ Ship

☐ P/U

☐ H/D

☐ CMI

☐ EE

Intake By: KTS Date: 6/8/2026		Quality Checks By: IAS Date: 6/11/2026		Flow Adjustment By: _____																																									
☒ Annual ☐ Dropped Off ☐ Registration ☒ Return from CMI / EE ☐ Training Instrument Visual Inspection ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/Accessories ☒ Power Cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes: CMI sent instrument to JSO instead of FDLE so JSO sent to FDLE for the required inspection. SLH 6/11/26		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value: 195 ☒ Flow Verification (L/s) Flow Column #: ATP102 32 mm .148 (.139-.169) 36 mm .167 (.156-.190) 53 mm .234 (.228-.278) 103 mm .492 (.447-.547) ☒ Barometric Pressure Check Gauge ID #: 28427 Gauge: 1011 Instrument: 1016 ☒ Stability Checks		Flow Column #: ☐ 5L/min – 17mm ☐ 15L/min – 53mm ☐ 30L/min – 103mm ☐ R-Value: _____ ☐ Post Adjustment Verification (L/S) Flow Column #: 32 mm (.139-.169) 36 mm (.156-.190) 53 mm (.228-.278) 103 mm (.447-.547)																																									
		<table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot#/Exp</th> </tr> </thead> <tbody> <tr> <td rowspan="2">0.050</td> <td rowspan="2">MP6291</td> <td>202603A</td> </tr> <tr> <td>3/3/2028</td> </tr> <tr> <td rowspan="2">0.080</td> <td rowspan="2">MP5089</td> <td>202603B</td> </tr> <tr> <td>3/3/2028</td> </tr> <tr> <td rowspan="2">0.200</td> <td rowspan="2">MP5090</td> <td>202603D</td> </tr> <tr> <td>3/4/2028</td> </tr> <tr> <td rowspan="2">0.080 DGS</td> <td rowspan="2">N/A</td> <td>23825080A1</td> </tr> <tr> <td>10/5/2027</td> </tr> </tbody> </table>		Simulator	Serial #	Lot#/Exp	0.050	MP6291	202603A	3/3/2028	0.080	MP5089	202603B	3/3/2028	0.200	MP5090	202603D	3/4/2028	0.080 DGS	N/A	23825080A1	10/5/2027	Maintenance By: _____ Date: _____ ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Tank Sensor Tare ☐ Breath Tube Replacement ☐ Other: _____																						
Simulator	Serial #	Lot#/Exp																																											
0.050	MP6291	202603A																																											
		3/3/2028																																											
0.080	MP5089	202603B																																											
		3/3/2028																																											
0.200	MP5090	202603D																																											
		3/4/2028																																											
0.080 DGS	N/A	23825080A1																																											
		10/5/2027																																											
Optical Bench Adjustment By: _____		Department Inspection By: SLH																																											
Barometric Pressure Gauge: _____ ID#: _____		Barometric Pressure ID#: 28427																																											
<table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>		Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Gauge: 1012 Instrument: 1017 Mouth Alcohol Solution Lot #: 2025-D Exp: 09/25/2027 Acetone Stock Solution Lot #: 2025-B Exp: 09/22/2027 <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>MP6294</td> </tr> <tr> <td>Interferent</td> <td>MP6290</td> </tr> <tr> <td>0.050</td> <td>MP6291</td> </tr> <tr> <td>0.080</td> <td>MP5089</td> </tr> <tr> <td>0.200</td> <td>MP5090</td> </tr> </tbody> </table>				Simulator	Serial Number	0.000	MP6294	Interferent	MP6290	0.050	MP6291	0.080	MP5089	0.200	MP5090
Simulator	Serial #	Lot #	Expiration																																										
0.000		N/A	N/A																																										
0.040																																													
0.100																																													
0.200																																													
0.300																																													
0.080 DGS	N/A																																												
Simulator	Serial Number																																												
0.000	MP6294																																												
Interferent	MP6290																																												
0.050	MP6291																																												
0.080	MP5089																																												
0.200	MP5090																																												
☐ Post Optical Bench Adjustment Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>		Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			Attachments ☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☐ Optical Bench Adjustment ☐ Post-Stability Checks ☐ Flow Adjustment ☒ Form 40 ☐ Other: _____																							
Simulator	Serial #	Lot #	Expiration																																										
0.050																																													
0.080																																													
0.200																																													
0.080 DGS	N/A																																												
Gauge ID #: _____ Gauge: _____ Instrument: _____																																													
Notes/Suggested Service: Instrument was able to read the pressure of the connected DGS. SLH 6/12/2026		☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use																																											
		Digitally signed by Wen-Chi Pierson Date: 2026.06.19 14:46:14 -04'00' Digitally signed by Kaitlyn Spearin Date: 2026.07.15 13:43:30 -04'00'																																											
		Tech Review Admin Review																																											

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: JACKSONVILLE SO
Time of Inspection: 10:52

Date of Inspection: 06/11/2026

Serial Number: 80-001045
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:

Number of Simulators Used: _____

Remarks:

BYPASS AI TO OPERATE INSTRUMENT. COMPLIANCE NOT DETERMINED.

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Isabella Smykowski ISABELLA A SMYKOWSKI
Signature and Printed Name

06/11/2026
Date

Stability Checks

0.050 g/210L	0.080 g/210L	0.200 g/210L	DGS 0.080 g/210L
0.047 to 0.053 g/210L	0.077 to 0.083 g/210L	0.194 to 0.206 g/210L	0.077 to 0.083 g/210L 50.003 g/210L of Wet
JACKSONVILLE SO Intoxilyzer - Alcohol Analyzer Model 8000 06/11/2026 Software: 8100.27 SN 80-001045 Test g/210L Time Air Blank 0.000 13:12 Control Test 0.050 13:12 Air Blank 0.000 13:13 Control Test 0.050 13:13 Air Blank 0.000 13:14 Control Test 0.050 13:14 Air Blank 0.000 13:15 Control Test 0.050 13:15 Control Test Stats Average 0.0500 Std Dev 0.0000 Rel Std Dev(%) 0.0000 <i>Rabeella Sypharishi</i> Operator's Signature	JACKSONVILLE SO Intoxilyzer - Alcohol Analyzer Model 8000 06/11/2026 Software: 8100.27 SN 80-001045 Test g/210L Time Air Blank 0.000 14:12 Control Test 0.078 14:13 Air Blank 0.000 14:14 Control Test 0.079 14:14 Air Blank 0.000 14:15 Control Test 0.079 14:16 Air Blank 0.000 14:16 Control Test Stats Average 0.0787 Std Dev 0.0006 Rel Std Dev(%) 0.7339 <i>Rabeella Sypharishi</i> Operator's Signature	JACKSONVILLE SO Intoxilyzer - Alcohol Analyzer Model 8000 06/11/2026 Software: 8100.27 SN 80-001045 Test g/210L Time Air Blank 0.000 14:16 Control Test 0.200 14:16 Air Blank 0.000 14:17 Control Test 0.202 14:18 Air Blank 0.000 14:19 Control Test 0.201 14:19 Air Blank 0.000 14:19 Control Test Stats Average 0.2010 Std Dev 0.0010 Rel Std Dev(%) 0.4975 <i>Rabeella Sypharishi</i> Operator's Signature	JACKSONVILLE SO Intoxilyzer - Alcohol Analyzer Model 8000 06/11/2026 Software: 8100.27 SN 80-001045 Test g/210L Time Air Blank 0.000 11:51 Control Test 0.081 11:51 Air Blank 0.000 11:52 Control Test 0.080 11:52 Air Blank 0.000 11:53 Control Test 0.079 11:53 Air Blank 0.000 11:54 Control Test Stats Average 0.0810 Std Dev 0.0010 Rel Std Dev(%) 1.2510 <i>Rabeella Sypharishi</i> Operator's Signature

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: JACKSONVILLE SO
Time of Inspection: 15:57

Date of Inspection: 06/12/2026

Serial Number: 80-001045
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202603A Exp: 03/03/2028	0.08g/210L Test (g/210L) Lot#:202603B Exp: 03/03/2028	0.20g/210L Test (g/210L) Lot#:202603D Exp: 03/04/2028	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:23825080A1 Exp: 10/05/2027
0.000	0.049	0.079	0.198	0.081
0.000	0.049	0.076	0.198	0.081
0.000	0.049	0.076	0.198	0.081
0.000	0.049	0.076	0.198	0.080
0.000	0.049	0.076	0.198	0.081
0.000	0.049	0.076	0.198	0.081
0.000	0.049	0.075	0.198	0.081
0.000	0.049	0.076	0.198	0.081
0.000	0.050	0.076	0.198	0.081
0.000	0.050	0.076	0.198	0.082

Standard Deviations	0.0004	0.0010	0.0000	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

REPEATED MIN SAMPLE VOL DUE TO DELAY PROVIDING SAMPLE



The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

LEANDRA HIGGINBOTHAM

Signature and Printed Name

06/12/2026
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001045, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001045</u>	UNCERTAINTY * $\pm$
Owning Agency:	<u>JACKSONVILLE SO</u>	0.050 g/210 L 0.004
Calibration Date:	<u>06/12/2026</u>	0.080 g/210 L 0.004
Calibration Time:	<u>15:57</u>	0.200 g/210 L 0.008
		0.080 g/210 L Dry Gas Control 0.005

All results are reported in g/210 L.
Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).
The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Date

06/12/2026

LEANDRA HIGINBOTHAM,

Department Inspector

FDLE/ATP Form 69 January 2026
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality