

## Instrument Processing Sheet

Agency: Tyndall AFB Instrument Serial Number: 80-001230  
 Date In: 10/27/2025 DI Completion Date: 10/30/2025  Ship  P/U  H/D  CMI  EE

<b>Intake</b> By: <u>SLH</u> Date: <u>10/27/25</u> <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE <input type="checkbox"/> Return unworked <input type="checkbox"/> Training Visual Inspection <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/Accessories <input type="checkbox"/> Power Cord <input type="checkbox"/> Printer Cable <input type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: Dropped off by agency.	<b>Quality Checks</b> By: <u>SLH</u> Date: <u>10/27/25</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value: <u>237</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column #: <u>ATP 102</u> 32 mm <u>0.144</u> (.139-.169) 36 mm <u>0.160</u> (.156-.190) 53 mm <u>0.230</u> (.228-.278) 103 mm <u>0.488</u> (.447-.547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID #: <u>28662</u> Gauge: <u>1006</u> Instrument: <u>1005</u> <input checked="" type="checkbox"/> Stability Checks	<b>Flow Adjustment</b> By: _____ Date: _____ Flow Column #: _____ <input type="checkbox"/> 5L/min – 17mm <input type="checkbox"/> 15L/min – 53mm <input type="checkbox"/> 30L/min – 103mm <input type="checkbox"/> R-Value: _____ <input type="checkbox"/> Post Adjustment Verification (L/S) Flow Column #: _____ 32 mm _____ (.139-.169) 36 mm _____ (.156-.190) 53 mm _____ (.228-.278) 103 mm _____ (.447-.547)
--	---	--

Simulator	Serial #	Lot#/Exp	Maintenance	By:	Date:
0.050	MP5088	202406K 6/19/2026	<input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement and Tank Sensor Tare <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other:		
0.080	MP5089/5089	202406L 6/19/2026			
0.200	MP5090	202406N 6/20/2026			
0.080 DGS	N/A	AG510701 4/17/2027			

<b>Optical Bench Adjustment</b> By: _____ Barometric Pressure Gauge: _____ ID#: _____ <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr> <tr><td>0.040</td><td></td><td></td><td></td></tr> <tr><td>0.100</td><td></td><td></td><td></td></tr> <tr><td>0.200</td><td></td><td></td><td></td></tr> <tr><td>0.300</td><td></td><td></td><td></td></tr> <tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr> </tbody> </table> <input type="checkbox"/> Post Optical Bench Adjustment Stability Checks <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr><td>0.050</td><td></td><td></td><td></td></tr> <tr><td>0.080</td><td></td><td></td><td></td></tr> <tr><td>0.200</td><td></td><td></td><td></td></tr> <tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr> </tbody> </table> Barometric Pressure Gauge: _____ ID#: _____	Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			<b>Department Inspection</b> By: <u>SLH</u> Barometric Pressure ID#: <u>28421 / 28421</u> Gauge: <u>1011 / 1010</u> Instrument: <u>1009 / 1009</u> Mouth Alcohol Solution Lot #: <u>2025-C</u> Exp: <u>9/25/2027</u> Acetone Stock Solution Lot #: <u>2025-B</u> Exp: <u>9/22/2027</u> <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr><td>0.000</td><td>MP5086 / 6294</td></tr> <tr><td>Interferent</td><td>MP5087 / 6290</td></tr> <tr><td>0.050</td><td>MP6291 / 5088</td></tr> <tr><td>0.080</td><td>MP6292 / 5089</td></tr> <tr><td>0.200</td><td>MP5090 / 5090</td></tr> </tbody> </table> <b>Attachments</b> <input checked="" type="checkbox"/> Form 41 (x2) <input type="checkbox"/> Post-Stability Checks <input checked="" type="checkbox"/> Stability Checks (x2) <input type="checkbox"/> Flow Adjustment <input checked="" type="checkbox"/> Calibration Certificate <input type="checkbox"/> Form 40 <input type="checkbox"/> Optical Bench Adjustment <input type="checkbox"/> Other:	Simulator	Serial Number	0.000	MP5086 / 6294	Interferent	MP5087 / 6290	0.050	MP6291 / 5088	0.080	MP6292 / 5089	0.200	MP5090 / 5090
Simulator	Serial #	Lot #	Expiration																																																										
0.000		N/A	N/A																																																										
0.040																																																													
0.100																																																													
0.200																																																													
0.300																																																													
0.080 DGS	N/A																																																												
Simulator	Serial #	Lot #	Expiration																																																										
0.050																																																													
0.080																																																													
0.200																																																													
0.080 DGS	N/A																																																												
Simulator	Serial Number																																																												
0.000	MP5086 / 6294																																																												
Interferent	MP5087 / 6290																																																												
0.050	MP6291 / 5088																																																												
0.080	MP6292 / 5089																																																												
0.200	MP5090 / 5090																																																												

Notes/Suggested Service: Inadvertently did not complete the 0.050 and 0.200 g/210L stability checks prior to the department inspection (DI) on 10/28/25. Repeated on 10/30/25 all stabilities and the DI. The lot/exp info for the stabilities same and SIM S/N are as indicated. For the DI, the mouth alcohol and acetone solutions are the same information for both DI's. SLH 10/30/25	<input checked="" type="checkbox"/> <b>Instrument Complies with Chapter 11D-8, FAC</b> <input type="checkbox"/> <b>Instrument Does Not Comply with Chapter 11D-8, FAC</b> <hr/> <input checked="" type="checkbox"/> <b>Return to/Place into Evidentiary Use</b> <input type="checkbox"/> <b>Remain Out of Evidentiary Use</b> <hr/> <input checked="" type="checkbox"/> <b>Conduct an Agency Inspection Before Evidentiary Use</b>
Taylor Gutschow <small>Digitally signed by Taylor Gutschow Date: 2025.11.03 12:20:23 -05'00'</small>	Shayla Platt <small>Digitally signed by Shayla Platt Date: 2025.11.14 09:55:12 -05'00'</small> <b>Tech Review</b> <span style="float: right;"><b>Admin Review</b></span>

# Stability Checks

80-001230  
Sub 10/27/2025

0.050 g/210L 0.047 to 0.053 g/210L	0.080 g/210L 0.077 to 0.083 g/210L	0.200 g/210L 0.194 to 0.206 g/210L	DGS 0.080 g/210L 0.077 to 0.083 g/210L 50.003 g/210L of Wet																																																																																																																																																
<p>Performed Root Cause Analysis <input type="checkbox"/></p> <p>TYNDALL AFB Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001230 10/27/2025 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Fir Blank</td><td>0.000</td><td>13:24</td></tr> <tr><td>Control Test</td><td>0.078</td><td>13:24</td></tr> <tr><td>Fir Blank</td><td>0.000</td><td>13:25</td></tr> <tr><td>Control Test</td><td>0.078</td><td>13:26</td></tr> <tr><td>Fir Blank</td><td>0.000</td><td>13:26</td></tr> <tr><td>Control Test</td><td>0.078</td><td>13:27</td></tr> <tr><td>Fir Blank</td><td>0.000</td><td>13:27</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0780</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> <p>Operator Signature <i>M. Myllett</i></p>	Test	g/210L	Time	Fir Blank	0.000	13:24	Control Test	0.078	13:24	Fir Blank	0.000	13:25	Control Test	0.078	13:26	Fir Blank	0.000	13:26	Control Test	0.078	13:27	Fir Blank	0.000	13:27	Control Test Stats			Average	0.0780		Std Dev	0.0000		Rel Std Dev(%)	0.0000		<p>Performed Root Cause Analysis <input checked="" type="checkbox"/></p> <p>TYNDALL AFB Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001230 10/27/2025 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Fir Blank</td><td>0.000</td><td>13:30</td></tr> <tr><td>Control Test</td><td>0.080</td><td>13:30</td></tr> <tr><td>Fir Blank</td><td>0.000</td><td>13:31</td></tr> <tr><td>Control Test</td><td>0.080</td><td>13:31</td></tr> <tr><td>Fir Blank</td><td>0.000</td><td>13:32</td></tr> <tr><td>Control Test</td><td>0.080</td><td>13:32</td></tr> <tr><td>Fir Blank</td><td>0.000</td><td>13:32</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0800</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> <p>Operator Signature <i>M. Myllett</i></p>	Test	g/210L	Time	Fir Blank	0.000	13:30	Control Test	0.080	13:30	Fir Blank	0.000	13:31	Control Test	0.080	13:31	Fir Blank	0.000	13:32	Control Test	0.080	13:32	Fir Blank	0.000	13:32	Control Test Stats			Average	0.0800		Std Dev	0.0000		Rel Std Dev(%)	0.0000		<p>Performed Root Cause Analysis <input checked="" type="checkbox"/></p> <p>TYNDALL AFB Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001230 10/27/2025 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Fir Blank</td><td>0.000</td><td>13:30</td></tr> <tr><td>Control Test</td><td>0.080</td><td>13:30</td></tr> <tr><td>Fir Blank</td><td>0.000</td><td>13:31</td></tr> <tr><td>Control Test</td><td>0.080</td><td>13:31</td></tr> <tr><td>Fir Blank</td><td>0.000</td><td>13:32</td></tr> <tr><td>Control Test</td><td>0.080</td><td>13:32</td></tr> <tr><td>Fir Blank</td><td>0.000</td><td>13:32</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0800</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> <p>Operator Signature <i>M. Myllett</i></p>	Test	g/210L	Time	Fir Blank	0.000	13:30	Control Test	0.080	13:30	Fir Blank	0.000	13:31	Control Test	0.080	13:31	Fir Blank	0.000	13:32	Control Test	0.080	13:32	Fir Blank	0.000	13:32	Control Test Stats			Average	0.0800		Std Dev	0.0000		Rel Std Dev(%)	0.0000		<p>Performed Root Cause Analysis <input checked="" type="checkbox"/></p> <p>TYNDALL AFB Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001230 10/27/2025 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Fir Blank</td><td>0.000</td><td>13:30</td></tr> <tr><td>Control Test</td><td>0.080</td><td>13:30</td></tr> <tr><td>Fir Blank</td><td>0.000</td><td>13:31</td></tr> <tr><td>Control Test</td><td>0.080</td><td>13:31</td></tr> <tr><td>Fir Blank</td><td>0.000</td><td>13:32</td></tr> <tr><td>Control Test</td><td>0.080</td><td>13:32</td></tr> <tr><td>Fir Blank</td><td>0.000</td><td>13:32</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0800</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> <p>Operator Signature <i>M. Myllett</i></p>	Test	g/210L	Time	Fir Blank	0.000	13:30	Control Test	0.080	13:30	Fir Blank	0.000	13:31	Control Test	0.080	13:31	Fir Blank	0.000	13:32	Control Test	0.080	13:32	Fir Blank	0.000	13:32	Control Test Stats			Average	0.0800		Std Dev	0.0000		Rel Std Dev(%)	0.0000	
Test	g/210L	Time																																																																																																																																																	
Fir Blank	0.000	13:24																																																																																																																																																	
Control Test	0.078	13:24																																																																																																																																																	
Fir Blank	0.000	13:25																																																																																																																																																	
Control Test	0.078	13:26																																																																																																																																																	
Fir Blank	0.000	13:26																																																																																																																																																	
Control Test	0.078	13:27																																																																																																																																																	
Fir Blank	0.000	13:27																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0780																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Fir Blank	0.000	13:30																																																																																																																																																	
Control Test	0.080	13:30																																																																																																																																																	
Fir Blank	0.000	13:31																																																																																																																																																	
Control Test	0.080	13:31																																																																																																																																																	
Fir Blank	0.000	13:32																																																																																																																																																	
Control Test	0.080	13:32																																																																																																																																																	
Fir Blank	0.000	13:32																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0800																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Fir Blank	0.000	13:30																																																																																																																																																	
Control Test	0.080	13:30																																																																																																																																																	
Fir Blank	0.000	13:31																																																																																																																																																	
Control Test	0.080	13:31																																																																																																																																																	
Fir Blank	0.000	13:32																																																																																																																																																	
Control Test	0.080	13:32																																																																																																																																																	
Fir Blank	0.000	13:32																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0800																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Fir Blank	0.000	13:30																																																																																																																																																	
Control Test	0.080	13:30																																																																																																																																																	
Fir Blank	0.000	13:31																																																																																																																																																	
Control Test	0.080	13:31																																																																																																																																																	
Fir Blank	0.000	13:32																																																																																																																																																	
Control Test	0.080	13:32																																																																																																																																																	
Fir Blank	0.000	13:32																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0800																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
<p>* Sub 10/27/2025 e-sub</p>																																																																																																																																																			

\* Inadvertently did not complete stabilities prior to department inspection

# Stability Checks #2

80-001230  
10/30/2025 SMT

0.050 g/210L 0.047 to 0.053 g/210L	0.080 g/210L 0.077 to 0.083 g/210L	0.200 g/210L 0.194 to 0.206 g/210L	DGS 0.080 g/210L 0.077 to 0.083 g/210L ±0.003 g/210L of Wet																																																																																																																																																
<p>TYNDALL AFB Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001230 10/30/2025 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.100</td><td>08:49</td></tr> <tr><td>Control Test</td><td>0.148</td><td>08:49</td></tr> <tr><td>Air Blank</td><td>0.100</td><td>08:50</td></tr> <tr><td>Control Test</td><td>0.148</td><td>08:51</td></tr> <tr><td>Air Blank</td><td>0.100</td><td>08:51</td></tr> <tr><td>Control Test</td><td>0.143</td><td>08:52</td></tr> <tr><td>Air Blank</td><td>0.100</td><td>08:52</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1480</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.100	08:49	Control Test	0.148	08:49	Air Blank	0.100	08:50	Control Test	0.148	08:51	Air Blank	0.100	08:51	Control Test	0.143	08:52	Air Blank	0.100	08:52	Control Test Stats			Average	0.1480		Std Dev	0.0000		Rel Std Dev(%)	0.0000		<p>TYNDALL AFB Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001230 10/30/2025 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>08:43</td></tr> <tr><td>Control Test</td><td>0.078</td><td>08:44</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>08:44</td></tr> <tr><td>Control Test</td><td>0.079</td><td>08:45</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>08:45</td></tr> <tr><td>Control Test</td><td>0.079</td><td>08:46</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>08:47</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0787</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7339</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	08:43	Control Test	0.078	08:44	Air Blank	0.000	08:44	Control Test	0.079	08:45	Air Blank	0.000	08:45	Control Test	0.079	08:46	Air Blank	0.000	08:47	Control Test Stats			Average	0.0787		Std Dev	0.0006		Rel Std Dev(%)	0.7339		<p>TYNDALL AFB Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001230 10/30/2025 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>09:04</td></tr> <tr><td>Control Test</td><td>0.196</td><td>09:05</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:16</td></tr> <tr><td>Control Test</td><td>0.196</td><td>09:16</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:17</td></tr> <tr><td>Control Test</td><td>0.195</td><td>09:18</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:18</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1957</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.2951</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	09:04	Control Test	0.196	09:05	Air Blank	0.000	09:16	Control Test	0.196	09:16	Air Blank	0.000	09:17	Control Test	0.195	09:18	Air Blank	0.000	09:18	Control Test Stats			Average	0.1957		Std Dev	0.0006		Rel Std Dev(%)	0.2951		<p>TYNDALL AFB Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001230 10/30/2025 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>09:23</td></tr> <tr><td>Control Test</td><td>0.081</td><td>09:23</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:24</td></tr> <tr><td>Control Test</td><td>0.080</td><td>09:24</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:24</td></tr> <tr><td>Control Test</td><td>0.080</td><td>09:25</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:25</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0803</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7187</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	09:23	Control Test	0.081	09:23	Air Blank	0.000	09:24	Control Test	0.080	09:24	Air Blank	0.000	09:24	Control Test	0.080	09:25	Air Blank	0.000	09:25	Control Test Stats			Average	0.0803		Std Dev	0.0006		Rel Std Dev(%)	0.7187	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.100	08:49																																																																																																																																																	
Control Test	0.148	08:49																																																																																																																																																	
Air Blank	0.100	08:50																																																																																																																																																	
Control Test	0.148	08:51																																																																																																																																																	
Air Blank	0.100	08:51																																																																																																																																																	
Control Test	0.143	08:52																																																																																																																																																	
Air Blank	0.100	08:52																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1480																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	08:43																																																																																																																																																	
Control Test	0.078	08:44																																																																																																																																																	
Air Blank	0.000	08:44																																																																																																																																																	
Control Test	0.079	08:45																																																																																																																																																	
Air Blank	0.000	08:45																																																																																																																																																	
Control Test	0.079	08:46																																																																																																																																																	
Air Blank	0.000	08:47																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0787																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7339																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	09:04																																																																																																																																																	
Control Test	0.196	09:05																																																																																																																																																	
Air Blank	0.000	09:16																																																																																																																																																	
Control Test	0.196	09:16																																																																																																																																																	
Air Blank	0.000	09:17																																																																																																																																																	
Control Test	0.195	09:18																																																																																																																																																	
Air Blank	0.000	09:18																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1957																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.2951																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	09:23																																																																																																																																																	
Control Test	0.081	09:23																																																																																																																																																	
Air Blank	0.000	09:24																																																																																																																																																	
Control Test	0.080	09:24																																																																																																																																																	
Air Blank	0.000	09:24																																																																																																																																																	
Control Test	0.080	09:25																																																																																																																																																	
Air Blank	0.000	09:25																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0803																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7187																																																																																																																																																		
<p>Operator Signature: </p>	<p>Operator Signature: </p>	<p>Operator Signature: </p>	<p>Operator Signature: </p>																																																																																																																																																

# Florida Department of Law Enforcement Alcohol Testing Program

## DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: TYNDALL AFB  
Time of Inspection: 12:36

Date of Inspection: 10/28/2025

Serial Number: 80-001230  
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202406K Exp: 06/19/2026	0.08g/210L Test (g/210L) Lot#:202406L Exp: 06/19/2026	0.20g/210L Test (g/210L) Lot#:202406N Exp: 06/20/2026	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG510701 Exp: 04/17/2027
0.000	0.047	0.078	0.194	0.079
0.000	0.047	0.078	0.194	0.080
0.000	0.047	0.077	0.194	0.079
0.000	0.047	0.078	0.195	0.079
0.000	0.047	0.077	0.195	0.080
0.000	0.047	0.077	0.195	0.080
0.000	0.047	0.078	0.195	0.080
0.000	0.048	0.078	0.195	0.080
0.000	0.048	0.078	0.196	0.080
0.000	0.048	0.078	0.195	0.080


Standard Deviations	0.0004	0.0004	0.0006	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



LEANDRA HIGGINBOTHAM

Signature and Printed Name

10/28/2025  
Date

# Florida Department of Law Enforcement Alcohol Testing Program

## DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: TYNDALL AFB  
Time of Inspection: 13:41

Date of Inspection: 10/30/2025

Serial Number: 80-001230  
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202406K Exp: 06/19/2026	0.08g/210L Test (g/210L) Lot#:202406L Exp: 06/19/2026	0.20g/210L Test (g/210L) Lot#:202406N Exp: 06/20/2026	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG510701 Exp: 04/17/2027
0.000	0.049	0.078	0.194	0.079
0.000	0.049	0.078	0.195	0.080
0.000	0.048	0.078	0.196	0.080
0.000	0.048	0.078	0.195	0.080
0.000	0.048	0.078	0.195	0.080
0.000	0.048	0.078	0.195	0.080
0.000	0.048	0.078	0.195	0.080
0.000	0.048	0.077	0.194	0.080
0.000	0.048	0.077	0.195	0.080
0.000	0.048	0.077	0.195	0.080


Standard Deviations	0.0004	0.0004	0.0005	0.0003
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5



Remarks:

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

  
 LEANDRA HIGGINBOTHAM  
 Signature and Printed Name

10/30/2025  
 Date



# Calibration Certificate

Florida Department of Law Enforcement  
Alcohol Testing Program  
2331 Phillips Road  
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001230, manufactured by CMI, Inc. was calibrated in accordance with FDLE/AIP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001230</u>	UNCERTAINTY* ±	
Owning Agency:	<u>TYNDALL AFB</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>10/30/2025</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>13:41</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

\*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

### TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

  
Date 10/30/2025  
LEANDRA HIGGINBOTHAM,  
Department Inspector

FDLE/AIP Form 69 October 2024  
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality