

Alcohol Testing Program - Instrument Processing Sheet

Agency: PASCO COUNTY SO Instrument Serial Number: 80-001074
 Date In: 1/8/2026 DI Completion Date: 1/15/26 Ship P/U H/D CMI EE

Intake By: <u>KTS</u> Date: <u>1/8/2026</u> <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Dropped Off <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE <input type="checkbox"/> Training Instrument Visual Inspection <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/Accessories <input type="checkbox"/> Power Cord <input type="checkbox"/> Printer Cable <input type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes:	Quality Checks By: <u>KTS</u> Date: <u>1/13/26</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value: <u>226</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column #: <u>ATP103</u> 32 mm <u>0.136</u> (.139-.169) 36 mm <u>0.148</u> (.156-.190) 53 mm <u>0.222</u> (.228-.278) 103 mm <u>0.476</u> (.447-.547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID #: <u>28427</u> Gauge: <u>1020</u> Instrument: <u>1019</u> <input checked="" type="checkbox"/> Stability Checks <table border="1" style="width:100%; border-collapse: collapse; margin-top: 5px;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot#/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>MP5088</td> <td>202406K 6/19/26</td> </tr> <tr> <td>0.080</td> <td>MP5089</td> <td>202406L 6/19/26</td> </tr> <tr> <td>0.200</td> <td>MP5090</td> <td>202406N 6/20/26</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG510701 4/17/27</td> </tr> </tbody> </table>	Simulator	Serial #	Lot#/Exp	0.050	MP5088	202406K 6/19/26	0.080	MP5089	202406L 6/19/26	0.200	MP5090	202406N 6/20/26	0.080 DGS	N/A	AG510701 4/17/27	Flow Adjustment By: <u>KTS</u> Flow Column #: <u>ATP102</u> <input checked="" type="checkbox"/> 5L/min – 17mm <input checked="" type="checkbox"/> 15L/min – 53mm <input checked="" type="checkbox"/> 30L/min – 103mm <input checked="" type="checkbox"/> R-Value: <u>229</u> <input checked="" type="checkbox"/> Post Adjustment Verification (L/S) Flow Column #: <u>ATP103</u> 32 mm <u>0.148</u> (.139-.169) 36 mm <u>0.167</u> (.156-.190) 53 mm <u>0.242</u> (.228-.278) 103 mm <u>0.507</u> (.447-.547)																									
Simulator	Serial #	Lot#/Exp																																								
0.050	MP5088	202406K 6/19/26																																								
0.080	MP5089	202406L 6/19/26																																								
0.200	MP5090	202406N 6/20/26																																								
0.080 DGS	N/A	AG510701 4/17/27																																								
<table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th colspan="2">Maintenance</th> <th>By:</th> <th>Date:</th> </tr> </thead> <tbody> <tr> <td><input type="checkbox"/> Battery Replacement</td> <td><input type="checkbox"/> Dry Gas Regulator Replacement</td> <td><u>KTS</u></td> <td></td> </tr> <tr> <td><input type="checkbox"/> Tank Sensor Tare</td> <td><input type="checkbox"/> Breath Tube Replacement</td> <td></td> <td></td> </tr> <tr> <td colspan="2"><input type="checkbox"/> Other:</td> <td></td> <td></td> </tr> </tbody> </table>			Maintenance		By:	Date:	<input type="checkbox"/> Battery Replacement	<input type="checkbox"/> Dry Gas Regulator Replacement	<u>KTS</u>		<input type="checkbox"/> Tank Sensor Tare	<input type="checkbox"/> Breath Tube Replacement			<input type="checkbox"/> Other:																											
Maintenance		By:	Date:																																							
<input type="checkbox"/> Battery Replacement	<input type="checkbox"/> Dry Gas Regulator Replacement	<u>KTS</u>																																								
<input type="checkbox"/> Tank Sensor Tare	<input type="checkbox"/> Breath Tube Replacement																																									
<input type="checkbox"/> Other:																																										
Optical Bench Adjustment By: _____ Barometric Pressure Gauge: _____ ID#: _____		Department Inspection By: <u>KTS</u> Barometric Pressure ID#: <u>28427</u> Gauge: <u>1011</u> Instrument: <u>1009</u> Mouth Alcohol Solution Lot #: <u>2025-D</u> Exp: <u>9/25/27</u> Acetone Stock Solution Lot #: <u>2025-B</u> Exp: <u>9/22/27</u>																																								
<table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>		Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			<table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>MP6289</td> </tr> <tr> <td>Interferent</td> <td>MP6290</td> </tr> <tr> <td>0.050</td> <td>MP5088</td> </tr> <tr> <td>0.080</td> <td>MP5089</td> </tr> <tr> <td>0.200</td> <td>MP5090</td> </tr> </tbody> </table>	Simulator	Serial Number	0.000	MP6289	Interferent	MP6290	0.050	MP5088	0.080	MP5089	0.200	MP5090
Simulator	Serial #	Lot #	Expiration																																							
0.000		N/A	N/A																																							
0.040																																										
0.100																																										
0.200																																										
0.300																																										
0.080 DGS	N/A																																									
Simulator	Serial Number																																									
0.000	MP6289																																									
Interferent	MP6290																																									
0.050	MP5088																																									
0.080	MP5089																																									
0.200	MP5090																																									
<input type="checkbox"/> Post Optical Bench Adjustment Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> Gauge ID #: _____ Gauge: _____ Instrument: _____		Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			Attachments <input checked="" type="checkbox"/> Form 41 <input type="checkbox"/> Post-Stability Checks <input checked="" type="checkbox"/> Stability Checks <input checked="" type="checkbox"/> Flow Adjustment <input checked="" type="checkbox"/> Calibration Certificate <input type="checkbox"/> Form 40 <input type="checkbox"/> Optical Bench Adjustment <input type="checkbox"/> Other:																				
Simulator	Serial #	Lot #	Expiration																																							
0.050																																										
0.080																																										
0.200																																										
0.080 DGS	N/A																																									
Notes/Suggested Service:		<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use																																								
		Digitally signed by Taylor Gutschow Date: 2026.01.23 11:27:46 -05'00' Digitally signed by Wen-Chi Pierson Date: 2026.01.26 15:14:24 -05'00'																																								
		Tech Review Admin Review																																								

PASCO COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001074
01/13/2026
Software: 8100.27

Flow Rate Calibration*****

1: Rate (Liters/min) = 5
SQRT(Diff)) = 6.926
2: Rate (Liters/min) = 15
SQRT(Diff)) = 11.660
3: Rate (Liters/min) = 30
SQRT(Diff)) = 21.117
Dependent Data Scale Factor = 100000 L/min
Independent Data Scale Factor = 256
Rounded Slope = 678
Rounded Intercept = -631621
Correlation = 0.99720

Kats
1/13/26 ~~12/~~ KTS

PASCO COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001074
01/13/2026
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:39
Control Test	0.079	10:39
Air Blank	0.000	10:40
Control Test	0.080	10:40
Air Blank	0.000	10:41
Control Test	0.080	10:41
Air Blank	0.000	10:42
Control Test Stats		
Average	0.0797	
Std Dev	0.0006	
Rel Std Dev(%)	0.7247	



Operator's Signature

Dry

PASCO COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001074
01/13/2026
Software: 8100.27


Test	g/210L	Time
Air Blank	0.000	10:24
Control Test	0.049	10:25
Air Blank	0.000	10:26
Control Test	0.049	10:26
Air Blank	0.000	10:27
Control Test	0.049	10:27
Air Blank	0.000	10:28
Control Test Stats		
Average	0.0490	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	



Operator's Signature

PASCO COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001074
01/13/2026
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:29
Control Test	0.079	10:30
Air Blank	0.000	10:31
Control Test	0.079	10:31
Air Blank	0.000	10:32
Control Test	0.079	10:33
Air Blank	0.000	10:33
Control Test Stats		
Average	0.0790	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	



Operator's Signature

PASCO COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001074
01/13/2026
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:34
Control Test	0.198	10:35
Air Blank	0.000	10:35
Control Test	0.197	10:36
Air Blank	0.000	10:37
Control Test	0.198	10:37
Air Blank	0.000	10:38
Control Test Stats		
Average	0.1977	
Std Dev	0.0006	
Rel Std Dev(%)	0.2921	



Operator's Signature

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: PASCO COUNTY SO

Time of Inspection: 14:45

Date of Inspection: 01/15/2026

Serial Number: 80-001074

Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202406K Exp: 06/19/2026	0.08g/210L Test (g/210L) Lot#:202406L Exp: 06/19/2026	0.20g/210L Test (g/210L) Lot#:202406N Exp: 06/20/2026	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG510701 Exp: 04/17/2027
0.000	0.049	0.078	0.201	0.080
0.000	0.050	0.079	0.202	0.080
0.000	0.049	0.079	0.202	0.080
0.000	0.050	0.079	0.202	0.080
0.000	0.049	0.079	0.202	0.080
0.000	0.050	0.079	0.203	0.080
0.000	0.050	0.079	0.202	0.080
0.000	0.050	0.079	0.202	0.080
0.000	0.050	0.079	0.202	0.080
0.000	0.050	0.079	0.202	0.080
0.000	0.050	0.079	0.202	0.080

Standard Deviations	0.0004	0.0003	0.0004	0.0000
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0002 Number of Simulators Used: 5


Remarks:

A F / M A:.

Alcohol free test detected a measurement of 0.008 due to residual mouth alcohol from previous inspection. Repeated AF/MA test. KTS 1/16/26


The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



KATIE T SPEARIN
Signature and Printed Name

01/15/2026
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001074, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001074</u>	UNCERTAINTY* ±	
Owning Agency:	<u>PASCO COUNTY SO</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>01/15/2026</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>14:45</u>	0.200 g/ 210 L	0.008
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

01/15/2026

Date


KATIE T SPEARIN,

Department Inspector

FDLE/ATP Form 69 January 2026

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality