

Alcohol Testing Program - Instrument Processing Sheet

Agency: FFWCC

Instrument Serial Number: 80-000900

Date In: 8/4/2026

DI Completion Date: 08/13/2026

☐ Ship ☒ P/U ☐ H/D ☐ CMI ☐ EE

Intake By: <u>SLH</u> Date: <u>8/4/2026</u>		Quality Checks By: <u>SLH</u> Date: <u>8/4/2026</u>		Flow Adjustment By: _____																													
☒ Annual ☒ Dropped Off ☐ Registration ☐ Return from CMI / EE ☐ Training Instrument Visual Inspection ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/Accessories ☒ Power Cord ☐ Printer Cable ☐ Static Bag ☐ 12V DC Cable Notes:		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value: <u>215</u> ☒ Flow Verification (L/s) Flow Column #: <u>ATP 105</u> 32 mm <u>0.140</u> (.139-.169) 36 mm <u>0.156</u> (.156-.190) 53 mm <u>0.230</u> (.228-.278) 103 mm <u>0.503</u> (.447-.547) ☒ Barometric Pressure Check Gauge ID #: <u>28427</u> Gauge: <u>1009</u> Instrument: <u>1010</u> ☒ Stability Checks		Flow Column #: _____ ☐ 5L/min – 17mm ☐ 15L/min – 53mm ☐ 30L/min – 103mm ☐ R-Value: _____ ☐ Post Adjustment Verification (L/S) Flow Column #: _____ 32 mm _____ (.139-.169) 36 mm _____ (.156-.190) 53 mm _____ (.228-.278) 103 mm _____ (.447-.547)																													
		<table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot#/Exp</th> </tr> </thead> <tbody> <tr> <td rowspan="2">0.050</td> <td rowspan="2">MP5088</td> <td>202603A</td> </tr> <tr> <td>03/03/2028</td> </tr> <tr> <td rowspan="2">0.080</td> <td rowspan="2">MP5089</td> <td>202603B</td> </tr> <tr> <td>03/03/2028</td> </tr> <tr> <td rowspan="2">0.200</td> <td rowspan="2">MP5090</td> <td>202603D</td> </tr> <tr> <td>03/04/2028</td> </tr> <tr> <td rowspan="2">0.080 DGS</td> <td rowspan="2">N/A</td> <td>23825080A1</td> </tr> <tr> <td>10/05/2027</td> </tr> </tbody> </table>		Simulator	Serial #	Lot#/Exp	0.050	MP5088	202603A	03/03/2028	0.080	MP5089	202603B	03/03/2028	0.200	MP5090	202603D	03/04/2028	0.080 DGS	N/A	23825080A1	10/05/2027	Maintenance By: <u>SLH</u> Date: <u>8/4/2026</u> ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☒ Tank Sensor Tare ☐ Breath Tube Replacement ☐ Other:										
Simulator	Serial #	Lot#/Exp																															
0.050	MP5088	202603A																															
		03/03/2028																															
0.080	MP5089	202603B																															
		03/03/2028																															
0.200	MP5090	202603D																															
		03/04/2028																															
0.080 DGS	N/A	23825080A1																															
		10/05/2027																															
Optical Bench Adjustment By: <u>IAS</u> Barometric Pressure Gauge: <u>1018</u> ID#: <u>34420</u>		Department Inspection By: <u>SLH</u> Barometric Pressure ID#: <u>28427</u>																															
<table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>MP6294</td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td>MP6295</td> <td>25090</td> <td>3/11/2027</td> </tr> <tr> <td>0.100</td> <td>MP6296</td> <td>24390</td> <td>10/29/2026</td> </tr> <tr> <td>0.200</td> <td>MP6297</td> <td>25020</td> <td>1/14/2027</td> </tr> <tr> <td>0.300</td> <td>MP6298</td> <td>24430</td> <td>12/10/2026</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG526603</td> <td>9/23/2027</td> </tr> </tbody> </table>		Simulator	Serial #	Lot #	Expiration	0.000	MP6294	N/A	N/A	0.040	MP6295	25090	3/11/2027	0.100	MP6296	24390	10/29/2026	0.200	MP6297	25020	1/14/2027	0.300	MP6298	24430	12/10/2026	0.080 DGS	N/A	AG526603	9/23/2027	Gauge: <u>1012</u> Instrument: <u>1014</u> Mouth Alcohol Solution Lot #: <u>2025-D</u> Exp: <u>09/25/2027</u> Acetone Stock Solution Lot #: <u>2025-B</u> Exp: <u>09/22/2027</u>			
Simulator	Serial #	Lot #	Expiration																														
0.000	MP6294	N/A	N/A																														
0.040	MP6295	25090	3/11/2027																														
0.100	MP6296	24390	10/29/2026																														
0.200	MP6297	25020	1/14/2027																														
0.300	MP6298	24430	12/10/2026																														
0.080 DGS	N/A	AG526603	9/23/2027																														
☒ Post Optical Bench Adjustment Stability Checks		<table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>MP5088</td> <td>202603A</td> <td>3/3/2028</td> </tr> <tr> <td>0.080</td> <td>MP5089</td> <td>202603B</td> <td>3/3/2028</td> </tr> <tr> <td>0.200</td> <td>MP5090</td> <td>202603D</td> <td>3/4/2028</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>23725080A1</td> <td>10/5/2027</td> </tr> </tbody> </table>				Simulator	Serial #	Lot #	Expiration	0.050	MP5088	202603A	3/3/2028	0.080	MP5089	202603B	3/3/2028	0.200	MP5090	202603D	3/4/2028	0.080 DGS	N/A	23725080A1	10/5/2027								
Simulator	Serial #	Lot #	Expiration																														
0.050	MP5088	202603A	3/3/2028																														
0.080	MP5089	202603B	3/3/2028																														
0.200	MP5090	202603D	3/4/2028																														
0.080 DGS	N/A	23725080A1	10/5/2027																														
Gauge ID #: <u>28427</u> Gauge: <u>1015</u> Instrument: <u>1016</u>		<table border="1" style="width: 100%; border-collapse: collapse;"> <tr> <td style="width: 50%; vertical-align: top;"> ☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☒ Optical Bench Adjustment </td> <td style="width: 50%; vertical-align: top;"> ☒ Post-Stability Checks ☐ Flow Adjustment ☐ Form 40 ☒ Other: Calculation Form </td> </tr> </table>				☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☒ Optical Bench Adjustment	☒ Post-Stability Checks ☐ Flow Adjustment ☐ Form 40 ☒ Other: Calculation Form																										
☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☒ Optical Bench Adjustment	☒ Post-Stability Checks ☐ Flow Adjustment ☐ Form 40 ☒ Other: Calculation Form																																
Notes/Suggested Service: <u>mark</u> Tech Review: Added check on stabilities box and performed tank sensor tare to reset reading to zero with no gas attached. SLH 8/20/2026		☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use																															
		<table style="width: 100%;"> <tr> <td style="width: 50%; text-align: center;"> Wen-Chi Pierson Digitally signed by Wen-Chi Pierson Date: 2026.08.20 10:06:28 -04'00' </td> <td style="width: 50%; text-align: center;"> Kaitlyn Spearin Digitally signed by Kaitlyn Spearin Date: 2026.08.20 10:22:14 -04'00' </td> </tr> <tr> <td style="text-align: center;">Tech Review</td> <td style="text-align: center;">Admin Review</td> </tr> </table>				Wen-Chi Pierson Digitally signed by Wen-Chi Pierson Date: 2026.08.20 10:06:28 -04'00'	Kaitlyn Spearin Digitally signed by Kaitlyn Spearin Date: 2026.08.20 10:22:14 -04'00'	Tech Review	Admin Review																								
Wen-Chi Pierson Digitally signed by Wen-Chi Pierson Date: 2026.08.20 10:06:28 -04'00'	Kaitlyn Spearin Digitally signed by Kaitlyn Spearin Date: 2026.08.20 10:22:14 -04'00'																																
Tech Review	Admin Review																																

Stability Checks

80-000900

0.050 g/210L 0.047 to 0.053 g/210L	0.080 g/210L 0.077 to 0.083 g/210L	0.200 g/210L 0.194 to 0.206 g/210L	DGS 0.080 g/210L 0.077 to 0.083 g/210L 50.003 g/210L of Wet																																																																																				
FFWCC Intoxilyzer - Alcohol Analyzer Model 8000 08/04/2026 Software: 8100.27 SN 80-000900 Test g/210L Time <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>17:22</td></tr> <tr><td>Control Test</td><td>0.048</td><td>17:22</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>17:23</td></tr> <tr><td>Control Test</td><td>0.047</td><td>17:24</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>17:24</td></tr> <tr><td>Control Test</td><td>0.048</td><td>17:25</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>17:25</td></tr> </table> Control Test Stats Average 0.0477 Std Dev 0.0005 Rel Std Dev(%) 1.2112 Operator Signature <i>Miguel</i>	Air Blank	0.000	17:22	Control Test	0.048	17:22	Air Blank	0.000	17:23	Control Test	0.047	17:24	Air Blank	0.000	17:24	Control Test	0.048	17:25	Air Blank	0.000	17:25	FFWCC Intoxilyzer - Alcohol Analyzer Model 8000 08/04/2026 Software: 8100.27 SN 80-000900 Test g/210L Time <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>17:43</td></tr> <tr><td>Control Test</td><td>0.076</td><td>17:44</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>17:44</td></tr> <tr><td>Control Test</td><td>0.077</td><td>17:45</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>17:46</td></tr> <tr><td>Control Test</td><td>0.076</td><td>17:46</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>17:47</td></tr> </table> Control Test Stats Average 0.0763 Std Dev 0.0006 Rel Std Dev(%) 0.7564 Operator Signature <i>Miguel</i>	Air Blank	0.000	17:43	Control Test	0.076	17:44	Air Blank	0.000	17:44	Control Test	0.077	17:45	Air Blank	0.000	17:46	Control Test	0.076	17:46	Air Blank	0.000	17:47	FFWCC Intoxilyzer - Alcohol Analyzer Model 8000 08/04/2026 Software: 8100.27 SN 80-000900 Test g/210L Time <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>17:15</td></tr> <tr><td>Control Test</td><td>0.193</td><td>17:17</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>17:17</td></tr> <tr><td>Control Test</td><td>0.193</td><td>17:18</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>17:18</td></tr> <tr><td>Control Test</td><td>0.192</td><td>17:19</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>17:20</td></tr> </table> Control Test Stats Average 0.1927 Std Dev 0.0006 Rel Std Dev(%) 0.2997 Operator Signature <i>Miguel</i>	Air Blank	0.000	17:15	Control Test	0.193	17:17	Air Blank	0.000	17:17	Control Test	0.193	17:18	Air Blank	0.000	17:18	Control Test	0.192	17:19	Air Blank	0.000	17:20	DGS FFWCC Intoxilyzer - Alcohol Analyzer Model 8000 08/04/2026 Software: 8100.27 SN 80-000900 Test g/210L Time <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>17:33</td></tr> <tr><td>Control Test</td><td>0.082</td><td>17:34</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>17:34</td></tr> <tr><td>Control Test</td><td>0.082</td><td>17:35</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>17:35</td></tr> <tr><td>Control Test</td><td>0.082</td><td>17:35</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>17:36</td></tr> </table> Control Test Stats Average 0.0820 Std Dev 0.0000 Rel Std Dev(%) 0.0000 Operator Signature <i>Miguel</i>	Air Blank	0.000	17:33	Control Test	0.082	17:34	Air Blank	0.000	17:34	Control Test	0.082	17:35	Air Blank	0.000	17:35	Control Test	0.082	17:35	Air Blank	0.000	17:36
Air Blank	0.000	17:22																																																																																					
Control Test	0.048	17:22																																																																																					
Air Blank	0.000	17:23																																																																																					
Control Test	0.047	17:24																																																																																					
Air Blank	0.000	17:24																																																																																					
Control Test	0.048	17:25																																																																																					
Air Blank	0.000	17:25																																																																																					
Air Blank	0.000	17:43																																																																																					
Control Test	0.076	17:44																																																																																					
Air Blank	0.000	17:44																																																																																					
Control Test	0.077	17:45																																																																																					
Air Blank	0.000	17:46																																																																																					
Control Test	0.076	17:46																																																																																					
Air Blank	0.000	17:47																																																																																					
Air Blank	0.000	17:15																																																																																					
Control Test	0.193	17:17																																																																																					
Air Blank	0.000	17:17																																																																																					
Control Test	0.193	17:18																																																																																					
Air Blank	0.000	17:18																																																																																					
Control Test	0.192	17:19																																																																																					
Air Blank	0.000	17:20																																																																																					
Air Blank	0.000	17:33																																																																																					
Control Test	0.082	17:34																																																																																					
Air Blank	0.000	17:34																																																																																					
Control Test	0.082	17:35																																																																																					
Air Blank	0.000	17:35																																																																																					
Control Test	0.082	17:35																																																																																					
Air Blank	0.000	17:36																																																																																					

requires optical bench adjustment *sent*

Verified instrument was able to read gas pressure. ☒

FFWCC
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000900
08/06/2026 08:54:23

SLA

Auto Calibration

pg 1 of 2

	<<<<<	3um	>>>>>	<<<<<	9um	>>>>>

Solution = 0.000 g/210L or 0.0000 mg/l, Samples = 4, Discarded = 1						
Sample	% Abs	(% Abs Ref)		% Abs	(% Abs Ref)	
Sample #1	0.0790	(-0.0430)		0.1380	(-0.0260)	
Sample #2	0.1130	(-0.0430)		0.1370	(-0.0250)	
Sample #3	0.0830	(-0.0290)		0.1280	(-0.0240)	
Sample #4	0.0840	(-0.0060)		0.1340	(-0.0070)	
Avg % Abs	0.0933	(-0.0260)		0.1330	(-0.0187)	
STD DEV	0.0170	(0.0187)		0.0046	(0.0101)	
REL STD DEV	18.256	(71.852)		3.446	(54.193)	

Solution = 0.040 g/210L or 0.1905 mg/l, Samples = 4, Discarded = 1						
Sample	% Abs	(% Abs Ref)		% Abs	(% Abs Ref)	
Sample #1	0.7490	(-0.0200)		1.5250	(-0.0110)	
Sample #2	0.7860	(-0.0350)		1.5380	(-0.0170)	
Sample #3	0.7550	(-0.0020)		1.5190	(0.0020)	
Sample #4	0.7620	(-0.0170)		1.5300	(-0.0090)	
Avg % Abs	0.7677	(-0.0180)		1.5290	(-0.0080)	
STD DEV	0.0163	(0.0165)		0.0095	(0.0095)	
REL STD DEV	2.118	(91.793)		0.624	(119.242)	

Solution = 0.100 g/210L or 0.4762 mg/l, Samples = 4, Discarded = 1						
Sample	% Abs	(% Abs Ref)		% Abs	(% Abs Ref)	
Sample #1	1.7430	(-0.0190)		3.5860	(-0.0150)	
Sample #2	1.7850	(-0.0290)		3.6090	(-0.0060)	
Sample #3	1.7710	(-0.0250)		3.5970	(-0.0110)	
Sample #4	1.7420	(-0.0120)		3.5680	(0.0050)	
Avg % Abs	1.7660	(-0.0220)		3.5913	(-0.0040)	
STD DEV	0.0219	(0.0089)		0.0211	(0.0082)	
REL STD DEV	1.242	(40.401)		0.587	(204.634)	

Solution = 0.200 g/210L or 0.9524 mg/l, Samples = 4, Discarded = 1						
Sample	% Abs	(% Abs Ref)		% Abs	(% Abs Ref)	
Sample #1	3.3710	(-0.0200)		6.8570	(-0.0180)	
Sample #2	3.3780	(-0.0040)		6.8360	(0.0050)	
Sample #3	3.3650	(0.0190)		6.8340	(0.0270)	
Sample #4	3.3510	(0.0080)		6.8400	(0.0130)	
Avg % Abs	3.3647	(0.0077)		6.8367	(0.0150)	
STD DEV	0.0135	(0.0115)		0.0031	(0.0111)	
REL STD DEV	0.401	(150.047)		0.045	(74.237)	

Solution = 0.300 g/210L or 1.4286 mg/l, Samples = 4, Discarded = 1						
Sample	% Abs	(% Abs Ref)		% Abs	(% Abs Ref)	
Sample #1	4.9720	(-0.0080)		10.0050	(-0.0220)	
Sample #2	4.9310	(0.0170)		9.9630	(0.0200)	
Sample #3	4.9400	(0.0130)		9.9620	(0.0180)	
Sample #4	4.9220	(0.0340)		9.9730	(0.0050)	
Avg % Abs	4.9310	(0.0213)		9.9660	(0.0143)	
STD DEV	0.0090	(0.0112)		0.0061	(0.0081)	
REL STD DEV	0.183	(52.268)		0.061	(56.822)	

FFWCC
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000900
08/06/2026 08:54:23

Auto Calibration

pg 2 of 2

<<<< 3um >>>>

Zero Order Coef -266.21
First Order Coef 2803.04
Second Order Coef 30.25

<<<< 9um >>>>

-179.13
1340.92
11.13

Act (g/210L)	Fit (g/210L)	Residual (g/210L)
0.000	-0.000	0.0001
0.040	0.040	0.0000
0.100	0.100	-0.0003
0.200	0.200	0.0003
0.300	0.300	-0.0001

Act (g/210L)	Fit (g/210L)	Residual (g/210L)
0.000	-0.000	0.0000
0.040	0.040	0.0002
0.100	0.100	-0.0004
0.200	0.200	0.0003
0.300	0.300	-0.0001

<<<< 3um >>>> <<<< 9um >>>>

Solution = 0.080 g/210L or 0.3810 mg/l, Samples = 4, Discarded = 1
Sample
Sample #1 3117.00 3363.00
Sample #2 3094.00 3325.00
Sample #3 3031.00 3308.00
Sample #4 3050.00 3302.00
Avg 3058.3333 3311.6667
STD DEV 32.3161 11.9304
REL STD DEV 1.057 0.360
H2O adjust (mg/l*10k) 751 498

Barometric Pressure = 1017

*****CALIBRATION SUCCESSFUL*****

SLH

Post-Calibration Adjustment Stability Checks

0.050 g/210L	0.080 g/210L	0.200 g/210L	DGS 0.080 g/210L
0.047 to 0.053 g/210L	0.077 to 0.083 g/210L	0.194 to 0.206 g/210L	0.077 to 0.083 g/210L ≤0.003 g/210L of Wet
FFIACC Intoxilyzer - Alcohol Analyzer Model 8000 08/06/2026 SN 80-000900 Software: 8100.27 Test g/210L Time Air Blank 0.000 15:44 Control Test 0.050 15:44 Air Blank 0.000 15:45 Control Test 0.050 15:46 Air Blank 0.000 15:46 Control Test 0.049 15:47 Air Blank 0.000 15:47 Control Test Stats Average 0.0497 Std Dev 0.0006 Rel Std Dev(%) 1.1625	FFIACC Intoxilyzer - Alcohol Analyzer Model 8000 08/06/2026 SN 80-000900 Software: 8100.27 Test g/210L Time Air Blank 0.000 15:32 Control Test 0.079 15:33 Air Blank 0.000 15:33 Control Test 0.079 15:34 Air Blank 0.000 15:35 Control Test 0.078 15:35 Air Blank 0.000 15:36 Control Test Stats Average 0.0787 Std Dev 0.0016 Rel Std Dev(%) 0.7339	FFIACC Intoxilyzer - Alcohol Analyzer Model 8000 08/06/2026 SN 80-000900 Software: 8100.27 Test g/210L Time Air Blank 0.000 15:37 Control Test 0.198 15:38 Air Blank 0.000 15:39 Control Test 0.198 15:39 Air Blank 0.000 15:40 Control Test 0.198 15:40 Air Blank 0.000 15:41 Control Test Stats Average 0.1980 Std Dev 0.0000 Rel Std Dev(%) 0.0000	FFIACC Intoxilyzer - Alcohol Analyzer Model 8000 08/15/2026 SN 80-000900 Software: 8100.27 Test g/210L Time Air Blank 0.000 14:37 Control Test 0.080 14:37 Air Blank 0.000 14:38 Control Test 0.079 14:38 Air Blank 0.000 14:38 Control Test 0.080 14:39 Air Blank 0.000 14:39 Control Test Stats Average 0.0797 Std Dev 0.0006 Rel Std Dev(%) 0.7247
 Operator's Signature	 Operator's Signature	 Operator's Signature	 Operator's Signature

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: FFWCC

Time of Inspection: 18:22

Date of Inspection: 08/13/2026

Serial Number: 80-000900

Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202603A Exp: 03/03/2028	0.08g/210L Test (g/210L) Lot#:202603B Exp: 03/03/2028	0.20g/210L Test (g/210L) Lot#:202603D Exp: 03/04/2028	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:23725080A1 Exp: 10/05/2027
0.000	0.049	0.079	0.197	0.080
0.000	0.050	0.079	0.197	0.080
0.000	0.050	0.079	0.197	0.080
0.000	0.050	0.079	0.198	0.079
0.000	0.050	0.079	0.198	0.079
0.000	0.050	0.080	0.197	0.080
0.000	0.050	0.079	0.198	0.079
0.000	0.050	0.079	0.198	0.079
0.000	0.050	0.079	0.198	0.080
0.000	0.050	0.079	0.199	0.079

Standard Deviations	0.0003	0.0003	0.0006	0.0005
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.


 LEANDRA HIGGINBOTHAM
 Signature and Printed Name

08/13/2026
Date

Florida Department of Law Enforcement
Alcohol Testing Program - Average Calculations for Calibration Certificate

Calculations By: SLH

Serial Number: 80-000900 Date: 8/19/2026

Solution:

0.05 g/210 L Test: 0.499 / 10 = 0.050 g/210L
(SUM) (AVERAGE)

0.08 g/210L Test: 0.791 / 10 = 0.079 g/210L
(SUM) (AVERAGE)

0.20 g/210L Test: 1.977 / 10 = 0.198 g/210L
(SUM) (AVERAGE)

0.08 g/210L Dry Gas Std Test: 0.795 / 10 = 0.080 g/210L
(SUM) (AVERAGE)

Comments:

Average = (Sum of observations) / (Total number of observations)

All results are reported in g/210L



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000900 , manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-000900</u>	Average of $y \pm$ Uncertainty*
Owning Agency:	<u>FFWCC</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>08/13/2026</u>	0.079 g/ 210 L 0.004
Calibration Time:	<u>18:22</u>	0.198 g/ 210 L 0.008
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

S. LeAndra
Higginbotham
Digitally signed by S. LeAndra
Higginbotham
Date: 2026.08.19 15:35:14 -04'00'

08/13/2026

Date

LEANDRA HIGGINBOTHAM,

Department Inspector

FDLE/ATP Form 69 August 2026
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality