

Alcohol Testing Program - Instrument Processing Sheet

Agency: GILCHRIST COUNTY SO Instrument Serial Number: 80-001285

Date In: 5/18/2026 DI Completion Date: _____ ☐ Ship ☐ P/U ☐ H/D ☒ CMI ☐ EE

Intake By: <u>SLH</u> Date: <u>5/18/2026</u> ☒ Annual ☐ Dropped Off ☐ Registration ☐ Return from CMI / EE ☐ Training Instrument Visual Inspection ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/Accessories ☐ Power Cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes: Replaced bent keyboard screw.	Quality Checks By: <u>IAS</u> Date: <u>5/22/2026</u> ☒ Breath Tube Screen <u>6/5/2026*</u> ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value: <u>129</u> ☒ Flow Verification (L/s) Flow Column #: <u>ATP105</u> 32 mm <u>.148</u> (.139-.169) 36 mm <u>.167</u> (.156-.190) 53 mm <u>.242</u> (.228-.278) 103 mm <u>.531</u> (.447-.547) ☒ Barometric Pressure Check Gauge ID #: <u>28427</u> Gauge: <u>1011</u> Instrument: <u>1006</u> ☒ Stability Checks <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot#/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050*</td> <td>MP6291</td> <td>202603A 3/3/2028</td> </tr> <tr> <td>0.080</td> <td>MP5089</td> <td>202603B 3/3/2028</td> </tr> <tr> <td>0.200*</td> <td>MP6293</td> <td>202406N 6/20/2026</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>23825080A1 10/5/2027</td> </tr> </tbody> </table>	Simulator	Serial #	Lot#/Exp	0.050*	MP6291	202603A 3/3/2028	0.080	MP5089	202603B 3/3/2028	0.200*	MP6293	202406N 6/20/2026	0.080 DGS	N/A	23825080A1 10/5/2027	Flow Adjustment By: _____ Flow Column #: _____ ☐ 5L/min – 17mm ☐ 15L/min – 53mm ☐ 30L/min – 103mm ☐ R-Value: _____ ☐ Post Adjustment Verification (L/S) Flow Column #: _____ 32 mm _____ (.139-.169) 36 mm _____ (.156-.190) 53 mm _____ (.228-.278) 103 mm _____ (.447-.547)																																	
Simulator	Serial #	Lot#/Exp																																																
0.050*	MP6291	202603A 3/3/2028																																																
0.080	MP5089	202603B 3/3/2028																																																
0.200*	MP6293	202406N 6/20/2026																																																
0.080 DGS	N/A	23825080A1 10/5/2027																																																
Optical Bench Adjustment By: _____ Department Inspection By: _____ Barometric Pressure Gauge: _____ ID#: _____ Barometric Pressure ID#: _____ <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> ☐ Post Optical Bench Adjustment Stability Checks <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> Gauge ID #: _____ Instrument: _____ Gauge: _____ Instrument: _____			Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A		
Simulator	Serial #	Lot #	Expiration																																															
0.000		N/A	N/A																																															
0.040																																																		
0.100																																																		
0.200																																																		
0.300																																																		
0.080 DGS	N/A																																																	
Simulator	Serial #	Lot #	Expiration																																															
0.050																																																		
0.080																																																		
0.200																																																		
0.080 DGS	N/A																																																	
Notes/Suggested Service: The tank sensor tare was performed after the repeated 0.08 control. SLH 5/22/2026 Troubleshooted the dry gas regulator and determined that the instrument was not reading the dry gas tank pressure when a tank with sufficient pressure was attached. WKP 5/22/2026 Troubleshooting was performed by attaching a working dry gas regulator and monitoring a change to the PSI reading on the instrument when a dry gas was attached. No change was observed so the original regulator was placed back onto the instrument. WKP 6/12/2026																																																		
☐ Instrument Complies with Chapter 11D-8, FAC ☒ Instrument Does Not Comply with Chapter 11D-8, FAC ☐ Return to/Place into Evidentiary Use ☒ Remain Out of Evidentiary Use ☐ Conduct an Agency Inspection Before Evidentiary Use <table border="1" style="width: 100%; border-collapse: collapse;"> <tr> <td style="width: 50%; text-align: center;"> Digitally signed by Taylor Gutschow Date: 2026.06.19 14:39:52 -04'00' </td> <td style="width: 50%; text-align: center;"> Digitally signed by Kaitlyn Spearin Date: 2026.06.22 10:16:40 -04'00' </td> </tr> </table>			Digitally signed by Taylor Gutschow Date: 2026.06.19 14:39:52 -04'00'	Digitally signed by Kaitlyn Spearin Date: 2026.06.22 10:16:40 -04'00'																																														
Digitally signed by Taylor Gutschow Date: 2026.06.19 14:39:52 -04'00'	Digitally signed by Kaitlyn Spearin Date: 2026.06.22 10:16:40 -04'00'																																																	
Tech Review Admin Review																																																		

Tech Correction: Elaborated upon the troubleshooting steps taken and added an additional reason for return in on the Return Material Authorization. WKP 6/12/2026

Admin Review: Clarifying that the asterisks by the quality check simulator concentrations were performed on the asterisked date. IAS 6/22/2026

Stability Checks

0.050 g/210L	0.080 g/210L	0.200 g/210L	DGS 0.080 g/210L
0.047 to 0.053 g/210L	0.077 to 0.083 g/210L	0.194 to 0.206 g/210L	0.077 to 0.083 g/210L 50.003 g/210L of Wet
CLARKIST COUNTY SD Intoxilizer - Alcohol Analyzer Model 8000 SN 80-001285 05/05/2026 Software: 8100.27 Test g/210L Time Air Blank 0.000 11:57 Control Test 0.000 11:57 Air Blank 0.000 11:58 Control Test 0.048 11:59 Air Blank 0.000 11:59 Control Test 0.048 12:00 Air Blank 0.000 12:00 Control Test Stats Average 0.048 Std Dev 0.000 Rel Std Dev (%) 0.000 *Interferent Detect <i>Radella-Smyth</i> Operator's Signature	CLARKIST COUNTY SD Intoxilizer - Alcohol Analyzer Model 8000 SN 80-001285 05/05/2026 Software: 8100.27 Test g/210L Time Air Blank 0.000 11:47 Control Test 0.194 11:47 Air Blank 0.000 11:48 Control Test 0.192 11:49 Air Blank 0.000 11:49 Control Test 0.191 11:50 Air Blank 0.000 11:50 Control Test Stats Average 0.1923 Std Dev 0.0015 Rel Std Dev (%) 0.7942 <i>Radella-Smyth</i> Operator's Signature	CLARKIST COUNTY SD Intoxilizer - Alcohol Analyzer Model 8000 SN 80-001285 05/05/2026 Software: 8100.27 Test g/210L Time Air Blank 0.000 11:47 Control Test 0.194 11:47 Air Blank 0.000 11:48 Control Test 0.192 11:49 Air Blank 0.000 11:49 Control Test 0.191 11:50 Air Blank 0.000 11:50 Control Test Stats Average 0.1923 Std Dev 0.0015 Rel Std Dev (%) 0.7942 <i>Radella-Smyth</i> Operator's Signature	See other attachment. IAS 6/5/26 See other attachment. IAS 6/5/26

.08 Dry Gas Stability Check

SILVERHIST COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model: 8100 SN: 80-001285
05/22/2026
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:16
Control Test	0.048	13:16
Air Blank	0.000	13:16
Control Test	0.047	13:17
Air Blank	0.000	13:17
Control Test	0.048	13:18
Air Blank	0.000	13:18
Control Test Stats		
Average	0.0477	
Std Dev	0.0006	
Rel Std Dev(%)	1.2112	

Isabella Smykowski
Operator's Signature

.08 Wet Bath Stability Check

SILVERHIST COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model: 8100 SN: 80-001285
05/22/2026
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:33
Control Test	INT*	13:34
Air Blank	0.000	13:35
Control Test	INT*	13:35
Air Blank	0.000	13:36
Control Test	0.076	13:37
Air Blank	0.000	13:37
Control Test Stats		
Average	0.0767	
Std Dev	0.0006	
Rel Std Dev(%)	0.7531	

*Interferent Detect

Isabella Smykowski
Operator's Signature

Dry Gas stability ran outside of the acceptable range. No user error found. Psi levels on instrument read high with or without dry gas cylinder attached. Ran .08 wet bath stability to compare results which also ran outside the acceptable range. Both dry gas and wet bath controls were not within .003 g/210L of each other.
IAS 6/2/26

Return Material Authorization

Ship to:☒ CMI, Inc.☐ Enforcement ElectronicsShipment to repair facility authorized by: Jayson Levy on 5/29/2026Items Returned: ☒ Instrument ☐ Supplies ☐ Other Describe: _____Instrument Model: Intoxilyzer 8000 Serial Number: 80-001285

Bill To Address:

Jayson Levy
8817 NW State Road 45
High Springs, FL 32643

Ship to Address:

FDLE Offsite Mail Facility
c/o FDLE HQ
813-B Lake Bradford Rd
Tallahassee, FL 32304

Reason for Return:

Instrument is not reading dry gas PSI when there is a full tank attached to regulator.
Instrument was intermittently producing a RFI detect exception message during wet bath stability checks. WKP 6/12/2026I require an estimate **BEFORE** any repairs will be authorized and/or conductedPlease contact: Jayson Levy Phone #: 352-339-0143Email: jlevy@alachuasheriff.org

ATP Contact Name: Wen-Chi Pierson

ATP Email: Wen-ChiPierson@fdle.state.fl.us