

Instrument Processing Sheet

Agency: ALACHUA PD Instrument Serial Number: 80-001720
 Date In: 12/19/2025 DI Completion Date: 12/31/2025 Ship P/U H/D CMI EE

Intake By: <u>WKP</u> Date: <u>12/19/25</u>	Quality Checks By: <u>WKP</u> Date: <u>12/22/25</u>	Flow Adjustment By: _____ Date: _____
<input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE <input type="checkbox"/> Return unworked <input type="checkbox"/> Training Visual Inspection <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/Accessories <input type="checkbox"/> Power Cord <input type="checkbox"/> Printer Cable <input type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable	<input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value: <u>161</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column #: <u>ATP 103</u> 32 mm <u>0.152</u> (.139-.169) 36 mm <u>0.167</u> (.156-.190) 53 mm <u>0.242</u> (.228-.278) 103 mm <u>0.500</u> (.447-.547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID #: <u>28662</u> Gauge: <u>1020</u> Instrument: <u>1020</u>	Flow Column #: _____ <input type="checkbox"/> 5L/min – 17mm <input type="checkbox"/> 15L/min – 53mm <input type="checkbox"/> 30L/min – 103mm <input type="checkbox"/> R-Value: _____ <input type="checkbox"/> Post Adjustment Verification (L/S) Flow Column #: _____ 32 mm _____ (.139-.169) 36 mm _____ (.156-.190) 53 mm _____ (.228-.278) 103 mm _____ (.447-.547)

Notes: 	<input checked="" type="checkbox"/> Stability Checks <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot#/Exp</th> </tr> </thead> <tbody> <tr> <td rowspan="2">0.050</td> <td rowspan="2">MP5088</td> <td>202406K</td> </tr> <tr> <td>06/19/2026</td> </tr> <tr> <td rowspan="2">0.080</td> <td rowspan="2">MP5089</td> <td>202406L</td> </tr> <tr> <td>06/19/2026</td> </tr> <tr> <td rowspan="2">0.200</td> <td rowspan="2">MP5090</td> <td>202406N</td> </tr> <tr> <td>06/20/2026</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG510701</td> </tr> <tr> <td></td> <td></td> <td>04/17/2027</td> </tr> </tbody> </table>	Simulator	Serial #	Lot#/Exp	0.050	MP5088	202406K	06/19/2026	0.080	MP5089	202406L	06/19/2026	0.200	MP5090	202406N	06/20/2026	0.080 DGS	N/A	AG510701			04/17/2027	<table border="1" style="width: 100%; border-collapse: collapse;"> <tr> <td style="width: 25%;">Maintenance</td> <td>By: _____</td> <td>Date: _____</td> </tr> <tr> <td colspan="3"> <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement and Tank Sensor Tare <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other: _____ </td> </tr> </table>	Maintenance	By: _____	Date: _____	<input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement and Tank Sensor Tare <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other: _____		
Simulator	Serial #	Lot#/Exp																											
0.050	MP5088	202406K																											
		06/19/2026																											
0.080	MP5089	202406L																											
		06/19/2026																											
0.200	MP5090	202406N																											
		06/20/2026																											
0.080 DGS	N/A	AG510701																											
		04/17/2027																											
Maintenance	By: _____	Date: _____																											
<input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement and Tank Sensor Tare <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other: _____																													

Optical Bench Adjustment By: <u>WKP</u>	Department Inspection By: <u>WKP</u>																																
Barometric Pressure Gauge: <u>1026/1026/1020</u> ID#: <u>28421 (x3)</u>	Barometric Pressure ID#: <u>28662</u>																																
<table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>MP6294 (x2) / MP6289</td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td>MP6295 (x3)</td> <td>25090</td> <td>03/11/2027</td> </tr> <tr> <td>0.100</td> <td>MP6296 (x3)</td> <td>24390/25282*</td> <td>10/29/2026 / 08/12/2027*</td> </tr> <tr> <td>0.200</td> <td>MP6297 (x3)</td> <td>25020</td> <td>01/14/2027</td> </tr> <tr> <td>0.300</td> <td>MP6298 (x3)</td> <td>25150</td> <td>05/06/2027</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>28424080A3</td> <td>11/05/2026</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #	Expiration	0.000	MP6294 (x2) / MP6289	N/A	N/A	0.040	MP6295 (x3)	25090	03/11/2027	0.100	MP6296 (x3)	24390/25282*	10/29/2026 / 08/12/2027*	0.200	MP6297 (x3)	25020	01/14/2027	0.300	MP6298 (x3)	25150	05/06/2027	0.080 DGS	N/A	28424080A3	11/05/2026	Gauge: <u>1020</u> Instrument: <u>1019</u> Mouth Alcohol Solution Lot #: <u>2025-D</u> Exp: <u>09/25/2027</u> Acetone Stock Solution Lot #: <u>2025-B</u> Exp: <u>09/22/2027</u>				
Simulator	Serial #	Lot #	Expiration																														
0.000	MP6294 (x2) / MP6289	N/A	N/A																														
0.040	MP6295 (x3)	25090	03/11/2027																														
0.100	MP6296 (x3)	24390/25282*	10/29/2026 / 08/12/2027*																														
0.200	MP6297 (x3)	25020	01/14/2027																														
0.300	MP6298 (x3)	25150	05/06/2027																														
0.080 DGS	N/A	28424080A3	11/05/2026																														
<input checked="" type="checkbox"/> Post Optical Bench Adjustment Stability Checks <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>MP5088 (x3)</td> <td>202406K</td> <td>06/19/2026</td> </tr> <tr> <td>0.080</td> <td>MP5089 (x3)</td> <td>202406L</td> <td>06/19/2026</td> </tr> <tr> <td>0.200</td> <td>MP5090 (x3)</td> <td>202406N</td> <td>06/20/2026</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG510701</td> <td>04/17/2027</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #	Expiration	0.050	MP5088 (x3)	202406K	06/19/2026	0.080	MP5089 (x3)	202406L	06/19/2026	0.200	MP5090 (x3)	202406N	06/20/2026	0.080 DGS	N/A	AG510701	04/17/2027	<table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>MP6289</td> </tr> <tr> <td>Interferent</td> <td>MP6290</td> </tr> <tr> <td>0.050</td> <td>MP5088</td> </tr> <tr> <td>0.080</td> <td>MP5089</td> </tr> <tr> <td>0.200</td> <td>MP5090</td> </tr> </tbody> </table>	Simulator	Serial Number	0.000	MP6289	Interferent	MP6290	0.050	MP5088	0.080	MP5089	0.200	MP5090
Simulator	Serial #	Lot #	Expiration																														
0.050	MP5088 (x3)	202406K	06/19/2026																														
0.080	MP5089 (x3)	202406L	06/19/2026																														
0.200	MP5090 (x3)	202406N	06/20/2026																														
0.080 DGS	N/A	AG510701	04/17/2027																														
Simulator	Serial Number																																
0.000	MP6289																																
Interferent	MP6290																																
0.050	MP5088																																
0.080	MP5089																																
0.200	MP5090																																
Barometric Pressure Gauge: <u>1024/1026/1017</u> ID#: <u>28427 (x3)</u>	<table border="1" style="width: 100%; border-collapse: collapse;"> <tr> <td style="width: 50%;">Attachments</td> <td style="width: 50%;"><input checked="" type="checkbox"/> Post-Stability Checks (x3)</td> </tr> <tr> <td><input checked="" type="checkbox"/> Form 41</td> <td><input type="checkbox"/> Flow Adjustment</td> </tr> <tr> <td><input checked="" type="checkbox"/> Stability Checks</td> <td><input type="checkbox"/> Form 40</td> </tr> <tr> <td><input checked="" type="checkbox"/> Calibration Certificate</td> <td><input checked="" type="checkbox"/> Other: Agency Memo</td> </tr> <tr> <td><input checked="" type="checkbox"/> Optical Bench Adjustment (x3)</td> <td></td> </tr> </table>	Attachments	<input checked="" type="checkbox"/> Post-Stability Checks (x3)	<input checked="" type="checkbox"/> Form 41	<input type="checkbox"/> Flow Adjustment	<input checked="" type="checkbox"/> Stability Checks	<input type="checkbox"/> Form 40	<input checked="" type="checkbox"/> Calibration Certificate	<input checked="" type="checkbox"/> Other: Agency Memo	<input checked="" type="checkbox"/> Optical Bench Adjustment (x3)																							
Attachments	<input checked="" type="checkbox"/> Post-Stability Checks (x3)																																
<input checked="" type="checkbox"/> Form 41	<input type="checkbox"/> Flow Adjustment																																
<input checked="" type="checkbox"/> Stability Checks	<input type="checkbox"/> Form 40																																
<input checked="" type="checkbox"/> Calibration Certificate	<input checked="" type="checkbox"/> Other: Agency Memo																																
<input checked="" type="checkbox"/> Optical Bench Adjustment (x3)																																	

Notes/Suggested Service:
 * 0.100 Lot # and Expiration date for optical bench adjustment #3. WKP 12/30/2025
 Optical Bench Adjustment #2 and #3 conducted to address the outside acceptable range values during the Post Optical Bench Adjustment Stability Checks #1 and #2, respectively. Prior to beginning Optical Bench Adjustment #3, ensured alignment of instrument ports and changed simulator O-rings. WKP 12/30/2025

<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use	Digitally signed by Shayla Platt Date: 2026.01.12 10:31:28 -05'00' Digitally signed by Shayla Platt Date: 2026.01.12 10:33:56 -05'00'
Shayla Platt Tech Review	Shayla Platt Admin Review

December 15, 2025

To: FDLE Off-Site Mail Facility
c/o Florida Department of Law Enforcement
Alcohol Testing Program
813 B Lake Brandford Road
Tallahassee, Fl. 32304

From: Field Service Technician/SCG Supervisor Toni Fulton
Alachua County Sheriff's Office
3371 NE 39th Ave
Gainesville, FL 32609

To whom it may concern,

Enclosed is Alachua Police Departments Intoxilyzer 8000 serial number 80-001720 for the annual inspection. If you have any questions or concerns, please contact me at 352-367-4100 or tfulton@acso.us.

Thank you,



FST Toni Fulton #549
Field Service Technician/SCG Supervisor

Stability Checks

0.05g/210L 0.047 to 0.053	0.08g/210L 0.077 to 0.083	0.20g/210L 0.194 to 0.206	DGS 0.08g/210L 0.077 to 0.083 ✓ ≤ 0.003 of Wet																																																																																																																																																
<p>ALC-10A PD Intoxilyzer - Alcolon Analyzer Model 8000 SN 80-001720 12/22/2025 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>15:27</td></tr> <tr><td>Control Test</td><td>0.042</td><td>15:28</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:29</td></tr> <tr><td>Control Test</td><td>0.042</td><td>15:29</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:30</td></tr> <tr><td>Control Test</td><td>0.042</td><td>15:30</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:31</td></tr> <tr><td>Control Test</td><td>0.042</td><td>15:31</td></tr> <tr><td>Average</td><td>0.0420</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel. Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> <p>Operator's Signature: <i>manpi</i></p>	Test	g/210L	Time	Air Blank	0.000	15:27	Control Test	0.042	15:28	Air Blank	0.000	15:29	Control Test	0.042	15:29	Air Blank	0.000	15:30	Control Test	0.042	15:30	Air Blank	0.000	15:31	Control Test	0.042	15:31	Average	0.0420		Std Dev	0.0000		Rel. Std Dev(%)	0.0000		<p>ALC-10A PD Intoxilyzer - Alcolon Analyzer Model 8000 SN 80-001720 12/22/2025 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>15:33</td></tr> <tr><td>Control Test</td><td>0.061</td><td>15:33</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:34</td></tr> <tr><td>Control Test</td><td>0.061</td><td>15:34</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:35</td></tr> <tr><td>Control Test</td><td>0.060</td><td>15:35</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:36</td></tr> <tr><td>Control Test</td><td>0.060</td><td>15:36</td></tr> <tr><td>Average</td><td>0.0607</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel. Std Dev(%)</td><td>0.9517</td><td></td></tr> </tbody> </table> <p>Operator's Signature: <i>manpi</i></p>	Test	g/210L	Time	Air Blank	0.000	15:33	Control Test	0.061	15:33	Air Blank	0.000	15:34	Control Test	0.061	15:34	Air Blank	0.000	15:35	Control Test	0.060	15:35	Air Blank	0.000	15:36	Control Test	0.060	15:36	Average	0.0607		Std Dev	0.0006		Rel. Std Dev(%)	0.9517		<p>ALC-10A PD Intoxilyzer - Alcolon Analyzer Model 8000 SN 80-011720 12/22/2025 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>15:36</td></tr> <tr><td>Control Test</td><td>0.175</td><td>15:36</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:37</td></tr> <tr><td>Control Test</td><td>0.175</td><td>15:37</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:38</td></tr> <tr><td>Control Test</td><td>0.176</td><td>15:38</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:39</td></tr> <tr><td>Control Test</td><td>0.176</td><td>15:39</td></tr> <tr><td>Average</td><td>0.1753</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel. Std Dev(%)</td><td>3.3293</td><td></td></tr> </tbody> </table> <p>Operator's Signature: <i>manpi</i></p>	Test	g/210L	Time	Air Blank	0.000	15:36	Control Test	0.175	15:36	Air Blank	0.000	15:37	Control Test	0.175	15:37	Air Blank	0.000	15:38	Control Test	0.176	15:38	Air Blank	0.000	15:39	Control Test	0.176	15:39	Average	0.1753		Std Dev	0.0006		Rel. Std Dev(%)	3.3293		<p>ALC-10A PD Intoxilyzer - Alcolon Analyzer Model 8000 SN 80-011720 12/22/2025 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>15:43</td></tr> <tr><td>Control Test</td><td>0.078</td><td>15:44</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:44</td></tr> <tr><td>Control Test</td><td>0.077</td><td>15:44</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:45</td></tr> <tr><td>Control Test</td><td>0.077</td><td>15:45</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:46</td></tr> <tr><td>Control Test</td><td>0.000</td><td>15:46</td></tr> <tr><td>Average</td><td>0.0773</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel. Std Dev(%)</td><td>0.7466</td><td></td></tr> </tbody> </table> <p>Operator's Signature: <i>manpi</i></p>	Test	g/210L	Time	Air Blank	0.000	15:43	Control Test	0.078	15:44	Air Blank	0.000	15:44	Control Test	0.077	15:44	Air Blank	0.000	15:45	Control Test	0.077	15:45	Air Blank	0.000	15:46	Control Test	0.000	15:46	Average	0.0773		Std Dev	0.0006		Rel. Std Dev(%)	0.7466	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	15:27																																																																																																																																																	
Control Test	0.042	15:28																																																																																																																																																	
Air Blank	0.000	15:29																																																																																																																																																	
Control Test	0.042	15:29																																																																																																																																																	
Air Blank	0.000	15:30																																																																																																																																																	
Control Test	0.042	15:30																																																																																																																																																	
Air Blank	0.000	15:31																																																																																																																																																	
Control Test	0.042	15:31																																																																																																																																																	
Average	0.0420																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel. Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	15:33																																																																																																																																																	
Control Test	0.061	15:33																																																																																																																																																	
Air Blank	0.000	15:34																																																																																																																																																	
Control Test	0.061	15:34																																																																																																																																																	
Air Blank	0.000	15:35																																																																																																																																																	
Control Test	0.060	15:35																																																																																																																																																	
Air Blank	0.000	15:36																																																																																																																																																	
Control Test	0.060	15:36																																																																																																																																																	
Average	0.0607																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel. Std Dev(%)	0.9517																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	15:36																																																																																																																																																	
Control Test	0.175	15:36																																																																																																																																																	
Air Blank	0.000	15:37																																																																																																																																																	
Control Test	0.175	15:37																																																																																																																																																	
Air Blank	0.000	15:38																																																																																																																																																	
Control Test	0.176	15:38																																																																																																																																																	
Air Blank	0.000	15:39																																																																																																																																																	
Control Test	0.176	15:39																																																																																																																																																	
Average	0.1753																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel. Std Dev(%)	3.3293																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	15:43																																																																																																																																																	
Control Test	0.078	15:44																																																																																																																																																	
Air Blank	0.000	15:44																																																																																																																																																	
Control Test	0.077	15:44																																																																																																																																																	
Air Blank	0.000	15:45																																																																																																																																																	
Control Test	0.077	15:45																																																																																																																																																	
Air Blank	0.000	15:46																																																																																																																																																	
Control Test	0.000	15:46																																																																																																																																																	
Average	0.0773																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel. Std Dev(%)	0.7466																																																																																																																																																		

ALACHUA PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001720
12/22/2025
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	15:21
Control Test	0.042	15:22
Air Blank	0.000	15:22
Control Test	0.041	15:23
Air Blank	0.000	15:24
Control Test	0.040	15:24
Air Blank	0.000	15:25
Control Test Stats		
Average	0.0410	
Std Dev	0.0010	
Rel Std Dev(%)	2.4390	


Operator's Signature

0.05g/210L stability #1
outside of acceptable
range, retightened
SM seal and
connection to instrument,
then repeated.
WKP 12/23/25

***** AUTO CAL DATA *****

<<<<< CHANNEL 1 >>>>>
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.115
Std Dev = 0.01 Rel Std Dev = 5.29
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 0.798
Std Dev = 0.01 Rel Std Dev = 0.80
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 1.498
Std Dev = 0.01 Rel Std Dev = 0.95
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 3.354
Std Dev = 0.06 Rel Std Dev = 1.90
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 4.418
Std Dev = 0.02 Rel Std Dev = 0.37
Zero Order Coef = -162.64
First Order Coef = 2666.94
Second Order Coef = 116.62
Standard Deviation = 486.745636

<<<<< CHANNEL 2 >>>>>
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.112
Std Dev = 0.01 Rel Std Dev = 5.44
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 1.430
Std Dev = 0.02 Rel Std Dev = 1.15
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 2.764
Std Dev = 0.03 Rel Std Dev = 1.00
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 6.203
Std Dev = 0.11 Rel Std Dev = 1.85
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 8.164
Std Dev = 0.03 Rel Std Dev = 0.33
Zero Order Coef = -23.66
First Order Coef = 1400.34
Second Order Coef = 38.70
Standard Deviation = 476.524689

<<<<< CHANNEL 2 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 6.1990 (-0.0150)
Sample #2 = 6.1350 (0.0150)
Sample #3 = 6.1390 (0.0210)
Sample #4 = 6.3360 (0.0170)
Avg % Abs = 6.2033 (0.0177)
STD DEV = 0.1149 (0.0231)
REL STD DEV = 1.852 (17.2933)

Sol Value = 0.300 g/210L ***
Fit Value = 1.4286 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12720, Sum Io = 12984
<<<<< CHANNEL 1 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 4.4340 (-0.0140)
Sample #2 = 4.4360 (0.0020)
Sample #3 = 4.4040 (0.0000)
Sample #4 = 4.4150 (0.0030)
Avg % Abs = 4.4183 (0.0017)
STD DEV = 0.0163 (0.0015)
REL STD DEV = 0.368 (91.652)

<<<<< CHANNEL 2 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 8.2100 (-0.0080)
Sample #2 = 8.1950 (0.0090)
Sample #3 = 8.1490 (0.0140)
Sample #4 = 8.1490 (0.0190)
Avg % Abs = 8.1643 (0.0140)
STD DEV = 0.0266 (0.0050)
REL STD DEV = 0.325 (35.714)

Optical
Bench Calibration
Adjustment #1
By: W.K.P

<<<<< CHANNEL 2 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 1.4250 (-0.0010)
Sample #2 = 1.4490 (-0.0100)
Sample #3 = 1.4210 (-0.0030)
Sample #4 = 1.4280 (-0.0020)
Avg % Abs = 1.4300 (-0.0050)
STD DEV = 0.0165 (0.0044)
REL STD DEV = 1.151 (87.178)

Sol Value = 0.180 g/210L ***
Fit Value = 0.4762 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12730, Sum Io = 12990
<<<<< CHANNEL 1 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 1.5940 (-0.0110)
Sample #2 = 1.5880 (-0.0080)
Sample #3 = 1.5050 (0.0030)
Sample #4 = 1.4820 (0.0070)
Avg % Abs = 1.4983 (0.0007)
STD DEV = 0.0142 (0.0078)
REL STD DEV = 0.949 (1165.118)

<<<<< CHANNEL 2 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 2.8400 (0.0000)
Sample #2 = 2.7810 (0.0180)
Sample #3 = 2.7790 (0.0200)
Sample #4 = 2.7320 (0.0160)
Avg % Abs = 2.7640 (0.0180)
STD DEV = 0.0277 (0.0020)
REL STD DEV = 1.003 (11.111)

Sol Value = 0.200 g/210L ***
Fit Value = 0.9524 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12723, Sum Io = 12985
<<<<< CHANNEL 1 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 3.3400 (-0.0370)
Sample #2 = 3.3140 (-0.0110)
Sample #3 = 3.3210 (-0.0030)
Sample #4 = 3.4280 (0.0000)
Avg % Abs = 3.3543 (-0.0047)
STD DEV = 0.0639 (0.0057)
REL STD DEV = 1.905 (121.848)

ALCACHUA PD
Intoxilizer - Alcohol Analyzer
Model 8000
12/23/2025
SN 80-001720
10:33:09
Auto Calibration
Max Power Res Value = 30
Auto Range Res Value = 18

Sol Value = 0.600 g/210L ***
Fit Value = 0.0000 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12760, Sum Io = 13005
<<<<< CHANNEL 1 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 0.1360 (0.0000)
Sample #2 = 0.1120 (0.0360)
Sample #3 = 0.1220 (0.0470)
Sample #4 = 0.1110 (0.0690)
Avg % Abs = 0.1150 (0.0507)
STD DEV = 0.0061 (0.0168)
REL STD DEV = 5.289 (333.163)

<<<<< CHANNEL 2 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 0.1430 (-0.0130)
Sample #2 = 0.1110 (0.0040)
Sample #3 = 0.1110 (0.0110)
Sample #4 = 0.1070 (0.0230)
Avg % Abs = 0.1123 (0.0127)
STD DEV = 0.0061 (0.0096)
REL STD DEV = 5.439 (75.861)

Sol Value = 0.040 g/210L ***
Fit Value = 0.1905 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12739, Sum Io = 12991
<<<<< CHANNEL 1 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 0.7870 (-0.0080)
Sample #2 = 0.7940 (0.0060)
Sample #3 = 0.8850 (0.0190)
Sample #4 = 0.7940 (0.0220)
Avg % Abs = 0.7977 (0.0157)
STD DEV = 0.0064 (0.0085)
REL STD DEV = 0.796 (54.287)

Optical
benon calorimeter
Adjustment #1

By: WKP

Solution Stats Quadratic Fit Chan 1

Act	Fit	Residual
g/210L	g/210L	g/210L
0.000	0.003	-0.0031
0.040	0.043	-0.0032
0.100	0.087	0.0134
0.200	0.213	-0.0134
0.300	0.294	0.0062

Solution Stats Quadratic Fit Chan 2

Act	Fit	Residual
g/210L	g/210L	g/210L
0.000	0.003	-0.0028
0.040	0.043	-0.0032
0.100	0.087	0.0130
0.200	0.213	-0.0132
0.300	0.294	0.0062

Sol Value = 0.080 g/210L ***
 Fit value = 0.3810 mg/l %%%
 Samples Taken = 4, Discarded = 1

***** CHANNEL 1
 Sample #1 = 3638.00
 Sample #2 = 3571.00
 Sample #3 = 3606.00
 Sample #4 = 3606.00
 Average Result = 3594.3333
 STD DEV = 20.2073
 REL STD DEV = 0.562

***** CHANNEL 2
 Sample #1 = 3857.00
 Sample #2 = 3826.00
 Sample #3 = 3835.00
 Sample #4 = 3820.00
 Average Result = 3827.0000
 STD DEV = 7.5498
 REL STD DEV = 0.197

 Dry Gas H2O Adjust Results *****
 Barometric Pressure = 1026
 3 um H2O Adjust (mg/l*10,000) = 215
 9 um H2O Adjust (mg/l*10,000) = -17

 **** AUTO CAL PASS

Sol Value = 0.040 g/210L ***
 Fit value = 0.1905 mg/l %%%
 Samples Taken = 4, Discarded = 1
 Sum 10 = 12709, Sum 10 = 12977
 <<<<< CHANNEL 1 >>>>>
 Sample % Abs (% Abs Ref)
 Sample #1 = 0.8100 (-0.0220)
 Sample #2 = 0.7780 (0.0000)
 Sample #3 = 0.7760 (0.0180)
 Sample #4 = 0.7660 (0.0200)
 Avg % Abs = 0.7733 (0.0127)
 STD DEV = 0.0164 (0.0110)
 REL STD DEV = 0.831 (86.962)

<<<<< CHANNEL 2 >>>>>
 Sample % Abs (% Abs Ref)
 Sample #1 = 1.4380 (-0.0130)
 Sample #2 = 1.4030 (0.0140)
 Sample #3 = 1.4250 (0.0150)
 Sample #4 = 1.4260 (0.0040)
 Avg % Abs = 1.4180 (0.0110)
 STD DEV = 0.0130 (0.0061)
 REL STD DEV = 0.917 (55.298)

Sol Value = 0.100 g/210L ***
 Fit value = 0.4762 mg/l %%%
 Samples Taken = 4, Discarded = 1
 Sum 10 = 12795, Sum 10 = 12973
 <<<<< CHANNEL 1 >>>>>
 Sample % Abs (% Abs Ref)
 Sample #1 = 1.8180 (-0.0080)
 Sample #2 = 1.8180 (0.0100)
 Sample #3 = 1.8290 (0.0120)
 Sample #4 = 1.8100 (0.0230)
 Avg % Abs = 1.8190 (0.0150)
 STD DEV = 0.0095 (0.0070)
 REL STD DEV = 0.524 (46.667)

<<<<< CHANNEL 2 >>>>>
 Sample % Abs (% Abs Ref)
 Sample #1 = 3.3610 (0.0010)
 Sample #2 = 3.4040 (0.0000)
 Sample #3 = 3.3960 (0.0030)
 Sample #4 = 3.3810 (0.0020)
 Avg % Abs = 3.3937 (0.0017)
 STD DEV = 0.0117 (0.0015)
 REL STD DEV = 0.344 (91.652)

Sol Value = 0.200 g/210L ***
 Fit value = 0.9524 mg/l %%%
 Samples Taken = 4, Discarded = 1
 Sum 10 = 12701, Sum 10 = 12970
 <<<<< CHANNEL 1 >>>>>
 Sample % Abs (% Abs Ref)
 Sample #1 = 3.6210 (-0.0150)
 Sample #2 = 3.5810 (0.0060)
 Sample #3 = 3.5900 (0.0080)
 Sample #4 = 3.6030 (0.0100)
 Avg % Abs = 3.5913 (0.0080)
 STD DEV = 0.0111 (0.0028)
 REL STD DEV = 0.308 (25.000)

<<<<< CHANNEL 2 >>>>>
 Sample % Abs (% Abs Ref)
 Sample #1 = 6.6580 (-0.0050)
 Sample #2 = 6.6690 (-0.0080)
 Sample #3 = 6.6130 (0.0170)
 Sample #4 = 6.6640 (-0.0030)
 Avg % Abs = 6.6487 (0.0020)
 STD DEV = 0.0319 (0.0132)
 REL STD DEV = 0.466 (66.1438)

Sol Value = 0.300 g/210L ***
 Fit value = 1.4286 mg/l %%%
 Samples Taken = 4, Discarded = 1
 Sum 10 = 12697, Sum 10 = 12969
 <<<<< CHANNEL 1 >>>>>
 Sample % Abs (% Abs Ref)
 Sample #1 = 5.1350 (-0.0160)
 Sample #2 = 5.1310 (-0.0020)
 Sample #3 = 5.1520 (0.0080)
 Sample #4 = 5.1360 (0.0060)
 Avg % Abs = 5.1397 (0.0040)
 STD DEV = 0.0110 (0.0053)
 REL STD DEV = 0.213 (132.288)

<<<<< CHANNEL 2 >>>>>
 Sample % Abs (% Abs Ref)
 Sample #1 = 9.4100 (-0.0010)
 Sample #2 = 9.4320 (0.0130)
 Sample #3 = 9.4690 (0.0150)
 Sample #4 = 9.4240 (0.0210)
 Avg % Abs = 9.4417 (0.0163)
 STD DEV = 0.0240 (0.0042)
 REL STD DEV = 0.254 (25.490)

***** AUTO CAL DATA *****
 <<<<< CHANNEL 1 >>>>>
 Sol Val = 0.0000 mg/l or 0.000 g/210L
 % Abs = 0.102
 Std Dev = 0.01 Rel Std Dev = 10.09
 Sol Val = 0.1905 mg/l or 0.040 g/210L
 % Abs = 0.773
 Std Dev = 0.01 Rel Std Dev = 0.83
 Sol Val = 0.4762 mg/l or 0.100 g/210L
 % Abs = 1.819
 Std Dev = 0.01 Rel Std Dev = 0.52
 Sol Val = 0.9524 mg/l or 0.200 g/210L
 % Abs = 3.591
 Std Dev = 0.01 Rel Std Dev = 0.31
 Sol Val = 1.4286 mg/l or 0.300 g/210L
 % Abs = 5.140
 Std Dev = 0.01 Rel Std Dev = 0.21
 Zero Order Coef = -188.20
 First Order Coef = 2593.50
 Second Order Coef = 41.32
 Standard Deviation = 99.839767

<<<<< CHANNEL 2 >>>>>
 Sol Val = 0.0000 mg/l or 0.000 g/210L
 % Abs = 0.101
 Std Dev = 0.01 Rel Std Dev = 12.17
 Sol Val = 0.1915 mg/l or 0.040 g/210L
 % Abs = 1.418
 Std Dev = 0.01 Rel Std Dev = 0.92
 Sol Val = 0.4762 mg/l or 0.100 g/210L
 % Abs = 3.394
 Std Dev = 0.01 Rel Std Dev = 0.34
 Sol Val = 0.9524 mg/l or 0.200 g/210L
 % Abs = 6.649
 Std Dev = 0.03 Rel Std Dev = 0.47
 Sol Val = 1.4286 mg/l or 0.300 g/210L
 % Abs = 9.442
 Std Dev = 0.02 Rel Std Dev = 0.25
 Zero Order Coef = -73.82
 First Order Coef = 1333.99
 Second Order Coef = 19.24
 Standard Deviation = 88.010094

Solution Stats Quadratic Fit Chan 1
 Act Fit Residual
 g/210L g/210L g/210L
 0.000 0.002 -0.0016
 0.040 0.039 0.0013
 0.100 0.098 0.0028
 0.200 0.203 -0.0028
 0.300 0.299 0.0011

Solution Stats Quadratic Fit Chan 2
 Act Fit Residual
 g/210L g/210L g/210L
 0.000 0.001 -0.0013
 0.040 0.039 0.0010
 0.100 0.098 0.0018
 0.200 0.203 -0.0026
 0.300 0.299 0.0010

Sol Value = 0.080 g/210L ***
 Fit value = 0.3810 mg/l %%%
 Samples Taken = 4, Discarded = 1
 ***** CHANNEL 1 *****
 Sample #1 = 3351.00
 Sample #2 = 3340.00
 Sample #3 = 3343.00
 Sample #4 = 3367.00
 Average Result = 3350.0000
 STD DEV = 14.7986
 REL STD DEV = 0.442
 ***** CHANNEL 2 *****
 Sample #1 = 3495.00
 Sample #2 = 3492.00
 Sample #3 = 3471.00
 Sample #4 = 3479.00
 Average Result = 3480.6667
 STD DEV = 10.5987
 REL STD DEV = 0.305

 Dry Gas H2O Adjust Results *****
 Barometric Pressure = 1024
 3 um H2O Adjust (mg/l*10,000) = 459
 9 um H2O Adjust (mg/l*10,000) = 329
 ***** AUTO CAL PASS *****

Optical Bench
 Calibration Adjust-
 ment By: WKP
 #2

ALCACHA PD
 Intoxilyzer - Alcohol Analyzer
 Model 8000
 12/23/2025
 SN 80-001720
 11:55:06
 Auto Calibration
 Max Power Res Value = 30
 Auto Range Res Value = 18

Post-Cal Stability Checks

0.05g/210L 0.047 to 0.053	0.08g/210L 0.077 to 0.083	0.20g/210L 0.194 to 0.206	DGS 0.08g/210L 0.077 to 0.083 ✓ ≤0.003 of Wet X																																																																																																																																																
<p>ALC-100 PD Intoxilyzer - Alcohol Analyzer Model: 8100 SN: 80-001720 12/23/2025 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>13:13</td></tr> <tr><td>Control Test</td><td>0.051</td><td>13:14</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:14</td></tr> <tr><td>Control Test</td><td>0.050</td><td>13:15</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:16</td></tr> <tr><td>Control Test</td><td>0.050</td><td>13:16</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:17</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0503</td><td></td></tr> <tr><td>Std Dev</td><td>0.0005</td><td></td></tr> <tr><td>Rel. Std Dev(%)</td><td>1.004</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	13:13	Control Test	0.051	13:14	Air Blank	0.000	13:14	Control Test	0.050	13:15	Air Blank	0.000	13:16	Control Test	0.050	13:16	Air Blank	0.000	13:17	Control Test Stats			Average	0.0503		Std Dev	0.0005		Rel. Std Dev(%)	1.004		<p>ALC-100 PD Intoxilyzer - Alcohol Analyzer Model: 8100 SN: 80-001720 12/23/2025 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>13:22</td></tr> <tr><td>Control Test</td><td>0.081</td><td>13:23</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:24</td></tr> <tr><td>Control Test</td><td>0.076</td><td>13:24</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:25</td></tr> <tr><td>Control Test</td><td>0.082</td><td>13:25</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:26</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0797</td><td></td></tr> <tr><td>Std Dev</td><td>0.0032</td><td></td></tr> <tr><td>Rel. Std Dev(%)</td><td>4.0350</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	13:22	Control Test	0.081	13:23	Air Blank	0.000	13:24	Control Test	0.076	13:24	Air Blank	0.000	13:25	Control Test	0.082	13:25	Air Blank	0.000	13:26	Control Test Stats			Average	0.0797		Std Dev	0.0032		Rel. Std Dev(%)	4.0350		<p>ALC-100 PD Intoxilyzer - Alcohol Analyzer Model: 8100 SN: 80-001720 12/23/2025 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>13:29</td></tr> <tr><td>Control Test</td><td>0.207</td><td>13:30</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:31</td></tr> <tr><td>Control Test</td><td>0.208</td><td>13:31</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:31</td></tr> <tr><td>Control Test</td><td>0.207</td><td>13:32</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:33</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.2073</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel. Std Dev(%)</td><td>0.2785</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	13:29	Control Test	0.207	13:30	Air Blank	0.000	13:31	Control Test	0.208	13:31	Air Blank	0.000	13:31	Control Test	0.207	13:32	Air Blank	0.000	13:33	Control Test Stats			Average	0.2073		Std Dev	0.0006		Rel. Std Dev(%)	0.2785		<p>ALC-100 PD Intoxilyzer - Alcohol Analyzer Model: 8100 SN: 80-001720 12/23/2025 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>13:34</td></tr> <tr><td>Control Test</td><td>0.080</td><td>13:34</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:35</td></tr> <tr><td>Control Test</td><td>0.079</td><td>13:35</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:36</td></tr> <tr><td>Control Test</td><td>0.080</td><td>13:36</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:36</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0797</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel. Std Dev(%)</td><td>0.7247</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	13:34	Control Test	0.080	13:34	Air Blank	0.000	13:35	Control Test	0.079	13:35	Air Blank	0.000	13:36	Control Test	0.080	13:36	Air Blank	0.000	13:36	Control Test Stats			Average	0.0797		Std Dev	0.0006		Rel. Std Dev(%)	0.7247	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	13:13																																																																																																																																																	
Control Test	0.051	13:14																																																																																																																																																	
Air Blank	0.000	13:14																																																																																																																																																	
Control Test	0.050	13:15																																																																																																																																																	
Air Blank	0.000	13:16																																																																																																																																																	
Control Test	0.050	13:16																																																																																																																																																	
Air Blank	0.000	13:17																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0503																																																																																																																																																		
Std Dev	0.0005																																																																																																																																																		
Rel. Std Dev(%)	1.004																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	13:22																																																																																																																																																	
Control Test	0.081	13:23																																																																																																																																																	
Air Blank	0.000	13:24																																																																																																																																																	
Control Test	0.076	13:24																																																																																																																																																	
Air Blank	0.000	13:25																																																																																																																																																	
Control Test	0.082	13:25																																																																																																																																																	
Air Blank	0.000	13:26																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0797																																																																																																																																																		
Std Dev	0.0032																																																																																																																																																		
Rel. Std Dev(%)	4.0350																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	13:29																																																																																																																																																	
Control Test	0.207	13:30																																																																																																																																																	
Air Blank	0.000	13:31																																																																																																																																																	
Control Test	0.208	13:31																																																																																																																																																	
Air Blank	0.000	13:31																																																																																																																																																	
Control Test	0.207	13:32																																																																																																																																																	
Air Blank	0.000	13:33																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.2073																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel. Std Dev(%)	0.2785																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	13:34																																																																																																																																																	
Control Test	0.080	13:34																																																																																																																																																	
Air Blank	0.000	13:35																																																																																																																																																	
Control Test	0.079	13:35																																																																																																																																																	
Air Blank	0.000	13:36																																																																																																																																																	
Control Test	0.080	13:36																																																																																																																																																	
Air Blank	0.000	13:36																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0797																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel. Std Dev(%)	0.7247																																																																																																																																																		
<p><i>Wenji</i> Operator's Signature</p>	<p><i>Wenji</i> Operator's Signature</p>	<p><i>Wenji</i> Operator's Signature</p>	<p><i>Wenji</i> Operator's Signature</p>																																																																																																																																																

0.07g/210L
 wet stability (post calibration #2)
 outside acceptable
 range, tightened
 sim head and ensured
 tight port connection,
 repeated.

WKP 12/23/25

ALACHUA PD
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001720
 12/23/2025
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:18
Control Test	0.065	13:19
Air Blank	0.000	13:19
Control Test	0.073	13:20
Air Blank	0.000	13:20
Control Test	0.082	13:21
Air Blank	0.000	13:22
Control Test Stats		
Average	0.0733	
Std Dev	0.0085	
Rel Std Dev(%)	11.5976	


 Operator's Signature

ALPHA 80
Intoxilyzer - Alcohol Analyzer
Model 8000
12/30/2025

SN 80-001720
11:06:54

Auto Calibration
Max Power Res Value = 31
Auto Range Res Value = 18

Sol Value = 0.000 g/210L ***
Fit Value = 0.0000 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12728, Sum Io = 12990
<<<<< CHANNEL 1 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 0.1370 (-0.0200)
Sample #2 = 0.1370 (-0.0120)
Sample #3 = 0.1170 (-0.0490)
Sample #4 = 0.1320 (-0.0640)
Avg % Abs = 0.1267 (-0.0417)
STD DEV = 0.0104 (-0.0268)
REL STD DEV = 8.089 (54.235)

<<<<< CHANNEL 2 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 0.1340 (-0.0150)
Sample #2 = 0.1350 (-0.0030)
Sample #3 = 0.1320 (-0.0210)
Sample #4 = 0.1380 (-0.0150)
Avg % Abs = 0.1350 (-0.0130)
STD DEV = 0.0030 (-0.0092)
REL STD DEV = 2.222 (70.501)

Sol Value = 0.040 g/210L ***
Fit Value = 0.1905 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12712, Sum Io = 12985
<<<<< CHANNEL 1 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 0.7520 (-0.0160)
Sample #2 = 0.7490 (-0.0020)
Sample #3 = 0.7460 (-0.0170)
Sample #4 = 0.7410 (-0.0290)
Avg % Abs = 0.7453 (-0.0147)
STD DEV = 0.0040 (-0.0156)
REL STD DEV = 0.542 (106.576)

<<<<< CHANNEL 2 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 1.3810 (0.0120)
Sample #2 = 1.3180 (0.0000)
Sample #3 = 1.3200 (0.0120)
Sample #4 = 1.2950 (0.0160)
Avg % Abs = 1.3110 (-0.0093)
STD DEV = 0.0139 (0.0083)
REL STD DEV = 1.060 (89.214)

Sol Value = 0.100 g/210L ***
Fit Value = 0.4762 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12704, Sum Io = 12975
<<<<< CHANNEL 1 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 1.8000 (-0.0140)
Sample #2 = 1.8250 (-0.0070)
Sample #3 = 1.8180 (0.0080)
Sample #4 = 1.8560 (0.0030)
Avg % Abs = 1.8330 (-0.0013)
STD DEV = 0.0202 (0.0076)
REL STD DEV = 1.103 (572.822)

<<<<< CHANNEL 2 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 3.2930 (-0.0170)
Sample #2 = 3.3100 (0.0030)
Sample #3 = 3.3300 (-0.0020)
Sample #4 = 3.3720 (-0.0110)
Avg % Abs = 3.3373 (-0.0033)
STD DEV = 0.0316 (0.0071)
REL STD DEV = 0.948 (212.838)

Sol Value = 0.200 g/210L ***
Fit Value = 0.9524 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12700, Sum Io = 12976
<<<<< CHANNEL 1 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 3.5350 (-0.0110)
Sample #2 = 3.5670 (0.0070)
Sample #3 = 3.5870 (0.0030)
Sample #4 = 3.5580 (0.0100)
Avg % Abs = 3.5707 (0.0067)
STD DEV = 0.0148 (0.0035)
REL STD DEV = 0.416 (52.678)

***** AUTO CAL DATA *****
<<<<< CHANNEL 1 >>>>>
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.129
Std Dev = 0.01 Rel Std Dev = 8.09
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 0.745
Std Dev = 0.00 Rel Std Dev = 0.54
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 1.833
Std Dev = 0.02 Rel Std Dev = 1.10
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 3.571
Std Dev = 0.01 Rel Std Dev = 0.42
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 5.184
Std Dev = 0.02 Rel Std Dev = 0.39
Zero Order Coef = -240.30
First Order Coef = 2677.58
Second Order Coef = 22.68
Standard Deviation = 98.263832

<<<<< CHANNEL 2 >>>>>
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.135
Std Dev = 0.00 Rel Std Dev = 2.22
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 1.311
Std Dev = 0.01 Rel Std Dev = 1.06
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 3.337
Std Dev = 0.03 Rel Std Dev = 0.95
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 6.566
Std Dev = 0.02 Rel Std Dev = 0.31
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 9.489
Std Dev = 0.04 Rel Std Dev = 0.47
Zero Order Coef = -88.00
First Order Coef = 1407.72
Second Order Coef = 10.82
Standard Deviation = 99.214989

<<<<< CHANNEL 1 >>>>>
Sol Value = 0.300 g/210L ***
Fit Value = 1.4286 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12701, Sum Io = 12975
<<<<< CHANNEL 1 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 5.1820 (0.0030)
Sample #2 = 5.1650 (0.0310)
Sample #3 = 5.1820 (0.0270)
Sample #4 = 5.2050 (0.0240)
Avg % Abs = 5.1840 (0.0273)
STD DEV = 0.0201 (0.0035)
REL STD DEV = 0.387 (12.848)

<<<<< CHANNEL 2 >>>>>
Sol Value = 0.300 g/210L ***
Fit Value = 1.4286 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12700, Sum Io = 12976
<<<<< CHANNEL 1 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 9.4730 (-0.0250)
Sample #2 = 9.4540 (0.0020)
Sample #3 = 9.4740 (-0.0050)
Sample #4 = 9.5390 (-0.0060)
Avg % Abs = 9.4890 (-0.0030)
STD DEV = 0.0444 (0.0044)
REL STD DEV = 0.468 (145.297)

Solution Stats Quadratic Fit Chan 1
Act Fit Residual
g/210L g/210L g/210L
0.000 0.002 -0.0022
0.040 0.037 0.0029
0.100 0.100 0.0004
0.200 0.202 -0.0018
0.300 0.299 0.0008

Solution Stats Quadratic Fit Chan 2
Act Fit Residual
g/210L g/210L g/210L
0.000 0.002 -0.0021
0.040 0.037 0.0027
0.100 0.099 0.0007
0.200 0.202 -0.0021
0.300 0.299 0.0019

Sol Value = 0.080 g/210L ***
Fit Value = 0.3810 mg/l %%%
Samples Taken = 4, Discarded = 1
***** CHANNEL 1 *****
Sample #1 = 3437.00
Sample #2 = 3378.00
Sample #3 = 3375.00
Sample #4 = 3416.00
Average Result = 3389.6667
STD DEV = 22.8546
REL STD DEV = 0.674
***** CHANNEL 2 *****
Sample #1 = 3638.00
Sample #2 = 3637.00
Sample #3 = 3609.00
Sample #4 = 3633.00
Average Result = 3626.3333
STD DEV = 15.1438
REL STD DEV = 0.418
***** CHANNEL 1 *****
Dry Gas H2O Adjust Results *****
Barometric Pressure = 1019
3 um H2O Adjust (mg/l*10,000) = 420
9 um H2O Adjust (mg/l*10,000) = 183
***** AUTO CAL PASS *****

Optical penek
Calibration
adjustment #3
By: WICKP

Post-Cal Stability Checks

0.05g/210L 0.047 to 0.053	0.08g/210L 0.077 to 0.083	0.20g/210L 0.194 to 0.206	DGS 0.08g/210L 0.077 to 0.083																																																																																																																																																
<p>ALACHUA PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001720 12/30/2025 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>11:53</td></tr> <tr><td>Control Test</td><td>0.051</td><td>11:54</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:55</td></tr> <tr><td>Control Test</td><td>0.051</td><td>11:55</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:56</td></tr> <tr><td>Control Test</td><td>0.051</td><td>11:57</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:57</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0510</td><td></td></tr> <tr><td>Std Dev</td><td>0.0010</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0010</td><td></td></tr> </tbody> </table> <p>Operator's Signature: <i>Wenji</i></p>	Test	g/210L	Time	Air Blank	0.000	11:53	Control Test	0.051	11:54	Air Blank	0.000	11:55	Control Test	0.051	11:55	Air Blank	0.000	11:56	Control Test	0.051	11:57	Air Blank	0.000	11:57	Control Test Stats			Average	0.0510		Std Dev	0.0010		Rel Std Dev(%)	0.0010		<p>ALACHUA PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001720 12/30/2025 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>11:59</td></tr> <tr><td>Control Test</td><td>0.080</td><td>12:00</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:00</td></tr> <tr><td>Control Test</td><td>0.079</td><td>12:01</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:01</td></tr> <tr><td>Control Test</td><td>0.081</td><td>12:02</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:03</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0800</td><td></td></tr> <tr><td>Std Dev</td><td>0.0010</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.2500</td><td></td></tr> </tbody> </table> <p>Operator's Signature: <i>Wenji</i></p>	Test	g/210L	Time	Air Blank	0.000	11:59	Control Test	0.080	12:00	Air Blank	0.000	12:00	Control Test	0.079	12:01	Air Blank	0.000	12:01	Control Test	0.081	12:02	Air Blank	0.000	12:03	Control Test Stats			Average	0.0800		Std Dev	0.0010		Rel Std Dev(%)	1.2500		<p>ALACHUA PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001720 12/30/2025 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>12:04</td></tr> <tr><td>Control Test</td><td>0.201</td><td>12:04</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:05</td></tr> <tr><td>Control Test</td><td>0.205</td><td>12:06</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:06</td></tr> <tr><td>Control Test</td><td>0.204</td><td>12:07</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:07</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.2033</td><td></td></tr> <tr><td>Std Dev</td><td>0.0021</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.0238</td><td></td></tr> </tbody> </table> <p>Operator's Signature: <i>Wenji</i></p>	Test	g/210L	Time	Air Blank	0.000	12:04	Control Test	0.201	12:04	Air Blank	0.000	12:05	Control Test	0.205	12:06	Air Blank	0.000	12:06	Control Test	0.204	12:07	Air Blank	0.000	12:07	Control Test Stats			Average	0.2033		Std Dev	0.0021		Rel Std Dev(%)	1.0238		<p>ALACHUA PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001720 12/30/2025 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.009</td><td>12:10</td></tr> <tr><td>Control Test</td><td>0.080</td><td>12:11</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:11</td></tr> <tr><td>Control Test</td><td>0.080</td><td>12:12</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:12</td></tr> <tr><td>Control Test</td><td>0.080</td><td>12:12</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:13</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0800</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> <p>Operator's Signature: <i>Wenji</i></p>	Test	g/210L	Time	Air Blank	0.009	12:10	Control Test	0.080	12:11	Air Blank	0.000	12:11	Control Test	0.080	12:12	Air Blank	0.000	12:12	Control Test	0.080	12:12	Air Blank	0.000	12:13	Control Test Stats			Average	0.0800		Std Dev	0.0000		Rel Std Dev(%)	0.0000	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	11:53																																																																																																																																																	
Control Test	0.051	11:54																																																																																																																																																	
Air Blank	0.000	11:55																																																																																																																																																	
Control Test	0.051	11:55																																																																																																																																																	
Air Blank	0.000	11:56																																																																																																																																																	
Control Test	0.051	11:57																																																																																																																																																	
Air Blank	0.000	11:57																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0510																																																																																																																																																		
Std Dev	0.0010																																																																																																																																																		
Rel Std Dev(%)	0.0010																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	11:59																																																																																																																																																	
Control Test	0.080	12:00																																																																																																																																																	
Air Blank	0.000	12:00																																																																																																																																																	
Control Test	0.079	12:01																																																																																																																																																	
Air Blank	0.000	12:01																																																																																																																																																	
Control Test	0.081	12:02																																																																																																																																																	
Air Blank	0.000	12:03																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0800																																																																																																																																																		
Std Dev	0.0010																																																																																																																																																		
Rel Std Dev(%)	1.2500																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	12:04																																																																																																																																																	
Control Test	0.201	12:04																																																																																																																																																	
Air Blank	0.000	12:05																																																																																																																																																	
Control Test	0.205	12:06																																																																																																																																																	
Air Blank	0.000	12:06																																																																																																																																																	
Control Test	0.204	12:07																																																																																																																																																	
Air Blank	0.000	12:07																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.2033																																																																																																																																																		
Std Dev	0.0021																																																																																																																																																		
Rel Std Dev(%)	1.0238																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.009	12:10																																																																																																																																																	
Control Test	0.080	12:11																																																																																																																																																	
Air Blank	0.000	12:11																																																																																																																																																	
Control Test	0.080	12:12																																																																																																																																																	
Air Blank	0.000	12:12																																																																																																																																																	
Control Test	0.080	12:12																																																																																																																																																	
Air Blank	0.000	12:13																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0800																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: ALACHUA PD
Time of Inspection: 13:47

Date of Inspection: 12/31/2025

Serial Number: 80-001720
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202406K Exp: 06/19/2026	0.08g/210L Test (g/210L) Lot#:202406L Exp: 06/19/2026	0.20g/210L Test (g/210L) Lot#:202406N Exp: 06/20/2026	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG510701 Exp: 04/17/2027
0.000	0.053	0.081	0.198	0.080
0.000	0.053	0.081	0.202	0.081
0.000	0.053	0.081	0.203	0.081
0.000	0.053	0.082	0.203	0.080
0.000	0.053	0.081	0.202	0.081
0.000	0.053	0.082	0.204	0.080
0.000	0.054	0.082	0.202	0.081
0.000	0.053	0.082	0.202	0.081
0.000	0.054	0.081	0.204	0.080
0.000	0.053	0.081	0.204	0.081

Standard Deviations	0.0004	0.0005	0.0017	0.0005
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0007 Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



WEN-CHI K PIERSON
Signature and Printed Name

12/31/2025
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001720, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001720</u>	UNCERTAINTY* ±	
Owning Agency:	<u>ALACHUA PD</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>12/31/2025</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>13:47</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Wen-Chi
Pierson
Digitally signed by Wen-Chi Pierson
Date: 2025.12.31 13:57:34 -05'00'

12/31/2025

Date

WEN-CHI K PIERSON,
Department Inspector

FDLE/ATP Form 69 October 2024

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality