

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006638

| | | |
|---|------------------------|-------------------------------|
| Agency FL HIGHWAY PATROL | Date 01/08/2026 | Type Agency Inspection |
| Inspector MONTGOMERY, KENNETH, R | Time 20:11:03 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|---------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 403135736 |
| Expiration 06/24/2027 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 08/29/2027 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.200 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 01/08/2026 20:12

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006638

| | | |
|----------------------------------|------------------------|-------------------------------|
| Agency FL HIGHWAY PATROL | Date 02/06/2026 | Type Agency Inspection |
| Inspector SEARS, COLBY, D | Time 17:47:10 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|---------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202510E | 0.20 Lot # 202410F | 0.08 Gas Lot # 403135736 |
| Expiration 06/24/2027 | Expiration 10/28/2027 | Expiration 10/30/2026 | Expiration 08/29/2027 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.081 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.200 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.051 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.081 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.081 |
| | | 0.20 g/210L Test # 3 | 0.201 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 02/06/2026 17:49

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006638

| | | |
|----------------------------------|------------------------|-------------------------------|
| Agency FL HIGHWAY PATROL | Date 03/06/2026 | Type Agency Inspection |
| Inspector SEARS, COLBY, D | Time 18:00:26 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|---------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202510E | 0.20 Lot # 202410F | 0.08 Gas Lot # 403135736 |
| Expiration 06/24/2027 | Expiration 10/28/2027 | Expiration 10/30/2026 | Expiration 08/29/2027 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.081 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.200 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.081 |
| | | 0.20 g/210L Test # 3 | 0.201 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 03/06/2026 18:11

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006638

| | | |
|----------------------------------|------------------------|-------------------------------|
| Agency FL HIGHWAY PATROL | Date 04/04/2026 | Type Agency Inspection |
| Inspector SEARS, COLBY, D | Time 19:53:10 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202510E | 0.20 Lot # 202410F | 0.08 Gas Lot # 19825080A1 |
| Expiration 06/24/2027 | Expiration 10/28/2027 | Expiration 10/30/2026 | Expiration 09/05/2027 |

Number of Simulators 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.200 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 04/04/2026 20:01

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006692

| | | |
|----------------------------------|------------------------|-------------------------------|
| Agency HOLMES BEACH PD | Date 02/04/2026 | Type Agency Inspection |
| Inspector SPEARIN,KATIE,T | Time 08:33:29 | Software 8100.27 |

| | | | |
|-------------------|-------------------|-------------------|-----------------------|
| 0.05 Lot # | 0.08 Lot # | 0.20 Lot # | 0.08 Gas Lot # |
| Expiration | Expiration | Expiration | Expiration |

Number of Simulators **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|------|--------|
| Diagnostic Check (Pre-Inspection): OK | No | | |
| Date and/or Time Adjusted | No | | |
| Minimum Sample Volume Check: OK | | | |
| Barometric Pressure Sensor Check: OK | | | |
| Alcohol Free Subject Test: 0.000 | No | | |
| Mouth Alcohol Test: Slope Not Met | No | | |
| Interferent Detect Test: Interferent Detect | No | | |
| Diagnostic Check (Post-Inspection): OK | No | | |
| Alcohol Free Test: OK | No | | |
| 0.05 g/210L Test: OK | No | | |
| 0.08 g/210L Test: OK | No | | |
| 0.20 g/210L Test: OK | No | | |
| 0.08 g/210L Dry Gas Std Test: OK | No | | |

Remarks:

BYPASSED TO OPERATE INSTRUMENT.COMPLIANCE NOT DETERMINED.

Data Download Date/Time 02/06/2026 09:23

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006692

Agency HOLMES BEACH PD **Date** 02/06/2026 **Type** Department Inspection
Inspector SPEARIN,KATIE,T **Time** 09:22:45 **Software** 8100.27

0.05 Lot # 202406K **0.08 Lot #** 202406L **0.20 Lot #** 202406N **0.08 Gas Lot #** AG510701
Expiration 06/19/2026 **Expiration** 06/19/2026 **Expiration** 06/20/2026 **Expiration** 04/17/2027

Number of Simulators 5 **Average Standard Deviation** 0.0002 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.052 |
| Minimum Sample Volume Check: OK | Yes | 0.08 g/210L Test # 1 | 0.083 |
| Barometric Pressure Sensor Check: OK | Yes | 0.20 g/210L Test # 1 | 0.201 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.052 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.083 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.202 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.052 |
| | | 0.08 g/210L Test # 3 | 0.083 |
| | | 0.20 g/210L Test # 3 | 0.202 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |
| | | Alcohol Free Test # 4 | 0.000 |
| | | 0.05 g/210L Test # 4 | 0.052 |
| | | 0.08 g/210L Test # 4 | 0.082 |
| | | 0.20 g/210L Test # 4 | 0.202 |
| | | 0.08 g/210L Dry Gas Std Test # 4 | 0.081 |
| | | Interferent Detect Test # 4 | INT |
| | | Alcohol Free Test # 5 | 0.000 |
| | | 0.05 g/210L Test # 5 | 0.052 |
| | | 0.08 g/210L Test # 5 | 0.083 |
| | | 0.20 g/210L Test # 5 | 0.202 |
| | | 0.08 g/210L Dry Gas Std Test # 5 | 0.081 |
| | | Interferent Detect Test # 5 | INT |

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006692

| | |
|--|--------|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.052 |
| 0.08 g/210L Test # 6 | 0.083 |
| 0.20 g/210L Test # 6 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.082 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.052 |
| 0.08 g/210L Test # 7 | 0.083 |
| 0.20 g/210L Test # 7 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.081 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.052 |
| 0.08 g/210L Test # 8 | 0.083 |
| 0.20 g/210L Test # 8 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.081 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.052 |
| 0.08 g/210L Test # 9 | 0.083 |
| 0.20 g/210L Test # 9 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.081 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.052 |
| 0.08 g/210L Test # 10 | 0.083 |
| 0.20 g/210L Test # 10 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.081 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0000 |
| 0.08 g/210L Test (Std Dev) | 0.0003 |
| 0.20 g/210L Test (Std Dev) | 0.0003 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0004 |

Remarks:

Data Download Date/Time 02/06/2026 09:23

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006692

| | | |
|-----------------------------------|------------------------|-------------------------------|
| Agency HOLMES BEACH PD | Date 02/13/2026 | Type Agency Inspection |
| Inspector HURT,ALEXANDER,C | Time 14:53:20 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|---------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 403255643 |
| Expiration 06/24/2027 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 01/29/2028 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.051 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.081 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.201 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.051 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.081 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.052 |
| | | 0.08 g/210L Test # 3 | 0.081 |
| | | 0.20 g/210L Test # 3 | 0.201 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

INSPECTED POST REPAIRS EVERYTHING IS IN WORKING ORDER.

Data Download Date/Time 02/13/2026 16:44

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006692

| | | |
|------------------------------------|------------------------|-------------------------------|
| Agency HOLMES BEACH PD | Date 03/23/2026 | Type Agency Inspection |
| Inspector DESANTIS, ADAM, B | Time 09:34:02 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|---------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 403255643 |
| Expiration 06/24/2027 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 01/29/2028 |

Number of Simulators 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.051 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.081 |
| Barometric Pressure Sensor Check: OK | | 0.08 g/210L Test # (Repeat) 1 | 0.081 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.20 g/210L Test # 1 | 0.198 |
| Mouth Alcohol Test: Slope Not Met | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test: Interferent Detect | Yes | Interferent Detect Test # 1 | INT |
| Diagnostic Check (Post-Inspection): OK | Yes | Alcohol Free Test # 2 | 0.000 |
| Alcohol Free Test: OK | Yes | 0.05 g/210L Test # 2 | 0.051 |
| 0.05 g/210L Test: OK | Yes | 0.08 g/210L Test # 2 | 0.081 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Test # (Repeat) 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| | | Interferent Detect Test # 2 | INT |
| | | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.052 |
| | | 0.08 g/210L Test # 3 | 0.081 |
| | | 0.08 g/210L Test # (Repeat) 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

08: RFI Detect. INSTRUMENT INSPECTED FOR DAMAGES.

Data Download Date/Time 03/23/2026 09:44

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006692

| | | |
|-----------------------------------|------------------------|-------------------------------|
| Agency HOLMES BEACH PD | Date 04/29/2026 | Type Agency Inspection |
| Inspector HURT,ALEXANDER,C | Time 08:57:45 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|---------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 403255643 |
| Expiration 06/24/2027 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 01/29/2028 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.051 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.081 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.200 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.052 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.081 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.051 |
| | | 0.08 g/210L Test # 3 | 0.081 |
| | | 0.20 g/210L Test # 3 | 0.201 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

CHECKED FOR DAMAGE WITH NEGATIVE RESULTS

Data Download Date/Time 04/29/2026 08:58

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

| | | |
|-----------------------------------|------------------------|-------------------------------|
| Agency MONROE COUNTY SO | Date 01/21/2026 | Type Agency Inspection |
| Inspector ZITO, ANTHONY, P | Time 15:31:40 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|---------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 403207508 |
| Expiration 06/24/2027 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 12/02/2027 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.200 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.202 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.201 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

PK MAIN

Data Download Date/Time 01/21/2026 15:29

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

| | | | |
|----------------------------|-----------------|------------------------|--|
| Agency MONROE COUNTY SO | Date 02/17/2026 | Type Agency Inspection | |
| Inspector ZITO, ANTHONY, P | Time 13:09:15 | Software 8100.27 | |

| | | | |
|-----------------------|-----------------------|-----------------------|--------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 403207508 |
| Expiration 06/24/2027 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 12/02/2027 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

PK SPARE

Data Download Date/Time 02/17/2026 13:07

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

| | | |
|-----------------------------------|------------------------|-------------------------------|
| Agency MONROE COUNTY SO | Date 03/12/2026 | Type Agency Inspection |
| Inspector ZITO, ANTHONY, P | Time 02:48:06 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|---------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 403411042 |
| Expiration 06/24/2027 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 08/02/2028 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

PK SPARE

Data Download Date/Time 03/12/2026 14:46

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

| | | |
|-----------------------------------|------------------------|-------------------------------|
| Agency MONROE COUNTY SO | Date 03/25/2026 | Type Agency Inspection |
| Inspector ZITO, ANTHONY, P | Time 14:35:17 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|---------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 403411042 |
| Expiration 06/24/2027 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 08/02/2028 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.047 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.076 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.192 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.047 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.077 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.193 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.047 |
| | | 0.08 g/210L Test # 3 | 0.077 |
| | | 0.20 g/210L Test # 3 | 0.194 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

POST EVIDENTIARY AL. GOING TO FDLE

Data Download Date/Time 03/25/2026 15:11

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

Agency MONROE COUNTY SO Date 04/28/2026 Type Department Inspection
 Inspector GUTSCHOW,TAYLOR,D Time 11:19:56 Software 8100.27

0.05 Lot # 202406K 0.08 Lot # 202603B 0.20 Lot # 202406N 0.08 Gas Lot # AG526603
 Expiration 06/19/2026 Expiration 03/03/2028 Expiration 06/20/2026 Expiration 09/23/2027

Number of Simulators 5 Average Standard Deviation 0.0002 Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | Yes | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | Yes | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.077 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.077 |
| | | 0.20 g/210L Test # 3 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |
| | | Alcohol Free Test # 4 | 0.000 |
| | | 0.05 g/210L Test # 4 | 0.048 |
| | | 0.08 g/210L Test # 4 | 0.077 |
| | | 0.20 g/210L Test # 4 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 4 | 0.078 |
| | | Interferent Detect Test # 4 | INT |
| | | Alcohol Free Test # 5 | 0.000 |
| | | 0.05 g/210L Test # 5 | 0.048 |
| | | 0.08 g/210L Test # 5 | 0.077 |
| | | 0.20 g/210L Test # 5 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 5 | 0.078 |
| | | Interferent Detect Test # 5 | INT |

**Florida Department of Law Enforcement
Alcohol Testing Program**

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

| | |
|--|--------|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.048 |
| 0.08 g/210L Test # 6 | 0.077 |
| 0.20 g/210L Test # 6 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.078 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.048 |
| 0.08 g/210L Test # 7 | 0.077 |
| 0.20 g/210L Test # 7 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.077 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.048 |
| 0.08 g/210L Test # 8 | 0.077 |
| 0.20 g/210L Test # 8 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.077 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.048 |
| 0.08 g/210L Test # 9 | 0.078 |
| 0.20 g/210L Test # 9 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.077 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.048 |
| 0.08 g/210L Test # 10 | 0.077 |
| 0.20 g/210L Test # 10 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.078 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Test (Std Dev) | 0.0003 |
| 0.20 g/210L Test (Std Dev) | 0.0000 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0004 |

Remarks:

Data Download Date/Time 04/28/2026 11:33

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

| | | |
|------------------------------------|------------------------|-------------------------------|
| Agency MANATEE COUNTY SO | Date 01/12/2026 | Type Agency Inspection |
| Inspector COLEMAN, WILLIAM, | Time 09:44:57 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|---------------------------------|
| 0.05 Lot # 202402E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 403141809 |
| Expiration 02/27/2026 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 09/10/2027 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 01/12/2026 09:40

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

Agency MANATEE COUNTY SO **Date** 01/28/2026 **Type** Department Inspection
Inspector SPEARIN,KATIE,T **Time** 11:35:01 **Software** 8100.27

0.05 Lot # 202406K **0.08 Lot #** 202406L **0.20 Lot #** 202406N **0.08 Gas Lot #** AG510701
Expiration 06/19/2026 **Expiration** 06/19/2026 **Expiration** 06/20/2026 **Expiration** 04/17/2027

Number of Simulators 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | Yes | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | Yes | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |
| | | Alcohol Free Test # 4 | 0.000 |
| | | 0.05 g/210L Test # 4 | 0.049 |
| | | 0.08 g/210L Test # 4 | 0.080 |
| | | 0.20 g/210L Test # 4 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 4 | 0.080 |
| | | Interferent Detect Test # 4 | INT |
| | | Alcohol Free Test # 5 | 0.000 |
| | | 0.05 g/210L Test # 5 | 0.050 |
| | | 0.08 g/210L Test # 5 | 0.080 |
| | | 0.20 g/210L Test # 5 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 5 | 0.080 |
| | | Interferent Detect Test # 5 | INT |

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

| | |
|--|--------|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.049 |
| 0.08 g/210L Test # 6 | 0.080 |
| 0.20 g/210L Test # 6 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.080 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.050 |
| 0.08 g/210L Test # 7 | 0.080 |
| 0.20 g/210L Test # 7 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.081 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.050 |
| 0.08 g/210L Test # 8 | 0.080 |
| 0.20 g/210L Test # 8 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.080 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.050 |
| 0.08 g/210L Test # 9 | 0.080 |
| 0.20 g/210L Test # 9 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.080 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.050 |
| 0.08 g/210L Test # 10 | 0.080 |
| 0.20 g/210L Test # 10 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.080 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Test (Std Dev) | 0.0000 |
| 0.20 g/210L Test (Std Dev) | 0.0005 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0004 |

Remarks:

Data Download Date/Time 01/28/2026 11:43

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

| | | |
|------------------------------------|------------------------|-------------------------------|
| Agency MANATEE COUNTY SO | Date 02/16/2026 | Type Agency Inspection |
| Inspector COLEMAN, WILLIAM, | Time 09:34:43 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|---------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202510E | 0.20 Lot # 202410F | 0.08 Gas Lot # 403141809 |
| Expiration 06/24/2027 | Expiration 10/28/2027 | Expiration 10/30/2026 | Expiration 09/10/2027 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 02/16/2026 09:31

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

| | | |
|------------------------------------|------------------------|-------------------------------|
| Agency MANATEE COUNTY SO | Date 03/16/2026 | Type Agency Inspection |
| Inspector COLEMAN, WILLIAM, | Time 10:15:58 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|---------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202510E | 0.20 Lot # 202410F | 0.08 Gas Lot # 403141809 |
| Expiration 06/24/2027 | Expiration 10/28/2027 | Expiration 10/30/2026 | Expiration 09/10/2027 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.051 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 03/16/2026 10:12

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

| | | |
|--------------------------------------|------------------------|-------------------------------|
| Agency MANATEE COUNTY SO | Date 04/22/2026 | Type Agency Inspection |
| Inspector NOVAK, CHRISTOPHER, | Time 10:22:33 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|---------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202510E | 0.20 Lot # 202410F | 0.08 Gas Lot # 403419188 |
| Expiration 06/24/2027 | Expiration 10/28/2027 | Expiration 10/30/2026 | Expiration 08/13/2028 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.081 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.081 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.051 |
| | | 0.08 g/210L Test # 3 | 0.081 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 04/22/2026 10:18

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006736

| | | | |
|---------------------------------|------------------------------|-------------------------------|---------------------------------|
| Agency FL HIGHWAY PATROL | Date 01/30/2026 | Type Agency Inspection | |
| Inspector KEMP,TREVOR,D | Time 17:06:02 | Software 8100.27 | |
| 0.05 Lot # 202506E | 0.08 Lot # 202510E | 0.20 Lot # 202410F | 0.08 Gas Lot # 403135737 |
| Expiration 06/24/2027 | Expiration 10/28/2027 | Expiration 10/30/2026 | Expiration 08/30/2027 |
| Number of Simulators 5 | | Compliance Yes | |

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.196 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.196 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

N/A

Data Download Date/Time 01/30/2026 18:10

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006736

| | | |
|---------------------------------|------------------------|-------------------------------|
| Agency FL HIGHWAY PATROL | Date 02/22/2026 | Type Agency Inspection |
| Inspector KEMP,TREVOR,D | Time 16:36:13 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|---------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202510E | 0.20 Lot # 202410F | 0.08 Gas Lot # 403135737 |
| Expiration 06/24/2027 | Expiration 10/28/2027 | Expiration 10/30/2026 | Expiration 08/30/2027 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | Yes | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Time-Date changed.

Data Download Date/Time 02/22/2026 17:16

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006736

| | | |
|---------------------------------|------------------------|-------------------------------|
| Agency FL HIGHWAY PATROL | Date 03/20/2026 | Type Agency Inspection |
| Inspector KEMP,TREVOR,D | Time 13:14:42 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|---------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202510E | 0.20 Lot # 202410F | 0.08 Gas Lot # 403135737 |
| Expiration 06/24/2027 | Expiration 10/28/2027 | Expiration 10/30/2026 | Expiration 08/30/2027 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | Yes | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.196 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Time-Date changed.

Data Download Date/Time 03/23/2026 16:17

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006736

| | | | |
|---------------------------------|------------------------------|-------------------------------|---------------------------------|
| Agency FL HIGHWAY PATROL | Date 04/26/2026 | Type Agency Inspection | |
| Inspector KEMP,TREVOR,D | Time 19:04:36 | Software 8100.27 | |
| 0.05 Lot # 202506E | 0.08 Lot # 202510E | 0.20 Lot # 202410F | 0.08 Gas Lot # 403268835 |
| Expiration 06/24/2027 | Expiration 10/28/2027 | Expiration 10/30/2026 | Expiration 02/13/2028 |
| Number of Simulators 5 | | Compliance Yes | |

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

NA

Data Download Date/Time 04/26/2026 19:43

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006758

| | | |
|-----------------------------------|------------------------|-------------------------------|
| Agency FL HIGHWAY PATROL | Date 01/12/2026 | Type Agency Inspection |
| Inspector LEDFORD,TAYLOR,M | Time 09:47:33 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|---------------------------------|
| 0.05 Lot # 202402E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 403411042 |
| Expiration 02/27/2026 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 08/02/2028 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 01/12/2026 09:45

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006758

| | | |
|----------------------------------|------------------------|-------------------------------|
| Agency FL HIGHWAY PATROL | Date 02/16/2026 | Type Agency Inspection |
| Inspector LEDFORD,TAYLOR, | Time 14:11:39 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|---------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202510E | 0.20 Lot # 202410F | 0.08 Gas Lot # 403411042 |
| Expiration 06/24/2027 | Expiration 10/28/2027 | Expiration 10/30/2026 | Expiration 08/02/2028 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 02/16/2026 14:09

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006758

| | | |
|----------------------------------|------------------------|-------------------------------|
| Agency FL HIGHWAY PATROL | Date 03/16/2026 | Type Agency Inspection |
| Inspector LEDFORD,TAYLOR, | Time 09:00:54 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|---------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202510E | 0.20 Lot # 202410F | 0.08 Gas Lot # 403411042 |
| Expiration 06/24/2027 | Expiration 10/28/2027 | Expiration 10/30/2026 | Expiration 08/02/2028 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 03/16/2026 08:58

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006758

| | | |
|----------------------------------|------------------------|-------------------------------|
| Agency FL HIGHWAY PATROL | Date 04/22/2026 | Type Agency Inspection |
| Inspector LEDFORD,TAYLOR, | Time 09:19:32 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|---------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202510E | 0.20 Lot # 202410F | 0.08 Gas Lot # 403411042 |
| Expiration 06/24/2027 | Expiration 10/28/2027 | Expiration 10/30/2026 | Expiration 08/02/2028 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 04/22/2026 09:16

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006759

| | | |
|------------------------------------|------------------------|-------------------------------|
| Agency KENNEDY SPACE CENTER | Date 01/28/2026 | Type Agency Inspection |
| Inspector CLARK,JOSIE,P | Time 02:57:49 | Software 8100.27 |

| | | | |
|-------------------|-------------------|-------------------|-----------------------|
| 0.05 Lot # | 0.08 Lot # | 0.20 Lot # | 0.08 Gas Lot # |
| Expiration | Expiration | Expiration | Expiration |

Number of Simulators 5 **Compliance** No

| Check or Test | Result | Test | Result |
|---|--------|------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | | |
| Date and/or Time Adjusted | No | | |
| Minimum Sample Volume Check: OK | | | |
| Barometric Pressure Sensor Check: OK | | | |
| Alcohol Free Subject Test: 0.000 | Yes | | |
| Mouth Alcohol Test: Slope Not Met | Yes | | |
| Interferent Detect Test: Interferent Detect | No | | |
| Diagnostic Check (Post-Inspection): OK | No | | |
| Alcohol Free Test: OK | No | | |
| 0.05 g/210L Test: OK | No | | |
| 0.08 g/210L Test: OK | No | | |
| 0.20 g/210L Test: OK | No | | |
| 0.08 g/210L Dry Gas Std Test: OK | No | | |

Remarks:

00: Ambient Fail. Non-compliance: .

Data Download Date/Time 01/28/2026 22:18

Florida Department of Law Enforcement
Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006759

| | | |
|------------------------------------|------------------------|-------------------------------|
| Agency KENNEDY SPACE CENTER | Date 01/28/2026 | Type Agency Inspection |
| Inspector CLARK,JOSIE,P | Time 03:05:09 | Software 8100.27 |

| | | | |
|-------------------|-------------------|-------------------|-----------------------|
| 0.05 Lot # | 0.08 Lot # | 0.20 Lot # | 0.08 Gas Lot # |
| Expiration | Expiration | Expiration | Expiration |

Number of Simulators 5

Compliance No

| Check or Test | Result | Test | Result |
|---|--------|------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | | |
| Date and/or Time Adjusted | No | | |
| Minimum Sample Volume Check: OK | | | |
| Barometric Pressure Sensor Check: OK | | | |
| Alcohol Free Subject Test: 0.000 | Yes | | |
| Mouth Alcohol Test: Slope Not Met | Yes | | |
| Interferent Detect Test: Interferent Detect | No | | |
| Diagnostic Check (Post-Inspection): OK | No | | |
| Alcohol Free Test: OK | No | | |
| 0.05 g/210L Test: OK | No | | |
| 0.08 g/210L Test: OK | No | | |
| 0.20 g/210L Test: OK | No | | |
| 0.08 g/210L Dry Gas Std Test: OK | No | | |

Remarks:

00: Ambient Fail. Non-compliance: .

Data Download Date/Time 01/28/2026 22:18

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006759

Agency KENNEDY SPACE CENTER **Date** 01/28/2026 **Type** Agency Inspection
Inspector ETTMAN,KRISTOFER,C **Time** 03:13:51 **Software** 8100.27

0.05 Lot # **0.08 Lot #** **0.20 Lot #** **0.08 Gas Lot #**
Expiration **Expiration** **Expiration** **Expiration**

Number of Simulators 5 **Compliance** No

| Check or Test | Result | Test | Result |
|---|--------|------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | | |
| Date and/or Time Adjusted | No | | |
| Minimum Sample Volume Check: OK | | | |
| Barometric Pressure Sensor Check: OK | | | |
| Alcohol Free Subject Test: 0.000 | No | | |
| Mouth Alcohol Test: Slope Not Met | No | | |
| Interferent Detect Test: Interferent Detect | No | | |
| Diagnostic Check (Post-Inspection): OK | No | | |
| Alcohol Free Test: OK | No | | |
| 0.05 g/210L Test: OK | No | | |
| 0.08 g/210L Test: OK | No | | |
| 0.20 g/210L Test: OK | No | | |
| 0.08 g/210L Dry Gas Std Test: OK | No | | |

Remarks:

A F / M A: Ambient Fail Ambient Fail Ambient Fail. Non-compliance: .

Data Download Date/Time 01/28/2026 22:18

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006759

| | | |
|---------------------------------------|------------------------|-------------------------------|
| Agency KENNEDY SPACE CENTER | Date 01/28/2026 | Type Agency Inspection |
| Inspector ETTMAN, KRISTOFER, C | Time 04:01:34 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 08425080A1 |
| Expiration 06/24/2027 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 05/05/2027 |

Number of Simulators 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | INT |
| Date and/or Time Adjusted | No | Alcohol Free Test # (Repeat) 1 | 0.000 |
| Minimum Sample Volume Check: OK | | 0.05 g/210L Test # 1 | 0.048 |
| Barometric Pressure Sensor Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Test # (Repeat) 1 | 0.078 |
| Mouth Alcohol Test: Slope Not Met | Yes | 0.20 g/210L Test # 1 | 0.193 |
| Interferent Detect Test: Interferent Detect | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Diagnostic Check (Post-Inspection): OK | Yes | Interferent Detect Test # 1 | INT |
| Alcohol Free Test: OK | Yes | Alcohol Free Test # (Repeat) 2 | 0.000 |
| 0.05 g/210L Test: OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | 0.08 g/210L Test # (Repeat) 2 | 0.078 |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | 0.20 g/210L Test # 2 | 0.194 |
| | | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| | | Interferent Detect Test # 2 | INT |
| | | Alcohol Free Test # (Repeat) 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | INT |
| | | 0.08 g/210L Test # (Repeat) 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.195 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

00: Interferent Detect. 08: Interferent Detect.

Data Download Date/Time 01/28/2026 22:18

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006759

| | | |
|-------------------------------------|------------------------|-------------------------------|
| Agency KENNEDY SPACE CENTER | Date 02/24/2026 | Type Agency Inspection |
| Inspector SAVARD, CHARLES, M | Time 12:37:25 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 08425080A1 |
| Expiration 06/24/2027 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 05/05/2027 |

Number of Simulators 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.193 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.077 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.077 |
| | | 0.20 g/210L Test # 3 | 0.194 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 02/24/2026 13:39

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006760

| | | |
|---------------------------------|------------------------|-------------------------------|
| Agency FL HIGHWAY PATROL | Date 01/09/2026 | Type Agency Inspection |
| Inspector RUBEL,NEIL,F | Time 11:56:14 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|---------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202510E | 0.20 Lot # 202410F | 0.08 Gas Lot # 403308953 |
| Expiration 06/24/2027 | Expiration 10/28/2027 | Expiration 10/30/2026 | Expiration 04/01/2028 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.201 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.203 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.203 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

N/A

Data Download Date/Time 01/09/2026 12:55

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006760

| | | |
|---------------------------------|------------------------|-------------------------------|
| Agency FL HIGHWAY PATROL | Date 02/10/2026 | Type Agency Inspection |
| Inspector ABRAHAM,ERIC,D | Time 09:28:13 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|---------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202510E | 0.20 Lot # 202410F | 0.08 Gas Lot # 403308953 |
| Expiration 06/24/2027 | Expiration 10/28/2027 | Expiration 10/30/2026 | Expiration 04/01/2028 |

Number of Simulators 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.200 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.202 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 02/10/2026 10:31

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006760

| | | |
|---------------------------------|------------------------|-------------------------------|
| Agency FL HIGHWAY PATROL | Date 03/11/2026 | Type Agency Inspection |
| Inspector RUBEL,NEIL,F | Time 13:38:20 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|---------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202510E | 0.20 Lot # 202410F | 0.08 Gas Lot # 403308953 |
| Expiration 06/24/2027 | Expiration 10/28/2027 | Expiration 10/30/2026 | Expiration 04/01/2028 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.081 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

A F / M A: RFI Detect. WIRELESS TRANSMITTER ON-TURNED OFF

Data Download Date/Time 03/11/2026 15:39

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006760

| | | |
|---------------------------------------|------------------------|-------------------------------|
| Agency FL HIGHWAY PATROL | Date 04/04/2026 | Type Agency Inspection |
| Inspector MCDOWELL, MICHAEL, J | Time 08:53:30 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|---------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202510E | 0.20 Lot # 202410F | 0.08 Gas Lot # 403308953 |
| Expiration 06/24/2027 | Expiration 10/28/2027 | Expiration 10/30/2026 | Expiration 04/01/2028 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

N/A

Data Download Date/Time 04/04/2026 10:04

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006761

| | | |
|-----------------------------------|------------------------|-------------------------------|
| Agency FL HIGHWAY PATROL | Date 01/12/2026 | Type Agency Inspection |
| Inspector LEDFORD,TAYLOR,M | Time 09:51:51 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|---------------------------------|
| 0.05 Lot # 202402E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 403141809 |
| Expiration 02/27/2026 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 09/10/2027 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 01/12/2026 09:51

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006761

| | | |
|----------------------------------|------------------------|-------------------------------|
| Agency FL HIGHWAY PATROL | Date 02/16/2026 | Type Agency Inspection |
| Inspector LEDFORD,TAYLOR, | Time 14:04:38 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|---------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202510E | 0.20 Lot # 202410F | 0.08 Gas Lot # 403141809 |
| Expiration 06/24/2027 | Expiration 10/28/2027 | Expiration 10/30/2026 | Expiration 09/10/2027 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 02/16/2026 14:01

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006761

| | | |
|----------------------------------|------------------------|-------------------------------|
| Agency FL HIGHWAY PATROL | Date 03/16/2026 | Type Agency Inspection |
| Inspector LEDFORD,TAYLOR, | Time 08:54:19 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|---------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202510E | 0.20 Lot # 202410F | 0.08 Gas Lot # 403141809 |
| Expiration 06/24/2027 | Expiration 10/28/2027 | Expiration 10/30/2026 | Expiration 09/10/2027 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.200 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 03/16/2026 08:50

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006761

| | | |
|----------------------------------|------------------------|-------------------------------|
| Agency FL HIGHWAY PATROL | Date 04/22/2026 | Type Agency Inspection |
| Inspector LEDFORD,TAYLOR, | Time 09:23:11 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|---------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202510E | 0.20 Lot # 202410F | 0.08 Gas Lot # 403141809 |
| Expiration 06/24/2027 | Expiration 10/28/2027 | Expiration 10/30/2026 | Expiration 09/10/2027 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.081 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 04/22/2026 09:19

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006762

| | | |
|---------------------------------|------------------------|-------------------------------|
| Agency FL HIGHWAY PATROL | Date 01/23/2026 | Type Agency Inspection |
| Inspector LIMB,HUNTER, | Time 19:01:54 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|---------------------------------|
| 0.05 Lot # 202402E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 403008479 |
| Expiration 02/27/2026 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 03/28/2027 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 01/23/2026 19:00

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006762

| | | |
|---------------------------------|------------------------|-------------------------------|
| Agency FL HIGHWAY PATROL | Date 02/26/2026 | Type Agency Inspection |
| Inspector LIMB,HUNTER, | Time 18:29:18 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|---------------------------------|
| 0.05 Lot # 202402E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 403008479 |
| Expiration 02/27/2026 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 03/28/2027 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 02/26/2026 18:27

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

| | | |
|---|------------------------|-------------------------------|
| Agency FL HIGHWAY PATROL | Date 01/08/2026 | Type Agency Inspection |
| Inspector MONTGOMERY, KENNETH, R | Time 20:03:18 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|---------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 403135736 |
| Expiration 06/24/2027 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 08/29/2027 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.201 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.201 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 01/08/2026 20:06

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

| | | |
|---|------------------------|-------------------------------|
| Agency FL HIGHWAY PATROL | Date 02/04/2026 | Type Agency Inspection |
| Inspector MONTGOMERY, KENNETH, R | Time 21:05:49 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|---------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202510E | 0.20 Lot # 202410F | 0.08 Gas Lot # 403135736 |
| Expiration 06/24/2027 | Expiration 10/28/2027 | Expiration 10/30/2026 | Expiration 08/29/2027 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.05 g/210L Test # (Repeat) 1 | 0.050 |
| Barometric Pressure Sensor Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.20 g/210L Test # 1 | 0.199 |
| Mouth Alcohol Test: Slope Not Met | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test: Interferent Detect | Yes | Interferent Detect Test # 1 | INT |
| Diagnostic Check (Post-Inspection): OK | Yes | Alcohol Free Test # 2 | 0.000 |
| Alcohol Free Test: OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| 0.05 g/210L Test: OK | Yes | 0.05 g/210L Test # (Repeat) 2 | 0.050 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| | | Interferent Detect Test # 2 | INT |
| | | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # (Repeat) 3 | 0.051 |
| | | 0.08 g/210L Test # 3 | 0.081 |
| | | 0.20 g/210L Test # 3 | 0.201 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

05:UNKNOWN RFI.

Data Download Date/Time 02/04/2026 21:07

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

| | | |
|---|------------------------|-------------------------------|
| Agency FL HIGHWAY PATROL | Date 03/05/2026 | Type Agency Inspection |
| Inspector MONTGOMERY, KENNETH, R | Time 16:57:48 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|---------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202510E | 0.20 Lot # 202410F | 0.08 Gas Lot # 403135736 |
| Expiration 06/24/2027 | Expiration 10/28/2027 | Expiration 10/30/2026 | Expiration 08/29/2027 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.193 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 03/05/2026 16:58

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

| | | |
|--|------------------------|-------------------------------|
| Agency FL HIGHWAY PATROL | Date 04/06/2026 | Type Agency Inspection |
| Inspector MONTOMERY, KENNETH, R | Time 21:00:27 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|---------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202510E | 0.20 Lot # 202410F | 0.08 Gas Lot # 403135736 |
| Expiration 06/24/2027 | Expiration 10/28/2027 | Expiration 10/30/2026 | Expiration 08/29/2027 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.202 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.201 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 04/07/2026 19:32

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006765

| | | |
|---|------------------------|-------------------------------|
| Agency FL HIGHWAY PATROL | Date 02/11/2026 | Type Agency Inspection |
| Inspector HIGGINBOTHAM, LEANDRA, | Time 10:47:20 | Software 8100.27 |

| | | | |
|-------------------|-------------------|-------------------|-----------------------|
| 0.05 Lot # | 0.08 Lot # | 0.20 Lot # | 0.08 Gas Lot # |
| Expiration | Expiration | Expiration | Expiration |

Number of Simulators **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|------|--------|
| Diagnostic Check (Pre-Inspection): OK | No | | |
| Date and/or Time Adjusted | No | | |
| Minimum Sample Volume Check: OK | | | |
| Barometric Pressure Sensor Check: OK | | | |
| Alcohol Free Subject Test: 0.000 | No | | |
| Mouth Alcohol Test: Slope Not Met | No | | |
| Interferent Detect Test: Interferent Detect | No | | |
| Diagnostic Check (Post-Inspection): OK | No | | |
| Alcohol Free Test: OK | No | | |
| 0.05 g/210L Test: OK | No | | |
| 0.08 g/210L Test: OK | No | | |
| 0.20 g/210L Test: OK | No | | |
| 0.08 g/210L Dry Gas Std Test: OK | No | | |

Remarks:

BYPASS AI FOR OPERATION, COMPLIANCE UNDETERMINED

Data Download Date/Time 03/04/2026 09:44

Florida Department of Law Enforcement
Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006766

| | | |
|------------------------------------|------------------------|-------------------------------|
| Agency TOXL | Date 02/04/2026 | Type Agency Inspection |
| Inspector GUTSCHOW,TAYLOR,D | Time 12:33:24 | Software 8100.27 |

| | | | |
|-------------------|-------------------|-------------------|-----------------------|
| 0.05 Lot # | 0.08 Lot # | 0.20 Lot # | 0.08 Gas Lot # |
| Expiration | Expiration | Expiration | Expiration |

Number of Simulators **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|------|--------|
| Diagnostic Check (Pre-Inspection): OK | No | | |
| Date and/or Time Adjusted | No | | |
| Minimum Sample Volume Check: OK | | | |
| Barometric Pressure Sensor Check: OK | | | |
| Alcohol Free Subject Test: 0.000 | No | | |
| Mouth Alcohol Test: Slope Not Met | No | | |
| Interferent Detect Test: Interferent Detect | No | | |
| Diagnostic Check (Post-Inspection): OK | No | | |
| Alcohol Free Test: OK | No | | |
| 0.05 g/210L Test: OK | No | | |
| 0.08 g/210L Test: OK | No | | |
| 0.20 g/210L Test: OK | No | | |
| 0.08 g/210L Dry Gas Std Test: OK | No | | |

Remarks:

AI NOT CONDUCTED. COMPLIANCE NOT DETERMINED.

Data Download Date/Time 02/05/2026 12:36

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006766

Agency FL HIGHWAY PATROL **Date** 02/05/2026 **Type** Department Inspection
Inspector GUTSCHOW,TAYLOR,D **Time** 12:00:46 **Software** 8100.27

0.05 Lot # 202406K **0.08 Lot #** 202406L **0.20 Lot #** 202406N **0.08 Gas Lot #** AG429602
Expiration 06/19/2026 **Expiration** 06/19/2026 **Expiration** 06/20/2026 **Expiration** 10/22/2026

Number of Simulators 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.047 |
| Minimum Sample Volume Check: OK | Yes | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | Yes | 0.20 g/210L Test # 1 | 0.200 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.077 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.047 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.076 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.047 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.077 |
| | | Interferent Detect Test # 3 | INT |
| | | Alcohol Free Test # 4 | 0.000 |
| | | 0.05 g/210L Test # 4 | 0.047 |
| | | 0.08 g/210L Test # 4 | 0.079 |
| | | 0.20 g/210L Test # 4 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 4 | 0.077 |
| | | Interferent Detect Test # 4 | INT |
| | | Alcohol Free Test # 5 | 0.000 |
| | | 0.05 g/210L Test # 5 | 0.047 |
| | | 0.08 g/210L Test # 5 | 0.079 |
| | | 0.20 g/210L Test # 5 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 5 | 0.077 |
| | | Interferent Detect Test # 5 | INT |

Florida Department of Law Enforcement
Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006766

| | |
|--|--------|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.048 |
| 0.08 g/210L Test # 6 | 0.079 |
| 0.20 g/210L Test # 6 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.077 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.047 |
| 0.08 g/210L Test # 7 | 0.079 |
| 0.20 g/210L Test # 7 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.077 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.048 |
| 0.08 g/210L Test # 8 | 0.079 |
| 0.20 g/210L Test # 8 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.077 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.047 |
| 0.08 g/210L Test # 9 | 0.079 |
| 0.20 g/210L Test # 9 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.077 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.047 |
| 0.08 g/210L Test # 10 | 0.079 |
| 0.20 g/210L Test # 10 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.077 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Test (Std Dev) | 0.0004 |
| 0.20 g/210L Test (Std Dev) | 0.0003 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0003 |

Remarks:

Data Download Date/Time 02/05/2026 12:36

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006767

| | | |
|-----------------------------------|------------------------|-------------------------------|
| Agency FL HIGHWAY PATROL | Date 02/16/2026 | Type Agency Inspection |
| Inspector LEDFORD,TAYLOR,M | Time 14:14:05 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|---------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202510E | 0.20 Lot # 202410F | 0.08 Gas Lot # 403419188 |
| Expiration 06/24/2027 | Expiration 10/28/2027 | Expiration 10/30/2026 | Expiration 08/13/2028 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

A F / M A: .

Data Download Date/Time 02/16/2026 14:13

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006767

| | | |
|---------------------------------|------------------------|-------------------------------|
| Agency FL HIGHWAY PATROL | Date 03/11/2026 | Type Agency Inspection |
| Inspector BLAKE,DAVID,A | Time 20:56:29 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 3024034191 |
| Expiration 06/24/2027 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 08/13/2028 |

Number of Simulators 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.045 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.194 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.047 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.076 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.047 |
| | | 0.08 g/210L Test # 3 | 0.076 |
| | | 0.20 g/210L Test # 3 | 0.194 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

INTOX FREE OF CRACKS OR DAMAGE

Data Download Date/Time 03/19/2026 10:30

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006767

Agency FL HIGHWAY PATROL **Date** 03/19/2026 **Type** Department Inspection
Inspector GUTSCHOW,TAYLOR,D **Time** 13:51:50 **Software** 8100.27

0.05 Lot # 202406K **0.08 Lot #** 202406L **0.20 Lot #** 202406N **0.08 Gas Lot #** AG429602
Expiration 06/19/2026 **Expiration** 06/19/2026 **Expiration** 06/20/2026 **Expiration** 10/22/2026

Number of Simulators 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | Yes | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | Yes | 0.20 g/210L Test # 1 | 0.196 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.077 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.196 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.077 |
| | | Interferent Detect Test # 3 | INT |
| | | Alcohol Free Test # 4 | 0.000 |
| | | 0.05 g/210L Test # 4 | 0.048 |
| | | 0.08 g/210L Test # 4 | 0.078 |
| | | 0.20 g/210L Test # 4 | 0.196 |
| | | 0.08 g/210L Dry Gas Std Test # 4 | 0.078 |
| | | Interferent Detect Test # 4 | INT |
| | | Alcohol Free Test # 5 | 0.000 |
| | | 0.05 g/210L Test # 5 | 0.048 |
| | | 0.08 g/210L Test # 5 | 0.079 |
| | | 0.20 g/210L Test # 5 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 5 | 0.078 |
| | | Interferent Detect Test # 5 | INT |

**Florida Department of Law Enforcement
Alcohol Testing Program**

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006767

| | |
|--|--------|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.048 |
| 0.08 g/210L Test # 6 | 0.078 |
| 0.20 g/210L Test # 6 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.078 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.048 |
| 0.08 g/210L Test # 7 | 0.078 |
| 0.20 g/210L Test # 7 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.078 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.048 |
| 0.08 g/210L Test # 8 | 0.079 |
| 0.20 g/210L Test # 8 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.078 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.048 |
| 0.08 g/210L Test # 9 | 0.079 |
| 0.20 g/210L Test # 9 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.078 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.048 |
| 0.08 g/210L Test # 10 | 0.078 |
| 0.20 g/210L Test # 10 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.078 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0000 |
| 0.08 g/210L Test (Std Dev) | 0.0005 |
| 0.20 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0004 |

Remarks:

Data Download Date/Time 03/19/2026 14:39

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006767

| | | |
|---------------------------------|------------------------|-------------------------------|
| Agency FL HIGHWAY PATROL | Date 04/02/2026 | Type Agency Inspection |
| Inspector BLAKE,DAVID,A | Time 01:28:05 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 3024034191 |
| Expiration 06/24/2027 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 08/13/2028 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # (Repeat) 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.046 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.076 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.192 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # (Repeat) 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.046 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.077 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.193 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # (Repeat) 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.047 |
| | | 0.08 g/210L Test # 3 | 0.076 |
| | | 0.20 g/210L Test # 3 | 0.194 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

00: RFI Detect.

Data Download Date/Time 04/03/2026 18:42

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006768

| | | |
|---------------------------------|------------------------|-------------------------------|
| Agency FL HIGHWAY PATROL | Date 01/19/2026 | Type Agency Inspection |
| Inspector RUBEL, NEIL, F | Time 10:28:43 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|---------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202510E | 0.20 Lot # 202410F | 0.08 Gas Lot # 403056709 |
| Expiration 06/24/2027 | Expiration 10/28/2027 | Expiration 10/30/2026 | Expiration 05/28/2027 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.051 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.081 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.201 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.051 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.081 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.202 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.051 |
| | | 0.08 g/210L Test # 3 | 0.081 |
| | | 0.20 g/210L Test # 3 | 0.201 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

N/A

Data Download Date/Time 01/19/2026 11:29

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006768

| | | |
|---------------------------------|------------------------|-------------------------------|
| Agency FL HIGHWAY PATROL | Date 02/06/2026 | Type Agency Inspection |
| Inspector RUBEL,NEIL,F | Time 08:41:26 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|---------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202510E | 0.20 Lot # 202410F | 0.08 Gas Lot # 403056709 |
| Expiration 06/24/2027 | Expiration 10/28/2027 | Expiration 10/30/2026 | Expiration 05/28/2027 |

Number of Simulators 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.200 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.051 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.201 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

N/A

Data Download Date/Time 02/06/2026 10:01

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006768

| | | |
|---------------------------------|------------------------|-------------------------------|
| Agency FL HIGHWAY PATROL | Date 03/16/2026 | Type Agency Inspection |
| Inspector ABRAHAM,ERIC,D | Time 03:59:07 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|---------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202510E | 0.20 Lot # 202410F | 0.08 Gas Lot # 403056709 |
| Expiration 06/24/2027 | Expiration 10/28/2027 | Expiration 10/30/2026 | Expiration 05/28/2027 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.081 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.200 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.082 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.082 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.077 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 03/16/2026 04:55

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006768

| | | |
|---------------------------------------|------------------------|-------------------------------|
| Agency FL HIGHWAY PATROL | Date 04/04/2026 | Type Agency Inspection |
| Inspector MCDOWELL, MICHAEL, J | Time 09:22:13 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|---------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202510E | 0.20 Lot # 202410F | 0.08 Gas Lot # 403056709 |
| Expiration 06/24/2027 | Expiration 10/28/2027 | Expiration 10/30/2026 | Expiration 05/28/2027 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.076 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.193 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.047 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

A F / M A: .MOUTH ALCOHOL DID NOT REACH SLOPE NOT MET

Data Download Date/Time 04/04/2026 10:18

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006770

| | | |
|---------------------------------|------------------------|-------------------------------|
| Agency FL HIGHWAY PATROL | Date 01/22/2026 | Type Agency Inspection |
| Inspector BLAKE,DAVID,A | Time 00:45:44 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 3024028434 |
| Expiration 06/24/2027 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 09/08/2026 |

Number of Simulators 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.196 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

INTOX FREE OF CRACKS OR DAMAGE

Data Download Date/Time 01/22/2026 01:58

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006770

| | | |
|----------------------------------|------------------------|-------------------------------|
| Agency FL HIGHWAY PATROL | Date 02/18/2026 | Type Agency Inspection |
| Inspector BLAKE, DAVID, A | Time 22:58:06 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 3024028434 |
| Expiration 06/24/2027 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 09/08/2026 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.047 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.000 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.20 g/210L Test # (Repeat) 1 | 0.196 |
| Mouth Alcohol Test: Slope Not Met | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test: Interferent Detect | Yes | Interferent Detect Test # 1 | INT |
| Diagnostic Check (Post-Inspection): OK | Yes | Alcohol Free Test # 2 | 0.000 |
| Alcohol Free Test: OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| 0.05 g/210L Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.08 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.000 |
| 0.20 g/210L Test: OK | Yes | 0.20 g/210L Test # (Repeat) 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| | | Interferent Detect Test # 2 | INT |
| | | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.000 |
| | | 0.20 g/210L Test # (Repeat) 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

FORGOT TO ATTACH HOSE ON .200

Data Download Date/Time 02/19/2026 01:58

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006770

| | | |
|----------------------------------|------------------------|-------------------------------|
| Agency FL HIGHWAY PATROL | Date 03/11/2026 | Type Agency Inspection |
| Inspector BLAKE, DAVID, A | Time 20:57:57 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 3024028434 |
| Expiration 06/24/2027 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 09/08/2026 |

Number of Simulators 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

INTOX FREE OF CRACKS AND DAMAGE

Data Download Date/Time 03/19/2026 10:26

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006770

Agency FL HIGHWAY PATROL **Date** 03/19/2026 **Type** Department Inspection
Inspector GUTSCHOW,TAYLOR,D **Time** 14:06:47 **Software** 8100.27

0.05 Lot # 202406K **0.08 Lot #** 202406L **0.20 Lot #** 202406N **0.08 Gas Lot #** AG526603
Expiration 06/19/2026 **Expiration** 06/19/2026 **Expiration** 06/20/2026 **Expiration** 09/23/2027

Number of Simulators 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | Yes | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | Yes | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |
| | | Alcohol Free Test # 4 | 0.000 |
| | | 0.05 g/210L Test # 4 | 0.049 |
| | | 0.08 g/210L Test # 4 | 0.078 |
| | | 0.20 g/210L Test # 4 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 4 | 0.080 |
| | | Interferent Detect Test # 4 | INT |
| | | Alcohol Free Test # 5 | 0.000 |
| | | 0.05 g/210L Test # 5 | 0.048 |
| | | 0.08 g/210L Test # 5 | 0.078 |
| | | 0.20 g/210L Test # 5 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 5 | 0.080 |
| | | Interferent Detect Test # 5 | INT |

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006770

| | |
|--|--------|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.048 |
| 0.08 g/210L Test # 6 | 0.078 |
| 0.20 g/210L Test # 6 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.080 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.049 |
| 0.08 g/210L Test # 7 | 0.078 |
| 0.20 g/210L Test # 7 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.080 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.049 |
| 0.08 g/210L Test # 8 | 0.078 |
| 0.20 g/210L Test # 8 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.080 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.049 |
| 0.08 g/210L Test # 9 | 0.078 |
| 0.20 g/210L Test # 9 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.080 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.049 |
| 0.08 g/210L Test # 10 | 0.078 |
| 0.20 g/210L Test # 10 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.080 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0005 |
| 0.08 g/210L Test (Std Dev) | 0.0003 |
| 0.20 g/210L Test (Std Dev) | 0.0005 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0003 |

Remarks:

Data Download Date/Time 03/19/2026 14:41

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006770

| | | |
|---------------------------------|------------------------|-------------------------------|
| Agency FL HIGHWAY PATROL | Date 04/02/2026 | Type Agency Inspection |
| Inspector BLAKE,DAVID,A | Time 01:36:25 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 3024028434 |
| Expiration 06/24/2027 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 09/08/2026 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.196 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

A F / M A: No Sample Provided.

Data Download Date/Time 04/03/2026 18:34

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006771

| | | | |
|---|--|-----------------------------------|----------------------------------|
| Agency FL HIGHWAY PATROL | Date 03/18/2026 | Type Department Inspection | |
| Inspector HIGGINBOTHAM, LEANDRA, | Time 16:30:57 | Software 8100.27 | |
| 0.05 Lot # 202406K | 0.08 Lot # 202406L | 0.20 Lot # 202406N | 0.08 Gas Lot # 23825080A1 |
| Expiration 06/19/2026 | Expiration 06/19/2026 | Expiration 06/20/2026 | Expiration 10/05/2027 |
| Number of Simulators 5 | Average Standard Deviation 0.0004 | Compliance Yes | |

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | Yes | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | Yes | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| | | Interferent Detect Test # 3 | INT |
| | | Alcohol Free Test # 4 | 0.000 |
| | | 0.05 g/210L Test # 4 | 0.048 |
| | | 0.08 g/210L Test # 4 | 0.080 |
| | | 0.20 g/210L Test # 4 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 4 | 0.081 |
| | | Interferent Detect Test # 4 | INT |
| | | Alcohol Free Test # 5 | 0.000 |
| | | 0.05 g/210L Test # 5 | 0.048 |
| | | 0.08 g/210L Test # 5 | 0.079 |
| | | 0.20 g/210L Test # 5 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 5 | 0.082 |
| | | Interferent Detect Test # 5 | INT |

Florida Department of Law Enforcement
Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006771

| | |
|--|--------|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.048 |
| 0.08 g/210L Test # 6 | 0.080 |
| 0.20 g/210L Test # 6 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.082 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.047 |
| 0.08 g/210L Test # 7 | 0.079 |
| 0.20 g/210L Test # 7 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.081 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.048 |
| 0.08 g/210L Test # 8 | 0.080 |
| 0.20 g/210L Test # 8 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.082 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.047 |
| 0.08 g/210L Test # 9 | 0.079 |
| 0.20 g/210L Test # 9 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.082 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.047 |
| 0.08 g/210L Test # 10 | 0.080 |
| 0.20 g/210L Test # 10 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.082 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Test (Std Dev) | 0.0005 |
| 0.20 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0005 |

Remarks:

Data Download Date/Time 03/18/2026 16:29

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

| | | | |
|--------------------------------|------------------------|-------------------------------|--|
| Agency MIAMI DADE SO | Date 01/16/2026 | Type Agency Inspection | |
| Inspector AZRAK,JOSEPH, | Time 00:17:41 | Software 8100.27 | |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 202402E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 28424080A4 |
| Expiration 02/27/2026 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 11/05/2026 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

AI R CLOSIUS/R MYRTIL PRESENT

Data Download Date/Time 01/16/2026 00:39

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

| | | |
|--------------------------------|------------------------|-------------------------------|
| Agency MIAMI DADE SO | Date 02/05/2026 | Type Agency Inspection |
| Inspector AZRAK,JOSEPH, | Time 02:06:59 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 202402E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 28424080A4 |
| Expiration 02/27/2026 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 11/05/2026 |

Number of Simulators 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

AI R CLOSIUS/R MYRTIL PRESENT

Data Download Date/Time 02/05/2026 02:16

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

| | | |
|--------------------------------|------------------------|-------------------------------|
| Agency MIAMI DADE SO | Date 03/05/2026 | Type Agency Inspection |
| Inspector AZRAK,JOSEPH, | Time 03:25:55 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 28424080A4 |
| Expiration 06/24/2027 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 11/05/2026 |

Number of Simulators 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.196 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.196 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

AI R CLOSIUS/R MYRTIL PRESENT

Data Download Date/Time 03/05/2026 03:42

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

| | | |
|--------------------------------|------------------------|-------------------------------|
| Agency MIAMI DADE SO | Date 04/16/2026 | Type Agency Inspection |
| Inspector AZRAK,JOSEPH, | Time 03:52:21 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 28424080A4 |
| Expiration 06/24/2027 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 11/05/2026 |

Number of Simulators 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.196 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.196 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

AI CLOSIUS/R MYRTIL PRESENT

Data Download Date/Time 04/16/2026 04:08

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

| | | |
|------------------------------------|------------------------|-------------------------------|
| Agency BROWARD COUNTY SO | Date 01/14/2026 | Type Agency Inspection |
| Inspector FRAZIER, ANAYA, S | Time 08:29:32 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|---------------------------------|
| 0.05 Lot # 202402E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 403345775 |
| Expiration 02/27/2026 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 05/13/2028 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.077 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

S. TILLMAN IN TRAINING

Data Download Date/Time 01/14/2026 08:39

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

| | | |
|------------------------------------|------------------------|-------------------------------|
| Agency BROWARD COUNTY SO | Date 02/04/2026 | Type Agency Inspection |
| Inspector FRAZIER, ANAYA, S | Time 08:25:58 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|---------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 403345775 |
| Expiration 06/24/2027 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 05/13/2028 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.082 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.051 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.081 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.081 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

S. TILLMAN IN TRAINING.

Data Download Date/Time 02/04/2026 08:35

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

| | | |
|------------------------------------|------------------------|-------------------------------|
| Agency BROWARD COUNTY SO | Date 03/04/2026 | Type Agency Inspection |
| Inspector FRAZIER, ANAYA, S | Time 08:03:32 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|---------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 403345775 |
| Expiration 06/24/2027 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 05/13/2028 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.200 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

S. TILLMAN IN TRAINING.

Data Download Date/Time 03/04/2026 08:04

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

Agency BROWARD COUNTY SO **Date** 04/02/2026 **Type** Department Inspection
Inspector GUTSCHOW,TAYLOR,D **Time** 13:32:10 **Software** 8100.27

0.05 Lot # 202406K **0.08 Lot #** 202406L **0.20 Lot #** 202406N **0.08 Gas Lot #** AG526603
Expiration 06/19/2026 **Expiration** 06/19/2026 **Expiration** 06/20/2026 **Expiration** 09/23/2027

Number of Simulators 5 **Average Standard Deviation** 0.0005 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | Yes | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | Yes | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.047 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.077 |
| | | Interferent Detect Test # 3 | INT |
| | | Alcohol Free Test # 4 | 0.000 |
| | | 0.05 g/210L Test # 4 | 0.048 |
| | | 0.08 g/210L Test # 4 | 0.079 |
| | | 0.20 g/210L Test # 4 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 4 | 0.078 |
| | | Interferent Detect Test # 4 | INT |
| | | Alcohol Free Test # 5 | 0.000 |
| | | 0.05 g/210L Test # 5 | 0.048 |
| | | 0.08 g/210L Test # 5 | 0.079 |
| | | 0.20 g/210L Test # 5 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 5 | 0.077 |
| | | Interferent Detect Test # 5 | INT |

Florida Department of Law Enforcement
Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

| | |
|--|--------|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.048 |
| 0.08 g/210L Test # 6 | 0.080 |
| 0.20 g/210L Test # 6 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.077 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.049 |
| 0.08 g/210L Test # 7 | 0.080 |
| 0.20 g/210L Test # 7 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.077 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.048 |
| 0.08 g/210L Test # 8 | 0.080 |
| 0.20 g/210L Test # 8 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.077 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.048 |
| 0.08 g/210L Test # 9 | 0.080 |
| 0.20 g/210L Test # 9 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.077 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.049 |
| 0.08 g/210L Test # 10 | 0.080 |
| 0.20 g/210L Test # 10 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.077 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0006 |
| 0.08 g/210L Test (Std Dev) | 0.0005 |
| 0.20 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0006 |

Remarks:

Data Download Date/Time 04/02/2026 14:07

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

| | | |
|------------------------------------|------------------------|-------------------------------|
| Agency BROWARD COUNTY SO | Date 04/17/2026 | Type Agency Inspection |
| Inspector FRAZIER, ANAYA, S | Time 23:14:22 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|---------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 403345775 |
| Expiration 06/24/2027 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 05/13/2028 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.200 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

S. TILLMAN IN TRAINING.

Data Download Date/Time 04/17/2026 23:16

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

| | | |
|------------------------------------|------------------------|-------------------------------|
| Agency BROWARD COUNTY SO | Date 01/14/2026 | Type Agency Inspection |
| Inspector FRAZIER, ANAYA, S | Time 09:19:41 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|---------------------------------|
| 0.05 Lot # 202402E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 402783528 |
| Expiration 02/27/2026 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 06/27/2026 |

Number of Simulators 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

S. TILLMAN IN TRAINING.

Data Download Date/Time 01/14/2026 09:28

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

| | | |
|------------------------------------|------------------------|-------------------------------|
| Agency BROWARD COUNTY SO | Date 02/27/2026 | Type Agency Inspection |
| Inspector GUTSCHOW,TAYLOR,D | Time 15:38:24 | Software 8100.27 |

| | | | |
|-------------------|-------------------|-------------------|-----------------------|
| 0.05 Lot # | 0.08 Lot # | 0.20 Lot # | 0.08 Gas Lot # |
| Expiration | Expiration | Expiration | Expiration |

Number of Simulators **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|------|--------|
| Diagnostic Check (Pre-Inspection): OK | No | | |
| Date and/or Time Adjusted | No | | |
| Minimum Sample Volume Check: OK | | | |
| Barometric Pressure Sensor Check: OK | | | |
| Alcohol Free Subject Test: 0.000 | No | | |
| Mouth Alcohol Test: Slope Not Met | No | | |
| Interferent Detect Test: Interferent Detect | No | | |
| Diagnostic Check (Post-Inspection): OK | No | | |
| Alcohol Free Test: OK | No | | |
| 0.05 g/210L Test: OK | No | | |
| 0.08 g/210L Test: OK | No | | |
| 0.20 g/210L Test: OK | No | | |
| 0.08 g/210L Dry Gas Std Test: OK | No | | |

Remarks:

AI NOT CONDUCTED. COMPLIANCE NOT DETERMINED.

Data Download Date/Time 02/27/2026 15:38

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

Agency BROWARD COUNTY SO **Date** 03/05/2026 **Type** Department Inspection
Inspector GUTSCHOW,TAYLOR,D **Time** 13:03:19 **Software** 8100.27

0.05 Lot # 202406K **0.08 Lot #** 202406L **0.20 Lot #** 202406N **0.08 Gas Lot #** AG526603
Expiration 06/19/2026 **Expiration** 06/19/2026 **Expiration** 06/20/2026 **Expiration** 09/23/2027

Number of Simulators 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | Yes | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | Yes | 0.20 g/210L Test # 1 | 0.196 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.196 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |
| | | Alcohol Free Test # 4 | 0.000 |
| | | 0.05 g/210L Test # 4 | 0.048 |
| | | 0.08 g/210L Test # 4 | 0.079 |
| | | 0.20 g/210L Test # 4 | 0.196 |
| | | 0.08 g/210L Dry Gas Std Test # 4 | 0.079 |
| | | Interferent Detect Test # 4 | INT |
| | | Alcohol Free Test # 5 | 0.000 |
| | | 0.05 g/210L Test # 5 | 0.048 |
| | | 0.08 g/210L Test # 5 | 0.079 |
| | | 0.20 g/210L Test # 5 | 0.196 |
| | | 0.08 g/210L Dry Gas Std Test # 5 | 0.078 |
| | | Interferent Detect Test # 5 | INT |

Florida Department of Law Enforcement
Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

| | |
|--|--------|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.048 |
| 0.08 g/210L Test # 6 | 0.078 |
| 0.20 g/210L Test # 6 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.079 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.049 |
| 0.08 g/210L Test # 7 | 0.079 |
| 0.20 g/210L Test # 7 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.078 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.048 |
| 0.08 g/210L Test # 8 | 0.079 |
| 0.20 g/210L Test # 8 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.078 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.048 |
| 0.08 g/210L Test # 9 | 0.079 |
| 0.20 g/210L Test # 9 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.079 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.048 |
| 0.08 g/210L Test # 10 | 0.079 |
| 0.20 g/210L Test # 10 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.079 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Test (Std Dev) | 0.0003 |
| 0.20 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0005 |

Remarks:

Data Download Date/Time 03/05/2026 13:04

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

| | | |
|------------------------------------|------------------------|-------------------------------|
| Agency BROWARD COUNTY SO | Date 03/31/2026 | Type Agency Inspection |
| Inspector FRAZIER, ANAYA, S | Time 23:06:36 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|---------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 402783528 |
| Expiration 06/24/2027 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 06/27/2026 |

Number of Simulators 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

S. TILLMAN IN TRAINING.

Data Download Date/Time 03/31/2026 23:08

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

| | | |
|------------------------------------|------------------------|-------------------------------|
| Agency BROWARD COUNTY SO | Date 04/01/2026 | Type Agency Inspection |
| Inspector FRAZIER, ANAYA, S | Time 10:13:29 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|---------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 402783528 |
| Expiration 06/24/2027 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 06/27/2026 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.075 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.075 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.075 |
| | | 0.20 g/210L Test # 3 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

S. TILLMAN IN TRAINING

Data Download Date/Time 04/27/2026 02:03

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006926

| | | |
|------------------------------|------------------------|-------------------------------|
| Agency TALLAHASSEE PD | Date 01/21/2026 | Type Agency Inspection |
| Inspector ROY,GARY,A | Time 20:05:45 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202510E | 0.20 Lot # 202410F | 0.08 Gas Lot # 31124080A2 |
| Expiration 06/24/2027 | Expiration 10/28/2027 | Expiration 10/30/2026 | Expiration 12/05/2026 |

Number of Simulators 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.051 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

IN COMPLIANCE

Data Download Date/Time 01/21/2026 20:04

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006926

| | | | |
|-----------------------|-----------------|------------------------|--|
| Agency TALLAHASSEE PD | Date 02/12/2026 | Type Agency Inspection | |
| Inspector ROY,GARY,A | Time 17:11:45 | Software 8100.27 | |

| | | | |
|-----------------------|-----------------------|-----------------------|---------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202510E | 0.20 Lot # 202410F | 0.08 Gas Lot # 31124080A2 |
| Expiration 06/24/2027 | Expiration 10/28/2027 | Expiration 10/30/2026 | Expiration 12/05/2026 |

Number of Simulators 5 Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.075 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.196 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.075 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.075 |
| | | 0.20 g/210L Test # 3 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 02/12/2026 17:11

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006926

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| Agency TALLAHASSEE PD | Date 03/19/2026 | Type Agency Inspection | |
| Inspector ROY,GARY,A | Time 14:32:19 | Software 8100.27 | |
| 0.05 Lot # 202506E | 0.08 Lot # 202510E | 0.20 Lot # 202410F | 0.08 Gas Lot # 31124080A2 |
| Expiration 06/24/2027 | Expiration 10/28/2027 | Expiration 10/30/2026 | Expiration 12/05/2026 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.195 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.077 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.083 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 03/19/2026 14:31

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006926

| | | | |
|-----------------------|-----------------|------------------------|--|
| Agency TALLAHASSEE PD | Date 04/09/2026 | Type Agency Inspection | |
| Inspector ROY,GARY,A | Time 14:44:40 | Software 8100.27 | |

| | | | |
|-----------------------|-----------------------|-----------------------|---------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202510E | 0.20 Lot # 202410F | 0.08 Gas Lot # 31124080A2 |
| Expiration 06/24/2027 | Expiration 10/28/2027 | Expiration 10/30/2026 | Expiration 12/05/2026 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.205 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.077 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.203 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.077 |
| | | 0.20 g/210L Test # 3 | 0.202 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 04/09/2026 14:43

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006938

| | | |
|----------------------------------|------------------------|-------------------------------|
| Agency FORT LAUDERDALE PD | Date 01/31/2026 | Type Agency Inspection |
| Inspector STONER,NATHAN,C | Time 21:11:29 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|---------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202510E | 0.20 Lot # 202510F | 0.08 Gas Lot # 402886996 |
| Expiration 06/24/2027 | Expiration 10/28/2027 | Expiration 10/29/2027 | Expiration 11/01/2026 |

Number of Simulators 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 02/03/2026 18:11

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006938

| | | |
|----------------------------------|------------------------|-------------------------------|
| Agency FORT LAUDERDALE PD | Date 02/28/2026 | Type Agency Inspection |
| Inspector STONER,NATHAN,C | Time 20:44:39 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|---------------------------------|
| 0.05 Lot # 2025006E | 0.08 Lot # 202510E | 0.20 Lot # 202510F | 0.08 Gas Lot # 402886996 |
| Expiration 06/24/2027 | Expiration 10/28/2027 | Expiration 10/29/2027 | Expiration 11/01/2026 |

Number of Simulators 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 03/01/2026 17:56

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006938

| | | |
|----------------------------------|------------------------|-------------------------------|
| Agency FORT LAUDERDALE PD | Date 03/31/2026 | Type Agency Inspection |
| Inspector STONER,NATHAN,C | Time 20:51:40 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|---------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202510E | 0.20 Lot # 202510F | 0.08 Gas Lot # 402886996 |
| Expiration 06/24/2027 | Expiration 10/28/2027 | Expiration 10/29/2027 | Expiration 11/01/2026 |

Number of Simulators 5

Compliance No

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

SGT THOMAS PRESENT AND OBSERVED Non-compliance: .

Data Download Date/Time 03/31/2026 21:50

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006939

| | | |
|----------------------------------|------------------------|-------------------------------|
| Agency FORT LAUDERDALE PD | Date 01/31/2026 | Type Agency Inspection |
| Inspector STONER,NATHAN,C | Time 21:12:11 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|---------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202510E | 0.20 Lot # 202510F | 0.08 Gas Lot # 402886996 |
| Expiration 06/24/2027 | Expiration 10/28/2027 | Expiration 10/29/2027 | Expiration 11/01/2026 |

Number of Simulators 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 02/03/2026 18:15

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006939

| | | |
|----------------------------------|------------------------|-------------------------------|
| Agency FORT LAUDERDALE PD | Date 02/28/2026 | Type Agency Inspection |
| Inspector STONER,NATHAN,C | Time 20:57:20 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|---------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202510E | 0.20 Lot # 202510F | 0.08 Gas Lot # 403476659 |
| Expiration 06/24/2027 | Expiration 10/28/2027 | Expiration 10/29/2027 | Expiration 10/28/2028 |

Number of Simulators 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.042 |
| Minimum Sample Volume Check: OK | | 0.05 g/210L Test # (Repeat) 1 | 0.051 |
| Barometric Pressure Sensor Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.20 g/210L Test # 1 | 0.199 |
| Mouth Alcohol Test: Slope Not Met | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test: Interferent Detect | Yes | Interferent Detect Test # 1 | INT |
| Diagnostic Check (Post-Inspection): OK | Yes | Alcohol Free Test # 2 | 0.000 |
| Alcohol Free Test: OK | Yes | 0.05 g/210L Test # 2 | 0.042 |
| 0.05 g/210L Test: OK | Yes | 0.05 g/210L Test # (Repeat) 2 | 0.051 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| | | Interferent Detect Test # 2 | INT |
| | | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.042 |
| | | 0.05 g/210L Test # (Repeat) 3 | 0.051 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

05: Control Outside Tolerance. LOOSE SIM

Data Download Date/Time 03/01/2026 18:02

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006939

| | | |
|----------------------------------|------------------------|-------------------------------|
| Agency FORT LAUDERDALE PD | Date 03/31/2026 | Type Agency Inspection |
| Inspector STONER,NATHAN,C | Time 20:44:14 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|---------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202510E | 0.20 Lot # 202510F | 0.08 Gas Lot # 403476659 |
| Expiration 06/24/2027 | Expiration 10/28/2027 | Expiration 10/29/2027 | Expiration 10/28/2028 |

Number of Simulators 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.201 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.081 |
| | | 0.20 g/210L Test # 3 | 0.201 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

SGT THOMAS PRESENT AND OBSERVED

Data Download Date/Time 03/31/2026 21:54

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

| | | |
|------------------------------------|------------------------|-------------------------------|
| Agency BROWARD COUNTY SO | Date 01/14/2026 | Type Agency Inspection |
| Inspector FRAZIER, ANAYA, S | Time 09:18:11 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|---------------------------------|
| 0.05 Lot # 202402E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 403345775 |
| Expiration 02/27/2026 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 05/13/2028 |

Number of Simulators 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.201 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.201 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

S. TILLMAN IN TRAINING.

Data Download Date/Time 01/14/2026 09:24

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

| | | |
|------------------------------------|------------------------|-------------------------------|
| Agency BROWARD COUNTY SO | Date 02/04/2026 | Type Agency Inspection |
| Inspector FRAZIER, ANAYA, S | Time 09:30:17 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|---------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 403345775 |
| Expiration 06/24/2027 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 05/13/2028 |

Number of Simulators 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.066 |
| Barometric Pressure Sensor Check: OK | | 0.08 g/210L Test # (Repeat) 1 | 0.081 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.20 g/210L Test # 1 | 0.201 |
| Mouth Alcohol Test: Slope Not Met | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test: Interferent Detect | Yes | Interferent Detect Test # 1 | INT |
| Diagnostic Check (Post-Inspection): OK | Yes | Alcohol Free Test # 2 | 0.000 |
| Alcohol Free Test: OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| 0.05 g/210L Test: OK | Yes | 0.08 g/210L Test # 2 | 0.060 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Test # (Repeat) 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.203 |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| | | Interferent Detect Test # 2 | INT |
| | | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.071 |
| | | 0.08 g/210L Test # (Repeat) 3 | 0.081 |
| | | 0.20 g/210L Test # 3 | 0.202 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

08: Control Outside Tolerance.SIM NOT SECURELY ATTACHED.

Data Download Date/Time 02/04/2026 09:34

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

| | | |
|------------------------------------|------------------------|-------------------------------|
| Agency BROWARD COUNTY SO | Date 03/04/2026 | Type Agency Inspection |
| Inspector FRAZIER, ANAYA, S | Time 09:36:13 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|---------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 403345775 |
| Expiration 06/24/2027 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 05/13/2028 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.202 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.201 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

S. TILLMAN IN TRAINING.

Data Download Date/Time 03/05/2026 02:48

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

Agency BROWARD COUNTY SO **Date** 03/20/2026 **Type** Department Inspection
Inspector GUTSCHOW,TAYLOR,D **Time** 14:27:26 **Software** 8100.27

0.05 Lot # 202406K **0.08 Lot #** 202406L **0.20 Lot #** 202406N **0.08 Gas Lot #** AG526603
Expiration 06/19/2026 **Expiration** 06/19/2026 **Expiration** 06/20/2026 **Expiration** 09/23/2027

Number of Simulators 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | Yes | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | Yes | 0.20 g/210L Test # 1 | 0.200 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.077 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.077 |
| | | Interferent Detect Test # 3 | INT |
| | | Alcohol Free Test # 4 | 0.000 |
| | | 0.05 g/210L Test # 4 | 0.050 |
| | | 0.08 g/210L Test # 4 | 0.079 |
| | | 0.20 g/210L Test # 4 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 4 | 0.078 |
| | | Interferent Detect Test # 4 | INT |
| | | Alcohol Free Test # 5 | 0.000 |
| | | 0.05 g/210L Test # 5 | 0.048 |
| | | 0.08 g/210L Test # 5 | 0.079 |
| | | 0.20 g/210L Test # 5 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 5 | 0.078 |
| | | Interferent Detect Test # 5 | INT |

Florida Department of Law Enforcement
Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

| | |
|--|--------|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.049 |
| 0.08 g/210L Test # 6 | 0.079 |
| 0.20 g/210L Test # 6 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.078 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.049 |
| 0.08 g/210L Test # 7 | 0.079 |
| 0.20 g/210L Test # 7 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.078 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.049 |
| 0.08 g/210L Test # 8 | 0.079 |
| 0.20 g/210L Test # 8 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.078 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.049 |
| 0.08 g/210L Test # 9 | 0.079 |
| 0.20 g/210L Test # 9 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.077 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.049 |
| 0.08 g/210L Test # 10 | 0.079 |
| 0.20 g/210L Test # 10 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.078 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Test (Std Dev) | 0.0000 |
| 0.20 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0004 |

Remarks:

Data Download Date/Time 03/20/2026 14:24

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

| | | |
|------------------------------------|------------------------|-------------------------------|
| Agency BROWARD COUNTY SO | Date 04/17/2026 | Type Agency Inspection |
| Inspector FRAZIER, ANAYA, S | Time 23:22:33 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|---------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202410E | 0.20 Lot # 202410E | 0.08 Gas Lot # 403345775 |
| Expiration 05/24/2027 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 05/13/2028 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.201 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.202 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.202 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

S. TILLMAN IN TRAINING.

Data Download Date/Time 04/17/2026 23:21

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007007

| | | |
|----------------------------------|------------------------|-------------------------------|
| Agency FORT LAUDERDALE PD | Date 02/28/2026 | Type Agency Inspection |
| Inspector STONER,NATHAN,C | Time 20:48:26 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|---------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202510E | 0.20 Lot # 202510F | 0.08 Gas Lot # 403476659 |
| Expiration 06/24/2027 | Expiration 10/28/2027 | Expiration 10/29/2027 | Expiration 10/28/2028 |

Number of Simulators 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 03/07/2026 15:23

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007007

| | | |
|----------------------------------|------------------------|-------------------------------|
| Agency FORT LAUDERDALE PD | Date 03/31/2026 | Type Agency Inspection |
| Inspector STONER,NATHAN,C | Time 20:42:29 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|---------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202510E | 0.20 Lot # 202510F | 0.08 Gas Lot # 403476659 |
| Expiration 06/24/2027 | Expiration 10/28/2027 | Expiration 10/29/2027 | Expiration 10/28/2028 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.081 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

SGT THOMAS PRESENT AND OBSERVED

Data Download Date/Time 03/31/2026 21:28

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007008

| | | |
|------------------------------------|------------------------|-------------------------------|
| Agency BROWARD COUNTY SO | Date 01/14/2026 | Type Agency Inspection |
| Inspector FRAZIER, ANAYA, S | Time 10:02:12 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|---------------------------------|
| 0.05 Lot # 202402E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 402783528 |
| Expiration 02/27/2026 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 01/14/2026 |

Number of Simulators 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.200 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

S. TILLMAN IN TRAINING.

Data Download Date/Time 01/14/2026 10:25

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007008

| | | |
|------------------------------------|------------------------|-------------------------------|
| Agency BROWARD COUNTY SO | Date 02/04/2026 | Type Agency Inspection |
| Inspector FRAZIER, ANAYA, S | Time 09:14:53 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|---------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 402783528 |
| Expiration 06/24/2027 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 06/27/2026 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.200 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.201 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

S. TILLMAN IN TRAINING.

Data Download Date/Time 02/04/2026 09:15

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007008

| | | |
|------------------------------------|------------------------|-------------------------------|
| Agency BROWARD COUNTY SO | Date 03/04/2026 | Type Agency Inspection |
| Inspector FRAZIER, ANAYA, S | Time 08:34:44 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|---------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 402783528 |
| Expiration 06/24/2027 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 06/27/2026 |

Number of Simulators 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.200 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.201 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

S. TILLMAN IN TRAINING.

Data Download Date/Time 03/05/2026 02:54

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007008

Agency BROWARD COUNTY SO **Date** 03/20/2026 **Type** Department Inspection
Inspector GUTSCHOW,TAYLOR,D **Time** 14:11:08 **Software** 8100.27

0.05 Lot # 202406K **0.08 Lot #** 202406L **0.20 Lot #** 202406N **0.08 Gas Lot #** AG526603
Expiration 06/19/2026 **Expiration** 06/19/2026 **Expiration** 06/20/2026 **Expiration** 09/23/2027

Number of Simulators 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | Yes | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | Yes | 0.20 g/210L Test # 1 | 0.196 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |
| | | Alcohol Free Test # 4 | 0.000 |
| | | 0.05 g/210L Test # 4 | 0.049 |
| | | 0.08 g/210L Test # 4 | 0.078 |
| | | 0.20 g/210L Test # 4 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 4 | 0.080 |
| | | Interferent Detect Test # 4 | INT |
| | | Alcohol Free Test # 5 | 0.000 |
| | | 0.05 g/210L Test # 5 | 0.048 |
| | | 0.08 g/210L Test # 5 | 0.078 |
| | | 0.20 g/210L Test # 5 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 5 | 0.079 |
| | | Interferent Detect Test # 5 | INT |

Florida Department of Law Enforcement
Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007008

| | |
|--|--------|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.048 |
| 0.08 g/210L Test # 6 | 0.079 |
| 0.20 g/210L Test # 6 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.079 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.048 |
| 0.08 g/210L Test # 7 | 0.079 |
| 0.20 g/210L Test # 7 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.080 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.048 |
| 0.08 g/210L Test # 8 | 0.077 |
| 0.20 g/210L Test # 8 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.079 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.048 |
| 0.08 g/210L Test # 9 | 0.078 |
| 0.20 g/210L Test # 9 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.079 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.048 |
| 0.08 g/210L Test # 10 | 0.078 |
| 0.20 g/210L Test # 10 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.080 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0003 |
| 0.08 g/210L Test (Std Dev) | 0.0006 |
| 0.20 g/210L Test (Std Dev) | 0.0005 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0005 |

Remarks:

Data Download Date/Time 03/20/2026 14:08

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007008

| | | |
|------------------------------------|------------------------|-------------------------------|
| Agency BROWARD COUNTY SO | Date 04/17/2026 | Type Agency Inspection |
| Inspector FRAZIER, ANAYA, S | Time 22:39:47 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|---------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 402783528 |
| Expiration 06/24/2027 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 06/27/2026 |

Number of Simulators 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.200 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

S. TILLMAN IN TRAINING.

Data Download Date/Time 04/17/2026 23:07

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

| | | |
|----------------------------------|------------------------|-------------------------------|
| Agency COLLIER COUNTY SO | Date 01/24/2026 | Type Agency Inspection |
| Inspector LANDAS, CARTER, | Time 13:45:52 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 202402E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 20524080A2 |
| Expiration 02/27/2026 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 09/05/2026 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.200 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

TESTED AT NJC

Data Download Date/Time 01/25/2026 07:15

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

| | | |
|------------------------------------|------------------------|-------------------------------|
| Agency COLLIER COUNTY SO | Date 03/06/2026 | Type Agency Inspection |
| Inspector GUTSCHOW,TAYLOR,D | Time 14:10:12 | Software 8100.27 |

| | | | |
|-------------------|-------------------|-------------------|-----------------------|
| 0.05 Lot # | 0.08 Lot # | 0.20 Lot # | 0.08 Gas Lot # |
| Expiration | Expiration | Expiration | Expiration |

Number of Simulators **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|------|--------|
| Diagnostic Check (Pre-Inspection): OK | No | | |
| Date and/or Time Adjusted | No | | |
| Minimum Sample Volume Check: OK | | | |
| Barometric Pressure Sensor Check: OK | | | |
| Alcohol Free Subject Test: 0.000 | No | | |
| Mouth Alcohol Test: Slope Not Met | No | | |
| Interferent Detect Test: Interferent Detect | No | | |
| Diagnostic Check (Post-Inspection): OK | No | | |
| Alcohol Free Test: OK | No | | |
| 0.05 g/210L Test: OK | No | | |
| 0.08 g/210L Test: OK | No | | |
| 0.20 g/210L Test: OK | No | | |
| 0.08 g/210L Dry Gas Std Test: OK | No | | |

Remarks:

AI NOT CONDUCTED. COMPLIANCE NOT DETERMINED.

Data Download Date/Time 03/09/2026 12:18

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

| | | |
|------------------------------------|------------------------|-----------------------------------|
| Agency COLLIER COUNTY SO | Date 03/09/2026 | Type Department Inspection |
| Inspector GUTSCHOW,TAYLOR,D | Time 11:50:49 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| 0.05 Lot # 202406K | 0.08 Lot # 202406L | 0.20 Lot # 202406N | 0.08 Gas Lot # AG429602 |
| Expiration 06/19/2026 | Expiration 06/19/2026 | Expiration 06/20/2026 | Expiration 10/22/2026 |

Number of Simulators 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | Yes | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | Yes | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |
| | | Alcohol Free Test # 4 | 0.000 |
| | | 0.05 g/210L Test # 4 | 0.048 |
| | | 0.08 g/210L Test # 4 | 0.079 |
| | | 0.20 g/210L Test # 4 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 4 | 0.079 |
| | | Interferent Detect Test # 4 | INT |
| | | Alcohol Free Test # 5 | 0.000 |
| | | 0.05 g/210L Test # 5 | 0.048 |
| | | 0.08 g/210L Test # 5 | 0.079 |
| | | 0.20 g/210L Test # 5 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 5 | 0.079 |
| | | Interferent Detect Test # 5 | INT |

**Florida Department of Law Enforcement
Alcohol Testing Program**

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

| | |
|--|--------|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.048 |
| 0.08 g/210L Test # 6 | 0.079 |
| 0.20 g/210L Test # 6 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.079 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.049 |
| 0.08 g/210L Test # 7 | 0.080 |
| 0.20 g/210L Test # 7 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.079 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.049 |
| 0.08 g/210L Test # 8 | 0.079 |
| 0.20 g/210L Test # 8 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.078 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.049 |
| 0.08 g/210L Test # 9 | 0.080 |
| 0.20 g/210L Test # 9 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.079 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.049 |
| 0.08 g/210L Test # 10 | 0.080 |
| 0.20 g/210L Test # 10 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.079 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0005 |
| 0.08 g/210L Test (Std Dev) | 0.0004 |
| 0.20 g/210L Test (Std Dev) | 0.0003 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0004 |

Remarks:

Data Download Date/Time 03/09/2026 12:18

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

| | | |
|----------------------------------|------------------------|-------------------------------|
| Agency COLLIER COUNTY SO | Date 03/27/2026 | Type Agency Inspection |
| Inspector LANDAS, CARTER, | Time 07:45:58 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 20524080A2 |
| Expiration 06/24/2027 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 09/05/2026 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

TESTED AT NJC

Data Download Date/Time 03/27/2026 23:05

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

| | | | |
|---------------------------------|------------------------------|-------------------------------|----------------------------------|
| Agency COLLIER COUNTY SO | Date 04/25/2026 | Type Agency Inspection | |
| Inspector LANDAS,CARTER, | Time 15:56:54 | Software 8100.27 | |
| 0.05 Lot # 202506E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 20524080A2 |
| Expiration 06/24/2027 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 09/05/2026 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.051 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.051 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.051 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 04/26/2026 08:21

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007078

| | | | |
|----------------------------------|------------------------|-------------------------------|--|
| Agency COLLIER COUNTY SO | Date 01/23/2026 | Type Agency Inspection | |
| Inspector LANDAS, CARTER, | Time 08:17:37 | Software 8100.27 | |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 202402E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 20524080A2 |
| Expiration 02/27/2026 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 09/05/2026 |

Number of Simulators 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.051 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

TESTED AT NJC

Data Download Date/Time 01/23/2026 12:49

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007078

| | | |
|---------------------------------|------------------------|-------------------------------|
| Agency COLLIER COUNTY SO | Date 02/21/2026 | Type Agency Inspection |
| Inspector LANDAS,CARTER, | Time 11:21:11 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 202402E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 20524080A2 |
| Expiration 02/27/2026 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 09/05/2026 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

TESTED AT NJC

Data Download Date/Time 02/21/2026 12:13

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007078

| | | |
|----------------------------------|------------------------|-------------------------------|
| Agency COLLIER COUNTY SO | Date 03/27/2026 | Type Agency Inspection |
| Inspector LANDAS, CARTER, | Time 21:10:18 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 20524080A2 |
| Expiration 06/24/2027 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 09/05/2026 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

TESTED AT NJC

Data Download Date/Time 03/27/2026 23:01

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007078

| | | |
|----------------------------------|------------------------|-------------------------------|
| Agency COLLIER COUNTY SO | Date 04/25/2026 | Type Agency Inspection |
| Inspector LANDAS, CARTER, | Time 11:20:42 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 20524080A2 |
| Expiration 06/24/2027 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 09/05/2026 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.051 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 04/25/2026 14:40

**Florida Department of Law Enforcement
Alcohol Testing Program**

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

| | | |
|--------------------------------------|------------------------|-------------------------------|
| Agency CHARLOTTE COUNTY SO | Date 02/27/2026 | Type Agency Inspection |
| Inspector GUTSCHOW, TAYLOR, D | Time 09:26:51 | Software 8100.27 |

| | | | |
|-------------------|-------------------|-------------------|-----------------------|
| 0.05 Lot # | 0.08 Lot # | 0.20 Lot # | 0.08 Gas Lot # |
| Expiration | Expiration | Expiration | Expiration |

Number of Simulators **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|------|--------|
| Diagnostic Check (Pre-Inspection): OK | No | | |
| Date and/or Time Adjusted | No | | |
| Minimum Sample Volume Check: OK | | | |
| Barometric Pressure Sensor Check: OK | | | |
| Alcohol Free Subject Test: 0.000 | No | | |
| Mouth Alcohol Test: Slope Not Met | No | | |
| Interferent Detect Test: Interferent Detect | No | | |
| Diagnostic Check (Post-Inspection): OK | No | | |
| Alcohol Free Test: OK | No | | |
| 0.05 g/210L Test: OK | No | | |
| 0.08 g/210L Test: OK | No | | |
| 0.20 g/210L Test: OK | No | | |
| 0.08 g/210L Dry Gas Std Test: OK | No | | |

Remarks:

AI NOT CONDUCTED. COMPLIANCE NOT DETERMINED.

Data Download Date/Time 02/27/2026 14:05

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

Agency CHARLOTTE COUNTY SO **Date** 02/27/2026 **Type** Department Inspection
Inspector GUTSCHOW,TAYLOR,D **Time** 13:47:48 **Software** 8100.27

0.05 Lot # 202406K **0.08 Lot #** 202406L **0.20 Lot #** 202406N **0.08 Gas Lot #** AG526603
Expiration 06/19/2026 **Expiration** 06/19/2026 **Expiration** 06/20/2026 **Expiration** 09/23/2027

Number of Simulators 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | Yes | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | Yes | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.077 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.077 |
| | | Interferent Detect Test # 3 | INT |
| | | Alcohol Free Test # 4 | 0.000 |
| | | 0.05 g/210L Test # 4 | 0.049 |
| | | 0.08 g/210L Test # 4 | 0.079 |
| | | 0.20 g/210L Test # 4 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 4 | 0.078 |
| | | Interferent Detect Test # 4 | INT |
| | | Alcohol Free Test # 5 | 0.000 |
| | | 0.05 g/210L Test # 5 | 0.048 |
| | | 0.08 g/210L Test # 5 | 0.079 |
| | | 0.20 g/210L Test # 5 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 5 | 0.078 |
| | | Interferent Detect Test # 5 | INT |

Florida Department of Law Enforcement
Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

| | |
|--|--------|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.049 |
| 0.08 g/210L Test # 6 | 0.079 |
| 0.20 g/210L Test # 6 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.078 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.049 |
| 0.08 g/210L Test # 7 | 0.080 |
| 0.20 g/210L Test # 7 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.078 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.049 |
| 0.08 g/210L Test # 8 | 0.080 |
| 0.20 g/210L Test # 8 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.078 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.049 |
| 0.08 g/210L Test # 9 | 0.080 |
| 0.20 g/210L Test # 9 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.077 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.049 |
| 0.08 g/210L Test # 10 | 0.079 |
| 0.20 g/210L Test # 10 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.078 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Test (Std Dev) | 0.0004 |
| 0.20 g/210L Test (Std Dev) | 0.0000 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0004 |

Remarks:

Data Download Date/Time 02/27/2026 14:05

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

| | | |
|-----------------------------------|------------------------|-------------------------------|
| Agency CHARLOTTE COUNTY SO | Date 03/29/2026 | Type Agency Inspection |
| Inspector LAVANCE,ALEYA,D | Time 18:33:42 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 03326080A2 |
| Expiration 06/24/2027 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 03/05/2028 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

CLEANED AND INSPECTED

Data Download Date/Time 03/29/2026 18:34

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

| | | |
|-------------------------------------|------------------------|-------------------------------|
| Agency CHARLOTTE COUNTY SO | Date 04/18/2026 | Type Agency Inspection |
| Inspector CHANDLER,VANESSA,N | Time 20:58:40 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 03326080A2 |
| Expiration 06/24/2027 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 03/05/2028 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

CLEANED AND INSPECTED

Data Download Date/Time 04/18/2026 20:56

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007084

| | | | |
|--------------------------------|------------------------------|-------------------------------|----------------------------------|
| Agency MIAMI DADE SO | Date 01/16/2026 | Type Agency Inspection | |
| Inspector AZRAK,JOSEPH, | Time 02:16:58 | Software 8100.27 | |
| 0.05 Lot # 202402E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 28424080A4 |
| Expiration 02/27/2026 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 11/05/2026 |
| Number of Simulators 5 | Compliance Yes | | |

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.200 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.051 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.201 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

AI R CLOSIUS/R MYRTIL PRESENT

Data Download Date/Time 01/16/2026 02:28

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007084

| | | |
|--------------------------------|------------------------|-------------------------------|
| Agency MIAMI DADE SO | Date 02/05/2026 | Type Agency Inspection |
| Inspector AZRAK,JOSEPH, | Time 02:06:50 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 202402E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 28424080A4 |
| Expiration 02/27/2026 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 11/05/2026 |

Number of Simulators 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

AI R CLOSIUS/R MYRTIL PRESENT

Data Download Date/Time 02/05/2026 02:10

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007084

| | | |
|--------------------------------|------------------------|-------------------------------|
| Agency MIAMI DADE SO | Date 03/05/2026 | Type Agency Inspection |
| Inspector AZRAK,JOSEPH, | Time 01:08:17 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 28424080A4 |
| Expiration 06/24/2027 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 11/05/2026 |

Number of Simulators 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.200 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.201 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

AI R CLOSIUS/R MYRTIL PRESENT

Data Download Date/Time 03/05/2026 01:35

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007084

| | | |
|--------------------------------|------------------------|-------------------------------|
| Agency MIAMI DADE SO | Date 04/16/2026 | Type Agency Inspection |
| Inspector AZRAK,JOSEPH, | Time 03:48:47 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 28424080A4 |
| Expiration 06/24/2027 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 11/05/2026 |

Number of Simulators 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.201 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.201 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

AI R CLOSIUS/R MYRTIL PRESENT

Data Download Date/Time 04/16/2026 04:00

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007107

| | | | |
|------------------------------------|------------------------------|-------------------------------|---------------------------------|
| Agency BROWARD COUNTY SO | Date 01/14/2026 | Type Agency Inspection | |
| Inspector FRAZIER, ANAYA, S | Time 10:48:56 | Software 8100.27 | |
| 0.05 Lot # 202402E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 402783528 |
| Expiration 02/27/2026 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 06/27/2026 |
| Number of Simulators 5 | Compliance Yes | | |

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

S. TILLMAN IN TRAINING.

Data Download Date/Time 01/14/2026 11:30

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007107

| | | |
|------------------------------------|------------------------|-------------------------------|
| Agency BROWARD COUNTY SO | Date 02/27/2026 | Type Agency Inspection |
| Inspector GUTSCHOW,TAYLOR,D | Time 15:46:12 | Software 8100.27 |

| | | | |
|-------------------|-------------------|-------------------|-----------------------|
| 0.05 Lot # | 0.08 Lot # | 0.20 Lot # | 0.08 Gas Lot # |
| Expiration | Expiration | Expiration | Expiration |

Number of Simulators **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|------|--------|
| Diagnostic Check (Pre-Inspection): OK | No | | |
| Date and/or Time Adjusted | No | | |
| Minimum Sample Volume Check: OK | | | |
| Barometric Pressure Sensor Check: OK | | | |
| Alcohol Free Subject Test: 0.000 | No | | |
| Mouth Alcohol Test: Slope Not Met | No | | |
| Interferent Detect Test: Interferent Detect | No | | |
| Diagnostic Check (Post-Inspection): OK | No | | |
| Alcohol Free Test: OK | No | | |
| 0.05 g/210L Test: OK | No | | |
| 0.08 g/210L Test: OK | No | | |
| 0.20 g/210L Test: OK | No | | |
| 0.08 g/210L Dry Gas Std Test: OK | No | | |

Remarks:

AI NOT CONDUCTED. COMPLIANCE NOT DETERMINED.

Data Download Date/Time 02/27/2026 15:44

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007107

Agency BROWARD COUNTY SO **Date** 03/05/2026 **Type** Department Inspection
Inspector GUTSCHOW,TAYLOR,D **Time** 13:25:07 **Software** 8100.27

0.05 Lot # 202406K **0.08 Lot #** 202406L **0.20 Lot #** 202406N **0.08 Gas Lot #** AG526603
Expiration 06/19/2026 **Expiration** 06/19/2026 **Expiration** 06/20/2026 **Expiration** 09/23/2027

Number of Simulators 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | Yes | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | Yes | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |
| | | Alcohol Free Test # 4 | 0.000 |
| | | 0.05 g/210L Test # 4 | 0.048 |
| | | 0.08 g/210L Test # 4 | 0.078 |
| | | 0.20 g/210L Test # 4 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 4 | 0.078 |
| | | Interferent Detect Test # 4 | INT |
| | | Alcohol Free Test # 5 | 0.000 |
| | | 0.05 g/210L Test # 5 | 0.049 |
| | | 0.08 g/210L Test # 5 | 0.079 |
| | | 0.20 g/210L Test # 5 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 5 | 0.078 |
| | | Interferent Detect Test # 5 | INT |

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007107

| | |
|--|--------|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.049 |
| 0.08 g/210L Test # 6 | 0.078 |
| 0.20 g/210L Test # 6 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.079 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.049 |
| 0.08 g/210L Test # 7 | 0.079 |
| 0.20 g/210L Test # 7 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.078 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.049 |
| 0.08 g/210L Test # 8 | 0.079 |
| 0.20 g/210L Test # 8 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.079 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.049 |
| 0.08 g/210L Test # 9 | 0.079 |
| 0.20 g/210L Test # 9 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.078 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.049 |
| 0.08 g/210L Test # 10 | 0.079 |
| 0.20 g/210L Test # 10 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.078 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Test (Std Dev) | 0.0004 |
| 0.20 g/210L Test (Std Dev) | 0.0005 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0004 |

Remarks:

Data Download Date/Time 03/05/2026 13:26

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007107

| | | |
|------------------------------------|------------------------|-------------------------------|
| Agency BROWARD COUNTY SO | Date 03/31/2026 | Type Agency Inspection |
| Inspector FRAZIER, ANAYA, S | Time 22:59:58 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|---------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 402783528 |
| Expiration 06/24/2027 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 06/27/2026 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.081 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.081 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

S. TILLMAN IN TRAINING.

Data Download Date/Time 03/31/2026 22:59

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007107

| | | |
|------------------------------------|------------------------|-------------------------------|
| Agency BROWARD COUNTY SO | Date 04/01/2026 | Type Agency Inspection |
| Inspector FRAZIER, ANAYA, S | Time 08:34:49 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|---------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 402783528 |
| Expiration 06/24/2027 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 06/27/2026 |

Number of Simulators 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

S. TILLMAN IN TRAINING.

Data Download Date/Time 04/01/2026 08:56

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007108

| | | |
|------------------------------------|------------------------|-------------------------------|
| Agency BROWARD COUNTY SO | Date 01/14/2026 | Type Agency Inspection |
| Inspector FRAZIER, ANAYA, S | Time 08:50:53 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|---------------------------------|
| 0.05 Lot # 202402E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 402783528 |
| Expiration 02/27/2026 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 06/27/2026 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

S. TILLMAN IN TRAINING.

Data Download Date/Time 01/14/2026 09:10

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007108

| | | |
|------------------------------------|------------------------|-------------------------------|
| Agency BROWARD COUNTY SO | Date 02/04/2026 | Type Agency Inspection |
| Inspector FRAZIER, ANAYA, S | Time 08:23:58 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|---------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 402783528 |
| Expiration 06/24/2027 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 06/27/2026 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

S. TILLMAN IN TRAINING.

Data Download Date/Time 02/04/2026 08:33

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007108

| | | |
|------------------------------------|------------------------|-------------------------------|
| Agency BROWARD COUNTY SO | Date 03/04/2026 | Type Agency Inspection |
| Inspector FRAZIER, ANAYA, S | Time 08:59:41 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|---------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 402783528 |
| Expiration 06/24/2027 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 06/27/2026 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

S. TILLMAN IN TRAINING.

Data Download Date/Time 03/10/2026 22:02

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007108

Agency BROWARD COUNTY SO **Date** 04/28/2026 **Type** Department Inspection
Inspector GUTSCHOW,TAYLOR,D **Time** 11:09:10 **Software** 8100.27

0.05 Lot # 202406K **0.08 Lot #** 202603B **0.20 Lot #** 202406N **0.08 Gas Lot #** AG429602
Expiration 06/19/2026 **Expiration** 03/03/2028 **Expiration** 06/20/2026 **Expiration** 10/22/2026

Number of Simulators 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | Yes | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | Yes | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |
| | | Alcohol Free Test # 4 | 0.000 |
| | | 0.05 g/210L Test # 4 | 0.048 |
| | | 0.08 g/210L Test # 4 | 0.079 |
| | | 0.20 g/210L Test # 4 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 4 | 0.080 |
| | | Interferent Detect Test # 4 | INT |
| | | Alcohol Free Test # 5 | 0.000 |
| | | 0.05 g/210L Test # 5 | 0.048 |
| | | 0.08 g/210L Test # 5 | 0.078 |
| | | 0.20 g/210L Test # 5 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 5 | 0.080 |
| | | Interferent Detect Test # 5 | INT |

Florida Department of Law Enforcement
Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007108

| | |
|--|--------|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.049 |
| 0.08 g/210L Test # 6 | 0.078 |
| 0.20 g/210L Test # 6 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.080 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.049 |
| 0.08 g/210L Test # 7 | 0.078 |
| 0.20 g/210L Test # 7 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.080 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.049 |
| 0.08 g/210L Test # 8 | 0.079 |
| 0.20 g/210L Test # 8 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.080 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.049 |
| 0.08 g/210L Test # 9 | 0.078 |
| 0.20 g/210L Test # 9 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.079 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.049 |
| 0.08 g/210L Test # 10 | 0.078 |
| 0.20 g/210L Test # 10 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.079 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0005 |
| 0.08 g/210L Test (Std Dev) | 0.0004 |
| 0.20 g/210L Test (Std Dev) | 0.0003 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0004 |

Remarks:

Data Download Date/Time 04/28/2026 11:27

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

| | | |
|----------------------------------|------------------------|-------------------------------|
| Agency PENSACOLA P.D. | Date 01/12/2026 | Type Agency Inspection |
| Inspector COLLINS,CHASE,M | Time 09:37:58 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|---------------------------------|
| 0.05 Lot # 202402E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 403342588 |
| Expiration 02/27/2026 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 05/09/2028 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.084 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.084 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.084 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 01/12/2026 10:41

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

Agency PENSACOLA P.D. **Date** 01/28/2026 **Type** Department Inspection
Inspector PIERSON, WEN-CHI, K **Time** 11:54:41 **Software** 8100.27

0.05 Lot # 202406K **0.08 Lot #** 202406L **0.20 Lot #** 202406N **0.08 Gas Lot #** AG510701
Expiration 06/19/2026 **Expiration** 06/19/2026 **Expiration** 06/20/2026 **Expiration** 04/17/2027

Number of Simulators 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | Yes | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | Yes | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |
| | | Alcohol Free Test # 4 | 0.000 |
| | | 0.05 g/210L Test # 4 | 0.048 |
| | | 0.08 g/210L Test # 4 | 0.079 |
| | | 0.20 g/210L Test # 4 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 4 | 0.080 |
| | | Interferent Detect Test # 4 | INT |
| | | Alcohol Free Test # 5 | 0.000 |
| | | 0.05 g/210L Test # 5 | 0.048 |
| | | 0.08 g/210L Test # 5 | 0.079 |
| | | 0.20 g/210L Test # 5 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 5 | 0.080 |
| | | Interferent Detect Test # 5 | INT |

Florida Department of Law Enforcement
Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

| | |
|--|--------|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.048 |
| 0.08 g/210L Test # 6 | 0.079 |
| 0.20 g/210L Test # 6 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.080 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.048 |
| 0.08 g/210L Test # 7 | 0.079 |
| 0.20 g/210L Test # 7 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.080 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.048 |
| 0.08 g/210L Test # 8 | 0.079 |
| 0.20 g/210L Test # 8 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.080 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.048 |
| 0.08 g/210L Test # 9 | 0.080 |
| 0.20 g/210L Test # 9 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.080 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.049 |
| 0.08 g/210L Test # 10 | 0.080 |
| 0.20 g/210L Test # 10 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.080 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0003 |
| 0.08 g/210L Test (Std Dev) | 0.0005 |
| 0.20 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0000 |

Remarks:

Data Download Date/Time 01/28/2026 13:02

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

| | | |
|----------------------------------|------------------------|-------------------------------|
| Agency PENSACOLA P.D. | Date 02/16/2026 | Type Agency Inspection |
| Inspector COLLINS,CHASE,M | Time 09:52:59 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|---------------------------------|
| 0.05 Lot # 202402E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 403135737 |
| Expiration 02/27/2026 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 08/30/2027 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 02/16/2026 10:50

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

| | | |
|----------------------------------|------------------------|-------------------------------|
| Agency PENSACOLA P.D. | Date 03/02/2026 | Type Agency Inspection |
| Inspector COLLINS,CHASE,M | Time 10:01:11 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|---------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 403135737 |
| Expiration 06/24/2027 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 08/30/2027 |

Number of Simulators 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.141 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.20 g/210L Test # (Repeat) 1 | 0.199 |
| Mouth Alcohol Test: Slope Not Met | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test: Interferent Detect | Yes | Interferent Detect Test # 1 | INT |
| Diagnostic Check (Post-Inspection): OK | Yes | Alcohol Free Test # 2 | 0.000 |
| Alcohol Free Test: OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| 0.05 g/210L Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.08 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.142 |
| 0.20 g/210L Test: OK | Yes | 0.20 g/210L Test # (Repeat) 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| | | Interferent Detect Test # 2 | INT |
| | | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.145 |
| | | 0.20 g/210L Test # (Repeat) 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

20: Control Outside Tolerance.

Data Download Date/Time 03/02/2026 10:59

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

| | | |
|----------------------------------|------------------------|-------------------------------|
| Agency PENSACOLA P.D. | Date 04/20/2026 | Type Agency Inspection |
| Inspector COLLINS,CHASE,M | Time 11:07:41 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|---------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 403342588 |
| Expiration 06/24/2027 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 05/09/2028 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 04/20/2026 12:09

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007160

| | | |
|--------------------------------------|------------------------|-------------------------------|
| Agency HARDEE COUNTY SO | Date 02/01/2026 | Type Agency Inspection |
| Inspector ALBRITTON,GARRETT,M | Time 00:55:57 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|---------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 403345775 |
| Expiration 06/24/2027 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 05/13/2028 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.194 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.196 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 02/01/2026 01:02

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007160

| | | |
|--------------------------------------|------------------------|-------------------------------|
| Agency HARDEE COUNTY SO | Date 02/28/2026 | Type Agency Inspection |
| Inspector ALBRITTON,GARRETT,M | Time 16:00:26 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|---------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202510E | 0.20 Lot # 202410F | 0.08 Gas Lot # 403345775 |
| Expiration 06/24/2027 | Expiration 10/28/2027 | Expiration 10/30/2026 | Expiration 05/13/2028 |

Number of Simulators 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.045 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.075 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.189 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.20 g/210L Test # (Repeat) 1 | 0.192 |
| Mouth Alcohol Test: Slope Not Met | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test: Interferent Detect | Yes | Interferent Detect Test # 1 | INT |
| Diagnostic Check (Post-Inspection): OK | Yes | Alcohol Free Test # 2 | 0.000 |
| Alcohol Free Test: OK | Yes | 0.05 g/210L Test # 2 | 0.045 |
| 0.05 g/210L Test: OK | Yes | 0.08 g/210L Test # 2 | 0.076 |
| 0.08 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.190 |
| 0.20 g/210L Test: OK | Yes | 0.20 g/210L Test # (Repeat) 2 | 0.194 |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| | | Interferent Detect Test # 2 | INT |
| | | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.046 |
| | | 0.08 g/210L Test # 3 | 0.076 |
| | | 0.20 g/210L Test # 3 | 0.192 |
| | | 0.20 g/210L Test # (Repeat) 3 | 0.194 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

20: Control Outside Tolerance.TEMPERATURE IN THE ROOM.

Data Download Date/Time 02/28/2026 15:58

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007160

| | | |
|--------------------------------------|------------------------|-------------------------------|
| Agency HARDEE COUNTY SO | Date 03/31/2026 | Type Agency Inspection |
| Inspector ALBRITTON,GARRETT,M | Time 23:43:21 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|---------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202510E | 0.20 Lot # 202410F | 0.08 Gas Lot # 403345775 |
| Expiration 06/24/2027 | Expiration 10/28/2027 | Expiration 10/30/2026 | Expiration 05/13/2028 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.046 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.076 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.192 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.046 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.194 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.047 |
| | | 0.08 g/210L Test # 3 | 0.077 |
| | | 0.20 g/210L Test # 3 | 0.195 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 03/31/2026 23:43

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

| | | |
|----------------------------------|------------------------|-------------------------------|
| Agency HILLSBOROUGH CSO | Date 01/15/2026 | Type Agency Inspection |
| Inspector BROWN, KEVIN, C | Time 10:30:06 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 04125080A6 |
| Expiration 06/24/2027 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 03/05/2027 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

AI B KROME #263389 PRESENT FOR INSPECTION

Data Download Date/Time 01/15/2026 11:27

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

| | | |
|--------------------------------|------------------------|-------------------------------|
| Agency HILLSBOROUGH CSO | Date 02/12/2026 | Type Agency Inspection |
| Inspector BROWN,KEVIN,C | Time 09:11:06 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 04125080A6 |
| Expiration 06/24/2027 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 03/05/2027 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.046 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.047 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.047 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

AI B KROME #263389 PRESENT FOR INSPECTION.

Data Download Date/Time 02/12/2026 11:02

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

| | | |
|--------------------------------|------------------------|-------------------------------|
| Agency HILLSBOROUGH CSO | Date 03/12/2026 | Type Agency Inspection |
| Inspector BROWN,KEVIN,C | Time 08:50:33 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 04125080A6 |
| Expiration 06/24/2027 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 03/05/2027 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.194 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.077 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.077 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.195 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.077 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

AI B KROME #263389 PRESENT FOR INSPECTION

Data Download Date/Time 03/12/2026 11:34

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

| | | |
|----------------------------------|------------------------|-------------------------------|
| Agency HILLSBOROUGH CSO | Date 04/09/2026 | Type Agency Inspection |
| Inspector BROWN, KEVIN, C | Time 08:31:34 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 04125080A6 |
| Expiration 06/24/2027 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 03/05/2027 |

Number of Simulators 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.195 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.196 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

AI B KROME #263389 PRESENT FOR INSPECTION

Data Download Date/Time 04/09/2026 09:51

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007165

| | | | |
|-----------------------------------|------------------------------|-------------------------------|---------------------------------|
| Agency FFWCC | Date 01/30/2026 | Type Agency Inspection | |
| Inspector CONROY,JESSICA,P | Time 20:26:03 | Software 8100.27 | |
| 0.05 Lot # 202506E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 403207508 |
| Expiration 06/24/2027 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 12/02/2027 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.196 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.196 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 02/05/2026 11:35

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007165

| | | |
|-----------------------------------|------------------------|-------------------------------|
| Agency FFWCC | Date 03/24/2026 | Type Agency Inspection |
| Inspector CONROY,JESSICA,P | Time 14:45:41 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|---------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 403207508 |
| Expiration 06/24/2027 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 12/02/2027 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 03/30/2026 11:17

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007165

| | | |
|------------------------------------|------------------------|-----------------------------------|
| Agency FFWCC | Date 04/13/2026 | Type Department Inspection |
| Inspector GUTSCHOW,TAYLOR,D | Time 14:08:34 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| 0.05 Lot # 202406K | 0.08 Lot # 202406L | 0.20 Lot # 202406N | 0.08 Gas Lot # AG526603 |
| Expiration 06/19/2026 | Expiration 06/19/2026 | Expiration 06/20/2026 | Expiration 09/23/2027 |

| | | |
|-------------------------------|--|-----------------------|
| Number of Simulators 5 | Average Standard Deviation 0.0004 | Compliance Yes |
|-------------------------------|--|-----------------------|

| Check or Test | Result | Test | Result |
|---|--------|---|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | Yes | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | Yes | 0.20 g/210L Test # 1 | 0.201 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.000 |
| Mouth Alcohol Test: Slope Not Met | Yes | 0.08 g/210L Dry Gas Std Test # (Repeat) 1 | 0.081 |
| Interferent Detect Test: Interferent Detect | Yes | Interferent Detect Test # 1 | INT |
| Diagnostic Check (Post-Inspection): OK | Yes | Alcohol Free Test # 2 | 0.000 |
| Alcohol Free Test: OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| 0.05 g/210L Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.08 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.200 |
| 0.20 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.000 |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # (Repeat) 2 | 0.080 |
| | | Interferent Detect Test # 2 | INT |
| | | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.000 |
| | | 0.08 g/210L Dry Gas Std Test # (Repeat) 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |
| | | Alcohol Free Test # 4 | 0.000 |
| | | 0.05 g/210L Test # 4 | 0.049 |
| | | 0.08 g/210L Test # 4 | 0.079 |
| | | 0.20 g/210L Test # 4 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 4 | 0.000 |
| | | 0.08 g/210L Dry Gas Std Test # (Repeat) 4 | 0.081 |
| | | Interferent Detect Test # 4 | INT |
| | | Alcohol Free Test # 5 | 0.000 |
| | | 0.05 g/210L Test # 5 | 0.050 |

Florida Department of Law Enforcement
Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007165

| | |
|--|-------|
| 0.08 g/210L Test # 5 | 0.079 |
| 0.20 g/210L Test # 5 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.000 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 5 | 0.080 |
| Interferent Detect Test # 5 | INT |
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.050 |
| 0.08 g/210L Test # 6 | 0.079 |
| 0.20 g/210L Test # 6 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.000 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 6 | 0.080 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.049 |
| 0.08 g/210L Test # 7 | 0.079 |
| 0.20 g/210L Test # 7 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.000 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 7 | 0.080 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.049 |
| 0.08 g/210L Test # 8 | 0.079 |
| 0.20 g/210L Test # 8 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.000 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 8 | 0.080 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.050 |
| 0.08 g/210L Test # 9 | 0.079 |
| 0.20 g/210L Test # 9 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.000 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 9 | 0.080 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.050 |
| 0.08 g/210L Test # 10 | 0.080 |
| 0.20 g/210L Test # 10 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.000 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 10 | 0.080 |
| Interferent Detect Test # 10 | INT |

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

**Florida Department of Law Enforcement
Alcohol Testing Program**

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007165

| | |
|--|--------|
| 0.05 g/210L Test (Std Dev) | 0.0005 |
| 0.08 g/210L Test (Std Dev) | 0.0003 |
| 0.20 g/210L Test (Std Dev) | 0.0006 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0000 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) (Repeat) | 0.0004 |

Remarks:

08: Control Outside Tolerance. DGS HOSE NOT PLUGGED IN..

Data Download Date/Time 04/13/2026 14:07

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007165

| | | | |
|-----------------------------------|------------------------------|-------------------------------|---------------------------------|
| Agency FFWCC | Date 04/22/2026 | Type Agency Inspection | |
| Inspector SAAVEDRA, OSCAR, | Time 16:07:07 | Software 8100.27 | |
| 0.05 Lot # 202506E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 403207508 |
| Expiration 06/24/2027 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 12/02/2027 |
| Number of Simulators 5 | | Compliance Yes | |

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.051 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 04/22/2026 16:49

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007166

| | | |
|------------------------------------|------------------------|-------------------------------|
| Agency FFWCC | Date 03/06/2026 | Type Agency Inspection |
| Inspector GUTSCHOW,TAYLOR,D | Time 09:20:10 | Software 8100.27 |

| | | | |
|-------------------|-------------------|-------------------|-----------------------|
| 0.05 Lot # | 0.08 Lot # | 0.20 Lot # | 0.08 Gas Lot # |
| Expiration | Expiration | Expiration | Expiration |

Number of Simulators **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|------|--------|
| Diagnostic Check (Pre-Inspection): OK | No | | |
| Date and/or Time Adjusted | No | | |
| Minimum Sample Volume Check: OK | | | |
| Barometric Pressure Sensor Check: OK | | | |
| Alcohol Free Subject Test: 0.000 | No | | |
| Mouth Alcohol Test: Slope Not Met | No | | |
| Interferent Detect Test: Interferent Detect | No | | |
| Diagnostic Check (Post-Inspection): OK | No | | |
| Alcohol Free Test: OK | No | | |
| 0.05 g/210L Test: OK | No | | |
| 0.08 g/210L Test: OK | No | | |
| 0.20 g/210L Test: OK | No | | |
| 0.08 g/210L Dry Gas Std Test: OK | No | | |

Remarks:

AI NOT CONDUCTED. COMPLIANCE NOT DETERMINED.

Data Download Date/Time 03/06/2026 14:10

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007166

| | | |
|------------------------------------|------------------------|-----------------------------------|
| Agency FFWCC | Date 03/06/2026 | Type Department Inspection |
| Inspector GUTSCHOW,TAYLOR,D | Time 14:05:43 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| 0.05 Lot # 202406K | 0.08 Lot # 202406L | 0.20 Lot # 202406N | 0.08 Gas Lot # AG526603 |
| Expiration 06/19/2026 | Expiration 06/19/2026 | Expiration 06/20/2026 | Expiration 09/23/2027 |

Number of Simulators 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | Yes | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | Yes | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |
| | | Alcohol Free Test # 4 | 0.000 |
| | | 0.05 g/210L Test # 4 | 0.049 |
| | | 0.08 g/210L Test # 4 | 0.079 |
| | | 0.20 g/210L Test # 4 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 4 | 0.078 |
| | | Interferent Detect Test # 4 | INT |
| | | Alcohol Free Test # 5 | 0.000 |
| | | 0.05 g/210L Test # 5 | 0.049 |
| | | 0.08 g/210L Test # 5 | 0.079 |
| | | 0.20 g/210L Test # 5 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 5 | 0.078 |
| | | Interferent Detect Test # 5 | INT |

Florida Department of Law Enforcement
Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007166

| | |
|--|--------|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.049 |
| 0.08 g/210L Test # 6 | 0.080 |
| 0.20 g/210L Test # 6 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.078 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.049 |
| 0.08 g/210L Test # 7 | 0.080 |
| 0.20 g/210L Test # 7 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.078 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.048 |
| 0.08 g/210L Test # 8 | 0.080 |
| 0.20 g/210L Test # 8 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.078 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.049 |
| 0.08 g/210L Test # 9 | 0.080 |
| 0.20 g/210L Test # 9 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.078 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.049 |
| 0.08 g/210L Test # 10 | 0.079 |
| 0.20 g/210L Test # 10 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.078 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Test (Std Dev) | 0.0005 |
| 0.20 g/210L Test (Std Dev) | 0.0003 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0000 |

Remarks:

Data Download Date/Time 03/06/2026 14:10

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

| | | |
|-----------------------------------|------------------------|-------------------------------|
| Agency MONROE COUNTY SO | Date 01/22/2026 | Type Agency Inspection |
| Inspector ZITO, ANTHONY, P | Time 12:26:27 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|---------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 402843437 |
| Expiration 06/24/2027 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 09/08/2026 |

Number of Simulators 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.195 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.077 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.196 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

POST AI. OOS GOING TO FDLE FOR DEP INSPECTION

Data Download Date/Time 01/22/2026 12:24

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

| | | |
|------------------------------------|------------------------|-------------------------------|
| Agency MONROE COUNTY SO | Date 03/06/2026 | Type Agency Inspection |
| Inspector GUTSCHOW,TAYLOR,D | Time 09:26:06 | Software 8100.27 |

| | | | |
|-------------------|-------------------|-------------------|-----------------------|
| 0.05 Lot # | 0.08 Lot # | 0.20 Lot # | 0.08 Gas Lot # |
| Expiration | Expiration | Expiration | Expiration |

Number of Simulators **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|------|--------|
| Diagnostic Check (Pre-Inspection): OK | No | | |
| Date and/or Time Adjusted | No | | |
| Minimum Sample Volume Check: OK | | | |
| Barometric Pressure Sensor Check: OK | | | |
| Alcohol Free Subject Test: 0.000 | No | | |
| Mouth Alcohol Test: Slope Not Met | No | | |
| Interferent Detect Test: Interferent Detect | No | | |
| Diagnostic Check (Post-Inspection): OK | No | | |
| Alcohol Free Test: OK | No | | |
| 0.05 g/210L Test: OK | No | | |
| 0.08 g/210L Test: OK | No | | |
| 0.20 g/210L Test: OK | No | | |
| 0.08 g/210L Dry Gas Std Test: OK | No | | |

Remarks:

AI NOT CONDUCTED. COMPLIANCE NOT DETERMINED.

Data Download Date/Time 03/06/2026 13:46

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

Agency MONROE COUNTY SO **Date** 03/06/2026 **Type** Department Inspection
Inspector GUTSCHOW,TAYLOR,D **Time** 13:48:42 **Software** 8100.27

0.05 Lot # 202406K **0.08 Lot #** 202406L **0.20 Lot #** 202406N **0.08 Gas Lot #** AG526603
Expiration 06/19/2026 **Expiration** 06/19/2026 **Expiration** 06/20/2026 **Expiration** 09/23/2027

Number of Simulators 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | Yes | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | Yes | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.047 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.077 |
| | | Interferent Detect Test # 3 | INT |
| | | Alcohol Free Test # 4 | 0.000 |
| | | 0.05 g/210L Test # 4 | 0.048 |
| | | 0.08 g/210L Test # 4 | 0.079 |
| | | 0.20 g/210L Test # 4 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 4 | 0.078 |
| | | Interferent Detect Test # 4 | INT |
| | | Alcohol Free Test # 5 | 0.000 |
| | | 0.05 g/210L Test # 5 | 0.048 |
| | | 0.08 g/210L Test # 5 | 0.079 |
| | | 0.20 g/210L Test # 5 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 5 | 0.078 |
| | | Interferent Detect Test # 5 | INT |

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

| | |
|--|--------|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.048 |
| 0.08 g/210L Test # 6 | 0.079 |
| 0.20 g/210L Test # 6 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.077 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.048 |
| 0.08 g/210L Test # 7 | 0.079 |
| 0.20 g/210L Test # 7 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.077 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.048 |
| 0.08 g/210L Test # 8 | 0.079 |
| 0.20 g/210L Test # 8 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.078 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.048 |
| 0.08 g/210L Test # 9 | 0.079 |
| 0.20 g/210L Test # 9 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.078 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.048 |
| 0.08 g/210L Test # 10 | 0.079 |
| 0.20 g/210L Test # 10 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.077 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0003 |
| 0.08 g/210L Test (Std Dev) | 0.0000 |
| 0.20 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0005 |

Remarks:

Data Download Date/Time 03/06/2026 13:46

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

| | | |
|-----------------------------------|------------------------|-------------------------------|
| Agency MONROE COUNTY SO | Date 03/25/2026 | Type Agency Inspection |
| Inspector ZITO, ANTHONY, P | Time 14:46:43 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|---------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 403411042 |
| Expiration 06/24/2027 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 08/02/2028 |

Number of Simulators 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|---|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.193 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.000 |
| Mouth Alcohol Test: Slope Not Met | Yes | 0.08 g/210L Dry Gas Std Test # (Repeat) 1 | 0.079 |
| Interferent Detect Test: Interferent Detect | Yes | Interferent Detect Test # 1 | INT |
| Diagnostic Check (Post-Inspection): OK | Yes | Alcohol Free Test # 2 | 0.000 |
| Alcohol Free Test: OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| 0.05 g/210L Test: OK | Yes | 0.08 g/210L Test # 2 | 0.077 |
| 0.08 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.194 |
| 0.20 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.000 |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # (Repeat) 2 | 0.079 |
| | | Interferent Detect Test # 2 | INT |
| | | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.195 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.000 |
| | | 0.08 g/210L Dry Gas Std Test # (Repeat) 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

08: Control Outsid.HOSE ON SIDE WAS NOT PLUGGED IN INTOX . PRE EVIDENTIARY AI. IN SERVICE

Data Download Date/Time 03/25/2026 15:09

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

| | | |
|-----------------------------------|------------------------|-------------------------------|
| Agency MONROE COUNTY SO | Date 04/08/2026 | Type Agency Inspection |
| Inspector ZITO, ANTHONY, P | Time 19:49:39 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|---------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202510E | 0.20 Lot # 202510F | 0.08 Gas Lot # 403207508 |
| Expiration 06/24/2027 | Expiration 10/28/2027 | Expiration 10/29/2027 | Expiration 12/02/2027 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.191 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.192 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.192 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.076 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

PRE EVIDENTIARY AI

Data Download Date/Time 04/08/2026 19:46

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007168

| | | |
|-----------------------------------|------------------------|-------------------------------|
| Agency FFWCC | Date 01/21/2026 | Type Agency Inspection |
| Inspector SAAVEDRA, OSCAR, | Time 17:14:11 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|---------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 403056709 |
| Expiration 06/24/2027 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 05/28/2027 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.196 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.077 |
| | | 0.20 g/210L Test # 3 | 0.196 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 01/21/2026 17:16

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007168

| | | | |
|-----------------------------------|------------------------------|-------------------------------|---------------------------------|
| Agency FFWCC | Date 02/12/2026 | Type Agency Inspection | |
| Inspector SAAVEDRA, OSCAR, | Time 17:34:23 | Software 8100.27 | |
| 0.05 Lot # 202506E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 403056709 |
| Expiration 06/24/2027 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 05/28/2027 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.196 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.196 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 02/12/2026 18:01

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007168

| | | |
|-------------------------------------|------------------------|-------------------------------|
| Agency FFWCC | Date 03/19/2026 | Type Agency Inspection |
| Inspector GORNEY, WILLIAM, M | Time 11:54:29 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|---------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 403056709 |
| Expiration 06/24/2027 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 05/28/2027 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.196 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.077 |
| | | 0.20 g/210L Test # 3 | 0.195 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 03/19/2026 12:11

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007168

| | | |
|-------------------------------------|------------------------|-------------------------------|
| Agency FFWCC | Date 04/21/2026 | Type Agency Inspection |
| Inspector GORNEY, WILLIAM, M | Time 12:26:57 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|---------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 403056709 |
| Expiration 06/24/2027 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 05/28/2027 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.195 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.196 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 04/21/2026 12:40

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007169

| | | |
|--------------------------------------|------------------------|-------------------------------|
| Agency FFWCC | Date 01/15/2026 | Type Agency Inspection |
| Inspector ARBOGAST, JOSHUA, S | Time 12:57:55 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|---------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202510E | 0.20 Lot # 202410F | 0.08 Gas Lot # 403008479 |
| Expiration 06/24/2027 | Expiration 10/28/2027 | Expiration 10/30/2026 | Expiration 03/28/2027 |

Number of Simulators 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.200 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.077 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.201 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.077 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 01/15/2026 13:08

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007169

| | | | |
|--------------------------------------|------------------------------|-------------------------------|---------------------------------|
| Agency FFWCC | Date 02/23/2026 | Type Agency Inspection | |
| Inspector ARBOGAST, JOSHUA, S | Time 11:31:54 | Software 8100.27 | |
| 0.05 Lot # 202506E | 0.08 Lot # 202510E | 0.20 Lot # 202410F | 0.08 Gas Lot # 403008479 |
| Expiration 06/24/2027 | Expiration 10/28/2027 | Expiration 10/30/2026 | Expiration 03/28/2027 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.081 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.201 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.051 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.081 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.202 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.051 |
| | | 0.08 g/210L Test # 3 | 0.082 |
| | | 0.20 g/210L Test # 3 | 0.203 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 02/23/2026 11:44

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007169

| | | | |
|--------------------------------------|------------------------------|-------------------------------|---------------------------------|
| Agency FFWCC | Date 03/28/2026 | Type Agency Inspection | |
| Inspector ARBOGAST, JOSHUA, S | Time 16:16:26 | Software 8100.27 | |
| 0.05 Lot # 202506E | 0.08 Lot # 202510E | 0.20 Lot # 202410F | 0.08 Gas Lot # 403008479 |
| Expiration 06/24/2027 | Expiration 10/28/2027 | Expiration 10/30/2026 | Expiration 03/28/2027 |
| Number of Simulators 5 | | Compliance Yes | |

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.201 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.202 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.202 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

IN COMPLIANCE

Data Download Date/Time 03/28/2026 16:28

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007169

| | | |
|--------------------------------------|------------------------|-------------------------------|
| Agency FFWCC | Date 04/22/2026 | Type Agency Inspection |
| Inspector ARBOGAST, JOSHUA, S | Time 11:33:51 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|---------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202510E | 0.20 Lot # 202410F | 0.08 Gas Lot # 403008479 |
| Expiration 06/24/2027 | Expiration 10/28/2027 | Expiration 10/30/2026 | Expiration 03/28/2027 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.200 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.201 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 04/22/2026 12:05

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

| | | |
|------------------------------------|------------------------|-------------------------------|
| Agency MONROE COUNTY SO | Date 02/27/2026 | Type Agency Inspection |
| Inspector GUTSCHOW,TAYLOR,D | Time 13:35:29 | Software 8100.27 |

| | | | |
|-------------------|-------------------|-------------------|-----------------------|
| 0.05 Lot # | 0.08 Lot # | 0.20 Lot # | 0.08 Gas Lot # |
| Expiration | Expiration | Expiration | Expiration |

Number of Simulators **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|------|--------|
| Diagnostic Check (Pre-Inspection): OK | No | | |
| Date and/or Time Adjusted | No | | |
| Minimum Sample Volume Check: OK | | | |
| Barometric Pressure Sensor Check: OK | | | |
| Alcohol Free Subject Test: 0.000 | No | | |
| Mouth Alcohol Test: Slope Not Met | No | | |
| Interferent Detect Test: Interferent Detect | No | | |
| Diagnostic Check (Post-Inspection): OK | No | | |
| Alcohol Free Test: OK | No | | |
| 0.05 g/210L Test: OK | No | | |
| 0.08 g/210L Test: OK | No | | |
| 0.20 g/210L Test: OK | No | | |
| 0.08 g/210L Dry Gas Std Test: OK | No | | |

Remarks:

AI NOT CONDUCTED. COMPLIANCE NOT DETERMINED.

Data Download Date/Time 02/27/2026 15:40

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

Agency MONROE COUNTY SO **Date** 03/05/2026 **Type** Department Inspection
Inspector GUTSCHOW,TAYLOR,D **Time** 13:06:26 **Software** 8100.27

0.05 Lot # 202406K **0.08 Lot #** 202406L **0.20 Lot #** 202406N **0.08 Gas Lot #** AG526603
Expiration 06/19/2026 **Expiration** 06/19/2026 **Expiration** 06/20/2026 **Expiration** 09/23/2027

Number of Simulators 5 **Average Standard Deviation** 0.0002 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | Yes | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | Yes | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |
| | | Alcohol Free Test # 4 | 0.000 |
| | | 0.05 g/210L Test # 4 | 0.050 |
| | | 0.08 g/210L Test # 4 | 0.080 |
| | | 0.20 g/210L Test # 4 | 0.196 |
| | | 0.08 g/210L Dry Gas Std Test # 4 | 0.079 |
| | | Interferent Detect Test # 4 | INT |
| | | Alcohol Free Test # 5 | 0.000 |
| | | 0.05 g/210L Test # 5 | 0.050 |
| | | 0.08 g/210L Test # 5 | 0.080 |
| | | 0.20 g/210L Test # 5 | 0.196 |
| | | 0.08 g/210L Dry Gas Std Test # 5 | 0.079 |
| | | Interferent Detect Test # 5 | INT |

Florida Department of Law Enforcement
Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

| | |
|--|--------|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.049 |
| 0.08 g/210L Test # 6 | 0.080 |
| 0.20 g/210L Test # 6 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.079 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.050 |
| 0.08 g/210L Test # 7 | 0.080 |
| 0.20 g/210L Test # 7 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.079 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.050 |
| 0.08 g/210L Test # 8 | 0.080 |
| 0.20 g/210L Test # 8 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.079 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.050 |
| 0.08 g/210L Test # 9 | 0.080 |
| 0.20 g/210L Test # 9 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.079 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.050 |
| 0.08 g/210L Test # 10 | 0.080 |
| 0.20 g/210L Test # 10 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.080 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Test (Std Dev) | 0.0000 |
| 0.20 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0003 |

Remarks:

Data Download Date/Time 03/05/2026 13:02

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

| | | |
|-----------------------------------|------------------------|-------------------------------|
| Agency MONROE COUNTY SO | Date 03/13/2026 | Type Agency Inspection |
| Inspector ZITO, ANTHONY, P | Time 18:13:34 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|---------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 403207508 |
| Expiration 06/24/2027 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 12/02/2027 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.194 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.051 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.052 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.196 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

PRE EVIDENTIARY AL. KV MAIN

Data Download Date/Time 03/13/2026 18:09

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

| | | |
|-----------------------------------|------------------------|-------------------------------|
| Agency MONROE COUNTY SO | Date 04/08/2026 | Type Agency Inspection |
| Inspector ZITO, ANTHONY, P | Time 19:34:41 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|---------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202510E | 0.20 Lot # 202510F | 0.08 Gas Lot # 403207508 |
| Expiration 06/24/2027 | Expiration 10/28/2027 | Expiration 10/29/2027 | Expiration 12/02/2027 |

Number of Simulators 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.076 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.188 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.20 g/210L Test # (Repeat) 1 | 0.192 |
| Mouth Alcohol Test: Slope Not Met | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test: Interferent Detect | Yes | Interferent Detect Test # 1 | INT |
| Diagnostic Check (Post-Inspection): OK | Yes | Alcohol Free Test # 2 | 0.000 |
| Alcohol Free Test: OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| 0.05 g/210L Test: OK | Yes | 0.08 g/210L Test # 2 | 0.077 |
| 0.08 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.191 |
| 0.20 g/210L Test: OK | Yes | 0.20 g/210L Test # (Repeat) 2 | 0.192 |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| | | Interferent Detect Test # 2 | INT |
| | | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.077 |
| | | 0.20 g/210L Test # 3 | 0.191 |
| | | 0.20 g/210L Test # (Repeat) 3 | 0.193 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

20: Control Outside Tolerance. SIM HOSE WAS LOOSE.

Data Download Date/Time 04/08/2026 19:30

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

| | | | |
|--------------------------------------|--|-----------------------------------|----------------------------------|
| Agency MARION COUNTY SO | Date 03/23/2026 | Type Department Inspection | |
| Inspector PIERSON, WEN-CHI, K | Time 13:56:38 | Software 8100.27 | |
| 0.05 Lot # 202406K | 0.08 Lot # 202406L | 0.20 Lot # 202406N | 0.08 Gas Lot # 23725080A1 |
| Expiration 06/19/2026 | Expiration 06/19/2026 | Expiration 06/20/2026 | Expiration 10/05/2027 |
| Number of Simulators 5 | Average Standard Deviation 0.0004 | Compliance Yes | |

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.047 |
| Minimum Sample Volume Check: OK | Yes | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | Yes | 0.20 g/210L Test # 1 | 0.200 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.047 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.077 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.047 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.077 |
| | | Interferent Detect Test # 3 | INT |
| | | Alcohol Free Test # 4 | 0.000 |
| | | 0.05 g/210L Test # 4 | 0.047 |
| | | 0.08 g/210L Test # 4 | 0.078 |
| | | 0.20 g/210L Test # 4 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 4 | 0.076 |
| | | Interferent Detect Test # 4 | INT |
| | | Alcohol Free Test # 5 | 0.000 |
| | | 0.05 g/210L Test # 5 | 0.047 |
| | | 0.08 g/210L Test # 5 | 0.078 |
| | | 0.20 g/210L Test # 5 | 0.201 |
| | | 0.08 g/210L Dry Gas Std Test # 5 | 0.076 |
| | | Interferent Detect Test # 5 | INT |

Florida Department of Law Enforcement
Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

| | |
|--|--------|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.047 |
| 0.08 g/210L Test # 6 | 0.078 |
| 0.20 g/210L Test # 6 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.076 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.047 |
| 0.08 g/210L Test # 7 | 0.078 |
| 0.20 g/210L Test # 7 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.076 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.047 |
| 0.08 g/210L Test # 8 | 0.078 |
| 0.20 g/210L Test # 8 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.075 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.047 |
| 0.08 g/210L Test # 9 | 0.078 |
| 0.20 g/210L Test # 9 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.076 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.047 |
| 0.08 g/210L Test # 10 | 0.079 |
| 0.20 g/210L Test # 10 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.075 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0000 |
| 0.08 g/210L Test (Std Dev) | 0.0003 |
| 0.20 g/210L Test (Std Dev) | 0.0005 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0009 |

Remarks:

Data Download Date/Time 03/23/2026 13:54

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

| | | |
|---------------------------------|------------------------|-------------------------------|
| Agency MARION COUNTY SO | Date 04/08/2026 | Type Agency Inspection |
| Inspector YOUNG,ROBERT,A | Time 20:27:58 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|---------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202510E | 0.20 Lot # 202510F | 0.08 Gas Lot # 403549457 |
| Expiration 06/24/2027 | Expiration 10/28/2027 | Expiration 10/29/2027 | Expiration 01/29/2029 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.046 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.075 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.047 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.076 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.076 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.077 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

IN COMPLIANCE

Data Download Date/Time 04/08/2026 20:26

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

| | | |
|---------------------------------|------------------------|-------------------------------|
| Agency MARION COUNTY SO | Date 01/08/2026 | Type Agency Inspection |
| Inspector YOUNG,ROBERT,A | Time 11:40:48 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|---------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 403369407 |
| Expiration 06/24/2027 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 06/23/2028 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # (Repeat) 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.047 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.193 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # (Repeat) 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # (Repeat) 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

IN COMPLIANCE

Data Download Date/Time 01/08/2026 11:52

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

| | | |
|---------------------------------|------------------------|-------------------------------|
| Agency MARION COUNTY SO | Date 02/20/2026 | Type Agency Inspection |
| Inspector YOUNG,ROBERT,A | Time 16:19:52 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|---------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 403369407 |
| Expiration 06/24/2027 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 06/23/2028 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.077 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.077 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

IN COMPLIANCE

Data Download Date/Time 02/23/2026 16:06

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

| | | |
|---------------------------------|------------------------|-------------------------------|
| Agency MARION COUNTY SO | Date 03/25/2026 | Type Agency Inspection |
| Inspector YOUNG,ROBERT,A | Time 20:47:56 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|---------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202410E | 0.20 Lot # 202510F | 0.08 Gas Lot # 403549457 |
| Expiration 06/24/2027 | Expiration 10/22/2026 | Expiration 10/29/2027 | Expiration 01/29/2029 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # (Repeat) 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.047 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.071 |
| Barometric Pressure Sensor Check: OK | | 0.08 g/210L Test # (Repeat) 1 | 0.079 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.20 g/210L Test # 1 | 0.195 |
| Mouth Alcohol Test: Slope Not Met | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test: Interferent Detect | Yes | Interferent Detect Test # 1 | INT |
| Diagnostic Check (Post-Inspection): OK | Yes | Alcohol Free Test # (Repeat) 2 | 0.000 |
| Alcohol Free Test: OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| 0.05 g/210L Test: OK | Yes | 0.08 g/210L Test # 2 | 0.071 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Test # (Repeat) 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| | | Interferent Detect Test # 2 | INT |
| | | Alcohol Free Test # (Repeat) 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.071 |
| | | 0.08 g/210L Test # (Repeat) 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.196 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

IN COMPLIANCE

Data Download Date/Time 03/25/2026 20:45

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

| | | |
|---------------------------------|------------------------|-------------------------------|
| Agency MARION COUNTY SO | Date 04/08/2026 | Type Agency Inspection |
| Inspector YOUNG,ROBERT,A | Time 20:01:29 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|---------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202510E | 0.20 Lot # 202510F | 0.08 Gas Lot # 403549457 |
| Expiration 06/24/2027 | Expiration 10/28/2027 | Expiration 10/29/2027 | Expiration 01/29/2029 |

Number of Simulators 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.047 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.195 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

IN COMPLIANCE

Data Download Date/Time 04/08/2026 19:58

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007173

| | | |
|----------------------------------|------------------------|-------------------------------|
| Agency WALTON COUNTY SO | Date 01/06/2026 | Type Agency Inspection |
| Inspector TULLEY,DUSTIN,R | Time 12:39:43 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|---------------------------------|
| 0.05 Lot # 202402E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 403135737 |
| Expiration 02/27/2026 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 08/30/2027 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.047 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.047 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 01/07/2026 13:31

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007173

| | | |
|----------------------------------|------------------------|-------------------------------|
| Agency WALTON COUNTY SO | Date 02/05/2026 | Type Agency Inspection |
| Inspector TULLEY,DUSTIN,R | Time 16:28:46 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|---------------------------------|
| 0.05 Lot # 202402E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 403135737 |
| Expiration 02/27/2026 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 08/30/2027 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|---|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.047 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.200 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.000 |
| Mouth Alcohol Test: Slope Not Met | Yes | 0.08 g/210L Dry Gas Std Test # (Repeat) 1 | 0.081 |
| Interferent Detect Test: Interferent Detect | Yes | Interferent Detect Test # 1 | INT |
| Diagnostic Check (Post-Inspection): OK | Yes | Alcohol Free Test # 2 | 0.000 |
| Alcohol Free Test: OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| 0.05 g/210L Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.08 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.200 |
| 0.20 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.000 |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # (Repeat) 2 | 0.081 |
| | | Interferent Detect Test # 2 | INT |
| | | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.000 |
| | | 0.08 g/210L Dry Gas Std Test # (Repeat) 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

08: Control Outside Tolerance.

Data Download Date/Time 02/05/2026 17:26

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007173

| | | |
|----------------------------------|------------------------|-------------------------------|
| Agency WALTON COUNTY SO | Date 03/11/2026 | Type Agency Inspection |
| Inspector TULLEY,DUSTIN,R | Time 16:04:27 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|---------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 403135737 |
| Expiration 06/24/2027 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 08/30/2027 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # (Repeat) 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.047 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.196 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # (Repeat) 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.047 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.077 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # (Repeat) 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

00: Ambient Fail. TURNED FAN ON. REPEATED TEST.

Data Download Date/Time 03/11/2026 17:01

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007173

| | | |
|----------------------------------|------------------------|-------------------------------|
| Agency WALTON COUNTY SO | Date 04/18/2026 | Type Agency Inspection |
| Inspector TULLEY,DUSTIN,R | Time 15:20:08 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|---------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 403135737 |
| Expiration 06/24/2027 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 08/30/2027 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 04/18/2026 16:17

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007252

| | | |
|----------------------------------|------------------------|-------------------------------|
| Agency GROVELAND PD | Date 01/18/2026 | Type Agency Inspection |
| Inspector MAVES,JEFFREY,M | Time 17:37:24 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 25024080A3 |
| Expiration 06/24/2027 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 10/05/2026 |

Number of Simulators 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.202 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.202 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.203 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 01/18/2026 17:35

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007252

| | | |
|----------------------------------|------------------------|-------------------------------|
| Agency GROVELAND PD | Date 02/15/2026 | Type Agency Inspection |
| Inspector MAVES,JEFFREY,M | Time 14:52:10 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 25024080A3 |
| Expiration 06/24/2027 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 10/05/2026 |

Number of Simulators 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.201 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 02/15/2026 14:50

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007252

| | | |
|----------------------------------|------------------------|-------------------------------|
| Agency GROVELAND PD | Date 03/16/2026 | Type Agency Inspection |
| Inspector MAVES,JEFFREY,M | Time 15:36:09 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 25024080A3 |
| Expiration 06/24/2027 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 10/05/2026 |

Number of Simulators 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|---|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.051 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.202 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.000 |
| Mouth Alcohol Test: Slope Not Met | Yes | 0.08 g/210L Dry Gas Std Test # (Repeat) 1 | 0.078 |
| Interferent Detect Test: Interferent Detect | Yes | Interferent Detect Test # 1 | INT |
| Diagnostic Check (Post-Inspection): OK | Yes | Alcohol Free Test # 2 | 0.000 |
| Alcohol Free Test: OK | Yes | 0.05 g/210L Test # 2 | 0.051 |
| 0.05 g/210L Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.08 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.204 |
| 0.20 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.000 |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # (Repeat) 2 | 0.078 |
| | | Interferent Detect Test # 2 | INT |
| | | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.051 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.203 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.000 |
| | | 0.08 g/210L Dry Gas Std Test # (Repeat) 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

08: Control Outside Tolerance.

Data Download Date/Time 03/16/2026 15:34

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007252

| | | |
|----------------------------------|------------------------|-------------------------------|
| Agency GROVELAND PD | Date 04/15/2026 | Type Agency Inspection |
| Inspector MAVES,JEFFREY,M | Time 17:03:56 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 25024080A3 |
| Expiration 06/24/2027 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 10/05/2026 |

Number of Simulators 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.200 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.201 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 04/15/2026 17:01

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

| | | | |
|----------------------------------|------------------------------|-------------------------------|---------------------------------|
| Agency PENSACOLA PD | Date 01/12/2026 | Type Agency Inspection | |
| Inspector COLLINS,CHASE,M | Time 09:46:31 | Software 8100.27 | |
| 0.05 Lot # 202402E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 403135737 |
| Expiration 02/27/2026 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 08/30/2027 |
| Number of Simulators 5 | | Compliance Yes | |

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.051 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.082 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.203 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.051 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.082 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.203 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.051 |
| | | 0.08 g/210L Test # 3 | 0.082 |
| | | 0.20 g/210L Test # 3 | 0.203 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 01/12/2026 10:44

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

| | | |
|----------------------------------|------------------------|-------------------------------|
| Agency PENSACOLA PD | Date 02/16/2026 | Type Agency Inspection |
| Inspector COLLINS,CHASE,M | Time 09:45:19 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|---------------------------------|
| 0.05 Lot # 202402E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 403342588 |
| Expiration 02/27/2026 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 05/09/2028 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.081 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.202 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.081 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.202 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.082 |
| | | 0.20 g/210L Test # 3 | 0.202 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 02/16/2026 10:47

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

| | | |
|----------------------------------|------------------------|-------------------------------|
| Agency PENSACOLA PD | Date 03/02/2026 | Type Agency Inspection |
| Inspector COLLINS,CHASE,M | Time 09:56:55 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|---------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 403342588 |
| Expiration 06/24/2027 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 05/09/2028 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # (Repeat) 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.051 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.081 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.202 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # (Repeat) 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.081 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.202 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # (Repeat) 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.081 |
| | | 0.20 g/210L Test # 3 | 0.201 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

00: Ambient Fail.

Data Download Date/Time 03/02/2026 10:55

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

| | | | |
|--------------------------------------|--|-----------------------------------|----------------------------------|
| Agency PENSACOLA PD | Date 03/23/2026 | Type Department Inspection | |
| Inspector PIERSON, WEN-CHI, K | Time 13:24:36 | Software 8100.27 | |
| 0.05 Lot # 202406K | 0.08 Lot # 202406L | 0.20 Lot # 202406N | 0.08 Gas Lot # 23725080A1 |
| Expiration 06/19/2026 | Expiration 06/19/2026 | Expiration 06/20/2026 | Expiration 10/05/2027 |
| Number of Simulators 5 | Average Standard Deviation 0.0003 | Compliance Yes | |

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | Yes | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | Yes | 0.20 g/210L Test # 1 | 0.200 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |
| | | Alcohol Free Test # 4 | 0.000 |
| | | 0.05 g/210L Test # 4 | 0.049 |
| | | 0.08 g/210L Test # 4 | 0.079 |
| | | 0.20 g/210L Test # 4 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 4 | 0.080 |
| | | Interferent Detect Test # 4 | INT |
| | | Alcohol Free Test # 5 | 0.000 |
| | | 0.05 g/210L Test # 5 | 0.049 |
| | | 0.08 g/210L Test # 5 | 0.079 |
| | | 0.20 g/210L Test # 5 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 5 | 0.080 |
| | | Interferent Detect Test # 5 | INT |

Florida Department of Law Enforcement
Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

| | |
|--|--------|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.049 |
| 0.08 g/210L Test # 6 | 0.079 |
| 0.20 g/210L Test # 6 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.080 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.050 |
| 0.08 g/210L Test # 7 | 0.078 |
| 0.20 g/210L Test # 7 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.080 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.049 |
| 0.08 g/210L Test # 8 | 0.079 |
| 0.20 g/210L Test # 8 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.080 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.049 |
| 0.08 g/210L Test # 9 | 0.079 |
| 0.20 g/210L Test # 9 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.080 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.049 |
| 0.08 g/210L Test # 10 | 0.079 |
| 0.20 g/210L Test # 10 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.080 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0003 |
| 0.08 g/210L Test (Std Dev) | 0.0003 |
| 0.20 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0004 |

Remarks:

Data Download Date/Time 03/23/2026 14:23

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

| | | | |
|----------------------------------|------------------------------|-------------------------------|---------------------------------|
| Agency PENSACOLA PD | Date 04/20/2026 | Type Agency Inspection | |
| Inspector COLLINS,CHASE,M | Time 11:14:33 | Software 8100.27 | |
| 0.05 Lot # 202506E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 403135737 |
| Expiration 06/24/2027 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 08/30/2027 |
| Number of Simulators 5 | | Compliance Yes | |

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.052 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.082 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.203 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.083 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.052 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.082 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.204 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.083 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.052 |
| | | 0.08 g/210L Test # 3 | 0.083 |
| | | 0.20 g/210L Test # 3 | 0.204 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.083 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 04/20/2026 12:11

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

| | | | |
|-------------------------------|------------------------------|-------------------------------|---------------------------------|
| Agency ST JOHNS SO | Date 01/29/2026 | Type Agency Inspection | |
| Inspector MILLS,S,T | Time 22:43:46 | Software 8100.27 | |
| 0.05 Lot # 202506E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 403374313 |
| Expiration 06/24/2027 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 06/25/2028 |
| Number of Simulators 5 | | Compliance Yes | |

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.000 |
| Minimum Sample Volume Check: OK | | 0.05 g/210L Test # (Repeat) 1 | 0.050 |
| Barometric Pressure Sensor Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.20 g/210L Test # 1 | 0.201 |
| Mouth Alcohol Test: Slope Not Met | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Interferent Detect Test: Interferent Detect | Yes | Interferent Detect Test # 1 | INT |
| Diagnostic Check (Post-Inspection): OK | Yes | Alcohol Free Test # 2 | 0.000 |
| Alcohol Free Test: OK | Yes | 0.05 g/210L Test # 2 | 0.000 |
| 0.05 g/210L Test: OK | Yes | 0.05 g/210L Test # (Repeat) 2 | 0.050 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.202 |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| | | Interferent Detect Test # 2 | INT |
| | | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.000 |
| | | 0.05 g/210L Test # (Repeat) 3 | 0.051 |
| | | 0.08 g/210L Test # 3 | 0.081 |
| | | 0.20 g/210L Test # 3 | 0.202 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

05: CONTROL OUTSIDE TOLERANCE SIM DISCONNECTED

Data Download Date/Time 02/02/2026 13:23

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

| | | | |
|--------------------------------|------------------------------|-------------------------------|---------------------------------|
| Agency ST JOHNS SO | Date 02/27/2026 | Type Agency Inspection | |
| Inspector MILLS,SCOTT,T | Time 20:40:45 | Software 8100.27 | |
| 0.05 Lot # 202506E | 0.08 Lot # 202510E | 0.20 Lot # 202510F | 0.08 Gas Lot # 403374313 |
| Expiration 06/24/2027 | Expiration 10/28/2027 | Expiration 10/29/2027 | Expiration 06/25/2028 |
| Number of Simulators 5 | Compliance Yes | | |

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.201 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.201 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 02/27/2026 20:44

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

| | | | |
|--------------------------------|------------------------------|-------------------------------|---------------------------------|
| Agency ST JOHNS SO | Date 03/26/2026 | Type Agency Inspection | |
| Inspector MILLS,SCOTT,T | Time 00:43:31 | Software 8100.27 | |
| 0.05 Lot # 202506E | 0.08 Lot # 202510E | 0.20 Lot # 202510F | 0.08 Gas Lot # 403374313 |
| Expiration 06/24/2027 | Expiration 10/28/2027 | Expiration 10/29/2027 | Expiration 06/25/2028 |
| Number of Simulators 5 | Compliance Yes | | |

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.200 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

A F / M A:RESIDUAL M/A.

Data Download Date/Time 03/26/2026 00:52

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

| | | | |
|--------------------------------|------------------------------|-------------------------------|---------------------------------|
| Agency ST JOHNS SO | Date 04/29/2026 | Type Agency Inspection | |
| Inspector MILLS,SCOTT,T | Time 22:21:56 | Software 8100.27 | |
| 0.05 Lot # 202506E | 0.08 Lot # 202510E | 0.20 Lot # 202510F | 0.08 Gas Lot # 403374313 |
| Expiration 06/24/2027 | Expiration 10/28/2027 | Expiration 10/29/2027 | Expiration 06/25/2028 |
| Number of Simulators 5 | | Compliance Yes | |

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.200 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 04/29/2026 22:39

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007373

| | | |
|------------------------------------|------------------------|-------------------------------|
| Agency BROWARD COUNTY SO | Date 01/14/2026 | Type Agency Inspection |
| Inspector FRAZIER, ANAYA, S | Time 08:31:30 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|---------------------------------|
| 0.05 Lot # 202402E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 402783528 |
| Expiration 02/27/2026 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 06/27/2026 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|---|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.200 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.000 |
| Mouth Alcohol Test: Slope Not Met | Yes | 0.08 g/210L Dry Gas Std Test # (Repeat) 1 | 0.079 |
| Interferent Detect Test: Interferent Detect | Yes | Interferent Detect Test # 1 | INT |
| Diagnostic Check (Post-Inspection): OK | Yes | Alcohol Free Test # 2 | 0.000 |
| Alcohol Free Test: OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| 0.05 g/210L Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.08 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.201 |
| 0.20 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.000 |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # (Repeat) 2 | 0.078 |
| | | Interferent Detect Test # 2 | INT |
| | | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.000 |
| | | 0.08 g/210L Dry Gas Std Test # (Repeat) 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

08: Control Outside Tolerance. DRY GAS NOT ATTACHED..

Data Download Date/Time 01/14/2026 09:03

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007373

| | | |
|------------------------------------|------------------------|-------------------------------|
| Agency BROWARD COUNTY SO | Date 02/04/2026 | Type Agency Inspection |
| Inspector FRAZIER, ANAYA, S | Time 09:19:40 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|---------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 402783528 |
| Expiration 06/24/2027 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 02/04/2026 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.081 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.202 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.051 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.081 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.203 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.051 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.202 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

S TILMAN IN TRAINING

Data Download Date/Time 02/04/2026 09:27

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007373

| | | |
|------------------------------------|------------------------|-------------------------------|
| Agency BROWARD COUNTY SO | Date 03/04/2026 | Type Agency Inspection |
| Inspector FRAZIER, ANAYA, S | Time 08:53:34 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|---------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 402783528 |
| Expiration 06/24/2027 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 06/27/2026 |

Number of Simulators 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.051 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.201 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.051 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.051 |
| | | 0.08 g/210L Test # 3 | 0.081 |
| | | 0.20 g/210L Test # 3 | 0.201 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

S. TILLMAN IN TRAINING.

Data Download Date/Time 03/05/2026 02:41

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007373

| | | |
|------------------------------------|------------------------|-------------------------------|
| Agency BROWARD COUNTY SO | Date 04/01/2026 | Type Agency Inspection |
| Inspector FRAZIER, ANAYA, S | Time 09:15:05 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|---------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 402783528 |
| Expiration 06/24/2027 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 06/27/2026 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.203 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.202 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.083 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.202 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.083 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

S. TILLMAN IN TRAINING

Data Download Date/Time 04/01/2026 09:27

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007380

| | | |
|-----------------------------------|------------------------|-------------------------------|
| Agency HOMESTEAD PD | Date 01/22/2026 | Type Agency Inspection |
| Inspector ACKERMAN,ELVIS,A | Time 19:45:38 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|---------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202510E | 0.20 Lot # 202410F | 0.08 Gas Lot # 403008479 |
| Expiration 06/24/2027 | Expiration 10/28/2027 | Expiration 10/30/2026 | Expiration 03/28/2027 |

Number of Simulators 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

N/A

Data Download Date/Time 01/22/2026 19:55

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007380

| | | |
|-----------------------------------|------------------------|-------------------------------|
| Agency HOMESTEAD PD | Date 02/16/2026 | Type Agency Inspection |
| Inspector ACKERMAN,ELVIS,A | Time 18:19:26 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|---------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202510E | 0.20 Lot # 202410F | 0.08 Gas Lot # 403008479 |
| Expiration 06/24/2027 | Expiration 10/28/2027 | Expiration 10/30/2026 | Expiration 03/28/2027 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 02/16/2026 18:18

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007380

| | | |
|-----------------------------------|------------------------|-------------------------------|
| Agency HOMESTEAD PD | Date 03/23/2026 | Type Agency Inspection |
| Inspector ACKERMAN,ELVIS,A | Time 19:59:00 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|---------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202510E | 0.20 Lot # 202410F | 0.08 Gas Lot # 400374313 |
| Expiration 06/24/2027 | Expiration 10/28/2027 | Expiration 10/30/2026 | Expiration 06/25/2028 |

Number of Simulators 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 03/23/2026 20:32

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

| | | |
|------------------------------------|------------------------|-------------------------------|
| Agency BROWARD COUNTY SO | Date 01/14/2026 | Type Agency Inspection |
| Inspector FRAZIER, ANAYA, S | Time 08:49:29 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|---------------------------------|
| 0.05 Lot # 202402E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 402783528 |
| Expiration 02/27/2026 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 06/26/2026 |

Number of Simulators 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

S. TILLMAN IN TRAINING.

Data Download Date/Time 01/14/2026 09:06

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

| | | |
|------------------------------------|------------------------|-------------------------------|
| Agency BROWARD COUNTY SO | Date 02/04/2026 | Type Agency Inspection |
| Inspector FRAZIER, ANAYA, S | Time 08:11:58 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|---------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 402783528 |
| Expiration 06/24/2027 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 06/27/2026 |

Number of Simulators 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

S. TILLMAN IN TRAINING.

Data Download Date/Time 02/04/2026 08:19

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

| | | |
|------------------------------------|------------------------|-------------------------------|
| Agency BROWARD COUNTY SO | Date 03/04/2026 | Type Agency Inspection |
| Inspector FRAZIER, ANAYA, S | Time 09:42:14 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|---------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 402783528 |
| Expiration 06/24/2027 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 06/27/2026 |

Number of Simulators 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.196 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.196 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

S. TILLMAN IN TRAINING.

Data Download Date/Time 03/05/2026 01:59

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

Agency BROWARD COUNTY SO **Date** 03/20/2026 **Type** Department Inspection
Inspector GUTSCHOW,TAYLOR,D **Time** 15:59:26 **Software** 8100.27

0.05 Lot # 202406K **0.08 Lot #** 202406L **0.20 Lot #** 202406N **0.08 Gas Lot #** AG429602
Expiration 06/19/2026 **Expiration** 06/19/2026 **Expiration** 06/20/2026 **Expiration** 10/22/2026

Number of Simulators 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | Yes | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | Yes | 0.20 g/210L Test # 1 | 0.196 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.196 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |
| | | Alcohol Free Test # 4 | 0.000 |
| | | 0.05 g/210L Test # 4 | 0.048 |
| | | 0.08 g/210L Test # 4 | 0.079 |
| | | 0.20 g/210L Test # 4 | 0.196 |
| | | 0.08 g/210L Dry Gas Std Test # 4 | 0.078 |
| | | Interferent Detect Test # 4 | INT |
| | | Alcohol Free Test # 5 | 0.000 |
| | | 0.05 g/210L Test # 5 | 0.048 |
| | | 0.08 g/210L Test # 5 | 0.078 |
| | | 0.20 g/210L Test # 5 | 0.196 |
| | | 0.08 g/210L Dry Gas Std Test # 5 | 0.078 |
| | | Interferent Detect Test # 5 | INT |

Florida Department of Law Enforcement
Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

| | |
|--|--------|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.049 |
| 0.08 g/210L Test # 6 | 0.078 |
| 0.20 g/210L Test # 6 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.078 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.048 |
| 0.08 g/210L Test # 7 | 0.079 |
| 0.20 g/210L Test # 7 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.078 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.049 |
| 0.08 g/210L Test # 8 | 0.079 |
| 0.20 g/210L Test # 8 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.078 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.049 |
| 0.08 g/210L Test # 9 | 0.079 |
| 0.20 g/210L Test # 9 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.077 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.049 |
| 0.08 g/210L Test # 10 | 0.079 |
| 0.20 g/210L Test # 10 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.078 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Test (Std Dev) | 0.0004 |
| 0.20 g/210L Test (Std Dev) | 0.0000 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0004 |

Remarks:

Data Download Date/Time 03/20/2026 15:57

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

| | | |
|------------------------------------|------------------------|-------------------------------|
| Agency BROWARD COUNTY SO | Date 04/17/2026 | Type Agency Inspection |
| Inspector FRAZIER, ANAYA, S | Time 22:34:43 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|---------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 402783528 |
| Expiration 06/24/2027 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 06/27/2026 |

Number of Simulators 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

S. TILLMAN IN TRAINING.

Data Download Date/Time 04/17/2026 22:39

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007382

| | | |
|-----------------------------------|------------------------|-------------------------------|
| Agency SUWANNEE COUNTY SO | Date 01/28/2026 | Type Agency Inspection |
| Inspector GRIFFIN,ANGELA,F | Time 05:39:04 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|---------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 403374313 |
| Expiration 06/24/2027 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 06/25/2028 |

Number of Simulators 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.051 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.200 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.083 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.051 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.081 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.083 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.051 |
| | | 0.08 g/210L Test # 3 | 0.081 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.083 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

NONE

Data Download Date/Time 01/28/2026 08:35

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007382

| | | | |
|---|--|-----------------------------------|----------------------------------|
| Agency SUWANNEE COUNTY SO | Date 03/18/2026 | Type Department Inspection | |
| Inspector HIGGINBOTHAM, LEANDRA, | Time 13:24:48 | Software 8100.27 | |
| 0.05 Lot # 202406K | 0.08 Lot # 202406L | 0.20 Lot # 202406N | 0.08 Gas Lot # 23825080A1 |
| Expiration 06/19/2026 | Expiration 06/19/2026 | Expiration 06/20/2026 | Expiration 10/05/2027 |
| Number of Simulators 5 | Average Standard Deviation 0.0004 | Compliance Yes | |

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | Yes | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | Yes | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.051 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |
| | | Alcohol Free Test # 4 | 0.000 |
| | | 0.05 g/210L Test # 4 | 0.051 |
| | | 0.08 g/210L Test # 4 | 0.080 |
| | | 0.20 g/210L Test # 4 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 4 | 0.081 |
| | | Interferent Detect Test # 4 | INT |
| | | Alcohol Free Test # 5 | 0.000 |
| | | 0.05 g/210L Test # 5 | 0.051 |
| | | 0.08 g/210L Test # 5 | 0.080 |
| | | 0.20 g/210L Test # 5 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 5 | 0.081 |
| | | Interferent Detect Test # 5 | INT |

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007382

| | |
|--|--------|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.051 |
| 0.08 g/210L Test # 6 | 0.080 |
| 0.20 g/210L Test # 6 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.081 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.051 |
| 0.08 g/210L Test # 7 | 0.081 |
| 0.20 g/210L Test # 7 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.081 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.051 |
| 0.08 g/210L Test # 8 | 0.080 |
| 0.20 g/210L Test # 8 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.081 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.051 |
| 0.08 g/210L Test # 9 | 0.080 |
| 0.20 g/210L Test # 9 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.081 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.051 |
| 0.08 g/210L Test # 10 | 0.080 |
| 0.20 g/210L Test # 10 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.081 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Test (Std Dev) | 0.0003 |
| 0.20 g/210L Test (Std Dev) | 0.0009 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0003 |

Remarks:

Data Download Date/Time 03/18/2026 13:22

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007433

| | | |
|------------------------------------|------------------------|-------------------------------|
| Agency BROWARD COUNTY SO | Date 01/14/2026 | Type Agency Inspection |
| Inspector FRAZIER, ANAYA, S | Time 10:09:36 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|---------------------------------|
| 0.05 Lot # 202402E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 402783528 |
| Expiration 02/27/2026 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 06/27/2026 |

Number of Simulators 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.200 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

S. TILLMAN IN TRAINING.

Data Download Date/Time 01/14/2026 10:37

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007433

| | | |
|------------------------------------|------------------------|-------------------------------|
| Agency BROWARD COUNTY SO | Date 03/09/2026 | Type Agency Inspection |
| Inspector GUTSCHOW,TAYLOR,D | Time 14:33:56 | Software 8100.27 |

| | | | |
|-------------------|-------------------|-------------------|-----------------------|
| 0.05 Lot # | 0.08 Lot # | 0.20 Lot # | 0.08 Gas Lot # |
| Expiration | Expiration | Expiration | Expiration |

Number of Simulators **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|------|--------|
| Diagnostic Check (Pre-Inspection): OK | No | | |
| Date and/or Time Adjusted | No | | |
| Minimum Sample Volume Check: OK | | | |
| Barometric Pressure Sensor Check: OK | | | |
| Alcohol Free Subject Test: 0.000 | No | | |
| Mouth Alcohol Test: Slope Not Met | No | | |
| Interferent Detect Test: Interferent Detect | No | | |
| Diagnostic Check (Post-Inspection): OK | No | | |
| Alcohol Free Test: OK | No | | |
| 0.05 g/210L Test: OK | No | | |
| 0.08 g/210L Test: OK | No | | |
| 0.20 g/210L Test: OK | No | | |
| 0.08 g/210L Dry Gas Std Test: OK | No | | |

Remarks:

AI NOT CONDUCTED. COMPLIANCE NOT DETERMINED.

Data Download Date/Time 03/09/2026 14:31

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007433

Agency BROWARD COUNTY SO **Date** 03/12/2026 **Type** Department Inspection
Inspector GUTSCHOW,TAYLOR,D **Time** 11:53:02 **Software** 8100.27

0.05 Lot # 202406K **0.08 Lot #** 202406L **0.20 Lot #** 202406N **0.08 Gas Lot #** AG526603
Expiration 06/19/2026 **Expiration** 06/19/2026 **Expiration** 06/20/2026 **Expiration** 09/23/2027

Number of Simulators 5 **Average Standard Deviation** 0.0002 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | Yes | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | Yes | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |
| | | Alcohol Free Test # 4 | 0.000 |
| | | 0.05 g/210L Test # 4 | 0.048 |
| | | 0.08 g/210L Test # 4 | 0.078 |
| | | 0.20 g/210L Test # 4 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 4 | 0.079 |
| | | Interferent Detect Test # 4 | INT |
| | | Alcohol Free Test # 5 | 0.000 |
| | | 0.05 g/210L Test # 5 | 0.048 |
| | | 0.08 g/210L Test # 5 | 0.078 |
| | | 0.20 g/210L Test # 5 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 5 | 0.079 |
| | | Interferent Detect Test # 5 | INT |

**Florida Department of Law Enforcement
Alcohol Testing Program**

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007433

| | |
|--|--------|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.048 |
| 0.08 g/210L Test # 6 | 0.078 |
| 0.20 g/210L Test # 6 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.079 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.048 |
| 0.08 g/210L Test # 7 | 0.078 |
| 0.20 g/210L Test # 7 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.079 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.048 |
| 0.08 g/210L Test # 8 | 0.078 |
| 0.20 g/210L Test # 8 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.079 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.048 |
| 0.08 g/210L Test # 9 | 0.078 |
| 0.20 g/210L Test # 9 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.079 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.048 |
| 0.08 g/210L Test # 10 | 0.078 |
| 0.20 g/210L Test # 10 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.079 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0000 |
| 0.08 g/210L Test (Std Dev) | 0.0003 |
| 0.20 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0003 |

Remarks:

Data Download Date/Time 03/12/2026 11:49

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007433

| | | |
|------------------------------------|------------------------|-------------------------------|
| Agency BROWARD COUNTY SO | Date 03/31/2026 | Type Agency Inspection |
| Inspector FRAZIER, ANAYA, S | Time 23:42:07 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|---------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 402783528 |
| Expiration 06/24/2027 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 06/27/2026 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.200 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

S. TILLMAN IN TRAINING.

Data Download Date/Time 03/31/2026 23:46

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007433

| | | |
|------------------------------------|------------------------|-------------------------------|
| Agency BROWARD COUNTY SO | Date 04/01/2026 | Type Agency Inspection |
| Inspector FRAZIER, ANAYA, S | Time 08:24:42 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|---------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 402783528 |
| Expiration 06/24/2027 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 06/27/2026 |

Number of Simulators 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.051 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

S. TILLMAN IN TRAINING

Data Download Date/Time 04/01/2026 08:25

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007434

| | | |
|------------------------------------|------------------------|-------------------------------|
| Agency BROWARD COUNTY SO | Date 01/14/2026 | Type Agency Inspection |
| Inspector FRAZIER, ANAYA, S | Time 09:50:34 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|---------------------------------|
| 0.05 Lot # 202402E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 403345775 |
| Expiration 02/27/2026 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 05/13/2028 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|---|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.201 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.000 |
| Mouth Alcohol Test: Slope Not Met | Yes | 0.08 g/210L Dry Gas Std Test # (Repeat) 1 | 0.079 |
| Interferent Detect Test: Interferent Detect | Yes | Interferent Detect Test # 1 | INT |
| Diagnostic Check (Post-Inspection): OK | Yes | Alcohol Free Test # 2 | 0.000 |
| Alcohol Free Test: OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| 0.05 g/210L Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.08 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.201 |
| 0.20 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.000 |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # (Repeat) 2 | 0.079 |
| | | Interferent Detect Test # 2 | INT |
| | | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.201 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.000 |
| | | 0.08 g/210L Dry Gas Std Test # (Repeat) 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

08: Control Outside Tolerance.DRY GAS NO ATTACHED.

Data Download Date/Time 01/14/2026 10:22

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007434

| | | |
|------------------------------------|------------------------|-------------------------------|
| Agency BROWARD COUNTY SO | Date 03/06/2026 | Type Agency Inspection |
| Inspector GUTSCHOW,TAYLOR,D | Time 09:34:16 | Software 8100.27 |

| | | | |
|-------------------|-------------------|-------------------|-----------------------|
| 0.05 Lot # | 0.08 Lot # | 0.20 Lot # | 0.08 Gas Lot # |
| Expiration | Expiration | Expiration | Expiration |

Number of Simulators **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|------|--------|
| Diagnostic Check (Pre-Inspection): OK | No | | |
| Date and/or Time Adjusted | No | | |
| Minimum Sample Volume Check: OK | | | |
| Barometric Pressure Sensor Check: OK | | | |
| Alcohol Free Subject Test: 0.000 | No | | |
| Mouth Alcohol Test: Slope Not Met | No | | |
| Interferent Detect Test: Interferent Detect | No | | |
| Diagnostic Check (Post-Inspection): OK | No | | |
| Alcohol Free Test: OK | No | | |
| 0.05 g/210L Test: OK | No | | |
| 0.08 g/210L Test: OK | No | | |
| 0.20 g/210L Test: OK | No | | |
| 0.08 g/210L Dry Gas Std Test: OK | No | | |

Remarks:

AI NOT CONDUCTED. COMPLIANCE NOT DETERMINED.

Data Download Date/Time 03/06/2026 12:51

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007434

Agency BROWARD COUNTY SO **Date** 03/06/2026 **Type** Department Inspection
Inspector GUTSCHOW,TAYLOR,D **Time** 12:53:54 **Software** 8100.27

0.05 Lot # 202406K **0.08 Lot #** 202406L **0.20 Lot #** 202406N **0.08 Gas Lot #** AG429602
Expiration 06/19/2026 **Expiration** 06/19/2026 **Expiration** 06/20/2026 **Expiration** 10/22/2026

Number of Simulators 5 **Average Standard Deviation** 0.0001 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | Yes | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | Yes | 0.20 g/210L Test # 1 | 0.200 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |
| | | Alcohol Free Test # 4 | 0.000 |
| | | 0.05 g/210L Test # 4 | 0.048 |
| | | 0.08 g/210L Test # 4 | 0.078 |
| | | 0.20 g/210L Test # 4 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 4 | 0.078 |
| | | Interferent Detect Test # 4 | INT |
| | | Alcohol Free Test # 5 | 0.000 |
| | | 0.05 g/210L Test # 5 | 0.048 |
| | | 0.08 g/210L Test # 5 | 0.078 |
| | | 0.20 g/210L Test # 5 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 5 | 0.078 |
| | | Interferent Detect Test # 5 | INT |

Florida Department of Law Enforcement
Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007434

| | |
|--|--------|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.048 |
| 0.08 g/210L Test # 6 | 0.078 |
| 0.20 g/210L Test # 6 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.078 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.048 |
| 0.08 g/210L Test # 7 | 0.078 |
| 0.20 g/210L Test # 7 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.078 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.048 |
| 0.08 g/210L Test # 8 | 0.078 |
| 0.20 g/210L Test # 8 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.078 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.048 |
| 0.08 g/210L Test # 9 | 0.078 |
| 0.20 g/210L Test # 9 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.078 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.048 |
| 0.08 g/210L Test # 10 | 0.079 |
| 0.20 g/210L Test # 10 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.078 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0000 |
| 0.08 g/210L Test (Std Dev) | 0.0003 |
| 0.20 g/210L Test (Std Dev) | 0.0000 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0003 |

Remarks:

Data Download Date/Time 03/06/2026 12:51

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007434

| | | |
|------------------------------------|------------------------|-------------------------------|
| Agency BROWARD COUNTY SO | Date 03/31/2026 | Type Agency Inspection |
| Inspector FRAZIER, ANAYA, S | Time 23:34:50 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|---------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 403345775 |
| Expiration 06/24/2027 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 05/13/2028 |

Number of Simulators 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.202 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.201 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

S. TILLMAN IN TRAINING.

Data Download Date/Time 04/01/2026 00:01

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007434

| | | |
|------------------------------------|------------------------|-------------------------------|
| Agency BROWARD COUNTY SO | Date 04/01/2026 | Type Agency Inspection |
| Inspector FRAZIER, ANAYA, S | Time 09:10:20 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|---------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 403345775 |
| Expiration 06/24/2027 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 05/13/2028 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.201 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.201 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

S. TILLMAN IN TRAINING

Data Download Date/Time 04/01/2026 10:51

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007439

| | | |
|------------------------------------|------------------------|-------------------------------|
| Agency BAKER COUNTY SO | Date 01/28/2026 | Type Agency Inspection |
| Inspector COLLINS, CALEB, A | Time 15:05:42 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 01524080A2 |
| Expiration 06/24/2027 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 03/05/2026 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.051 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.201 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.051 |
| | | 0.08 g/210L Test # 3 | 0.081 |
| | | 0.20 g/210L Test # 3 | 0.201 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 01/29/2026 14:58

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007439

| | | |
|---|------------------------|-------------------------------|
| Agency BAKER COUNTY SO | Date 02/24/2026 | Type Agency Inspection |
| Inspector DEREK WATSON, CALEB COLLINS, / | Time 13:52:02 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 01524O80A2 |
| Expiration 06/24/2027 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 03/05/2026 |

Number of Simulators 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.052 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.081 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.200 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.052 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.081 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.052 |
| | | 0.08 g/210L Test # 3 | 0.081 |
| | | 0.20 g/210L Test # 3 | 0.201 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 02/24/2026 13:48

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007439

| | | |
|------------------------------------|------------------------|-------------------------------|
| Agency BAKER COUNTY SO | Date 03/24/2026 | Type Agency Inspection |
| Inspector COLLINS, CALEB, A | Time 13:27:11 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 03326080A2 |
| Expiration 06/24/2027 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 03/05/2028 |

Number of Simulators 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.201 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.202 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.201 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 03/24/2026 13:25

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007439

| | | |
|---|------------------------|-------------------------------|
| Agency BAKER COUNTY SO | Date 04/16/2026 | Type Agency Inspection |
| Inspector CALEB COLLINS,DEREK WATSON, / | Time 14:19:54 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 03326080A2 |
| Expiration 06/24/2027 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 03/05/2028 |

Number of Simulators 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.081 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.200 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.081 |
| | | 0.20 g/210L Test # 3 | 0.201 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 04/16/2026 14:16

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007440

| | | |
|-------------------------------------|------------------------|-------------------------------|
| Agency FL HIGHWAY PATROL | Date 02/11/2026 | Type Agency Inspection |
| Inspector PALESE, ANTHONY, W | Time 18:31:31 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 202402E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 2403135737 |
| Expiration 02/27/2026 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 08/30/2027 |

Number of Simulators 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 03/26/2026 17:28

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007440

| | | |
|-------------------------------------|------------------------|-------------------------------|
| Agency FL HIGHWAY PATROL | Date 03/26/2026 | Type Agency Inspection |
| Inspector PALESE, ANTHONY, W | Time 17:20:10 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 2403135737 |
| Expiration 06/24/2027 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 08/30/2027 |

Number of Simulators 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

M/A.

Data Download Date/Time 03/26/2026 17:28

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007440

| | | |
|-------------------------------------|------------------------|-------------------------------|
| Agency FL HIGHWAY PATROL | Date 04/15/2026 | Type Agency Inspection |
| Inspector PALESE, ANTHONY, W | Time 15:58:42 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 2403135737 |
| Expiration 06/24/2027 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 08/30/2027 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | Yes | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Time-Date changed.

Data Download Date/Time 04/15/2026 16:01

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007441

| | | | |
|---------------------------------|------------------------------|-------------------------------|---------------------------------|
| Agency FHP | Date 01/19/2026 | Type Agency Inspection | |
| Inspector RUBEL, NEIL, F | Time 10:23:50 | Software 8100.27 | |
| 0.05 Lot # 202506E | 0.08 Lot # 202510E | 0.20 Lot # 202410F | 0.08 Gas Lot # 402843437 |
| Expiration 06/24/2027 | Expiration 10/28/2027 | Expiration 10/30/2026 | Expiration 09/08/2026 |
| Number of Simulators 5 | | Compliance Yes | |

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.081 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.203 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.051 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.082 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.204 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.052 |
| | | 0.08 g/210L Test # 3 | 0.083 |
| | | 0.20 g/210L Test # 3 | 0.205 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

N/A

Data Download Date/Time 01/19/2026 11:21

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007441

| | | | |
|--------------------------|-----------------------|------------------------|--------------------------|
| Agency FHP | Date 02/06/2026 | Type Agency Inspection | |
| Inspector RUBEL, NEIL, F | Time 08:36:24 | Software 8100.27 | |
| 0.05 Lot # 202506E | 0.08 Lot # 202510E | 0.20 Lot # 202410F | 0.08 Gas Lot # 402843437 |
| Expiration 06/24/2027 | Expiration 10/28/2027 | Expiration 10/30/2026 | Expiration 09/08/2026 |
| Number of Simulators 5 | | Compliance Yes | |

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.051 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.200 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.051 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.081 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.202 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.051 |
| | | 0.08 g/210L Test # 3 | 0.081 |
| | | 0.20 g/210L Test # 3 | 0.203 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

N/A

Data Download Date/Time 02/06/2026 10:03

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007441

| | | | |
|--------------------------|-----------------------|------------------------|--------------------------|
| Agency FHP | Date 03/12/2026 | Type Agency Inspection | |
| Inspector RUBEL, NEIL, F | Time 12:15:38 | Software 8100.27 | |
| 0.05 Lot # 202506E | 0.08 Lot # 202510E | 0.20 Lot # 202410F | 0.08 Gas Lot # 402843437 |
| Expiration 06/24/2027 | Expiration 10/28/2027 | Expiration 10/30/2026 | Expiration 09/08/2026 |
| Number of Simulators 5 | | Compliance Yes | |

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.081 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.081 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.082 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

N/A

Data Download Date/Time 03/12/2026 14:00

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007441

| | | | |
|---------------------------------------|------------------------------|-------------------------------|---------------------------------|
| Agency FHP | Date 04/04/2026 | Type Agency Inspection | |
| Inspector MCDOWELL, MICHAEL, J | Time 10:16:51 | Software 8100.27 | |
| 0.05 Lot # 202506E | 0.08 Lot # 202510E | 0.20 Lot # 202410F | 0.08 Gas Lot # 402843437 |
| Expiration 06/24/2027 | Expiration 10/28/2027 | Expiration 10/30/2026 | Expiration 09/08/2026 |
| Number of Simulators 5 | | Compliance Yes | |

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

N/A

Data Download Date/Time 04/04/2026 11:13

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007444

| | | |
|-------------------------------------|------------------------|-------------------------------|
| Agency FL HIGHWAY PATROL | Date 02/11/2026 | Type Agency Inspection |
| Inspector PALESE, ANTHONY, W | Time 19:38:54 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 202402E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 2403135737 |
| Expiration 02/27/2026 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 08/30/2027 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.201 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.202 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.201 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

M/A .

Data Download Date/Time 02/14/2026 17:42

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007444

| | | |
|-------------------------------------|------------------------|-------------------------------|
| Agency FL HIGHWAY PATROL | Date 03/26/2026 | Type Agency Inspection |
| Inspector PALESE, ANTHONY, W | Time 15:51:56 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 2403135737 |
| Expiration 06/24/2027 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 08/30/2027 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.201 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.201 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 03/26/2026 15:51

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007444

| | | |
|-------------------------------------|------------------------|-------------------------------|
| Agency FL HIGHWAY PATROL | Date 04/15/2026 | Type Agency Inspection |
| Inspector PALESE, ANTHONY, W | Time 16:33:29 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 2403135737 |
| Expiration 06/24/2027 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 08/30/2027 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.051 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.202 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.051 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.202 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.051 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.202 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 04/15/2026 16:33

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007455

| | | |
|---|------------------------|-------------------------------|
| Agency FL HIGHWAY PATROL | Date 02/11/2026 | Type Agency Inspection |
| Inspector HIGGINBOTHAM, LEANDRA, | Time 10:39:04 | Software 8100.27 |

| | | | |
|-------------------|-------------------|-------------------|-----------------------|
| 0.05 Lot # | 0.08 Lot # | 0.20 Lot # | 0.08 Gas Lot # |
| Expiration | Expiration | Expiration | Expiration |

Number of Simulators **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|------|--------|
| Diagnostic Check (Pre-Inspection): OK | No | | |
| Date and/or Time Adjusted | No | | |
| Minimum Sample Volume Check: OK | | | |
| Barometric Pressure Sensor Check: OK | | | |
| Alcohol Free Subject Test: 0.000 | No | | |
| Mouth Alcohol Test: Slope Not Met | No | | |
| Interferent Detect Test: Interferent Detect | No | | |
| Diagnostic Check (Post-Inspection): OK | No | | |
| Alcohol Free Test: OK | No | | |
| 0.05 g/210L Test: OK | No | | |
| 0.08 g/210L Test: OK | No | | |
| 0.20 g/210L Test: OK | No | | |
| 0.08 g/210L Dry Gas Std Test: OK | No | | |

Remarks:

BYPASS AI FOR OPERATION, COMPLIANCE UNDETERMINED

Data Download Date/Time 02/12/2026 17:58

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007455

Agency FL HIGHWAY PATROL **Date** 02/12/2026 **Type** Department Inspection
Inspector HIGGINBOTHAM, LEANDRA, **Time** 18:00:43 **Software** 8100.27

0.05 Lot # 202406K **0.08 Lot #** 202406L **0.20 Lot #** 202406N **0.08 Gas Lot #** AG510701
Expiration 06/19/2026 **Expiration** 06/19/2026 **Expiration** 06/20/2026 **Expiration** 04/17/2027

Number of Simulators 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.047 |
| Minimum Sample Volume Check: OK | Yes | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | Yes | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.047 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.047 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |
| | | Alcohol Free Test # 4 | 0.000 |
| | | 0.05 g/210L Test # 4 | 0.047 |
| | | 0.08 g/210L Test # 4 | 0.078 |
| | | 0.20 g/210L Test # 4 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 4 | 0.080 |
| | | Interferent Detect Test # 4 | INT |
| | | Alcohol Free Test # 5 | 0.000 |
| | | 0.05 g/210L Test # 5 | 0.047 |
| | | 0.08 g/210L Test # 5 | 0.078 |
| | | 0.20 g/210L Test # 5 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 5 | 0.079 |
| | | Interferent Detect Test # 5 | INT |

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007455

| | |
|--|--------|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.047 |
| 0.08 g/210L Test # 6 | 0.079 |
| 0.20 g/210L Test # 6 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.080 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.047 |
| 0.08 g/210L Test # 7 | 0.079 |
| 0.20 g/210L Test # 7 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.079 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.047 |
| 0.08 g/210L Test # 8 | 0.079 |
| 0.20 g/210L Test # 8 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.079 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.048 |
| 0.08 g/210L Test # 9 | 0.079 |
| 0.20 g/210L Test # 9 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.079 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.048 |
| 0.08 g/210L Test # 10 | 0.079 |
| 0.20 g/210L Test # 10 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.080 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Test (Std Dev) | 0.0004 |
| 0.20 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0007 |

Remarks:

Data Download Date/Time 02/12/2026 17:58

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007455

| | | |
|----------------------------------|------------------------|-------------------------------|
| Agency FL HIGHWAY PATROL | Date 04/04/2026 | Type Agency Inspection |
| Inspector SEARS, COLBY, D | Time 19:54:04 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202510E | 0.20 Lot # 202410F | 0.08 Gas Lot # 19825080A1 |
| Expiration 06/24/2027 | Expiration 10/28/2027 | Expiration 10/30/2026 | Expiration 09/05/2027 |

Number of Simulators 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 04/04/2026 20:06

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007456

| | | |
|--------------------------------------|------------------------|-------------------------------|
| Agency FHP | Date 01/07/2026 | Type Agency Inspection |
| Inspector MOSCHETTO,ANTHONY,C | Time 07:43:25 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 202402E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 10124080A3 |
| Expiration 02/27/2026 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 06/05/2026 |

Number of Simulators 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # (Repeat) 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.202 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # (Repeat) 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # (Repeat) 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

00: Ambient Fail.STANDING NEXT TO INSTRUMENT AFTER MA TEST

Data Download Date/Time 01/07/2026 08:51

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007456

| | | |
|--------------------------------------|------------------------|-------------------------------|
| Agency FHP | Date 02/04/2026 | Type Agency Inspection |
| Inspector MOSCHETTO,ANTHONY,C | Time 07:57:55 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 202402E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 10124080A3 |
| Expiration 02/27/2026 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 06/05/2026 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

NONE

Data Download Date/Time 02/04/2026 08:58

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007456

| | | | |
|--------------------------------------|--|-----------------------------------|----------------------------------|
| Agency FHP | Date 03/02/2026 | Type Department Inspection | |
| Inspector PIERSON, WEN-CHI, K | Time 14:10:43 | Software 8100.27 | |
| 0.05 Lot # 202406K | 0.08 Lot # 202406L | 0.20 Lot # 202406N | 0.08 Gas Lot # 28424080A3 |
| Expiration 06/19/2026 | Expiration 06/19/2026 | Expiration 06/20/2026 | Expiration 11/05/2026 |
| Number of Simulators 5 | Average Standard Deviation 0.0002 | Compliance Yes | |

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | Yes | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | Yes | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |
| | | Alcohol Free Test # 4 | 0.000 |
| | | 0.05 g/210L Test # 4 | 0.048 |
| | | 0.08 g/210L Test # 4 | 0.078 |
| | | 0.20 g/210L Test # 4 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 4 | 0.081 |
| | | Interferent Detect Test # 4 | INT |
| | | Alcohol Free Test # 5 | 0.000 |
| | | 0.05 g/210L Test # 5 | 0.048 |
| | | 0.08 g/210L Test # 5 | 0.078 |
| | | 0.20 g/210L Test # 5 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 5 | 0.081 |
| | | Interferent Detect Test # 5 | INT |

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007456

| | |
|--|--------|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.048 |
| 0.08 g/210L Test # 6 | 0.078 |
| 0.20 g/210L Test # 6 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.081 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.048 |
| 0.08 g/210L Test # 7 | 0.079 |
| 0.20 g/210L Test # 7 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.081 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.049 |
| 0.08 g/210L Test # 8 | 0.078 |
| 0.20 g/210L Test # 8 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.082 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.048 |
| 0.08 g/210L Test # 9 | 0.078 |
| 0.20 g/210L Test # 9 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.081 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.049 |
| 0.08 g/210L Test # 10 | 0.078 |
| 0.20 g/210L Test # 10 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.081 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Test (Std Dev) | 0.0004 |
| 0.20 g/210L Test (Std Dev) | 0.0000 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0003 |

Remarks:

Data Download Date/Time 03/02/2026 14:12

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007456

| | | |
|--------------------------------------|------------------------|-------------------------------|
| Agency FHP | Date 03/11/2026 | Type Agency Inspection |
| Inspector MOSCHETTO,ANTHONY,C | Time 07:03:47 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202510E | 0.20 Lot # 202410F | 0.08 Gas Lot # 10124080A3 |
| Expiration 06/24/2027 | Expiration 10/28/2027 | Expiration 10/30/2026 | Expiration 06/05/2026 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

NONE

Data Download Date/Time 03/11/2026 07:07

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007456

| | | |
|--------------------------------------|------------------------|-------------------------------|
| Agency FHP | Date 04/01/2026 | Type Agency Inspection |
| Inspector MOSCHETTO,ANTHONY,C | Time 06:51:37 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202510E | 0.20 Lot # 202410F | 0.08 Gas Lot # 10124080A3 |
| Expiration 06/24/2027 | Expiration 10/28/2027 | Expiration 10/30/2026 | Expiration 06/05/2026 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.196 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.196 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

NONE

Data Download Date/Time 04/01/2026 07:04

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007457

| | | |
|---|------------------------|-------------------------------|
| Agency FL HIGHWAY PATROL | Date 01/01/2026 | Type Agency Inspection |
| Inspector WILDEMAN-STUART,JORDAN,R | Time 18:40:41 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|---------------------------------|
| 0.05 Lot # 202402E | 0.08 Lot # 202402E | 0.20 Lot # 202410F | 0.08 Gas Lot # 403516666 |
| Expiration 02/27/2026 | Expiration 02/27/2026 | Expiration 10/30/2026 | Expiration 12/16/2028 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.202 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.202 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.202 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 01/01/2026 18:51

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007457

Agency FL HIGHWAY PATROL **Date** 01/28/2026 **Type** Department Inspection
Inspector PIERSON, WEN-CHI, K **Time** 13:22:01 **Software** 8100.27

0.05 Lot # 202406K **0.08 Lot #** 202406L **0.20 Lot #** 202406N **0.08 Gas Lot #** AG510701
Expiration 06/19/2026 **Expiration** 06/19/2026 **Expiration** 04/17/2027 **Expiration** 04/17/2027

Number of Simulators 5 **Average Standard Deviation** 0.0001 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | Yes | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | Yes | 0.20 g/210L Test # 1 | 0.200 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |
| | | Alcohol Free Test # 4 | 0.000 |
| | | 0.05 g/210L Test # 4 | 0.048 |
| | | 0.08 g/210L Test # 4 | 0.080 |
| | | 0.20 g/210L Test # 4 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 4 | 0.079 |
| | | Interferent Detect Test # 4 | INT |
| | | Alcohol Free Test # 5 | 0.000 |
| | | 0.05 g/210L Test # 5 | 0.048 |
| | | 0.08 g/210L Test # 5 | 0.080 |
| | | 0.20 g/210L Test # 5 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 5 | 0.079 |
| | | Interferent Detect Test # 5 | INT |

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007457

| | |
|--|--------|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.048 |
| 0.08 g/210L Test # 6 | 0.080 |
| 0.20 g/210L Test # 6 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.079 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.048 |
| 0.08 g/210L Test # 7 | 0.080 |
| 0.20 g/210L Test # 7 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.079 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.048 |
| 0.08 g/210L Test # 8 | 0.080 |
| 0.20 g/210L Test # 8 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.079 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.048 |
| 0.08 g/210L Test # 9 | 0.080 |
| 0.20 g/210L Test # 9 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.079 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.048 |
| 0.08 g/210L Test # 10 | 0.080 |
| 0.20 g/210L Test # 10 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.079 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0000 |
| 0.08 g/210L Test (Std Dev) | 0.0000 |
| 0.20 g/210L Test (Std Dev) | 0.0003 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0003 |

Remarks:

Data Download Date/Time 01/28/2026 13:36

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007457

| | | |
|---|------------------------|-------------------------------|
| Agency FL HIGHWAY PATROL | Date 02/26/2026 | Type Agency Inspection |
| Inspector WILDEMAN-STUART,JORDAN,R | Time 17:30:25 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|---------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 403516666 |
| Expiration 06/24/2027 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 12/16/2028 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.202 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.201 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 03/02/2026 13:47

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007457

| | | |
|---|------------------------|-------------------------------|
| Agency FL HIGHWAY PATROL | Date 03/02/2026 | Type Agency Inspection |
| Inspector WILDEMAN-STUART,JORDAN,R | Time 13:42:25 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|---------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 403516666 |
| Expiration 06/24/2027 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 12/16/2028 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.000 |
| Minimum Sample Volume Check: OK | | 0.05 g/210L Test # (Repeat) 1 | 0.049 |
| Barometric Pressure Sensor Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.20 g/210L Test # 1 | 0.200 |
| Mouth Alcohol Test: Slope Not Met | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test: Interferent Detect | Yes | Interferent Detect Test # 1 | INT |
| Diagnostic Check (Post-Inspection): OK | Yes | Alcohol Free Test # 2 | 0.000 |
| Alcohol Free Test: OK | Yes | 0.05 g/210L Test # 2 | 0.000 |
| 0.05 g/210L Test: OK | Yes | 0.05 g/210L Test # (Repeat) 2 | 0.049 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| | | Interferent Detect Test # 2 | INT |
| | | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.000 |
| | | 0.05 g/210L Test # (Repeat) 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.201 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

05:Control Outside Tolerance.

Data Download Date/Time 03/02/2026 13:47

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007478

| | | |
|--------------------------------------|------------------------|-------------------------------|
| Agency BAY COUNTY SO | Date 03/23/2026 | Type Agency Inspection |
| Inspector PIERSON, WEN-CHI, K | Time 07:54:59 | Software 8100.27 |

| | | | |
|-------------------|-------------------|-------------------|-----------------------|
| 0.05 Lot # | 0.08 Lot # | 0.20 Lot # | 0.08 Gas Lot # |
| Expiration | Expiration | Expiration | Expiration |

Number of Simulators **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|------|--------|
| Diagnostic Check (Pre-Inspection): OK | No | | |
| Date and/or Time Adjusted | No | | |
| Minimum Sample Volume Check: OK | | | |
| Barometric Pressure Sensor Check: OK | | | |
| Alcohol Free Subject Test: 0.000 | No | | |
| Mouth Alcohol Test: Slope Not Met | No | | |
| Interferent Detect Test: Interferent Detect | No | | |
| Diagnostic Check (Post-Inspection): OK | No | | |
| Alcohol Free Test: OK | No | | |
| 0.05 g/210L Test: OK | No | | |
| 0.08 g/210L Test: OK | No | | |
| 0.20 g/210L Test: OK | No | | |
| 0.08 g/210L Dry Gas Std Test: OK | No | | |

Remarks:

BYPASS AI TO OPERATE INSTRUMENT.COMPLIANCE NOT DETERMINED.

Data Download Date/Time 03/23/2026 13:14

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007478

| | | | |
|--------------------------------------|--|-----------------------------------|----------------------------------|
| Agency BAY COUNTY SO | Date 03/23/2026 | Type Department Inspection | |
| Inspector PIERSON, WEN-CHI, K | Time 12:13:54 | Software 8100.27 | |
| 0.05 Lot # 202406K | 0.08 Lot # 202406L | 0.20 Lot # 202406N | 0.08 Gas Lot # 23725080A1 |
| Expiration 06/19/2026 | Expiration 06/19/2026 | Expiration 06/20/2026 | Expiration 10/05/2027 |
| Number of Simulators 5 | Average Standard Deviation 0.0006 | Compliance Yes | |

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | Yes | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | Yes | 0.20 g/210L Test # 1 | 0.202 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.077 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.076 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.203 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.077 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.077 |
| | | 0.20 g/210L Test # 3 | 0.202 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.077 |
| | | Interferent Detect Test # 3 | INT |
| | | Alcohol Free Test # 4 | 0.000 |
| | | 0.05 g/210L Test # 4 | 0.049 |
| | | 0.08 g/210L Test # 4 | 0.078 |
| | | 0.20 g/210L Test # 4 | 0.202 |
| | | 0.08 g/210L Dry Gas Std Test # 4 | 0.077 |
| | | Interferent Detect Test # 4 | INT |
| | | Alcohol Free Test # 5 | 0.000 |
| | | 0.05 g/210L Test # 5 | 0.049 |
| | | 0.08 g/210L Test # 5 | 0.079 |
| | | 0.20 g/210L Test # 5 | 0.202 |
| | | 0.08 g/210L Dry Gas Std Test # 5 | 0.077 |
| | | Interferent Detect Test # 5 | INT |

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007478

| | |
|--|--------|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.049 |
| 0.08 g/210L Test # 6 | 0.079 |
| 0.20 g/210L Test # 6 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.077 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.048 |
| 0.08 g/210L Test # 7 | 0.079 |
| 0.20 g/210L Test # 7 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.077 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.049 |
| 0.08 g/210L Test # 8 | 0.079 |
| 0.20 g/210L Test # 8 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.077 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.049 |
| 0.08 g/210L Test # 9 | 0.080 |
| 0.20 g/210L Test # 9 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.077 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.049 |
| 0.08 g/210L Test # 10 | 0.079 |
| 0.20 g/210L Test # 10 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.077 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Test (Std Dev) | 0.0011 |
| 0.20 g/210L Test (Std Dev) | 0.0009 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0000 |

Remarks:

Data Download Date/Time 03/23/2026 13:14

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007478

| | | | |
|---------------------------------|------------------------|-------------------------------|--|
| Agency BAY COUNTY SO | Date 03/27/2026 | Type Agency Inspection | |
| Inspector MILLER,BRIAN,D | Time 12:35:35 | Software 8100.27 | |

| | | | |
|------------------------------|------------------------------|------------------------------|---------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 402886996 |
| Expiration 06/24/2027 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 11/01/2026 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.201 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.051 |
| | | 0.08 g/210L Test # 3 | 0.081 |
| | | 0.20 g/210L Test # 3 | 0.202 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 03/27/2026 14:33

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007478

| | | | |
|-----------------------------------|------------------------|-------------------------------|--|
| Agency BAY COUNTY SO | Date 04/29/2026 | Type Agency Inspection | |
| Inspector MILLER, BRIAN, D | Time 09:55:33 | Software 8100.27 | |

| | | | |
|------------------------------|------------------------------|------------------------------|---------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 402886996 |
| Expiration 06/24/2027 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 11/01/2026 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.201 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.081 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.202 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.081 |
| | | 0.20 g/210L Test # 3 | 0.202 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 04/29/2026 12:01

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007480

| | | |
|----------------------------------|------------------------|-------------------------------|
| Agency HILLSBOROUGH CSO | Date 01/15/2026 | Type Agency Inspection |
| Inspector BROWN, KEVIN, C | Time 10:43:52 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 04125080A6 |
| Expiration 06/24/2027 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 03/05/2027 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

AI B KROME #263389 PRESENT FOR INSPECTION

Data Download Date/Time 01/15/2026 10:59

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007480

| | | | |
|-------------------------|-----------------|------------------------|--|
| Agency HILLSBOROUGH CSO | Date 02/12/2026 | Type Agency Inspection | |
| Inspector BROWN,KEVIN,C | Time 08:31:00 | Software 8100.27 | |

| | | | |
|-----------------------|------------|------------|----------------|
| 0.05 Lot # 202506E | 0.08 Lot # | 0.20 Lot # | 0.08 Gas Lot # |
| Expiration 06/24/2027 | Expiration | Expiration | Expiration |

Number of Simulators 5 Compliance No

| Check or Test | Result | Test | Result |
|---|--------|-------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.000 |
| Minimum Sample Volume Check: OK | | 0.05 g/210L Test # (Repeat) 1 | 0.000 |
| Barometric Pressure Sensor Check: OK | | Interferent Detect Test # 1 | INT |
| Alcohol Free Subject Test: 0.000 | Yes | Alcohol Free Test # 2 | 0.000 |
| Mouth Alcohol Test: Slope Not Met | Yes | 0.05 g/210L Test # 2 | 0.000 |
| Interferent Detect Test: Interferent Detect | Yes | 0.05 g/210L Test # (Repeat) 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | No | Interferent Detect Test # 2 | INT |
| Alcohol Free Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test: OK | No | 0.05 g/210L Test # 3 | 0.000 |
| 0.08 g/210L Test: OK | No | 0.05 g/210L Test # (Repeat) 3 | 0.000 |
| 0.20 g/210L Test: OK | No | Interferent Detect Test # 3 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | No | | |

Remarks:

05: Control Outside Tolerance, Control Outside Tolerance. Non-compliance:CONTROL OUTSIDE TOLERANCE FOR .05 SOLUTION .

Data Download Date/Time 02/12/2026 11:32

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007480

| | | | |
|----------------------------------|--|-----------------------------------|----------------------------------|
| Agency HILLSBOROUGH CSO | Date 02/26/2026 | Type Department Inspection | |
| Inspector SPEARIN,KATIE,T | Time 08:48:22 | Software 8100.27 | |
| 0.05 Lot # 202406K | 0.08 Lot # 202406L | 0.20 Lot # 202406N | 0.08 Gas Lot # 28424080A3 |
| Expiration 06/19/2026 | Expiration 06/19/2026 | Expiration 06/20/2026 | Expiration 11/05/2026 |
| Number of Simulators 5 | Average Standard Deviation 0.0003 | Compliance Yes | |

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.047 |
| Minimum Sample Volume Check: OK | Yes | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | Yes | 0.20 g/210L Test # 1 | 0.196 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |
| | | Alcohol Free Test # 4 | 0.000 |
| | | 0.05 g/210L Test # 4 | 0.048 |
| | | 0.08 g/210L Test # 4 | 0.078 |
| | | 0.20 g/210L Test # 4 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 4 | 0.079 |
| | | Interferent Detect Test # 4 | INT |
| | | Alcohol Free Test # 5 | 0.000 |
| | | 0.05 g/210L Test # 5 | 0.048 |
| | | 0.08 g/210L Test # 5 | 0.078 |
| | | 0.20 g/210L Test # 5 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 5 | 0.079 |
| | | Interferent Detect Test # 5 | INT |

**Florida Department of Law Enforcement
Alcohol Testing Program**

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007480

| | |
|--|--------|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.048 |
| 0.08 g/210L Test # 6 | 0.078 |
| 0.20 g/210L Test # 6 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.079 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.048 |
| 0.08 g/210L Test # 7 | 0.078 |
| 0.20 g/210L Test # 7 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.079 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.048 |
| 0.08 g/210L Test # 8 | 0.078 |
| 0.20 g/210L Test # 8 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.079 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.048 |
| 0.08 g/210L Test # 9 | 0.078 |
| 0.20 g/210L Test # 9 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.079 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.048 |
| 0.08 g/210L Test # 10 | 0.079 |
| 0.20 g/210L Test # 10 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.078 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Test (Std Dev) | 0.0003 |
| 0.20 g/210L Test (Std Dev) | 0.0003 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0005 |

Remarks:

Data Download Date/Time 02/26/2026 09:51

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007480

| | | |
|----------------------------------|------------------------|-------------------------------|
| Agency HILLSBOROUGH CSO | Date 04/09/2026 | Type Agency Inspection |
| Inspector BROWN, KEVIN, C | Time 08:17:57 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 29425080A2 |
| Expiration 06/24/2027 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 12/05/2027 |

Number of Simulators 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.195 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.196 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

AI B KROME #263389 PRESENT FOR INSPECTION

Data Download Date/Time 04/09/2026 09:42

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007481

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| Agency HILLSBOROUGH CSO | Date 01/15/2026 | Type Agency Inspection | |
| Inspector BROWN,KEVIN,C | Time 10:36:03 | Software 8100.27 | |
| 0.05 Lot # 202506E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 29425080A2 |
| Expiration 06/24/2027 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 12/05/2027 |

Number of Simulators 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.201 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.202 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.202 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

AI B KROME #263389 PRESENT FOR INSPECTION

Data Download Date/Time 01/15/2026 11:22

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007481

| | | |
|--------------------------------|------------------------|-------------------------------|
| Agency HILLSBOROUGH CSO | Date 02/12/2026 | Type Agency Inspection |
| Inspector BROWN,KEVIN,C | Time 09:27:37 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 29425080A2 |
| Expiration 06/24/2027 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 12/05/2027 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.000 |
| Minimum Sample Volume Check: OK | | 0.05 g/210L Test # (Repeat) 1 | 0.048 |
| Barometric Pressure Sensor Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.20 g/210L Test # 1 | 0.199 |
| Mouth Alcohol Test: Slope Not Met | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test: Interferent Detect | Yes | Interferent Detect Test # 1 | INT |
| Diagnostic Check (Post-Inspection): OK | Yes | Alcohol Free Test # 2 | 0.000 |
| Alcohol Free Test: OK | Yes | 0.05 g/210L Test # 2 | 0.000 |
| 0.05 g/210L Test: OK | Yes | 0.05 g/210L Test # (Repeat) 2 | 0.049 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| | | Interferent Detect Test # 2 | INT |
| | | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.000 |
| | | 0.05 g/210L Test # (Repeat) 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

AI B KROME #263389 PRESENT FOR INSPECTION.

Data Download Date/Time 02/12/2026 11:37

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007481

| | | |
|--------------------------------|------------------------|-------------------------------|
| Agency HILLSBOROUGH CSO | Date 03/12/2026 | Type Agency Inspection |
| Inspector BROWN,KEVIN,C | Time 12:18:50 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 04125080A6 |
| Expiration 06/24/2027 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 03/05/2027 |

Number of Simulators 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.047 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.195 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.077 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.077 |
| | | 0.20 g/210L Test # 3 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

AI B KROME #263389 PRESENT FOR INSPECTION

Data Download Date/Time 03/12/2026 12:19

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007481

| | | |
|----------------------------------|------------------------|-------------------------------|
| Agency HILLSBOROUGH CSO | Date 04/09/2026 | Type Agency Inspection |
| Inspector BROWN, KEVIN, C | Time 11:44:30 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 04125080A6 |
| Expiration 06/24/2027 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 03/05/2027 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

AI B KROME #263389 PRESENT FOR INSPECTION

Data Download Date/Time 04/09/2026 11:48

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

| | | |
|--------------------------------|------------------------|-------------------------------|
| Agency HILLSBOROUGH CSO | Date 01/15/2026 | Type Agency Inspection |
| Inspector BROWN,KEVIN,C | Time 08:15:11 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 04125080A6 |
| Expiration 06/24/2027 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 03/05/2027 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

AI B KROME #263389 PRESENT FOR INSPECTION

Data Download Date/Time 01/15/2026 09:17

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007483

| | | | |
|----------------------------------|--|-----------------------------------|----------------------------------|
| Agency HILLSBOROUGH CSO | Date 02/26/2026 | Type Department Inspection | |
| Inspector SPEARIN,KATIE,T | Time 09:01:03 | Software 8100.27 | |
| 0.05 Lot # 202406K | 0.08 Lot # 202406N | 0.20 Lot # 202406N | 0.08 Gas Lot # 28424080A3 |
| Expiration 06/19/2026 | Expiration 06/20/2026 | Expiration 06/20/2026 | Expiration 11/05/2026 |
| Number of Simulators 5 | Average Standard Deviation 0.0003 | Compliance Yes | |

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.047 |
| Minimum Sample Volume Check: OK | Yes | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | Yes | 0.20 g/210L Test # 1 | 0.202 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.047 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.202 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.047 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.202 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.077 |
| | | Interferent Detect Test # 3 | INT |
| | | Alcohol Free Test # 4 | 0.000 |
| | | 0.05 g/210L Test # 4 | 0.047 |
| | | 0.08 g/210L Test # 4 | 0.078 |
| | | 0.20 g/210L Test # 4 | 0.201 |
| | | 0.08 g/210L Dry Gas Std Test # 4 | 0.078 |
| | | Interferent Detect Test # 4 | INT |
| | | Alcohol Free Test # 5 | 0.000 |
| | | 0.05 g/210L Test # 5 | 0.047 |
| | | 0.08 g/210L Test # 5 | 0.077 |
| | | 0.20 g/210L Test # 5 | 0.202 |
| | | 0.08 g/210L Dry Gas Std Test # 5 | 0.078 |
| | | Interferent Detect Test # 5 | INT |

Florida Department of Law Enforcement
Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007483

| | |
|--|--------|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.047 |
| 0.08 g/210L Test # 6 | 0.078 |
| 0.20 g/210L Test # 6 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.078 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.047 |
| 0.08 g/210L Test # 7 | 0.078 |
| 0.20 g/210L Test # 7 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.078 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.047 |
| 0.08 g/210L Test # 8 | 0.078 |
| 0.20 g/210L Test # 8 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.077 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.047 |
| 0.08 g/210L Test # 9 | 0.078 |
| 0.20 g/210L Test # 9 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.077 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.047 |
| 0.08 g/210L Test # 10 | 0.078 |
| 0.20 g/210L Test # 10 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.077 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0000 |
| 0.08 g/210L Test (Std Dev) | 0.0003 |
| 0.20 g/210L Test (Std Dev) | 0.0005 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0005 |

Remarks:

Data Download Date/Time 02/26/2026 09:52

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007483

| | | |
|--------------------------------|------------------------|-------------------------------|
| Agency HILLSBOROUGH CSO | Date 04/09/2026 | Type Agency Inspection |
| Inspector BROWN,KEVIN,C | Time 11:58:52 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 04125080A6 |
| Expiration 06/24/2027 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 03/05/2027 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.200 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.077 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.077 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.202 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.076 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.077 |
| | | 0.20 g/210L Test # 3 | 0.202 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.077 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

AI B KROME #263389 PRESENT FOR INSPECTION

Data Download Date/Time 04/09/2026 12:44

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007484

| | | |
|----------------------------------|------------------------|-------------------------------|
| Agency HILLSBOROUGH CSO | Date 01/15/2026 | Type Agency Inspection |
| Inspector BROWN, KEVIN, C | Time 10:20:58 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 04125080A6 |
| Expiration 06/24/2027 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 03/05/2027 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.200 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.201 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

AI B KROME #263389 PRESENT FOR INSPECTION

Data Download Date/Time 01/15/2026 11:34

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007484

| | | |
|----------------------------------|------------------------|-------------------------------|
| Agency HILLSBOROUGH CSO | Date 02/12/2026 | Type Agency Inspection |
| Inspector BROWN, KEVIN, C | Time 12:14:38 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 202410F | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 04125080A6 |
| Expiration 10/30/2026 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 03/05/2027 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.201 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

AI B KROME #263389 PRESENT FOR INSPECTION.

Data Download Date/Time 02/12/2026 12:43

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007484

| | | |
|----------------------------------|------------------------|-------------------------------|
| Agency HILLSBOROUGH CSO | Date 03/12/2026 | Type Agency Inspection |
| Inspector BROWN, KEVIN, C | Time 08:42:51 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 04125080A6 |
| Expiration 06/24/2027 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 03/05/2027 |

Number of Simulators 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.047 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.193 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.047 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.194 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.194 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

AI B KROME #263389 PRESENT FOR INSPECTION

Data Download Date/Time 03/12/2026 11:28

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007484

| | | |
|----------------------------------|------------------------|-------------------------------|
| Agency HILLSBOROUGH CSO | Date 04/09/2026 | Type Agency Inspection |
| Inspector BROWN, KEVIN, C | Time 11:42:32 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 29425080A2 |
| Expiration 06/24/2027 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 12/05/2027 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

AI B KROME #263389 PRESENT FOR INSPECTION

Data Download Date/Time 04/09/2026 12:28

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007492

| | | |
|----------------------------------|------------------------|-------------------------------|
| Agency FFWCC | Date 01/08/2026 | Type Agency Inspection |
| Inspector PULASKI,DAMON,J | Time 14:11:42 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202510E | 0.20 Lot # 202410F | 0.08 Gas Lot # 10124080A2 |
| Expiration 06/24/2027 | Expiration 10/28/2027 | Expiration 10/30/2026 | Expiration 06/05/2026 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 02/03/2026 18:43

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007492

| | | |
|----------------------------------|------------------------|-------------------------------|
| Agency FFWCC | Date 02/02/2026 | Type Agency Inspection |
| Inspector PULASKI,DAMON,J | Time 21:09:52 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202510E | 0.20 Lot # 202410F | 0.08 Gas Lot # 10124080A2 |
| Expiration 06/24/2027 | Expiration 10/28/2027 | Expiration 10/30/2026 | Expiration 06/05/2026 |

Number of Simulators 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.081 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.081 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 02/03/2026 18:43

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007492

| | | | |
|----------------------------------|--|-----------------------------------|----------------------------------|
| Agency FFWCC | Date 02/26/2026 | Type Department Inspection | |
| Inspector SPEARIN,KATIE,T | Time 10:52:38 | Software 8100.27 | |
| 0.05 Lot # 202406K | 0.08 Lot # 202406L | 0.20 Lot # 202406N | 0.08 Gas Lot # 28424080A3 |
| Expiration 06/19/2026 | Expiration 06/19/2026 | Expiration 06/20/2026 | Expiration 11/05/2026 |
| Number of Simulators 5 | Average Standard Deviation 0.0003 | Compliance Yes | |

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | Yes | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | Yes | 0.20 g/210L Test # 1 | 0.196 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.196 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |
| | | Alcohol Free Test # 4 | 0.000 |
| | | 0.05 g/210L Test # 4 | 0.048 |
| | | 0.08 g/210L Test # 4 | 0.078 |
| | | 0.20 g/210L Test # 4 | 0.196 |
| | | 0.08 g/210L Dry Gas Std Test # 4 | 0.079 |
| | | Interferent Detect Test # 4 | INT |
| | | Alcohol Free Test # 5 | 0.000 |
| | | 0.05 g/210L Test # 5 | 0.048 |
| | | 0.08 g/210L Test # 5 | 0.079 |
| | | 0.20 g/210L Test # 5 | 0.196 |
| | | 0.08 g/210L Dry Gas Std Test # 5 | 0.079 |
| | | Interferent Detect Test # 5 | INT |

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007492

| | |
|--|--------|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.048 |
| 0.08 g/210L Test # 6 | 0.078 |
| 0.20 g/210L Test # 6 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.079 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.048 |
| 0.08 g/210L Test # 7 | 0.078 |
| 0.20 g/210L Test # 7 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.079 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.048 |
| 0.08 g/210L Test # 8 | 0.078 |
| 0.20 g/210L Test # 8 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.079 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.048 |
| 0.08 g/210L Test # 9 | 0.079 |
| 0.20 g/210L Test # 9 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.078 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.048 |
| 0.08 g/210L Test # 10 | 0.079 |
| 0.20 g/210L Test # 10 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.078 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0000 |
| 0.08 g/210L Test (Std Dev) | 0.0004 |
| 0.20 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0004 |

Remarks:

Data Download Date/Time 02/26/2026 11:19

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007492

| | | |
|----------------------------------|------------------------|-------------------------------|
| Agency FFWCC | Date 03/16/2026 | Type Agency Inspection |
| Inspector PULASKI,DAMON,J | Time 16:53:57 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202510E | 0.20 Lot # 202410F | 0.08 Gas Lot # 10124080A2 |
| Expiration 06/24/2027 | Expiration 10/28/2027 | Expiration 10/30/2026 | Expiration 06/05/2026 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

IN COMPLIANCE

Data Download Date/Time 03/17/2026 19:41

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007494

| | | |
|--------------------------------------|------------------------|-------------------------------|
| Agency FFWCC | Date 01/15/2026 | Type Agency Inspection |
| Inspector ARBOGAST, JOSHUA, S | Time 13:02:14 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|---------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202510E | 0.20 Lot # 202410F | 0.08 Gas Lot # 403008479 |
| Expiration 06/24/2027 | Expiration 10/28/2027 | Expiration 10/30/2026 | Expiration 03/28/2027 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 01/15/2026 13:11

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007494

| | | |
|--------------------------------------|------------------------|-------------------------------|
| Agency FFWCC | Date 02/23/2026 | Type Agency Inspection |
| Inspector ARBOGAST, JOSHUA, S | Time 11:34:22 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|---------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202510E | 0.20 Lot # 202410F | 0.08 Gas Lot # 403008479 |
| Expiration 06/24/2027 | Expiration 10/28/2027 | Expiration 10/30/2026 | Expiration 03/28/2027 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.051 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.081 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.052 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.081 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.052 |
| | | 0.08 g/210L Test # 3 | 0.082 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 02/23/2026 11:54

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007494

| | | |
|--------------------------------------|------------------------|-------------------------------|
| Agency FFWCC | Date 03/28/2026 | Type Agency Inspection |
| Inspector ARBOGAST, JOSHUA, S | Time 16:15:46 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|---------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202510E | 0.20 Lot # 202410F | 0.08 Gas Lot # 403008479 |
| Expiration 06/24/2027 | Expiration 10/28/2027 | Expiration 10/30/2026 | Expiration 03/28/2027 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.081 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.081 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.081 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

IN COMPLIANCE

Data Download Date/Time 03/28/2026 16:34

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007494

| | | |
|--------------------------------------|------------------------|-------------------------------|
| Agency FFWCC | Date 04/22/2026 | Type Agency Inspection |
| Inspector ARBOGAST, JOSHUA, S | Time 11:31:36 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|---------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202510E | 0.20 Lot # 202410F | 0.08 Gas Lot # 403008479 |
| Expiration 06/24/2027 | Expiration 10/28/2027 | Expiration 10/30/2026 | Expiration 03/28/2027 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.081 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.051 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.081 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.051 |
| | | 0.08 g/210L Test # 3 | 0.081 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 04/22/2026 12:02

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007495

| | | |
|--------------------------------------|------------------------|-------------------------------|
| Agency FFWCC | Date 01/30/2026 | Type Agency Inspection |
| Inspector MARROQUIN,RICHARD,J | Time 16:41:05 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 202402E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 00925080A3 |
| Expiration 02/27/2026 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 02/05/2027 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # (Repeat) 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # (Repeat) 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # (Repeat) 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

00: Ambient Fail.

Data Download Date/Time 02/11/2026 13:52

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007495

| | | | |
|--------------------------------------|--|-----------------------------------|----------------------------------|
| Agency FFWCC | Date 02/26/2026 | Type Department Inspection | |
| Inspector PIERSON, WEN-CHI, K | Time 16:04:00 | Software 8100.27 | |
| 0.05 Lot # 202406K | 0.08 Lot # 202406L | 0.20 Lot # 202406N | 0.08 Gas Lot # 28424080A3 |
| Expiration 06/19/2026 | Expiration 06/19/2026 | Expiration 06/20/2026 | Expiration 11/05/2026 |
| Number of Simulators 5 | Average Standard Deviation 0.0004 | Compliance No | |

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | Yes | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | Yes | 0.20 g/210L Test # 1 | 0.196 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | No | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.196 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |
| | | Alcohol Free Test # 4 | 0.000 |
| | | 0.05 g/210L Test # 4 | 0.049 |
| | | 0.08 g/210L Test # 4 | 0.079 |
| | | 0.20 g/210L Test # 4 | 0.196 |
| | | 0.08 g/210L Dry Gas Std Test # 4 | 0.081 |
| | | Interferent Detect Test # 4 | INT |
| | | Alcohol Free Test # 5 | 0.000 |
| | | 0.05 g/210L Test # 5 | 0.049 |
| | | 0.08 g/210L Test # 5 | 0.080 |
| | | 0.20 g/210L Test # 5 | 0.196 |
| | | 0.08 g/210L Dry Gas Std Test # 5 | 0.081 |
| | | Interferent Detect Test # 5 | INT |

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007495

| | |
|--|--------|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.050 |
| 0.08 g/210L Test # 6 | 0.080 |
| 0.20 g/210L Test # 6 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.081 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.050 |
| 0.08 g/210L Test # 7 | 0.079 |
| 0.20 g/210L Test # 7 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.081 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.050 |
| 0.08 g/210L Test # 8 | 0.079 |
| 0.20 g/210L Test # 8 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.081 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.050 |
| 0.08 g/210L Test # 9 | 0.079 |
| 0.20 g/210L Test # 9 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.081 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.050 |
| 0.08 g/210L Test # 10 | 0.079 |
| 0.20 g/210L Test # 10 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.082 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0005 |
| 0.08 g/210L Test (Std Dev) | 0.0004 |
| 0.20 g/210L Test (Std Dev) | 0.0005 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0003 |

Remarks:

Non-compliance: .

Data Download Date/Time 03/02/2026 15:22

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007495

| | | | |
|--------------------------------------|--|-----------------------------------|----------------------------------|
| Agency FFWCC | Date 03/02/2026 | Type Department Inspection | |
| Inspector PIERSON, WEN-CHI, K | Time 15:18:56 | Software 8100.27 | |
| 0.05 Lot # 202406K | 0.08 Lot # 202406L | 0.20 Lot # 202406N | 0.08 Gas Lot # 28424080A3 |
| Expiration 06/19/2026 | Expiration 06/19/2026 | Expiration 06/20/2026 | Expiration 11/05/2026 |
| Number of Simulators 5 | Average Standard Deviation 0.0005 | Compliance Yes | |

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | Yes | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | Yes | 0.20 g/210L Test # 1 | 0.196 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |
| | | Alcohol Free Test # 4 | 0.000 |
| | | 0.05 g/210L Test # 4 | 0.049 |
| | | 0.08 g/210L Test # 4 | 0.079 |
| | | 0.20 g/210L Test # 4 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 4 | 0.081 |
| | | Interferent Detect Test # 4 | INT |
| | | Alcohol Free Test # 5 | 0.000 |
| | | 0.05 g/210L Test # 5 | 0.049 |
| | | 0.08 g/210L Test # 5 | 0.079 |
| | | 0.20 g/210L Test # 5 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 5 | 0.082 |
| | | Interferent Detect Test # 5 | INT |

Florida Department of Law Enforcement
Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007495

| | |
|--|--------|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.050 |
| 0.08 g/210L Test # 6 | 0.079 |
| 0.20 g/210L Test # 6 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.081 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.050 |
| 0.08 g/210L Test # 7 | 0.079 |
| 0.20 g/210L Test # 7 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.082 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.050 |
| 0.08 g/210L Test # 8 | 0.080 |
| 0.20 g/210L Test # 8 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.082 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.050 |
| 0.08 g/210L Test # 9 | 0.080 |
| 0.20 g/210L Test # 9 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.082 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.050 |
| 0.08 g/210L Test # 10 | 0.079 |
| 0.20 g/210L Test # 10 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.082 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0005 |
| 0.08 g/210L Test (Std Dev) | 0.0004 |
| 0.20 g/210L Test (Std Dev) | 0.0006 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0005 |

Remarks:

Data Download Date/Time 03/02/2026 15:22

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007495

| | | |
|--------------------------------------|------------------------|-------------------------------|
| Agency FFWCC | Date 03/13/2026 | Type Agency Inspection |
| Inspector MARROQUIN,RICHARD,J | Time 16:56:29 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 00925080A3 |
| Expiration 06/24/2027 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 02/05/2027 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.081 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 03/13/2026 17:03

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007495

| | | |
|--------------------------------------|------------------------|-------------------------------|
| Agency FFWCC | Date 04/19/2026 | Type Agency Inspection |
| Inspector LAWRENCE, JOSHUA, S | Time 15:18:44 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202510E | 0.20 Lot # 202510F | 0.08 Gas Lot # 00925080A3 |
| Expiration 06/24/2027 | Expiration 10/28/2027 | Expiration 10/29/2027 | Expiration 02/05/2027 |

Number of Simulators 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.200 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.051 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.051 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 04/20/2026 09:25

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007496

| | | |
|--------------------------------------|------------------------|-------------------------------|
| Agency FFWCC | Date 01/30/2026 | Type Agency Inspection |
| Inspector MARROQUIN,RICHARD,J | Time 16:05:15 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 202402E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 00925080A3 |
| Expiration 02/27/2026 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 02/05/2027 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 02/11/2026 13:46

Florida Department of Law Enforcement
Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007496

| | | |
|--------------------------------------|------------------------|-------------------------------|
| Agency FFWCC | Date 03/02/2026 | Type Agency Inspection |
| Inspector PIERSON, WEN-CHI, K | Time 09:44:46 | Software 8100.27 |

| | | | |
|-------------------|-------------------|-------------------|-----------------------|
| 0.05 Lot # | 0.08 Lot # | 0.20 Lot # | 0.08 Gas Lot # |
| Expiration | Expiration | Expiration | Expiration |

| | |
|-----------------------------|-----------------------|
| Number of Simulators | Compliance Yes |
|-----------------------------|-----------------------|

| Check or Test | Result | Test | Result |
|---|--------|------|--------|
| Diagnostic Check (Pre-Inspection): OK | No | | |
| Date and/or Time Adjusted | No | | |
| Minimum Sample Volume Check: OK | | | |
| Barometric Pressure Sensor Check: OK | | | |
| Alcohol Free Subject Test: 0.000 | No | | |
| Mouth Alcohol Test: Slope Not Met | No | | |
| Interferent Detect Test: Interferent Detect | No | | |
| Diagnostic Check (Post-Inspection): OK | No | | |
| Alcohol Free Test: OK | No | | |
| 0.05 g/210L Test: OK | No | | |
| 0.08 g/210L Test: OK | No | | |
| 0.20 g/210L Test: OK | No | | |
| 0.08 g/210L Dry Gas Std Test: OK | No | | |

Remarks:

| |
|---|
| BYPASS AI TO OPERATE INSTRUMENT. COMPLIANCE NOT DETERMINED. |
|---|

Data Download Date/Time 03/02/2026 12:59

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007496

| | | | |
|--------------------------------------|--|-----------------------------------|----------------------------------|
| Agency FFWCC | Date 03/02/2026 | Type Department Inspection | |
| Inspector PIERSON, WEN-CHI, K | Time 12:51:35 | Software 8100.27 | |
| 0.05 Lot # 202406K | 0.08 Lot # 202406L | 0.20 Lot # 202406N | 0.08 Gas Lot # 28424080A3 |
| Expiration 06/19/2026 | Expiration 06/19/2026 | Expiration 06/20/2026 | Expiration 11/05/2026 |
| Number of Simulators 5 | Average Standard Deviation 0.0003 | Compliance Yes | |

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | Yes | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | Yes | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |
| | | Alcohol Free Test # 4 | 0.000 |
| | | 0.05 g/210L Test # 4 | 0.049 |
| | | 0.08 g/210L Test # 4 | 0.079 |
| | | 0.20 g/210L Test # 4 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 4 | 0.081 |
| | | Interferent Detect Test # 4 | INT |
| | | Alcohol Free Test # 5 | 0.000 |
| | | 0.05 g/210L Test # 5 | 0.048 |
| | | 0.08 g/210L Test # 5 | 0.079 |
| | | 0.20 g/210L Test # 5 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 5 | 0.081 |
| | | Interferent Detect Test # 5 | INT |

Florida Department of Law Enforcement
Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007496

| | |
|--|--------|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.049 |
| 0.08 g/210L Test # 6 | 0.079 |
| 0.20 g/210L Test # 6 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.081 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.049 |
| 0.08 g/210L Test # 7 | 0.079 |
| 0.20 g/210L Test # 7 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.081 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.049 |
| 0.08 g/210L Test # 8 | 0.079 |
| 0.20 g/210L Test # 8 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.081 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.049 |
| 0.08 g/210L Test # 9 | 0.079 |
| 0.20 g/210L Test # 9 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.082 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.049 |
| 0.08 g/210L Test # 10 | 0.079 |
| 0.20 g/210L Test # 10 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.081 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Test (Std Dev) | 0.0004 |
| 0.20 g/210L Test (Std Dev) | 0.0000 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0004 |

Remarks:

Data Download Date/Time 03/02/2026 12:59

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007496

| | | |
|--------------------------------------|------------------------|-------------------------------|
| Agency FFWCC | Date 03/13/2026 | Type Agency Inspection |
| Inspector MARROQUIN,RICHARD,J | Time 16:52:46 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 00925080A3 |
| Expiration 06/24/2027 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 02/05/2027 |

Number of Simulators 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 03/13/2026 17:06

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007496

| | | |
|--------------------------------------|------------------------|-------------------------------|
| Agency FFWCC | Date 04/19/2026 | Type Agency Inspection |
| Inspector LAWRENCE, JOSHUA, S | Time 14:58:06 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202510E | 0.20 Lot # 202510F | 0.08 Gas Lot # 00925080A3 |
| Expiration 06/24/2027 | Expiration 10/28/2027 | Expiration 10/29/2027 | Expiration 02/05/2027 |

Number of Simulators 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.081 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 04/20/2026 09:31

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007498

| | | | |
|----------------------------------|------------------------------|-------------------------------|---------------------------------|
| Agency FFWCC | Date 01/20/2026 | Type Agency Inspection | |
| Inspector BROOKS,JOSEPH,P | Time 14:55:04 | Software 8100.27 | |
| 0.05 Lot # 202506E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 403321942 |
| Expiration 06/24/2027 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 04/14/2028 |
| Number of Simulators 5 | | Compliance Yes | |

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.196 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.051 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 01/22/2026 13:58

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007498

| | | | |
|----------------------------------|------------------------------|-------------------------------|---------------------------------|
| Agency FFWCC | Date 02/02/2026 | Type Agency Inspection | |
| Inspector BROOKS,JOSEPH,P | Time 10:29:10 | Software 8100.27 | |
| 0.05 Lot # 202506E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 403321942 |
| Expiration 06/24/2027 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 04/14/2028 |
| Number of Simulators 5 | | Compliance Yes | |

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.051 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.051 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.081 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.051 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 02/04/2026 10:42

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007498

| | | | |
|----------------------------------|------------------------------|-------------------------------|---------------------------------|
| Agency FFWCC | Date 03/23/2026 | Type Agency Inspection | |
| Inspector BROOKS,JOSEPH,P | Time 12:39:26 | Software 8100.27 | |
| 0.05 Lot # 202506E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 403321942 |
| Expiration 06/24/2027 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 04/14/2028 |
| Number of Simulators 5 | | Compliance Yes | |

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 03/23/2026 14:58

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007498

| | | | |
|----------------------------------|------------------------------|-------------------------------|---------------------------------|
| Agency FFWCC | Date 04/23/2026 | Type Agency Inspection | |
| Inspector BROOKS,JOSEPH,P | Time 17:09:16 | Software 8100.27 | |
| 0.05 Lot # 202506E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 403321942 |
| Expiration 06/24/2027 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 04/14/2028 |
| Number of Simulators 4 | | Compliance Yes | |

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.051 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 04/23/2026 19:17

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007499

| | | | |
|----------------------------------|------------------------------|-------------------------------|---------------------------------|
| Agency FFWCC | Date 01/20/2026 | Type Agency Inspection | |
| Inspector BROOKS,JOSEPH,P | Time 16:07:01 | Software 8100.27 | |
| 0.05 Lot # 202506E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 403207508 |
| Expiration 06/24/2027 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 12/02/2027 |
| Number of Simulators 5 | | Compliance Yes | |

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 01/22/2026 13:53

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007499

| | | | |
|----------------------------------|------------------------------|-------------------------------|---------------------------------|
| Agency FFWCC | Date 02/02/2026 | Type Agency Inspection | |
| Inspector BROOKS,JOSEPH,P | Time 11:34:29 | Software 8100.27 | |
| 0.05 Lot # 202506E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 403207508 |
| Expiration 06/24/2027 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 12/02/2027 |
| Number of Simulators 5 | | Compliance Yes | |

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.051 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.051 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.051 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 02/04/2026 10:45

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007499

| | | |
|----------------------------------|------------------------|-------------------------------|
| Agency FFWCC | Date 03/23/2026 | Type Agency Inspection |
| Inspector BROOKS,JOSEPH,P | Time 13:46:33 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|---------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 403207508 |
| Expiration 06/24/2027 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 12/02/2027 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 03/23/2026 14:50

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007499

| | | |
|----------------------------------|------------------------|-------------------------------|
| Agency FFWCC | Date 04/23/2026 | Type Agency Inspection |
| Inspector BROOKS,JOSEPH,P | Time 18:14:51 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|---------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 403207508 |
| Expiration 06/24/2027 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 12/02/2027 |

Number of Simulators 4

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 04/23/2026 19:10

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007500

| | | | |
|------------------------------------|------------------------------|-------------------------------|---------------------------------|
| Agency FFWCC | Date 01/16/2026 | Type Agency Inspection | |
| Inspector BURNSED, DAVID, J | Time 11:06:29 | Software 8100.27 | |
| 0.05 Lot # 202506E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 403268835 |
| Expiration 06/24/2027 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 02/13/2028 |
| Number of Simulators 4 | | Compliance Yes | |

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.196 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.196 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

IN COMP

Data Download Date/Time 01/16/2026 11:05

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007500

| | | | |
|------------------------------------|------------------------------|-------------------------------|---------------------------------|
| Agency FFWCC | Date 03/13/2026 | Type Agency Inspection | |
| Inspector BURNSED, DAVID, J | Time 11:35:14 | Software 8100.27 | |
| 0.05 Lot # 202506E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 403268835 |
| Expiration 06/24/2027 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 02/13/2028 |
| Number of Simulators 4 | | Compliance Yes | |

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.196 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.195 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

IN COMP

Data Download Date/Time 03/13/2026 11:43

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007500

| | | | |
|------------------------------------|------------------------------|-------------------------------|---------------------------------|
| Agency FFWCC | Date 04/21/2026 | Type Agency Inspection | |
| Inspector BURNSED, DAVID, J | Time 10:47:42 | Software 8100.27 | |
| 0.05 Lot # 202506E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 403268835 |
| Expiration 06/24/2027 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 02/13/2028 |
| Number of Simulators 4 | | Compliance Yes | |

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.195 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.196 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

IN COMP

Data Download Date/Time 04/21/2026 10:59

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007501

| | | | |
|------------------------------------|------------------------------|-------------------------------|---------------------------------|
| Agency FFWCC | Date 01/16/2026 | Type Agency Inspection | |
| Inspector BURNSED, DAVID, J | Time 11:02:43 | Software 8100.27 | |
| 0.05 Lot # 202506E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 403268835 |
| Expiration 06/24/2027 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 02/13/2028 |
| Number of Simulators 4 | | Compliance Yes | |

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

IN COMP

Data Download Date/Time 01/16/2026 11:01

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007501

| | | | |
|------------------------------------|------------------------------|-------------------------------|---------------------------------|
| Agency FFWCC | Date 03/13/2026 | Type Agency Inspection | |
| Inspector GILMORE,RICHARD,L | Time 11:40:11 | Software 8100.27 | |
| 0.05 Lot # 202506E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 403268835 |
| Expiration 06/24/2027 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 02/13/2028 |
| Number of Simulators 4 | | Compliance Yes | |

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

A F / M A:RESIDUAL MOUTH ALCOHOL.

Data Download Date/Time 03/13/2026 11:52

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007501

| | | | |
|------------------------------------|------------------------------|-------------------------------|---------------------------------|
| Agency FFWCC | Date 04/21/2026 | Type Agency Inspection | |
| Inspector GILMORE,RICHARD,L | Time 10:57:19 | Software 8100.27 | |
| 0.05 Lot # 202506E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 403268835 |
| Expiration 06/24/2027 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 02/13/2028 |
| Number of Simulators 4 | | Compliance Yes | |

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.200 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

IN COMP

Data Download Date/Time 04/21/2026 11:06

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007502

| | | | |
|-----------------------------------|------------------------------|-------------------------------|---------------------------------|
| Agency FFWCC | Date 01/21/2026 | Type Agency Inspection | |
| Inspector SAAVEDRA, OSCAR, | Time 17:31:26 | Software 8100.27 | |
| 0.05 Lot # 202506E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 403374313 |
| Expiration 06/24/2027 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 06/25/2028 |
| Number of Simulators 5 | | Compliance Yes | |

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.196 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

A F / M A: No Sample Provided.

Data Download Date/Time 01/21/2026 17:29

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007502

| | | |
|-----------------------------------|------------------------|-------------------------------|
| Agency FFWCC | Date 02/12/2026 | Type Agency Inspection |
| Inspector SAAVEDRA, OSCAR, | Time 17:50:38 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|---------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 403374313 |
| Expiration 06/24/2027 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 06/25/2028 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.193 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 02/12/2026 19:06

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007502

| | | |
|-------------------------------------|------------------------|-------------------------------|
| Agency FFWCC | Date 03/19/2026 | Type Agency Inspection |
| Inspector GORNEY, WILLIAM, M | Time 12:00:49 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|---------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 403374313 |
| Expiration 06/24/2027 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 06/25/2028 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.196 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 03/19/2026 12:15

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007502

| | | |
|-------------------------------------|------------------------|-------------------------------|
| Agency FFWCC | Date 04/21/2026 | Type Agency Inspection |
| Inspector GORNEY, WILLIAM, M | Time 12:36:35 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|---------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 403374313 |
| Expiration 06/24/2027 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 06/25/2028 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.196 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.196 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 04/21/2026 12:49

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007502

| | | |
|------------------------------------|------------------------|-----------------------------------|
| Agency FFWCC | Date 04/28/2026 | Type Department Inspection |
| Inspector GUTSCHOW,TAYLOR,D | Time 12:24:27 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| 0.05 Lot # 202406K | 0.08 Lot # 202603B | 0.20 Lot # 202406N | 0.08 Gas Lot # AG526603 |
| Expiration 06/19/2026 | Expiration 03/03/2028 | Expiration 06/20/2026 | Expiration 09/23/2027 |

| | | |
|-------------------------------|--|-----------------------|
| Number of Simulators 5 | Average Standard Deviation 0.0003 | Compliance Yes |
|-------------------------------|--|-----------------------|

| Check or Test | Result | Test | Result |
|---|--------|---|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.047 |
| Minimum Sample Volume Check: OK | Yes | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | Yes | 0.20 g/210L Test # 1 | 0.196 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.000 |
| Mouth Alcohol Test: Slope Not Met | Yes | 0.08 g/210L Dry Gas Std Test # (Repeat) 1 | 0.079 |
| Interferent Detect Test: Interferent Detect | Yes | Interferent Detect Test # 1 | INT |
| Diagnostic Check (Post-Inspection): OK | Yes | Alcohol Free Test # 2 | 0.000 |
| Alcohol Free Test: OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| 0.05 g/210L Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.08 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.196 |
| 0.20 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.000 |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # (Repeat) 2 | 0.079 |
| | | Interferent Detect Test # 2 | INT |
| | | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.077 |
| | | 0.20 g/210L Test # 3 | 0.196 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.000 |
| | | 0.08 g/210L Dry Gas Std Test # (Repeat) 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |
| | | Alcohol Free Test # 4 | 0.000 |
| | | 0.05 g/210L Test # 4 | 0.047 |
| | | 0.08 g/210L Test # 4 | 0.077 |
| | | 0.20 g/210L Test # 4 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 4 | 0.000 |
| | | 0.08 g/210L Dry Gas Std Test # (Repeat) 4 | 0.079 |
| | | Interferent Detect Test # 4 | INT |
| | | Alcohol Free Test # 5 | 0.000 |
| | | 0.05 g/210L Test # 5 | 0.048 |

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007502

| | |
|--|-------|
| 0.08 g/210L Test # 5 | 0.078 |
| 0.20 g/210L Test # 5 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.000 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 5 | 0.079 |
| Interferent Detect Test # 5 | INT |
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.048 |
| 0.08 g/210L Test # 6 | 0.077 |
| 0.20 g/210L Test # 6 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.000 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 6 | 0.079 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.048 |
| 0.08 g/210L Test # 7 | 0.077 |
| 0.20 g/210L Test # 7 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.000 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 7 | 0.079 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.048 |
| 0.08 g/210L Test # 8 | 0.077 |
| 0.20 g/210L Test # 8 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.000 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 8 | 0.079 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.048 |
| 0.08 g/210L Test # 9 | 0.078 |
| 0.20 g/210L Test # 9 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.000 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 9 | 0.079 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.048 |
| 0.08 g/210L Test # 10 | 0.078 |
| 0.20 g/210L Test # 10 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.000 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 10 | 0.079 |
| Interferent Detect Test # 10 | INT |

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

**Florida Department of Law Enforcement
Alcohol Testing Program**

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007502

| | |
|--|--------|
| 0.05 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Test (Std Dev) | 0.0005 |
| 0.20 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0000 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) (Repeat) | 0.0000 |

Remarks:

08: Control Outside Tolerance. EMPTY DGS.

Data Download Date/Time 04/28/2026 12:24

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007503

| | | |
|-----------------------------------|------------------------|-------------------------------|
| Agency POLK COUNTY SO | Date 02/24/2026 | Type Agency Inspection |
| Inspector WALLISER,SHAUN,J | Time 09:00:23 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202510E | 0.20 Lot # 202410F | 0.08 Gas Lot # 15025080A1 |
| Expiration 06/24/2027 | Expiration 10/28/2027 | Expiration 10/30/2026 | Expiration 07/05/2027 |

Number of Simulators 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.074 |
| Barometric Pressure Sensor Check: OK | | 0.08 g/210L Test # (Repeat) 1 | 0.081 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.20 g/210L Test # 1 | 0.200 |
| Mouth Alcohol Test: Slope Not Met | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test: Interferent Detect | Yes | Interferent Detect Test # 1 | INT |
| Diagnostic Check (Post-Inspection): OK | Yes | Alcohol Free Test # 2 | 0.000 |
| Alcohol Free Test: OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| 0.05 g/210L Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Test # (Repeat) 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| | | Interferent Detect Test # 2 | INT |
| | | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.08 g/210L Test # (Repeat) 3 | 0.081 |
| | | 0.20 g/210L Test # 3 | 0.201 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

08: Control Outside Tolerance.

Data Download Date/Time 02/24/2026 08:59

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007503

| | | |
|-----------------------------------|------------------------|-------------------------------|
| Agency POLK COUNTY SO | Date 03/24/2026 | Type Agency Inspection |
| Inspector WALLISER,SHAUN,J | Time 10:44:15 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202510E | 0.20 Lot # 202410F | 0.08 Gas Lot # 15025080A1 |
| Expiration 06/24/2027 | Expiration 10/28/2027 | Expiration 10/30/2026 | Expiration 07/05/2027 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

A F / M A: REPEAT TEST/MOUTH ALCOHOL ON A/F

Data Download Date/Time 03/24/2026 11:21

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007503

| | | |
|-----------------------------------|------------------------|-------------------------------|
| Agency POLK COUNTY SO | Date 04/28/2026 | Type Agency Inspection |
| Inspector LUCAS,NATHANIEL, | Time 09:34:00 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202510E | 0.20 Lot # 202410F | 0.08 Gas Lot # 15025080A1 |
| Expiration 06/24/2027 | Expiration 10/28/2027 | Expiration 10/30/2026 | Expiration 07/05/2027 |

Number of Simulators 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.077 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.077 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.077 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

N/A

Data Download Date/Time 04/28/2026 09:31

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007505

| | | |
|-----------------------------------|------------------------|-------------------------------|
| Agency HIGHLANDS COUNTY SO | Date 02/04/2026 | Type Agency Inspection |
| Inspector SPEARIN,KATIE,T | Time 08:42:28 | Software 8100.27 |

| | | | |
|-------------------|-------------------|-------------------|-----------------------|
| 0.05 Lot # | 0.08 Lot # | 0.20 Lot # | 0.08 Gas Lot # |
| Expiration | Expiration | Expiration | Expiration |

Number of Simulators **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|------|--------|
| Diagnostic Check (Pre-Inspection): OK | No | | |
| Date and/or Time Adjusted | No | | |
| Minimum Sample Volume Check: OK | | | |
| Barometric Pressure Sensor Check: OK | | | |
| Alcohol Free Subject Test: 0.000 | No | | |
| Mouth Alcohol Test: Slope Not Met | No | | |
| Interferent Detect Test: Interferent Detect | No | | |
| Diagnostic Check (Post-Inspection): OK | No | | |
| Alcohol Free Test: OK | No | | |
| 0.05 g/210L Test: OK | No | | |
| 0.08 g/210L Test: OK | No | | |
| 0.20 g/210L Test: OK | No | | |
| 0.08 g/210L Dry Gas Std Test: OK | No | | |

Remarks:

BYPASSED TO OPERATE INSTRUMENT.COMPLIANCE NOT DETERMINED.

Data Download Date/Time 02/06/2026 09:37

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007505

Agency HIGHLANDS COUNTY SO **Date** 02/06/2026 **Type** Department Inspection
Inspector SPEARIN,KATIE,T **Time** 09:31:56 **Software** 8100.27

0.05 Lot # 202406K **0.08 Lot #** 202406L **0.20 Lot #** 202406N **0.08 Gas Lot #** AG510701
Expiration 06/19/2026 **Expiration** 06/19/2026 **Expiration** 06/20/2026 **Expiration** 04/17/2027

Number of Simulators 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | Yes | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | Yes | 0.20 g/210L Test # 1 | 0.200 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.201 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| | | Interferent Detect Test # 3 | INT |
| | | Alcohol Free Test # 4 | 0.000 |
| | | 0.05 g/210L Test # 4 | 0.049 |
| | | 0.08 g/210L Test # 4 | 0.080 |
| | | 0.20 g/210L Test # 4 | 0.201 |
| | | 0.08 g/210L Dry Gas Std Test # 4 | 0.082 |
| | | Interferent Detect Test # 4 | INT |
| | | Alcohol Free Test # 5 | 0.000 |
| | | 0.05 g/210L Test # 5 | 0.050 |
| | | 0.08 g/210L Test # 5 | 0.080 |
| | | 0.20 g/210L Test # 5 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 5 | 0.082 |
| | | Interferent Detect Test # 5 | INT |

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007505

| | |
|--|--------|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.050 |
| 0.08 g/210L Test # 6 | 0.080 |
| 0.20 g/210L Test # 6 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.082 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.050 |
| 0.08 g/210L Test # 7 | 0.080 |
| 0.20 g/210L Test # 7 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.082 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.050 |
| 0.08 g/210L Test # 8 | 0.080 |
| 0.20 g/210L Test # 8 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.081 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.050 |
| 0.08 g/210L Test # 9 | 0.080 |
| 0.20 g/210L Test # 9 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.082 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.050 |
| 0.08 g/210L Test # 10 | 0.080 |
| 0.20 g/210L Test # 10 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.082 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0003 |
| 0.08 g/210L Test (Std Dev) | 0.0003 |
| 0.20 g/210L Test (Std Dev) | 0.0005 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0003 |

Remarks:

Data Download Date/Time 02/06/2026 09:37

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007505

| | | |
|-----------------------------------|------------------------|-------------------------------|
| Agency HIGHLANDS COUNTY SO | Date 03/23/2026 | Type Agency Inspection |
| Inspector GLASCO,GREGORY,S | Time 20:58:44 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 23725080A2 |
| Expiration 06/24/2027 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 10/05/2027 |

Number of Simulators 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.047 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.194 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.20 g/210L Test # (Repeat) 1 | 0.196 |
| Mouth Alcohol Test: Slope Not Met | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test: Interferent Detect | Yes | Interferent Detect Test # 1 | INT |
| Diagnostic Check (Post-Inspection): OK | Yes | Alcohol Free Test # 2 | 0.000 |
| Alcohol Free Test: OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| 0.05 g/210L Test: OK | Yes | 0.08 g/210L Test # 2 | 0.077 |
| 0.08 g/210L Test: OK | Yes | 0.20 g/210L Test # (Repeat) 2 | 0.197 |
| 0.20 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Interferent Detect Test # 2 | INT |
| | | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # (Repeat) 3 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

20: RFI Detect.

Data Download Date/Time 03/24/2026 02:08

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007505

| | | |
|-----------------------------------|------------------------|-------------------------------|
| Agency HIGHLANDS COUNTY SO | Date 04/26/2026 | Type Agency Inspection |
| Inspector ABELN, SHON, W | Time 21:43:34 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 23725080A2 |
| Expiration 06/24/2027 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 10/05/2027 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.195 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.077 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.077 |
| | | 0.20 g/210L Test # 3 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 04/26/2026 22:24

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007506

| | | |
|-----------------------------------|------------------------|-------------------------------|
| Agency POLK COUNTY SO | Date 04/28/2026 | Type Agency Inspection |
| Inspector LUCAS,NATHANIEL, | Time 09:38:57 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202510E | 0.20 Lot # 202410F | 0.08 Gas Lot # 15025080A1 |
| Expiration 06/24/2027 | Expiration 10/28/2027 | Expiration 10/30/2026 | Expiration 07/05/2027 |

Number of Simulators 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.047 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.196 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.196 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

A F / M A: No Sample Provided.TIMED OUT

Data Download Date/Time 04/28/2026 09:38

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

| | | |
|-----------------------------------|------------------------|-------------------------------|
| Agency HIGHLANDS COUNTY SO | Date 01/03/2026 | Type Agency Inspection |
| Inspector GLASCO,GREGORY,S | Time 21:57:56 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 23725080A2 |
| Expiration 06/24/2027 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 10/05/2027 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.047 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.076 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.192 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.047 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.077 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.194 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.077 |
| | | 0.20 g/210L Test # 3 | 0.195 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 01/03/2026 22:05

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

| | | |
|-----------------------------------|------------------------|-------------------------------|
| Agency HIGHLANDS COUNTY SO | Date 02/23/2026 | Type Agency Inspection |
| Inspector GLASCO,GREGORY,S | Time 21:23:17 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 23725080A2 |
| Expiration 06/24/2027 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 10/05/2027 |

Number of Simulators 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.195 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 02/23/2026 21:41

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

| | | |
|-----------------------------------|------------------------|-------------------------------|
| Agency HIGHLANDS COUNTY SO | Date 03/23/2026 | Type Agency Inspection |
| Inspector GLASCO,GREGORY,S | Time 21:55:44 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|---------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 2372080A2 |
| Expiration 06/24/2027 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 10/05/2027 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.047 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.195 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.047 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.077 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.047 |
| | | 0.08 g/210L Test # 3 | 0.077 |
| | | 0.20 g/210L Test # 3 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 03/24/2026 01:58

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

| | | |
|-----------------------------------|------------------------|-------------------------------|
| Agency HIGHLANDS COUNTY SO | Date 04/26/2026 | Type Agency Inspection |
| Inspector ABELN, SHON, W | Time 21:00:13 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 23725080A2 |
| Expiration 06/24/2027 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 10/05/2027 |

Number of Simulators 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 04/26/2026 21:27

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007509

| | | |
|---|------------------------|-------------------------------|
| Agency DESOTO COUNTY SO | Date 01/31/2026 | Type Agency Inspection |
| Inspector BELLFLOWER, MICHAEL, D | Time 12:03:23 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 04125080A5 |
| Expiration 06/24/2027 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 03/05/2027 |

Number of Simulators 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|---|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.046 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.193 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.000 |
| Mouth Alcohol Test: Slope Not Met | Yes | 0.08 g/210L Dry Gas Std Test # (Repeat) 1 | 0.079 |
| Interferent Detect Test: Interferent Detect | Yes | Interferent Detect Test # 1 | INT |
| Diagnostic Check (Post-Inspection): OK | Yes | Alcohol Free Test # 2 | 0.000 |
| Alcohol Free Test: OK | Yes | 0.05 g/210L Test # 2 | 0.047 |
| 0.05 g/210L Test: OK | Yes | 0.08 g/210L Test # 2 | 0.077 |
| 0.08 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.193 |
| 0.20 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # (Repeat) 2 | 0.079 |
| | | Interferent Detect Test # 2 | INT |
| | | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.046 |
| | | 0.08 g/210L Test # 3 | 0.077 |
| | | 0.20 g/210L Test # 3 | 0.193 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | 0.08 g/210L Dry Gas Std Test # (Repeat) 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

08: Control Outside Tolerance.

Data Download Date/Time 01/31/2026 12:02

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007509

| | | |
|-------------------------------------|------------------------|-------------------------------|
| Agency DESOTO COUNTY SO | Date 02/28/2026 | Type Agency Inspection |
| Inspector WEMHEUER,MICHAEL,W | Time 11:05:18 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202510E | 0.20 Lot # 202410F | 0.08 Gas Lot # 04125080A5 |
| Expiration 06/24/2027 | Expiration 10/28/2027 | Expiration 10/30/2026 | Expiration 03/05/2027 |

Number of Simulators 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.047 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.074 |
| Barometric Pressure Sensor Check: OK | | 0.08 g/210L Test # (Repeat) 1 | 0.076 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.20 g/210L Test # 1 | 0.191 |
| Mouth Alcohol Test: Slope Not Met | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test: Interferent Detect | Yes | Interferent Detect Test # 1 | INT |
| Diagnostic Check (Post-Inspection): OK | Yes | Alcohol Free Test # 2 | 0.000 |
| Alcohol Free Test: OK | Yes | 0.05 g/210L Test # 2 | 0.047 |
| 0.05 g/210L Test: OK | Yes | 0.08 g/210L Test # 2 | 0.075 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Test # (Repeat) 2 | 0.077 |
| 0.20 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.191 |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| | | Interferent Detect Test # 2 | INT |
| | | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.047 |
| | | 0.08 g/210L Test # 3 | 0.076 |
| | | 0.08 g/210L Test # (Repeat) 3 | 0.077 |
| | | 0.20 g/210L Test # 3 | 0.191 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

08: Control Outside Tolerance.

Data Download Date/Time 02/28/2026 11:05

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007509

| | | |
|-------------------------------------|------------------------|-------------------------------|
| Agency DESOTO COUNTY SO | Date 04/03/2026 | Type Agency Inspection |
| Inspector WEMHEUER,MICHAEL,W | Time 18:58:35 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202510E | 0.20 Lot # 202410F | 0.08 Gas Lot # 04125080A5 |
| Expiration 06/24/2027 | Expiration 10/28/2027 | Expiration 10/30/2026 | Expiration 03/05/2027 |

Number of Simulators 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.047 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.074 |
| Barometric Pressure Sensor Check: OK | | 0.08 g/210L Test # (Repeat) 1 | 0.076 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.20 g/210L Test # 1 | 0.191 |
| Mouth Alcohol Test: Slope Not Met | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test: Interferent Detect | Yes | Interferent Detect Test # 1 | INT |
| Diagnostic Check (Post-Inspection): OK | Yes | Alcohol Free Test # 2 | 0.000 |
| Alcohol Free Test: OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| 0.05 g/210L Test: OK | Yes | 0.08 g/210L Test # 2 | 0.075 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Test # (Repeat) 2 | 0.077 |
| 0.20 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.193 |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| | | Interferent Detect Test # 2 | INT |
| | | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.047 |
| | | 0.08 g/210L Test # 3 | 0.076 |
| | | 0.08 g/210L Test # (Repeat) 3 | 0.077 |
| | | 0.20 g/210L Test # 3 | 0.193 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

08: Control Outside Tolerance.

Data Download Date/Time 04/03/2026 18:56

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

| | | | |
|----------------------------------|------------------------|-------------------------------|--|
| Agency ST JOHNS COUNTY SO | Date 01/29/2026 | Type Agency Inspection | |
| Inspector MILLS,S,T | Time 22:34:14 | Software 8100.27 | |

| | | | |
|------------------------------|------------------------------|------------------------------|---------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 403374313 |
| Expiration 06/24/2027 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 06/25/2028 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.051 |
| | | 0.08 g/210L Test # 3 | 0.081 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 02/02/2026 13:14

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

| | | | |
|----------------------------------|------------------------|-------------------------------|--|
| Agency ST JOHNS COUNTY SO | Date 02/27/2026 | Type Agency Inspection | |
| Inspector MILLS,SCOTT,T | Time 17:38:27 | Software 8100.27 | |

| | | | |
|------------------------------|------------------------------|------------------------------|---------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202510E | 0.20 Lot # 202510F | 0.08 Gas Lot # 403374313 |
| Expiration 06/24/2027 | Expiration 10/28/2027 | Expiration 10/29/2027 | Expiration 06/25/2028 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 02/27/2026 17:39

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

| | | | |
|----------------------------------|------------------------|-------------------------------|--|
| Agency ST JOHNS COUNTY SO | Date 03/25/2026 | Type Agency Inspection | |
| Inspector MILLS,SCOTT,T | Time 23:57:31 | Software 8100.27 | |

| | | | |
|------------------------------|------------------------------|------------------------------|---------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202510E | 0.20 Lot # 202510F | 0.08 Gas Lot # 403374313 |
| Expiration 06/24/2027 | Expiration 10/28/2027 | Expiration 10/29/2027 | Expiration 06/25/2028 |

Number of Simulators 5

Compliance No

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | No | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

A F / M A: No Sample Provided. Non-compliance:DIAG FAILURE PRINTER UNPLUGGED ACCIDENTAL .

Data Download Date/Time 03/26/2026 00:48

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

| | | | |
|----------------------------------|------------------------|-------------------------------|--|
| Agency ST JOHNS COUNTY SO | Date 03/26/2026 | Type Agency Inspection | |
| Inspector MILLS,SCOTT,T | Time 00:40:39 | Software 8100.27 | |

| | | | |
|------------------------------|------------------------------|------------------------------|---------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202510E | 0.20 Lot # 202510F | 0.08 Gas Lot # 403374313 |
| Expiration 06/24/2027 | Expiration 10/28/2027 | Expiration 10/29/2027 | Expiration 06/25/2028 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 03/26/2026 00:48

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

| | | | |
|----------------------------------|------------------------|-------------------------------|--|
| Agency ST JOHNS COUNTY SO | Date 04/29/2026 | Type Agency Inspection | |
| Inspector MILLS,SCOTT,T | Time 22:42:37 | Software 8100.27 | |

| | | | |
|------------------------------|------------------------------|------------------------------|---------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202510E | 0.20 Lot # 202510F | 0.08 Gas Lot # 403525131 |
| Expiration 06/24/2027 | Expiration 10/28/2027 | Expiration 10/29/2027 | Expiration 01/02/2029 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 04/29/2026 22:42

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007516

| | | |
|-----------------------------------|------------------------|-------------------------------|
| Agency MIAMI PD | Date 01/24/2026 | Type Agency Inspection |
| Inspector FERNANDEZ,ANGEL, | Time 14:23:15 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|---------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 402732434 |
| Expiration 06/24/2027 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 04/28/2026 |

| | |
|-------------------------------|-----------------------|
| Number of Simulators 5 | Compliance Yes |
|-------------------------------|-----------------------|

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.046 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.076 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.194 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.047 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.076 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.047 |
| | | 0.08 g/210L Test # 3 | 0.076 |
| | | 0.20 g/210L Test # 3 | 0.196 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 01/24/2026 14:20

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007516

| | | | |
|-----------------------------------|------------------------------|-------------------------------|---------------------------------|
| Agency MIAMI PD | Date 02/16/2026 | Type Agency Inspection | |
| Inspector FERNANDEZ,ANGEL, | Time 12:01:03 | Software 8100.27 | |
| 0.05 Lot # 202506E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 402732434 |
| Expiration 06/24/2027 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 04/28/2026 |
| Number of Simulators 5 | | Compliance Yes | |

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 02/16/2026 12:01

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007516

| | | |
|-----------------------------------|------------------------|-------------------------------|
| Agency MIAMI PD | Date 03/23/2026 | Type Agency Inspection |
| Inspector FERNANDEZ,ANGEL, | Time 16:25:57 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|---------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 403318190 |
| Expiration 06/24/2027 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 04/11/2028 |

| | |
|-------------------------------|-----------------------|
| Number of Simulators 5 | Compliance Yes |
|-------------------------------|-----------------------|

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.047 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.075 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.196 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.076 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.076 |
| | | 0.20 g/210L Test # 3 | 0.196 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 03/23/2026 16:23

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007516

| | | | |
|-----------------------------------|------------------------------|-------------------------------|---------------------------------|
| Agency MIAMI PD | Date 04/24/2026 | Type Agency Inspection | |
| Inspector FERNANDEZ,ANGEL, | Time 10:52:15 | Software 8100.27 | |
| 0.05 Lot # 202506E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 403318190 |
| Expiration 06/24/2027 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 04/11/2028 |
| Number of Simulators 5 | | Compliance Yes | |

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.051 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.202 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.053 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.205 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.053 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.207 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 04/24/2026 10:50

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007517

| | | |
|------------------------------------|------------------------|-------------------------------|
| Agency MIAMI PD | Date 02/27/2026 | Type Agency Inspection |
| Inspector GUTSCHOW,TAYLOR,D | Time 14:40:52 | Software 8100.27 |

| | | | |
|-------------------|-------------------|-------------------|-----------------------|
| 0.05 Lot # | 0.08 Lot # | 0.20 Lot # | 0.08 Gas Lot # |
| Expiration | Expiration | Expiration | Expiration |

Number of Simulators **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|------|--------|
| Diagnostic Check (Pre-Inspection): OK | No | | |
| Date and/or Time Adjusted | No | | |
| Minimum Sample Volume Check: OK | | | |
| Barometric Pressure Sensor Check: OK | | | |
| Alcohol Free Subject Test: 0.000 | No | | |
| Mouth Alcohol Test: Slope Not Met | No | | |
| Interferent Detect Test: Interferent Detect | No | | |
| Diagnostic Check (Post-Inspection): OK | No | | |
| Alcohol Free Test: OK | No | | |
| 0.05 g/210L Test: OK | No | | |
| 0.08 g/210L Test: OK | No | | |
| 0.20 g/210L Test: OK | No | | |
| 0.08 g/210L Dry Gas Std Test: OK | No | | |

Remarks:

AI NOT CONDUCTED. COMPLIANCE NOT DETERMINED.

Data Download Date/Time 03/09/2026 14:04

Florida Department of Law Enforcement
Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007517

| | |
|--|--------|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.049 |
| 0.08 g/210L Test # 6 | 0.080 |
| 0.20 g/210L Test # 6 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.078 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.049 |
| 0.08 g/210L Test # 7 | 0.080 |
| 0.20 g/210L Test # 7 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.078 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.049 |
| 0.08 g/210L Test # 8 | 0.080 |
| 0.20 g/210L Test # 8 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.078 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.050 |
| 0.08 g/210L Test # 9 | 0.080 |
| 0.20 g/210L Test # 9 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.078 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.050 |
| 0.08 g/210L Test # 10 | 0.080 |
| 0.20 g/210L Test # 10 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.078 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Test (Std Dev) | 0.0005 |
| 0.20 g/210L Test (Std Dev) | 0.0006 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0005 |

Remarks:

Data Download Date/Time 03/12/2026 11:21

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007517

| | | |
|-----------------------------------|------------------------|-------------------------------|
| Agency MIAMI PD | Date 04/24/2026 | Type Agency Inspection |
| Inspector FERNANDEZ,ANGEL, | Time 11:02:28 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|---------------------------------|
| 0.05 Lot # 202506E | 0.08 Lot # 202410E | 0.20 Lot # 202410F | 0.08 Gas Lot # 403318190 |
| Expiration 06/24/2027 | Expiration 10/22/2026 | Expiration 10/30/2026 | Expiration 04/11/2028 |

Number of Simulators 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.054 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.081 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.194 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.054 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.082 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.055 |
| | | 0.08 g/210L Test # 3 | 0.082 |
| | | 0.20 g/210L Test # 3 | 0.206 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 04/24/2026 11:02