

Alcohol Testing Program - Instrument Processing Sheet

Agency: FHP

Instrument Serial Number: 80-006761

Date In: 5/4/2026

DI Completion Date: 5/7/26

☒ Ship

☐ P/U

☐ H/D

☐ CMI

☐ EE

Intake By: <u>SLH</u> Date: <u>5/4/2026</u>		Quality Checks By: <u>KTS</u> Date: <u>5/5/2026</u>		Flow Adjustment By: <u>KTS</u>																
☒ Annual ☐ Dropped Off ☐ Registration ☐ Return from CMI / EE ☐ Training Instrument Visual Inspection ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/Accessories ☐ Power Cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes:		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value: <u>124</u> ☒ Flow Verification (L/s) Flow Column #: <u>ATP102</u> 32 mm <u>0.136</u> (.139-.169) 36 mm <u>0.156</u> (.156-.190) 53 mm <u>0.222</u> (.228-.278) 103 mm <u>0.476</u> (.447-.547) ☒ Barometric Pressure Check Gauge ID #: <u>28427</u> Gauge: <u>1011</u> Instrument: <u>1012</u> ☒ Stability Checks <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot#/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>MP5088</td> <td>202603A 3/3/2028</td> </tr> <tr> <td>0.080</td> <td>MP5089</td> <td>202603B 3/3/2028</td> </tr> <tr> <td>0.200</td> <td>MP5090</td> <td>202603D 3/4/2028</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>23825080A1 10/5/2027</td> </tr> </tbody> </table>		Simulator	Serial #	Lot#/Exp	0.050	MP5088	202603A 3/3/2028	0.080	MP5089	202603B 3/3/2028	0.200	MP5090	202603D 3/4/2028	0.080 DGS	N/A	23825080A1 10/5/2027	Flow Column #: <u>ATP103</u> ☒ 5L/min – 17mm ☒ 15L/min – 53mm ☒ 30L/min – 103mm ☒ R-Value: <u>125</u> ☒ Post Adjustment Verification (L/S) Flow Column #: <u>ATP102</u> 32 mm <u>0.152</u> (.139-.169) 36 mm <u>0.160</u> (.156-.190) 53 mm <u>0.238</u> (.228-.278) 103 mm <u>0.480</u> (.447-.547)	
Simulator	Serial #	Lot#/Exp																		
0.050	MP5088	202603A 3/3/2028																		
0.080	MP5089	202603B 3/3/2028																		
0.200	MP5090	202603D 3/4/2028																		
0.080 DGS	N/A	23825080A1 10/5/2027																		
Optical Bench Adjustment By: _____		Department Inspection By: <u>KTS</u>																		
Barometric Pressure Gauge: _____ ID#: _____		Barometric Pressure ID#: <u>28427</u>																		
Simulator	Serial #	Lot #	Expiration	Gauge: <u>1009</u> Instrument: <u>1011</u>																
0.000		N/A	N/A	Mouth Alcohol Solution Lot #: <u>2025-D</u> Exp: <u>9/25/27</u>																
0.040				Acetone Stock Solution Lot #: <u>2025-B</u> Exp: <u>9/22/27</u>																
0.100				Simulator	Serial Number															
0.200				0.000	MP5086															
0.300				Interferent	MP5087															
0.080 DGS	N/A			0.050	MP5088															
☐ Post Optical Bench Adjustment Stability Checks				0.080	MP5089															
Simulator	Serial #	Lot #	Expiration	0.200	MP5090															
0.050				Attachments																
0.080				☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☐ Optical Bench Adjustment ☐ Post-Stability Checks ☒ Flow Adjustment ☐ Form 40 ☐ Other:																
0.200																				
0.080 DGS	N/A																			
Gauge ID #: _____ Gauge: _____ Instrument: _____																				
Notes/Suggested Service:				☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use																
				Digitally signed by LeAndra Higginbotham LeAndra Higginbotham Date: 2026.05.15 10:15:38 -04'00' Digitally signed by Wen-Chi Pierson Wen-Chi Pierson Date: 2026.05.20 12:23:11 -04'00'																
				Tech Review																
				Admin Review																

FL Auto W/ PATROL
Intro Analyzer - Alcohol Analyzer
Model 9100 SN 81-00076
15/05/2006
Software: 9100.07

Flow Rate Calibration*****

1: Rate (Liters/min) = 5

SCRT(Diff) = 0.000

2: Rate (Liters/min) = 15

SCRT(Diff) = 0.000

3: Rate (Liters/min) = 30

SCRT(Diff) = 0.000

Dependent Data Scale Factor = 100000

Independent Data Scale Factor = 255

Rounded Slope = 0

Rounded Intercept = 0

Correlation = 0.0000

80-006761

Flow Adjustment #1

Column not plugged in.

Repeated after connecting.

[Signature]

FL Auto W/ PATROL
Intro Analyzer - Alcohol Analyzer
Model 9100 SN 81-00076
15/05/2006
Software: 9100.07

Flow Rate Calibration*****

1: Rate (Liters/min) = 5

SCRT(Diff) = 6.240

2: Rate (Liters/min) = 15

SCRT(Diff) = 11.288

3: Rate (Liters/min) = 30

SCRT(Diff) = 20.855

Dependent Data Scale Factor = 100000

Independent Data Scale Factor = 255

Rounded Slope = 561

Rounded Intercept = -495203

Correlation = 0.99788

Flow Adjustment #2

Stability Checks

0.050 g/210L 0.047 to 0.053 g/210L	0.080 g/210L 0.077 to 0.083 g/210L	0.200 g/210L 0.194 to 0.206 g/210L	DGS 0.080 g/210L 0.077 to 0.083 g/210L ≤0.003 g/210L of Wet																																																																																																																																																
FL-4000M PATROL Intoximeter - Alcotest Analyser Model 8010 SW 00-00000 SN 00-00000 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>%</th> </tr> </thead> <tbody> <tr><td>Ref Blank</td><td>0.000</td><td>0.00</td></tr> <tr><td>Control Test</td><td>0.050</td><td>100.0</td></tr> <tr><td>Ref Blank</td><td>0.000</td><td>0.00</td></tr> <tr><td>Control Test</td><td>0.048</td><td>96.0</td></tr> <tr><td>Ref Blank</td><td>0.000</td><td>0.00</td></tr> <tr><td>Control Test</td><td>0.049</td><td>98.0</td></tr> <tr><td>Ref Blank</td><td>0.000</td><td>0.00</td></tr> <tr><td>Control Test</td><td>0.049</td><td>98.0</td></tr> <tr><td>Average</td><td>0.049</td><td></td></tr> <tr><td>Std Dev</td><td>0.000</td><td></td></tr> <tr><td>Ref Std Dev</td><td>0.000</td><td></td></tr> </tbody> </table> Operator's Signature: <i>Katz</i>	Test	g/210L	%	Ref Blank	0.000	0.00	Control Test	0.050	100.0	Ref Blank	0.000	0.00	Control Test	0.048	96.0	Ref Blank	0.000	0.00	Control Test	0.049	98.0	Ref Blank	0.000	0.00	Control Test	0.049	98.0	Average	0.049		Std Dev	0.000		Ref Std Dev	0.000		FL-4000M PATROL Intoximeter - Alcotest Analyser Model 8010 SW 00-00000 SN 00-00000 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>%</th> </tr> </thead> <tbody> <tr><td>Ref Blank</td><td>0.000</td><td>0.00</td></tr> <tr><td>Control Test</td><td>0.080</td><td>100.0</td></tr> <tr><td>Ref Blank</td><td>0.000</td><td>0.00</td></tr> <tr><td>Control Test</td><td>0.079</td><td>98.8</td></tr> <tr><td>Ref Blank</td><td>0.000</td><td>0.00</td></tr> <tr><td>Control Test</td><td>0.077</td><td>96.3</td></tr> <tr><td>Ref Blank</td><td>0.000</td><td>0.00</td></tr> <tr><td>Control Test</td><td>0.077</td><td>96.3</td></tr> <tr><td>Average</td><td>0.077</td><td></td></tr> <tr><td>Std Dev</td><td>0.000</td><td></td></tr> <tr><td>Ref Std Dev</td><td>0.000</td><td></td></tr> </tbody> </table> Operator's Signature: <i>Katz</i>	Test	g/210L	%	Ref Blank	0.000	0.00	Control Test	0.080	100.0	Ref Blank	0.000	0.00	Control Test	0.079	98.8	Ref Blank	0.000	0.00	Control Test	0.077	96.3	Ref Blank	0.000	0.00	Control Test	0.077	96.3	Average	0.077		Std Dev	0.000		Ref Std Dev	0.000		FL-4000M PATROL Intoximeter - Alcotest Analyser Model 8010 SW 00-00000 SN 00-00000 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>%</th> </tr> </thead> <tbody> <tr><td>Ref Blank</td><td>0.000</td><td>0.00</td></tr> <tr><td>Control Test</td><td>0.200</td><td>100.0</td></tr> <tr><td>Ref Blank</td><td>0.000</td><td>0.00</td></tr> <tr><td>Control Test</td><td>0.199</td><td>99.5</td></tr> <tr><td>Ref Blank</td><td>0.000</td><td>0.00</td></tr> <tr><td>Control Test</td><td>0.198</td><td>99.0</td></tr> <tr><td>Ref Blank</td><td>0.000</td><td>0.00</td></tr> <tr><td>Control Test</td><td>0.198</td><td>99.0</td></tr> <tr><td>Average</td><td>0.198</td><td></td></tr> <tr><td>Std Dev</td><td>0.000</td><td></td></tr> <tr><td>Ref Std Dev</td><td>0.000</td><td></td></tr> </tbody> </table> Operator's Signature: <i>Katz</i>	Test	g/210L	%	Ref Blank	0.000	0.00	Control Test	0.200	100.0	Ref Blank	0.000	0.00	Control Test	0.199	99.5	Ref Blank	0.000	0.00	Control Test	0.198	99.0	Ref Blank	0.000	0.00	Control Test	0.198	99.0	Average	0.198		Std Dev	0.000		Ref Std Dev	0.000		FL-4000M PATROL Intoximeter - Alcotest Analyser Model 8010 SW 00-00000 SN 00-00000 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>%</th> </tr> </thead> <tbody> <tr><td>Ref Blank</td><td>0.000</td><td>0.00</td></tr> <tr><td>Control Test</td><td>0.080</td><td>100.0</td></tr> <tr><td>Ref Blank</td><td>0.000</td><td>0.00</td></tr> <tr><td>Control Test</td><td>0.079</td><td>98.8</td></tr> <tr><td>Ref Blank</td><td>0.000</td><td>0.00</td></tr> <tr><td>Control Test</td><td>0.077</td><td>96.3</td></tr> <tr><td>Ref Blank</td><td>0.000</td><td>0.00</td></tr> <tr><td>Control Test</td><td>0.077</td><td>96.3</td></tr> <tr><td>Average</td><td>0.077</td><td></td></tr> <tr><td>Std Dev</td><td>0.000</td><td></td></tr> <tr><td>Ref Std Dev</td><td>0.000</td><td></td></tr> </tbody> </table> Operator's Signature: <i>Katz</i>	Test	g/210L	%	Ref Blank	0.000	0.00	Control Test	0.080	100.0	Ref Blank	0.000	0.00	Control Test	0.079	98.8	Ref Blank	0.000	0.00	Control Test	0.077	96.3	Ref Blank	0.000	0.00	Control Test	0.077	96.3	Average	0.077		Std Dev	0.000		Ref Std Dev	0.000	
Test	g/210L	%																																																																																																																																																	
Ref Blank	0.000	0.00																																																																																																																																																	
Control Test	0.050	100.0																																																																																																																																																	
Ref Blank	0.000	0.00																																																																																																																																																	
Control Test	0.048	96.0																																																																																																																																																	
Ref Blank	0.000	0.00																																																																																																																																																	
Control Test	0.049	98.0																																																																																																																																																	
Ref Blank	0.000	0.00																																																																																																																																																	
Control Test	0.049	98.0																																																																																																																																																	
Average	0.049																																																																																																																																																		
Std Dev	0.000																																																																																																																																																		
Ref Std Dev	0.000																																																																																																																																																		
Test	g/210L	%																																																																																																																																																	
Ref Blank	0.000	0.00																																																																																																																																																	
Control Test	0.080	100.0																																																																																																																																																	
Ref Blank	0.000	0.00																																																																																																																																																	
Control Test	0.079	98.8																																																																																																																																																	
Ref Blank	0.000	0.00																																																																																																																																																	
Control Test	0.077	96.3																																																																																																																																																	
Ref Blank	0.000	0.00																																																																																																																																																	
Control Test	0.077	96.3																																																																																																																																																	
Average	0.077																																																																																																																																																		
Std Dev	0.000																																																																																																																																																		
Ref Std Dev	0.000																																																																																																																																																		
Test	g/210L	%																																																																																																																																																	
Ref Blank	0.000	0.00																																																																																																																																																	
Control Test	0.200	100.0																																																																																																																																																	
Ref Blank	0.000	0.00																																																																																																																																																	
Control Test	0.199	99.5																																																																																																																																																	
Ref Blank	0.000	0.00																																																																																																																																																	
Control Test	0.198	99.0																																																																																																																																																	
Ref Blank	0.000	0.00																																																																																																																																																	
Control Test	0.198	99.0																																																																																																																																																	
Average	0.198																																																																																																																																																		
Std Dev	0.000																																																																																																																																																		
Ref Std Dev	0.000																																																																																																																																																		
Test	g/210L	%																																																																																																																																																	
Ref Blank	0.000	0.00																																																																																																																																																	
Control Test	0.080	100.0																																																																																																																																																	
Ref Blank	0.000	0.00																																																																																																																																																	
Control Test	0.079	98.8																																																																																																																																																	
Ref Blank	0.000	0.00																																																																																																																																																	
Control Test	0.077	96.3																																																																																																																																																	
Ref Blank	0.000	0.00																																																																																																																																																	
Control Test	0.077	96.3																																																																																																																																																	
Average	0.077																																																																																																																																																		
Std Dev	0.000																																																																																																																																																		
Ref Std Dev	0.000																																																																																																																																																		

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: FL HIGHWAY PATROL

Serial Number: 80-006761

Time of Inspection: 10:23

Date of Inspection: 05/07/2026

Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 202603A Exp: 03/03/2028	0.08g/210L Test (g/210L) Lot#: 202603B Exp: 03/03/2028	0.20g/210L Test (g/210L) Lot#: 202603D Exp: 03/04/2028	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 23825080A1 Exp: 10/05/2027
0.000	0.049	0.078	0.197	0.080
0.000	0.049	0.078	0.198	0.080
0.000	0.049	0.078	0.197	0.080
0.000	0.049	0.078	0.198	0.080
0.000	0.049	0.078	0.197	0.080
0.000	0.049	0.078	0.197	0.080
0.000	0.049	0.078	0.197	0.080
0.000	0.049	0.078	0.197	0.080
0.000	0.049	0.078	0.198	0.080
0.000	0.049	0.078	0.197	0.080

Standard Deviations	0.0000	0.0000	0.0004	0.0000
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0001 Number of Simulators Used: 5

Remarks: Reprinted due to printer issues. *Katie T. Spearin* 5/7/26

The above instrument complies (☒) does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Katie T. Spearin

KATIE T SPEARIN

Signature and Printed Name

05/07/2026

Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-006761, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-006761</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>FL HIGHWAY PATROL</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>05/07/2026</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>10:23</u>	0.200 g/ 210 L	0.008
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

05/07/2026

Date

KATIE T SPEARIN,
Department Inspector

FDLE/ATP Form 69 January 2026

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Page 1 of 1