

Alcohol Testing Program - Instrument Processing Sheet

Agency: MANATEE COUNTY SO Instrument Serial Number: 80-006559

Date In: 8/17/2026 DI Completion Date: N/A ☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake By: <u>WKP</u> Date: <u>8/18/2026</u>		Quality Checks By: _____ Date: _____		Flow Adjustment By: _____																							
☒ Annual ☐ Dropped Off ☐ Registration ☐ Return from CMI / EE ☐ Training Instrument Visual Inspection ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/Accessories ☐ Power Cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes:		☐ Breath Tube Screen ☐ Replace External O-Rings ☐ Instrument Set Up Verified ☐ R-Value: _____ ☐ Flow Verification (L/s) Flow Column #: _____ 32 mm _____ (.139-.169) 36 mm _____ (.156-.190) 53 mm _____ (.228-.278) 103 mm _____ (.447-.547) ☐ Barometric Pressure Check Gauge ID #: _____ Gauge: _____ Instrument: _____ ☐ Stability Checks		Flow Column #: _____ ☐ 5L/min – 17mm ☐ 15L/min – 53mm ☐ 30L/min – 103mm ☐ R-Value: _____ ☐ Post Adjustment Verification (L/S) Flow Column #: _____ 32 mm _____ (.139-.169) 36 mm _____ (.156-.190) 53 mm _____ (.228-.278) 103 mm _____ (.447-.547)																							
						<table border="1" style="width: 100%; border-collapse: collapse;"> <tr> <th style="width: 15%;">Simulator</th> <th style="width: 20%;">Serial #</th> <th style="width: 25%;">Lot#/Exp</th> <th style="width: 40%;">Maintenance By: _____ Date: _____</th> </tr> <tr> <td>0.050</td> <td></td> <td></td> <td rowspan="5"> ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Tank Sensor Tare ☐ Breath Tube Replacement ☐ Other: </td> </tr> <tr><td>0.080</td><td></td><td></td></tr> <tr><td>0.200</td><td></td><td></td></tr> <tr><td>0.080 DGS</td><td>N/A</td><td></td></tr> <tr><td> </td><td></td><td></td></tr> </table>		Simulator	Serial #	Lot#/Exp	Maintenance By: _____ Date: _____	0.050			☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Tank Sensor Tare ☐ Breath Tube Replacement ☐ Other:	0.080			0.200			0.080 DGS	N/A				
						Simulator	Serial #	Lot#/Exp	Maintenance By: _____ Date: _____																		
						0.050			☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Tank Sensor Tare ☐ Breath Tube Replacement ☐ Other:																		
						0.080																					
0.200																											
0.080 DGS	N/A																										
Optical Bench Adjustment By: _____		Department Inspection By: _____																									
Barometric Pressure Gauge: _____ ID#: _____		Barometric Pressure ID#: _____																									
Simulator	Serial #	Lot #	Expiration	Gauge: _____ Instrument: _____																							
0.000		N/A	N/A	Mouth Alcohol Solution Lot #: _____ Exp: _____																							
0.040				Acetone Stock Solution Lot #: _____ Exp: _____																							
0.100				Simulator	Serial Number																						
0.200				0.000																							
0.300				Interferent																							
0.080 DGS	N/A			0.050																							
☐ Post Optical Bench Adjustment Stability Checks				0.080																							
Simulator	Serial #	Lot #	Expiration	0.200																							
0.050				Attachments																							
0.080				☐ Form 41 ☐ Post-Stability Checks ☐ Stability Checks ☐ Flow Adjustment ☐ Calibration Certificate ☐ Form 40 ☐ Optical Bench Adjustment ☐ Other:																							
0.200																											
0.080 DGS	N/A																										
Gauge ID #: _____																											
Gauge: _____ Instrument: _____																											
Notes/Suggested Service: Instrument was sent to us inadvertently by agency inspector. KTS 8/20/26 Admin review: Instrument compliance box checked due to instrument receiving a Department Inspection on 6/17/26. KTS 8/26/26				☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC																							
				☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use																							
				☒ Conduct an Agency Inspection Before Evidentiary Use																							
				S. LeAndra Higginbotham Digitally signed by S. LeAndra Higginbotham Date: 2026.08.26 10:19:58 -04'00'																							
				Wen-Chi Pierson Digitally signed by Wen-Chi Pierson Date: 2026.08.27 13:29:39 -04'00'																							
Tech Review				Admin Review																							

Alcohol Testing Program - Instrument Processing Sheet

Agency: MANATEE COUNTY SO

Instrument Serial Number: 80-006559

Date In: 6/11/2026

DI Completion Date: 6/17/26

☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake By: <u>SLH</u> Date: <u>6/11/2026</u>				Quality Checks By: <u>KTS</u> Date: <u>6/12/2026</u>				Flow Adjustment By: _____																																			
☒ Annual ☐ Dropped Off ☐ Registration ☒ Return from CMI / EE ☐ Training Instrument Visual Inspection ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/Accessories ☐ Power Cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes: Missing plastic joint connecting DGS hose to instrument.				☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value: <u>205</u> ☒ Flow Verification (L/s) Flow Column #: <u>ATP103</u> 32 mm <u>0.148</u> (.139-.169) 36 mm <u>0.167</u> (.156-.190) 53 mm <u>0.238</u> (.228-.278) 103 mm <u>0.492</u> (.447-.547) ☒ Barometric Pressure Check Gauge ID #: <u>28427</u> Gauge: <u>1012</u> Instrument: <u>1008</u> ☒ Stability Checks				Flow Column #: _____ ☐ 5L/min – 17mm ☐ 15L/min – 53mm ☐ 30L/min – 103mm ☐ R-Value: _____ ☐ Post Adjustment Verification (L/S) Flow Column #: _____ 32 mm _____ (.139-.169) 36 mm _____ (.156-.190) 53 mm _____ (.228-.278) 103 mm _____ (.447-.547)																																			
				<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot#/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP5088</td><td>202603A 3/3/2028</td></tr><tr><td>0.080</td><td>MP5089</td><td>202603B 3/3/2028</td></tr><tr><td>0.200</td><td>MP5090</td><td>202603D 3/4/2028</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>23825080A1 10/5/2027</td></tr></tbody></table>				Simulator	Serial #	Lot#/Exp	0.050	MP5088	202603A 3/3/2028	0.080	MP5089	202603B 3/3/2028	0.200	MP5090	202603D 3/4/2028	0.080 DGS	N/A	23825080A1 10/5/2027	Maintenance By: <u>KTS</u> Date: <u>6/12/26</u> ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Tank Sensor Tare ☐ Breath Tube Replacement ☒ Other: Replaced dry gas elbow. Put electrical tape on breath hose.																				
				Simulator	Serial #	Lot#/Exp																																					
				0.050	MP5088	202603A 3/3/2028																																					
				0.080	MP5089	202603B 3/3/2028																																					
0.200	MP5090	202603D 3/4/2028																																									
0.080 DGS	N/A	23825080A1 10/5/2027																																									
Optical Bench Adjustment By: <u>KTS</u>				Department Inspection By: <u>KTS</u>																																							
Barometric Pressure Gauge: <u>1011</u> ID#: <u>34420</u>				Barometric Pressure ID#: <u>28427</u>																																							
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td>MP6294</td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td>MP6295</td><td>25090</td><td>3/11/27</td></tr><tr><td>0.100</td><td>MP6296</td><td>24390</td><td>10/29/26</td></tr><tr><td>0.200</td><td>MP6297</td><td>25020</td><td>1/14/27</td></tr><tr><td>0.300</td><td>MP6298</td><td>24430</td><td>12/10/26</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG510701</td><td>4/17/27</td></tr></tbody></table>				Simulator	Serial #	Lot #	Expiration	0.000	MP6294	N/A	N/A	0.040	MP6295	25090	3/11/27	0.100	MP6296	24390	10/29/26	0.200	MP6297	25020	1/14/27	0.300	MP6298	24430	12/10/26	0.080 DGS	N/A	AG510701	4/17/27	Gauge: <u>1009</u> Instrument: <u>1011</u> Mouth Alcohol Solution Lot #: <u>2025-D</u> Exp: <u>9/25/27</u> Acetone Stock Solution Lot #: <u>2025-B</u> Exp: <u>9/22/27</u>											
Simulator	Serial #	Lot #	Expiration																																								
0.000	MP6294	N/A	N/A																																								
0.040	MP6295	25090	3/11/27																																								
0.100	MP6296	24390	10/29/26																																								
0.200	MP6297	25020	1/14/27																																								
0.300	MP6298	24430	12/10/26																																								
0.080 DGS	N/A	AG510701	4/17/27																																								
☒ Post Optical Bench Adjustment Stability Checks				<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td>MP5088</td><td>202603A</td><td>3/3/28</td></tr><tr><td>0.080</td><td>MP5089</td><td>202603B</td><td>3/3/28</td></tr><tr><td>0.200</td><td>MP5090</td><td>202603D</td><td>3/4/28</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>23825080A1</td><td>10/5/27</td></tr></tbody></table>				Simulator	Serial #	Lot #	Expiration	0.050	MP5088	202603A	3/3/28	0.080	MP5089	202603B	3/3/28	0.200	MP5090	202603D	3/4/28	0.080 DGS	N/A	23825080A1	10/5/27	<table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>MP6289</td></tr><tr><td>Interferent</td><td>MP5087</td></tr><tr><td>0.050</td><td>MP5088</td></tr><tr><td>0.080</td><td>MP5089</td></tr><tr><td>0.200</td><td>MP5090</td></tr></tbody></table>				Simulator	Serial Number	0.000	MP6289	Interferent	MP5087	0.050	MP5088	0.080	MP5089	0.200	MP5090
Simulator	Serial #	Lot #	Expiration																																								
0.050	MP5088	202603A	3/3/28																																								
0.080	MP5089	202603B	3/3/28																																								
0.200	MP5090	202603D	3/4/28																																								
0.080 DGS	N/A	23825080A1	10/5/27																																								
Simulator	Serial Number																																										
0.000	MP6289																																										
Interferent	MP5087																																										
0.050	MP5088																																										
0.080	MP5089																																										
0.200	MP5090																																										
Gauge ID #: <u>28427</u> Gauge: <u>1008</u> Instrument: <u>1011</u>				Attachments ☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☒ Optical Bench Adjustment																																							
Notes/Suggested Service: Receipt paper got jammed during optical bench adjustment. KTS 6/16/26 Put electrical tape on fraying area of breath hose. KTS 6/16/26				☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC																																							
				☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use																																							
				☒ Conduct an Agency Inspection Before Evidentiary Use																																							
				<table><tr><td>Wen-Chi Pierson</td><td>Digitally signed by Wen-Chi Pierson Date: 2026.06.22 10:14:26 -04'00'</td><td>LeAndra Higginbotham</td><td>Digitally signed by LeAndra Higginbotham Date: 2026.06.28 15:57:02 -04'00'</td></tr></table>				Wen-Chi Pierson	Digitally signed by Wen-Chi Pierson Date: 2026.06.22 10:14:26 -04'00'	LeAndra Higginbotham	Digitally signed by LeAndra Higginbotham Date: 2026.06.28 15:57:02 -04'00'																																
Wen-Chi Pierson	Digitally signed by Wen-Chi Pierson Date: 2026.06.22 10:14:26 -04'00'	LeAndra Higginbotham	Digitally signed by LeAndra Higginbotham Date: 2026.06.28 15:57:02 -04'00'																																								
Tech Review				Admin Review																																							

Stability Checks

0.050 g/210L 0.047 to 0.053 g/210L	0.080 g/210L 0.077 to 0.083 g/210L	0.200 g/210L 0.194 to 0.206 g/210L	DGS 0.080 g/210L 0.077 to 0.083 g/210L ≤0.003 g/210L of Wet																																																																																																																																																
✓ PARATEE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-006559 06/12/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>08:59</td></tr> <tr><td>Control Test</td><td>0.049</td><td>09:00</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:01</td></tr> <tr><td>Control Test</td><td>0.049</td><td>09:01</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:02</td></tr> <tr><td>Control Test</td><td>0.049</td><td>09:02</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:03</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0490</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> Operator's Signature _____	Test	g/210L	Time	Air Blank	0.000	08:59	Control Test	0.049	09:00	Air Blank	0.000	09:01	Control Test	0.049	09:01	Air Blank	0.000	09:02	Control Test	0.049	09:02	Air Blank	0.000	09:03	Control Test Stats			Average	0.0490		Std Dev	0.0000		Rel Std Dev(%)	0.0000		✓ PARATEE COUNTY SC Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-006559 06/12/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.001</td><td>09:04</td></tr> <tr><td>Control Test</td><td>0.078</td><td>09:05</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:05</td></tr> <tr><td>Control Test</td><td>0.077</td><td>09:06</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:06</td></tr> <tr><td>Control Test</td><td>0.078</td><td>09:07</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:08</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0777</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7434</td><td></td></tr> </tbody> </table> Operator's Signature _____	Test	g/210L	Time	Air Blank	0.001	09:04	Control Test	0.078	09:05	Air Blank	0.000	09:05	Control Test	0.077	09:06	Air Blank	0.000	09:06	Control Test	0.078	09:07	Air Blank	0.000	09:08	Control Test Stats			Average	0.0777		Std Dev	0.0006		Rel Std Dev(%)	0.7434		✓ PARATEE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-006559 06/12/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>09:14</td></tr> <tr><td>Control Test</td><td>0.196</td><td>09:14</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:15</td></tr> <tr><td>Control Test</td><td>0.196</td><td>09:16</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:16</td></tr> <tr><td>Control Test</td><td>0.196</td><td>09:17</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:17</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1960</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> Operator's Signature _____	Test	g/210L	Time	Air Blank	0.000	09:14	Control Test	0.196	09:14	Air Blank	0.000	09:15	Control Test	0.196	09:16	Air Blank	0.000	09:16	Control Test	0.196	09:17	Air Blank	0.000	09:17	Control Test Stats			Average	0.1960		Std Dev	0.0000		Rel Std Dev(%)	0.0000		✓ PARATEE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-006559 06/12/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>08:53</td></tr> <tr><td>Control Test</td><td>0.081</td><td>08:53</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>08:53</td></tr> <tr><td>Control Test</td><td>0.081</td><td>08:54</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>08:54</td></tr> <tr><td>Control Test</td><td>0.081</td><td>08:55</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>08:55</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0810</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> Operator's Signature _____	Test	g/210L	Time	Air Blank	0.000	08:53	Control Test	0.081	08:53	Air Blank	0.000	08:53	Control Test	0.081	08:54	Air Blank	0.000	08:54	Control Test	0.081	08:55	Air Blank	0.000	08:55	Control Test Stats			Average	0.0810		Std Dev	0.0000		Rel Std Dev(%)	0.0000	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	08:59																																																																																																																																																	
Control Test	0.049	09:00																																																																																																																																																	
Air Blank	0.000	09:01																																																																																																																																																	
Control Test	0.049	09:01																																																																																																																																																	
Air Blank	0.000	09:02																																																																																																																																																	
Control Test	0.049	09:02																																																																																																																																																	
Air Blank	0.000	09:03																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0490																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.001	09:04																																																																																																																																																	
Control Test	0.078	09:05																																																																																																																																																	
Air Blank	0.000	09:05																																																																																																																																																	
Control Test	0.077	09:06																																																																																																																																																	
Air Blank	0.000	09:06																																																																																																																																																	
Control Test	0.078	09:07																																																																																																																																																	
Air Blank	0.000	09:08																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0777																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7434																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	09:14																																																																																																																																																	
Control Test	0.196	09:14																																																																																																																																																	
Air Blank	0.000	09:15																																																																																																																																																	
Control Test	0.196	09:16																																																																																																																																																	
Air Blank	0.000	09:16																																																																																																																																																	
Control Test	0.196	09:17																																																																																																																																																	
Air Blank	0.000	09:17																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1960																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	08:53																																																																																																																																																	
Control Test	0.081	08:53																																																																																																																																																	
Air Blank	0.000	08:53																																																																																																																																																	
Control Test	0.081	08:54																																																																																																																																																	
Air Blank	0.000	08:54																																																																																																																																																	
Control Test	0.081	08:55																																																																																																																																																	
Air Blank	0.000	08:55																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0810																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		

Optical Adjustment 80-006559 Katz 6/16/26

MANATEE COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006559
06/16/2026 09:27:27

Auto Calibration
Max Power Res Value = 100
Auto Range Res Value = 72

Sol Value = 0.000 g/210L ***
Fit value = 0.0000 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12509, Sum Io = 12609

Channel 1 Data:
Sample % Abs (% Abs Ref)
Sample #1 = -0.0150 (0.0060)
Sample #2 = 0.0170 (0.0060)
Sample #3 = 0.0260 (0.0210)
Sample #4 = 0.0060 (0.0650)
Avg % Abs = 0.0163 (0.0307)
STD DEV = 0.0100 (0.0307)
REL STD DEV = 61.326 (99.994)

Channel 2 Data:
Sample % Abs (% Abs Ref)
Sample #1 = 0.0850 (-0.0060)
Sample #2 = 0.1190 (-0.0080)
Sample #3 = 0.1250 (-0.0070)
Sample #4 = 0.1040 (0.0140)
Avg % Abs = 0.1160 (-0.0003)
STD DEV = 0.0108 (0.0124)
REL STD DEV = 9.325 (3726.928)

Sol Value = 0.040 g/210L ***
Fit value = 0.1905 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12493, Sum Io = 12600

Channel 1 Data:
Sample % Abs (% Abs Ref)
Sample #1 = 0.6900 (-0.0100)
Sample #2 = 0.7510 (-0.0230)
Sample #3 = 0.7240 (0.0200)
Sample #4 = 0.7130 (0.0410)
Avg % Abs = 0.7293 (0.0127)
STD DEV = 0.0196 (0.0326)
REL STD DEV = 2.681 (257.559)

Channel 2 Data:
Sample % Abs (% Abs Ref)
Sample #1 = 1.4430 (-0.0230)
Sample #2 = 1.4780 (-0.0260)
Sample #3 = 1.4560 (-0.0100)
Sample #4 = 1.4650 (-0.0050)
Avg % Abs = 1.4663 (-0.0137)
STD DEV = 0.0111 (0.0110)
REL STD DEV = 0.754 (80.266)

Sol Value = 0.100 g/210L ***
Fit value = 0.4762 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12481, Sum Io = 12595

Channel 1 Data:
Sample % Abs (% Abs Ref)
Sample #1 = 1.7290 (-0.0120)
Sample #2 = 1.7530 (-0.0070)
Sample #3 = 1.7760 (0.0020)
Sample #4 = 1.7370 (0.0410)
Avg % Abs = 1.7553 (0.0120)
STD DEV = 0.0196 (0.0255)
REL STD DEV = 1.117 (212.622)

Channel 2 Data:
Sample % Abs (% Abs Ref)
Sample #1 = 3.3970 (-0.0020)
Sample #2 = 3.3950 (0.0000)
Sample #3 = 3.4130 (0.0010)
Sample #4 = 3.3940 (0.0200)
Avg % Abs = 3.4007 (0.0070)
STD DEV = 0.0107 (0.0113)
REL STD DEV = 0.314 (160.992)

Sol Value = 0.200 g/210L ***
Fit value = 0.9524 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12472, Sum Io = 12590

Channel 1 Data:
Sample % Abs (% Abs Ref)
Sample #1 = 3.4680 (-0.0050)
Sample #2 = 3.4590 (0.0200)
Sample #3 = 3.4830 (0.0120)
Sample #4 = 3.4740 (0.0230)
Avg % Abs = 3.4720 (0.0183)
STD DEV = 0.0121 (0.0057)
REL STD DEV = 0.349 (31.016)

Channel 2 Data:
Sample % Abs (% Abs Ref)
Sample #1 = 6.5560 (-0.0180)
Sample #2 = 6.5530 (-0.0010)
Sample #3 = 6.5630 (0.0020)
Sample #4 = 6.5460 (-0.0020)
Avg % Abs = 6.5540 (-0.0003)
STD DEV = 0.0085 (0.0021)
REL STD DEV = 0.130 (624.500)

Sol Value = 0.300 g/210L ***
Fit value = 1.4286 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12464, Sum Io = 12587

Channel 1 Data:
Sample % Abs (% Abs Ref)
Sample #1 = 5.1100 (0.0000)
Sample #2 = 5.1310 (-0.0050)
Sample #3 = 5.0890 (0.0290)
Sample #4 = 5.1030 (0.0230)
Avg % Abs = 5.1077 (0.0157)
STD DEV = 0.0214 (0.0181)
REL STD DEV = 0.419 (115.835)

Channel 2 Data:

Sample % Abs (% Abs Ref)
Sample #1 = 9.5200 (-0.0060)
Sample #2 = 9.5340 (-0.0040)
Sample #3 = 9.5050 (0.0200)
Sample #4 = 9.5360 (0.0020)
Avg % Abs = 9.5250 (0.0060)
STD DEV = 0.0173 (0.0125)
REL STD DEV = 0.182 (208.167)

***** AUTO CAL DATA *****

Channel 1 Data:
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.016
Std Dev = 0.01 Rel Std Dev = 61.33
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 0.729
Std Dev = 0.02 Rel Std Dev = 2.68
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 1.755
Std Dev = 0.02 Rel Std Dev = 1.12
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 3.472
Std Dev = 0.01 Rel Std Dev = 0.35
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 5.108
Std Dev = 0.02 Rel Std Dev = 0.42
Zero Order Coef = -43.95
First Order Coef = 2675.23
Second Order Coef = 25.20
Standard Deviation = 22.103565

Channel 2 Data:
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.116
Std Dev = 0.01 Rel Std Dev = 9.32
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 1.466
Std Dev = 0.01 Rel Std Dev = 0.75
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 3.401
Std Dev = 0.01 Rel Std Dev = 0.31
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 6.554
Std Dev = 0.01 Rel Std Dev = 0.13
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 9.525
Std Dev = 0.02 Rel Std Dev = 0.18
Zero Order Coef = -157.04
First Order Coef = 1401.08
Second Order Coef = 12.11
Standard Deviation = 14.573538



Post-Optical Adj.

Stability Checks

0.050 g/210L	0.080 g/210L	0.200 g/210L	DGS 0.080 g/210L
0.047 to 0.053 g/210L	0.077 to 0.083 g/210L	0.194 to 0.206 g/210L	0.077 to 0.083 g/210L
☒	☒	☒	☒
☒	☒	☒	☒

Test	g/210L	Time
Air Blank	0.000	12:15
Control Test	0.050	12:16
Air Blank	0.010	12:17
Control Test	0.049	12:18
Air Blank	0.000	12:19
Control Test	0.050	12:19
Air Blank	0.000	12:19
Control Test Stats		
Average	0.0497	
Std Dev	0.0006	
Rel Std Dev(%)	1.1625	

MANATEE COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000
06/16/2026
Software: 8100.27
SN 80-06559

Test	g/210L	Time
Air Blank	0.000	12:20
Control Test	0.079	12:21
Air Blank	0.000	12:21
Control Test	0.079	12:22
Air Blank	0.000	12:23
Control Test	0.078	12:23
Air Blank	0.000	12:24
Control Test Stats		
Average	0.0787	
Std Dev	0.0006	
Rel Std Dev(%)	0.7339	

MANATEE COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000
06/16/2026
Software: 8100.27
SN 80-06559

Test	g/210L	Time
Air Blank	0.000	12:27
Control Test	0.199	12:28
Air Blank	0.000	12:29
Control Test	0.199	12:29
Air Blank	0.000	12:30
Control Test	0.199	12:31
Air Blank	0.000	12:31
Control Test Stats		
Average	0.1950	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

MANATEE COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000
06/16/2026
Software: 8100.27
SN 80-06559

Test	g/210L	Time
Air Blank	0.000	12:12
Control Test	0.180	12:12
Air Blank	0.000	12:13
Control Test	0.080	12:13
Air Blank	0.000	12:14
Control Test	0.081	12:14
Air Blank	0.000	12:15
Control Test Stats		
Average	0.0803	
Std Dev	0.0006	
Rel Std Dev(%)	0.7187	

MANATEE COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000
06/16/2026
Software: 8100.27
SN 80-06559

[Signature]

Operator's Signature

[Signature]

Operator's Signature

[Signature]

Operator's Signature

[Signature]

Operator's Signature

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: MANATEE COUNTY SO
Time of Inspection: 13:23

Date of Inspection: 06/17/2026

Serial Number: 80-006559
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202603A Exp: 03/03/2028	0.08g/210L Test (g/210L) Lot#:202603B Exp: 03/03/2028	0.20g/210L Test (g/210L) Lot#:202603D Exp: 03/04/2028	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:23825080A1 Exp: 10/05/2027
0.000	0.049	0.078	0.198	0.080
0.000	0.050	0.078	0.198	0.080
0.000	0.050	0.078	0.198	0.080
0.000	0.050	0.078	0.198	0.080
0.000	0.050	0.079	0.198	0.080
0.000	0.050	0.078	0.198	0.080
0.000	0.050	0.079	0.198	0.080
0.000	0.050	0.079	0.198	0.081
0.000	0.050	0.079	0.198	0.080
0.000	0.050	0.078	0.198	0.080


Standard Deviations	0.0003	0.0005	0.0000	0.0003
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0002 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



KATIE T SPEARIN

Signature and Printed Name

06/17/2026
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-006559, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-006559</u>	UNCERTAINTY* $\pm$
Owning Agency:	<u>MANATEE COUNTY SO</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>06/17/2026</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>13:23</u>	0.200 g/ 210 L 0.008
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

FDLE/ATP Form 69 January 2026

Issuing Authority: Alcohol Testing Program

Date

06/17/2026

KATIE T SPEARIN,

Department Inspector

Service • Integrity • Respect • Quality